

# RF Exposure Evaluation Report

**Product** : Wrist-type Fully Automatic Digital  
Blood Pressure Monitor

**Trade mark** : N/A

**Model/Type reference** : SBC45

**Serial Number** : N/A

**Report Number** : EED32Q80075802

**FCC ID** : 2AQVU0047

**Date of Issue** : Feb. 06, 2024

**Test Standards** : 47 CFR Part 1.1307  
47 CFR Part 1.1310  
47 CFR Part 2.1091  
47 CFR Part 2.1093  
KDB 447498 D04 Interim General RF  
Exposure Guidance v01

**Test result** : PASS

Prepared for:

**JOYTECH HEALTHCARE CO., LTD.**

**No.365, Wuzhou Road, Yuhang Economic Development Zone,  
Hangzhou City, 311100 Zhejiang, China**

Prepared by:

**Centre Testing International Group Co., Ltd.**

**Hongwei Industrial Zone, Bao'an 70 District,  
Shenzhen, Guangdong, China**

**TEL: +86-755-3368 3668**

**FAX: +86-755-3368 3385**



Compiled by:

*mark.chen.*

Mark Chen

Approved by:

*Aaron Ma*

Aaron Ma

Reviewed by:

*Frazer. Li*

Frazer Li

Date:

Feb. 06, 2024



## 2 Version

Version No.	Date	Description
00	Feb. 06, 2024	Original

## 3 Contents

	Page
<b>1 COVER PAGE</b> .....	<b>1</b>
<b>2 VERSION</b> .....	<b>3</b>
<b>3 CONTENTS</b> .....	<b>4</b>
<b>4 GENERAL INFORMATION</b> .....	<b>5</b>
4.1 CLIENT INFORMATION .....	5
4.2 GENERAL DESCRIPTION OF EUT .....	5
4.3 PRODUCT SPECIFICATION SUBJECTIVE TO THIS STANDARD .....	5
4.4 TEST LOCATION .....	6
4.5 DEVIATION FROM STANDARDS .....	6
4.6 ABNORMALITIES FROM STANDARD CONDITIONS .....	6
4.7 OTHER INFORMATION REQUESTED BY THE CUSTOMER .....	6
<b>5 SAR EVALUATION</b> .....	<b>7</b>
5.1 RF EXPOSURE COMPLIANCE REQUIREMENT .....	7
5.1.1 <i>Limits</i> .....	7
5.1.2 <i>Test Procedure</i> .....	7
5.1.3 <i>EUT RF Exposure Evaluation</i> .....	8

## 4 General Information

### 4.1 Client Information

Applicant:	JOYTECH HEALTHCARE CO., LTD.
Address of Applicant:	No.365, Wuzhou Road, Yuhang Economic Development Zone, Hangzhou City, 311100 Zhejiang, China
Manufacturer:	JOYTECH HEALTHCARE CO., LTD.
Address of Manufacturer:	No.365, Wuzhou Road, Yuhang Economic Development Zone, Hangzhou City, 311100 Zhejiang, China
Factory:	JOYTECH HEALTHCARE CO., LTD.
Address of Factory:	No.365, Wuzhou Road, Yuhang Economic Development Zone, Hangzhou City, 311100 Zhejiang, China

### 4.2 General Description of EUT

Product Name:	Wrist-type Fully Automatic Digital Blood Pressure Monitor
Model No.(EUT):	SBC45
Trade Mark:	N/A

### 4.3 Product Specification subjective to this standard

Frequency Range:	2402MHz~2480MHz
Modulation Type:	GFSK
Test Power Grade:	Default
Test Software of EUT:	PhyPlusKit
Antenna Type:	PCB Antenna
Antenna Gain:	-1.37612dBi
Power Supply:	Battery DC 3V
Sample Received Date:	Jan. 17, 2024
Sample tested Date:	Jan. 17, 2024 to Jan. 29, 2024
Remark:	Company Name and Address shown on Report, the sample(s) and sample Information was/ were provided by the applicant who should be responsible for the authenticity which CTI hasn't verified.

## 4.4 Test Location

All tests were performed at:

Centre Testing International Group Co., Ltd

Building C, Hongwei Industrial Park Block 70, Bao'an District, Shenzhen, China

Telephone: +86 (0) 755 33683668 Fax:+86 (0) 755 33683385

No tests were sub-contracted.

FCC Designation No.: CN1164

## 4.5 Deviation from Standards

None.

## 4.6 Abnormalities from Standard Conditions

None.

## 4.7 Other Information Requested by the Customer

None.

## 5 SAR Evaluation

### 5.1 RF Exposure Compliance Requirement

#### 5.1.1 Limits

The SAR-based exemption formula of § 1.1307(b)(3)(i)(B), repeated here as Formula (B.2), applies for single fixed, mobile, and portable RF sources with available maximum time-averaged power or effective radiated power (ERP), whichever is greater, of less than or equal to the threshold  $P_{th}$  (mW).

This method shall only be used at separation distances from 0.5 cm to 40 cm and at frequencies from 0.3 GHz to 6 GHz (inclusive).  $P_{th}$  is given by Formula

$$P_{th} \text{ (mW)} = \begin{cases} ERP_{20 \text{ cm}} (d/20 \text{ cm})^x & d \leq 20 \text{ cm} \\ ERP_{20 \text{ cm}} & 20 \text{ cm} < d \leq 40 \text{ cm} \end{cases}$$

where

$$x = -\log_{10} \left( \frac{60}{ERP_{20 \text{ cm}} \sqrt{f}} \right)$$

and  $f$  is in GHz,  $d$  is the separation distance (cm), and  $ERP_{20 \text{ cm}}$  is per Formula (B.1).

$$P_{th} \text{ (mW)} = ERP_{20 \text{ cm}} \text{ (mW)} = \begin{cases} 2040f & 0.3 \text{ GHz} \leq f < 1.5 \text{ GHz} \\ 3060 & 1.5 \text{ GHz} \leq f \leq 6 \text{ GHz} \end{cases} \quad (\text{B.1})$$

The 1 mW Blanket Exemption of § 1.1307(b)(3)(i)(A) applies for single fixed, mobile, and portable RF sources with available maximum time-averaged power of no more than 1 mW, regardless of separation distance.

#### 5.1.2 Test Procedure

Software provided by client enabled the EUT to transmit and receive data at lowest, middle and highest channel individually.

## 5.1.3 EUT RF Exposure Evaluation

For Stand alone:

For Bluetooth LE:

Frequency (MHz)	Available maximum time-averaged power (dBm)	Antenna gain (dBi)	ERP (dBm)	Available maximum time-averaged power (mW)	Limit (mW)	Result
2402	-0.34	-1.37612	-3.86612	0.925	≤2.7880	PASS

**Note:**

- ① EIRP=Available maximum time-averaged power+Antenna gain;
- ② ERP=EIRP-2.15;
- ③ According to § 1.1307(b)(3)(i)(B), RF sources with available maximum time-averaged power or effective radiated power (ERP), whichever is greater, of less than or equal to the threshold Pth (mW).
- ④ The estimation distance is 0.5cm.
- ⑤ The test data please refer to the report of EED32Q80075801 and only the worst case data was recorded in the report.

The test report is effective only with both signature and specialized stamp, The result(s) shown in this report refer only to the sample(s) tested. Without written approval of CTI, this report can't be reproduced except in full.

\*\*\* End of Report \*\*\*