

Intertek 731 Enterprise Drive Lexington, KY 40510

Tel 859 226 1000 Fax 859 226 1040

www.intertek.com

eNeura, Inc. SAR TEST REPORT

SCOPE OF WORK

SPECIFIC ABSORBTION RATE - STMS MINI TRANSCRANIAL MAGNETIC STIMULATOR

REPORT NUMBER

103714598LEX-008

ISSUE DATE

1/31/2019

PAGES

35

DOCUMENT CONTROL NUMBER

Non-Specific EMC Report Shell Rev. December 2017 © 2017 INTERTEK





SPECIFIC ABSORBTION RATE TEST REPORT

Report Number: 103714598LEX-008

Project Number: G103714598

Report Issue Date: 1/31/2019

Product Name: sTMS Mini Transcranial Magnetic

Stimulator

Model Number: sTMS Mini

Standards: FCC Part 2.1093

RSS-102 Issue 5

Tested by:

Intertek Testing Services NA, Inc. 731 Enterprise Drive Lexington, KY 40510 Client:

eNeura, Inc. 715 N Pastoria Avenue

Sunnyvale, CA 94085

USA

Report prepared by

Report reviewed by

Bryan Taylor, Team Leader

James Sudduth, Senior Staff Engineer

James T. Saduet

This report is for the exclusive use of Intertek's Client and is provided pursuant to the agreement between Intertek and its Client. Intertek's responsibility and liability are limited to the terms and conditions of the agreement. Intertek assumes no liability to any party, other than to the Client in accordance with the agreement, for any loss, expense or damage occasioned by the use of this report. Only the Client is authorized to permit copying or distribution of this report and then only in its entirety. Any use of the Intertek name or one of its marks for the sale or advertisement of the tested material, product or service must first be approved in writing by Intertek. The observations and test results in this report are relevant only to the sample tested. This report by itself does not imply that the material, product, or service is or has ever been under an Intertek certification program.

Evaluation For: eNeura, Inc. Product: sTMS Mini Transcranial Magnetic Stimulator

Date: 1/31/2019

TABLE OF CONTENTS

1	INTRODUCTION	4
2	TEST SITE DESCRIPTION	5
3	JOB DESCRIPTION	10
4	SYSTEM VERIFICATION	11
5	EVALUATION PROCEDURES	16
6	CRITERIA	20
7	TEST CONFIGURATION	20
8	TEST RESULTS	21
9	SAR DATA:	21
1.0	REFERENCES	26
APP	ENDIX A – SYSTEM VALIDATION SUMMARY	27
APP	ENDIX B – WORST CASE SAR PLOTS	28
APP	ENDIX C – DIPOLE VALIDATION PLOTS	32

Evaluation For: eNeura, Inc.

Product: sTMS Mini Transcranial Magnetic Stimulator

Date: 1/31/2019

1 INTRODUCTION

SAR Test Report

At the request of eNeura, Inc. the sTMS Mini Transcranial Magnetic Stimulator was evaluated for SAR in accordance with the requirements for FCC Part 2.1093 and RSS-102 Issue 5. Testing was performed in accordance with IEEE Std 1528:2013, IEC62209-2:2010, and the Office of Engineering and Technology KDB 447498. Testing was performed at the Intertek facility in Lexington, Kentucky.

For the evaluation, the dosimetric assessment system DASY52 was used. The total uncertainty for the evaluation of the spatial peak SAR values averaged over a cube of 1g tissue mass had been assessed for this system to be $\pm 22.2\%$ from 300MHz – 3GHz and 24.6% from 3GHz – 6GHz.

The sTMS Mini Transcranial Magnetic Stimulator was tested at the maximum output power measured by Intertek. Maximum output power measurements are tabulated under Section 8 Test Results. The maximum spatial peak SAR value for the sample device averaged over 1g (for body worn mode) and 10g (for hand held mode) is shown below.

Based on the worst-case data presented below, the sTMS Mini Transcranial Magnetic Stimulator was found to be **compliant** with the 1.6 W/kg and 4W/kg requirements for general population / uncontrolled exposure.

Table 1: Worst Case Reported SAR per Exposure Condition

Device Position	Transmit Mode	Separation Distance	Channel / Frequency (MHz)	Maximum Conducted Output Power (dBm)	Reported SAR (1g) (W/kg)	Limit (W/kg)
Back Side	QPSK, 1RB	0mm	LTE Band 13 Ch. 23230	24.29dBm	0.3556W/kg	1.6W/kg
Back Side	QPSK, 1RB	0mm	LTE Band 4 Ch. 20175	24.31dBm	0.9049W/kg	1.6W/kg

Evaluation For: eNeura, Inc. Product: sTMS Mini Transcranial Magnetic Stimulator **SAR Test Report**

Date: 1/31/2019

TEST SITE DESCRIPTION 2

The SAR test site located at 731 Enterprise Drive, Lexington KY 40510 is comprised of the SPEAG model DASY 5.2 automated near-field scanning system, which is a package, optimized for dosimetric evaluation of mobile radios [3]. This system is installed in an ambient-free shielded chamber. The ambient temperature is controlled to 22.0 $\pm 2^{\circ}$ C. During the SAR evaluations, the RF ambient conditions are monitored continuously for signals that might interfere with the test results. The tissue simulating liquid is also stored in this area in order to keep it at the same constant ambient temperature as the room.




Figure 1: Intertek SAR Test Site

Evaluation For: eNeura, Inc. Product: sTMS Mini Transcranial Magnetic Stimulator SAR Test Report

Date: 1/31/2019

Measurement Equipment

The following major equipment/components were used for the SAR evaluation:

Table 2: Test Equipment Used for SAR Evaluation

Description	Serial Number	Manufacturer	Model	Cal. Date	Cal. Due
SAR Probe	3516	Speag	EXDV3	11/12/2018	11/12/2019
1800MHz Dipole	224	Speag	D1800V2	11/7/2018	11/7/2019
750MHz Dipole	1042	Speag	D750V3	11/6/2018	11/6/2019
DAE	358	Speag	DAE4	11/6/2018	11/6/2019
Vector Signal Generator	257708	Rohde & Schwarz	SMBV100A	9/21/2018	9/21/2019
Network Analyzer	US39173983	Agilent	8753ES	3/6/2018	3/6/2019
USB Power Sensor	100155	Rohde & Schwarz	NRP-Z81	9/21/2018	9/21/2019
USB Power Sensor	100705	Rohde & Schwarz	NRP-Z51	9/21/2018	9/21/2019
Dielectric Probe Kit	1111	Speag	DAK-3.5	NCR	NCR
Spectrum Analyzer	3099	Rohde & Schwarz	FSP7	9/20/2018	9/20/2019
SAM Twin Phantom	1663	Speag	QD 000 P40 C	NCR	NCR
6-axis robot	F11/5H1YA/A/01	Staubli	RX-90	NCR	NCR

^{*}NCR - No Calibration Required



Evaluation For: eNeura, Inc. Product: sTMS Mini Transcranial Magnetic Stimulator

SAR Test Report Date: 1/31/2019

2.2 Measurement Uncertainty

The Tables below includes the uncertainty budget suggested by the IEEE Std 1528-2013 and IEC62209-2: 2010 as determined by SPEAG for the DASY5 measurement System.

		Prob.				Std.Unc.	Std.Unc.	(v _i)
Error Description	Uncertainty Value	Dist.	Div.	c _i (1g)	c; (10g)	(1g)	(10g)	(Vi) V _{eff}
			Measureme	nt System				
Probe Calibration	±6.0%	N	1	1	1	±6.0%	±6.0%	∞
Axial Isotropy	±4.7%	R	√3	0.7	0.7	±1.9%	±1.9%	∞
Hemispherical Isotropy	±9.6%	R	√3	0.7	0.7	±3.9%	±3.9%	∞
Boundary Effect	±1.0%	R	√3	1	1	±0.6%	±0.6%	∞
Linearity	±4.7%	R	√3	1	1	±2.7%	±2.7%	∞
System Detection Limits	±1.0%	R	√3	1	1	±0.6%	±0.6%	∞
Modulation Response	±2.4%	R	√3	1	1	±1.4%	±1.4%	∞
Readout Electronics	±0.3%	N	1	1	1	±0.3%	±0.3%	∞
Response Time	±0.8%	R	√3	1	1	±0.5%	±0.5%	∞
Integration Time	±2.6%	R	√3	1	1	±1.5%	±1.5%	∞
RF Ambient Noise	±3.0%	R	√3	1	1	±1.7%	±1.7%	∞
RF Ambient Reflections	±3.0%	R	√3	1	1	±1.7%	±1.7%	∞
Probe Positioner	±0.4%	R	√3	1	1	±0.2%	±0.2%	∞
Probe Positioning	±2.9%	R	√3	1	1	±1.7%	±1.7%	∞
Max. SAR Eval.	±2.0%	R	√3	1	1	±1.2%	±1.2%	∞
			Test sample	Related			I	ll .
Device Positioning	±2.9%	N	1	1	1	±2.9%	±2.9%	145
Device Holder	±3.6%	N	1	1	1	±3.6%	±3.6%	5
Power Drift	±5.0%	R	√3	1	1	±2.9%	±2.9%	∞
Power Scaling	±0.0%	R	√3	1	1	±0%	±0%	∞
			Phantom a	nd Setup				
Phantom Uncertainty	±6.1%	R	√3	1	1	±3.5%	±3.5%	∞
SAR Correction	±1.9%	R	√3	1	0.84	±1.1%	±0.9%	∞
Liquid Conductivity (mea.)	±2.5%	R	√3	0.78	0.71	±1.1%	±1.0%	∞
Liquid Permittivity(mea.)	±2.5%	R	√3	0.26	0.26	±0.3%	±0.4%	∞
Temp unc Conductivity	±3.4%	R	√3	0.78	0.71	±1.5%	±1.4%	∞
Temp unc Permittivity	±0.4%	R	√3	0.78	0.71	±0.1%	±0.1%	∞
Combined Standard	20.470	15	,,,	0.23	0.20	±0.1/0	20.170	-
Uncertainty						±11.2%	±11.1%	361
Expanded STD Uncertainty						±22.3%	±22.2%	

Notes:

Worst Case uncertainty budget for DASY5 assessed according to IEEE 1528-2013. The budget is valid for the frequency range 300 MHz – 3 GHz and represents a worst-case analysis. For specific tests and configurations, the uncertainty could be considerably smaller.



Evaluation For: eNeura, Inc. Product: sTMS Mini Transcranial Magnetic Stimulator Date: 1/31/2019

		Prob.				Std.Unc.	Std.Unc.	(,,)
Error Description	Uncertainty Value	Dist.	Div.	c _i (1g)	c _i (10g)	(1g)	(10g)	(v _i) v _{eff}
			Measuremer	nt System				
Probe Calibration	±6.55%	N	1	1	1	±6.55%	±6.55%	∞
Axial Isotropy	±4.7%	R	√3	0.7	0.7	±1.9%	±1.9%	∞
Hemispherical Isotropy	±9.6%	R	√3	0.7	0.7	±3.9%	±3.9%	∞
Boundary Effect	±2.0%	R	√3	1	1	±1.2%	±1.2%	∞
Linearity	±4.7%	R	√3	1	1	±2.7%	±2.7%	∞
System Detection Limits	±1.0%	R	√3	1	1	±0.6%	±0.6%	∞
Modulation Response	±2.4%	R	√3	1	1	±1.4%	±1.4%	∞
Readout Electronics	±0.3%	N	1	1	1	±0.3%	±0.3%	∞
Response Time	±0.8%	R	√3	1	1	±0.5%	±0.5%	∞
Integration Time	±2.6%	R	√3	1	1	±1.5%	±1.5%	∞
RF Ambient Noise	±3.0%	R	√3	1	1	±1.7%	±1.7%	∞
RF Ambient Reflections	±3.0%	R	√3	1	1	±1.7%	±1.7%	∞
Probe Positioner	±0.8%	R	√3	1	1	±0.5%	±0.5%	∞
Probe Positioning	±6.7%	R	√3	1	1	±3.9%	±3.9%	∞
Max. SAR Eval.	±4.0%	R	√3	1	1	±2.3%	±2.3%	∞
			Test sample	Related				
Device Positioning	±2.9%	N	1	1	1	±2.9%	±2.9%	145
Device Holder	±3.6%	N	1	1	1	±3.6%	±3.6%	5
Power Drift	±5.0%	R	√3	1	1	±2.9%	±2.9%	∞
Power Scaling	±0.0%	R	√3	1	1	±0%	±0%	∞
			Phantom ar	nd Setup				
Phantom Uncertainty	±6.6%	R	√3	1	1	±3.8%	±3.8%	∞
SAR Correction	±1.9%	R	√3	1	0.84	±1.1%	±0.9%	∞
Liquid Conductivity (mea.)	±2.5%	R	√3	0.78	0.71	±1.1%	±1.0%	∞
			la .					
Liquid Permittivity(mea.)	±2.5%	R	√3	0.26	0.26	±0.3%	±0.4%	∞
Temp unc Conductivity	±3.4%	R	√3	0.78	0.71	±1.5%	±1.4%	∞
Temp unc Permittivity	±0.4%	R	√3	0.23	0.26	±0.1%	±0.1%	∞
Combined Standard Uncertainty						±12.3%	±12.2%	748
Expanded STD Uncertainty						±24.6%	±24.5%	
	1					1	1	

Notes.

Worst Case uncertainty budget for DASY5 assessed according to IEEE 1528-2013. The budget is valid for the frequency range 3 GHz – 6 GHz and represents a worst-case analysis. Probe calibration error reflects uncertainty of the EX3D probe. For specific tests and configurations, the uncertainty could be considerably smaller.



Evaluation For: eNeura, Inc.
Product: sTMS Mini Transcranial Magnetic Stimulator
Date: 1/31/2019

Prob. Std.Unc. Std.Unc. (v_i) **Error Description Uncertainty Value** Dist. Div. c_i (1g) c_i (10g) (1g) (10g) Veff Measurement System **Probe Calibration** ±6.55% ±6.55% ±6.55% Ν 1 1 1 ∞ ±4.7% √3 0.7 0.7 ±1.9% ±1.9% **Axial Isotropy** R Hemispherical Isotropy ±9.6% R √3 0.7 0.7 ±3.9% ±3.9% ∞ **Boundary Effect** √3 ±1.2% ±2.0% R 1 1 ±1.2% ∞ ±4.7% R √3 1 1 ±2.7% ±2.7% Linearity ∞ System Detection Limits ±1.0% R √3 1 1 ±0.6% ±0.6% Modulation Response ±2.4% R √3 1 1 ±1.4% ±1.4% ∞ ±0.3% ±0.3% Readout Electronics ±0.3% Ν 1 1 1 ∞ √3 1 ±0.5% Response Time ±0.8% R 1 ±0.5% Integration Time ±2.6% R √3 1 1 ±1.5% ±1.5% 8 RF Ambient Noise ±3.0% R √3 1 1 ±1.7% ±1.7% ∞ RF Ambient Reflections √3 1 1 ±1.7% ±1.7% ±3.0% R ∞ √3 **Probe Positioner** ±0.8% R 1 1 ±0.5% ±0.5% Probe Positioning ±6.7% R √3 1 1 ±3.9% ±3.9% 8 Post-Processing ±4.0% R √3 1 1 ±2.3% ±2.3% Test sample Related Device Positioning ±2.9% ±2.9% 145 ±2.9% Ν 1 1 1 Device Holder ±3.6% Ν 1 1 1 ±3.6% ±3.6% 5 √3 Power Drift ±5.0% R 1 1 ±2.9% ±2.9% ∞ Power Scaling ±0.0% R √3 1 ±0% ±0% Phantom and Setup Phantom Uncertainty R √3 ±4.6% ±4.6% ±7.9% 1 1 √3 SAR Correction ±1.9% R 1 0.84 ±1.1% ±0.9% Liquid Conductivity (mea.) R √3 0.78 0.71 ±1.1% ±1.0% 8 ±2.5% Liquid Permittivity(mea.) ±2.5% R √3 0.26 0.26 ±0.3% ±0.4% √3 Temp unc. - Conductivity ±3.4% R 0.78 0.71 ±1.5% ±1.4% 000 √3 0.23 0.26 ±0.1% Temp unc. - Permittivity ±0.4% R ±0.1% 000 Combined Standard

Notes.

Uncertainty

Expanded STD Uncertainty

Worst Case uncertainty budget for DASY5 assessed according to IEC62209-2: 2010. The budget is valid for the frequency range 30MHz – 6 GHz and represents a worst-case analysis. Probe calibration error reflects uncertainty of the EX3D probe. For specific tests and configurations, the uncertainty could be considerably smaller.

+12.5%

±25.1%

748

+12.5%

±25.0%

Evaluation For: eNeura, Inc. Product: sTMS Mini Transcranial Magnetic Stimulator

Date: 1/31/2019

3 JOB DESCRIPTION

At the request eNeura, Inc., SAR testing was performed on the sTMS Mini Transcranial Magnetic Stimulator (Model: sTMS Mini.)

The sTMS Mini Transcranial Magnetic Stimulator is a therapeutic device used in the treatment of migraine headaches. It applies magnetic stimulation therapy and then transmits data back to the supervising clinician via an embedded cellular module (Sierra Wireless Model HL7618RD). The data transmission occurs after the therapy has completed and the device could then be lying in the users lap or in a purse or bag. The data transmission does not occur whilst the device is being held to the head.

Table 3: Product Information

Test Sample Information
eNeura, Inc.
sTMS Mini Transcranial Magnetic Stimulator
sTMS Mini
20001152
12/5/2018
Good
12/6/2018 to 12/14/2018
Portable
General Population/Uncontrolled Environment
Internal
None

Table 4: Operating Bands

Operating Bands	Technology	Modulation	Frequency Range (MHz)	Max Upper Tolerance Output Power (dBm)	Duty Cycle
LTE Band 4	LTE	QPSK, 16QAM	1710 to 1755MHz	25dBm	1:1
LTE Band 13	LTE	QPSK, 16QAM	777 to 787MHz	25dBm	1:1

Evaluation For: eNeura, Inc. Product: sTMS Mini Transcranial Magnetic Stimulator SAR Test Report

Date: 1/31/2019

SYSTEM VERIFICATION

System Validation

Prior to the assessment, the system was verified to be within ±10% of the specifications by using the system validation kit. The system validation procedure tests the system against reference SAR values and the performance of probe, readout electronics and software. The test setup utilizes a phantom and reference dipole.




Figure 2: System Verification Setup



Evaluation For: eNeura, Inc.
Product: sTMS Mini Transcranial Magnetic Stimulator

Date: 1/31/2019

Table 5: Dipole Validations (1g)

			Refe	erence Dipol	e Validati	on			
						Cal.		%	
Ambient	Fluid				Dipole	Lab		Error	
Temp	Temp	Frequency		Fluid	Power	SAR	Measured	SAR	
(°C)	(ºC)	(MHz)	Dipole	Type	Input	(1g)	SAR (1g)	(1g)	Date
23.2	23.1	750	D750V3	MSL750	1W	8.64	8.47	1.97	12/6/2018
23.2	23.1	1800	D1800V2	MSL1800	1W	38.9	41.9	7.71	12/12/2018

Table 6: Dipole Validations (10g)

	Reference Dipole Validation												
						Cal.		%					
Ambient	Fluid				Dipole	Lab		Error					
Temp	Temp	Frequency		Fluid	Power	SAR	Measured	SAR					
(ºC)	(ºC)	(MHz)	Dipole	Туре	Input	(10g)	SAR (10g)	(10g)	Date				
23.2	23.1	750	D750V3	MSL750	1W	5.72	5.69	0.52	12/6/2018				
23.2	23.1	1800	D1800V2	MSL1800	1W	20.3	22.1	8.87	12/12/2018				



Date: 1/31/2019

Measurement Uncertainty for System Validation

Source of Uncertainty	Value(dB)	Probability Distribution	Divisor	Ci	u _i (y)	(u _i (y))^2
Measurement System						
Probe Calibration	5.50	n1	1	1	5.50	30.250
Axial Isotropy	4.70	r	1.732	0.7	2.71	7.364
Hemispherical Isotropy	9.60	r	1.732	0.7	5.54	30.722
Boundary Effect	1.00	r	1.732	1	0.58	0.333
Linearity	4.70	r	1.732	1	2.71	7.364
System Detection Limits	1.00	r	1.732	1	0.58	0.333
Readout Electronics	0.30	n1	1	1	0.30	0.090
Response Time	0.80	r	1.732	1	0.46	0.213
Integration Time	2.60	r	1.732	1	1.50	2.253
RF Ambient Noise	3.00	r	1.732	1	1.73	3.000
RF Ambient Reflections	3.00	r	1.732	1	1.73	3.000
Probe Positioner	0.40	r	1.732	1	0.23	0.053
Probe Positioning	2.90	r	1.732	1	1.67	2.803
Max. SAR Eval.	1.00	r	1.732	1	0.58	0.333
Dipole / Generator / Power Meter Related						
Dipole positioning	2.90	n1	1	1	2.90	8.410
Dipole Calibration Uncertainty	0.68	r	1.732	1	0.39	0.154
Power Meter 1 Uncertainty (+20C to +25C)	0.13	n1	1	2	0.13	0.017
Power Meter 2 Uncertainty (+20C to +25C)	0.04	n1	1	3	0.04	0.002
Sig Gen VSWR Mismatch Error	1.80	n1	1	5	1.80	3.240
Sig Gen Resolution Error	0.01	n1	1	6	0.01	0.000
Sig Gen Level Error	0.90	n1	1	1	0.90	0.810
Phantom and Setup						
Phantom Uncertainty	4.00	r	1.732	1	2.31	5.334
Liquid Conductivity (target)	5.00	r	1.732	0.43	2.89	8.334
Liquid Conductivity (meas.)	2.50	n1	1	0.43	2.50	6.250
Liquid Permittivity (target)	5.00	r	1.732	0.49	2.89	8.334
Liquid Permittivity (meas.)	2.50	n1	1	0.49	2.50	6.250
Combined Standard Uncertainty		N1	1	1	11.63	135.247
Expanded Uncertainty		Normal k=	2		23.26	
Expanded Uncertainty	is	23.3	for	Normal	k=	2

Evaluation For: eNeura, Inc.

Product: sTMS Mini Transcranial Magnetic Stimulator
Date: 1/31/2019

Tissue Simulating Liquid Description and Validation

The dielectric parameters were verified to be within 5% of the target values prior to assessment. The dielectric parameters (ε_r , σ) are shown in Table 7. A recipe for the tissue simulating fluid used is shown in Table 8.

Table 7: Dielectric Parameter Validations

	Measured Tissue Properties											
	Frequen	Dielectri										
	су	С		Dielectric			Dielectric					
Tissue	Measure	Constant	Conductivity	Constant	Imaginary	Conductivity	%	Conductivity				
_												
Type	(MHz)	Target	Target	Measure	Part	Measure	Deviation	% Deviation	Date			
Туре	(MHz) 750	Target 56	Target 0.95	Measure 53.9	Part 21.96	Measure 0.92	Deviation 3.75	% Deviation 3.61	Date 12/6/2018			
Туре												

	Measured Tissue Properties											
	Frequen	Dielectri										
	су	С		Dielectric			Dielectric					
Tissue	Measure	Constant	Conductivity	Constant	Imaginary	Conductivity	%	Conductivity				
Type	(MHz)	Target	Target	Measure	Part	Measure	Deviation	% Deviation	Date			
	1710	53.53	1.46	53.12	15.7	1.49	0.77	2.23	12/12/2018			
1750	1710 1732	53.53 53.5	1.46 1.48	53.12 53.08	15.7 15.6	1.49 1.5022	0.77 0.79	2.23 1.50	12/12/2018 12/12/2018			

Evaluation For: eNeura, Inc. Product: sTMS Mini Transcranial Magnetic Stimulator

Date: 1/31/2019

Table 8: Tissue Simulating Fluid Recipe

Comp	Composition of Ingredients for Liquid Tissue Phantoms (450MHz to 2450 MHz data only)											
Ingredient	f (MHz)											
(% by weight)	450		835		915		1900		2450		5500	
Tissue Type	Head	Body	Head	Body	Head	Body	Head	Body	Head	Body	Head	Body
Water	38.56	51.16	41.45	52.4	41.05	56	54.9	70.45	62.7	68.64	65.53	78.67
Salt (NaCl)	3.95	1.49	1.45	1.4	1.35	0.76	0.18	0.36	0.5	0	0	0
Sugar	56.32	46.78	56	45	56.5	41.76	0	0	0	0	0	0
HEC	0.98	0.52	1	1	1	1.21	0	0	0	0	0	0
Bactericide	0.19	0.05	0.1	0.1	0.1	0.27	0	0	0	0	0	0
Triton X-100	0	0	0	0	0	0	0	0	36.8	0	17.235	10.665
DGBE	0	0	0	0	0	0	44.92	29.18	0	31.37	0	0
DGHE	0	0	0	0	0	0	0	0	0	0	17.235	10.665
Dielectric Constant	43.42	58	42.54	56.1	42	56.8	39.9	53.3	39.8	52.7		
Conductivity (S/m)	0.85	0.83	0.91	0.95	1	1.07	1.42	1.52	1.88	1.95		

Tissue Simulating Liquid for 5GHz, MBBL3500-5800V5 Manufactured by SPEAG (proprietary mixture)

Ingredients	(% by weight)
Water	78
Mineral oil	11
Emulsifiers	9
Additives and Salt	2



Evaluation For: eNeura, Inc. Product: sTMS Mini Transcranial Magnetic Stimulator

Date: 1/31/2019

5 EVALUATION PROCEDURES

Prior to any testing, the appropriate fluid was used to fill the phantom to a depth of 15 cm \pm 0.2cm. The fluid parameters were verified and the dipole validation was performed as described in the previous sections.

Test Positions:

The Device was positioned against the SAM and flat phantom using the exact procedure described in IEEE Std 1528:2013, IEC62209-2:2010, and the Office of Engineering and Technology KDB 447498.

Reference Power Measurement:

The measurement probe was positioned at a fixed location above the reference point. A power measurement was made with the probe above this reference position so it could used for the assessing the power drift later in the test procedure.

Area Scan:

A coarse area scan was performed in order to find the approximate location of the peak SAR value. This scan was performed with the measurement probe at a constant height in the simulating fluid. A two dimensional spline interpolation algorithm was then used to determine the peaks and gradients within the scanned area. The area scan resolution conformed to the requirements of KDB 865664 as shown in Table 9.

Zoom Scan:

A zoom scan was performed around the approximate location of the peak SAR as determined from the area scan. On the basis of this data set, the spatial peak SAR value was evaluated with the following procedure. The zoom scan resolution conformed to the requirements of KDB 865664 as shown in Table 9.

Evaluation For: eNeura, Inc.
Product: sTMS Mini Transcranial Magnetic Stimulator

Date: 1/31/2019

Table 9: SAR Area and Zoom Scan Resolutions

			≤ 3 GHz	> 3 GHz		
Maximum distance fro (geometric center of pr			5 ± 1 mm	$\frac{1}{2} \cdot \delta \cdot \ln(2) \pm 0.5 \text{ mm}$		
Maximum probe angle surface normal at the n			30° ± 1°	20° ± 1°		
			≤ 2 GHz: ≤ 15 mm 2 – 3 GHz: ≤ 12 mm	3 – 4 GHz: ≤ 12 mm 4 – 6 GHz: ≤ 10 mm		
Maximum area scan sp	atial resol	ution: Δx_{Area} , Δy_{Area}	When the x or y dimension of the test device, in the measurement plane orientation, is smaller than the above the measurement resolution must be ≤ the corresponding x or y dimension of the test device with at least one measurement point on the test device.			
Maximum zoom scan s	spatial reso	lution: Δx _{Zoom} , Δy _{Zoom}	≤ 2 GHz: ≤ 8 mm 2 – 3 GHz: ≤ 5 mm*	3 – 4 GHz: ≤ 5 mm* 4 – 6 GHz: ≤ 4 mm*		
	uniform	grid: Δz _{Zoom} (n)	≤ 5 mm	3 – 4 GHz: ≤ 4 mm 4 – 5 GHz: ≤ 3 mm 5 – 6 GHz: ≤ 2 mm		
Maximum zoom scan spatial resolution, normal to phantom surface	graded	Δz _{Zoom} (1): between 1 st two points closest to phantom surface	≤ 4 mm	3 – 4 GHz: ≤ 3 mm 4 – 5 GHz: ≤ 2.5 mm 5 – 6 GHz: ≤ 2 mm		
	grid	axis to phantom location $30^{\circ} \pm 1^{\circ}$ 20° $\leq 2 \text{ GHz:} \leq 15 \text{ mm}$ $3-4 \text{ GHz}$ $2-3 \text{ GHz:} \leq 12 \text{ mm}$ $4-6 \text{ GHz}$ $2-3 \text{ GHz:} \leq 12 \text{ mm}$ When the x or y dimension of the test device measurement plane orientation, is smaller that the measurement resolution must be \leq the convergence of the test device with at less than the measurement point on the test device. The measurement point on the test device with at less than the measurement point on the test device. The measurement point on the test device with at less than the measurement point on the test device. The measurement point on the test device with at less than the measurement point on the test device. The measurement point on the test device with at less than the measurement point on the test device. The measurement point on the test device with at less than the measurement point on the test device. The measurement point on the test device with at less than the measurement point on the test device with at less than the measurement point on the test device with at less than the measurement point on the test device with at less than the measurement point on the test device. The measurement point on the test device with at less than the measurement point on the test device. The measurement point on the test device with at less than the measurement point on the test device. The measurement point on the test device with at less than the measurement point on the test device. The measurement point on the test device with at less than the measurement point on the test device with at less than the measurement point on the test device with at less than the measurement point on the test device with at less than the measurement point on the test device with at less than the measurement point on the test device with at less than the measurement point on the test device with at less than the measurement point on the test device with at less than the measurement point on the test device. The measurement point on the test device with at less	z _{Zoom} (n-1)			
Minimum zoom scan volume	x, y, z		≥ 30 mm	3 – 4 GHz: ≥ 28 mm 4 – 5 GHz: ≥ 25 mm 5 – 6 GHz: ≥ 22 mm		

Note: δ is the penetration depth of a plane-wave at normal incidence to the tissue medium; see draft standard IEEE P1528-2011 for details.

When zoom scan is required and the <u>reported</u> SAR from the <u>area scan based 1-g SAR estimation</u> procedures of KDB 447498 is ≤ 1.4 W/kg, ≤ 8 mm, ≤ 7 mm and ≤ 5 mm zoom scan resolution may be applied, respectively, for 2 GHz to 3 GHz, 3 GHz to 4 GHz and 4 GHz to 6 GHz.



Evaluation For: eNeura, Inc.

Product: sTMS Mini Transcranial Magnetic Stimulator
Date: 1/31/2019

Interpolation, Extrapolation and Detection of Maxima:

The probe is calibrated at the center of the dipole sensors which is located 1 to 2.7 mm away from the probe tip. During measurements, the probe stops shortly above the phantom surface, depending on the probe and the surface detecting system. Both distances are included as parameters in the probe configuration file. The software always knows exactly how far away the measured point is from the surface. As the probe cannot directly measure at the surface, the values between the deepest measured point and the surface must be extrapolated.

In DASY5, the choice of the coordinate system defining the location of the measurement points has no influence on the uncertainty of the interpolation, Maxima Search and extrapolation routines. The interpolation, extrapolation and maximum search routines are all based on the modified Quadratic Shepard's method.

Thereby, the interpolation scheme combines a least-square fitted function method and a weighted average method which are the two basic types of computational interpolation and approximation. The DASY5 routines construct a once-continuously differentiable function that interpolates the measurement values as follows:

- For each measurement point a trivariate (3-D) / bivariate (2-D) quadratic is computed. It interpolates the measurement values at the data point and forms a least-square fit to neighboring measurement values.
- The spatial location of the quadratic with respect to the measurement values is attenuated by an inverse distance weighting. This is performed since the calculated quadratic will fit measurement values at nearby points more accurate than at points located further away.
- After the quadratics are calculated for at all measurement points, the interpolating function is calculated as a weighted average of the quadratics.

There are two control parameters that govern the behavior of the interpolation method. One specifies the number of measurement points to be used in computing the least-square fits for the local quadratics. These measurement points are the ones nearest the input point for which the quadratic is being computed. The second parameter specifies the number of measurement points that will be used in calculating the weights for the quadratics to produce the final function. The input data points used there are the ones nearest the point at which the interpolation is desired. Appropriate defaults are chosen for each of the control parameters.

The trivariate quadratics that have been previously computed for the 3-D interpolation and whose input data are at the closest distance from the phantom surface, are used in order to extrapolate the fields to the surface of the phantom.

In order to determine all the field maxima in 2-D (Area Scan) and 3-D (Zoom Scan), the measurement grid is refined by a default factor of 10 and the interpolation function is used to evaluate all field values between corresponding measurement points. Subsequently, a linear search is applied to find all the candidate maxima. In a last step, non-physical maxima are removed and only those maxima which are within 2 dB of the global maximum value are retained.

Report Number: 103714598LEX-008



Evaluation For: eNeura, Inc. Product: sTMS Mini Transcranial Magnetic Stimulator

Date: 1/31/2019

Averaging and Determination of Spatial Peak SAR

The interpolated data is used to average the SAR over the 1g and 10g cubes by spatially discretizing the entire measured volume. The resolution of this spatial grid used to calculate the averaged SAR is 1mm or about 42875 interpolated points. The resulting volumes are defined as cubical volumes containing the appropriate tissue parameters that are centered at the location. The location is defined as the center of the incremental volume.

The spatial-peak SAR must be evaluated in cubical volumes containing a mass that is within 5% of the required mass. The cubical volume centered at each location, as defined above, should be expanded in all directions until the desired value for the mass is reached, with no surface boundaries of the averaging volume extending beyond the outermost surface of the considered region. In addition, the cubical volume should not consist of more than 10% of air. If these conditions are not satisfied then the center of the averaging volume is moved to the next location. Otherwise, the exact size of the final sampling cube is found using an inverse polynomial approximation algorithm, leading to results with improved accuracy. If one boundary of the averaging volume reaches the boundary of the measured volume during its expansion, it will not be evaluated at all. Reference is kept of all locations used and those not used for averaging the SAR. All average SAR values are finally assigned to the centered location in each valid averaging volume.

All locations included in an averaging volume are marked to indicate that they have been used at least once. If a location has been marked as used, but has never been assigned to the center of a cube, the highest averaged SAR value of all other cubical volumes which have used this location for averaging is assigned to this location. Only those locations that are not part of any valid averaging volume should be marked as unused. For the case of an unused location, a new averaging volume must be constructed which will have the unused location centered at one surface of the cube. The remaining five surfaces are expanded evenly in all directions until the required mass is enclosed, regardless of the amount of included air. Of the six possible cubes with one surface centered on the unused location, the smallest cube is used, which still contains the required mass.

If the final cube containing the highest averaged SAR touches the surface of the measured volume, an appropriate warning is issued within the post processing engine.

Power Drift Measurement:

The probe was positioned at precisely the same reference point and the reference power measurement was repeated. The difference between the initial reference power and the final one is referred to as the power drift. This value should not exceed 5%. The power drift measurement was used to assess the output power stability of the test sample throughout the SAR scan.

RF Ambient Activity:

During the entire SAR evaluation, the RF ambient activity was monitored using a spectrum analyzer with an antenna connected to it. The spectrum analyzer was tuned to the frequency of measurement and with one trace set to max hold mode. In this way, it was possible to determine if at any point during the SAR measurement there was an interfering ambient signal. If an ambient signal was detected, then the SAR measurement was repeated.

Report Number: 103714598LEX-008



Evaluation For: eNeura, Inc. Product: sTMS Mini Transcranial Magnetic Stimulator

Date: 1/31/2019

6 CRITERIA

SAR Test Report

The following ANSI/IEEE C95.1 – 1992 limits for SAR apply to portable devices operating in the General Population/Uncontrolled Exposure environment. Uncontrolled environments are defined as locations where there is the exposure of individuals who have no knowledge or control of their exposure.

Exposure Type	SAR Limit
(General Population/Uncontrolled Exposure environment)	(W/kg or mW/g)
Average over the whole body	0.08
Spatial Peak (1g)	1.60
Spatial Peak for hands, wrists, feet and ankles (10g)	4.00

7 TEST CONFIGURATION

The sTMS Mini Transcranial Magnetic Stimulator could be used against the body or hand held. According to the manufacturer there are no against the head usage conditions. Therefore it was evaluated in body and extremity (hand held) positions. Due to the curved nature of the device the testing on the inside was performed against a curved section of the phantom that matched the curvature of the device.

The device was evaluated according to the specific requirements found in the following KDBs and Standards:

- FCC KDB 447498D01 v06, General RF Exposure Guidance
- FCC KDB 865664D01 v01r04, SAR Measurement Requirements for 100MHz to 6GHz
- FCC KDB 941225 D05, SAR for LTE Devices v02r05
- RSS-102 Issue 5, Radio Frequency (RF) Exposure Compliance of Radiocommunication Apparatus (All Frequency Bands)



Product: sTMS Mini Transcranial Magnetic Stimulator **SAR Test Report**

Date: 1/31/2019

Evaluation For: eNeura, Inc.

TEST RESULTS 8

The worst case 1g SAR value for body exposure was less than the 1.6W/kg limit. The worst case 10g SAR value for extremity exposure was less than the 4W/kg limit.

9 **SAR DATA:**

The results on the following page(s) were obtained when the device was transmitting at maximum output power. The worst case plots, which reveal information about the location of the maximum SAR with respect to the device, are referenced are shown in APPENDIX B - Worst Case SAR Plots. The measured conducted output power was compared to the power declared by the manufacturer and used for scaling the measured SAR values.

Evaluation For: eNeura, Inc. Product: sTMS Mini Transcranial Magnetic Stimulator **SAR Test Report** Date: 1/31/2019

Table 10: Body Worn SAR Results, (LTE Band 4)

				US / Canad	a Body S	AR Resul	ts Using 175	OMHz MSL		
Date	TX Mode	Spacing	RB	Position	RB Offset	Power Drift (dB)	Measured SAR 1g (W/kg)	Reported SAR 1g (W/kg)	Measured Conducted Output Power (dBm)	Maximum Conducted Output Power (dBm)
					low	-0.03	0.7720	0.9049	24.31	25.00
	LTE,			Back	mid	-0.01	0.7560	0.8821	24.33	25.00
12/12/2018	Band 4, QPSK,	0mm	1		high	-0.19	0.6610	0.7802	24.28	25.00
12/12/2010	Channel	Omm			low	-0.04	0.1800	0.2110	24.31	25.00
	20175			Front	mid	-0.09	0.1670	0.1949	24.33	25.00
					high	0.06	0.1530	0.1806	24.28	25.00
	LTE,			Back	low	-0.02	0.7150	0.8439	24.28	25.00
					mid	0.02	0.6830	0.8117	24.25	25.00
12/14/2018	Band 4, QPSK,	0mm	50%		high	0.50	0.6710	0.7975	24.25	25.00
12/14/2010	Channel	Omm	3070		low	0.05	0.1430	0.1688	24.28	25.00
	20175			Front	mid	0.03	0.1386	0.1647	24.25	25.00
					high	-0.11	0.1410	0.1676	24.25	25.00
					low	-0.04	0.6714	0.7961	24.26	25.00
	LTE,			Back	mid	-0.03	0.6722	0.8044	24.22	25.00
12/14/2018	Band 4,	0mm	100%		high	0.04	0.6534	0.7766	24.25	25.00
12/14/2018 QPSK, Channel 20175	Ollilli	10078		low				24.26	25.00	
			Front	mid	Reduced			24.22	25.00	
					high				24.25	25.00
				1g SAR Lin	nit (Head	& Body) =	= 1.6W/kg			ļ

Test Personnel: **Bryan Taylor** Test Date: 12/12/2018 - 12/14/2018 Supervising/Reviewing Engineer: (Where Applicable) Tissue Depth: 15cm Base Station Simulator Ambient Temperature: 22.8C Signal Setup: Power Method: **Fully Charged Battery** Relative Humidity: 41.6% Pretest Verification w / Ambient Signals or BB Source: Yes Atmospheric Pressure: 989.4mbar

Deviations, Additions, or Exclusions:

- (1) Test reduction was applied. Per KDB941225D05 v02r05 Section 5.2.3, testing 1ith 100% RB allocation is not required when the highest maximum output power is less than that which was measured in 50% and 1RB allocations and
- (2) Test reduction was applied. Per KDB941225D05 v02r05 Section 5.2.4 SAR for higher order modulations is only required when the maximum output power for the configuration with the higher order modulation is >1/2dB higher than the same configuration in QPSK or when the reported SAR for QPSK is >1.45W/kg.

Evaluation For: eNeura, Inc. Product: sTMS Mini Transcranial Magnetic Stimulator **SAR Test Report**

Date: 1/31/2019

Table 11: Body Worn SAR Results, (LTE Band 13)

				US / Canad	la Body S	AR Resu	Its Using 750	MHz MSL		
Date	TX Mode	Spacing	RB	Position	RB Offset	Power Drift (dB)	Measured SAR 1g (W/kg)	Reported SAR 1g (W/kg)	Measured Conducted Output Power (dBm)	Maximum Conducted Output Power (dBm)
					low	-0.13	0.3020	0.3556	24.29	25.00
	LTE Band			Back	mid	-0.12	0.2570	0.3047	24.26	25.00
12/6/2018	13,	0mm	1		high	-0.01	0.2290	0.2722	24.25	25.00
12/0/2010	QPSK, Channel	Omm	'		low	-0.09	0.0917	0.1080	24.29	25.00
	23230			Front	mid	-0.05	0.1040	0.1233	24.26	25.00
					high	0.08	0.0881	0.1047	24.25	25.00
				Back	low	0.02	0.2570	0.3069	24.23	25.00
	LTE Band				mid	-0.04	0.2490	0.2987	24.21	25.00
12/9/2018	13,	0mm	50%		high	-0.06	0.2420	0.2896	24.22	25.00
12/3/2010	QPSK, Channel	Offiliti	30 /6		low	0.05	0.0874	0.1044	24.23	25.00
	23230			Front	mid	-0.02	0.0813	0.0975	24.21	25.00
					high	0.08	0.0872	0.1044	24.22	25.00
					low				24.18	25.00
	LTE Band			Back	mid				24.20	25.00
12/9/2018	13,	0mm	100%		high		Reduced		24.17	25.00
12/3/2010	QPSK,	Offiliti	100 /6		low		Neudoed		24.18	25.00
	Channel 23230			Front	mid				24.20	25.00
					high				24.17	25.00
				lg SAR Lim	it (Head a	and Body) = 1.6W/kg			

Test Personnel:	Bryan Taylor	Test Date:	12/6/2018 – 12/9/2018
Supervising/Reviewing Engineer:			
(Where Applicable)	NA	Tissue Depth:	15cm
Signal Setup:	Base Station Simulator	Ambient Temperature:	23.1C
Power Method:	Fully Charged Battery	Relative Humidity:	38.7%
Pretest Verification w / Ambient Signals or BB Source:	Yes	Atmospheric Pressure:	988.7mbar

Deviations, Additions, or Exclusions:

- (1) Test reduction was applied. Per KDB941225D05 v02r05 Section 5.2.3, testing 1ith 100% RB allocation is not required when the highest maximum output power is less than that which was measured in 50% and 1RB allocations and
- (2) Test reduction was applied. Per KDB941225D05 v02r05 Section 5.2.4 SAR for higher order modulations is only required when the maximum output power for the configuration with the higher order modulation is >1/2dB higher than the same configuration in QPSK or when the reported SAR for QPSK is >1.45W/kg.



Evaluation For: eNeura, Inc. Product: sTMS Mini Transcranial Magnetic Stimulator

Date: 1/31/2019

Table 12: Extremity SAR Results, (LTE Band 4)

			U	S / Canada	Extremity	/ SAR Res	sults Using 1	750MHz MSI	L	
Date	TX Mode	Spacing	RB	Position	RB Offset	Power Drift (dB)	Measured SAR 10g (W/kg)	Reported SAR 10g (W/kg)	Measured Conducted Output Power (dBm)	Maximum Conducted Output Power (dBm)
					low	0.07	0.8590	1.0069	24.31	25.00
	LTE,			Back	mid	-0.24	0.8560	0.9988	24.33	25.00
12/12/2018 Band 4, QPSK, Channel 20175	0mm	1		high	0.06	0.8290	0.9785	24.28	25.00	
			Front	low	-0.04	0.1100	0.1289	24.31	25.00	
				mid	-0.09	0.1010	0.1178	24.33	25.00	
					high	0.06	0.0939	0.1108	24.28	25.00
					low	-0.04	0.7500	0.8852	24.28	25.00
	LTE,			Back	mid	-0.02	0.7310	0.8648	24.27	25.00
12/14/2018	Band 4, QPSK,	0mm	50%		high	-0.05	0.7140	0.8486	24.25	25.00
12/14/2010	Channel	Omm	JU /6		low	0.05	0.0910	0.1074	24.28	25.00
20175	20175			Front	mid	-0.14	0.0891	0.1054	24.27	25.00
				high	-0.11	0.0900	0.1070	24.25	25.00	
				10g SAR	Limit (Ex	ktremity) :	= 4W/kg			

Test Personnel:	Bryan Taylor	Test Date:	12/12/2018 – 12/14/2018
Supervising/Reviewing Engineer:		-	
(Where Applicable)	NA	Tissue Depth:	15cm
Signal Setup:	Base Station Simulator	Ambient Temperature:	22.8C
Power Method:	Fully Charged Battery	Relative Humidity:	41.6%
Pretest Verification w / Ambient Signals or BB Source:	Yes	Atmospheric Pressure:	989.4mbar

Deviations, Additions, or Exclusions:

(1) Test reduction was applied. Per KDB447498 D01v06 when the mid channel is more than 3dB below the limit testing at the remaining channels is not required.



Evaluation For: eNeura, Inc.
Product: sTMS Mini Transcranial Magnetic Stimulator

Date: 1/31/2019

Table 13: Extremity SAR Results, (LTE Band 13)

			U	S / Canada	Extremit	y SAR Re	sults Using 7	750MHz MSL		
Date	TX Mode	Spacing	RB	Position	RB Offset	Power Drift (dB)	Measured SAR 10g (W/kg)	Reported SAR 10g (W/kg)	Measured Conducted Output Power (dBm)	Maximum Conducted Output Power (dBm)
					low	-0.13	0.2030	0.2391	24.29	25.00
	LTE Band			Back	mid	-0.12	0.1710	0.2028	24.26	25.00
12/6/2018 13, QPSK,	0mm	1		high	-0.01	0.1530	0.1818	24.25	25.00	
				low	-0.09	0.0580	0.0683	24.29	25.00	
	Channel 23230			Front	mid	-0.05	0.0670	0.0794	24.26	25.00
					high	0.08	0.0558	0.0663	24.25	25.00
					low	0.02	0.1650	0.1970	24.23	25.00
	LTE Band			Back	mid	-0.02	0.1590	0.1907	24.21	25.00
12/9/2018	13,	0mm	50%		high	-0.06	0.1540	0.1843	24.22	25.00
12/3/2010	QPSK,	OHIIII	30 /6		low	0.05	0.0555	0.0663	24.23	25.00
	Channel 23230			Front	mid	-0.03	0.0534	0.0641	24.21	25.00
20200				high	0.08	0.0555	0.0664	24.22	25.00	
				10g SAF	R Limit (E	xtremity)	= 4W/kg			

Test Personnel:	Bryan Taylor	Test Date:	12/6/2018 – 12/9/2018
Supervising/Reviewing Engineer:		-	
(Where Applicable)	NA	Tissue Depth:	15cm
Signal Setup:	Base Station Simulator	Ambient Temperature:	23.1C
Power Method:	Fully Charged Battery	Relative Humidity:	38.7%
Pretest Verification w / Ambient Signals or BB Source:	Yes	Atmospheric Pressure:	988.7mbar

Deviations, Additions, or Exclusions:

(1) Test reduction was applied. Per KDB447498 D01v06 when the mid channel is more than 3dB below the limit testing at the remaining channels is not required.

Evaluation For: eNeura, Inc.

Product: sTMS Mini Transcranial Magnetic Stimulator
Date: 1/31/2019

1.0 REFERENCES

[1]ANSI, ANSI/IEEE C95.1-1992: IEEE Standard for Safety Levels with Respect to Human Exposure to Radio Frequency Electromagnetic Fields, 3kHz to 300 GHz, The Institute of electrical and Electronics Engineers, Inc., New York, NY 10017, 1992

[2]Federal Communications Commission, "Evaluating Compliance with FCC Guidelines for Human Exposure to Radiofrequency Electromagnetic Fields", Supplement C (Edition 01-01) to OET Bulletin 65 (Edition 97-01), FCC, Washington, D.C. 20554, 1997

[3] Thomas Schmid, Oliver Egger, and Niels Kuster, "Automated E-field scanning system for dosimetric assessments", *IEEE Transaction on Microwave Theory and Techniques*, vol. 44, pp. 105-113, Jan. 1996.

[4] Niels Kuster, Ralph Kastle, and Thomas Schmid, "Dosimetic evaluation of mobile communications equipment with know precision", IEICE Transactions on Communications, vol. E80-B, no. 5, pp.645-652, May 1997.

[5]NIS81, NAMAS, "The treatment of uncertainty in EMC measurement", Tech. Rep., NAMAS Executive, National Physical Laboratory, Teddinton, Middlesex, England, 1994.

[6]Barry N. Tayor and Chris E. Kuyatt, "Guidelines for evaluating and expressing the uncertainty of NIST measurement results", Tech. Rep., National Institute of Standards and Technology, 1994.

[7]Federal Communications Commission, KDB 248227 - "SAR Measurement Procedures for 802.11 a/b/g Transmitters"

- [8] Federal Communications Commission, KDB 648474 "SAR Evaluation Considerations for Handsets with Multiple Transmitters and Antennas".
- [9] Federal Communications Commission, KDB 447498 "Mobile and Portable Device RF Exposure Procedures and Equipment Authorization Policies".
- [10] Federal Communications Commission, KDB 616217 "SAR Evaluation Considerations for Laptop Computers with Antennas Built-in on Display Screens".
- [11] Federal Communications Commission, KDB 450824 "SAR Probe Calibration and System Verification Considerations for Measurements at 150MHz 3GHz".
- [12] Federal Communications Commission, KDB 865664 "SAR Measurement Requirements for 3-6GHz".
- [13] Federal Communications Commission, KDB 941225 "SAR Measurement Procedures for 3G Devices".
- [14] ANSI, ANSI/IEEE C63.10-2009: American National Standard for Testing Unlicensed Wireless Devices.

Evaluation For: eNeura, Inc.

Product: sTMS Mini Transcranial Magnetic Stimulator

Date: 1/31/2019

APPENDIX A – SYSTEM VALIDATION SUMMARY

Per FCC KDB 865664, a tabulated summary of the system validation status including the validation date(s), measurement frequencies, SAR probes and tissue dielectric parameters have been included in the summary table below. The validation was performed with reference dipoles using the required tissue equivalent media for system validation according to KDB 865664. Each probe calibration point was validated at a frequency within the valid frequency range of the probe calibration point. All measurements were performed using probes calibrated for CW signals. Modulations in the table above represent test configurations for which the SAR system has been validated. The SAR system was also validated with modulated signals per KDB 865664.

				Probe Calibi	Probe Calibration Point		Dielectric Properties CV		CW Validation		Modulation Validation		ation
Frequency (MHz)	Date	Probe (SN#)	Probe (Model #)	Frequency (MHz)	Fluid Type	σ	€r	Sensitivity	Probe Linearity	Probe Isotropy	Mod. Type	Duty Factor	PAR
2450	1/13/2017	3516	EX3DV3	2450	Body	50.65	2.02	Pass	Pass	Pass	OFDM	N/A	Pass
5200	1/13/2017	3516	EX3DV3	5200	Body	48.71	5.54	Pass	Pass	Pass	OFDM	N/A	Pass
5500	1/13/2017	3516	EX3DV3	5500	Body	47.68	6.29	Pass	Pass	Pass	OFDM	N/A	Pass
5800	1/13/2017	3516	EX3DV3	5800	Body	48.71	5.54	Pass	Pass	Pass	OFDM	N/A	Pass

				Probe Calibration Point		Dielectric Properties		CW Validation			Modulation Validation		
Frequency		Probe	Probe	Frequency					Probe	Probe		Duty	
(MHz)	Date	(SN#)	(Model #)	(MHz)	Fluid Type	σ	€r	Sensitivity	Linearity	Isotropy	Mod. Type	Factor	PAR
835	1/14/2017	3516	EX3DV3	835	Body	54.2	0.98	Pass	Pass	Pass	GMSK	Pass	N/A
900	1/14/2017	3516	EX3DV3	900	Body	54	1.02	Pass	Pass	Pass	GMSK	Pass	N/A
1750	1/14/2017	3516	EX3DV3	1800	Body	52.9	1.41	Pass	Pass	Pass	GMSK	Pass	N/A
1900	1/14/2017	3516	EX3DV3	1900	Body	52.7	1.48	Pass	Pass	Pass	GMSK	Pass	N/A

Table 14: SAR System Validation Summary



Evaluation For: eNeura, Inc. Product: sTMS Mini Transcranial Magnetic Stimulator

Date: 1/31/2019

APPENDIX B – WORST CASE SAR PLOTS

Date/Time: 12/12/2018 12:01:13 PM

Test Laboratory: Intertek

File Name: SAR LTE Band 4.da52:4

9.1.1 SAR_LTE Band 4

Procedure Notes:

DUT: eNeura sTMS mini;

Communication System: UID 0, Generic LTE 20 MHz Bandwidth (0); Communication System Band: Band 4;

Frequency: 1732.5 MHz; Duty Cycle: 1:1

Medium parameters used (interpolated): f = 1732.5 MHz; $\sigma = 1.459 \text{ S/m}$; $\epsilon_r = 53.239$; $\rho = 1000 \text{ kg/m}^3$

Phantom section: Flat Section

Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)

DASY5 Configuration:

Probe: EX3DV3 - SN3516; ConvF(8.97, 8.97, 8.97); Calibrated: 11/12/2018;

Sensor-Surface: 1.4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn358; Calibrated: 11/6/2018

Phantom: SAM 2 with CRP v5.0; Type: QD000P40CD; Serial: TP:1663

DASY52 52.8.7(1137); SEMCAD X 14.6.10(7164)

WWAN Flat-Section MSL Testing/Body LTE Band 4, 1RB Low, Back Side, 5mm/Area Scan 2 (51x71x1):

Interpolated grid: dx=1.500 mm, dy=1.500 mm

Maximum value of SAR (interpolated) = 1.16 W/kg

WWAN Flat-Section MSL Testing/Body LTE Band 4, 1RB Low, Back Side, 5mm/Zoom Scan (9x9x7)/Cube 0:


Measurement grid: dx=5mm, dy=5mm, dz=5mm Reference Value = 26.677 V/m; Power Drift = -0.03 dB Peak SAR (extrapolated) = 1.47 W/kg

SAR(1 g) = 0.772 W/kg; SAR(10 g) = 0.472 W/kg

Maximum value of SAR (measured) = 1.10 W/kg



Evaluation For: eNeura, Inc. Product: sTMS Mini Transcranial Magnetic Stimulator



Evaluation For: eNeura, Inc. Product: sTMS Mini Transcranial Magnetic Stimulator

Date: 1/31/2019

Date/Time: 12/6/2018 12:35:05 PM

Test Laboratory: Intertek

File Name: SAR_LTE Band 13.da52:4

SAR_LTE Band 13

Procedure Notes:

DUT: eNeura sTMS mini; Serial:

Communication System: UID 0, Generic LTE 10 MHz Bandwidth (0); Communication System Band: Band 13;

Frequency: 782 MHz; Duty Cycle: 1:1

Medium parameters used (interpolated): f = 782 MHz; σ = 0.926 S/m; ε_r = 41.412; ρ = 1000 kg/m³

Phantom section: Flat Section

Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)

DASY5 Configuration:

Probe: EX3DV3 - SN3516; ConvF(11.02, 11.02, 11.02); Calibrated: 11/12/2018;

Sensor-Surface: 1.4mm (Mechanical Surface Detection), Sensor-Surface: 1.4mm (Fix Surface)

• Electronics: DAE4 Sn358; Calibrated: 11/6/2018

Phantom: SAM 2 with CRP v5.0; Type: QD000P40CD; Serial: TP:1663

DASY52 52.8.7(1137); SEMCAD X 14.6.10(7164)

WWAN Flat-Section MSL Testing/LTE Band 13, 1RB Low, Back Side, Direct Contact 2/Area Scan 2 (51x71x1):

Interpolated grid: dx=1.500 mm, dy=1.500 mm

Maximum value of SAR (interpolated) = 0.391 W/kg

WWAN Flat-Section MSL Testing/LTE Band 13, 1RB Low, Back Side, Direct Contact 2/Zoom Scan (12x13x7)/Cube

0: Measurement grid: dx=5mm, dy=5mm, dz=5mm Reference Value = 19.628 V/m; Power Drift = -0.13 dB Peak SAR (extrapolated) = 0.539 W/kg

SAR(1 g) = 0.302 W/kg; SAR(10 g) = 0.203 W/kg

Maximum value of SAR (measured) = 0.419 W/kg



Evaluation For: eNeura, Inc. Product: sTMS Mini Transcranial Magnetic Stimulator Date: 1/31/2019

SAR Test Report W/kg 0.391 0.315 0.239 0.162 0.086 0.0099

Evaluation For: eNeura, Inc. Product: sTMS Mini Transcranial Magnetic Stimulator

Date: 1/31/2019

APPENDIX C - DIPOLE VALIDATION PLOTS

Date/Time: 12/6/2018 8:59:12 AM

Test Laboratory: Intertek

File Name: Dipole 750MHz.da52:0

Dipole_750MHz

Procedure Notes: Ambient Temp: 22.8C, Fluid Temp: 22.2C

DUT: Dipole 750 MHz D750V3; Serial: D750V3 - SN:xxx

Communication System: UID 0, CW (0); Communication System Band: D835 (835.0 MHz); Frequency: 835

MHz; Duty Cycle: 1:1

Medium parameters used (extrapolated): f = 835 MHz; $\sigma = 0.968 \text{ S/m}$; $\epsilon_r = 40.67$; $\rho = 1000 \text{ kg/m}^3$

Phantom section: Flat Section

Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)

DASY5 Configuration:

Probe: EX3DV3 - SN3516; ConvF(10.53, 10.53, 10.53); Calibrated: 11/12/2018;

- Sensor-Surface: 2mm (Mechanical Surface Detection), Sensor-Surface: 4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn358; Calibrated: 11/6/2018
- Phantom: SAM 2 with CRP v5.0; Type: QD000P40CD; Serial: TP:1663
- DASY52 52.8.7(1137); SEMCAD X 14.6.10(7164)

System Performance Check at Frequencies below 1 GHz/d=10mm, Pin=100 mW, dist=2.0mm (EX-Probe)/Area Scan (31x121x1): Interpolated grid: dx=1.500 mm, dy=1.500 mm

Maximum value of SAR (interpolated) = 0.104 W/kg

System Performance Check at Frequencies below 1 GHz/d=10mm, Pin=100 mW, dist=2.0mm (EX-Probe)/Zoom Scan (7x7x7) (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 10.580 V/m; Power Drift = 0.00 dB

Peak SAR (extrapolated) = 12.3 W/kg


SAR(1 g) = 8.47 W/kg; SAR(10 g) = 5.69 W/kg

Normalized to target power = 1 W and actual power = 0.01 W

Maximum value of SAR (measured) = 9.07 W/kg



Evaluation For: eNeura, Inc.
Product: sTMS Mini Transcranial Magnetic Stimulator





Evaluation For: eNeura, Inc. Product: sTMS Mini Transcranial Magnetic Stimulator

Date: 1/31/2019

Date/Time: 12/12/2018 9:00:30 AM

Test Laboratory: Intertek

File Name: Dipole 1800MHz.da52:0

Dipole_1800MHz

Procedure Notes: Ambient Temp: 22.8C, Fluid Temp: 22.2C

DUT: Dipole 1800 MHz D1800V2; Serial: D1800V2 - SN:xxx

Communication System: UID 0, CW (0); Communication System Band: D1800 (1800.0 MHz); Frequency: 1800

MHz; Duty Cycle: 1:1

Medium parameters used: f = 1800 MHz; $\sigma = 1.53 \text{ S/m}$; $\varepsilon_r = 52.89$; $\rho = 1000 \text{ kg/m}^3$

Phantom section: Flat Section

Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)

DASY5 Configuration:

Probe: EX3DV3 - SN3516; ConvF(8.97, 8.97, 8.97); Calibrated: 11/12/2018;

Sensor-Surface: 2mm (Mechanical Surface Detection), Sensor-Surface: 4mm (Mechanical Surface Detection)

• Electronics: DAE4 Sn358; Calibrated: 11/6/2018

• Phantom: SAM 2 with CRP v5.0; Type: QD000P40CD; Serial: TP:1663

DASY52 52.8.7(1137); SEMCAD X 14.6.10(7164)

System Performance Check at Frequencies below 1 GHz/d=10mm, Pin=100 mW, dist=2.0mm (EX-Probe) 2/Area Scan (31x121x1): Interpolated grid: dx=1.500 mm, dy=1.500 mm

Maximum value of SAR (interpolated) = 0.575 W/kg

System Performance Check at Frequencies below 1 GHz/d=10mm, Pin=100 mW, dist=2.0mm (EX-Probe) 2/Zoom Scan (7x7x7) (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 17.877 V/m; Power Drift = 0.13 dB

Peak SAR (extrapolated) = 70.0 W/kg

SAR(1 g) = 41.9 W/kg; SAR(10 g) = 22.1 W/kg

Normalized to target power = 1 W and actual power = 0.01 W

Maximum value of SAR (measured) = 46.5 W/kg



Evaluation For: eNeura, Inc.
Product: sTMS Mini Transcranial Magnetic Stimulator

