

**Symbols:**


- Operating Instructions (ISO 7000-1641).
- Defibrillation-proof type BF applied part (IEC 60417-5334).
- Manufacturer (ISO 7000-3082).
- CE Mark.

**Text:**

- Use CR3032 Coin Cell Batteries only. Remove Batteries when not in use. FCC ID: 2ARNZ-1001. (Arial, font size 4)
- Notified Body Number (TGA: 0805). (Arial, font size 7)
- Manufacturer's details. (Arial, font size 4)
- Vpod Fixed Label Part Number: "WRH-N-06270-XX", where "XX" is the revision number. (Arial, font size 4)

**Note:**

- The outline (border) of the label is not printed.
- The corner radius of the label is 1.5mm.
- The artwork in this document must not be used for printing.
- All labels are printed directly from the BarTender software.



**Material:**

- Asset Label
- Silver Thermal Transfer Polyester Film
- Width: 30mm
- Length: 16mm
- Label Roll Core Size: 76mm
- To be printed on a thermal printer from the label stock "WRH-S-06316-01 Blank Asset Label Roll (W30 x L16)".

The label material has to meet the requirements from "AS NZS IEC 60601.1-2015, Medical electrical equipment - Part 1- General requirements for basic safety and essential performance":

- 7.1.2 Legibility of markings.
- 7.1.3 Durability of markings.

**Vendor:**

Any supplier who can meet the above specifications and requirements.



PART NUMBER: WRH-N-06270		DESCRIPTION: Vpod Fixed Label	REVISION: 02	SCALE: N/A
ORIGINATOR: Pascal Yee <i>Pascal Yee</i>		CHECKED BY: <i>[Signature]</i>	APPROVED BY: <i>[Signature]</i>	
<b>REVISION:</b>	<b>CHANGE DETAILS:</b>			<b>DATE:</b> <i>02 November 2018</i>
01	First Issue			31-Jan-2018
02	Added "FCC ID: 2ARNZ-1001." at the top of the label.			



Use CR3032 Coin Cell Batteries only. Remove batteries when not in use. FCC ID: 2ARNZ-1001

0805



Medical Manufacturers Pty Ltd  
Unit 131, 45 Gilby Road,  
Mt. Waverley, VIC 3149, Australia.

WRH-N-06270-02



2017-12-19

GTIN(01) 0 9352877 00000 5

(11) 171219

**SN** (21) 101-003020

WRH-N-06313-01