








## PRODUCT LABELING

### Proposed Label Location on EUT

FCC ID Label Location



<p><b>Nerve and Muscle Stimulator (Foot Drop System)</b> XFT-2001EA Input DC 3.7V, rechargeable lithium battery Wave form: Biphasic asymmetric balanced wave Frequency: 16-50Hz Pulse Width: 100-300µs</p> <p><b>IP67</b>    <b>Rx Only</b></p>	<p>Output: 0-90mA Size: 413×104×17mm Weigh: 139g Service Life: 5 years</p> 	<p> <b>Shenzhen XFT Medical Limited</b> Room203, Building 1, Biomedicine Innovations Industrial Park, #14 Jinhai Road, Pingshan New District, Shenzhen, China</p> <p><b>US Agent</b>    YangLingYun 605 Sherwood Forest Place, Cary, North Carolina, Cary, NC 27519 Phone: 919 4751044 Ext</p>
<p><b>FCC ID: 2AS2I2001EA</b> <b>UDI</b></p>	  	<p> Federal law restricts this device to sale by or on the order of a practitioner licensed by the law of the State in which he/she practices to use or order the use of the device.</p> 