



Date (26/05/2020)

Federal Communications Commission  
Equipment Authorization Branch  
7435 Oakland Mills Road  
Columbia, MD 21046

**Modular Approval Request**

**FCC ID:** 2ATDA-WMODHPA

**The following attestation addresses the requirements to support modular approval:**

| Modular approval requirement  | Yes (provide brief statement)   | No *   |
|---|---|--|
| (a) The radio elements must have the radio frequency circuitry shielded. Physical components and tuning capacitor(s) may be located external to the shield, but must be on the module assembly                                |   | No<br>No shield is provided due to limited modular application. The device is for internal use. Host will perform radiated emission according to published module integration guidance to verify that spurious emissions meets applicable rules. |
| (b) The module must have buffered modulation/data inputs to ensure that the device will comply with Part 15 requirements with any type of input signal  | Yes<br>The 2.4 GHz SoC CC8530RHAR have buffering capability   |  |
| (c) The module must contain power supply regulation on the module   | Yes<br>A DC regulator is implement on the module.   |  |
| (d) The module must contain a permanently attached antenna, or contain a unique antenna connector, and be marketed and operated only with specific antenna(s), per Sections 15.203, 15.204(b), 15.204(c), 15.212(a), 2.929(b) | Yes,<br>The module may be configured with an internal antenna or with two specific external antennas (Taoglas WPC5A and FCP70).<br>The internal and external antennas are mutually exclusive. |  |
| (e) The module must demonstrate compliance in a stand-alone configuration   |   | No.<br>Tested in a host device due to not having RF shielding.   |
| (f) The module must be labelled with its permanently affixed FCC ID label, or use an electronic display (See KDB Publication 784748 about labelling requirements)   | Yes,<br>FCC ID silk screen on back of module.   |  |
| (g) The module must comply with all specific rules applicable to the transmitter. The grantee must provide comprehensive instructions to explain  | Yes.<br>Manual contains installation instruction.   |  |



|                         |  |  |
|-------------------------|--|--|
| compliance requirements |  |  |
|-------------------------|--|--|

| Modular approval requirement                             | Yes (provide brief statement) | No * |
|--|-------------------------------|------|
| (h) The module must comply with RF exposure requirements | Yes.<br>SAR exempted.         |      |

\* Please provide a detailed explanation if the answer is "No."

Yours sincerely,

A handwritten signature in black ink, appearing to read "K. Anderson", with a long, sweeping horizontal stroke at the end.

Name: Kevan Anderson, CTO

Company: Innovere Medical Inc.