

## RF Exposure Statement

Product Name	MultiSure GK Link Blood Glucose and Ketone Monitoring System
Model No.	MTM309
FCC ID	2AV24MTM309

Device Operation Range	
Frequency Low (GHz)	Frequency High (GHz)
2.402	2.480

### 1. Standard Applicable

According to 1.1307 (b)(1), systems operating under the provisions of this section shall be operated in a manner that ensure that the public is not exposed to radio frequency energy level in excess of the Commission's guideline.

### 2. Measurement Result:

According to KDB 447498 D01 Mobile Portable RF Exposure v06,  
Operation frequency = 2402MHz and antenna separation distance = 5mm,  
SAR Test Exclusion Threshold = 10mW

The Device max peak output power = -0.75dBm, Equivalent to 0.8414mW and less than 10 mW

The SAR/MPE measurement is not necessary.