

D1750V2, Serial No. 1137 Extended Dipole Calibrations

Referring to KDB 865664 D01 v01r02, if dipoles are verified in return loss ($< -20\text{dB}$, within 20% of prior calibration), and in impedance (within 5 ohm of prior calibration), the annual calibration is not necessary and the calibration interval can be extended.


D1750V2 – serial no. 1137												
	1750 Head						1750 Body					
Date of Measurement	Return-Loss (dB)	Delta (%)	Real Impedance (ohm)	Delta (ohm)	Imaginary Impedance (ohm)	Delta (ohm)	Return-Loss (dB)	Delta (%)	Real Impedance (ohm)	Delta (ohm)	Imaginary Impedance (ohm)	Delta (ohm)
2018.07.30	-40.7		50.3		-0.87		-24.3		44.8		-2.59	
2019.10.23	-40.4	0.7	51	0.7	-0.15	0.72	-24.7	-1.6	46.1	1.3	-2.1	0.49

<Justification of the extended calibration>


The return loss is $< -20\text{dB}$, within 20% of prior calibration; the impedance is within 5 ohm of prior calibration. Therefore the verification result should support extended calibration.

Dipole Verification Data> D1750V2, serial no. 1137

1750MHz - Head



1750MHz – Body





In Collaboration with
s p e a g
CALIBRATION LABORATORY



中国认可
国际互认
校准
CALIBRATION
CNAS L0570

Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62304633-2079 Fax: +86-10-62304633-2504
E-mail: cttl@chinattl.com http://www.chinattl.cn

Certificate No: **Z18-60536**

Client **Sporton**

CALIBRATION CERTIFICATE

Object **D1900V2 - SN: 5d182**

Calibration Procedure(s) **FF-Z11-003-01**
Calibration Procedures for dipole validation kits

Calibration date: **December 7, 2018**

This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements(SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature(22±3)°C and humidity<70%.

Calibration Equipment used (M&TE critical for calibration)

Primary Standards	ID #	Cal Date(Calibrated by, Certificate No.)	Scheduled Calibration
Power Meter NRVD	102196	07-Mar-18 (CTTL, No.J18X01510)	Mar-19
Power sensor NRV-Z5	100596	07-Mar-18 (CTTL, No.J18X01510)	Mar-19
Reference Probe EX3DV4	SN 7514	27-Aug-18(SPEAG,No.EX3-7514_Aug18)	Aug-19
DAE4	SN 1555	20-Aug-18(SPEAG,No.DAE4-1555_Aug18)	Aug-19
Secondary Standards	ID #	Cal Date(Calibrated by, Certificate No.)	Scheduled Calibration
Signal Generator E4438C	MY49071430	23-Jan-18 (CTTL, No.J18X00560)	Jan-19
NetworkAnalyzer E5071C	MY46110673	24-Jan-18 (CTTL, No.J18X00561)	Jan-19

	Name	Function
Calibrated by:	Zhao Jing	SAR Test Engineer
Reviewed by:	Lin Hao	SAR Test Engineer
Approved by:	Qi Dianyuan	SAR Project Leader

Signature

Issued: December 10, 2018

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.



In Collaboration with

s p e a g
CALIBRATION LABORATORY

Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62304633-2079 Fax: +86-10-62304633-2504
E-mail: cttl@chinattl.com http://www.chinattl.cn

lossary:

TSL	tissue simulating liquid
ConvF	sensitivity in TSL / NORM _{x,y,z}
N/A	not applicable or not measured

Calibration is Performed According to the Following Standards:

- IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013
- IEC 62209-1, "Measurement procedure for assessment of specific absorption rate of human exposure to radio frequency fields from hand-held and body-mounted wireless communication devices- Part 1: Device used next to the ear (Frequency range of 300MHz to 6GHz)", July 2016
- IEC 62209-2, "Procedure to measure the Specific Absorption Rate (SAR) For wireless communication devices used in close proximity to the human body (frequency range of 30MHz to 6GHz)", March 2010
- KDB865664, SAR Measurement Requirements for 100 MHz to 6 GHz

Additional Documentation:

- DASY4/5 System Handbook

Methods Applied and Interpretation of Parameters:

- Measurement Conditions:** Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL:** The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss:** These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay:** One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured:** SAR measured at the stated antenna input power.
- SAR normalized:** SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters:** The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor $k=2$, which for a normal distribution Corresponds to a coverage probability of approximately 95%.



Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62304633-2079 Fax: +86-10-62304633-2504
E-mail: cttl@chinattl.com http://www.chinattl.cn

Measurement Conditions

DASY system configuration, as far as not given on page 1.

DASY Version	DASY52	52.10.2.1495
Extrapolation	Advanced Extrapolation	
Phantom	Triple Flat Phantom 5.1C	
Distance Dipole Center - TSL	10 mm	with Spacer
Zoom Scan Resolution	dx, dy, dz = 5 mm	
Frequency	1900 MHz $\pm$ 1 MHz	

Head TSL parameters

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	40.0	1.40 mho/m
Measured Head TSL parameters	(22.0 $\pm$ 0.2) °C	39.6 $\pm$ 6 %	1.44 mho/m $\pm$ 6 %
Head TSL temperature change during test	<1.0 °C	----	----

SAR result with Head TSL

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	
SAR measured	250 mW input power	10.1 mW / g
SAR for nominal Head TSL parameters	normalized to 1W	39.6 mW / g $\pm$ 18.8 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition	
SAR measured	250 mW input power	5.25 mW / g
SAR for nominal Head TSL parameters	normalized to 1W	20.7 mW / g $\pm$ 18.7 % (k=2)

Body TSL parameters

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Body TSL parameters	22.0 °C	53.3	1.52 mho/m
Measured Body TSL parameters	(22.0 $\pm$ 0.2) °C	51.8 $\pm$ 6 %	1.56 mho/m $\pm$ 6 %
Body TSL temperature change during test	<1.0 °C	----	----

SAR result with Body TSL

SAR averaged over 1 cm ³ (1 g) of Body TSL	Condition	
SAR measured	250 mW input power	10.2 mW / g
SAR for nominal Body TSL parameters	normalized to 1W	39.9 mW / g $\pm$ 18.8 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Body TSL	Condition	
SAR measured	250 mW input power	5.31 mW / g
SAR for nominal Body TSL parameters	normalized to 1W	20.9 mW / g $\pm$ 18.7 % (k=2)



Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62304633-2079 Fax: +86-10-62304633-2504
E-mail: cttl@chinattl.com http://www.chinattl.cn

Appendix (Additional assessments outside the scope of CNAS L0570)

Antenna Parameters with Head TSL

Impedance, transformed to feed point	$52.1\Omega + 5.35j\Omega$
Return Loss	- 25.0dB

Antenna Parameters with Body TSL

Impedance, transformed to feed point	$48.9\Omega + 6.19j\Omega$
Return Loss	- 24.0dB

General Antenna Parameters and Design

Electrical Delay (one direction)	1.067 ns
----------------------------------	----------

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

Additional EUT Data

Manufactured by	SPEAG
-----------------	-------



Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62304633-2079 Fax: +86-10-62304633-2504
E-mail: cttl@chinattl.com http://www.chinattl.cn

Date: 12.06.2018

DASY5 Validation Report for Head TSL

Test Laboratory: CTTL, Beijing, China

DUT: Dipole 1900 MHz; Type: D1900V2; Serial: D1900V2 - SN: 5d182

Communication System: UID 0, CW; Frequency: 1900 MHz; Duty Cycle: 1:1

Medium parameters used: $f = 1900$ MHz; $\sigma = 1.441$ S/m; $\epsilon_r = 39.59$; $\rho = 1000$ kg/m³

Phantom section: Center Section

DASY5 Configuration:

- Probe: EX3DV4 - SN7514; ConvF(7.73, 7.73, 7.73) @ 1900 MHz; Calibrated: 8/27/2018
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn1555; Calibrated: 8/20/2018
- Phantom: MFP_V5.1C ; Type: QD 000 P51CA; Serial: 1062
- Measurement SW: DASY52, Version 52.10 (2); SEMCAD X Version 14.6.12 (7450)

System Performance Check/Zoom Scan (7x7x7) (7x7x7)/Cube 0: Measurement grid:


$dx=5$ mm, $dy=5$ mm, $dz=5$ mm

Reference Value = 95.91 V/m; Power Drift = 0.00 dB

Peak SAR (extrapolated) = 19.3 W/kg

SAR(1 g) = 10.1 W/kg; SAR(10 g) = 5.25 W/kg

Maximum value of SAR (measured) = 15.8 W/kg




0 dB = 15.8 W/kg = 11.99 dBW/kg



In Collaboration with
s p e a g
CALIBRATION LABORATORY

Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62304633-2079 Fax: +86-10-62304633-2504
E-mail: cttl@chinattl.com http://www.chinattl.cn

Impedance Measurement Plot for Head TSL





In Collaboration with

s p e a g
CALIBRATION LABORATORY

Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62304633-2079 Fax: +86-10-62304633-2504
E-mail: cttl@chinattl.com http://www.chinattl.cn

DASY5 Validation Report for Body TSL

Date: 12.05.2018

Test Laboratory: CTTL, Beijing, China

DUT: Dipole 1900 MHz; Type: D1900V2; Serial: D1900V2 - SN: 5d182

Communication System: UID 0, CW; Frequency: 1900 MHz; Duty Cycle: 1:1

Medium parameters used: $f = 1900$ MHz; $\sigma = 1.564$ S/m; $\epsilon_r = 51.82$; $\rho = 1000$ kg/m³

Phantom section: Right Section

DASY5 Configuration:

- Probe: EX3DV4 - SN7514; ConvF(7.53, 7.53, 7.53) @ 1900 MHz; Calibrated: 8/27/2018
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn1555; Calibrated: 8/20/2018
- Phantom: MFP_V5.1C ; Type: QD 000 P51CA; Serial: 1062
- Measurement SW: DASY52, Version 52.10 (2); SEMCAD X Version 14.6.12 (7450)

System Performance Check/Zoom Scan (7x7x7) (7x7x7)/Cube 0: Measurement grid:


$dx=5$ mm, $dy=5$ mm, $dz=5$ mm

Reference Value = 84.07 V/m; Power Drift = -0.04 dB

Peak SAR (extrapolated) = 18.9 W/kg

SAR(1 g) = 10.2 W/kg; SAR(10 g) = 5.31 W/kg

Maximum value of SAR (measured) = 15.7 W/kg






In Collaboration with

s p e a g
CALIBRATION LABORATORY

Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62304633-2079 Fax: +86-10-62304633-2504
E-mail: cttl@chinattl.com http://www.chinattl.cn

Impedance Measurement Plot for Body TSL



D1900V2, Serial No. 5d182 Extended Dipole Calibrations

Referring to KDB 865664 D01 v01r02, if dipoles are verified in return loss ($< -20\text{dB}$, within 20% of prior calibration), and in impedance (within 5 ohm of prior calibration), the annual calibration is not necessary and the calibration interval can be extended.


D1900V2 – serial no. 5d182												
	1900 Head						1900 Body					
Date of Measurement	Return-Loss (dB)	Delta (%)	Real Impedance (ohm)	Delta (ohm)	Imaginary Impedance (ohm)	Delta (ohm)	Return-Loss (dB)	Delta (%)	Real Impedance (ohm)	Delta (ohm)	Imaginary Impedance (ohm)	Delta (ohm)
2018.12.7	-25		52.1		5.35		-24		48.9		6.19	
2019.11.25	-25.2	-0.8	53.9	1.8	5.15	-0.2	-24.2	-0.8	48.7	-0.2	5.93	-0.26

<Justification of the extended calibration>


The return loss is $< -20\text{dB}$, within 20% of prior calibration; the impedance is within 5 ohm of prior calibration. Therefore the verification result should support extended calibration.

Dipole Verification Data> D1900V2, serial no. 5d182

1900MHz - Head



1900MHz - Body





In Collaboration with
s p e a g
CALIBRATION LABORATORY



中国认可
国际互认
校准
CALIBRATION
CNAS L0570

Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62304633-2079 Fax: +86-10-62304633-2504
E-mail: cttl@chinattl.com http://www.chinattl.cn

Client

Sporton

Certificate No: **Z19-60134**

CALIBRATION CERTIFICATE

Object **D2450V2 - SN: 924**

Calibration Procedure(s) **FF-Z11-003-01**
Calibration Procedures for dipole validation kits

Calibration date: **April 15, 2019**

This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements(SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature(22 ± 3)°C and humidity<70%.

Calibration Equipment used (M&TE critical for calibration)

Primary Standards	ID #	Cal Date(Calibrated by, Certificate No.)	Scheduled Calibration
Power Meter NRP2	106277	20-Aug-18 (CTTL, No.J18X06862)	Aug-19
Power sensor NRP8S	104291	20-Aug-18 (CTTL, No.J18X06862)	Aug-19
Reference Probe EX3DV4	SN 3617	31-Jan-19(SPEAG,No.EX3-3617_Jan19)	Jan-20
DAE4	SN 1331	06-Feb-19(SPEAG,No.DAE4-1331_Feb19)	Feb-20
Secondary Standards	ID #	Cal Date(Calibrated by, Certificate No.)	Scheduled Calibration
Signal Generator E4438C	MY49071430	23-Jan-19 (CTTL, No.J19X00336)	Jan-20
NetworkAnalyzer E5071C	MY46110673	24-Jan-19 (CTTL, No.J19X00547)	Jan-20

	Name	Function
Calibrated by:	Zhao Jing	SAR Test Engineer
Reviewed by:	Lin Hao	SAR Test Engineer
Approved by:	Qi Dianyuan	SAR Project Leader

Signature

Issued: April 20, 2019

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.



In Collaboration with

s p e a g
CALIBRATION LABORATORY

Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62304633-2079 Fax: +86-10-62304633-2504
E-mail: cttl@chinattl.com http://www.chinattl.cn

Glossary:

TSL	tissue simulating liquid
ConvF	sensitivity in TSL / NORM _{x,y,z}
N/A	not applicable or not measured

Calibration is Performed According to the Following Standards:

- IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013
- IEC 62209-1, "Measurement procedure for assessment of specific absorption rate of human exposure to radio frequency fields from hand-held and body-mounted wireless communication devices- Part 1: Device used next to the ear (Frequency range of 300MHz to 6GHz)", July 2016
- IEC 62209-2, "Procedure to measure the Specific Absorption Rate (SAR) For wireless communication devices used in close proximity to the human body (frequency range of 30MHz to 6GHz)", March 2010
- KDB865664, SAR Measurement Requirements for 100 MHz to 6 GHz

Additional Documentation:

- DASY4/5 System Handbook

Methods Applied and Interpretation of Parameters:

- Measurement Conditions:** Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL:** The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss:** These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay:** One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured:** SAR measured at the stated antenna input power.
- SAR normalized:** SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters:** The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor $k=2$, which for a normal distribution Corresponds to a coverage probability of approximately 95%.



In Collaboration with

s p e a g**CALIBRATION LABORATORY**

Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China

Tel: +86-10-62304633-2079

Fax: +86-10-62304633-2504

E-mail: cttl@chinattl.com

http://www.chinattl.cn

Measurement Conditions

DASY system configuration, as far as not given on page 1.

DASY Version	DASY52	52.10.2.1495
Extrapolation	Advanced Extrapolation	
Phantom	Triple Flat Phantom 5.1C	
Distance Dipole Center - TSL	10 mm	with Spacer
Zoom Scan Resolution	dx, dy, dz = 5 mm	
Frequency	2450 MHz $\pm$ 1 MHz	

Head TSL parameters

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	39.2	1.80 mho/m
Measured Head TSL parameters	(22.0 $\pm$ 0.2) °C	40.4 $\pm$ 6 %	1.85 mho/m $\pm$ 6 %
Head TSL temperature change during test	<1.0 °C	----	----

SAR result with Head TSL

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	
SAR measured	250 mW input power	13.1 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	52.1 W/kg $\pm$ 18.8 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition	
SAR measured	250 mW input power	5.99 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	23.9 W/kg $\pm$ 18.7 % (k=2)

Body TSL parameters

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Body TSL parameters	22.0 °C	52.7	1.95 mho/m
Measured Body TSL parameters	(22.0 $\pm$ 0.2) °C	54.3 $\pm$ 6 %	2.01 mho/m $\pm$ 6 %
Body TSL temperature change during test	<1.0 °C	----	----

SAR result with Body TSL

SAR averaged over 1 cm ³ (1 g) of Body TSL	Condition	
SAR measured	250 mW input power	12.6 W/kg
SAR for nominal Body TSL parameters	normalized to 1W	50.1 W/kg $\pm$ 18.8 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Body TSL	Condition	
SAR measured	250 mW input power	5.83 W/kg
SAR for nominal Body TSL parameters	normalized to 1W	23.3 W/kg $\pm$ 18.7 % (k=2)



In Collaboration with

s p e a g
CALIBRATION LABORATORY

Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62304633-2079 Fax: +86-10-62304633-2504
E-mail: cttl@chinattl.com http://www.chinattl.cn

Appendix (Additional assessments outside the scope of CNAS L0570)

Antenna Parameters with Head TSL

Impedance, transformed to feed point	51.9Ω+ 2.68 jΩ
Return Loss	- 29.9dB

Antenna Parameters with Body TSL

Impedance, transformed to feed point	48.8Ω+ 4.17 jΩ
Return Loss	- 27.2dB

General Antenna Parameters and Design

Electrical Delay (one direction)	1.019 ns
----------------------------------	----------

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.
No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

Additional EUT Data

Manufactured by	SPEAG
-----------------	-------



In Collaboration with

s p e a g
CALIBRATION LABORATORY

Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62304633-2079 Fax: +86-10-62304633-2504
E-mail: cttl@chinattl.com http://www.chinattl.cn

DASY5 Validation Report for Head TSL

Date: 04.15.2019

Test Laboratory: CTTL, Beijing, China

DUT: Dipole 2450 MHz; Type: D2450V2; Serial: D2450V2 - SN: 924

Communication System: UID 0, CW; Frequency: 2450 MHz; Duty Cycle: 1:1

Medium parameters used: $f = 2450$ MHz; $\sigma = 1.85$ S/m; $\epsilon_r = 40.35$; $\rho = 1000$ kg/m³

Phantom section: Right Section

DASY5 Configuration:

- Probe: EX3DV4 - SN3617; ConvF(7.62, 7.62, 7.62) @ 2450 MHz; Calibrated: 1/31/2019
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn1331; Calibrated: 2/6/2019
- Phantom: MFP_V5.1C ; Type: QD 000 P51CA; Serial: 1062
- Measurement SW: DASY52, Version 52.10 (2); SEMCAD X Version 14.6.12 (7450)


Dipole Calibration/Zoom Scan (7x7x7) (7x7x7)/Cube 0: Measurement grid: $dx=5$ mm, $dy=5$ mm, $dz=5$ mm

Reference Value = 86.73 V/m; Power Drift = 0.05 dB

Peak SAR (extrapolated) = 28.0 W/kg

SAR(1 g) = 13.1 W/kg; SAR(10 g) = 5.99 W/kg

Maximum value of SAR (measured) = 22.2 W/kg



0 dB = 22.2 W/kg = 13.46 dBW/kg




In Collaboration with

s p e a g
CALIBRATION LABORATORY

Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62304633-2079 Fax: +86-10-62304633-2504
E-mail: cttl@chinattl.com http://www.chinattl.cn

Impedance Measurement Plot for Head TSL





In Collaboration with

s p e a g
CALIBRATION LABORATORY

Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62304633-2079 Fax: +86-10-62304633-2504
E-mail: cttl@chinattl.com http://www.chinattl.cn

DASY5 Validation Report for Body TSL

Date: 04.15.2019

Test Laboratory: CTTL, Beijing, China

DUT: Dipole 2450 MHz; Type: D2450V2; Serial: D2450V2 - SN: 924

Communication System: UID 0, CW; Frequency: 2450 MHz; Duty Cycle: 1:1

Medium parameters used: $f = 2450$ MHz; $\sigma = 2.005$ S/m; $\epsilon_r = 54.25$; $\rho = 1000$ kg/m³

Phantom section: Center Section

DASY5 Configuration:

- Probe: EX3DV4 - SN3617; ConvF(7.79, 7.79, 7.79) @ 2450 MHz; Calibrated: 1/31/2019
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn1331; Calibrated: 2/6/2019
- Phantom: MFP_V5.1C ; Type: QD 000 P51CA; Serial: 1062
- Measurement SW: DASY52, Version 52.10 (2); SEMCAD X Version 14.6.12 (7450)


Dipole Calibration/Zoom Scan (7x7x7) (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 95.46 V/m; Power Drift = -0.02 dB

Peak SAR (extrapolated) = 26.3 W/kg

SAR(1 g) = 12.6 W/kg; SAR(10 g) = 5.83 W/kg

Maximum value of SAR (measured) = 20.9 W/kg



0 dB = 20.9 W/kg = 13.20 dBW/kg




In Collaboration with

s p e a g
CALIBRATION LABORATORY

Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62304633-2079 Fax: +86-10-62304633-2504
E-mail: cttl@chinattl.com http://www.chinattl.cn

Impedance Measurement Plot for Body TSL





In Collaboration with
s p e a g
CALIBRATION LABORATORY



中国认可
国际互认
校准
CALIBRATION
CNAS L0570

Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62304633-2512 Fax: +86-10-62304633-2504
E-mail: cttl@chinattl.com http://www.chinattl.cn

Client

Sporton

Certificate No: **Z18-60259**

CALIBRATION CERTIFICATE

Object **D5GHzV2 - SN: 1167**

Calibration Procedure(s) **FF-Z11-003-01**
Calibration Procedures for dipole validation kits

Calibration date: **August 03, 2018**

This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements(SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature(22±3)°C and humidity<70%.

Calibration Equipment used (M&TE critical for calibration)

Primary Standards	ID #	Cal Date(Calibrated by, Certificate No.)	Scheduled Calibration
Power Meter NRP2	102083	01-Nov-17 (CTTL, No.J17X08756)	Oct-18
Power sensor NRP-Z91	100542	01-Nov-17 (CTTL, No.J17X08756)	Oct-18
ReferenceProbe EX3DV4	SN 7464	12-Sep-17(SPEAG,No.EX3-7464_Sep17)	Sep-18
DAE4	SN 1524	13-Sep-17(SPEAG,No.DAE4-1524_Sep17)	Sep-18
Secondary Standards	ID #	Cal Date(Calibrated by, Certificate No.)	Scheduled Calibration
Signal Generator E4438C	MY49071430	23-Jan-18 (CTTL, No.J18X00560)	Jan-19
NetworkAnalyzerE5071C	MY46110673	24-Jan-18 (CTTL, No.J18X00561)	Jan-19

	Name	Function	Signature
Calibrated by:	Zhao Jing	SAR Test Engineer	
Reviewed by:	Lin Hao	SAR Test Engineer	
Approved by:	Qi Dianyuan	SAR Project Leader	

Issued: August 6, 2018

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.



In Collaboration with

s p e a g
CALIBRATION LABORATORY

Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62304633-2512 Fax: +86-10-62304633-2504
E-mail: cttl@chinattl.com http://www.chinattl.cn

Glossary:

TSL	tissue simulating liquid
ConvF	sensitivity in TSL / NORM _{x,y,z}
N/A	not applicable or not measured

Calibration is Performed According to the Following Standards:

- IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013
- IEC 62209-1, "Measurement procedure for assessment of specific absorption rate of human exposure to radio frequency fields from hand-held and body-mounted wireless communication devices- Part 1: Device used next to the ear (Frequency range of 300MHz to 6GHz)", July 2016
- IEC 62209-2, "Procedure to measure the Specific Absorption Rate (SAR) For wireless communication devices used in close proximity to the human body (frequency range of 30MHz to 6GHz)", March 2010
- KDB865664, SAR Measurement Requirements for 100 MHz to 6 GHz

Additional Documentation:

- DASY4/5 System Handbook

Methods Applied and Interpretation of Parameters:

- Measurement Conditions:** Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL:** The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss:** These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay:** One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured:** SAR measured at the stated antenna input power.
- SAR normalized:** SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters:** The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor $k=2$, which for a normal distribution Corresponds to a coverage probability of approximately 95%.



In Collaboration with

s p e a g
CALIBRATION LABORATORY

Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-62304633-2512 Fax: +86-10-62304633-2504
 E-mail: cttl@chinattl.com http://www.chinattl.cn

Measurement Conditions

DASY system configuration, as far as not given on page 1.

DASY Version	DASY52	52.10.1.1476
Extrapolation	Advanced Extrapolation	
Phantom	Triple Flat Phantom 5.1C	
Distance Dipole Center - TSL	10 mm	with Spacer
Zoom Scan Resolution	dx, dy = 4 mm, dz = 1.4 mm	Graded Ratio = 1.4 (Z direction)
Frequency	5250 MHz $\pm$ 1 MHz 5600 MHz $\pm$ 1 MHz 5750 MHz $\pm$ 1 MHz	

Head TSL parameters at 5250 MHz

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	35.9	4.71 mho/m
Measured Head TSL parameters	(22.0 $\pm$ 0.2) °C	35.9 $\pm$ 6 %	4.82 mho/m $\pm$ 6 %
Head TSL temperature change during test	<1.0 °C	----	----

SAR result with Head TSL at 5250 MHz

SAR averaged over 1 cm³ (1 g) of Head TSL	Condition	
SAR measured	100 mW input power	7.69 mW / g
SAR for nominal Head TSL parameters	normalized to 1W	77.0 mW /g $\pm$ 24.4 % (k=2)
SAR averaged over 10 cm³ (10 g) of Head TSL	Condition	
SAR measured	100 mW input power	2.20 mW / g
SAR for nominal Head TSL parameters	normalized to 1W	22.0 mW /g $\pm$ 24.2 % (k=2)



Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62304633-2512 Fax: +86-10-62304633-2504
E-mail: cttl@chinattl.com http://www.chinattl.cn

Head TSL parameters at 5600 MHz

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	35.5	5.07 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	35.1 ± 6 %	5.18 mho/m ± 6 %
Head TSL temperature change during test	<1.0 °C	----	----

SAR result with Head TSL at 5600 MHz

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	
SAR measured	100 mW input power	8.09 mW / g
SAR for nominal Head TSL parameters	normalized to 1W	80.8 mW / g ± 24.4 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition	
SAR measured	100 mW input power	2.32 mW / g
SAR for nominal Head TSL parameters	normalized to 1W	23.2 mW / g ± 24.2 % (k=2)

Head TSL parameters at 5750 MHz

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	35.4	5.22 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	34.9 ± 6 %	5.37 mho/m ± 6 %
Head TSL temperature change during test	<1.0 °C	----	----

SAR result with Head TSL at 5750 MHz

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	
SAR measured	100 mW input power	7.70 mW / g
SAR for nominal Head TSL parameters	normalized to 1W	76.9 mW / g ± 24.4 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition	
SAR measured	100 mW input power	2.17 mW / g
SAR for nominal Head TSL parameters	normalized to 1W	21.6 mW / g ± 24.2 % (k=2)



In Collaboration with

s p e a g
CALIBRATION LABORATORY

Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62304633-2512 Fax: +86-10-62304633-2504
E-mail: cttl@chinattl.com http://www.chinattl.cn

Body TSL parameters at 5250 MHz

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Body TSL parameters	22.0 °C	48.9	5.36 mho/m
Measured Body TSL parameters	(22.0 ± 0.2) °C	48.4 ± 6 %	5.32 mho/m ± 6 %
Body TSL temperature change during test	<1.0 °C	----	----

SAR result with Body TSL at 5250 MHz

SAR averaged over 1 cm ³ (1 g) of Body TSL	Condition	
SAR measured	100 mW input power	7.46 mW / g
SAR for nominal Body TSL parameters	normalized to 1W	74.4 mW /g ± 24.4 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Body TSL	Condition	
SAR measured	100 mW input power	2.10 mW / g
SAR for nominal Body TSL parameters	normalized to 1W	20.9 mW /g ± 24.2 % (k=2)

Body TSL parameters at 5600 MHz

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Body TSL parameters	22.0 °C	48.5	5.77 mho/m
Measured Body TSL parameters	(22.0 ± 0.2) °C	47.7 ± 6 %	5.79 mho/m ± 6 %
Body TSL temperature change during test	<1.0 °C	----	----

SAR result with Body TSL at 5600 MHz

SAR averaged over 1 cm ³ (1 g) of Body TSL	Condition	
SAR measured	100 mW input power	7.73 mW / g
SAR for nominal Body TSL parameters	normalized to 1W	77.1 mW /g ± 24.4 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Body TSL	Condition	
SAR measured	100 mW input power	2.16 mW / g
SAR for nominal Body TSL parameters	normalized to 1W	21.5 mW /g ± 24.2 % (k=2)



In Collaboration with

s p e a g
CALIBRATION LABORATORY

Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62304633-2512 Fax: +86-10-62304633-2504
E-mail: cttl@chinattl.com http://www.chinattl.cn

Body TSL parameters at 5750 MHz

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Body TSL parameters	22.0 °C	48.3	5.94 mho/m
Measured Body TSL parameters	(22.0 ± 0.2) °C	48.5 ± 6 %	5.93 mho/m ± 6 %
Body TSL temperature change during test	<1.0 °C	----	----

SAR result with Body TSL at 5750 MHz

SAR averaged over 1 cm ³ (1 g) of Body TSL	Condition	
SAR measured	100 mW input power	7.43 mW / g
SAR for nominal Body TSL parameters	normalized to 1W	74.3 mW /g ± 24.4 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Body TSL	Condition	
SAR measured	100 mW input power	2.08 mW / g
SAR for nominal Body TSL parameters	normalized to 1W	20.8 mW /g ± 24.2 % (k=2)



In Collaboration with

s p e a g
CALIBRATION LABORATORY

Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62304633-2512 Fax: +86-10-62304633-2504
E-mail: cttl@chinattl.com http://www.chinattl.cn

Appendix (Additional assessments outside the scope of CNAS L0570)

Antenna Parameters with Head TSL at 5250 MHz

Impedance, transformed to feed point	50.3 Ω - 9.42j Ω
Return Loss	- 20.6dB

Antenna Parameters with Head TSL at 5600 MHz

Impedance, transformed to feed point	58.1 Ω - 7.15j Ω
Return Loss	- 20.0dB

Antenna Parameters with Head TSL at 5750 MHz

Impedance, transformed to feed point	53.5 Ω - 7.66j Ω
Return Loss	- 21.8dB

Antenna Parameters with Body TSL at 5250 MHz

Impedance, transformed to feed point	49.5 Ω - 7.40j Ω
Return Loss	- 22.6dB

Antenna Parameters with Body TSL at 5600 MHz

Impedance, transformed to feed point	58.0 Ω - 6.37j Ω
Return Loss	- 20.5dB

Antenna Parameters with Body TSL at 5750 MHz

Impedance, transformed to feed point	54.5 Ω - 7.07j Ω
Return Loss	- 21.9dB



In Collaboration with

s p e a g
CALIBRATION LABORATORY

Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62304633-2512 Fax: +86-10-62304633-2504
E-mail: cttl@chinattl.com <http://www.chinattl.cn>

General Antenna Parameters and Design

Electrical Delay (one direction)	1.065 ns
----------------------------------	----------

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

Additional EUT Data

Manufactured by	SPEAG
-----------------	-------



In Collaboration with

s p e a g
CALIBRATION LABORATORY

Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62304633-2512 Fax: +86-10-62304633-2504
E-mail: cttl@chinattl.com http://www.chinattl.cn

DASY5 Validation Report for Head TSL

Date: 07.27.2018

Test Laboratory: CTTL, Beijing, China

DUT: Dipole 5GHz; Type: D5GHzV2; Serial: D5GHzV2 - SN: 1167

Communication System: CW; Frequency: 5250 MHz, Frequency: 5600 MHz,
Frequency: 5750 MHz,

Medium parameters used: $f = 5250$ MHz; $\sigma = 4.822$ S/m; $\epsilon_r = 35.92$; $\rho = 1000$ kg/m³, Medium parameters used: $f = 5600$ MHz; $\sigma = 5.184$ S/m; $\epsilon_r = 35.14$; $\rho = 1000$ kg/m³, Medium parameters used: $f = 5750$ MHz; $\sigma = 5.365$ S/m; $\epsilon_r = 34.88$; $\rho = 1000$ kg/m³,

Phantom section: Center Section

DASY5 Configuration:

- Probe: EX3DV4 - SN7464; ConvF(5.68, 5.68, 5.68) @ 5250 MHz; Calibrated: 9/12/2017, ConvF(4.98, 4.98, 4.98) @ 5600 MHz; Calibrated: 9/12/2017, ConvF(5.04, 5.04, 5.04) @ 5750 MHz; Calibrated: 9/12/2017,
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn1524; Calibrated: 9/13/2017
- Phantom: MFP_V5.1C ; Type: QD 000 P51CA; Serial: 1062
- Measurement SW: DASY52, Version 52.10 (1); SEMCAD X Version 14.6.11 (7439)

Dipole Calibration /Pin=100mW, d=10mm, f=5250 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm
Reference Value = 65.09 V/m; Power Drift = -0.02 dB
Peak SAR (extrapolated) = 32.4 W/kg
SAR(1 g) = 7.69 W/kg; SAR(10 g) = 2.2 W/kg
Maximum value of SAR (measured) = 18.0 W/kg


Dipole Calibration /Pin=100mW, d=10mm, f=5600 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm
Reference Value = 63.53 V/m; Power Drift = -0.01 dB
Peak SAR (extrapolated) = 36.2 W/kg
SAR(1 g) = 8.09 W/kg; SAR(10 g) = 2.32 W/kg
Maximum value of SAR (measured) = 19.7 W/kg

Dipole Calibration /Pin=100mW, d=10mm, f=5750 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm
Reference Value = 63.79 V/m; Power Drift = 0.01 dB
Peak SAR (extrapolated) = 36.2 W/kg
SAR(1 g) = 7.7 W/kg; SAR(10 g) = 2.17 W/kg
Maximum value of SAR (measured) = 19.0 W/kg



In Collaboration with
s p e a g
CALIBRATION LABORATORY

Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62304633-2512 Fax: +86-10-62304633-2504
E-mail: cttl@chinattl.com <http://www.chinattl.cn>




0 dB = 19.0 W/kg = 12.79 dBW/kg



Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62304633-2512 Fax: +86-10-62304633-2504
E-mail: cttl@chinattl.com http://www.chinattl.cn


Impedance Measurement Plot for Head TSL

Tr1 S11 Log Mag 10.00dB/ Ref 0.000dB [F1]



Tr2 S11 Smith (R+jX) scale 1.000U [F1 Del]

>1	5.250000 GHz	50.314 Ω	-9.4158 Ω	3.2196 pF
2	5.600000 GHz	58.136 Ω	-7.1538 Ω	3.9728 pF
3	5.750000 GHz	53.485 Ω	-7.6562 Ω	3.6153 pF





In Collaboration with

s p e a g
CALIBRATION LABORATORY

Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62304633-2512 Fax: +86-10-62304633-2504
E-mail: cttl@chinattl.com http://www.chinattl.cn

Date: 08.02.2018

DASY5 Validation Report for Body TSL

Test Laboratory: CTTL, Beijing, China

DUT: Dipole 5GHz; Type: D5GHzV2; Serial: D5GHzV2 - SN: 1167

Communication System: CW; Frequency: 5250 MHz, Frequency: 5600 MHz,
Frequency: 5750 MHz,

Medium parameters used: $f = 5250$ MHz; $\sigma = 5.316$ S/m; $\epsilon_r = 48.42$; $\rho = 1000$ kg/m³, Medium parameters used: $f = 5600$ MHz; $\sigma = 5.789$ S/m; $\epsilon_r = 47.7$; $\rho = 1000$ kg/m³, Medium parameters used: $f = 5750$ MHz; $\sigma = 5.926$ S/m; $\epsilon_r = 48.45$; $\rho = 1000$ kg/m³,

Phantom section: Right Section

DASY5 Configuration:

- Probe: EX3DV4 - SN7464; ConvF(5.29, 5.29, 5.29) @ 5250 MHz; Calibrated: 9/12/2017, ConvF(4.5, 4.5, 4.5) @ 5600 MHz; Calibrated: 9/12/2017, ConvF(4.59, 4.59, 4.59) @ 5750 MHz; Calibrated: 9/12/2017,
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn1524; Calibrated: 9/13/2017
- Phantom: MFP_V5.1C ; Type: QD 000 P51CA; Serial: 1062
- Measurement SW: DASY52, Version 52.10 (1); SEMCAD X Version 14.6.11 (7439)

Dipole Calibration /Pin=100mW, d=10mm, f=5250 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm
Reference Value = 64.14 V/m; Power Drift = 0.02 dB
Peak SAR (extrapolated) = 31.9 W/kg
SAR(1 g) = 7.46 W/kg; SAR(10 g) = 2.1 W/kg
Maximum value of SAR (measured) = 17.6 W/kg

Dipole Calibration /Pin=100mW, d=10mm, f=5600 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm
Reference Value = 62.32 V/m; Power Drift = -0.01 dB
Peak SAR (extrapolated) = 36.3 W/kg
SAR(1 g) = 7.73 W/kg; SAR(10 g) = 2.16 W/kg
Maximum value of SAR (measured) = 19.1 W/kg


Dipole Calibration /Pin=100mW, d=10mm, f=5750 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm
Reference Value = 63.99 V/m; Power Drift = 0.02 dB
Peak SAR (extrapolated) = 35.2 W/kg
SAR(1 g) = 7.43 W/kg; SAR(10 g) = 2.08 W/kg
Maximum value of SAR (measured) = 18.0 W/kg



In Collaboration with

s p e a k
CALIBRATION LABORATORY

Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62304633-2512 Fax: +86-10-62304633-2504
E-mail: cttl@chinattl.com <http://www.chinattl.cn>



0 dB = 18.0 W/kg = 12.55 dBW/kg



In Collaboration with

s p e a g
CALIBRATION LABORATORY

Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China


Tel: +86-10-62304633-2512

Fax: +86-10-62304633-2504

E-mail: cttl@chinattl.com

http://www.chinattl.cn

Impedance Measurement Plot for Body TSL




<Justification of the extended calibration>


The return loss is $< -20\text{dB}$, within 20% of prior calibration; the impedance is within 5 ohm of prior calibration. Therefore the verification result should support extended calibration.

Dipole Verification Data> D5GHzV3, serial no. 1167


5250MHz - Head




5250MHz – Body




5600MHz – Head




5600MHz – Body



5750MHz – Head



5750MHz – Body





In Collaboration with
s p e a g
CALIBRATION LABORATORY



中国认可
国际互认
校准
CALIBRATION
CNAS L0570

Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62304633-2512 Fax: +86-10-62304633-2504
E-mail: cttl@chinattl.com [Http://www.chinattl.cn](http://www.chinattl.cn)

Client : **Sporton**

Certificate No: **Z19-60436**

CALIBRATION CERTIFICATE

Object

DAE4 - SN: 1437

Calibration Procedure(s)

FF-Z11-002-01

Calibration Procedure for the Data Acquisition Electronics
(DAEx)

Calibration date:

November 19, 2019

This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements(SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature(22 ± 3) $^{\circ}\text{C}$ and humidity<70%.

Calibration Equipment used (M&TE critical for calibration)

Primary Standards	ID #	Cal Date(Calibrated by, Certificate No.)	Scheduled Calibration
Process Calibrator 753	1971018	24-Jun-19 (CTTL, No.J19X05126)	Jun-20

	Name	Function
Calibrated by:	Yu Zongying	SAR Test Engineer
Reviewed by:	Zhao Jing	SAR Test Engineer
Approved by:	Qi Dianyuan	SAR Project Leader

Signature

Issued: November 21, 2019

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.



In Collaboration with

s p e a g

CALIBRATION LABORATORY

Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62304633-2512 Fax: +86-10-62304633-2504
E-mail: cttl@chinattl.com Http://www.chinattl.cn

Glossary:

DAE data acquisition electronics
Connector angle information used in DASY system to align probe sensor X to the robot coordinate system.

Methods Applied and Interpretation of Parameters:

- *DC Voltage Measurement:* Calibration Factor assessed for use in DASY system by comparison with a calibrated instrument traceable to national standards. The figure given corresponds to the full scale range of the voltmeter in the respective range.
- *Connector angle:* The angle of the connector is assessed measuring the angle mechanically by a tool inserted. Uncertainty is not required.
- The report provide only calibration results for DAE, it does not contain other performance test results.



In Collaboration with

s p e a g
CALIBRATION LABORATORY

Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62304633-2512 Fax: +86-10-62304633-2504
E-mail: cttl@chinattl.com Http://www.chinattl.cn

DC Voltage Measurement

A/D - Converter Resolution nominal

High Range: 1LSB = $6.1\mu\text{V}$, full range = $-100\dots+300\text{ mV}$
Low Range: 1LSB = 61nV , full range = $-1\dots+3\text{mV}$

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

Calibration Factors	X	Y	Z
High Range	$404.110 \pm 0.15\% (k=2)$	$403.634 \pm 0.15\% (k=2)$	$404.056 \pm 0.15\% (k=2)$
Low Range	$3.95185 \pm 0.7\% (k=2)$	$3.93955 \pm 0.7\% (k=2)$	$3.90561 \pm 0.7\% (k=2)$

Connector Angle

Connector Angle to be used in DASY system	$65^\circ \pm 1^\circ$
---	------------------------



Accredited by the Swiss Accreditation Service (SAS)

The Swiss Accreditation Service is one of the signatories to the EA
 Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 0108**

Client

Sporton

Certificate No: **EX3-3819 Mar19**

CALIBRATION CERTIFICATE

Object

EX3DV4 - SN 3819

Calibration procedure(s)

QA CAL-01.v9, QA CAL-14.v5, QA CAL-23.v5, QA CAL-25.v7
Calibration procedure for dosimetric E-field probes

Calibration date:

March 1, 2019

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI).
 The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature (22 ± 3)°C and humidity < 70%.

Calibration Equipment used (M&TE critical for calibration)

Primary Standards	ID	Cal Date (Certificate No.)	Scheduled Calibration
Power meter NRP	SN: 104778	04-Apr-18 (No. 217-02672/02673)	Apr-19
Power sensor NRP-Z91	SN: 103244	04-Apr-18 (No. 217-02672)	Apr-19
Power sensor NRP-Z91	SN: 103245	04-Apr-18 (No. 217-02673)	Apr-19
Reference 20 dB Attenuator	SN: S5277 (20x)	04-Apr-18 (No. 217-02682)	Apr-19
DAE4	SN: 660	19-Dec-18 (No. DAE4-660_Dec18)	Dec-19
Reference Probe ES3DV2	SN: 3013	31-Dec-18 (No. ES3-3013_Dec18)	Dec-19
Secondary Standards	ID	Check Date (in house)	Scheduled Check
Power meter E4419B	SN: GB41293874	06-Apr-16 (in house check Jun-18)	In house check: Jun-20
Power sensor E4412A	SN: MY41498087	06-Apr-16 (in house check Jun-18)	In house check: Jun-20
Power sensor E4412A	SN: 000110210	06-Apr-16 (in house check Jun-18)	In house check: Jun-20
RF generator HP 8648C	SN: US3642U01700	04-Aug-99 (in house check Jun-18)	In house check: Jun-20
Network Analyzer E8358A	SN: US41080477	31-Mar-14 (in house check Oct-18)	In house check: Oct-19

	Name	Function	Signature
Calibrated by:	Michael Weber	Laboratory Technician	
Approved by:	Katja Pokovic	Technical Manager	
Issued: March 2, 2019			
This calibration certificate shall not be reproduced except in full without written approval of the laboratory.			



Accredited by the Swiss Accreditation Service (SAS)
 The Swiss Accreditation Service is one of the signatories to the EA
 Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 0108**

Glossary:

TSL	tissue simulating liquid
NORM _{x,y,z}	sensitivity in free space
ConvF	sensitivity in TSL / NORM _{x,y,z}
DCP	diode compression point
CF	crest factor (1/duty_cycle) of the RF signal
A, B, C, D	modulation dependent linearization parameters
Polarization ϕ	ϕ rotation around probe axis
Polarization ϑ	ϑ rotation around an axis that is in the plane normal to probe axis (at measurement center), i.e., $\vartheta = 0$ is normal to probe axis
Connector Angle	information used in DASY system to align probe sensor X to the robot coordinate system

Calibration is Performed According to the Following Standards:

- IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013
- IEC 62209-1, "Measurement procedure for the assessment of Specific Absorption Rate (SAR) from hand-held and body-mounted devices used next to the ear (frequency range of 300 MHz to 6 GHz)", July 2016
- IEC 62209-2, "Procedure to determine the Specific Absorption Rate (SAR) for wireless communication devices used in close proximity to the human body (frequency range of 30 MHz to 6 GHz)", March 2010
- KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

Methods Applied and Interpretation of Parameters:

- NORM_{x,y,z}**: Assessed for E-field polarization $\vartheta = 0$ ($f \leq 900$ MHz in TEM-cell; $f > 1800$ MHz: R22 waveguide). NORM_{x,y,z} are only intermediate values, i.e., the uncertainties of NORM_{x,y,z} does not affect the E^2 -field uncertainty inside TSL (see below ConvF).
- NORM(f)_{x,y,z}** = NORM_{x,y,z} * frequency_response (see Frequency Response Chart). This linearization is implemented in DASY4 software versions later than 4.2. The uncertainty of the frequency response is included in the stated uncertainty of ConvF.
- DCP_{x,y,z}**: DCP are numerical linearization parameters assessed based on the data of power sweep with CW signal (no uncertainty required). DCP does not depend on frequency nor media.
- PAR**: PAR is the Peak to Average Ratio that is not calibrated but determined based on the signal characteristics
- A_{x,y,z}; B_{x,y,z}; C_{x,y,z}; D_{x,y,z}; VR_{x,y,z}**: A, B, C, D are numerical linearization parameters assessed based on the data of power sweep for specific modulation signal. The parameters do not depend on frequency nor media. VR is the maximum calibration range expressed in RMS voltage across the diode.
- ConvF and Boundary Effect Parameters**: Assessed in flat phantom using E-field (or Temperature Transfer Standard for $f \leq 800$ MHz) and inside waveguide using analytical field distributions based on power measurements for $f > 800$ MHz. The same setups are used for assessment of the parameters applied for boundary compensation (alpha, depth) of which typical uncertainty values are given. These parameters are used in DASY4 software to improve probe accuracy close to the boundary. The sensitivity in TSL corresponds to NORM_{x,y,z} * ConvF whereby the uncertainty corresponds to that given for ConvF. A frequency dependent ConvF is used in DASY version 4.4 and higher which allows extending the validity from ± 50 MHz to ± 100 MHz.
- Spherical isotropy (3D deviation from isotropy)**: in a field of low gradients realized using a flat phantom exposed by a patch antenna.
- Sensor Offset**: The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No tolerance required.
- Connector Angle**: The angle is assessed using the information gained by determining the NORM_x (no uncertainty required).

DASY/EASY - Parameters of Probe: EX3DV4 - SN:3819

Basic Calibration Parameters

	Sensor X	Sensor Y	Sensor Z	Unc (k=2)
Norm ($\mu\text{V}/(\text{V}/\text{m})^2$) ^A	0.46	0.40	0.46	± 10.1 %
DCP (mV) ^B	101.7	100.6	101.3	

Calibration Results for Modulation Response

UID	Communication System Name		A dB	B dB/ μV	C	D dB	VR mV	Max dev.	Unc ^E (k=2)
0	CW	X	0.0	0.0	1.0	0.00	149.0	±3.0 %	± 4.7 %
		Y	0.0	0.0	1.0		142.6		
		Y	0.0	0.0	1.0		155.7		

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

^A The uncertainties of Norm X,Y,Z do not affect the E²-field uncertainty inside TSL (see Pages 5 and 6).

^B Numerical linearization parameter: uncertainty not required.

^E Uncertainty is determined using the max. deviation from linear response applying rectangular distribution and is expressed for the square of the field value.

DASY/EASY - Parameters of Probe: EX3DV4 - SN:3819

Other Probe Parameters

Sensor Arrangement	Triangular
Connector Angle (°)	112.8
Mechanical Surface Detection Mode	enabled
Optical Surface Detection Mode	disabled
Probe Overall Length	337 mm
Probe Body Diameter	10 mm
Tip Length	9 mm
Tip Diameter	2.5 mm
Probe Tip to Sensor X Calibration Point	1 mm
Probe Tip to Sensor Y Calibration Point	1 mm
Probe Tip to Sensor Z Calibration Point	1 mm
Recommended Measurement Distance from Surface	1.4 mm

DASY/EASY - Parameters of Probe: EX3DV4 - SN:3819

Calibration Parameter Determined in Head Tissue Simulating Media

f (MHz) ^C	Relative Permittivity ^F	Conductivity (S/m) ^F	ConvF X	ConvF Y	ConvF Z	Alpha ^G	Depth ^G (mm)	Unc (k=2)
750	41.9	0.89	10.00	10.00	10.00	0.42	1.05	± 12.0 %
835	41.5	0.90	9.57	9.57	9.57	0.55	0.89	± 12.0 %
900	41.5	0.97	9.43	9.43	9.43	0.41	1.05	± 12.0 %
1450	40.5	1.20	8.68	8.68	8.68	0.29	0.80	± 12.0 %
1750	40.1	1.37	8.54	8.54	8.54	0.40	0.89	± 12.0 %
1900	40.0	1.40	8.27	8.27	8.27	0.23	0.99	± 12.0 %
2000	40.0	1.40	8.20	8.20	8.20	0.35	0.86	± 12.0 %
2300	39.5	1.67	7.64	7.64	7.64	0.37	0.86	± 12.0 %
2450	39.2	1.80	7.21	7.21	7.21	0.34	0.92	± 12.0 %
2600	39.0	1.96	7.06	7.06	7.06	0.38	0.89	± 12.0 %
3300	38.2	2.71	6.91	6.91	6.91	0.29	1.20	± 14.0 %
3500	37.9	2.91	6.89	6.89	6.89	0.25	1.20	± 14.0 %
3700	37.7	3.12	6.67	6.67	6.67	0.25	1.25	± 14.0 %
5250	35.9	4.71	5.07	5.07	5.07	0.40	1.80	± 14.0 %
5600	35.5	5.07	4.70	4.70	4.70	0.40	1.80	± 14.0 %
5750	35.4	5.22	4.77	4.77	4.77	0.40	1.80	± 14.0 %

^C Frequency validity above 300 MHz of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to ± 50 MHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is ± 10, 25, 40, 50 and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Validity of ConvF assessed at 6 MHz is 4-9 MHz, and ConvF assessed at 13 MHz is 9-19 MHz. Above 5 GHz frequency validity can be extended to ± 110 MHz.

^F At frequencies below 3 GHz, the validity of tissue parameters (ϵ and σ) can be relaxed to ± 10% if liquid compensation formula is applied to measured SAR values. At frequencies above 3 GHz, the validity of tissue parameters (ϵ and σ) is restricted to ± 5%. The uncertainty is the RSS of the ConvF uncertainty for indicated target tissue parameters.

^G Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than ± 1% for frequencies below 3 GHz and below ± 2% for frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the boundary.

DASY/EASY - Parameters of Probe: EX3DV4 - SN:3819

Calibration Parameter Determined in Body Tissue Simulating Media


f (MHz) ^C	Relative Permittivity ^F	Conductivity (S/m) ^F	ConvF X	ConvF Y	ConvF Z	Alpha ^G	Depth ^G (mm)	Unc (k=2)
750	55.5	0.96	9.68	9.68	9.68	0.69	0.80	± 12.0 %
835	55.2	0.97	9.40	9.40	9.40	0.49	0.97	± 12.0 %
900	55.0	1.05	9.36	9.36	9.36	0.50	0.92	± 12.0 %
1750	53.4	1.49	8.06	8.06	8.06	0.33	0.85	± 12.0 %
1900	53.3	1.52	7.66	7.66	7.66	0.25	1.11	± 12.0 %
2300	52.9	1.81	7.49	7.49	7.49	0.32	0.96	± 12.0 %
2450	52.7	1.95	7.32	7.32	7.32	0.37	0.89	± 12.0 %
2600	52.5	2.16	7.04	7.04	7.04	0.34	0.95	± 12.0 %
3300	51.6	3.08	6.60	6.60	6.60	0.28	1.20	± 14.0 %
3500	51.3	3.31	6.57	6.57	6.57	0.25	1.20	± 14.0 %
3700	51.0	3.55	6.37	6.37	6.37	0.30	1.25	± 14.0 %
5250	48.9	5.36	4.46	4.46	4.46	0.50	1.90	± 14.0 %
5600	48.5	5.77	3.92	3.92	3.92	0.50	1.90	± 14.0 %
5750	48.3	5.94	4.07	4.07	4.07	0.50	1.90	± 14.0 %

^C Frequency validity above 300 MHz of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to ± 50 MHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is ± 10, 25, 40, 50 and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Validity of ConvF assessed at 6 MHz is 4-9 MHz, and ConvF assessed at 13 MHz is 9-19 MHz. Above 5 GHz frequency validity can be extended to ± 110 MHz.

^F At frequencies below 3 GHz, the validity of tissue parameters (ϵ and σ) can be relaxed to ± 10% if liquid compensation formula is applied to measured SAR values. At frequencies above 3 GHz, the validity of tissue parameters (ϵ and σ) is restricted to ± 5%. The uncertainty is the RSS of the ConvF uncertainty for indicated target tissue parameters.

^G Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than ± 1% for frequencies below 3 GHz and below ± 2% for frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the boundary.


Frequency Response of E-Field
(TEM-Cell:ifi110 EXX, Waveguide: R22)





Uncertainty of Frequency Response of E-field: $\pm 6.3\%$ (k=2)

Receiving Pattern (ϕ), $\vartheta = 0^\circ$

f=600 MHz, TEM





f=1800 MHz, R22




Uncertainty of Axial Isotropy Assessment: $\pm 0.5\%$ ($k=2$)

Dynamic Range $f(\text{SAR}_{\text{head}})$
(TEM cell , $f_{\text{eval}} = 1900 \text{ MHz}$)




Uncertainty of Linearity Assessment: $\pm 0.6\%$ ($k=2$)

Conversion Factor Assessment



Deviation from Isotropy in Liquid Error (ϕ, θ), f = 900 MHz





Appendix E. Conducted RF Output Power Table

The detailed power table are shown as follows.

UMTS/LTE Full Power

Band		WCDMA II			Turn-up Limit (dBm)	WCDMA V			Turn-up Limit (dBm)
TX Channel		9262	9400	9538		4132	4182	4233	
Rx Channel		9662	9800	9938		4357	4407	4458	
Frequency (MHz)		1852.4	1880	1907.6		826.4	836.4	846.6	
3GPP Rel 9a	RM-12.2Kbps	22.92	22.86	22.89	23.50	23.28	23.23	23.02	23.50
3GPP Rel 6	HSDPA Subtest-1	21.82	21.97	21.92	22.50	22.28	22.25	22.04	22.50
3GPP Rel 6	HSDPA Subtest-2	21.94	22.11	22.04	22.50	22.46	22.41	22.13	22.50
3GPP Rel 6	HSDPA Subtest-3	21.56	21.57	21.60	22.00	21.90	21.85	21.70	22.00
3GPP Rel 6	HSDPA Subtest-4	21.55	21.57	21.59	22.00	21.90	21.85	21.70	22.00
3GPP Rel 8	DC-HSDPA Subtest-1	21.73	21.96	21.87	22.50	22.13	22.23	21.88	22.50
3GPP Rel 8	DC-HSDPA Subtest-2	21.88	21.91	21.94	22.50	22.37	22.34	22.06	22.50
3GPP Rel 8	DC-HSDPA Subtest-3	21.48	21.49	21.45	22.00	21.73	21.76	21.65	22.00
3GPP Rel 8	DC-HSDPA Subtest-4	21.49	21.40	21.54	22.00	21.83	21.85	21.55	22.00
3GPP Rel 6	HSUPA Subtest-1	21.07	21.23	21.75	22.50	21.91	21.96	21.87	22.50
3GPP Rel 6	HSUPA Subtest-2	20.66	20.78	20.39	21.50	20.83	21.21	20.62	21.50
3GPP Rel 6	HSUPA Subtest-3	20.42	20.56	20.56	21.00	20.90	20.48	20.70	21.50
3GPP Rel 6	HSUPA Subtest-4	20.67	21.42	21.08	21.50	21.42	21.46	20.96	21.50
3GPP Rel 6	HSUPA Subtest-5	21.70	21.60	21.90	22.50	22.20	22.10	22.00	22.50

Band 2 (1900MHz Band)									
Part 24E									
BW [MHz]	Modulation	RB Size	RB Offset	Power Low Ch. / Freq.	Power Middle Ch. / Freq.	Power High Ch. / Freq.	Tune-up limit (dBm)	MPR (dB)	
Channel				18700	18900	19100			
Frequency (MHz)				1860	1880	1900			
20	QPSK	1	0	22.56	22.59	22.92	23.5	0	
20	QPSK	1	49	22.48	22.58	22.93			
20	QPSK	1	99	22.30	22.37	22.51			
20	QPSK	50	0	21.67	21.64	21.68	22.5	1	
20	QPSK	50	24	21.56	21.60	21.49			
20	QPSK	50	50	21.57	21.59	21.54			
20	QPSK	100	0	21.67	21.41	21.68	22.5	1	
20	16QAM	1	0	21.50	21.57	21.52			
20	16QAM	1	49	21.45	21.43	21.42			
20	16QAM	1	99	21.28	21.36	21.37	21.5	2	
20	16QAM	50	0	20.70	20.49	20.72			
20	16QAM	50	24	20.60	20.55	20.58			
20	16QAM	50	50	20.57	20.62	20.61	21.5	2	
20	16QAM	100	0	20.65	20.47	20.60			
Channel				18675	18900	19125	Tune-up limit (dBm)	MPR (dB)	
Frequency (MHz)				1857.5	1880	1902.5			
15	QPSK	1	0	22.57	22.45	22.63	23.5	0	
15	QPSK	1	37	22.91	22.57	22.75			
15	QPSK	1	74	22.64	22.70	22.64			
15	QPSK	36	0	21.75	21.52	21.57	22.5	1	
15	QPSK	36	20	21.73	21.46	21.51			
15	QPSK	36	39	21.60	21.47	21.56			
15	QPSK	75	0	21.73	21.51	21.55	22.5	1	
15	16QAM	1	0	21.47	21.43	21.50			
15	16QAM	1	37	21.68	21.21	21.42			
15	16QAM	1	74	21.41	21.38	21.37	21.5	2	
15	16QAM	36	0	20.68	20.48	20.62			
15	16QAM	36	20	20.70	20.59	20.62			
15	16QAM	36	39	20.70	20.51	20.63	21.5	2	
15	16QAM	75	0	20.93	20.54	20.68			
Channel				18650	18900	19150	Tune-up limit (dBm)	MPR (dB)	
Frequency (MHz)				1855	1880	1905			
10	QPSK	1	0	22.45	22.42	22.35	23.5	0	
10	QPSK	1	25	22.83	22.46	22.66			
10	QPSK	1	49	22.48	22.61	22.45			
10	QPSK	25	0	21.53	21.51	21.60	22.5	1	
10	QPSK	25	12	21.66	21.49	21.53			
10	QPSK	25	25	21.61	21.64	21.56			
10	QPSK	50	0	21.60	21.59	21.55	22.5	1	
10	16QAM	1	0	21.46	21.46	21.34			
10	16QAM	1	25	21.61	21.46	21.61			
10	16QAM	1	49	21.42	21.79	21.44	21.5	2	
10	16QAM	25	0	20.66	20.51	20.66			
10	16QAM	25	12	20.70	20.70	20.59			
10	16QAM	25	25	20.70	20.77	20.69	21.5	2	
10	16QAM	50	0	20.74	20.61	20.70			
Channel				18625	18900	19175	Tune-up limit (dBm)	MPR (dB)	
Frequency (MHz)				1852.5	1880	1907.5			
5	QPSK	1	0	22.47	22.42	22.48	23.5	0	
5	QPSK	1	12	22.58	22.53	22.63			
5	QPSK	1	24	22.60	22.43	22.38			
5	QPSK	12	0	21.46	21.48	21.56	22.5	1	
5	QPSK	12	7	21.58	21.50	21.51			
5	QPSK	12	13	21.57	21.52	21.48			
5	QPSK	25	0	21.55	21.57	21.56	22.5	1	
5	16QAM	1	0	21.68	21.28	21.38			
5	16QAM	1	12	21.43	21.23	21.16			
5	16QAM	1	24	21.29	21.20	21.26	21.5	2	
5	16QAM	12	0	20.51	20.44	20.62			
5	16QAM	12	7	20.61	20.57	20.55			
5	16QAM	12	13	20.64	20.54	20.47	21.5	2	
5	16QAM	25	0	20.59	20.53	20.72			
Channel				18615	18900	19185	Tune-up limit (dBm)	MPR (dB)	
Frequency (MHz)				1851.5	1880	1908.5			
3	QPSK	1	0	22.49	22.47	22.54	23.5	0	
3	QPSK	1	8	22.45	22.47	22.28			
3	QPSK	1	14	22.52	22.56	22.32			
3	QPSK	8	0	21.67	21.67	21.75	22.5	1	
3	QPSK	8	4	21.72	21.61	21.62			
3	QPSK	8	7	21.77	21.59	21.53			
3	QPSK	15	0	21.75	21.67	21.58	22.5	1	
3	16QAM	1	0	21.12	21.47	21.31			
3	16QAM	1	8	21.37	21.30	21.26			
3	16QAM	1	14	21.45	21.17	21.35	21.5	2	
3	16QAM	8	0	20.69	20.58	20.68			
3	16QAM	8	4	20.78	20.71	20.65			
3	16QAM	8	7	20.75	20.86	20.54	21.5	2	
3	16QAM	15	0	20.75	20.71	20.73			
Channel				18607	18900	19193	Tune-up limit (dBm)	MPR (dB)	
Frequency (MHz)				1850.7	1880	1909.3			
1.4	QPSK	1	0	22.31	22.53	22.60	23.5	0	
1.4	QPSK	1	3	22.49	22.68	22.66			
1.4	QPSK	1	5	22.64	22.61	22.61			
1.4	QPSK	3	0	22.63	22.66	22.74	22.5	1	
1.4	QPSK	3	1	22.78	22.84	22.64			
1.4	QPSK	3	3	22.73	22.77	22.65			
1.4	QPSK	6	0	21.74	21.72	21.64	22.5	1	
1.4	16QAM	1	0	21.77	21.78	21.47			
1.4	16QAM	1	3	21.90	21.85	21.83			
1.4	16QAM	1	5	21.60	21.46	21.65	22.5	1	
1.4	16QAM	3	0	21.73	21.57	21.84			
1.4	16QAM	3	1	21.80	21.69	21.82			
1.4	16QAM	3	3	21.82	21.57	21.67	21.5	2	
1.4	16QAM	6	0	20.70	20.69	20.64			

Band 4 (AWS Band)									
Part 27L (only on channel required)									
BW [MHz]	Modulation	RB Size	RB Offset	Power Low Ch. / Freq.	Power Middle Ch. / Freq.	Power High Ch. / Freq.	Tune-up limit (dBm)	MPR (dB)	
Channel				20050	20175	20300			
Frequency (MHz)				1720	1732.5	1745			
20	QPSK	1	0	22.04	22.28	22.95	23.5	0	
20	QPSK	1	49	22.34	22.26	22.47			
20	QPSK	1	99	22.31	22.11	22.34			
20	QPSK	50	0	21.47	21.49	21.65	22.5	1	
20	QPSK	50	24	21.43	21.42	21.55			
20	QPSK	50	50	21.43	21.28	21.54			
20	QPSK	100	0	21.42	21.49	21.68	22.5	1	
20	16QAM	1	0	21.17	21.21	21.45			
20	16QAM	1	49	21.31	21.44	21.42			
20	16QAM	1	99	21.22	21.31	21.29	21.5	2	
20	16QAM	50	0	20.35	20.47	20.65			
20	16QAM	50	24	20.57	20.49	20.64			
20	16QAM	50	50	20.47	20.27	20.64	21.5	2	
20	16QAM	100	0	20.45	20.42	20.61			
Channel				20025	20175	20325	Tune-up limit (dBm)	MPR (dB)	
Frequency (MHz)				1717.5	1732.5	1747.5			
15	QPSK	1	0	22.23	22.39	22.71	23.5	0	
15	QPSK	1	37	22.32	22.58	22.60			
15	QPSK	1	74	22.35	22.33	22.53			
15	QPSK	36	0	21.48	21.58	21.78	22.5	1	
15	QPSK	36	20	21.39	21.44	21.62			
15	QPSK	36	39	21.58	21.39	21.61			
15	QPSK	75	0	21.49	21.52	21.76	22.5	1	
15	16QAM	1	0	21.32	21.35	21.76			
15	16QAM	1	37	21.37	21.42	21.36			
15	16QAM	1	74	21.42	20.87	21.52	21.5	2	
15	16QAM	36	0	20.56	20.58	20.87			
15	16QAM	36	20	20.51	20.46	20.58			
15	16QAM	36	39	20.55	20.43	20.67	21.5	2	
15	16QAM	75	0	20.57	20.53	20.73			
Channel				20000	20175	20350	Tune-up limit (dBm)	MPR (dB)	
Frequency (MHz)				1715	1732.5	1750			
10	QPSK	1	0	22.16	22.36	22.64	23.5	0	
10	QPSK	1	25	22.41	22.55	22.92			
10	QPSK	1	49	22.65	22.33	22.52			
10	QPSK	25	0	21.47	21.54	21.82	22.5	1	
10	QPSK	25	12	21.39	21.41	21.71			
10	QPSK	25	25	21.44	21.44	21.78			
10	QPSK	50	0	21.55	21.54	21.82	22.5	1	
10	16QAM	1	0	21.23	21.28	21.64			
10	16QAM	1	25	21.29	21.79	21.63			
10	16QAM	1	49	21.33	21.57	21.54	21.5	2	
10	16QAM	25	0	20.67	20.58	20.96			
10	16QAM	25	12	20.49	20.52	20.77			
10	16QAM	25	25	20.46	20.48	20.84	21.5	2	
10	16QAM	50	0	20.57	20.58	20.91			
Channel				19975	20175	20375	Tune-up limit (dBm)	MPR (dB)	
Frequency (MHz)				1712.5	1732.5	1752.5			
5	QPSK	1	0	22.21	22.24	22.65	23.5	0	
5	QPSK	1	12	22.41	22.56	22.81			
5	QPSK	1	24	22.48	22.31	22.76			
5	QPSK	12	0	21.44	21.61	21.84	22.5	1	
5	QPSK	12	7	21.25	21.50	21.60			
5	QPSK	12	13	21.40	21.47	21.83			
5	QPSK	25	0	21.40	21.58	21.85	22.5	1	
5	16QAM	1	0	21.13	21.70	21.56			
5	16QAM	1	12	20.76	21.36	21.41			
5	16QAM	1	24	21.15	21.23	21.50	21.5	2	
5	16QAM	12	0	20.48	20.65	21.00			
5	16QAM	12	7	20.38	20.56	20.50			
5	16QAM	12	13	20.55	20.48	20.85	21.5	2	
5	16QAM	25	0	20.37	20.54	20.96			
Channel				19965	20175	20385	Tune-up limit (dBm)	MPR (dB)	
Frequency (MHz)				1711.5	1732.5	1753.5			
3	QPSK	1	0	22.29	22.56	22.48	23.5	0	
3	QPSK	1	8	22.17	22.40	22.24			
3	QPSK	1	14	22.36	22.48	22.54			
3	QPSK	8	0	21.28	21.42	21.48	22.5	1	
3	QPSK	8	4	21.38	21.48	21.57			
3	QPSK	8	7	21.45	21.40	21.61			
3	QPSK	15	0	21.31	21.44	21.61	22.5	1	
3	16QAM	1	0	21.04	21.19	21.36			
3	16QAM	1	8	20.97	21.22	21.22			
3	16QAM	1	14	20.74	21.38	21.29	21.5	2	
3	16QAM	8	0	20.41	20.57	20.84			
3	16QAM	8	4	20.25	20.62	20.72			
3	16QAM	8	7	20.31	20.53	20.69	21.5	2	
3	16QAM	15	0	20.34	20.55	20.66			
Channel				19957	20175	20393	Tune-up limit (dBm)	MPR (dB)	
Frequency (MHz)				1710.7	1732.5	1754.3			
1.4	QPSK	1	0	22.11	22.59	22.51	23.5	0	
1.4	QPSK	1	3	22.34	22.50	22.54			
1.4	QPSK	1	5	22.15	22.30	22.48			
1.4	QPSK	3	0	22.30	22.47	22.57	22.5	1	
1.4	QPSK	3	1	22.49	22.52	22.60			
1.4	QPSK	3	3	22.41	22.42	22.56			
1.4	QPSK	6	0	21.34	21.54	21.62	22.5	1	
1.4	16QAM	1	0	21.38	21.62	21.75			
1.4	16QAM	1	3	21.60	21.79	21.90			
1.4	16QAM	1	5	21.08	21.58	21.76	22.5	1	
1.4	16QAM	3	0	21.18	21.46	21.56			
1.4	16QAM	3	1	21.38	21.49	21.59			
1.4	16QAM	3	3	21.38	21.47	21.55	21.5	2	
1.4	16QAM	6	0	20.36	20.48	20.66			

Band 12 (700MHz Low Band) Part 27F(only on channel required)									
BW [MHz]	Modulation	RB Size	RB Offset	Power Low Ch. / Freq.	Power Middle Ch. / Freq.	Power High Ch. / Freq.	Tune-up limit (dBm)	MPR (dB)	
Channel				23060	23095	23130	24	0	
Frequency (MHz)				704	707.5	711			
10	QPSK	1	0	22.76	23.06	22.99	24	0	
10	QPSK	1	25	23.42	23.38	23.63			
10	QPSK	1	49	22.86	23.16	23.24			
10	QPSK	25	0	22.16	22.25	22.32			
10	QPSK	25	12	22.30	22.15	22.35	23	1	
10	QPSK	25	25	22.16	22.27	22.37			
10	QPSK	50	0	22.30	22.25	22.34			
10	16QAM	1	0	21.72	21.78	21.90			
10	16QAM	1	25	22.27	22.37	22.32	23	1	
10	16QAM	1	49	21.95	22.46	22.02			
10	16QAM	25	0	21.21	21.03	21.28			
10	16QAM	25	12	21.40	21.02	21.30			
10	16QAM	25	25	21.11	21.22	21.30	22	2	
10	16QAM	50	0	21.24	21.20	21.33			
Channel				23035	23095	23155	Tune-up limit (dBm)	MPR (dB)	
Frequency (MHz)				701.5	707.5	713.5	24	0	
5	QPSK	1	0	23.00	23.12	23.01			
5	QPSK	1	12	23.52	23.27	23.19	23	1	
5	QPSK	1	24	23.20	23.09	23.24			
5	QPSK	12	0	22.20	22.36	22.45			
5	QPSK	12	7	22.34	22.36	22.45			
5	QPSK	12	13	22.30	22.29	22.35	23	1	
5	QPSK	25	0	22.30	22.33	22.39			
5	16QAM	1	0	22.04	21.93	22.33			
5	16QAM	1	12	22.10	21.84	22.65			
5	16QAM	1	24	22.23	22.00	22.20	23	1	
5	16QAM	12	0	21.30	21.20	21.47			
5	16QAM	12	7	21.34	21.19	21.51			
5	16QAM	12	13	21.40	21.30	21.47			
5	16QAM	25	0	21.37	21.33	21.33	22	2	
5	16QAM	25	0	21.37	21.33	21.33			
Channel				23025	23095	23165	Tune-up limit (dBm)	MPR (dB)	
Frequency (MHz)				700.5	707.5	714.5	24	0	
3	QPSK	1	0	23.25	23.13	23.23			
3	QPSK	1	8	23.11	23.10	23.28	23	1	
3	QPSK	1	14	23.21	23.13	23.57			
3	QPSK	8	0	22.34	22.43	22.37			
3	QPSK	8	4	22.35	22.34	22.37			
3	QPSK	8	7	22.30	22.29	22.37	23	1	
3	QPSK	15	0	22.27	22.32	22.41			
3	16QAM	1	0	22.07	21.95	22.12			
3	16QAM	1	8	22.02	22.16	22.18			
3	16QAM	1	14	22.35	22.17	22.37	22	2	
3	16QAM	8	0	21.30	21.30	21.48			
3	16QAM	8	4	21.22	21.13	21.32			
3	16QAM	8	7	21.45	21.26	21.43			
3	16QAM	15	0	21.28	21.25	21.38	Tune-up limit (dBm)	MPR (dB)	
Channel				23017	23095	23173	24	0	
Frequency (MHz)				699.7	707.5	715.3			
1.4	QPSK	1	0	22.92	23.20	23.28	24	0	
1.4	QPSK	1	3	23.03	23.22	23.34			
1.4	QPSK	1	5	22.97	23.19	23.46			
1.4	QPSK	3	0	23.13	23.26	23.37			
1.4	QPSK	3	1	23.25	23.37	23.54	23	1	
1.4	QPSK	3	3	23.31	23.39	23.61			
1.4	QPSK	6	0	22.21	22.29	22.40			
1.4	16QAM	1	0	22.11	22.12	22.26			
1.4	16QAM	1	3	22.24	22.15	22.44	23	1	
1.4	16QAM	1	5	22.20	22.07	22.21			
1.4	16QAM	3	0	22.22	22.20	22.36			
1.4	16QAM	3	1	22.30	22.30	22.38			
1.4	16QAM	3	3	22.26	22.29	22.51	22	2	
1.4	16QAM	6	0	21.13	21.01	21.43			

Band 13(700MHz Band) Part 27F									
BW [MHz]	Modulation	RB Size	RB Offset	Power Low Ch. / Freq.	Power Middle Ch. / Freq.	Power High Ch. / Freq.	Tune-up limit (dBm)	MPR (dB)	
Channel				23230			24	0	
Frequency (MHz)				782					
10	QPSK	1	0		23.34		24	0	
10	QPSK	1	25		23.73				
10	QPSK	1	49		23.51				
10	QPSK	25	0		22.73				
10	QPSK	25	12		22.48		23	1	
10	QPSK	25	25		22.49				
10	QPSK	50	0		22.54				
10	16QAM	1	0		21.69				
10	16QAM	1	25		22.51		23	1	
10	16QAM	1	49		22.27				
10	16QAM	25	0		21.80				
10	16QAM	25	12		21.62				
10	16QAM	25	25		21.56		22	2	
10	16QAM	50	0		21.50				
Channel				23205	23230	23255	Tune-up limit (dBm)	MPR (dB)	
Frequency (MHz)				779.5	782	784.5	24	0	
5	QPSK	1	0	23.29	23.72	23.47			
5	QPSK	1	12	23.72	23.53	23.70	23	1	
5	QPSK	1	24	23.55	23.25	23.58			
5	QPSK	12	0	22.79	22.73	22.64			
5	QPSK	12	7	22.82	22.58	22.71			
5	QPSK	12	13	22.81	22.60	22.82	23	1	
5	QPSK	25	0	22.72	22.67	22.72			
5	16QAM	1	0	21.98	22.59	22.46			
5	16QAM	1	12	22.66	22.36	22.82			
5	16QAM	1	24	22.37	22.21	22.55	22	2	
5	16QAM	12	0	21.67	21.83	21.50			
5	16QAM	12	7	21.79	21.49	21.83			
5	16QAM	12	13	21.80	21.39	21.88			
5	16QAM	25	0	21.81	21.61	21.71	Tune-up limit (dBm)	MPR (dB)	

Band 17 (700MHz Band) Part 27H(only on channel required)									
BW [MHz]	Modulation	RB Size	RB Offset	Power Low Ch. / Freq.	Power Middle Ch. / Freq.	Power High Ch. / Freq.	Tune-up limit (dBm)	MPR (dB)	
Channel				23780	23790	23800	24	0	
Frequency (MHz)				709	710	711			
10	QPSK	1	0	23.62	23.27	23.18	24	0	
10	QPSK	1	25	23.57	23.58	23.35			
10	QPSK	1	49	23.44	23.52	23.51			
10	QPSK	25	0	22.58	22.45	22.48			
10	QPSK	25	12	22.50	22.45	22.51	23	1	
10	QPSK	25	25	22.56	22.51	22.55			
10	QPSK	50	0	22.54	22.51	22.53			
10	16QAM	1	0	22.31	22.21	22.24			
10	16QAM	1	25	22.46	22.34	22.38	23	1	
10	16QAM	1	49	22.41	22.18	22.32			
10	16QAM	25	0	21.72	21.34	21.63			
10	16QAM	25	12	21.53	21.50	21.56			
10	16QAM	25	25	21.54	21.48	21.42	22	2	
10	16QAM	50	0	21.52	21.37	21.49			
Channel				23755	23790	23825	Tune-up limit (dBm)	MPR (dB)	
Frequency (MHz)				706.5	710	713.5	24	0	
5	QPSK	1	0	23.38	23.28	23.25			
5	QPSK	1	12	23.44	23.24	23.28	23	1	
5	QPSK	1	24	23.24	23.46	23.50			
5	QPSK	12	0	22.52	22.41	22.39			
5	QPSK	12	7	22.50	22.44	22.32			
5	QPSK	12	13	22.53	22.38	22.52	23	1	
5	QPSK	25	0	22.52	22.51	22.47			
5	16QAM	1	0	22.23	22.20	22.34	23	1	
5	16QAM	1	12	22.07	22.25	22.21			
5	16QAM	1	24	22.20	22.09	22.26			
5	16QAM	12	0	21.47	21.15	21.15			
5	16QAM	12	7	21.32	21.39	21.19	22	2	
5	16QAM	12	13	21.40	21.40	21.37			
5	16QAM	25	0	21.58	21.42	21.20			

Band 25 (1900MHz Band)										
Part 24E										
BW [MHz]	Modulation	RB Size	RB Offset	Power Low Ch. / Freq.	Power Middle Ch. / Freq.	Power High Ch. / Freq.	Tune-up limit (dBm)	MPR (dB)		
Channel				26140	26340	26590				
Frequency (MHz)				1860	1880	1905				
20	QPSK	1	0	22.52	22.54	22.53	23.5	0		
20	QPSK	1	49	22.91	22.63	22.67				
20	QPSK	1	99	22.35	22.57	21.92				
20	QPSK	50	0	21.84	21.69	21.55	22.5	1		
20	QPSK	50	24	21.62	21.56	21.48				
20	QPSK	50	50	21.60	21.51	21.44				
20	QPSK	100	0	21.63	21.54	21.45	22.5	1		
20	16QAM	1	0	21.31	21.41	21.32				
20	16QAM	1	49	21.53	21.42	21.36				
20	16QAM	1	99	21.27	21.34	20.90	21.5	2		
20	16QAM	50	0	20.89	20.64	20.65				
20	16QAM	50	24	20.77	20.67	20.67				
20	16QAM	50	50	20.70	20.69	20.64	21.5	2		
20	16QAM	100	0	20.71	20.56	20.43				
Channel				26115	26340	26615	Tune-up limit (dBm)	MPR (dB)		
Frequency (MHz)				1857.5	1880	1907.5				
15	QPSK	1	0	22.62	22.57	22.34	23.5	0		
15	QPSK	1	37	22.81	22.54	22.78				
15	QPSK	1	74	22.70	22.69	21.91				
15	QPSK	36	0	21.80	21.63	21.59	22.5	1		
15	QPSK	36	20	21.66	21.55	21.40				
15	QPSK	36	39	21.66	21.58	21.39				
15	QPSK	75	0	21.59	21.51	21.44	22.5	1		
15	16QAM	1	0	21.64	21.48	21.39				
15	16QAM	1	37	21.68	21.60	21.34				
15	16QAM	1	74	21.49	21.31	20.85	21.5	2		
15	16QAM	36	0	20.77	20.74	20.64				
15	16QAM	36	20	20.75	20.66	20.53				
15	16QAM	36	39	20.73	20.67	20.42	21.5	2		
15	16QAM	75	0	20.75	20.62	20.65				
Channel				26090	26340	26640	Tune-up limit (dBm)	MPR (dB)		
Frequency (MHz)				1855	1880	1910				
10	QPSK	1	0	22.48	22.58	22.41	23.5	0		
10	QPSK	1	25	22.89	22.45	22.33				
10	QPSK	1	49	22.69	22.56	21.96				
10	QPSK	25	0	21.71	21.58	21.44	22.5	1		
10	QPSK	25	12	21.75	21.49	21.38				
10	QPSK	25	25	21.64	21.55	21.39				
10	QPSK	50	0	21.73	21.54	21.41	22.5	1		
10	16QAM	1	0	21.61	21.67	21.40				
10	16QAM	1	25	21.73	21.77	21.46				
10	16QAM	1	49	21.58	21.46	20.97	21.5	2		
10	16QAM	25	0	20.85	20.71	20.58				
10	16QAM	25	12	21.04	20.62	20.52				
10	16QAM	25	25	20.76	20.75	20.50	21.5	2		
10	16QAM	50	0	20.94	20.58	20.44				
Channel				26065	26340	26665	Tune-up limit (dBm)	MPR (dB)		
Frequency (MHz)				1852.5	1880	1912.5				
5	QPSK	1	0	22.48	22.58	22.54	23.5	0		
5	QPSK	1	12	22.64	22.72	22.53				
5	QPSK	1	24	22.76	22.64	21.93				
5	QPSK	12	0	21.58	21.53	21.39	22.5	1		
5	QPSK	12	7	21.68	21.52	21.43				
5	QPSK	12	13	21.71	21.57	21.65				
5	QPSK	25	0	21.68	21.48	21.47	22.5	1		
5	16QAM	1	0	21.79	21.31	21.22				
5	16QAM	1	12	21.83	21.36	21.21				
5	16QAM	1	24	21.45	21.25	21.25	21.5	2		
5	16QAM	12	0	20.64	20.44	20.46				
5	16QAM	12	7	20.64	20.44	20.33				
5	16QAM	12	13	20.82	20.47	20.33	21.5	2		
5	16QAM	25	0	20.75	20.62	20.38				
Channel				26055	26340	26675	Tune-up limit (dBm)	MPR (dB)		
Frequency (MHz)				1851.5	1880	1913.5				
3	QPSK	1	0	22.63	22.62	22.69	23.5	0		
3	QPSK	1	8	22.76	22.56	22.25				
3	QPSK	1	14	22.76	22.68	21.93				
3	QPSK	8	0	21.69	21.56	21.69	22.5	1		
3	QPSK	8	4	21.78	21.67	21.50				
3	QPSK	8	7	21.85	21.65	21.65				
3	QPSK	15	0	21.75	21.55	21.51	22.5	1		
3	16QAM	1	0	21.73	21.40	21.66				
3	16QAM	1	8	21.44	21.35	21.41				
3	16QAM	1	14	21.57	21.44	21.20	21.5	2		
3	16QAM	8	0	20.79	20.71	20.43				
3	16QAM	8	4	20.80	20.70	20.62				
3	16QAM	8	7	20.88	20.68	20.43	21.5	2		
3	16QAM	15	0	20.72	20.62	20.61				
Channel				26047	26340	26683	Tune-up limit (dBm)	MPR (dB)		
Frequency (MHz)				1850.7	1880	1914.3				
1.4	QPSK	1	0	22.38	22.37	22.48	23.5	0		
1.4	QPSK	1	3	22.52	22.53	22.48				
1.4	QPSK	1	5	22.47	22.46	22.41				
1.4	QPSK	3	0	22.70	22.67	22.54	22.5	1		
1.4	QPSK	3	1	22.82	22.71	22.52				
1.4	QPSK	3	3	22.78	22.76	22.40				
1.4	QPSK	6	0	21.66	21.55	21.22	22.5	1		
1.4	16QAM	1	0	21.46	21.34	21.34				
1.4	16QAM	1	3	21.61	21.58	21.60				
1.4	16QAM	1	5	21.52	21.28	21.35	22.5	1		
1.4	16QAM	3	0	21.59	21.53	21.51				
1.4	16QAM	3	1	21.89	21.66	21.53				
1.4	16QAM	3	3	21.91	21.55	21.52	21.5	2		
1.4	16QAM	6	0	20.79	20.68	20.47				

Band 26 for FCC									
(only on channel required)									
BW [MHz]	Modulation	RB Size	RB Offset	Power Low Ch. / Freq.	Power Middle Ch. / Freq.	Power High Ch. / Freq.	Tune-up limit (dbm)	MPR (dB)	
Channel				26765	26865	26965			
Frequency (MHz)				821.5	831.5	841.5			
15	QPSK	1	0	23.07	23.17	23.09	24	0	
15	QPSK	1	37	23.19	23.48	23.15			
15	QPSK	1	74	23.15	23.17	22.55			
15	QPSK	36	0	22.34	22.38	22.35	23	1	
15	QPSK	36	20	22.33	22.30	22.31			
15	QPSK	36	39	22.25	22.29	22.31			
15	QPSK	75	0	22.31	22.37	22.34	23	1	
15	16QAM	1	0	22.12	22.26	22.23			
15	16QAM	1	37	22.00	22.16	22.18			
15	16QAM	1	74	22.12	22.27	21.79	22	2	
15	16QAM	36	0	21.28	21.35	21.32			
15	16QAM	36	20	21.31	21.36	21.39			
15	16QAM	36	39	21.32	21.33	21.23	22	2	
15	16QAM	75	0	21.39	21.41	21.36			
Channel				26740	26865	26990	Tune-up limit (dbm)	MPR (dB)	
Frequency (MHz)				819	831.5	844			
10	QPSK	1	0	23.10	23.01	23.20	24	0	
10	QPSK	1	25	23.23	23.38	23.21			
10	QPSK	1	49	23.02	23.06	22.43			
10	QPSK	25	0	22.41	22.38	22.38	23	1	
10	QPSK	25	12	22.36	22.42	22.33			
10	QPSK	25	25	22.38	22.39	22.29			
10	QPSK	50	0	22.48	22.40	22.43	23	1	
10	16QAM	1	0	22.23	22.27	22.27			
10	16QAM	1	25	22.31	22.33	22.39			
10	16QAM	1	49	22.19	22.26	21.56	22	2	
10	16QAM	25	0	21.50	21.41	21.46			
10	16QAM	25	12	21.54	21.62	21.45			
10	16QAM	25	25	21.45	21.44	21.35	22	2	
10	16QAM	50	0	21.60	21.51	21.47			
Channel				26715	26865	27015	Tune-up limit (dbm)	MPR (dB)	
Frequency (MHz)				816.5	831.5	846.5			
5	QPSK	1	0	23.10	22.92	23.16	24	0	
5	QPSK	1	12	23.26	23.31	23.12			
5	QPSK	1	24	22.97	23.28	22.50			
5	QPSK	12	0	22.40	22.44	22.33	23	1	
5	QPSK	12	7	22.37	22.41	22.25			
5	QPSK	12	13	22.29	22.23	22.24			
5	QPSK	25	0	22.38	22.34	22.38	23	1	
5	16QAM	1	0	22.17	22.12	22.12			
5	16QAM	1	12	22.13	22.17	21.92			
5	16QAM	1	24	22.04	22.14	21.90	22	2	
5	16QAM	12	0	21.42	21.33	21.53			
5	16QAM	12	7	21.38	21.44	21.37			
5	16QAM	12	13	21.30	21.39	21.05	22	2	
5	16QAM	25	0	21.43	21.37	21.37			
Channel				26705	26865	27025	Tune-up limit (dbm)	MPR (dB)	
Frequency (MHz)				815.5	831.5	847.5			
3	QPSK	1	0	23.30	23.32	23.12	24	0	
3	QPSK	1	8	23.13	23.18	22.60			
3	QPSK	1	14	23.16	23.20	22.60			
3	QPSK	8	0	22.43	22.39	22.47	23	1	
3	QPSK	8	4	22.43	22.38	22.33			
3	QPSK	8	7	22.31	22.38	22.31			
3	QPSK	15	0	22.39	22.35	22.35	23	1	
3	16QAM	1	0	22.21	22.26	22.13			
3	16QAM	1	8	22.13	22.06	22.09			
3	16QAM	1	14	22.22	22.09	21.95	22	2	
3	16QAM	8	0	21.42	21.55	21.49			
3	16QAM	8	4	21.24	21.55	21.39			
3	16QAM	8	7	21.46	21.54	21.34	22	2	
3	16QAM	15	0	21.42	21.48	21.35			
Channel				26697	26865	27033	Tune-up limit (dbm)	MPR (dB)	
Frequency (MHz)				814.7	831.5	848.3			
1.4	QPSK	1	0	23.12	23.22	22.92	24	0	
1.4	QPSK	1	3	23.20	23.37	22.82			
1.4	QPSK	1	5	23.16	23.17	22.70			
1.4	QPSK	3	0	23.34	23.47	22.84	23	1	
1.4	QPSK	3	1	23.38	23.34	22.84			
1.4	QPSK	3	3	23.30	23.32	22.74			
1.4	QPSK	6	0	22.37	22.43	22.37	23	1	
1.4	16QAM	1	0	22.21	22.50	22.24			
1.4	16QAM	1	3	22.37	22.60	22.20			
1.4	16QAM	1	5	22.46	22.47	22.08	22	2	
1.4	16QAM	3	0	22.49	22.43	22.05			
1.4	16QAM	3	1	22.55	22.38	22.05			
1.4	16QAM	3	3	22.61	22.27	21.94	22	2	
1.4	16QAM	6	0	21.51	21.38	21.52			

WLAN Full Power

	Mode	Channel	Frequency (MHz)	Average power (dBm)	Tune-Up Limit	Duty Cycle %
2.4GHz WLAN	802.11b 1Mbps	1	2412	19.40	20.00	98.96
		6	2437	19.50	20.00	
		11	2462	19.40	20.00	
	802.11g 6Mbps	1	2412	14.50	16.00	94.68
		6	2437	14.80	16.00	
		11	2462	14.80	16.00	
	802.11n-HT20 MCS0	1	2412	14.00	16.00	91.38
		6	2437	14.00	16.00	
		11	2462	14.30	16.00	
	802.11n-HT40 MCS0	3	2422	11.60	13.00	90.35
		6	2437	14.60	16.00	
		9	2452	13.30	14.50	

	Mode	Channel	Frequency (MHz)	Average power (dBm)	Tune-Up Limit	Duty Cycle %
5.2GHz WLAN	802.11a 6Mbps	36	5180	15.39	17.00	92.53
		40	5200	15.33	17.00	
		44	5220	15.24	17.00	
		48	5240	15.15	17.00	
	802.11n-HT20 MCS0	36	5180	15.02	16.00	95.00
		40	5200	14.66	16.00	
		44	5220	14.82	16.00	
		48	5240	14.73	16.00	
	802.11n-HT40 MCS0	38	5190	15.79	17.00	90.46
		46	5230	15.47	17.00	
	802.11ac-VHT20 MCS0	36	5180	12.82	13.00	94.66
		40	5200	12.70	13.00	
		44	5220	12.74	13.00	
		48	5240	12.64	13.00	
	802.11ac-VHT40 MCS0	38	5190	12.66	13.00	89.77
		46	5230	12.31	13.00	
	802.11ac-VHT80 MCS0	42	5210	12.08	13.00	82.24

	Mode	Channel	Frequency (MHz)	Average power (dBm)	Tune-Up Limit	Duty Cycle %
5.8GHz WLAN	802.11a 6Mbps	149	5745	14.36	16.00	92.53
		157	5785	14.31	16.00	
		165	5825	14.35	16.00	
	802.11n-HT20 MCS0	149	5745	14.00	15.50	95.00
		157	5785	13.97	15.50	
		165	5825	13.96	15.50	
	802.11n-HT40 MCS0	151	5755	14.70	16.00	90.46
		159	5795	14.68	16.00	
	802.11ac-VHT20 MCS0	149	5745	12.04	13.00	94.66
		157	5785	12.02	13.00	
		165	5825	11.00	12.00	
	802.11ac-VHT40 MCS0	151	5755	11.80	13.00	89.77
		159	5795	11.78	13.00	
	802.11ac-VHT80 MCS0	155	5775	11.41	13.00	82.24