

**User Manual**

Jinmu Health Technology Co., LTD.

Model : JL02



## Includes

| Item                | No |
|---------------------|----|
| PulesCheckUp Sensor | x1 |
| User Manual         | x1 |
| Leather Case        | x1 |

## Specification

|                        |                                       |
|------------------------|---------------------------------------|
| Product Name           | PulesCheckUp Sensor                   |
| Dimension              | 62mm X35mm x 31mm                     |
| Display                | 0.96 in.OLED Resolution 128X64 pixels |
| Communication Standard | Bluetooth BLE4.2                      |
| Idle Time              | 7 Days                                |
| Voltage                | 3.7V DC                               |
| Capacity               | 600mAh                                |
| Charging Voltage       | 5V DC                                 |
| Charger                | Micro USB                             |
| Operating Condition    | Temperature 50~104°F Humidity<85%     |
| Storage Condition      | Temperature 14~104°F Humidity<85%     |
| Age Limit              | 16 or older                           |

## **Safety Information**

- **Please read this manual thoroughly before using the device!**
- **Warnings alert the user to potential serious outcomes, such as injury or adverse events to the patient or user.**
- **Cautions alert the user to exercise care necessary for the safe and effective use of the device.**
- **Notes contain important information that may be overlooked or missed.**

## **Disclaimer**

**This device is not intended to be a substitute for professional medical advice, diagnosis or treatment. Always seek the advice of your physician or other qualified health provider with any questions about your medical condition DO NOT disregard professional medical advice or delay seeking advice or treatment because of something you have read here.**

## **Warnings**

- **Keep away from source of fire and/or heat.**
- **DO NOT use the device in an MRI or a CT environment.**
- **DO NOT use the device in the presence of flammable anesthetics.**
- **Explosion hazard: DO NOT use the device in an explosive atmosphere.**
- **Chemicals from a broken OLED panel are toxic when ingested Use caution when the device has a broken display screen.**
- **The device is intended only as an adjunct in patient assessment It must be used in conjunction with other methods of assessing clinical signs and symptoms.**
- **Check the device application site frequently to determine the positioning of measurement and circulation and skin sensitivity of the patient.**

## Cautions

- Inaccurate readings for people who are 16 years old or younger.
- Inaccurate measurements may be caused by autoclaving, ethylene oxide sterilizing, or immersing the sensors in liquid may cause inaccurate readings.
- Significant levels of dysfunctional hemoglobin (such as carbonyl-hemoglobin or methemoglobin) may cause inaccurate readings.
- Intravascular dyes may cause inaccurate readings.
- Measurements may be adversely affected in the presence of upper ambient light.
- Excessive patient movement may cause inaccurate readings.
- Upper-frequency electro-surgical interference may cause inaccurate readings.
- Venous pulsations may cause inaccurate readings.
- Placement of a sensor on an extremity with a blood pressure cuff, arterial catheter, or intravascular line may cause inaccurate readings.
- The patient has hypotension, severe vasoconstriction, severe anemia, or hypothermia may cause inaccurate readings.


## **Cleaning**

- Always power off the system first.
- Never spray cleaning fluids directly onto the device. Always spray the cleaner onto a cloth then wipe.
- Do not allow any cleaner to run or drip near circuit boards, or near seams on device casings.
- Never allow circuits or electronics to become damp or wet.

## App installation

### ① Turn On Bluetooth

Please turn on the Bluetooth in your device Settings



### ② Download and Install App

- For iOS devices, search "PulseCheckUp" at App Store
- For Android devices, search "PulseCheckUp" at Google Play Store





## **Warranty**

### **General**

All products sold by JimLab Corp. are warranted against defects in material and workmanship for a period one (1) year from the date of shipment.

### **Limitations**

This warranty does not apply to defects resulting from any Customer actions, such as mishandling, improper interfacing, operation outside of design limits, misapplication, improper repair, or unauthorized modification. No other warranties are expressed or implied.

### **Contacts**

JimLab wants to hear from you. Tell us about your experiences with our app. We are always looking to improve and appreciate any feedback.

Please send us your comment at [service@jinmu.com.tw](mailto:service@jinmu.com.tw)

## FCC Statement

This equipment has been tested and found to comply with the limits for a Class B digital device, pursuant to Part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference in a residential installation.

This equipment generates, uses and can radiate radio frequency energy and, if not installed and used in accordance with the instructions, may cause harmful interference to radio communications.

However, there is no guarantee that interference will not occur in a particular installation.

If this equipment does cause harmful interference to radio or television reception, which can be determined by turning the equipment off and on, the user is encouraged to try to correct the interference by one or more of the following measures:

Reorient or relocate the receiving antenna.

Increase the separation between the equipment and receiver.

Connect the equipment into an outlet on a circuit different from that to which the receiver is connected.

Consult the dealer or an experienced radio/TV technician for help.

## FCC Statement

### Caution:

Any changes or modifications in construction of this device which are not expressly approved by the party responsible for compliance could void the user's authority to operate the equipment.

This device complies with Part 15 of the FCC Rules. Operation is subject to the following two conditions:

- 1) this device may not cause harmful interference and
- 2) this device must accept any interference received, including interference that may cause undesired operation of the device.