



| | | |
|---|--|--|
| BONTECH www.bontech1.com | BONTECH Co., Ltd. Digital Empire D Building #1201,1202,1203 16, Deogyong-daero 1556 beon-gil, Yeongtong-gu, Suwon-si, Gyunggi-do, Korea TEL : 82-31-303-5252 FAX : 82-31-303-5255 | |
| | | |
| REF BT-DA22W | SN XXXXXXXX | YYYY/MM/DD |
| Rating : 12V 3.75A (Adapter : MPU51-105) | | |
| FCC ID : 2AVRHBT-DA22W | | |
| CLASSIFIED c UL us E510892 | MEDICAL - GENERAL MEDICAL EQUIPMENT ANSIAAMI ES60601-1 (2005) + AMD 1 (2012) CAN/CSA-C22.2 No. 60601-1 (2014) | (01) 08800069500068 (11) 200213 (21) BY0202001 |
| "This product complies with 21 CFR Subchapter J" | | |
| EC REP | CMC Medical Devices & Drugs S.L. C/Horacio Lengo Nº 18, CP 29006, Málaga, Spain. | Made in Korea |
| This device complies with part 15 of the FCC Rules. Operation is subject to the following two conditions: (1) This device may not cause harmful interference, and (2) this device must accept any interference received, including interference that may cause undesired operation. | | |