

comments and response:

- b) Given the size/nature of the device, was it rotated about 3 axis in effort to obtain worse case results?

IMQ [EF]: we confirm, it has been rotated about 3 axis in effort to obtain worse case results.