

To:  
Federal Communications Commission  
Equipment Authorization Division  
7435 Oakland Mills Road  
Columbia, MD 21048  
U.S.A.

Date: 22 November 2020

### **FCC Labeling of Widex model MBB3D**

FCC ID: 2AXDT-MBB3D

#### To Whom It May Concern:

The dimensions of the hearing aid are too small to allow the FCC ID to be printed on the device in a font size that is readable. Consequently, the FCC ID is printed in the user instructions together with the FCC regulatory statements.

Justification is provided by 3 photographs below that show the MBB3D in 3 different angles together with a printout of the FCC ID in font size 4 point that is the minimum readable font size as defined in FCC OET Labelling Guidelines.

Bottom view



Side view



Top view



Thank you for your attention to this matter.

Sincerely,  
*WSAUD A/S*

Farooq Jawed Shaikh  
RF Design Engineer, Wireless system integration