



Date: 04/19/2022
Nemko North America, Inc.
303 River Road
Ottawa, Ontario, Canada

Dear Sir or Madam,

We, Advanced TeleSensors, Inc, 5114 Balcones Woods Dr. Suite 307-472, Austin, TX 78759 hereby request a Class 2 Permissive Change / Reassessment exercise for our non-contact vital sign monitor, FCC ID: 2AXL8-ATX2410V4P1. The objective is to include an additional variation of the device with the following changes:

- Depopulated HTS, IR Sens and DAC components
- Antenna traces moved slightly on PCB
- Inclusion of WiFi collocated radio

Yours sincerely,

A handwritten signature in black ink that reads "Robert W. Hohman".

Name: Robert W. Hohman
Title: Engineering Mgr