

Hunan Vathin Medical Instrument Co., Ltd.

July 29, 2024

Federal Communications Commission
Equipment Authorization Division
7435 Oakland Mills Road
Columbia, MD 21046
USA

Attn: OET Dept.

Ref: FCC Class II Permissive change for FCC ID: 2AY4E-DVMD

Applicant: Hunan Vathin Medical Instrument Co., Ltd.

Dear Examiner,

This is to request a Class II Permissive change for FCC ID: 2AY4E-DVMD originally granted on 09/28/2023 (date).

The change under this application is to

1. The battery model used in the original product: 08DQ0017-03(Rated capacity:12Ah/129.6Wh)

The battery model used in the new product: 08DQ0017-02(Rated capacity:9000mAh/97.2Wh)



Original Product



New Product

2. Add series model DVM-D2, the difference between DVM-D1 and DVM-D2 models lies in the video endoscope interface.

DVM-D1 model video endoscope interface: I (using a 26 pin connector) II (using a 14 pin connector) III (using a 14 pin connector)

DVM-D2 model video endoscope interface: I (using a 26 pin connector) II (using a 16 pin connector) III (using a 16 pin connector)



DVM-D1



DVM-D2

I attest that the certified device will not be capable of ad-hoc mode operation outside of the grant conditions.

Sincerely,

Name: Lily Fei

Date: 2024-7-29

Title: QA&RA Manager

Signature of applicant

FCC Class II PC request letter Rev1.1