



US Federal Communications Commission (FCC)

- **FCC ID: 2AYAJ-BS1**
- **This transmitter is authorized by rule under the Medical Device Radiocommunication Service (in part 95 of the FCC Rules) and must not cause harmful interference to stations operating in the 400.150–406.000 MHz band in the Meteorological Aids (i.e., transmitters and receivers used to communicate weather data), the Meteorological Satellite, or the Earth Exploration Satellite Services and must accept interference that may be caused by such stations, including interference that may cause undesired operation. This transmitter shall be used only in accordance with the FCC Rules governing the Medical Device Radiocommunication Service. Analog and digital voice communications are prohibited. Although this transmitter has been approved by the Federal Communications Commission, there is no guarantee that it will not receive interference or that any particular transmission from this transmitter will be free from interference. This equipment complies with FCC radiation exposure limits set forth for an uncontrolled environment.**
- **This device complies with part 15 of the FCC Rules. Operation is subject to the following two conditions: (1) This device may not cause harmful interference, and (2) this device must accept any interference received, including interference that may cause undesired operation.**
- **IMPORTANT: Changes or modifications to this product not authorized by Canary Medical could void the FCC Certification and negate your authority to operate this product.**

FCC RF Exposure statement

The device must be installed/operated to provide a separation distance of at least 20 cm from all persons and must not be colocated or operating in conjunction with any other transmitter, except in accordance with FCC multi-transmitter product RF exposure procedures.

Manufactured for:

Canary Medical, LLC
2710 Loker Ave. West
Suite 350
Carlsbad, CA 92010

www.canarymedical.com

E-mail: support@canarymedical.com

Customer Assistance: 1-800-123-9876

PN. K01-HBS-300003, Rev2

