

# **Appendix B - DAE & Probe Calibration Certificate**

ccredited by the Swiss Accredit the Swiss Accreditation Servic ultilateral Agreement for the r	e is one of the signatories	to the EA	No.: SCS 0108
lient SGS - TW (Au	den)	Certificate No	: DAE4-1665_Feb22
ALIBRATION	CERTIFICATE		
Dbject.	DAE4 - SD 000 D	04 BO - SN: 1665	
Calibration procedure(s)	QA CAL-06.v30 Calibration proced	dure for the data acquisition elec	tronics (DAE)
Calibration date:	February 28, 2022	2	
he measurements and the unc	ertainties with confidence pro	nal standards, which realize the physical un bability are given on the following pages an	d are part of the certificate.
The measurements and the unc	ertainties with confidence pro		d are part of the certificate.
The measurements and the uncould calibrations have been conducted and conducted the conducted (M8	ertainties with confidence pro incted in the closed laboratory ITE critical for calibration)	obability are given on the following pages an facility: environment temperature $(22 \pm 3)^\circ$	id are part of the certificate. C and humidity < 70%.
The measurements and the unc All calibrations have been condu- Calibration Equipment used (M8 Primary Standards	ertainties with confidence pro	obability are given on the following pages an	d are part of the certificate.
The measurements and the unc All calibrations have been condu Calibration Equipment used (M& Primary Standards Keithley Multimeter Type 2001	ertainties with confidence pro- incted in the closed laboratory ATE critical for calibration) ID # SN: 0810278	2020 Date (Certificate No.) 31-Aug-21 (No:31368)	d are part of the certificate. C and humidity < 70%. Scheduled Calibration Aug-22
The measurements and the unc All calibrations have been condu Calibration Equipment used (M& Primary Standards Keithley Multimeter Type 2001 Secondary Standards	ertainties with confidence pro- incted in the closed laboratory (TE critical for calibration) ID # SN: 0810278 ID #	2552 2014 2014 2014 2014 2014 2014 2014 201	d are part of the certificate. C and humidity < 70%. Scheduled Calibration Aug-22 Scheduled Check
The measurements and the unc All calibrations have been condu- Calibration Equipment used (M8 Primary Standards Keithley Multimeter Type 2001 Secondary Standards Auto DAE Calibration Unit	ertainties with confidence pro- incted in the closed laboratory ATE critical for calibration) ID # SN: 0810278	bability are given on the following pages an facility: environment temperature (22 ± 3)°( Cal Date (Certificate No.) 31-Aug-21 (No:31368) Check Date (in house) 24-Jan-22 (in house check)	d are part of the certificate. C and humidity < 70%. Scheduled Calibration Aug-22
The measurements and the unc All calibrations have been condu- Calibration Equipment used (M8 Primary Standards Keithley Multimeter Type 2001 Secondary Standards Auto DAE Calibration Unit	ertainties with confidence pro- incred in the closed laboratory TE critical for calibration) ID # SN: 0810278 ID # SE UWS 053 AA 1001	bability are given on the following pages an facility: environment temperature (22 ± 3)°( Cal Date (Certificate No.) 31-Aug-21 (No:31368) Check Date (in house) 24-Jan-22 (in house check)	d are part of the certificate. C and humidity < 70%. Scheduled Calibration Aug-22 Scheduled Check In house check: Jan-23
The measurements and the unc All calibrations have been condu- Calibration Equipment used (M8 Primary Standards Geithley Multimeter Type 2001 Secondary Standards Auto DAE Calibration Unit	ertainties with confidence pro- incted in the closed laboratory TE critical for calibration) ID # SN: 0810278 ID # SE UWS 053 AA 1001 SE UMS 006 AA 1002	bability are given on the following pages an facility: environment temperature (22 ± 3)°C Cal Date (Certificate No.) 31-Aug-21 (No:31368) Check Date (in house) 24-Jan-22 (in house check) 24-Jan-22 (in house check)	d are part of the certificate. C and humidity < 70%. Scheduled Calibration Aug-22 Scheduled Check In house check: Jan-23 In house check: Jan-23
The measurements and the unc All calibrations have been condu- Calibration Equipment used (M8 Primary Standards Keithley Multimeter Type 2001 Secondary Standards Auto DAE Calibration Unit Calibrator Box V2.1	ertainties with confidence pro- incred in the closed laboratory TE critical for calibration) ID # SN: 0810278 ID # SE UWS 053 AA 1001	bability are given on the following pages an facility: environment temperature (22 ± 3)°( Cal Date (Certificate No.) 31-Aug-21 (No:31368) Check Date (in house) 24-Jan-22 (in house check)	d are part of the certificate. C and humidity < 70%. Scheduled Calibration Aug-22 Scheduled Check In house check: Jan-23
The measurements and the unc All calibrations have been condu- calibration Equipment used (M& Primary Standards Geithley Multimeter Type 2001 Secondary Standards Auto DAE Calibration Unit Calibrator Box V2.1	ertainties with confidence pro- acted in the closed laboratory TE critical for calibration) ID # SN: 0810278 ID # SE UWS 053 AA 1001 SE UMS 006 AA 1002	bability are given on the following pages an facility: environment temperature (22 ± 3)°( Cal Date (Certificate No.) 31-Aug-21 (No:31368) Check Date (in house) 24-Jan-22 (in house check) 24-Jan-22 (in house check) Enction	d are part of the certificate. C and humidity < 70%. Scheduled Calibration Aug-22 Scheduled Check In house check: Jan-23 In house check: Jan-23
The measurements and the unc	ertainties with confidence pro- acted in the closed laboratory TE critical for calibration) ID # SN: 0810278 ID # SE UWS 053 AA 1001 SE UMS 006 AA 1002	bability are given on the following pages an facility: environment temperature (22 ± 3)°( Cal Date (Certificate No.) 31-Aug-21 (No:31368) Check Date (in house) 24-Jan-22 (in house check) 24-Jan-22 (in house check) Enction	d are part of the certificate. C and humidity < 70%. Scheduled Calibration Aug-22 Scheduled Check In house check: Jan-23 In house check: Jan-23

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 90 days only.

除非另有說明,此報告結果僅對測試之樣品負責,同時此樣品僅保留的天。本報告未經本公司書面許可,不可部份複製。 This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <a href="http://www.sgs.com.tw/Terms-and-Conditions">http://www.sgs.com.tw/Terms-and-Conditions</a> and for electronic format documents issued by the Company subject to its General Conductors and every printed overlear, available on request of accessible at <u>http://www.sgs.com.tw/Terms-and-Conditions</u>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

No.134,Wu Kung Road, New Taipei Industrial Park, Wuku District, New Taipei City, Taiwan/新北市五股區新北產業園區五工路 134 號

t (886-2) 2299-3279 台灣檢驗科技股份有限公司

f (886-2) 2298-0488

```
www.sqs.com.tw
```



Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland





S

C

S

Schweizerischer Kalibrierdienst Service suisse d'étalonnage Servizio svizzero di taratura **Swiss Calibration Service** 

Accreditation No.: SCS 0108

Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

#### Glossary

DAE Connector angle

data acquisition electronics information used in DASY system to align probe sensor X to the robot coordinate system.

### Methods Applied and Interpretation of Parameters

- DC Voltage Measurement: Calibration Factor assessed for use in DASY system by comparison with a calibrated instrument traceable to national standards. The figure given corresponds to the full scale range of the voltmeter in the respective range.
- Connector angle: The angle of the connector is assessed measuring the angle mechanically by a tool inserted. Uncertainty is not required.
- The following parameters as documented in the Appendix contain technical information as a result from the performance test and require no uncertainty.
  - DC Voltage Measurement Linearity: Verification of the Linearity at +10% and -10% of the nominal calibration voltage. Influence of offset voltage is included in this measurement.
  - Common mode sensitivity: Influence of a positive or negative common mode voltage on the differential measurement.
  - Channel separation: Influence of a voltage on the neighbor channels not subject to an input voltage.
  - AD Converter Values with inputs shorted: Values on the internal AD converter corresponding to zero input voltage
  - Input Offset Measurement: Output voltage and statistical results over a large number of zero voltage measurements.
  - Input Offset Current: Typical value for information; Maximum channel input offset current, not considering the input resistance.
  - Input resistance: Typical value for information: DAE input resistance at the connector, . during internal auto-zeroing and during measurement.
  - Low Battery Alarm Voltage: Typical value for information. Below this voltage, a battery alarm signal is generated.
  - Power consumption: Typical value for information. Supply currents in various operating modes.

Certificate No: DAE4-1665\_Feb22

Page 2 of 5

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 90 days only.

除非另有說明,此報告結果僅對測試之樣品負責,同時此樣品僅保留90天。本報告未經本公司書面許可,不可部份複製。 This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <u>http://www.sgs.com.tw/Terms-and-Conditions</u> and for electronic format documents, subject to Terms and Conditions for Electronic Documents at <a href="http://www.sgs.com.tw/Terms-and-Conditions">http://www.sgs.com.tw/Terms-and-Conditions</a>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction form exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law

No.134,Wu Kung Road, New Taipei Industrial Park, Wuku District, New Taipei City, Taiwan/新北市五股區新北產業園區五工路 134 號

t (886-2) 2299-3279 台灣檢驗科技股份有限公司

f (886-2) 2298-0488

www.sqs.com.tw



### **DC Voltage Measurement**

A/D - Converter Reso High Range:	1LSB =	6.1µV,	full range =	-100+300 mV
Low Range:	1LSB =	61nV,	full range =	-1+3mV
DASY measurement	parameters: Au	to Zero Time: 3	sec; Measuring	time: 3 sec

Calibration Factors	X	Y	Z
High Range	404.538 ± 0.02% (k=2)	404.846 ± 0.02% (k=2)	404.799 ± 0.02% (k=2)
Low Range	3.97984 ± 1.50% (k=2)	4.00706 ± 1.50% (k=2)	3.97892 ± 1.50% (k=2)

**Connector Angle** 

Connector Angle to be used in DASY system	67.5 ° ± 1 °
---	--------------

Certificate No: DAE4-1665\_Feb22

Page 3 of 5

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 90 days only. 除非另有說明,此報告結果僅對測試之樣品負責,同時此樣品僅保留90天。本報告未經本公司書面許可,不可部份複製。 This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <u>http://www.sgs.com.tw/Terms-and-Conditions</u> and for electronic format documents subject to Terms and Conditions for Electronic Documents at <u>http://www.sgs.com.tw/Terms-and-Conditions</u>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

No.134,Wu Kung Road, New Taipei Industrial Park, Wuku District, New Taipei City, Taiwan/新北市五股區新北產業園區五工路 134 號

t (886-2) 2299-3279 台灣檢驗科技股份有限公司

f (886-2) 2298-0488

www.sgs.com.tw



### Appendix (Additional assessments outside the scope of SCS0108)

# 1. DC Voltage Linearity

High Range	Reading (µV)	Difference (µV)	Error (%)
Channel X + Input	199994.01	0.74	0.00
Channel X + Input	20001.99	0.13	0.00
Channel X - Input	-20001.31	0.21	-0.00
Channel Y + Input	199989.61	-3.62	-0.00
Channel Y + Input	20000.37	-1.46	-0.01
Channel Y - Input	-20002.22	-0.64	0.00
Channel Z + Input	199995.34	2.10	0.00
Channel Z + Input	19997.55	-4.30	-0.02
Channel Z - Input	-20003.98	-2.43	0.01

Low Range	Reading (µV)	Difference (µV)	Error (%)
Channel X + Input	2000.61	-0.18	-0.01
Channel X + Input	201.51	0.25	0.13
Channel X - Input	-198.43	-0,05	0.02
Channel Y + Input	2001.19	0.29	0.01
Channel Y + Input	200,89	-0.45	-0.22
Channel Y - Input	-198.94	-0.49	0.24
Channel Z + Input	2000.93	0.12	0.01
Channel Z + Input	199.86	-1.37	-0.68
Channel Z - Input	-199.91	-1.34	0.68

### 2. Common mode sensitivity

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

	Common mode Input Voltage (mV)	High Range Average Reading (μV)	Low Range Average Reading (µV)
Channel X	200	-2.56	-4.08
	- 200	5.04	3,18
Channel Y	200	1.55	0.85
	- 200	-2.56	-2.50
Channel Z	200	-14.89	-14.80
	- 200	13.36	12.52

### 3. Channel separation

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

	Input Voltage (mV)	Channel X (µV)	Channel Y (µV)	Channel Z (µV)
Channel X	200		0.23	-2.92
Channel Y	200	4.73	-	1.35
Channel Z	200	8.05	2.29	

Certificate No: DAE4-1665\_Feb22

Page 4 of 5

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 90 days only.

除非另有說明,此報告結果僅對測試之樣品負責,同時此樣品僅保留90天。本報告未經本公司書面許可,不可部份複製。 This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <u>http://www.sgs.com.tw/Terms-and-Conditions</u> and for electronic format documents issued by the Company subject to its General Conductors and every printed overlear, available on request of accessible at <u>http://www.sgs.com.tw/Terms-and-Conditions</u>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

No.134,Wu Kung Road, New Taipei Industrial Park, Wuku District, New Taipei City, Taiwan/新北市五股區新北產業園區五工路 134 號

t (886-2) 2299-3279 台灣檢驗科技股份有限公司

f (886-2) 2298-0488

www.sqs.com.tw



### 4. AD-Converter Values with inputs shorted

	High Range (LSB)	Low Range (LSB)
Channel X	16097	16090
Channel Y	16137	13762
Channel Z	16309	15672

#### 5. Input Offset Measurement

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec Input 10MΩ

	Average (µV)	mîn. Offset (μV)	max. Offset (μV)	Std. Deviation (µV)
Channel X	0.70	-0.02	1.57	0.30
Channel Y	-0.24	-1.38	0.80	0.42
Channel Z	-0.59	-1.93	0.05	0.35

### 6. Input Offset Current

Nominal Input circuitry offset current on all channels: <25fA

7. Input Resistance (Typical values for information)

	Zeroing (kOhm)	Measuring (MOhm)
Channel X	200	200
Channel Y	200	200
Channel Z	200	200

# 8. Low Battery Alarm Voltage (Typical values for information)

Typical values	Alarm Level (VDC)	
Supply (+ Vcc)	+7.9	
Supply (- Vcc)	-7.6	

#### 9. Power Consumption (Typical values for information)

Typical values	Switched off (mA)	Stand by (mA)	Transmitting (mA)
Supply (+ Vcc)	+0.01	+6	+14
Supply (- Vcc)	-0.01	-8	-9

Certificate No: DAE4-1665\_Feb22

Page 5 of 5

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 90 days only.

除非另有說明,此報告結果僅對測試之樣品負責,同時此樣品僅保留90天。本報告未經本公司書面許可,不可部份複製。 This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <u>http://www.sgs.com.tw/Terms-and-Conditions</u> and for electronic format documents, subject to Terms and Conditions for Electronic Documents at <a href="http://www.sgs.com.tw/Terms-and-Conditions">http://www.sgs.com.tw/Terms-and-Conditions</a>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

No.134,Wu Kung Road, New Taipei Industrial Park, Wuku District, New Taipei City, Taiwan/新北市五股區新北產業園區五工路 134 號

t (886-2) 2299-3279 台灣檢驗科技股份有限公司



credited by the Swiss Accred	itation Service (SAS)	Acci	reditation No.: SCS 0108
	ice is one of the signatories		
ultilateral Agreement for the	recognition of calibration of	artificates	
lient SGS-TW (Aug	den)	Certificate No:	EX3-7686 Oct21
The second second second			CONTRACTOR OF THE OWNER
CALIBRATION	CERTIFICATE		
Object	EX3DV4 - SN:768	6	
Calibration procedure(s)		A CAL-14.v6, QA CAL-23.v5, QA	CAL-25.v7
	Calibration proced	ure for dosimetric E-field probes	
Calibration date:	October 05, 2021		
		al standards, which realize the physical units	
he measurements and the un	certainties with confidence pro		are part of the certificate
	the second s	county are given on the following pages and a	and hear of the entitledity
		causing and given on the following pages and a	
All calibrations have been cond		facility: environment temperature $(22 \pm 3)$ °C a	
All calibrations have been cond			
	lucted in the closed laboratory		
	lucted in the closed laboratory		
Calibration Equipment used (M	lucted in the closed laboratory	facility: environment temperature (22 ± 3)°C a	and humidily < 70%.
Calibration Equipment used (M Primary Standards	lucted in the closed laboratory		
Calibration Equipment used (M Primary Standards Power meter NRP	Nucled in the closed laboratory &TE critical for calibration)	facility: environment temperature (22 ± 3)°C a	and humidity < 70%.
Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-291	Aucted in the closed laboratory &TE critical for calibration)	facility: environment temperature (22 ± 3)°C a Cal Date (Certificate No.) 88-Apr/21 (No. 217-03291/03262)	and humidity < 70%. Scheduled Calibration Apr-21
Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator	Aucted in the closed laboratory &TE critical for calibration) ID SN: 104778 SN: 104244	facility: environment temperature (22 ± 3)°C a Cal Date (Certificate No.) 06-Apr-21 (No. 217-03291/03292) 08-Apr-21 (No. 217-03291)	Scheduled Calibration Apr-21 Apr-21
Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator DAE4	Iucted in the closed laboratory &TE critical for calibration) ID SN: 104778 SN: 103244 SN: 103245 SN: C22552 (20x) SN: 660	facility: environment temperature (22 ± 3)°C a           Cal Date (Certificate No.)           09-Apr/21 (No. 217-0328103282)           09-Apr/21 (No. 217-03281)           09-Apr/21 (No. 217-03282)           09-Apr/21 (No. 217-03283)           23-Dec/90 (No. DAFA-660 Dec/20)	Scheduled Calibration Apr-21 Apr-21 Apr-21
Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator DAE4	Auded in the closed laboratory &TE critical for calibration) ID SN: 104778 SN: 103244 SN: 103245 SN: C02562 (20x)	facility: environment temperature (22 ± 3)°C a	Scheduled Calibration Apr.21 Apr.21 Apr.21 Apr.21 Apr.21 Apr.21
Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator DAE4 Reference Probe ES3DV2	Ucted in the closed laboratory           &TE critical for calibration)           ID           SN: 104778           SN: 102244           SN: 102245           SN: 022552 (20x)           SN: 600           SN: 3013	facility: environment temperature (22 ± 3)°C a           Cal Date (Certificate No.)           09-Apr21 (No. 277-03281/03292)           09-Apr21 (No. 277-032821)           09-Apr21 (No. 277-032821)           09-Apr21 (No. 277-032821)           09-Apr21 (No. 277-032821)           09-Apr22 (No. 277-032821)           09-Apr21 (No. 277-032821)           09-Apr22 (No. 276-032831)           09-Apr21 (No. 277-032821)           09-Apr22 (No. 276-032831)           09-Apr22 (No. 276-032821)           09-Apr23 (No. 276-032821)           09-Apr24 (No. 277-032821)           09-Apr27 (No. 276-032821)           09-Apr27 (No	Scheduled Calibration           Apr.21           Apr.21           Apr.21           Dec.21           Dec.21
Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator DAEA Reference Probe ES3DV2 Secondary Standards	Nucled in the closed laboratory           &TE critical for calibration)           ID           SN: 104778           SN: 103244           SN: 00245           SN: 00245           SN: 00245           SN: 30324           SN: 60245           SN: 303244           ID           SN: 303244           ID           ID	Cal Date (Certificate No.)           06-Apr-21 (No. 217-03291/03292)           09-Apr-21 (No. 217-03291)           09-Apr-21 (No. 217-03292)           09-Apr-21 (No. 217-03292)	Scheduled Calibration Apr-21 Apr-21 Apr-21 Apr-21 Dec-21 Dec-21 Scheduled Chick
Calibration Equipment used (M Primary Standards Power mater NRP Power sensor NRP-291 Power sensor NRP-291 Power sensor NRP-294 Reference Probe ES3DV2 Secondary Standards Power mater E4108	ID         SN: 10276           SN: 104778         SN: 10274           SN: 10274         SN: 10224           SN: 05245         SN: 05245           SN: 05245         SN: 0541253874	Cal Date (Certificate No.)           08-Apr-21 (No. 217-03281/03262)           09-Apr-21 (No. 217-032821)           09-Apr-21 (No. 217-032843)           23-Dec-20 (No. DE4-680 Dec20)           30-Dec-20 (No. E83-3013 Dec20)           Check Date (n house)           08-Apr-16 (in house)	Scheduled Calibration Apr-21 Apr-21 Apr-21 Apr-21 Dec-21 Dec-21 Dec-21 Scheduled Check In house check: Jur-22
Calibration Equipment used (M Primary Standards Power enter NRP Power sensor NRP-291 Power sensor NRP-291 Power sensor NRP-292 Reference 20 dB Attenuator DAE4 Reference Probe ES3DV2 Secondary Standards Power matter E44198 Power sensor E4412A	Nucled in the closed laboratory           &TE critical for calibration)           ID           SN: 104778           SN: 104778           SN: 103244           SN: 103244           SN: 02244           SN: 3013           ID           SN: 0841293874           SN: Mr44496087	facility: environment temperature (22 ± 3)°C a           Cal Date (Certificate No.)           09-Apr-21 (No. 217-03291)           09-Apr-21 (No. 217-03291)           09-Apr-21 (No. 217-03291)           09-Apr-21 (No. 217-03292)           09-Apr-16 (In house check Jun-20)           09-Apr-16 (In house check Jun-20)           09-Apr-16 (In house check Jun-20)	In humidity < 70%. Scheduled Calibration Apr-21 Apr-21 Apr-21 Dec-21 Dec-21 Dec-21 Scheduled Check In house check. Jun-22 In house check. Jun-22
Calibration Equipment used (M Primary Slandards Power mater NRP Power sensor NRP-231 Power sensor NRP-231 Reference 20 de Atenuator DAE4 Secondary Slandards Power matter E44198 Power sensor E412A Power sensor E412A	ID         ID           ID         SN: 104778           SN: 104778         SN: 10244           SN: 102244         SN: 102245           SN: 05244         SN: 05246           SN: 05245         SN: 05245           SN: 05245         SN: 05245           SN: 05245         SN: 0513           ID         SN: 0541293874           SN: 0541293874         SN: 0710210	Cal Date (Certificate No.)           08-Apr-21 (No. 217-03291/03292)           09-Apr-21 (No. 217-03291)           09-Apr-21 (No. 217-03291)           09-Apr-21 (No. 217-03292)           09-Apr-21 (No. 217-03293)           09-Apr-21 (No. 217-03294)           09-Apr-21 (No. 217-03293)           09-Apr-21 (No. 217-03294)           09-Apr-21 (No. 217-03294)           09-Apr-21 (No. 217-03294)           09-Apr-21 (No. 217-03294)           09-Apr-16 (In house check Jun-20)	Scheduled Calibration Apr-21 Apr-21 Apr-21 Der-21 Der-21 Der-21 Der-21 In house check. Jun-22 In house check. Jun-22 In house check. Jun-22
Calibration Equipment used (M Primary Standards Power enter NRP Power sensor NRP-291 Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator DAE4 Secondards Power metre E430V2 Secondards Power sensor E4412A Power sensor E4412A Power sensor E4412A Power sensor E4412A	Nucled in the closed laboratory           &TE critical for calibration)           ID           SN: 104778           SN: 104778           SN: 103244           SN: 103244           SN: 02244           SN: 3013           ID           SN: 0841293874           SN: Mr44496087	facility: environment temperature (22 ± 3)°C a           Cal Date (Certificate No.)           09-Apr21 (No. 217-03291/03292)           09-Apr21 (No. 217-03291)           09-Apr21 (No. 217-03291)           09-Apr21 (No. 217-03343)           23-Dee-20 (No. DAE-460 Dec20)           09-Apr21 (No. 217-03343)           Check Date (In house heck Jun-20)           08-Apr16 (In house check Jun-20)	In humidity < 70%. Scheduled Calibration Apr-21 Apr-21 Apr-21 Dec-21 Dec-21 Dec-21 Dec-21 In house check: Jun-22 In house check: Jun-22 In house check: Jun-22
Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-291 Power sensor NRP-291 Power sensor NRP-294 Reference Probe ES3DV2 Secondary Standards Power meter E44198 Power sensor E4412A	ID         ID           ID         SN: 104778           SN: 10244         SN: 10244           SN: 10244         SN: 10245           SN: 02244         SN: 02244           SN: 02245         SN: 02245           SN: 02245         SN: 02455           SN: 0245         SN: 02455           SN: 0241293874         SN: MY41498087           SN: 000110210         SN: 000110210           SN: US41080477         SN: US41080477	Cal Date (Certificate No.)           08-Apr-21 (No. 217-03291/03292)           09-Apr-21 (No. 217-03291)           09-Apr-21 (No. 217-03291)           09-Apr-21 (No. 217-03292)           09-Apr-21 (No. 217-03293)           09-Apr-21 (No. 217-03294)           09-Apr-21 (No. 217-03293)           09-Apr-21 (No. 217-03294)           09-Apr-21 (No. 217-03294)           09-Apr-21 (No. 217-03294)           09-Apr-21 (No. 217-03294)           09-Apr-16 (In house check Jun-20)	Scheduled Calibration Apr-21 Apr-21 Apr-21 Der-21 Der-21 Der-21 Der-21 In house check. Jun-22 In house check. Jun-22 In house check. Jun-22
Calibration Equipment used (M Primary Slandards Power mater NRP Power mater NRP-291 Power sensor NRP-291 Power sensor NRP-291 Reference 7:04 Retunutor DAE4 Reference 7:04 Retunutor Reference 7:04 Retunutor Power sensor 2:412A Power sensor 2:412A RF generation HP 8648C Network Analyzer E3358A	ID         ID           ID         SN: 104778           SN: 104778         SN: 10244           SN: 102244         SN: 102245           SN: 05244         SN: 05245           SN: 05245         SN: 00110210           SN: 054109007         SN: US5662001700           SN: 00110210         SN: US5642001700           SN: 041080477         Name	Cal Date (Certificate No.)           08-Apr/21 (No. 217-03291/03292)           09-Apr/21 (No. 217-03291)           09-Apr/21 (No. 217-03292)           09-Apr/21 (No. 217-03293)           09-Apr/21 (No. 217-03294)           09-Apr/21 (In house check Jun-20)           09-Apr/21 (In house check Jun-20)           09-Apr/21 (In house check Jun-20)           31-Mar-14 (In house check Oct-20)           Function	In humidity < 70%. Scheduled Calibration Apr-21 Apr-21 Apr-21 Dec-21 Dec-21 Dec-21 Dec-21 In house check: Jun-22 In house check: Jun-22 In house check: Jun-22
Calibration Equipment used (M Primary Slandards Power mater NRP Power mater NRP-291 Power sensor NRP-291 Power sensor NRP-291 Reference 7:04 Retunutor DAE4 Reference 7:04 Retunutor Reference 7:04 Retunutor Power sensor 2:412A Power sensor 2:412A RF generation HP 8648C Network Analyzer E3358A	ID         ID           ID         SN: 104778           SN: 10244         SN: 10244           SN: 10244         SN: 10245           SN: 02244         SN: 02244           SN: 02245         SN: 02245           SN: 02245         SN: 02455           SN: 0245         SN: 02455           SN: 0241293874         SN: MY41498087           SN: 000110210         SN: 000110210           SN: US41080477         SN: US41080477	facility: environment temperature (22 ± 3)°C a           Cal Date (Certificate No.)           09-Apr21 (No. 217-03291/03292)           09-Apr21 (No. 217-032931)           09-Apr21 (No. 217-032931)           09-Apr21 (No. 217-032931)           09-Apr22 (No. 217-032931)           09-Apr22 (No. 217-032931)           09-Apr21 (No. 217-032931)           09-Apr22 (No. 553-3013, Dec20)           Check Date (In house heck, Jun-20)           08-Apr16 (In house heck, Jun-20)           08-Apr16 (In house heck, Jun-20)           08-Apr16 (In house heck, Jun-20)           08-Apr14 (In house heck, Jun-20)           08-Apr14 (In house heck, Jun-20)           31-Mar-14 (In house check Oct-20)	In humidity < 70%. Scheduled Calibration Apr-21 Apr-21 Apr-21 Dec-21 Dec-21 Dec-21 Dec-21 In house check: Jun-22 In
Calibration Equipment used (M Primary Slandards Power mater NRP Power mater NRP-291 Power sensor NRP-291 Power sensor NRP-291 Reference 7:04 Retunutor DAE4 Reference 7:04 Retunutor Reference 7:04 Retunutor Power sensor 2:412A Power sensor 2:412A RF generation HP 8648C Network Analyzer E3358A	ID         ID           ID         SN: 104778           SN: 104778         SN: 10244           SN: 102244         SN: 102245           SN: 05244         SN: 05245           SN: 05245         SN: 00110210           SN: 054109007         SN: US5662001700           SN: 00110210         SN: US5642001700           SN: 041080477         Name	Cal Date (Certificate No.)           08-Apr/21 (No. 217-03291/03292)           09-Apr/21 (No. 217-03291)           09-Apr/21 (No. 217-03292)           09-Apr/21 (No. 217-03293)           09-Apr/21 (No. 217-03294)           09-Apr/21 (In house check Jun-20)           09-Apr/21 (In house check Jun-20)           09-Apr/21 (In house check Jun-20)           31-Mar-14 (In house check Oct-20)           Function	In humidity < 70%. Scheduled Calibration Apr-21 Apr-21 Apr-21 Dec-21 Dec-21 Dec-21 Dec-21 In house check: Jun-22 In
Calibration Equipment used (M Primary Standards Power mater NRP- Power sensor NRP-291 Power sensor NRP-291 Power sensor RRP-294 Reference Probe ES3DV/2 Secondary Standards Power sensor E44128 Power sensor E44128 Power sensor E44128 Power sensor E44128 Network Analyzer E8358A Dathrated by:	ID         ID           ID         SN: 10276           SN: 10278         SN: 10274           SN: 10224         SN: 10224           SN: 02244         SN: 10224           SN: 02245         SN: 680           SN: 02552(20x)         SN: 680           SN: 05245         SN: 680           SN: 05342(20x)         SN: 6841293874           SN: 0541203874         SN: VIS342(201700)           SN: US342(201700)         SN: US342(201700)           SN: US342(201700)         SN: US342(201700)           Amme         Jeforn Kastrati	facility: environment temperature (22 ± 3)°C a           Cal Date (Certificate No.)           08-Apr21 (No. 217-03291/03292)           09-Apr21 (No. 217-03292)           09-Apr21 (Ino. 217-03292)           09-Apr21 (Ino. 217-03292)           09-Apr21 (In house beck Jun-20)           09-Apr16 (In house beck Jun-20)           09-Apr16 (In house beck Jun-20)           31-Mar-14 (In house beck Cdr-20)           Function           Laboratory Technician	In humidity < 70%. Scheduled Calibration Apr-21 Apr-21 Apr-21 Dec-21 Dec-21 Dec-21 Dec-21 In house check: Jun-22 In
Calibration Equipment used (M Primary Standards Power mater NRP- Power sensor NRP-291 Power sensor NRP-291 Power sensor RRP-294 Reference Probe ES3DV/2 Secondary Standards Power sensor E44128 Power sensor E44128 Power sensor E44128 Power sensor E44128 Network Analyzer E8358A Dathrated by:	ID         ID           ID         SN: 104778           SN: 104778         SN: 10244           SN: 102244         SN: 102245           SN: 05244         SN: 05245           SN: 05245         SN: 00110210           SN: 054109007         SN: US5662001700           SN: 1080477         Name	Cal Date (Certificate No.)           Cal Date (Certificate No.)           08-40pr21 (No. 217-03291)           09-40pr21 (No. 217-03291)           09-40pr21 (No. 217-03291)           09-40pr21 (No. 217-03291)           09-40pr21 (No. 217-03292)           09-40pr21 (No. 217-03293)           09-40pr21 (No. 217-03293)           09-40pr21 (No. 217-03293)           09-40pr20 (No. DAte-460, Dec20)           Check Date (in house check Jun-20)           06-40pr16 (in house check Jun-20)           06-40pr16 (in house check Jun-20)           03-40pr20 (in house check Jun-20)           31-40ar-14 (in house check Jun-20)           31-40ar-14 (in house check Oct-20)           Function	In humidity < 70%. Scheduled Calibration Apr-21 Apr-21 Apr-21 Dec-21 Dec-21 Dec-21 Dec-21 In house check: Jun-22 In
Calibration Equipment used (M Primary Standards Power mater NRP- Power sensor NRP-291 Power sensor NRP-291 Power sensor RRP-294 Reference Probe ES3DV/2 Secondary Standards Power sensor E44128 Power sensor E44128 Power sensor E44128 Power sensor E44128 Network Analyzer E8358A Datbrated by:	ID         ID           ID         SN: 10276           SN: 10278         SN: 10274           SN: 10224         SN: 10224           SN: 02244         SN: 10224           SN: 02245         SN: 60           SN: 02552(20x)         SN: 60           SN: 05245         SN: 60           SN: 05245         SN: 60           SN: 051020         SN: 98           SN: 05110210         SN: US362U01700           SN: US34620477         Name           Johon Kastrali         Johon Kastrali	facility: environment temperature (22 ± 3)°C a           Cal Date (Certificate No.)           08-Apr21 (No. 217-03291/03292)           09-Apr21 (No. 217-03292)           09-Apr21 (Ino. 217-03292)           09-Apr21 (Ino. 217-03292)           09-Apr21 (In house beck Jun-20)           09-Apr16 (In house beck Jun-20)           09-Apr16 (In house beck Jun-20)           31-Mar-14 (In house beck Cdr-20)           Function           Laboratory Technician	In humidity < 70%. Scheduled Calibration Apr-21 Apr-21 Apr-21 Dec-21 Dec-21 Dec-21 Dec-21 In house check: Jun-22 In
Calibration Equipment used (M Primary Standards Power enter NRP Power sensor NRP-291 Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator DAE4 Secondards Power metre E430V2 Secondards Power sensor E4412A Power sensor E4412A Power sensor E4412A Power sensor E4412A	ID         ID           ID         SN: 10276           SN: 10278         SN: 10274           SN: 10224         SN: 10224           SN: 02244         SN: 10224           SN: 02245         SN: 60           SN: 02552(20x)         SN: 60           SN: 05245         SN: 60           SN: 05245         SN: 60           SN: 051020         SN: 98           SN: 05110210         SN: US362U01700           SN: US34620477         Name           Johon Kastrali         Johon Kastrali	facility: environment temperature (22 ± 3)°C a           Cal Date (Certificate No.)           08-Apr21 (No. 217-03291/03292)           09-Apr21 (No. 217-03292)           09-Apr21 (Ino. 217-03292)           09-Apr21 (Ino. 217-03292)           09-Apr21 (In house beck Jun-20)           09-Apr16 (In house beck Jun-20)           09-Apr16 (In house beck Jun-20)           31-Mar-14 (In house beck Cdr-20)           Function           Laboratory Technician	In humidity < 70%. Scheduled Calibration Apr-21 Apr-21 Apr-21 Dec-21 Dec-21 Dec-21 Dec-21 In house check: Jun-22 In

Certificate No: EX3-7686\_Oct21

Page 1 of 23

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 90 days only.

t (886-2) 2299-3279

Unless otherwise stated the results shown in this test report reter only to the sample(s) tested and such sample(s) are retained tor 90 days only. Mir#JAfabity i, the state methy mining test report reter only to the sample(s) tested and such sample(s) are retained tor 90 days only. Mir#JAfabity i, the state methy mining test report reter only to the sample(s) tested and such sample(s) are retained tor 90 days only. This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <a href="http://www.sgs.com.tw/Terms-and-Conditions">http://www.sgs.com.tw/Terms-and-Conditions</a> and for electronic format documents, subject to Terms and Conditions for Electronic Documents at <a href="http://www.sgs.com.tw/Terms-and-Conditions">http://www.sgs.com.tw/Terms-and-Conditions</a> and for electronic format defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fulleest extent of the law. prosecuted to the fullest extent of the law.

No.134,Wu Kung Road, New Taipei Industrial Park, Wuku District, New Taipei City, Taiwan/新北市五股區新北產業園區五工路 134 號



#### Calibration Laboratory of Schmid & Partne Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland



Schweizerischer Kallbrierdienst Service suisse d'étalonnage Servizio svizzero di taratura Swiss Calibration Service

S

C

s

Accreditation No.: SCS 0108

Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration cartificates

Glossary:	
TSL	tissue simulating liquid
NORMx,y,z	sensitivity in free space
ConvF	sensitivity in TSL / NORMx.y.z
DCP	diode compression point
CF	crest factor (1/duty_cycle) of the RF signal
A, B, C, D	modulation dependent linearization parameters
Polarization (	φ rotation around probe axis
Polarization 9	9 rotation around an axis that is in the plane normal to probe axis (at measurement center),
	i.e., 9 = 0 is normal to probe axis
Connector Angle	information used in DASY system to align probe sensor X to the robot coordinate system

information used in DASY system to align probe sensor X to the robot coordinate system

- Calibration is Performed According to the Following Standards: a) IEC/IEEE 62209-1528, "Measurement Procedure For The Assessment Of Specific Absorption Rate Of Human Exposure To Radio Frequency Fields From Hand-Held And Body-Worn Wireless Communication Devices -Part 1528: Human Models, Instrumentation And Procedures (Frequency Range of 4 MHz to 10 GHz)", October Procession (Contemportation Contemportation (Contemportation Contemportation) (Contemportation) (Contemportati
  - b) KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

Methods Applied and Interpretation of Parameters:

- ods Applied and Interpretation of Parameters: NORMX, yz. Assessed for E-field polarization  $\vartheta = 0$  ( $\vartheta = 900$  MHz in TEM-cell; f > 1800 MHz: R22 waveguide). NORMX, yz are only intermediate values, i.e., the uncertainties of NORMx, yz does not affect the E<sup>3</sup>-field uncertainty inside TSL (see below *ConvF*). NORM(hx, yz = NORMx, yz \* frequency, response (see Frequency Response Charl). This linearization is implemented in DASY4 software versions later than 4.2. The uncertainty of the frequency response is included in the stated uncertainty of ConvF. DCPX, yz: DCP are numerical linearization parameters assessed based on the data of power sweep with CW signal (no uncertainty required). DCP does not depend on frequency nor media. PAR: PAR is the Peak to Average Ratio that is not calibrated but determined based on the signal
- Ξ. .
- .
- ω.
- PAR: PAR is the Peak to Average Ratio that is not calibrated but determined based on the signal characteristics Axy,z; Bxy,z; Cxy,z; Dxy,z; VRx,y; Z; A, B, C, D are numerical linearization parameters assessed based on the data of power sweep for specific modulation signal. The parameters do not depend on frequency nor media. VR is the maximum calibration range expressed in RMS voltage across the clode. ConvF and Boundary Effect Parameters: Assessed In RIA voltage across the clode. ConvF and Boundary Effect Parameters: Assessed in RIA voltage across the clode. ConvF and Boundary Effect Parameters: Assessed in RIA voltage across the clode. ConvF and Boundary Effect Parameters: Assessed In RIA voltage across the diode. ConvF and Boundary Effect Parameters: Assessed In RIA voltage across the diode. ConvF and Boundary Effect Parameters: Assessed In RIA voltage across the diode. ConvF and Boundary Effect Parameters: Assessed In RIA voltage are used for assessment of the parameters applied for boundary compensation (alpha, depth) of which typical uncertainly voltages are given. These parameters are used in DASY4 software to improve probe accuracy close to the boundary. The sensitivity in TSL corresponds to NORMkyz; \* ConvF whereby the uncertainly corresponds to that given for ConvF. A frequency dependent ConvF is used in DASY version 4.4 and higher which allows extending the validity from  $\pm$  50 MHz to  $\pm$  100 MHz. Spherical isofropy (3D deviation from isofropy): in a field of low gradients realized using a flat phantom exposed by a patch antenna. Sensor Offset: The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No loterance required. Connector Angle: The angle is assessed using the information gained by determining the NORMx (no uncertainty required).
- .
- .

Certificate No: EX3-7686 Oct21

Page 2 of 23

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 90 days only.

t (886-2) 2299-3279

除非另有說明,此報告結果僅對測試之樣品負責,同時此樣品僅保留90天。本報告未經本公司書面許可,不可部份複製。 This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <u>http://www.sgs.com.tw/Terms-and-Conditions</u> and for electronic format documents, subject to Terms and Conditions for Electronic Documents at <a href="http://www.sgs.com.tw/Terms-and-Conditions">http://www.sgs.com.tw/Terms-and-Conditions</a>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law



October 05, 2021

#### EX3DV4 - SN-7686

### DASY/EASY - Parameters of Probe: EX3DV4 - SN:7686

		Senso	r X		Sens	or Y	1.1	Sensor Z	2000	Unc (k=2
Norm (u)	//(V/m) <sup>2</sup> ) <sup>A</sup>	0.69	3		0.6	5		0.53		± 10.1 %
DCP (m)	/) <sup>B</sup>	101.	7		100		-	101.2	-	
Calibrat	ion Results for I		Res		в	-	-			
		stem Name		A dB	dBõV	с	D dB	VR mV	Max dev.	Max Unc <sup>E</sup> (k=2)
0	CW		X	0.00	0.00	1.00	0.00	139.9	± 3.5 %	± 4.7 %
	121		Y	0.00	0.00	1.00		144.6	1	1000
			Z	0.00	0.00	1.00		148.7		
10352-	Pulse Waveform (20	0Hz, 10%)	X	1.41	60.22	6.16	10.00	60.0	± 2.5 %	± 9.6 %
AAA	Dute 10 - (00011 - 2020)		Y	1.60	61.17	6.74		60.0	1	1.1.1
	Pulse Waveform (200Hz, 20%)		Z	1.47	60.50	6.36		60.0	1.00	1.00
10353-	Pulse Waveform (200Hz, 20%)		X	0.80	60.00	4.94	6.99	80.0	± 2.2 %	± 9.6 %
AAA			Y	22.00	78.00	11.00		80.0		1.1
	Pulse Waveform (200Hz, 40%)		Z	0.77	60.00	4.91		80.0		
10354-	Pulse Waveform (200Hz, 40%)		X	0.05	124.90	0.21	3.98	95.0	±2.6 %	± 9.6 %
AAA			Y	8.00	70.00	7.00		95.0		
10000			Z	0.01	122,46	0.50		95.0	-	
10355-	Pulse Waveform (20	0Hz, 60%)	X	9.66	156.50	22.91	2.22	120.0	±1.6 %	± 9.6 %
AAA			Y	11.16	134.19	3.73		120.0		1.00
10387-	ODDK Waysta	6411-	Z	8.10	158.58	27.75	7.00	120.0		
10387- AAA	QPSK Waveform, 1	MHZ	X	0.60	62.96	11.89	1.00	150.0	± 3.8 %	± 9.6 %
MMM	and the second s			08.0	66.90	14.35		150.0	1.000	1.1.1.1
10388-	000000000000000000000000000000000000000	Lat.	Z	0.70	65.20	13.33		150.0		1000
10388- AAA	QPSK Waveform, 10	MHZ	X	1.35	64.80	13.58	0.00	150.0	±1.3%	± 9.6 %
mm			Z	1.55	67.01	14.95		150.0		1.0.1
10396-	64-QAM Waveform.	100 kHz	X	1.47	66.19 62.90	14.43	0.04	150.0		
AAA	04-02-10 Wavelorm,	TOU KHZ	Y	1.55	63.22	15.19 15.60	3.01	150.0	±1.7%	± 9.6 %
	Charles and and and		Z	1.55	63.86	15.60				1.00
10399-	64-QAM Waveform.	40 MH-	X	2.83	65.68	15.63	0.00	150.0	14000	
AAA	or control wavelonni,	40 MI12	Y	2.83	66.59	14.83	0.00	150.0	± 1.5 %	± 9.6 %
			Z	2.90	66.30	15.44		150.0		
10414-	WLAN CCDF, 64-QA	M AOMHZ	X	4.06	66.18	15.45	0.00	150.0	± 2.8 %	1000
AAA	TERM CODI , 04-62	ani, aominiz	Ŷ	4.00	66.01	15.49	0.00	150.0	IZ.8 %	± 9.6 %
			7	3.97	65.86	15.49		150.0	1	

Note: For details on UID parameters see Appendix

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

The uncertainties of Norm X,Y,Z do not affect the E<sup>5</sup> field uncertainty inside TSL (see Pages 5 and 6). Numerical linearization parameter: uncertainty not required. Uncertainty & documented using the max deviation form linear response applying rectangular distribution and is expressed for the square of the

Certificate No: EX3-7686\_Oct21

Page 3 of 23

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 90 days only.

t (886-2) 2299-3279

除非另有說明,此報告結果僅對測試之樣品負責,同時此樣品僅保留90天。本報告未經本公司書面許可,不可部份複製。 This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <a href="http://www.sgs.com.tw/Terms-and-Conditions">http://www.sgs.com.tw/Terms-and-Conditions</a> and for electronic format documents, subject to Terms and Conditions for Electronic Documents at <a href="http://www.sgs.com.tw/Terms-and-Conditions">http://www.sgs.com.tw/Terms-and-Conditions</a>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

No.134,Wu Kung Road, New Taipei Industrial Park, Wuku District, New Taipei City, Taiwan/新北市五股區新北產業園區五工路 134 號

www.sgs.com.tw



#### EX3DV4- SN:7686

#### October 05, 2021

### DASY/EASY - Parameters of Probe: EX3DV4 - SN:7686

	C1 fF	C2 fF	α V <sup>-1</sup>	T1 ms.V <sup>-2</sup>	T2 ms.V <sup>-1</sup>	T3 ms	T4 V <sup>-2</sup>	T5 V-1	T6
Х	11.7	85.69	34.21	3.35	0.00	4.90	0.12	0.00	1.00
Y	11.7	85.51	34.20	3.27	0.00	4.90	0.00	0.00	1.00
Z	11.5	85.16	34.91	1.66	0.00	4.90	0.38	0.00	1.00

#### Other Probe Parameters

Sensor Arrangement	Triangular
Connector Angle (°)	-57.9
Mechanical Surface Detection Mode	enabled
Optical Surface Detection Mode	disabled
Probe Overall Length	337 mm
Probe Body Diameter	10 mm
Tip Length	9 mm
Tip Diameter	2.5 mm
Probe Tip to Sensor X Calibration Point	1 mm
Probe Tip to Sensor Y Calibration Point	1 mm
Probe Tip to Sensor Z Calibration Point	1 mm
Recommended Measurement Distance from Surface	1.4 mm

Note: Measurement distance from surface can be increased to 3-4 mm for an Area Scan job.

Certificate No: EX3-7686 Oct21

Page 4 of 23

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 90 days only.

除非另有說明,此報告結果僅對測試之樣品負責,同時比樣品僅保留90天。本報告未經本公司書面許可,不可部份複製。 This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <a href="http://www.sgs.com.tw/Terms-and-Conditions">http://www.sgs.com.tw/Terms-and-Conditions</a> and for electronic format documents subject to Terms and Conditions for Electronic Documents at <u>http://www.sgs.com.tw/Terms-and-Conditions</u>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

No.134,Wu Kung Road, New Taipei Industrial Park, Wuku District, New Taipei City, Taiwan/新北市五股區新北產業園區五工路 134 號

www.sgs.com.tw



EX3DV4-SN:7686

October 05, 2021

# DASY/EASY - Parameters of Probe: EX3DV4 - SN:7686

Calibration	Daramator	Determined	I too I down a	Timerra	Simulating Media	

f (MHz) <sup>c</sup>	Relative Permittivity <sup>F</sup>	Conductivity (S/m) F	ConvF X	ConvF Y	ConvF Z	Alpha <sup>G</sup>	Depth <sup>G</sup> (mm)	Unc (k=2)
750	41.9	0.89	10.73	10.73	10.73	0.47	0.80	± 12.0 %
835	41.5	0.90	10.36	10.36	10.36	0.49	0.80	± 12.0 %
900	41.5	0.97	10.09	10.09	10.09	0.41	0.90	± 12.0 %
1450	40.5	1.20	9.37	9.37	9.37	0.38	0.80	± 12.0 %
1750	40.1	1.37	9.16	9.16	9.16	0.39	0.86	± 12.0 9
1900	40.0	1.40	8.83	8.83	8.83	0.30	0.86	± 12.0 %
2000	40.0	1.40	8.73	8.73	8.73	0.29	0.86	± 12.0 %
2300	39.5	1.67	8.55	8.55	8.55	0.34	0.90	± 12.0 %
2450	39.2	1.80	8.32	8.32	8.32	0.31	0.90	± 12.0 %
2600	39.0	1.96	8.02	8.02	8.02	0.27	1.00	± 12.0 %
3300	38.2	2.71	7.40	7.40	7.40	0.30	1.35	± 13.1 9
3500	37.9	2.91	7.35	7.35	7.35	0.30	1.35	± 13.1 9
3700	37.7	3.12	7.25	7.25	7.25	0.30	1.35	± 13.1 9
3900	37.5	3.32	6.90	6.90	6.90	0.40	1.60	± 13.1 %
4100	37.2	3.53	6.78	6.78	6.78	0.40	1.60	± 13.1 9
4200	37.1	3.63	6.45	6.45	6.45	0.40	1.70	± 13.1 9
4400	36.9	3.84	6.39	6.39	6.39	0.40	1.70	± 13.1 9
4600	36.7	4.04	6.38	6.38	6.38	0.40	1.70	± 13.1 %
4800	36.4	4.25	6.32	6.32	6.32	0.40	1.80	± 13.1 9
4950	36.3	4.40	6.29	6.29	6.29	0.40	1.80	± 13.1 %
5250	35.9	4.71	5.81	5.81	5.81	0.40	1.80	± 13.1 %
5600	35.5	5.07	5.16	5.16	5.16	0.40	1.80	± 13.1 %
5750	35.4	5.22	5.30	5.30	5.30	0.40	1.80	± 13.1 %

Prequency validity above 300 MHz of ± 100 MHz only applies for DASY V4.4 and higher (see Page 2), else it is restricted to ± 50 MHz. The nucertainty is the RSS of the ConvF uncertainty at calibration frequency and the nucertainty for the indicated frequency band low 300 MHz is 10.25, 40, 60 and 70 MHz to ConvF assessments at 30, 64, 78, 150 and 220 MHz respectively. Validity of ConvF assessed at MHz is 4-4 MHz, and ConvF assessed at 13 MHz is 6-18 MHz have 50 GHz frequency validity and 20 MHz respectively. Validity of ConvF assessments at MHz is 4-4 MHz, and ConvF assessed at 13 MHz is 6-18 MHz have 5 GHz frequency validity can be extended to ± 10 MHz. Alt frequencies before 0.5 GHz, the validity of lossue parameters (and o) for 6 Highed compensation formula is applied to measured 50 KHz have 5 GHz, the validity of lossue parameters (and o) is restricted to ± 5%. The uncertainty is the RSS of Altrbalepht and elemmined during elibration. SER-4 GHz ensating during during elibration elibration of the probe tip and elibration of the obstration of the probe tip aniset from the boundary.

Certificate No: EX3-7686 Oct21

Page 5 of 23

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 90 days only.

t (886-2) 2299-3279

除非另有說明,此報告結果僅對測試之樣品負責,同時此樣品僅保留90天。本報告未經本公司書面許可,不可部份複製。 This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <u>http://www.sgs.com.tw/Terms-and-Conditions</u> and for electronic format documents subject to Terms and Conditions for Electronic Documents at <u>http://www.sgs.com.tw/Terms-and-Conditions</u>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

No.134,Wu Kung Road, New Taipei Industrial Park, Wuku District, New Taipei City, Taiwan/新北市五股區新北產業園區五工路 134 號

f (886-2) 2298-0488

www.sgs.com.tw



EX3DV4-SN:7686

October 05, 2021

### DASY/EASY - Parameters of Probe: EX3DV4 - SN:7686

Calibration Parameter Determined in Head Tissue Simulating Media

f (MHz) <sup>c</sup>	Relative Permittivity <sup>F</sup>	Conductivity (S/m) <sup>F</sup>	ConvF X	ConvF Y	ConvF Z	Alpha <sup>G</sup>	Depth <sup>G</sup> (mm)	Unc (k=2)
6500	34.5	6.07	6.20	6.20	6.20	0.20	2.50	± 18.6 %
7000	33,9	6.65	6.14	6.14	6.14	0.25	2.50	± 18.6 %
8000	32.7	7.84	6.08	6.08	6.08	0.50	1.50	± 18.6 %
9000	31.5	9.08	5.95	5.95	5.95	0.50	1.70	± 18.6 %

<sup>6</sup> Frequency validity above 6GHz is ± 700 MHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band.
<sup>7</sup> All requencies et OGHz, the validity of fissue parameters (c and q) can be relaxed to ± 10% if liquid compensation formula is applied to measured SAR values. The uncertainty is the RSS of the ConvF uncertainty for indicated arget tissue parameters.
<sup>8</sup> Alpha/Depth and defamined during diffrastion SPECAW warrants that the remaining deviation due to the boundary effect after compensation is always less than ± 1% for frequencies below a 5Hz below ± 2% for frequencies between 3-6 GHz; and below ± 4% for frequencies between 8-10 GHz at any distance larger than fail the probe tig dimeter from the boundary.

Certificate No: EX3-7686 Oct21

Page 6 of 23

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 90 days only.

t (886-2) 2299-3279

除非另有說明,此報告結果僅對測試之樣品負責,同時此樣品僅保留90天。本報告未經本公司書面許可,不可部份複製。 This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <u>http://www.sgs.com.tw/Terms-and-Conditions</u> and for electronic format documents, subject to Terms and Conditions for Electronic Documents at <a href="http://www.sgs.com.tw/Terms-and-Conditions">http://www.sgs.com.tw/Terms-and-Conditions</a>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.


No.134,Wu Kung Road, New Taipei Industrial Park, Wuku District, New Taipei City, Taiwan/新北市五股區新北產業園區五工路 134 號



EX3DV4- SN:7686

October 05, 2021

#### Frequency Response of E-Field (TEM-Cell:ifi110 EXX, Waveguide: R22)



Certificate No: EX3-7686 Oct21

Page 7 of 23

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 90 days only.

除非另有說明,此報告結果僅對測試之樣品負責,同時比樣品僅保留90天。本報告未經本公司書面許可,不可部份複製。 This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <a href="http://www.sgs.com.tw/Terms-and-Conditions">http://www.sgs.com.tw/Terms-and-Conditions</a> and for electronic format documents, subject to Terms and Conditions for Electronic Documents at <a href="http://www.sqs.com.tw/Terms-and-Conditions">http://www.sqs.com.tw/Terms-and-Conditions</a>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.


No.134,Wu Kung Road, New Taipei Industrial Park, Wuku District, New Taipei City, Taiwan/新北市五股區新北產業園區五工路 134 號

f (886-2) 2298-0488

www.sgs.com.tw



Report No: ES/2022/30012 Rev: 01 Page: 13 of 28



Certificate No: EX3-7686\_Oct21

Page 8 of 23

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 90 days only.

除非另有說明,此報告結果僅對測試之樣品負責,同時比樣品僅保留90天。本報告未經本公司書面許可,不可部份複製。 This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <a href="http://www.sgs.com.tw/Terms-and-Conditions">http://www.sgs.com.tw/Terms-and-Conditions</a> and for electronic format documents, subject to Terms and Conditions for Electronic Documents at <a href="http://www.sgs.com.tw/Terms-and-Conditions">http://www.sgs.com.tw/Terms-and-Conditions</a>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law


No.134,Wu Kung Road, New Taipei Industrial Park, Wuku District, New Taipei City, Taiwan/新北市五股區新北產業園區五工路 134 號

t (886-2) 2299-3279 台灣檢驗科技股份有限公司

f (886-2) 2298-0488

www.sgs.com.tw





Certificate No: EX3-7686\_Oct21

Page 9 of 23

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 90 days only.


除非另有說明,此報告結果僅對測試之樣品負責,同時此樣品僅保留90天。本報告未經本公司書面許可,不可部份複製。 This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <u>http://www.sgs.com.tw/Terms-and-Conditions</u> and for electronic format documents subject to Terms and Conditions for Electronic Documents at <u>http://www.sgs.com.tw/Terms-and-Conditions</u>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law

No.134,Wu Kung Road, New Taipei Industrial Park, Wuku District, New Taipei City, Taiwan/新北市五股區新北產業園區五工路 134 號

```
www.sgs.com.tw
```



Report No: ES/2022/30012 Rev: 01 Page: 15 of 28



Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 90 days only.

除非另有說明,此報告結果僅對測試之樣品負責,同時此樣品僅保留90天。本報告未經本公司書面許可,不可部份複製。 This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <u>http://www.sgs.com.tw/Terms-and-Conditions</u> and for electronic format documents subject to Terms and Conditions for Electronic Documents at <u>http://www.sgs.com.tw/Terms-and-Conditions</u>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law

No.134,Wu Kung Road, New Taipei Industrial Park, Wuku District, New Taipei City, Taiwan/新北市五股區新北產業園區五工路 134 號

f (886-2) 2298-0488

www.sgs.com.tw



# Report No: ES/2022/30012 Rev: 01 Page: 16 of 28

		odulation Calibration Parameters			
UID	Rev	Communication System Name	Group	PAR (dB)	Unc (k=2
0	4	CW	CW	0.00	± 4.7
10010	CAA	SAR Validation (Square, 100ms, 10ms)	Test	10.00	± 9.6 9
10011	CAB	UMTS-FDD (WCDMA)	WCDMA	2.91	± 9.6
10012	CAB	IEEE 802.11b WiFi 2.4 GHz (DSSS, 1 Mbps)	WLAN	1.87	± 9.6 9
10013	CAB	IEEE 802.11g WIFI 2.4 GHz (DSSS-OFDM, 6 Mbps)	WLAN	9.46	± 9.6 °
10021	DAC	GSM-FDD (TDMA, GMSK)	GSM	9.39	± 9.6 9
10023	DAC	GPRS-FDD (TDMA, GMSK, TN 0)	GSM	9.57	± 9.6 %
10024	DAC	GPRS-FDD (TDMA, GMSK, TN 0-1)	GSM	6.56	± 9.6 9
10025	DAC	EDGE-FDD (TDMA, 8PSK, TN 0)	GSM	12.62	± 9.6 9
10026	DAC	EDGE-FDD (TDMA, 8PSK, TN 0-1)	GSM	9.55	± 9.6 9
10027	DAC	GPRS-FDD (TDMA, GMSK, TN 0-1-2)	GSM	4.80	± 9.6 5
10028	DAC	GPRS-FDD (TDMA, GMSK, TN 0-1-2-3)	GSM	3.55	± 9.6 9
10029	DAC	EDGE-FDD (TDMA, 8PSK, TN 0-1-2)	GSM	7.78	± 9.6 9
10030	CAA	IEEE 802.15.1 Bluetooth (GFSK, DH1)	Bluetooth	5.30	± 9.6 9
10031	CAA	IEEE 802.15.1 Bluetooth (GFSK, DH3)	Bluetooth	1.87	± 9.6 %
10032	CAA	IEEE 802.15.1 Bluetooth (GFSK, DH5)	Bluetooth	1.16	± 9.6 %
10033	CAA	IEEE 802.15.1 Bluetooth (PI/4-DQPSK, DH1)	Bluetooth	7.74	± 9.6 9
10034	CAA	IEEE 802.15.1 Bluetooth (PI/4-DQPSK, DH3)	Bluetooth	4.53	± 9.6 9
10035	CAA	IEEE 802.15.1 Bluetooth (PI/4-DQPSK, DH5)	Bluetooth	3.83	± 9.6 9
10036	CAA	IEEE 802,15.1 Bluetooth (8-DPSK, DH1)	Bluetooth	8.01	± 9.6 5
10037	CAA	IEEE 802.15.1 Bluetooth (8-DPSK, DH3)	Bluetooth	4.77	± 9.6 9
10038	CAA	IEEE 802.15.1 Bluetooth (8-DPSK, DH5)	Bluetooth	4.10	± 9.6 9
10039	CAB	CDMA2000 (1xRTT, RC1)	CDMA2000	4.57	± 9.6 9
10042	CAB	IS-54 / IS-136 FDD (TDMA/FDM, PI/4-DQPSK, Halfrate)	AMPS	7.78	± 9.6
10044	CAA	IS-91/EIA/TIA-553 FDD (FDMA, FM)	AMPS	0.00	± 9.6 9
10048	CAA	DECT (TDD, TDMA/FDM, GFSK, Full Slot, 24)	DECT	13.80	± 9.6 9
10049	CAA	DECT (TDD, TDMA/FDM, GFSK, Double Slot, 12)	DECT	10.79	± 9.6 9
10056	CAA	UMTS-TDD (TD-SCDMA, 1.28 Mcps)	TD-SCDMA	11.01	± 9.6 9
10058	DAC	EDGE-FDD (TDMA, 8PSK, TN 0-1-2-3)	GSM	6.52	± 9.6 9
10059	CAB	IEEE 802.11b WIFI 2.4 GHz (DSSS, 2 Mbps)	WLAN	2.12	± 9.6 9
10060	CAB	IEEE 802.11b WiFi 2.4 GHz (DSSS, 5.5 Mbps)	WLAN	2.83	± 9.6 9
10061	CAB	IEEE 802.11b WiFi 2.4 GHz (DSSS. 11 Mbps)	WLAN	3.60	± 9.6 9
10062	CAD	IEEE 802.11a/h WIFI 5 GHz (OFDM, 6 Mbps)	WLAN	8.68	± 9.6
10063	CAD	IEEE 802.11a/h WIFi 5 GHz (OFDM, 9 Mbps)	WLAN	8.63	± 9.6 5
10064	CAD	IEEE 802.11a/h WiFi 5 GHz (OFDM, 12 Mbps)	WLAN	9.09	± 9.6 5
10065	CAD	IEEE 802.11a/h WIFI 5 GHz (OFDM, 18 Mbps)	WLAN	9.00	± 9.6 5
10066	CAD	IEEE 802.11a/h WIFi 5 GHz (OFDM, 24 Mbps)	WLAN	9.38	± 9.6
10067	CAD	IEEE 802.11a/h WIFi 5 GHz (OFDM, 36 Mbps)	WLAN	10.12	± 9.6 9
10068	CAD	IEEE 802,11a/h WIFi 5 GHz (OFDM, 48 Mbps)	WLAN	10.12	± 9.6 9
10069	CAD	IEEE 802.11a/h WiFi 5 GHz (OFDM, 54 Mbps)	WLAN	10.56	± 9.6 9
10071	CAB	IEEE 802.11g WiFI 2.4 GHz (DSSS/OFDM, 9 Mbps)	WLAN	9.83	± 9.6 9
10072	CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 12 Mbps)	WLAN	9.62	± 9.6 9
10073	CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 18 Mbps)	WLAN	9.94	± 9.6 9
10074	CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 24 Mbps)	WLAN	10.30	± 9.6 °
10075	CAB	IEEE 802.11g WiFI 2.4 GHz (DSSS/OFDM, 36 Mbps)	WLAN	10.30	± 9.6 9
10076	CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 48 Mbps)	WLAN		
10070	CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 48 Mbps)		10.94	± 9.6 9
10077	CAB	CDMA2000 (1xRTT, RC3)	WLAN	11.00	± 9.6 9
10081	CAB	IS-54 / IS-136 FDD (TDMA/FDM, PI/4-DQPSK, Fulirate)	CDMA2000	3.97	± 9.6 9
10082	DAC	GPRS-FDD (TDMA, GMSK, TN 0-4)	AMPS	4.77	± 9.6 9
			GSM	6.56	± 9.6 9
10097	CAB	UMTS-FDD (HSDPA)	WCDMA	3.98	± 9.6 °
	CAB	UMTS-FDD (HSUPA, Subtest 2)	WCDMA	3.98	± 9.6 9

Certificate No: EX3-7686\_Oct21

Page 11 of 23

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 90 days only.

t (886-2) 2299-3279

Unless other west stated the results structure the state report results in provide the structure of the state and state of the state of the state and st prosecuted to the fullest extent of the law.

No.134,Wu Kung Road, New Taipei Industrial Park, Wuku District, New Taipei City, Taiwan/新北市五股區新北產業園區五工路 134 號

f (886-2) 2298-0488

www.sgs.com.tw



# Report No: ES/2022/30012 Rev: 01 Page: 17 of 28

10100	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	5.67	± 9.6
10101	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, 16-QAM)	LTE-FDD	6.42	-
10102	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, 64-QAM)	LTE-FDD	6.60	± 9.6
10103	CAG	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-TDD	9.29	± 9.
10104	CAG	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, 16-QAM)	LTE-TDD	9.29	
10105	CAG	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, 64-QAM)	LTE-TDD	10.01	± 9.
10108	CAG	LTE-FDD (SC-FDMA, 100% RB, 10 MHz, QPSK)	LTE-FDD	5.80	
10109	CAG	LTE-FDD (SC-FDMA, 100% RB, 10 MHz, 16-QAM)	LTE-FDD	6.43	± 9.
10110	CAG	LTE-FDD (SC-FDMA, 100% RB, 5 MHz, QPSK)			± 9.
10111	CAG	LTE-FDD (SC-FDMA, 100% RB, 5 MHz, 16-QAM)	LTE-FDD	5.75	± 9.
10112	CAG	LTE-FDD (SC-FDMA, 100% RB, 10 MHz, 16-QAM)	LTE-FDD	6.44	± 9.
10112	CAG	LTE-FDD (SC-FDMA, 100% RB, 5 MHz, 64-QAM)	LTE-FDD	6.59	± 9.
10114	CAD	IEEE 802.11n (HT Greenfield, 13.5 Mbps, BPSK)	LTE-FDD	6.62	± 9.
10114	CAD	IEEE 802.11n (HT Greenfield, 13.5 Mbps, BPSK)	WLAN	8.10	± 9.0
	CAD		WLAN	8.46	± 9.
10116		IEEE 802.11n (HT Greenfield, 135 Mbps, 64-QAM)	WLAN	8.15	± 9.
10117	CAD	IEEE 802.11n (HT Mixed, 13.5 Mbps, BPSK)	WLAN	8.07	± 9.
10118		IEEE 802.11n (HT Mixed, 81 Mbps, 16-QAM)	WLAN	8.59	± 9.
10119	CAD	IEEE 802.11n (HT Mixed, 135 Mbps, 64-QAM)	WLAN	8.13	± 9.
10140	CAE	LTE-FDD (SC-FDMA, 100% RB, 15 MHz, 16-QAM)	LTE-FDD	6.49	± 9.6
10141	CAE	LTE-FDD (SC-FDMA, 100% RB, 15 MHz, 64-QAM)	LTE-FDD	6.53	± 9.
10142	CAE	LTE-FDD (SC-FDMA, 100% RB, 3 MHz, QPSK)	LTE-FDD	5.73	± 9.
10143	CAE	LTE-FDD (SC-FDMA, 100% RB, 3 MHz, 16-QAM)	LTE-FDD	6.35	± 9.
10144	CAE	LTE-FDD (SC-FDMA, 100% RB, 3 MHz, 64-QAM)	LTE-FDD	6.65	± 9.
10145	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.4 MHz, QPSK)	LTE-FDD	5.76	± 9.6
10146	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.4 MHz, 16-QAM)	LTE-FDD	6.41	± 9.
10147	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.4 MHz, 64-QAM)	LTE-FDD	6.72	± 9.
10149	CAE	LTE-FDD (SC-FDMA, 50% RB, 20 MHz, 16-QAM)	LTE-FDD	6.42	± 9.
10150	CAE	LTE-FDD (SC-FDMA, 50% RB, 20 MHz, 64-QAM)	LTE-FDD	6.60	± 9.
10151	CAG	LTE-TDD (SC-FDMA, 50% RB, 20 MHz, QPSK)	LTE-TDD	9.28	± 9.
10152	CAG	LTE-TDD (SC-FDMA, 50% RB, 20 MHz, 16-QAM)	LTE-TDD	9.92	± 9.
10153	CAG	LTE-TDD (SC-FDMA, 50% RB, 20 MHz, 64-QAM)	LTE-TDD	10.05	± 9.
10154	CAG	LTE-FDD (SC-FDMA, 50% RB, 10 MHz, QPSK)	LTE-FDD	5.75	± 9.
10155	CAG	LTE-FDD (SC-FDMA, 50% RB, 10 MHz, 16-QAM)	LTE-FDD	6.43	± 9.
10156	CAG	LTE-FDD (SC-FDMA, 50% RB, 5 MHz, QPSK)	LTE-FDD	5.79	± 9.
10157	CAG	LTE-FDD (SC-FDMA, 50% RB, 5 MHz, 16-QAM)	LTE-FDD	6.49	± 9.
10158	CAG	LTE-FDD (SC-FDMA, 50% RB, 10 MHz, 64-QAM)	LTE-FDD	6.62	± 9.0
10159	CAG	LTE-FDD (SC-FDMA, 50% RB, 5 MHz, 64-QAM)	LTE-FDD	6.56	± 9.4
10160	CAE	LTE-FDD (SC-FDMA, 50% RB, 15 MHz, OPSK)	LTE-FDD	5.82	± 9.
10161	CAE	LTE-FDD (SC-FDMA, 50% RB, 15 MHz, 16-QAM)	LTE-FDD	6.43	± 9.
10162	CAE	LTE-FDD (SC-FDMA, 50% RB, 15 MHz, 64-QAM)	LTE-FDD	6.58	± 9.
10166	CAF	LTE-FDD (SC-FDMA, 50% RB, 1.4 MHz, QPSK)	LTE-FDD	5.46	± 9.
10167	CAF	LTE-FDD (SC-FDMA, 50% RB, 1.4 MHz, 16-QAM)	LTE-FDD	6.21	± 9.
10168	CAF	LTE-FDD (SC-FDMA, 50% RB, 1.4 MHz, 64-QAM)	LTE-FDD	6.79	± 9.
10169	CAE	LTE-FDD (SC-FDMA, 1 RB, 20 MHz, QPSK)	LTE-FDD	5.73	± 9,
10109	CAE	LTE-FDD (SC-FDMA, 1 RB, 20 MHz, 16-QAM)			
10171	AAE	LTE-FDD (SC-FDMA, 1 RB, 20 MHz, 18-QAM)	LTE-FDD LTE-FDD	6.52	± 9.
10171	CAG	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, 64-GAM)		6.49	± 9.
10172	CAG	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, QPSK) LTE-TDD (SC-FDMA, 1 RB, 20 MHz, 16-QAM)	LTE-TDD	9.21	± 9.
10173	CAG	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, 16-QAM) LTE-TDD (SC-FDMA, 1 RB, 20 MHz, 64-QAM)	LTE-TDD	9.48	± 9.
	CAG		LTE-TDD	10.25	± 9.
10175		LTE-FDD (SC-FDMA, 1 RB, 10 MHz, QPSK)	LTE-FDD	5.72	± 9.
10176	CAG	LTE-FDD (SC-FDMA, 1 RB, 10 MHz, 16-QAM)	LTE-FDD	6.52	± 9.
10177	CAI	LTE-FDD (SC-FDMA, 1 RB, 5 MHz, QPSK)	LTE-FDD	5.73	± 9.
10178	CAG	LTE-FDD (SC-FDMA, 1 RB, 5 MHz, 16-QAM)	LTE-FDD	6.52	± 9.
10179	CAG	LTE-FDD (SC-FDMA, 1 RB, 10 MHz, 64-QAM)	LTE-FDD	6.50	± 9,
10180	CAG	LTE-FDD (SC-FDMA, 1 RB, 5 MHz, 64-QAM)	LTE-FDD	6.50	± 9.1
10181	CAE	LTE-FDD (SC-FDMA, 1 RB, 15 MHz, QPSK)	LTE-FDD	5.73	± 9,

Certificate No: EX3-7686\_Oct21

Page 12 of 23

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 90 days only.

t (886-2) 2299-3279

Unless otherwise stated the results shown in this test report reter only to the sample(s) tested and such sample(s) are retained tor 90 days only. Mir#JAfabity i, the state methy mining test report reter only to the sample(s) tested and such sample(s) are retained tor 90 days only. Mir#JAfabity i, the state methy mining test report reter only to the sample(s) tested and such sample(s) are retained tor 90 days only. This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <a href="http://www.sgs.com.tw/Terms-and-Conditions">http://www.sgs.com.tw/Terms-and-Conditions</a> and for electronic format documents, subject to Terms and Conditions for Electronic Documents at <a href="http://www.sgs.com.tw/Terms-and-Conditions">http://www.sgs.com.tw/Terms-and-Conditions</a> and for electronic format defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fulleest extent of the law. prosecuted to the fullest extent of the law.

No.134,Wu Kung Road, New Taipei Industrial Park, Wuku District, New Taipei City, Taiwan/新北市五股區新北產業園區五工路 134 號

f (886-2) 2298-0488

```
www.sgs.com.tw
```



# Report No: ES/2022/30012 Rev: 01 Page: 18 of 28

10182	CAE	LTE-FDD (SC-FDMA, 1 RB, 15 MHz, 16-QAM)	LTE-FDD	6.52	± 9.6
10183	AAD	LTE-FDD (SC-FDMA, 1 RB, 15 MHz, 64-QAM)	LTE-FDD	6.50	± 9.6
10184	CAE	LTE-FDD (SC-FDMA, 1 RB, 3 MHz, QPSK)	LTE-FDD	5.73	± 9.6
10185	CAE	LTE-FDD (SC-FDMA, 1 RB, 3 MHz, 16-QAM)	LTE-FDD	6.51	± 9.6
10186	AAE	LTE-FDD (SC-FDMA, 1 RB, 3 MHz, 64-QAM)	LTE-FDD	6.50	± 9.6
10187	CAF	LTE-FDD (SC-FDMA, 1 RB, 1.4 MHz, QPSK)	LTE-FDD	5.73	± 9.
10188	CAF	LTE-FDD (SC-FDMA, 1 RB, 1.4 MHz, 16-QAM)	LTE-FDD	6.52	± 9.0
10189	AAF	LTE-FDD (SC-FDMA, 1 RB, 1.4 MHz, 64-QAM)	LTE-FDD	6.50	± 9.6
10193	CAD	IEEE 802.11n (HT Greenfield, 6.5 Mbps, BPSK)	WLAN	8.09	± 9.6
10194	CAD	IEEE 802.11n (HT Greenfield, 39 Mbps, 16-QAM)	WLAN	8.12	± 9.6
10195	CAD	IEEE 802.11n (HT Greenfield, 65 Mbps, 64-QAM)	WLAN	8.21	± 9.0
10196	CAD	IEEE 802.11n (HT Mixed, 6.5 Mbps, BPSK)	WLAN	8.10	± 9.0
10197	CAD	IEEE 802.11n (HT Mixed. 39 Mbps, 16-QAM)	WLAN	8.13	± 9.6
10198	CAD	IEEE 802.11n (HT Mixed, 65 Mbps, 64-QAM)	WLAN	8.27	± 9.0
10219	CAD	IEEE 802.11n (HT Mixed, 7.2 Mbps, BPSK)	WLAN	8.03	± 9.0
10220	CAD	IEEE 802.11n (HT Mixed, 43.3 Mbps, 16-QAM)			
10221	CAD	IEEE 802.11n (HT Mixed, 43.3 Mbps, 1940AW)	WLAN	8.13	± 9.6
10221	CAD	IEEE 802.11n (HT Mixed, 12 Mbps, 64-QAM)	WLAN	8.27	± 9.6
10222	CAD	IEEE 802.11n (HT Mixed, 90 Mbps, 16-QAM)	WLAN	8.06	± 9,6
10223	CAD	IEEE 802.11n (HT Mixed, 50 Mbps, 64-QAM)	WLAN	8.48	± 9.6
10224	CAB	UMTS-FDD (HSPA+)	WLAN	8.08	± 9.6
10225	CAB	LTE-TDD (SC-FDMA, 1 RB, 1.4 MHz, 16-QAM)	WCDMA	5.97	± 9.6
10226	CAB	LTE-TDD (SC-FDMA, 1 RB, 1.4 MHz, 16-QAM) LTE-TDD (SC-FDMA, 1 RB, 1.4 MHz, 64-QAM)	LTE-TDD	9.49	± 9.6
10227	CAB	LTE-TDD (SC-FDMA, 1 RB, 1.4 MHz, 64-QAM) LTE-TDD (SC-FDMA, 1 RB, 1.4 MHz, QPSK)	LTE-TDD	10.26	± 9.6
10228	CAB	LTE-TDD (SC-FDMA, 1 RB, 1.4 MHz, QPSK) LTE-TDD (SC-FDMA, 1 RB, 3 MHz, 16-QAM)	LTE-TDD	9.22	± 9.6
10229	CAD	LTE-TDD (SC-FDMA, 1 RB, 3 MHz, 16-QAM) LTE-TDD (SC-FDMA, 1 RB, 3 MHz, 64-QAM)	LTE-TDD	9.48	± 9.
10230	CAD	LTE-TDD (SC-FDMA, 1 RB, 3 MHz, 04-QAM)	LTE-TDD	10.25	± 9.
10231	CAG		LTE-TDD	9.19	± 9.
10232	CAG	LTE-TDD (SC-FDMA, 1 RB, 5 MHz, 18-QAM) LTE-TDD (SC-FDMA, 1 RB, 5 MHz, 64-QAM)	LTE-TDD	9.48	± 9.
	CAG		LTE-TDD	10.25	± 9.
10234	CAG	LTE-TDD (SC-FDMA, 1 RB, 5 MHz, QPSK) LTE-TDD (SC-FDMA, 1 RB, 10 MHz, 16-QAM)	LTE-TDD	9.21	± 9.6
10235	CAG		LTE-TDD	9.48	± 9.0
10230	CAG	LTE-TDD (SC-FDMA, 1 RB, 10 MHz, 64-QAM) LTE-TDD (SC-FDMA, 1 RB, 10 MHz, QPSK)	LTE-TDD	10.25	± 9.0
10237	CAG		LTE-TDD	9.21	± 9.
10230	CAF	LTE-TDD (SC-FDMA, 1 RB, 15 MHz, 16-QAM)	LTE-TDD	9.48	± 9.
10239	CAF	LTE-TDD (SC-FDMA, 1 RB, 15 MHz, 64-QAM) LTE-TDD (SC-FDMA, 1 RB, 15 MHz, QPSK)	LTE-TDD	10.25	± 9.6
10240	CAP	LTE-TDD (SC-FDMA, 1 KB, 15 MHz, QPSK) LTE-TDD (SC-FDMA, 50% RB, 1.4 MHz, 16-QAM)	LTE-TDD	9.21	± 9.
10241	CAB	LTE-TDD (SC-FDMA, 50% RB, 1,4 MHz, 16-QAM)	LTE-TDD	9.82	± 9.
10242	CAB	LTE-TDD (SC-FDMA, 50% RB, 1,4 MHz, 64-QAM) LTE-TDD (SC-FDMA, 50% RB, 1,4 MHz, QPSK)	LTE-TDD	9.86	± 9.6
10243	CAD	LTE-TDD (SC-FDMA, 50% RB, 1.4 MHz, QPSK) LTE-TDD (SC-FDMA, 50% RB, 3 MHz, 16-QAM)	LTE-TDD	9,46	± 9.6
10244	CAD	LTE-TDD (SC-FDMA, 50% RB, 3 MHz, 16-QAM) LTE-TDD (SC-FDMA, 50% RB, 3 MHz, 64-QAM)	LTE-TDD	10.06	± 9.1
10245	CAD	LTE-TDD (SC-FDMA, 50% RB, 3 MHZ, 64-QAM) LTE-TDD (SC-FDMA, 50% RB, 3 MHZ, QPSK)	LTE-TDD	10.06	± 9.
10246	CAG	LTE-TDD (SC-FDMA, 50% RB, 5 MHz, 0FSK)	LTE-TDD	9.30	± 9.6
10247	CAG	LTE-TDD (SC-FDMA, 50% RB, 5 MHz, 16-QAM) LTE-TDD (SC-FDMA, 50% RB, 5 MHz, 64-QAM)	LTE-TDD	9.91	± 9.
10248	CAG	LTE-TDD (SC-FDMA, 50% RB, 5 MHz, 64-QAM) LTE-TDD (SC-FDMA, 50% RB, 5 MHz, QPSK)	LTE-TDD	10.09	± 9.
10249	CAG	LTE-TDD (SC-FDMA, 50% RB, 5 MHz, QPSK) LTE-TDD (SC-FDMA, 50% RB, 10 MHz, 16-QAM)	LTE-TDD	9.29	± 9.
10250	CAG	LTE-TDD (SC-FDMA, 50% RB, 10 MHz, 16-QAM) LTE-TDD (SC-FDMA, 50% RB, 10 MHz, 64-QAM)	LTE-TDD	9.81	± 9.
10251	CAG	LTE-TDD (SC-FDMA, 50% RB, 10 MHZ, 84-QAM) LTE-TDD (SC-FDMA, 50% RB, 10 MHZ, QPSK)	LTE-TDD	10.17	± 9.
	CAG		LTE-TDD	9.24	± 9.
10253	CAF	LTE-TDD (SC-FDMA, 50% RB, 15 MHz, 16-QAM)	LTE-TDD	9.90	± 9.
10254		LTE-TDD (SC-FDMA, 50% RB, 15 MHz, 64-QAM)	LTE-TDD	10.14	± 9.
10255	CAF	LTE-TDD (SC-FDMA, 50% RB, 15 MHz, QPSK)	LTE-TDD	9.20	± 9.
10256	CAB	LTE-TDD (SC-FDMA, 100% RB, 1.4 MHz, 16-QAM)	LTE-TDD	9.96	± 9.
10257	CAB	LTE-TDD (SC-FDMA, 100% RB, 1.4 MHz, 64-QAM)	LTE-TDD	10.08	± 9.
10258	CAB	LTE-TDD (SC-FDMA, 100% RB, 1,4 MHz, QPSK)	LTE-TDD	9.34	± 9,
10259	CAD	LTE-TDD (SC-FDMA, 100% RB, 3 MHz, 16-QAM)	LTE-TDD	9.98	± 9.
10260	CAD	LTE-TDD (SC-FDMA, 100% RB, 3 MHz, 64-QAM)	LTE-TDD	9,97	± 9.

Certificate No: EX3-7686\_Oct21

Page 13 of 23

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 90 days only.

除非另有說明,此報告結果僅對測試之樣品負責,同時此樣品僅保留90天。本報告未經本公司書面許可,不可部份複製。 This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <a href="http://www.sgs.com.tw/Terms-and-Conditions">http://www.sgs.com.tw/Terms-and-Conditions</a> and for electronic format documents subject to Terms and Conditions for Electronic Documents at <u>http://www.sgs.com.tw/Terms-and-Conditions</u>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

No.134,Wu Kung Road, New Taipei Industrial Park, Wuku District, New Taipei City, Taiwan/新北市五股區新北產業園區五工路 134 號

f (886-2) 2298-0488

www.sgs.com.tw



# Report No: ES/2022/30012 Rev: 01 Page: 19 of 28

10261	CAD	LTE-TDD (SC-FDMA, 100% RB, 3 MHz, QPSK)	LTE-TDD	9.24	1.00
10262	CAG	LTE-TDD (SC-FDMA, 100% RB, 5 MHz, 16-QAM)	LTE-TDD	9.24	± 9.6
10263	CAG	LTE-TDD (SC-FDMA, 100% RB, 5 MHz, 64-QAM)	LTE-TDD		± 9.6
10264	CAG	LTE-TDD (SC-FDMA, 100% RB, 5 MHz, QPSK)	LTE-TDD	9,23	± 9.6
10265	CAG	LTE-TDD (SC-FDMA, 100% RB, 10 MHz, 16-OAM)	LTE-TDD	9.23	± 9.6
10266	CAG	LTE-TDD (SC-FDMA, 100% RB, 10 MHz, 64-QAM)	LTE-TDD	10.07	± 9.6
10267	CAG	LTE-TDD (SC-FDMA, 100% RB, 10 MHz, QPSK)	LTE-TDD	9.30	± 9.6 ± 9.6
10268	CAF	LTE-TDD (SC-FDMA, 100% RB, 15 MHz, 16-QAM)	LTE-TDD	10.06	± 9.6
10269	CAF	LTE-TDD (SC-FDMA, 100% RB, 15 MHz, 64-QAM)	LTE-TDD	10.08	± 9.6
10270	CAF	LTE-TDD (SC-FDMA, 100% RB, 15 MHz, QPSK)	LTE-TDD	9.58	± 9.6
10274	CAB	UMTS-FDD (HSUPA, Subtest 5, 3GPP Rel8.10)	WCDMA	4.87	± 9.6
10275	CAB	UMTS-FDD (HSUPA, Subtest 5, 3GPP Rel8.4)	WCDMA	3.96	± 9.6
10277	CAA	PHS (QPSK)	PHS	11.81	± 9.6
10278	CAA	PHS (QPSK, BW 884MHz, Rolloff 0.5)	PHS	11.81	± 9.6
10279	CAA	PHS (QPSK, BW 884MHz, Rolloff 0.38)	PHS	12.18	± 9.6
10290	AAB	CDMA2000, RC1, SO55, Full Rate	CDMA2000	3.91	± 9.6
10291	AAB	CDMA2000, RC3, SO55, Full Rate	CDMA2000	3.46	± 9.6
10292	AAB	CDMA2000, RC3, SO32, Full Rate	CDMA2000	3.39	± 9.6
10293	AAB	CDMA2000, RC3, SO3, Full Rate	CDMA2000	3.50	± 9.6
10295	AAB	CDMA2000, RC1, SO3, 1/8th Rate 25 fr.	CDMA2000	12.49	± 9.6
10297	AAD	LTE-FDD (SC-FDMA, 50% RB, 20 MHz, QPSK)	LTE-FDD	5.81	± 9.6
10298	AAD	LTE-FDD (SC-FDMA, 50% RB, 3 MHz, QPSK)	LTE-FDD	5.72	± 9.6
10299	AAD	LTE-FDD (SC-FDMA, 50% RB, 3 MHz, 16-QAM)	LTE-FDD	6.39	± 9.6
10300	AAD	LTE-FDD (SC-FDMA, 50% RB, 3 MHz, 64-QAM)	LTE-FDD	6.60	± 9.6
10301	AAA	IEEE 802.16e WIMAX (29:18, 5ms, 10MHz, QPSK, PUSC)	WIMAX	12.03	± 9.6
10302	AAA	IEEE 802.16e WIMAX (29:18, 5ms, 10MHz, QPSK, PUSC, 3CTRL)	WIMAX	12.03	± 9.6
10303	AAA	IEEE 802.16e WIMAX (31:15, 5ms, 10MHz, 64QAM, PUSC)	WIMAX	12.57	± 9.6
10304	AAA	IEEE 802.16e WIMAX (29:18, 5ms, 10MHz, 64QAM, PUSC)	WIMAX	12.52	
10305	AAA	IEEE 802.16e WIMAX (31:15, 10ms, 10MHz, 64QAM, PUSC)	WIMAX	15.24	± 9.6
10306	AAA	IEEE 802.16e WIMAX (29:18, 10ms, 10MHz, 64QAM, PUSC)	WIMAX	14.67	± 9.6 ± 9.6
10307	AAA	IEEE 802.16e WIMAX (29:18, 10ms, 10MHz, QPSK, PUSC)	WIMAX	14.67	± 9.6
10308	AAA	IEEE 802.16e WIMAX (29:18, 10ms, 10MHz, 16QAM, PUSC)	WIMAX	14.49	± 9.6
10309	AAA	IEEE 802.16e WIMAX (29:18, 10ms, 10MHz, 16QAM,AMC 2x3)	WIMAX	14.40	± 9.6
10310	AAA	IEEE 802.16e WIMAX (29:18, 10ms, 10MHz, QPSK, AMC 2x3)	WIMAX	14.50	± 9.6
10311	AAD	LTE-FDD (SC-FDMA, 100% RB, 15 MHz, QPSK)	LTE-FDD	6.06	± 9.6
10313	AAA	IDEN 1:3	IDEN	10.51	± 9.6
10314	AAA	IDEN 1:6	IDEN		
10315	AAB	IEEE 802.11b WIFI 2.4 GHz (DSSS, 1 Mbps, 96pc dc)	WLAN	13.48	± 9.6 ± 9.6
10316	AAB	IEEE 802.11g WiFi 2.4 GHz (ERP-OFDM, 6 Mbps, 96pc dc)	WLAN	8.36	± 9.6
10317	AAD	IEEE 802,11a WIFI 5 GHz (OFDM, 6 Mbps, 96pc dc)	WLAN	8.36	± 9.6
10352	AAA	Pulse Waveform (200Hz, 10%)	Generic	10.00	± 9.6
10353	AAA	Pulse Waveform (200Hz, 20%)	Generic	6.99	± 9.6
10354	AAA	Pulse Waveform (200Hz, 40%)	Generic	3.98	± 9.6
10355	AAA	Pulse Waveform (200Hz, 60%)	Generic	2.22	± 9.6
10356	AAA	Pulse Waveform (200Hz, 80%)	Generic	0.97	± 9.6
10387	AAA	QPSK Waveform, 1 MHz	Generic	5.10	± 9.6
10388	AAA	QPSK Waveform, 10 MHz	Generic	5.10	± 9.6
10396	AAA	64-QAM Waveform, 100 kHz	Generic	6.27	± 9.6
10399	AAA	64-QAM Waveform, 40 MHz	Generic	6.27	± 9.6
10400	AAE	IEEE 802.11ac WiFi (20MHz, 64-QAM, 99pc dc)	WLAN	8.37	± 9.6
10400	AAE	IEEE 802.11ac WiFi (200Hz, 64-QAM, 99pc dc)	WLAN	8.37	± 9.6
	AAE	IEEE 802.11ac WIFI (80MHz, 64-QAM, 99pc dc)	WLAN	8.60	± 9.6
10402	AAB	CDMA2000 (1xEV-DO, Rev. 0)			
10403	AAB	CDMA2000 (1xEV-DO, Rev. d)	CDMA2000	3.76	± 9.6
10404	AAB	CDMA2000 (TAE V-DO, Rev. A) CDMA2000, RC3, SO32, SCH0, Full Rate	CDMA2000	3.77	±9.6
		ourin acout, rico, court, corro, rui Rate	CDMA2000	5.22	± 9.6

Certificate No: EX3-7686\_Oct21

Page 14 of 23

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 90 days only.

t (886-2) 2299-3279

Unless other west stated the results structure the state report results in provide the structure of the state and state of the state of the state and st prosecuted to the fullest extent of the law.

No.134,Wu Kung Road, New Taipei Industrial Park, Wuku District, New Taipei City, Taiwan/新北市五股區新北產業園區五工路 134 號

f (886-2) 2298-0488

www.sqs.com.tw



# Report No: ES/2022/30012 Rev: 01 Page: 20 of 28

10414	AAA	WLAN CCDF, 64-QAM, 40MHz	Generic	8.54	± 9.6
10415	AAA	IEEE 802.11b WIFI 2.4 GHz (DSSS, 1 Mbps, 99pc dc)	WLAN	1.54	± 9.
10416	AAA	IEEE 802,11g WiFi 2.4 GHz (ERP-OFDM, 6 Mbps, 99pc dc)	WLAN	8.23	± 9.
10417	AAC.	IEEE 802,11a/h WIFI 5 GHz (OFDM, 6 Mbps, 99pc dc)	WLAN	8.23	± 9.
10418	AAA	IEEE 802.11g WIFI 2.4 GHz (DSSS-OFDM, 6 Mbps, 99pc, Long)	WLAN	8.14	± 9.
10419	AAA	IEEE 802.11g WIFI 2.4 GHz (DSSS-OFDM, 6 Mbps, 99pc, Short)	WLAN	8.19	± 9.
10422	AAC	IEEE 802.11n (HT Greenfield, 7.2 Mbps, BPSK)	WLAN	8.32	± 9.
10423	AAC	IEEE 802.11n (HT Greenfield, 43.3 Mbps, 16-QAM)	WLAN	8.47	± 9.
10424	AAC	IEEE 802.11n (HT Greenfield, 72.2 Mbps, 64-QAM)	WLAN	8.40	± 9.1
10425	AAC	IEEE 802.11n (HT Greenfield, 15 Mbps, BPSK)	WLAN	8.41	± 9.
10426	AAC	IEEE 802.11n (HT Greenfield, 90 Mbps, 16-QAM)	WLAN	8.45	± 9.
10427	AAC	IEEE 802.11n (HT Greenfield, 150 Mbps, 64-QAM)	WLAN	8.41	± 9.
10430	AAD	LTE-FDD (OFDMA, 5 MHz, E-TM 3.1)	LTE-FDD	8.28	± 9.
10431	AAD	LTE-FDD (OFDMA, 10 MHz, E-TM 3.1)	LTE-FDD	8.38	± 9.
10432	AAC	LTE-FDD (OFDMA, 15 MHz, E-TM 3.1)	LTE-FDD	8.34	± 9.
10433	AAC	LTE-FDD (OFDMA, 20 MHz, E-TM 3,1)	LTE-FDD	8.34	± 9.
10434	AAA	W-CDMA (BS Test Model 1, 64 DPCH)	WCDMA	8.60	± 9.
10435	AAF	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, QPSK, UL Sub)	LTE-TDD	7.82	± 9.
10447	AAD	LTE-FDD (OFDMA, 5 MHz, E-TM 3.1, Clipping 44%)	LTE-FDD	7.56	± 9. ± 9.
10448	AAD	LTE-FDD (OFDMA, 10 MHz, E-TM 3.1, Clippin 44%)	LTE-FDD	7.53	± 9.
10449	AAC	LTE-FDD (OFDMA, 15 MHz, E-TM 3.1, Clipting 44%)	LTE-FDD	7.51	± 9.
10445	AAC	LTE-FDD (OFDMA, 20 MHz, E-TM 3.1, Clipping 44%)			
10451	AAA	W-CDMA (BS Test Model 1, 64 DPCH, Clipping 44%)	LTE-FDD	7.48	±9.
10451	AAD	Validation (Square, 10ms, 1ms)	WCDMA	7.59	± 9.
10456	AAC	IEEE 802.11ac WiFi (160MHz, 64-QAM, 99pc dc)	Test	10.00	± 9.
10450	AAA	UMTS-FDD (DC-HSDPA)	WLAN	8.63	± 9.
10457	AAA	CDMA2000 (1xEV-DO, Rev. B, 2 carriers)	WCDMA	6.62	± 9.
10458	AAA		CDMA2000	6.55	± 9.
10459	AAA	CDMA2000 (1xEV-DO, Rev. B, 3 carriers)	CDMA2000	8.25	± 9.
		UMTS-FDD (WCDMA, AMR)	WCDMA	2.39	± 9.
10461	AAB	LTE-TDD (SC-FDMA, 1 RB, 1.4 MHz, QPSK, UL Sub)	LTE-TDD	7.82	± 9.
10462	AAB	LTE-TDD (SC-FDMA, 1 RB, 1.4 MHz, 16-QAM, UL Sub)	LTE-TDD	8.30	± 9.
10463	AAB	LTE-TDD (SC-FDMA, 1 RB, 1.4 MHz, 64-QAM, UL Sub)	LTE-TDD	8.56	± 9.
10464	AAC	LTE-TDD (SC-FDMA, 1 RB, 3 MHz, QPSK, UL Sub)	LTE-TDD	7.82	± 9.
10465	AAC	LTE-TDD (SC-FDMA, 1 RB, 3 MHz, 16-QAM, UL Sub)	LTE-TDD	8.32	± 9.
10466	AAC	LTE-TDD (SC-FDMA, 1 RB, 3 MHz, 64-QAM, UL Sub)	LTE-TDD	8.57	± 9.
10467	AAF	LTE-TDD (SC-FDMA, 1 RB, 5 MHz, QPSK, UL Sub)	LTE-TDD	7.82	± 9.
10468	AAF	LTE-TDD (SC-FDMA, 1 RB, 5 MHz, 16-QAM, UL Sub)	LTE-TDD	8.32	± 9.
10469	AAF	LTE-TDD (SC-FDMA, 1 RB, 5 MHz, 64-QAM, UL Sub)	LTE-TDD	8.56	± 9.
10470	AAF	LTE-TDD (SC-FDMA, 1 RB, 10 MHz, QPSK, UL Sub)	LTE-TDD	7.82	± 9.
10471	AAF	LTE-TDD (SC-FDMA, 1 RB, 10 MHz, 16-QAM, UL Sub)	LTE-TDD	8.32	± 9.
10472	AAF	LTE-TDD (SC-FDMA, 1 RB, 10 MHz, 64-QAM, UL Sub)	LTE-TDD	8.57	± 9.
10473	AAE	LTE-TDD (SC-FDMA, 1 RB, 15 MHz, QPSK, UL Sub)	LTE-TDD	7.82	± 9.
10474	AAE	LTE-TDD (SC-FDMA, 1 RB, 15 MHz, 16-QAM, UL Sub)	LTE-TDD	8.32	± 9.
10475	AAE	LTE-TDD (SC-FDMA, 1 RB, 15 MHz, 64-QAM, UL Sub)	LTE-TDD	8.57	± 9.
10477	AAF	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, 16-QAM, UL Sub)	LTE-TDD	8.32	± 9.
10478	AAF	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, 64-QAM, UL Sub)	LTE-TDD	8.57	± 9.
10479	AAB	LTE-TDD (SC-FDMA, 50% RB, 1.4 MHz, QPSK, UL Sub)	LTE-TDD	7.74	± 9.
10480	AAB	LTE-TDD (SC-FDMA, 50% RB, 1.4 MHz, 16-QAM, UL Sub)	LTE-TDD	8.18	± 9.
10481	AAB	LTE-TDD (SC-FDMA, 50% RB, 1.4 MHz, 64-QAM, UL Sub)	LTE-TDD	8.45	± 9.
10482	AAC	LTE-TDD (SC-FDMA, 50% RB, 3 MHz, QPSK, UL Sub)	LTE-TDD	7.71	± 9.
10483	AAC	LTE-TDD (SC-FDMA, 50% RB, 3 MHz, 18-QAM, Sub)	LTE-TDD	8.39	± 9.
10484	AAC	LTE-TDD (SC-FDMA, 50% RB, 3 MHz, 64-QAM, UL Sub)	LTE-TDD	8.47	± 9.
10485	AAF	LTE-TDD (SC-FDMA, 50% RB, 5 MHz, QPSK, UL Sub)	LTE-TDD	7.59	± 9.
10486	AAF	LTE-TDD (SC-FDMA, 50% RB, 5 MHz, 16-QAM, UL Sub)	LTE-TDD	8.38	± 9.
10487	AAF	LTE-TDD (SC-FDMA, 50% RB, 5 MHz, 64-QAM, UL Sub)	LTE-TDD	8.60	± 9.
10488	AAF	LTE-TDD (SC-FDMA, 50% RB, 10 MHz, QPSK, UL Sub)	LTE-TDD	7.70	± 9.

Certificate No: EX3-7686\_Oct21

Page 15 of 23

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 90 days only.

t (886-2) 2299-3279

Unless other west stated the results structure the state report results in provide the structure of the state and state of the state of the state and st prosecuted to the fullest extent of the law.

No.134,Wu Kung Road, New Taipei Industrial Park, Wuku District, New Taipei City, Taiwan/新北市五股區新北產業園區五工路 134 號

f (886-2) 2298-0488

www.sgs.com.tw



# Report No: ES/2022/30012 Rev: 01 Page: 21 of 28

10489	AAF	LTE-TDD (SC-FDMA, 50% RB, 10 MHz, 16-QAM, UL Sub)	LTE-TDD	8.31	± 9.6
10490	AAF	LTE-TDD (SC-FDMA, 50% RB, 10 MHz, 64-QAM, UL Sub)	LTE-TDD	8.54	
10491	AAE	LTE-TDD (SC-FDMA, 50% RB, 15 MHz, QPSK, UL Sub)	LTE-TDD	7,74	± 9.6
10492	AAE	LTE-TDD (SC-FDMA, 50% RB, 15 MHz, 16-QAM, UL Sub)	LTE-TDD	8.41	± 9.6
10493	AAE	LTE-TDD (SC-FDMA, 50% RB, 15 MHz, 64-QAM, UL Sub)	LTE-TDD	8.55	± 9.6
10494	AAF	LTE-TDD (SC-FDMA, 50% RB, 20 MHz, QPSK, UL Sub)	LTE-TDD	7.74	± 9.6
10495	AAF	LTE-TDD (SC-FDMA, 50% RB, 20 MHz, 16-QAM, UL Sub)	LTE-TDD	8.37	± 9.6
10496	AAF	LTE-TDD (SC-FDMA, 50% RB, 20 MHz, 64-QAM, UL Sub)	LTE-TDD	8.54	± 9.6
10497	AAB	LTE-TDD (SC-FDMA, 100% RB, 1.4 MHz, QPSK, UL Sub)	LTE-TDD	7.67	± 9.6
10498	AAB	LTE-TDD (SC-FDMA, 100% RB, 1.4 MHz, 16-QAM, UL Sub)	LTE-TDD	8.40	± 9.6
10499	AAB	LTE-TDD (SC-FDMA, 100% RB, 1.4 MHz, 64-QAM, UL Sub)	LTE-TDD	8.68	± 9.6
10500	AAC	LTE-TDD (SC-FDMA, 100% RB, 3 MHz, QPSK, UL Sub)	LTE-TDD	7.67	± 9.6
10501	AAC	LTE-TDD (SC-FDMA, 100% RB, 3 MHz, 16-QAM, UL Sub)	LTE-TDD	8.44	± 9.6
10502	AAC	LTE-TDD (SC-FDMA, 100% RB, 3 MHz, 64-QAM, UL Sub)	LTE-TDD	8.52	± 9.6
10503	AAF	LTE-TDD (SC-FDMA, 100% RB, 5 MHz, QPSK, UL Sub)	LTE-TDD	7.72	± 9.6
10504	AAF	LTE-TDD (SC-FDMA, 100% RB, 5 MHz, 16-QAM, UL Sub)	LTE-TDD	8.31	± 9.6
10505	AAF	LTE-TDD (SC-FDMA, 100% RB, 5 MHz, 64-QAM, UL Sub)	LTE-TDD	8.54	± 9.6
10506	AAF	LTE-TDD (SC-FDMA, 100% RB, 10 MHz, QPSK, UL Sub)			
10507	AAF	LTE-TDD (SC-FDMA, 100% RB, 10 MHz, 16-QAM, UL Sub)	LTE-TDD	7.74	± 9.6
10508	AAF	LTE-TDD (SC-FDMA, 100% RB, 10 MHz, 10-QAM, 0E SUB)	LTE-TDD	8.35	± 9.6
10509	AAE	LTE-TDD (SC-FDMA, 100% RB, 15 MHz, QPSK, UL Sub)	LTE-TOD	7.99	
10510	AAE	LTE-TDD (SC-FDMA, 100% RB, 15 MHz, 16-QAM, UL Sub)	LTE-TDD	8.49	± 9.6
10511	AAE	LTE-TDD (SC-FDMA, 100% RB, 15 MHz, 64-QAM, UL Sub)			± 9.6
10512	AAF	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, QPSK, UL Sub)	LTE-TDD LTE-TDD	8.51	± 9.6
10513	AAF	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, 16-QAM, UL Sub)			± 9.6
10514	AAF	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, 10-QAM, 0L Sub)	LTE-TDD	8.42	± 9.6
10515	AAA	IEEE 802.11b WiFi 2.4 GHz (DSSS, 2 Mbps, 99pc dc)	WLAN	8.45	± 9.6
10516	AAA	IEEE 802.11b WiFi 2.4 GHz (DSSS, 5.5 Mbps, 99pc dc)	WLAN	1.58	± 9.6
10517	AAA	IEEE 802.11b WIFI 2.4 GHz (DSSS, 3.5 Mbps, 99pc dc)	WLAN		± 9.6
10518	AAC	IEEE 802.11a/h WIFI 5 GHz (DS33, 11 Mibps, 99pc dc)		1.58	± 9.6
10519	AAC	IEEE 802.11a/h WiFi 5 GHz (OFDM, 12 Mbps, 99pc dc)	WLAN	8.23	± 9.6
10520	AAC	IEEE 802.11a/h WiFi 5 GHz (OFDM, 12 Mbps, 99pc dc)		8.39	± 9.6
10521	AAC	IEEE 802.11a/h WiFi 5 GHz (OFDM, 18 Mbps, 99pc dc)	WLAN	8.12	± 9.6
10522	AAC	IEEE 802.11a/n WIFI'S GHz (OFDM, 24 Mbps, 99pc dc)	WLAN	7.97	± 9.6
10522	AAC	IEEE 802.11a/h WiFi 5 GHz (OFDM, 36 Mbps, 99pc dc)		8.45	± 9.6
10523	AAC	IEEE 802.11a/h WiFi 5 GHz (OFDM, 48 Mbps, 99pc dc)	WLAN	8.08	± 9.6
10524	AAC	IEEE 802.11a/r WiFI'S GH2 (OFDM, 54 Mbps, 99pc dc)	WLAN	8.27	± 9.6
10525	AAC	IEEE 802.11ac WiFI (20MHz, MCS1, 99pc dc)	WLAN	8.36	± 9.6
10520	AAC	IEEE 802.11ac WiFI (20MHz, MCS1, 99pc dc)	WLAN	8.42	± 9.6
10527	AAC	IEEE 802.11ac WiFi (20MHz, MCS2, 99pc dc)	WLAN	8.21	± 9.6
10528	AAC	IEEE 802.11ac WiFI (20MHz, MCS3, 99pc dc) IEEE 802.11ac WiFI (20MHz, MCS4, 99pc dc)	WLAN	8.36	± 9.6
10529	AAC	IEEE 802.11ac WIFI (20MHz, MCS4, 99pc dc) IEEE 802.11ac WIFI (20MHz, MCS6, 99pc dc)	WLAN	8.36	± 9.6
10531	AAC	IEEE 802.11ac WiFi (20MHz, MCS6, 99pc dc) IEEE 802.11ac WiFi (20MHz, MCS7, 99pc dc)	WLAN	8.43	± 9.6
10532	AAC	IEEE 802.11ac WIFI (20MHz, MCS7, 99pc dc) IEEE 802.11ac WIFI (20MHz, MCS8, 99pc dc)	WLAN	8.29	± 9.6
10533	AAC	IEEE 802.11ac WiFi (20MHz, MCS8, 99pc dc) IEEE 802.11ac WiFi (40MHz, MCS0, 99pc dc)	WLAN	8.38	± 9.6
10534	AAC		WLAN	8.45	± 9.6
		IEEE 802.11ac WiFi (40MHz, MCS1, 99pc dc)	WLAN	B.45	± 9.6
10536	AAC	IEEE 802.11ac WiFi (40MHz, MCS2, 99pc dc)	WLAN	8.32	± 9.6
10537	AAC	IEEE 802.11ac WiFi (40MHz, MCS3, 99pc dc)	WLAN	8.44	± 9.6
10538	AAC	IEEE 802.11ac WiFi (40MHz, MCS4, 99pc dc)	WLAN	8.54	± 9.6
10540	AAC	IEEE 802.11ac WiFi (40MHz, MCS6, 99pc dc)	WLAN	8.39	± 9.6
10541	AAC	IEEE 802.11ac WiFi (40MHz, MCS7, 99pc dc)	WLAN	8.46	± 9.6
10542	AAC	IEEE 802.11ac WiFi (40MHz, MCS8, 99pc dc)	WLAN	8.65	± 9.6
10543	AAC	IEEE 802.11ac WiFi (40MHz, MCS9, 99pc dc)	WLAN	8.65	± 9.6
10544	AAC	IEEE 802.11ac WiFi (80MHz, MCS0, 99pc dc)	WLAN	8.47	± 9.6
10545	AAC	IEEE 802.11ac WiFI (80MHz, MCS1, 99pc dc)	WLAN	8.55	± 9.6
10546	AAC	IEEE 802.11ac WiFi (80MHz, MCS2, 99pc dc)	WLAN	8.35	± 9.6

Certificate No: EX3-7686\_Oct21

Page 16 of 23

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 90 days only.

t (886-2) 2299-3279

Unless other west stated the results structure the state report results in provide the structure of the state and state of the state of the state and st prosecuted to the fullest extent of the law.

No.134,Wu Kung Road, New Taipei Industrial Park, Wuku District, New Taipei City, Taiwan/新北市五股區新北產業園區五工路 134 號

f (886-2) 2298-0488

www.sqs.com.tw



# Report No: ES/2022/30012 Rev: 01 Page: 22 of 28

10547	AAC	IEEE 802.11ac WiFi (80MHz, MCS3, 99pc dc)	WLAN	8.49	± 9.6
10548	AAC	IEEE 802.11ac WiFi (80MHz, MCS4, 99pc dc)	WLAN	8.37	± 9.6
10550	AAC	IEEE 802.11ac WiFi (80MHz, MCS6, 99pc dc)	WLAN	8.39	± 9.6
10551	AAC	IEEE 802.11ac WIFI (80MHz, MCS7, 99pc dc)	WLAN	8.50	± 9.6
10552	AAC	IEEE 802.11ac WiFi (80MHz, MCS8, 99pc dc)	WLAN	8.42	± 9.6
10553	AAC	IEEE 802.11ac WiFI (80MHz, MCS9, 99pc dc)	WLAN	8.45	± 9.6
10554	AAD	IEEE 802.11ac WiFi (160MHz, MCS0, 99pc dc)	WLAN	8.48	± 9.6
10555	AAD	IEEE 802.11ac WiFi (160MHz, MCS1, 99pc dc)	WLAN	8.47	± 9.6
10556	AAD	IEEE 802.11ac WiFi (160MHz, MCS2, 99pc dc)	WLAN	8.50	± 9.6
10557	AAD	IEEE 802.11ac WiFI (160MHz, MCS3, 99pc dc)	WLAN	8.52	± 9.6
10558	AAD	IEEE 802.11ac WIFI (160MHz, MCS4, 99pc dc)	WLAN	8.61	± 9.6
10560	AAD	IEEE 802.11ac WiFi (160MHz, MCS6, 99pc dc)	WLAN	8.73	± 9.6
10561	AAD	IEEE 802.11ac WiFi (160MHz, MCS7, 99pc dc)	WLAN	8.56	± 9.6
10562	AAD	IEEE 802.11ac WiFi (160MHz, MCS8, 99pc dc)	WLAN	8.69	± 9.6
10563	AAD	IEEE 802.11ac WIFi (160MHz, MCS9, 99pc dc)	WLAN	8.77	± 9.6
10564	AAA	IEEE 802.11g WiFI 2.4 GHz (DSSS-OFDM, 9 Mbps, 99pc dc)	WLAN	8.25	± 9.6
10565	AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 5 Mbps, 56pc dc)	WLAN	8.45	± 9.6
10566	AAA	IEEE 802.11g WIFI 2.4 GHz (DSSS-OFDM, 12 Mbps, 99pc dc)	WLAN	8.13	± 9.6
10567	AAA	IEEE 802.11g WIFI 2.4 GHz (DSSS-OFDM, 18 Mbps, 99pc dc)	WLAN	8.00	± 9.6
10568	AAA	IEEE 802.11g WIFI 2.4 GHz (DSSS-OFDM, 36 Mbps, 99pc dc)	WLAN	8.37	± 9.6
10569	AAA	IEEE 802.11g WIFI 2.4 GHz (DSSS-OFDM, 48 Mbps, 99pc dc)	WLAN	8.10	± 9.6
10570	AAA	IEEE 802.11g WIFI 2.4 GHz (DSSS-OFDM, 48 Mbps, 39pc dc)	WLAN	8.30	± 9.6
10571	AAA	IEEE 802.11b WIFi 2.4 GHz (DSSS, 1 Mbps, 90pc dc)	WLAN	1.99	± 9.6
10572	AAA	IEEE 802.11b WiFi 2.4 GHz (DSSS, 2 Mbps, 90pc dc)	WLAN	1.99	± 9.6
10573	AAA	IEEE 802.11b WiFi 2.4 GHz (DSSS, 5.5 Mbps, 90pc dc)	WLAN	1,98	± 9.6
10574	AAA	IEEE 802.11b WiFi 2.4 GHz (DSSS, 11 Mbps, 90pc dc)	WLAN	1.98	± 9.6
10575	AAA	IEEE 802.11g WIFI 2.4 GHz (DSSS-OFDM, 6 Mbps, 90pc dc)	WLAN	8.59	± 9.6
10576	AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 9 Mbps, 90pc dc)	WLAN	8.60	± 9.6
10577	AAA	IEEE 802,11g WIFI 2.4 GHz (DSSS-OFDM, 12 Mbps, 90pc dc)	WLAN	8.70	± 9.6
10578	AAA	IEEE 802.11g WIFI 2.4 GHz (DSSS-OFDM, 12 Mops, sope dc)	WLAN	8.49	± 9.6
10579	AAA	IEEE 802.11g WiFI 2.4 GHz (DSSS-OFDM, 24 Mbps, 90pc dc)	WLAN	8.36	± 9.6
10580	AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 36 Mbps, 90pc dc)	WLAN	8.76	± 9.6
10581	AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 30 Mbps, 30pc dc)	WLAN	8.35	± 9.6
10582	AAA	IEEE 802.11g WIFI 2.4 GHz (DSSS-OFDM, 54 Mbps, 90pc dc)	WLAN	8.67	± 9.6
10583	AAC	IEEE 802.11a/h WiFi 5 GHz (OFDM, 6 Mbps, 90pc dc)	WLAN	8.59	± 9.6
10584	AAC	IEEE 802.11a/h WiFi 5 GHz (OFDM, 9 Mbps, 90pc dc)	WLAN	8.60	± 9.6
10585	AAC	IEEE 802.11a/h WiFi 5 GHz (OFDM, 12 Mbps, 90pc dc)	WLAN	8.70	± 9.6
10586	AAC	IEEE 802.11a/h WiFI 5 GHz (OFDM, 18 Mbps, 90pc dc)	WLAN	8.49	± 9.6
10587	AAC	IEEE 802.11a/h WiFi 5 GHz (OFDM, 14 Mbps, 90pc dc)	WLAN	8.36	± 9.6
10588	AAC	IEEE 802.11a/h WiFi 5 GHz (OFDM, 36 Mbps, 90pc dc)	WLAN	8.76	± 9.6
10589	AAC	IEEE 802.11a/n WiFi 5 GHz (OFDM, 48 Mbps, 90pc dc)	WLAN	8.35	± 9.6
10590	AAC	IEEE 802.11a/h WiFI 5 GHz (OFDM, 54 Mbps, 90pc dc)	WLAN	8.67	± 9.6
10591	AAC	IEEE 802.11n (HT Mixed, 20MHz, MCS0, 90pc dc)	WLAN	8.63	± 9.6
10592	AAC	IEEE 802.11n (HT Mixed, 20MHz, MCS1, 90pc dc)	WLAN	8.79	± 9.6
10592	AAC	IEEE 802.11n (HT Mixed, 20MHz, MCS1, 20pc dc)	WLAN	8.64	± 9.6
10594	AAC	IEEE 802.11n (HT Mixed, 20MHz, MCS3, 90pc dc)	WLAN	8.74	± 9.6
10595	AAC	IEEE 802.11n (HT Mixed, 20MHz, MCS3, 90pc dc)	WLAN	8.74	± 9.6
10595	AAC	IEEE 802.11n (HT Mixed, 20MHz, MCS4, 90pc dc)	WLAN	8.71	± 9.6
10596	AAC	IEEE 802.11n (HT Mixed, 20MHz, MCS5, 90pc dc)	WLAN	8.72	± 9.0
10597	AAC	IEEE 802.11n (HT Mixed, 20MHz, MCS0, 90pc dc)	WLAN	8.50	± 9.6
10599	AAC	IEEE 802.11n (HT Mixed, 20MHz, MCS7, 50pc dc)	WLAN	8.79	± 9,6
10600	AAC	IEEE 802.11n (HT Mixed, 40MHz, MCS0, 90pc dc)	WLAN	8.88	± 9.0
	AAC	IEEE 802.11n (HT Mixed, 40MHz, MCS1, 90pc dc)	WLAN	8.88	± 9.6
10601	AAC	IEEE 802.11n (HT Mixed, 40MHz, MCS2, 90pc dc) IEEE 802.11n (HT Mixed, 40MHz, MCS3, 90pc dc)	WLAN		± 9.0
	AAC	IEEE 802.11n (HT Mixed, 40MHz, MCS3, 90pc dc)		8.94	
10603	MAG	TILLE OVE. THINT MIXED, 40MILE, MIGO4, SUDC DC)	WLAN	9.03	± 9.6

Certificate No: EX3-7686\_Oct21

Page 17 of 23

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 90 days only.

t (886-2) 2299-3279

Unless other west stated the results structure the state report results in provide the structure of the state and state of the state of the state and st prosecuted to the fullest extent of the law.

No.134,Wu Kung Road, New Taipei Industrial Park, Wuku District, New Taipei City, Taiwan/新北市五股區新北產業園區五工路 134 號

f (886-2) 2298-0488

www.sgs.com.tw



# Report No: ES/2022/30012 Rev: 01 Page: 23 of 28

10605	AAC	IEEE 802.11n (HT Mixed, 40MHz, MCS6, 90pc dc)	WLAN	8.97	± 9.6
10606	AAC	IEEE 802.11n (HT Mixed, 40MHz, MCS7, 90pc dc)	WLAN	8.82	± 9.0
10607	AAC	IEEE 802.11ac WiFi (20MHz, MCS0, 90pc dc)	WLAN	8.64	± 9.
10608	AAC	IEEE 802,11ac WiFi (20MHz, MCS1, 90pc dc)	WLAN	8.77	± 9.
10609	AAC	IEEE 802.11ac WiFi (20MHz, MCS2, 90pc dc)	WLAN	8.57	± 9.
10610	AAC	JEEE 802.11ac WiFI (20MHz, MCS3, 90pc dc)	WLAN	8.78	± 9.
10611	AAC	IEEE 802.11ac WiFi (20MHz, MCS4, 90pc dc)	WLAN	8.70	± 9.
10612	AAC	IEEE 802.11ac WiFi (20MHz, MCS5, 90pc dc)	WLAN	8.77	± 9.
10613	AAC	IEEE 802.11ac WiFi (20MHz, MCS6, 90pc dc)	WLAN	8.94	± 9,
10614	AAC	IEEE 802.11ac WIFI (20MHz, MCS7, 90pc dc)	WLAN	8.59	± 9.
10615	AAC	IEEE 802.11ac WiFi (20MHz, MCS8, 90pc dc)	WLAN	8.82	± 9.
10616	AAC	IEEE 802.11ac WiFi (40MHz, MCS0, 90pc dc)	WLAN	8.82	± 9.
10617	AAC	IEEE 802.11ac WiFi (40MHz, MCS1, 90pc dc)	WLAN	8.81	± 9.
10618	AAC	IEEE 802.11ac WiFi (40MHz, MCS2, 90pc dc)	WLAN	8.58	± 9.
10619	AAC	IEEE 802.11ac WiFi (40MHz, MCS3, 90pc dc)	WLAN	8.86	± 9.
10620	AAC	IEEE 802.11ac WiFi (40MHz, MCS4, 90pc dc)	WLAN	8.87	± 9.
10621	AAC	IEEE 802.11ac WiFi (40MHz, MCS5, 90pc dc)	WLAN		
10622	AAC	IEEE 802.11ac WiFi (40MHz, MCS6, 90pc dc)	WLAN	8.77	± 9.
10623	AAC	IEEE 802.11ac WiFi (40MHz, MCS0, S0pc dc)	WLAN	8.68	± 9.0
10624	AAC	IEEE 802,11ac WiFi (40MHz, MCS8, 90pc dc)	WLAN	8.96	
10625	AAC	IEEE 802.11ac WiFI (40MHz, MCS9, 90pc dc)	WLAN		± 9.
10626	AAC	IEEE 802.11ac WIFI (80MHz, MCS0, 90pc dc)	WLAN	8.96	± 9.
10627	AAC	IEEE 802.11ac WiFi (80MHz, MCS0, 90pc dc)	WLAN	8.83	± 9.
10628	AAC	IEEE 802.11ac WiFi (80MHz, MCS1, 90pc dc)		8.88	± 9.
10629	AAC	IEEE 802.11ac WiFi (80MHz, MCS3, 90pc dc)	WLAN	8.71	± 9.
10630	AAC	IEEE 802.11ac WiFi (80MHz, MCS4, 90pc dc)	WLAN	8.85	± 9.
10631	AAC	IEEE 802.11ac WiFi (80MHz, MCS4, Sopc dc)	WLAN	8.72	± 9.
10632	AAC	IEEE 802.11ac WiFi (80MHz, MCS6, 90pc dc)	WLAN	8.81	± 9.
10633	AAC	IEEE 802.11ac WiFI (80MHz, MCS0, 90pc dc)	WLAN	8.74	± 9.
10634	AAC	IEEE 802.11ac WIFI (80MHz, MCS8, 90pc dc)	WLAN	8.83	± 9.
10635	AAC	IEEE 802.11ac WiFi (80MHz, MCS8, 90pc dc)		8.80	± 9.
10636	AAD	IEEE 802.11ac WiFi (160MHz, MCSS, 90pc dc)	WLAN	8.81	± 9,
10637	AAD	IEEE 802.11ac WiFi (160MHz, MCS0, S0pc dc)	and the second se	8.83	± 9.
10638	AAD	IEEE 802.11ac WiFi (160MHz, MCS1, 90pc dc)	WLAN	8.79	± 9.
10639	AAD	IEEE 802.11ac WiFi (160MHz, MCS2, 90pc dc)		8.86	± 9.
10640	AAD	IEEE 802.11ac WiFi (160MHz, MCS3, 90pc dc)	WLAN	8.85	± 9.
10641	AAD	IEEE 802.11ac WiFi (160MHz, MCS5, 90pc dc)	WLAN	8.98	±.9.
10641	AAD	IEEE 802.11ac WiFi (160MHz, MCS6, 90pc dc)	WLAN	9.06	± 9,
10643	AAD	IEEE 802.11ac WiFi (160MHz, MCS7, 90pc dc)	WLAN	9.06	± 9.
10644	AAD	IEEE 802.11ac WiFi (160MHz, MCS8, 90pc dc)	WLAN	8.89	± 9.
10645	AAD	IEEE 802.11ac WiFi (160MHz, MCS9, 90pc dc)	WLAN	9.05	± 9.
10646	AAG	LTE-TDD (SC-FDMA, 1 RB, 5 MHz, QPSK, UL Sub=2,7)	WLAN	9.11	± 9.
10647	AAF	LTE-TDD (SC-FDMA, 1 RB, 5 WHZ, QPSK, 0L Sub=2,7)	LTE-TDD	11.96	± 9.
10648	AAA	CDMA2000 (1x Advanced)	LTE-TDD	11.96	± 9.
10646	AAE		CDMA2000	3.45	± 9,
10653	AAE	LTE-TDD (OFDMA, 5 MHz, E-TM 3.1, Clipping 44%) LTE-TDD (OFDMA, 10 MHz, E-TM 3.1, Clipping 44%)	LTE-TDD	6.91	±9,
10654	AAD		LTE-TDD	7.42	± 9.
10654	AAD	LTE-TDD (OFDMA, 15 MHz, E-TM 3.1, Clipping 44%)	LTE-TDD	6.96	± 9.
10655	AAE	LTE-TDD (OFDMA, 20 MHz, E-TM 3.1, Clipping 44%) Pulse Waveform (200Hz, 10%)	LTE-TDD	7.21	± 9.
10658	AAA	Pulse Waveform (200Hz, 10%) Pulse Waveform (200Hz, 20%)	Test	10.00	± 9.
			Test	6.99	±9.
10660	AAA	Pulse Waveform (200Hz, 40%)	Test	3.98	± 9,
10661	AAA	Pulse Waveform (200Hz, 60%)	Test	2.22	± 9.
10662	AAA	Pulse Waveform (200Hz, 80%)	Test	0.97	± 9.
10670	AAA	Bluetooth Low Energy	Bluetooth	2:19	± 9.
		EEE 802.11ax (20MHz, MCS0, 90pc dc)	WLAN	9.09	± 9.

Certificate No: EX3-7686\_Oct21

Page 18 of 23

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 90 days only.

t (886-2) 2299-3279

除非另有說明,此報告結果僅對測試之樣品負責,同時此樣品僅保留90天。本報告未經本公司書面許可,不可部份複製。 This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <a href="http://www.sgs.com.tw/Terms-and-Conditions">http://www.sgs.com.tw/Terms-and-Conditions</a> and for electronic format documents subject to Terms and Conditions for Electronic Documents at <u>http://www.sgs.com.tw/Terms-and-Conditions</u>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

No.134,Wu Kung Road, New Taipei Industrial Park, Wuku District, New Taipei City, Taiwan/新北市五股區新北產業園區五工路 134 號

f (886-2) 2298-0488

www.sgs.com.tw



10673	AAC	IEEE 802.11ax (20MHz, MCS2, 90pc dc)	WLAN	8.78	± 9.6
10674	AAC	IEEE 802.11ax (20MHz, MCS3, 90pc dc)	WLAN	8.74	± 9.6
10675	AAC	IEEE 802.11ax (20MHz, MCS4, 90pc dc)	WLAN	8.90	± 9.6
10676	AAC	IEEE 802.11ax (20MHz, MCS5, 90pc dc)	WLAN	8.77	± 9.6
10677	AAC	IEEE 802.11ax (20MHz, MCS6, 90pc dc)	WLAN	8.73	± 9.6
10678	AAC	IEEE 802.11ax (20MHz, MCS7, 90pc dc)	WLAN	8.78	± 9.6
10679	AAC	IEEE 802,11ax (20MHz, MCS8, 90pc dc)	WLAN	8.89	± 9.6
10680	AAC	IEEE 802.11ax (20MHz, MCS9, 90pc dc)	WLAN	8.80	± 9.6
10681	AAC	IEEE 802.11ax (20MHz, MCS10, 90pc dc)	WLAN	8.62	± 9.6
10682	AAC	IEEE 802.11ax (20MHz, MCS11, 90pc dc)	WLAN	8.83	± 9.6
10683	AAC	IEEE 802.11ax (20MHz, MCS0, 99pc dc)	WLAN	8.42	± 9.6
10684	AAC	IEEE 802.11ax (20MHz, MCS1, 99pc dc)	WLAN	8.26	± 9.6
10685	AAC	IEEE 802.11ax (20MHz, MCS2, 99pc dc)	WLAN	8.33	± 9.6
10686	AAC	IEEE 802.11ax (20MHz, MCS3, 99pc dc)	WLAN	8.28	± 9.6
10687	AAC	IEEE 802.11ax (20MHz, MCS4, 99pc dc)	WLAN	8.45	± 9.6
10688	AAC	IEEE 802.11ax (20MHz, MCS5, 99pc dc)	WLAN	8.29	± 9.6
10689	AAC	IEEE 802.11ax (20MHz, MCS6, 99pc dc)	WLAN	8.55	± 9.6
10690	AAC	IEEE 802.11ax (20MHz, MCS7, 99pc dc)	WLAN	8.00	± 9.6
10691	AAC	IEEE 802.11ax (20MHz, MCS8, 99pc dc)	WLAN		
10692	AAC	IEEE 802.11ax (20MHz, MCS9, 99pc dc)	WLAN	8.25	± 9.6
10693	AAC	IEEE 802.11ax (20MHz, MCS10, 99pc dc)	WLAN	8.29	± 9.6
10694	AAC	IEEE 802.11ax (20MHz, MCS10, 99pc dc)	WLAN		
10695	AAC	IEEE 802.11ax (40MHz, MCS0, 90pc dc)		8.57	± 9.6
10696	AAC	IEEE 802.11ax (40MHz, MCS1, 90pc dc)	WLAN	8.78	± 9.6
10696	AAC	IEEE 802.11ax (40MHz, MCS1, 90pc dc)	WLAN	8.91	± 9.6
10698	AAC	IEEE 802.11ax (40MHz, MCS2, 90pc 6c)	WLAN	8.61	± 9.6
10699	AAC	IEEE 802.11ax (40MHz, MCS3, 90pc dc)	WLAN	8.89	± 9.6
10700	AAC	IEEE 802.11ax (40MHz, MCS4, 90pc 6c) IEEE 802.11ax (40MHz, MCS5, 90pc 6c)	WLAN	8.82	± 9.6
10700	AAC		WLAN	8.73	± 9.6
10701	AAC	IEEE 802.11ax (40MHz, MCS6, 90pc dc) IEEE 802.11ax (40MHz, MCS7, 90pc dc)	WLAN	8.86	± 9.6
			WLAN	8.70	± 9.6
10703	AAC	IEEE 802.11ax (40MHz, MCS8, 90pc dc)	WLAN	8.82	± 9.6
10704	AAC	IEEE 802.11ax (40MHz, MCS9, 90pc dc)	WLAN	8,56	± 9.6
10705	AAC	IEEE 802.11ax (40MHz, MCS10, 90pc dc)	WLAN	8.69	± 9.6
10706	AAC	IEEE 802.11ax (40MHz, MCS11, 90pc dc)	WLAN	8.66	± 9.6
10707	AAC	IEEE 802.11ax (40MHz, MCS0, 99pc dc)	WLAN	8.32	± 9.6
10708	AAC	IEEE 802.11ax (40MHz. MCS1. 99pc dc)	WLAN	8.55	± 9.6
10709	AAC	IEEE 802.11ax (40MHz, MCS2, 99pc dc)	WLAN	8.33	± 9.6
10710	AAC	IEEE 802.11ax (40MHz, MCS3, 99pc dc)	WLAN	8.29	± 9.6
10711	AAC	IEEE 802.11ax (40MHz, MCS4, 99pc dc)	WLAN	8.39	± 9.6
10712	AAC	IEEE 802.11ax (40MHz, MCS5, 99pc dc)	WLAN	8.67	± 9,6
10713	AAC	IEEE 802.11ax (40MHz, MCS6, 99pc dc)	WLAN	8.33	± 9.6
10714	AAC	IEEE 802.11ax (40MHz, MCS7, 99pc dc)	WLAN	8.26	± 9.6
10715	AAC	IEEE 802.11ax (40MHz, MCS8, 99pc dc)	WLAN	8.45	± 9.6
10716	AAC	IEEE 802.11ax (40MHz, MCS9, 99pc dc)	WLAN	8.30	± 9.6
10717	AAC	IEEE 802.11ax (40MHz, MCS10, 99pc dc)	WLAN	8.48	± 9.6
10718	AAC	IEEE 802.11ax (40MHz, MCS11, 99pc dc)	WLAN	8.24	± 9.6
10719	AAC	IEEE 802.11ax (80MHz, MCS0, 90pc dc)	WLAN	8.81	± 9.6
10720	AAC	IEEE 802.11ax (80MHz, MCS1, 90pc dc)	WLAN	8.87	± 9.6
10721	AAC	IEEE 802.11ax (80MHz, MCS2, 90pc dc)	WLAN	8.76	± 9.6
10722	AAC	IEEE 802.11ax (80MHz, MCS3, 90pc dc)	WLAN	8.55	± 9.6
10723	AAC	IEEE 802.11ax (80MHz, MCS4, 90pc dc)	WLAN	8.70	± 9.6
10724	AAC	IEEE 802.11ax (80MHz, MCS5, 90pc dc)	WLAN	8.90	± 9.6
10725	AAC	IEEE 802.11ax (80MHz, MCS6, 90pc dc)	WLAN	8.74	± 9.6
10726	AAC	IEEE 802.11ax (80MHz, MCS7, 90pc dc)	WLAN	8.72	± 9.6
10727	AAC	IEEE 802.11ax (80MHz, MCS8, 90pc dc)	WLAN	8.66	± 9.6
10728	AAC	IEEE 802.11ax (80MHz, MCS9, 90pc dc)	WLAN	8.65	± 9.6

Certificate No: EX3-7686\_Oct21

Page 19 of 23

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 90 days only.

t (886-2) 2299-3279

除非另有說明,此報告結果僅對測試之樣品負責,同時此樣品僅保留的天。本報告未經本公司書面許可,不可部份複製。 This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <a href="http://www.sgs.com.tw/Terms-and-Conditions">http://www.sgs.com.tw/Terms-and-Conditions</a> and for electronic format documents subject to Terms and Conditions for Electronic Documents at <u>http://www.sgs.com.tw/Terms-and-Conditions</u>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

No.134,Wu Kung Road, New Taipei Industrial Park, Wuku District, New Taipei City, Taiwan/新北市五股區新北產業園區五工路 134 號

f (886-2) 2298-0488

www.sgs.com.tw



# Report No: ES/2022/30012 Rev: 01 Page: 25 of 28

10729	AAC	IEEE 802.11ax (80MHz, MCS10, 90pc dc)	WLAN	8.64	± 9.6
10730	AAC	IEEE 802.11ax (80MHz, MCS11, 90pc dc)	WLAN	8.67	± 9.6
10731	AAC	IEEE 802.11ax (80MHz, MCS0, 99pc dc)	WLAN	8.42	± 9.6
10732	AAC	IEEE 802.11ax (80MHz, MCS1, 99pc dc)	WLAN	8.46	± 9.6
10733	AAC	IEEE 802.11ax (80MHz, MCS2, 99pc dc)	WLAN	8.40	± 9.6
10734	AAC	IEEE 802.11ax (80MHz, MCS3, 99pc dc)	WLAN	8.25	± 9.6
10735	AAC	IEEE 802.11ax (80MHz, MCS4, 99pc dc)	WLAN	8.33	± 9.6
10736	AAC	IEEE 802.11ax (80MHz, MCS5, 99pc dc)	WLAN	8.27	± 9.6
10737	AAC	IEEE 802.11ax (80MHz, MCS6, 99pc dc)	WLAN	8.36	± 9.6
10738	AAC	IEEE 802.11ax (80MHz, MCS7, 99pc dc)	WLAN	8.42	± 9.6
10739	AAC	IEEE 802.11ax (80MHz, MCS8, 99pc dc)	WLAN	8.29	± 9.6
10740	AAC	IEEE 802.11ax (80MHz, MCS9, 99pc dc)	WLAN	8.48	± 9.6
10741	AAC	IEEE 802.11ax (80MHz, MCS10, 99pc dc)	WLAN	8.40	± 9.6
10742	AAC	IEEE 802.11ax (80MHz, MCS11, 99pc dc)	WLAN	8.43	
10743		IEEE 802.11ax (160MHz, MCS0, 90pc dc)	WLAN	8.94	± 9.6
10744	AAC	IEEE 802.11ax (160MHz, MCS1, 90pc dc)	WLAN		± 9.6
10745	AAC	IEEE 802.11ax (160MHz, MCS1, 90pc dc)	WLAN	9.16	± 9.6
10746	AAC	IEEE 802.11ax (160MHz, MCS2, 90pc dc)	WLAN	8.93	± 9.6
10740		IEEE 802.11ax (160MHz, MCS3, 90pc dc)	WLAN	9.11	± 9.6
10748	AAC	IEEE 802,11ax (160MHz, MCS4, sope de)		9.04	± 9.6
10748	AAC	IEEE 802.11ax (160MHz, MCS6, 90pc dc)	WLAN	8.93	± 9.6
10749	AAC	IEEE 802.11ax (160MHz, MCS6, 90pc dc)	WLAN	8.90	± 9.6
10750	AAC	IEEE 802.118X (160MHz, MCS7, 90pc dc)	WLAN	8.79	± 9.6
10751	AAC		WLAN	8.82	± 9.0
10753	AAC	IEEE 802.11ax (160MHz, MCS9, 90pc dc)	WLAN	8.81	± 9.0
		IEEE 802.11ax (160MHz, MCS10, 90pc dc)	WLAN	9.00	± 9.0
10754	AAC	IEEE 802.11ax (160MHz, MCS11, 90pc dc)	WLAN	8.94	± 9.6
		IEEE 802.11ax (160MHz, MCS0, 99pc dc)	WLAN	8.64	± 9.6
10756	AAC	IEEE 802.11ax (160MHz, MCS1, 99pc dc)	WLAN	8.77	± 9.6
10757	AAC	IEEE 802,11ax (160MHz, MCS2, 99pc dc)	WLAN	8.77	± 9.6
10758	AAC	IEEE 802.11ax (160MHz, MCS3, 99pc dc)	WLAN	8.69	± 9.6
10759	AAC	IEEE 802.11ax (160MHz, MCS4, 99pc dc)	WLAN	8.58	± 9.6
10760	AAC	IEEE 802.11ax (160MHz, MCS5, 99pc dc)	WLAN	8.49	± 9.6
10761	AAC	IEEE 802.11ax (160MHz, MCS6, 99pc dc)	WLAN	8.58	± 9.6
10762	AAC	IEEE 802.11ax (160MHz, MCS7, 99pc dc)	WLAN	8.49	± 9,6
10763	AAC	IEEE 802.11ax (160MHz, MCS8, 99pc dc)	WLAN	8.53	± 9.6
10764	AAC	IEEE 802.11ax (160MHz, MCS9, 99pc dc)	WLAN	8.54	± 9.6
10765	AAC	IEEE 802.11ax (160MHz, MCS10. 99pc dc)	WLAN	8.54	± 9.6
10766	AAC	IEEE 802.11ax (160MHz, MCS11, 99pc dc)	WLAN	8.51	± 9.6
10767	AAE	5G NR (CP-OFDM, 1 RB, 5 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	7.99	± 9.6
10768	AAD	5G NR (CP-OFDM, 1 RB, 10 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.01	± 9.6
10769	AAD	5G NR (CP-OFDM, 1 RB, 15 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.01	± 9.6
10770	AAD	5G NR (CP-OFDM, 1 RB, 20 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.02	± 9.6
10771	AAD	5G NR (CP-OFDM, 1 RB, 25 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.02	± 9.6
10772	AAD	5G NR (CP-OFDM, 1 RB, 30 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.23	± 9.6
10773	AAD	5G NR (CP-OFDM, 1 RB, 40 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.03	± 9.6
10774	AAD	5G NR (CP-OFDM, 1 RB, 50 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.02	± 9,6
10775	AAD	5G NR (CP-OFDM, 50% RB, 5 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.31	± 9.0
10776	AAD	5G NR (CP-OFDM, 50% RB, 10 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.30	± 9.6
10777	AAC	5G NR (CP-OFDM, 50% RB, 15 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.30	± 9.6
10778	AAD	5G NR (CP-OFDM, 50% RB, 20 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.34	± 9.6
10779	AAC	5G NR (CP-OFDM, 50% RB, 25 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.42	± 9.0
10780	AAD	5G NR (CP-OFDM, 50% RB, 30 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.38	± 9.0
10781	AAD	5G NR (CP-OFDM, 50% RB, 40 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.38	± 9.1
10782	AAD	5G NR (CP-OFDM, 50% RB, 50 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.43	± 9,1
10783	AAE	5G NR (CP-OFDM, 100% RB, 5 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.31	± 9.1
10784	AAD	5G NR (CP-OFDM, 100% RB, 10 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.29	± 9.1

Certificate No: EX3-7686\_Oct21

Page 20 of 23

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 90 days only.

t (886-2) 2299-3279

Unless other west stated the results structure the state report results in provide the structure of the state and state of the state of the state and st prosecuted to the fullest extent of the law.

No.134,Wu Kung Road, New Taipei Industrial Park, Wuku District, New Taipei City, Taiwan/新北市五股區新北產業園區五工路 134 號



10785	AAD	5G NR (CP-OFDM, 100% RB, 15 MHz, QPSK, 15 kHz)		0.40	1.00
10786	AAD	5G NR (CP-OFDM, 100% RB, 20 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.40	± 9.6
10787	AAD	5G NR (CP-OFDM, 100% RB, 25 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.35	± 9.6
10788	AAD	5G NR (CP-OFDM, 100% RB, 30 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.44	± 9.1
10789	AAD	5G NR (CP-OFDM, 100% RB, 40 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.39	± 9,1
10790	AAD	5G NR (CP-OFDM, 100% RB, 50 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.37	± 9.1
10791	AAE	5G NR (CP-OFDM, 1 RB, 5 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.39	± 9.1
10792	AAD	5G NR (CP-OFDM, 1 RB, 10 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	7.83	± 9.1
10793	AAD	5G NR (CP-OFDM, 1 RB, 15 MHz, QPSK, 30 kHz)	5G NR FR1 TDD 5G NR FR1 TDD	7.92	± 9.1
10794	AAD	5G NR (CP-OFDM, 1 RB, 20 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	7.95	± 9,1
10795	AAD	5G NR (CP-OFDM, 1 RB, 25 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	7.84	± 9,1
10796	AAD	5G NR (CP-OFDM, 1 RB, 30 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	7.82	± 9.6
10797	AAD	5G NR (CP-OFDM, 1 RB, 40 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.01	± 9.1
10798	AAD	5G NR (CP-OFDM, 1 RB, 50 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	7.89	± 9.0
10799	AAD	5G NR (CP-OFDM, 1 RB, 60 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	7.93	± 9.0
10801	AAD	5G NR (CP-OFDM, 1 RB, 80 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	7.89	-
10802	AAD	5G NR (CP-OFDM, 1 RB, 90 MHz, OPSK, 30 kHz)	5G NR FR1 TDD	7.87	± 9.1
10803	AAD	5G NR (CP-OFDM, 1 RB, 100 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	7.93	± 9,1 ± 9,1
10805	AAD	5G NR (CP-OFDM, 50% RB, 10 MHz, QPSK, 30 KHz)	5G NR FR1 TDD	8.34	± 9,1
10806	AAD	5G NR (CP-OFDM, 50% RB, 15 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.37	± 9.1
10809	AAD	5G NR (CP-OFDM, 50% RB, 30 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.34	± 9.0
10810	AAD	5G NR (CP-OFDM, 50% RB, 40 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.34	± 9.0
10812	AAD	5G NR (CP-OFDM, 50% RB, 60 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.35	± 9.1
10817	AAE	5G NR (CP-OFDM, 100% RB, 5 MHz, QPSK, 30 KHz)	5G NR FR1 TDD	8.35	± 9.1
10818	AAD	5G NR (CP-OFDM, 100% RB, 10 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.34	± 9.1
10819	AAD	5G NR (CP-OFDM, 100% RB, 15 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.33	± 9.1
10820	AAD	5G NR (CP-OFDM, 100% RB, 20 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.30	± 9.1
10821	AAD	5G NR (CP-OFDM, 100% RB, 25 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.41	± 9.
10822	AAD	5G NR (CP-OFDM, 100% RB, 30 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.41	
10823	AAD	5G NR (CP-OFDM, 100% RB, 40 MHz, QPSK, 30 kHz)	5G NR FR1 TDD		± 9.1
10824	AAD	5G NR (CP-OFDM, 100% RB, 50 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.36	± 9.0
10825	AAD	5G NR (CP-OFDM, 100% RB, 60 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.41	
10827	AAD	5G NR (CP-OFDM, 100% RB, 80 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.42	± 9.0
10828	AAD	5G NR (CP-OFDM, 100% RB, 90 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.43	
10829	AAD	5G NR (CP-OFDM, 100% RB, 100 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.40	± 9.0
10830	AAD	5G NR (CP-OFDM, 1 RB, 10 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	7.63	± 9.
10831	AAD	5G NR (CP-OFDM, 1 RB, 15 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	7.73	± 9.
10832	AAD	5G NR (CP-OFDM, 1 RB, 20 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	7.74	± 9.
10833	AAD	5G NR (CP-OFDM, 1 RB, 25 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	7.70	± 9.
10834	AAD	5G NR (CP-OFDM, 1 RB, 30 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	7.75	± 9.
10835	AAD	5G NR (CP-OFDM, 1 RB, 40 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	7.70	± 9.
10836	AAD	5G NR (CP-OFDM, 1 RB, 50 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	7.66	± 9.
10837	AAD	5G NR (CP-OFDM, 1 RB, 60 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	7.68	± 9.
10839	AAD	5G NR (CP-OFDM, 1 RB, 80 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	7.70	± 9.
10840	AAD	5G NR (CP-OFDM, 1 RB, 90 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	7.67	± 9.
10841	AAD	5G NR (CP-OFDM, 1 RB, 100 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	7.71	± 9.
10843	AAD	5G NR (CP-OFDM, 50% RB, 15 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	8.49	± 9.
10844	AAD	5G NR (CP-OFDM, 50% RB, 20 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	8.34	± 9.
10846	AAD	5G NR (CP-OFDM, 50% RB, 30 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	8.41	± 9.
10854	AAD	5G NR (CP-OFDM, 100% RB, 10 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	8.34	± 9.
10855	AAD	5G NR (CP-OFDM, 100% RB, 15 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	8.36	± 9.
10856	AAD	5G NR (CP-OFDM, 100% RB, 20 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	8.30	± 9.
10857	AAD	5G NR (CP-OFDM, 100% RB, 25 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	8.35	± 9.
10858	AAD	5G NR (CP-OFDM, 100% RB, 30 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	8.36	± 9.
10859	AAD	5G NR (CP-OFDM, 100% RB, 40 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	8.36	± 9. ± 9.
10860	AAD	5G NR (CP-OFDM, 100% RB, 50 MHz, CPSK, 60 kHz)	5G NR FR1 TDD	8.41	± 9.

Certificate No: EX3-7686\_Oct21

Page 21 of 23

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 90 days only.

t (886-2) 2299-3279

Unless other west stated the results structure the state report results in provide the structure of the state and state of the state of the state and st prosecuted to the fullest extent of the law.

No.134,Wu Kung Road, New Taipei Industrial Park, Wuku District, New Taipei City, Taiwan/新北市五股區新北產業園區五工路 134 號

f (886-2) 2298-0488

www.sgs.com.tw



# Report No: ES/2022/30012 Rev: 01 Page: 27 of 28

10861	AAD	5G NR (CP-OFDM, 100% RB, 60 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	8.40	± 9.1
10863	AAD	5G NR (CP-OFDM, 100% RB, 80 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	8.41	± 9.1
10864	AAD	5G NR (CP-OFDM, 100% RB, 90 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	8.37	± 9.0
10865	AAD	5G NR (CP-OFDM, 100% RB, 100 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	8.41	± 9.
10866	AAD	5G NR (DFT-s-OFDM, 1 RB, 100 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.68	± 9.
10868	AAD	5G NR (DFT-s-OFDM, 100% RB, 100 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.89	± 9.
10869	AAD	5G NR (DFT-s-OFDM, 1 RB, 100 MHz, QPSK, 120 kHz)	5G NR FR2 TDD	5.75	±9.
10870	AAD	5G NR (DFT-s-OFDM, 100% RB, 100 MHz, QPSK, 120 kHz)	5G NR FR2 TDD	5.86	± 9.
10871	AAD	5G NR (DFT-s-OFDM, 1 RB, 100 MHz, 16QAM, 120 kHz)	5G NR FR2 TDD	5.75	± 9.
10872	AAD	5G NR (DFT-s-OFDM, 100% RB, 100 MHz, 16QAM, 120 kHz)	5G NR FR2 TDD	6.52	± 9.
10873	AAD	5G NR (DFT-s-OFDM, 1 RB, 100 MHz, 64QAM, 120 kHz)	5G NR FR2 TDD	6.61	± 9.
10874	AAD	5G NR (DFT-s-OFDM, 100% RB, 100 MHz, 64QAM, 120 kHz)	5G NR FR2 TDD	6.65	± 9.
10875	AAD	5G NR (CP-OFDM, 1 RB, 100 MHz, OPSK, 120 kHz)	5G NR FR2 TDD	7.78	± 9.
10876	AAD	5G NR (CP-OFDM, 100% RB, 100 MHz, QPSK, 120 kHz)	5G NR FR2 TDD	8.39	± 9.
10877	AAD	5G NR (CP-OFDM, 1 RB, 100 MHz, 16QAM, 120 kHz)	5G NR FR2 TDD	7.95	± 9.
10878	AAD	5G NR (CP-OFDM, 100% RB, 100 MHz, 16QAM, 120 kHz)	5G NR FR2 TDD	8.41	± 9.
10879	AAD	5G NR (CP-OFDM, 1 RB, 100 MHz, 64QAM, 120 kHz)	5G NR FR2 TDD	8.12	± 9.
10880	AAD	5G NR (CP-OFDM, 100% RB, 100 MHz, 64QAM, 120 kHz)	5G NR FR2 TDD	8.38	± 9.
10881	AAD	5G NR (DFT-s-OFDM, 1 RB, 50 MHz, QPSK, 120 kHz)	5G NR FR2 TDD	5.75	± 9.
10882	AAD	5G NR (DFT-s-OFDM, 100% RB, 50 MHz, QPSK, 120 kHz)	5G NR FR2 TDD	5.96	± 9.
10883	AAD	5G NR (DFT-s-OFDM, 1 RB, 50 MHz, 16QAM, 120 kHz)	5G NR FR2 TDD	6.57	± 9.
10884	AAD	5G NR (DFT-s-OFDM, 100% RB, 50 MHz, 16QAM, 120 kHz)	5G NR FR2 TDD	6.53	± 9.
10885	AAD	5G NR (DFT-s-OFDM, 1 RB, 50 MHz, 64QAM, 120 kHz)	5G NR FR2 TDD	6.61	± 9.
10886	AAD	5G NR (DFT-s-OFDM, 100% RB, 50 MHz, 64QAM, 120 kHz)	5G NR FR2 TDD	6.65	±9.
10887	AAD	5G NR (CP-OFDM, 1 RB, 50 MHz, QPSK, 120 kHz)	5G NR FR2 TDD	7.78	± 9.
10888	AAD	5G NR (CP-OFDM, 100% RB, 50 MHz, QPSK, 120 kHz)	5G NR FR2 TDD	8.35	± 9.
10889	AAD	5G NR (CP-OFDM, 1 RB, 50 MHz, 16QAM, 120 KHz)	5G NR FR2 TDD	8.02	± 9.
10890	AAD	5G NR (CP-OFDM, 100% RB, 50 MHz, 16QAM, 120 kHz)	5G NR FR2 TDD	8.40	± 9.
10891	AAD	5G NR (CP-OFDM, 1 RB, 50 MHz, 64QAM, 120 kHz)	5G NR FR2 TDD	8.13	± 9.
10892	AAD	5G NR (CP-OFDM, 100% RB, 50 MHz, 64QAM, 120 kHz)	5G NR FR2 TDD	8.41	± 9.
10897	AAC	5G NR (DFT-s-OFDM, 1 RB, 5 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.66	± 9.
10898	AAB	5G NR (DFT-s-OFDM, 1 RB, 10 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.67	± 9.
10899	AAB	5G NR (DFT-s-OFDM, 1 RB, 15 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.67	± 9.
10900	AAB	5G NR (DFT-s-OFDM, 1 RB, 20 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.68	±9.
10901	AAB	5G NR (DFT-s-OFDM, 1 RB, 25 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.68	± 9.
10902	AAB	5G NR (DFT-s-OFDM, 1 RB, 30 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.68	±9.
10903	AAB	5G NR (DFT-s-OFDM, 1 RB, 40 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.68	±9.
10904	AAB	5G NR (DFT-s-OFDM, 1 RB, 50 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.68	± 9.
10905	AAB	5G NR (DFT-s-OFDM, 1 RB, 60 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.68	± 9.
10906	AAB	5G NR (DFT-s-OFDM, 1 RB, 80 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.68	± 9.
10907	AAC	5G NR (DFT-s-OFDM, 50% RB, 5 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.78	± 9.
10908	AAB	5G NR (DFT-s-OFDM, 50% RB, 10 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.93	± 9.
10909	AAB	5G NR (DFT-s-OFDM, 50% RB, 15 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.96	± 9.
10910	AAB	5G NR (DFT-s-OFDM, 50% RB, 20 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.83	± 9.
10911	AAB	5G NR (DFT-s-OFDM, 50% RB, 25 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.93	± 9.
10912	AAB	5G NR (DFT-s-OFDM, 50% RB, 30 MHz, QPSK, 30 KHz)	5G NR FR1 TDD	5.84	
10913	AAB	5G NR (DFT-s-OFDM, 50% RB, 40 MHz, QPSK, 30 kHz)		5.84	± 9.
10914	AAB	5G NR (DFT-s-OFDM, 50% RB, 50 MHz, QPSK, 30 KHz)	5G NR FR1 TDD	5.84	± 9.
10914	AAB	5G NR (DFT-s-OFDM, 50% RB, 50 MHz, QPSK, 30 KHz)	5G NR FR1 TDD		± 9.
10915	AAB	5G NR (DFT-s-OFDM, 50% RB, 80 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.83	± 9.
10917	AAB	5G NR (DFT-s-OFDM, 50% RB, 100 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.87	± 9.
10918	AAC	5G NR (DFT-s-OFDM, 50% RB, 100 MHz, QPSK, 30 kHz) 5G NR (DFT-s-OFDM, 100% RB, 5 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.94	± 9.
10918	AAB	5G NR (DFT-s-OFDM, 100% RB, 5 MHz, QPSK, 30 KHz) 5G NR (DFT-s-OFDM, 100% RB, 10 MHz, QPSK, 30 KHz)	5G NR FR1 TDD	5.86	± 9.
10919	AAB	5G NR (DFT-s-OFDM, 100% RB, 10 MHz, QPSK, 30 KHz) 5G NR (DFT-s-OFDM, 100% RB, 15 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.86	±9.
10920	AAB		5G NR FR1 TDD	5.87	± 9.
	AND	5G NR (DFT-s-OFDM, 100% RB, 20 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.84	± 9.

Certificate No: EX3-7686 Oct21

Page 22 of 23

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 90 days only.

t (886-2) 2299-3279

Unless other west stated the results structure the state report results in provide the structure of the state and state of the state of the state and st prosecuted to the fullest extent of the law.

No.134,Wu Kung Road, New Taipei Industrial Park, Wuku District, New Taipei City, Taiwan/新北市五股區新北產業園區五工路 134 號

f (886-2) 2298-0488

www.sgs.com.tw



10923	AAB	5G NR (DFT-s-OFDM, 100% RB, 30 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.84	± 9.6
10924	AAB	5G NR (DFT-s-OFDM, 100% RB, 40 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.84	± 9.6
10925	AAB	5G NR (DFT-s-OFDM, 100% RB, 50 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.95	± 9.6
10926	AAB	5G NR (DFT-s-OFDM, 100% RB, 60 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.84	± 9.6
10927	AAB	5G NR (DFT-s-OFDM, 100% RB, 80 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.94	± 9.6
10928	AAC	5G NR (DFT-s-OFDM, 1 RB, 5 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.52	± 9.6
10929	AAC	5G NR (DFT-s-OFDM, 1 RB, 10 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.52	± 9.0
10930	AAC	5G NR (DFT-s-OFDM, 1 RB, 15 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.52	± 9.6
10931	AAC	5G NR (DFT-s-OFDM, 1 RB, 20 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.51	± 9.6
10932	AAC	5G NR (DFT-s-OFDM, 1 RB, 25 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.51	± 9.6
10933	AAC	5G NR (DFT-s-OFDM, 1 RB, 30 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.51	± 9.6
10934	AAC	5G NR (DFT-s-OFDM, 1 RB, 40 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.51	± 9.6
10935	AAD	5G NR (DFT-s-OFDM, 1 RB, 50 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.51	± 9.6
10936	AAC	5G NR (DFT-s-OFDM, 50% RB, 5 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.90	± 9.1
10937	AAC	5G NR (DFT-s-OFDM, 50% RB, 10 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.77	± 9.1
10938	AAC	5G NR (DFT-s-OFDM, 50% RB, 15 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.90	± 9.
10939	AAC	5G NR (DFT-s-OFDM, 50% RB, 20 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.82	± 9.1
10940	AAC	5G NR (DFT-s-OFDM, 50% RB, 25 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.89	± 9.1
10941	AAC	5G NR (DFT-s-OFDM, 50% RB, 30 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.83	± 9.1
10942	AAC	5G NR (DFT-s-OFDM, 50% RB, 40 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.85	± 9.6
10943	AAD	5G NR (DFT-s-OFDM, 50% RB, 50 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.95	± 9.1
10944	AAC	5G NR (DFT-s-OFDM, 100% RB, 5 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.81	± 9.1
10945	AAC	5G NR (DFT-s-OFDM, 100% RB, 10 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.85	± 9.
10946	AAC	5G NR (DFT-s-OFDM, 100% RB, 15 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.83	± 9.1
10947	AAC	5G NR (DFT-s-OFDM, 100% RB, 20 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.87	± 9.0
10948	AAC	5G NR (DFT-s-OFDM, 100% RB, 25 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.94	± 9.1
10949	AAC	5G NR (DFT-s-OFDM, 100% RB, 30 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.87	± 9.1
10950	AAC	5G NR (DFT-s-OFDM, 100% RB, 40 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.94	± 9.1
10951	AAD	5G NR (DFT-s-OFDM, 100% RB, 50 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.92	± 9.
10952	AAA	5G NR DL (CP-OFDM, TM 3.1. 5 MHz, 64-QAM, 15 kHz)	5G NR FR1 FDD	8.25	± 9.
10953	AAA	5G NR DL (CP-OFDM, TM 3.1, 10 MHz, 64-QAM, 15 kHz)	5G NR FR1 FDD	8.15	± 9.1
10954	AAA	5G NR DL (CP-OFDM, TM 3.1, 15 MHz, 64-QAM, 15 kHz)	5G NR FR1 FDD	8.23	± 9.
10955	AAA	5G NR DL (CP-OFDM, TM 3.1, 20 MHz, 64-QAM, 15 kHz)	5G NR FR1 FDD	8.42	± 9.
10956	AAA	5G NR DL (CP-OFDM, TM 3.1, 5 MHz, 64-QAM, 30 kHz)	5G NR FR1 FDD	8.14	± 9.1
10957	AAA	5G NR DL (CP-OFDM, TM 3.1, 10 MHz, 64-QAM, 30 kHz)	5G NR FR1 FDD	8.31	± 9.
10958	AAA	5G NR DL (CP-OFDM, TM 3.1, 15 MHz, 64-QAM, 30 kHz)	5G NR FR1 FDD	8.61	± 9.
10959	AAA	5G NR DL (CP-OFDM, TM 3.1, 20 MHz, 64-QAM, 30 KHz)	5G NR FR1 FDD	8.33	± 9.
10960	AAC	5G NR DL (CP-OFDM, TM 3.1, 5 MHz, 64-QAM, 15 kHz)	5G NR FR1 TDD	9.32	± 9.
10961	AAB	5G NR DL (CP-OFDM, TM 3.1, 10 MHz, 64-QAM, 15 kHz)	5G NR FR1 TDD	9.36	± 9.
10962	AAB	5G NR DL (CP-OFDM, TM 3.1, 15 MHz, 64-QAM, 15 KHz)	5G NR FR1 TDD	9.30	± 9.
10963	AAB	5G NR DL (CP-OFDM, TM 3.1, 20 MHz, 64-QAM, 15 kHz)	5G NR FR1 TDD	9.40	± 9.
10964	AAC	5G NR DL (CP-OFDM; TM 3.1, 5 MHz, 64-QAM; 30 kHz)	5G NR FR1 TDD	9.00	± 9.
10965	AAB	5G NR DL (CP-OFDM, TM 3.1, 10 MHz, 64-QAM, 30 kHz)	5G NR FR1 TDD	9.29	± 9,0
10966	AAB	5G NR DL (CP-OFDM, TM 3.1, 15 MHz, 64-QAM, 30 KHz)	5G NR FR1 TDD	9.37	± 9.
10967	AAB	5G NR DL (CP-OFDM, TM 3.1, 20 MHz, 64-QAM, 30 KHz)	5G NR FR1 TDD	9.55	± 9.
10968	AAB	5G NR DL (CP-OFDM, TM 3.1, 100 MHz, 64-QAM, 30 kHz)		9.42	± 9,
10908	AAB	5G NR (CP-OFDM, 1 RB, 20 MHz, QPSK, 15 kHz)	5G NR FR1 TDD		
10972	AAB	5G NR (DFT-6-OFDM, 1 RB, 20 MHz, QPSK, 15 kHz) 5G NR (DFT-6-OFDM, 1 RB, 100 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	11.59	± 9,
10974	AAB		5G NR FR1 TDD	9.06	± 9.
10974	AAA	5G NR (CP-OFDM, 100% RB, 100 MHz, 256-QAM, 30 kHz)	5G NR FR1 TDD	10.28	± 9.
10978	AAA	ULLA HDR4	ULLA	2.23	± 9.
10979	AAA		ULLA	7.02	± 9.
	AAA	ULLA HDR8	ULLA	8.82	± 9.
10981	10 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1	ULLA HDRp4	ULLA	1.50	± 9,
10982	AAA	ULLA HDRp8	ULLA	1.44	± 9,

Certificate No: EX3-7686 Oct21

Page 23 of 23

- End of report -

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 90 days only.

t (886-2) 2299-3279

除非另有說明,此報告結果僅對測試之樣品負責,同時此樣品僅保留的天。本報告未經本公司書面許可,不可部份複製。 This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <a href="http://www.sgs.com.tw/Terms-and-Conditions">http://www.sgs.com.tw/Terms-and-Conditions</a> and for electronic format documents subject to Terms and Conditions for Electronic Documents at <u>http://www.sgs.com.tw/Terms-and-Conditions</u>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

No.134,Wu Kung Road, New Taipei Industrial Park, Wuku District, New Taipei City, Taiwan/新北市五股區新北產業園區五工路 134 號

f (886-2) 2298-0488

```
www.sgs.com.tw
```