

Validation band: CW, FRONT

Exposure Conditions

Band	Validation band	Phantom Section	5G
Frequency [MHz] Channel Number	60000.0 60000	Conversion Factor	1.0
Group UID	CW, 0--	Position Test Distance [mm]	FRONT 5.55

Hardware Setup

Probe Calibration Date	EUmmWV4 - SN9589_F55-110GHz 2023-09-05	Phantom	mmWave xxxx
DAE Calibration Date	DAE4 Sn1239 2023-03-16	Medium	Air -
Software Version	3.2.2.2358		

Scan Setup

Scan Type	5G Scan	Grid Extents [mm]	32.5 x 32.5
Grid Steps [lambda]	0.25 x 0.25	Sensor Surface [mm]	5.55

Measurement Results

Avg. Area [cm ²]	4.00
psPDn+ [W/m ²]	239
psPDtot+ [W/m ²]	242
psPDmod+ [W/m ²]	243
E _{max} [V/m]	428
H _{max} [A/m]	1.16
Power Drift [dB]	-0.03

