


APPENDIX REPORT

Project No.	SHT2303063702EW	Radio Specification	DFS
Test sample No.	YPHT23030637001	Model No.	EVOLUTION MEDICAL
Start test date	2023-06-06	Finish date	2023-06-06
Temperature	24.8℃	Humidity	53%
Test Engineer	Xiaoqin Li	Auditor	<i>Xiaodong Zheo</i>


Appendix clause	Test item	Result
A	Radar Waveform Calibration	PASS
B	Channel Move Time, Channel Closing Transmission Time	PASS

Appendix A: Radar Waveform Calibration




Appendix C: Channel Move Time, Channel Closing Transmission Time

Test item	Result
Single pulse	430µs



Test Mode	Channel	Frequency (MHz)	Test item	Result (s)	Limit (s)	Verdict
802.11ac (VHT80)	58	5290	Channel Move Time	2.15	< 10	PASS
			Channel Closing Transmission Time	0.0129	< 1	PASS

Note: Channel Closing Transmission Time=Single pulse*Pulse count



-----End of Report-----