

Doc ID: G02-OPL-0137	Rev.: 03	Reference: G02 Patient Kit US_CM DMR	Page 1 of 2
Title: Revi™ Wearable Device Label for US Commercial Use			


Revision History

Revision	ECO/PDP Change no.	Description of Change	Position	Revision By	Date
01	2206002	First Issue	Director Regulatory Affairs	Belia Juarez	August 2022
02	2212003	- Changed name to Revi - updated date format to YYYY-MM-DD - added FCC symbol and ID	Product Manager	Carry Schlaff	May 2023
03	2307001	- Added FCC conditions - Added Rx Symbol to meet US FDA requirements	Product Manager	Carry Schlaff	August 2023

Doc ID: G02-OPL-0137	Rev.: 03	Reference: G02 Patient Kit US_CM DMR	Page 2 of 2
Title: Revi™ Wearable Device Label for US Commercial Use			

1. Purpose

This document specifies the format of the Revi™ Wearable Device label.



Label Size: Width: 77.98mm, Height: 30mm.

Material: Polyester.

Logo Color: Light Blue with Black. Light Blue Pantone: 300C

Note: the UDI code appears in the stick for illustration only