



6513 Windcrest Drive
Suite #100
Plano, TX 75024
Tel: 888-447-6673
www.somnomed.com

Federal Communications Commission
Authorization and Evaluation Division
1435 Oakland Mills Road
Columbia, MD 21046

Date: 28/11/2023

SUBJECT: FCC Application Class II Permissive Change FCC ID: 2BBTC-1154412

To Whom It May Concern:

Pursuant to CFR 2.1043, SomnoMed Inc. requests a Class II Permissive change for the following device:

FCC ID: 2BBTC-1154412 – Original grant date: 23 October 2023

Description of changes/differences:

The differences between the Rest Assure Standard PCBA (Elite Standard) and the short PCBA (Elite Small) are defined in the documents in the following table.

Exhibit Type	Avant (Original Certification)	Elite Standard (Original Certification)	Elite Small (C2PC)
Block Diagram	Document 472_1388_01 (same for all devices)		
Operational Description	Document 472_1387_01 (same for all devices)		
Schematics	PI-0079482-G01*	PI-0086353-A01*	

The small PCBA is 29mm shorter than the standard PCBA. Components and functionality are the otherwise the same between the small and standard PCBA.

Regards,

Dorene Markwiese
Global Director RA