

APPENDIX C: SAR TISSUE SPECIFICATIONS

Measurement Procedure for Tissue verification:

- 1) The network analyzer and probe system was configured and calibrated.
- 2) The probe was immersed in the tissue. The tissue was placed in a nonmetallic container. Trapped air bubbles beneath the flange were minimized by placing the probe at a slight angle.
- 3) The complex admittance with respect to the probe aperture was measured
- 4) The complex relative permittivity ε can be calculated from the below equation (Pournaropoulos and Misra):

$$Y = \frac{j2\omega\varepsilon_{r}\varepsilon_{0}}{\left[\ln(b/a)\right]^{2}} \int_{a}^{b} \int_{a}^{b} \int_{0}^{\pi} \cos\phi' \frac{\exp\left[-j\omega r(\mu_{0}\varepsilon_{r}'\varepsilon_{0})^{1/2}\right]}{r} d\phi' d\rho' d\rho$$

where Y is the admittance of the probe in contact with the sample, the primed and unprimed coordinates refer to source and observation points, respectively, $r^2 = \rho^2 + \rho'^2 - 2\rho\rho'\cos\phi'$, ω is the angular frequency, and $j = \sqrt{-1}$.

3 Composition / Information on ingredients

3.2 Mixtures

Description: Aqueous solution with surfactants and inhibitors

Decialable, of flazardous compone	ents.	
CAS: 107-21-1	Ethanediol	>1.0-4.9%
EINECS: 203-473-3	STOT RE 2, H373;	
Reg.nr.: 01-2119456816-28-0000	Acute Tox. 4, H302	
CAS: 68608-26-4	Sodium petroleum sulfonate	< 2.9%
EINECS: 271-781-5	Eye Irrit. 2, H319	
Reg.nr.: 01-2119527859-22-0000		
CAS: 107-41-5	Hexylene Glycol / 2-Methyl-pentane-2,4-diol	< 2.9%
EINECS: 203-489-0	Skin Irrit. 2, H315; Eye Irrit. 2, H319	
Reg.nr.: 01-2119539582-35-0000		
CAS: 68920-66-1	Alkoxylated alcohol, > C ₁₆	< 2.0%
NLP: 500-236-9	Aquatic Chronic 2, H411;	
Reg.nr.: 01-2119489407-26-0000	Skin Irrit. 2, H315; Eye Irrit. 2, H319	

Additional information:

withheld as a trade secret.

For the wording of the listed risk phrases refer to section 16.

Not mentioned CAS-, EINECS- or registration numbers are to be regarded as Proprietary/Confidential. The specific chemical identity and/or exact percentage concentration of proprietary components is

Figure C-1

Note: Liquid recipes are proprietary SPEAG. Since the composition is approximate to the actual liquids utilized, the manufacturer tissue-equivalent liquid data sheets are provided below.

FCC ID A3LSMF936JPN	SAR EVALUATION REPORT	Approved by: Technical Manager	
DUT Type: Portable Handset		APPENDIX C: Page 1 of 4	



Schmid & Partner Engineering AG

s p e a g

5500

4500 5500

-0.9 -6.3

-1.0 -6.2

-2.6 1.7

-2.7 1.9 -2.8 2.2

-3.0 3.2 -3.2 3.6

-3.3 4.2

-3.5 4.6

-3.7 5.5

Zeughausstrasse 43, 8004 Zurich, Switzerland Phone +41 44 245 9700, Fax +41 44 245 9779 www.speag.swiss, info@speag.swiss

Measurement Certificate / Material Test

Item Name Body Tissue Simulating Liquid (MBBL600-6000V6)

Product No. SL AAM U16 BC (Batch: 210621-3)

Manufacturer SPEAG

Measurement Method

TSL dielectric parameters measured using calibrated DAK probe.

Target Parameters

Target parameters as defined in the KDB 865664 compliance standard.

Test Condition

Ambient Condition 22°C; 30% humidity

TSL Temperature 22°C
Test Date 23-June

Test Date 23-Jun-21 Operator WM

Additional Information

TSL Density

TSL Heat-capacity

Results

	Measu	red	1911	Targe	t	Diff.to Tar	get [%]	15.0					
f [MHz]	e'	e"	sigma	eps	sigma	Δ-eps	∆-sigma	10.0			TO LOT		
600	55.7	26.7	0.89	56.1	0.95	-0.7	-6.3	% ≥ 5.0					
750	55.3	22.5	0.94	55.5	0.96	-0.4	-2.1	0.0 5.0					
800	55.1	21.5	0.96	55.3	0.97	-0.4	-1.0	-5.0					
825	55.1	21.1	0.97	55.2	0.98	-0.3	-1.0	2					
835	55.1	20.8	0.97	55.1	0.99	0.0	-1.5	16750			1117		117
850	55.0	20.6	0.97	55.2	0.99	-0.3	-2.0	-15.0	00	1500	2500	3500	4500
900	54.9	19.9	0.99	55.0	1.05	-0.2	-5.7			1500	2500 Freque	ncy MHz	1000
1400	54.1	15.9	1.24	54.1	1.28	0.0	-3.1	15.0	_	-	111120		
1450	54.0	15.7	1.27	54.0	1.30	0.0	-2.3	10.0					
1600	53.8	15.3	1.36	53.8	1.39	0.0	-2.2	» ≥ 5.0		1	A.		
1625	53.8	15.2	1.38	53.8	1.41	0.1	-2.1	Conductivity 5.0 5.0			1		
1640	53.8	15.2	1.39	53.7	1.42	0.1	-2.1	onp.	Λ	~	1		
1650	53.7	15.1	1.39	53.7	1.43	0.0	-2.8		10				
1700	53.7	15.0	1.42	53.6	1.46	0.3	-2.7	5-10.0 G		HILE		1,13 (6)	
1750	53.6	14.9	1.45	53.4	1.49	0.3	-2.7	-15.0	500	1500	2500	3500	4500
1800	53.5	14.9	1.49	53.3	1.52	0.4	-2.0			1000	Freque	3500 ncy MHz	
1810	53.5	14.9	1.50	53.3	1.52	0.4	-1.3	3500	50.9	15.9	3.10	51.3	3.31
1825	53.5	14.8	1.51	53.3	1.52	0.4	-0.7	3700	50.6	16.2	3.33	51.1	3.55
1850	53.5	14.8	1.52	53.3	1.52	0.4	0.0	5200	47.7	18.6	5.39	49.0	5.30
1900	53.4	14.8	1.56	53.3	1.52	0.2	2.6	5250	47.6	18.7	5.46	49.0	5.36
1950	53.4	14.7	1.60	53.3	1.52	0.2	5.3	5300	47.5	18.8	5.54	48.9	5.42
2000	53.3	14.7	1.63	53.3	1.52	0.0	7.2	5500	47.1	19.1	5.83	48.6	5.65
2050	53.3	14.7	1.67	53.2	1.57	0.1	6.4	5600	46.9	19.2	5.98	48.5	5.77
2100	53.2	14.7	1.71	53.2	1.62	0.1	5.6	5700	46.7	19.3	6.13	48.3	5.88
2150	53.1	14.7	1.75	53.1	1.66	0.0	5.4	5800	46.5	19.4	6.27	48.2	6.00
2200	53.1	14.7	1.80	53.0	1.71	0.1	5.3	6000	46.1	19.7	6.57	47.9	6.23
2250	53.0	14.7	1.84	53.0	1.76	0.1	4.5	6500					
2300	52.9	14.7	1.88	52.9	1.81	0.0	3.9	7000			H		
2350	52.9	14.8	1.93	52.8	1.85	0.1	4.3	7500					
2400	52.8	14.8	1.98	52.8	1.90	0.1	4.2	8000			100		
2450	52.7	14.8	2.02	52.7	1.95	0.0	3.6	8500			900		
2500	52.6	14.9	2.07	52.6	2.02	-0.1	2.5	9000			12 1		
2550	52.5	14.9	2.12	52.6	2.09	-0.1	1.4	9500					
	52.5	15.0	2.16	52.5	2.16	0.0	0.0	10000					

Figure C-2 600 – 6000 MHz Body Tissue Equivalent Matter

FCC ID A3LSMF936JPN	SAR EVALUATION REPORT	Approved by: Technical Manager	
DUT Type: Portable Handset		APPENDIX C: Page 2 of 4	



Schmid & Partner Engineering AG

s peag

Zeughausstrasse 43, 8004 Zurich, Switzerland Phone +41 44 245 9700, Fax +41 44 245 9779 www.speag.swiss, info@speag.swiss

Measurement Certificate / Material Test

Item Name Head Tissue Simulating Liquid (HBBL600-10000V6)

Product No. SL AAH U16 BC (Batch: 210629-3)

Manufacturer SPEAG

Measurement Method

TSL dielectric parameters measured using calibrated DAK probe.

Target Parameters

Target parameters as defined in the IEEE 1528 and IEC 62209 compliance standards.

Test Condition

Ambient Condition 22°C; 30% humidity

TSL Temperature 22°C
Test Date 1-Jul-21
Operator WM

Additional Information TSL Density

TSL Heat-capacity

Results

	Measu	easured			et	Diff.to Target [%]		
f [MHz]	e'	e"	sigma	eps	sigma	∆-eps	∆-sigma	
600	44.7	25.5	0.85	42.7	0.88	4.6	-3.6	
750	44.1	21.6	0.90	41.9	0.89	5.1	0.7	
800	44.0	20.6	0.92	41.7	0.90	5.6	2.5	
825	44.0	20.2	0.93	41.6	0.91	5.8	2.6	
835	44.0	20.0	0.93	41.5	0.91	5.9	2.0	
850	43.9	19.8	0.93	41.5	0.92	5.8	1.5	
900	43.8	19.0	0.95	41.5	0.97	5.5	-2.1	
1400	42.8	15.1	1.18	40.6	1.18	5.4	0.0	
1450	42.7	14.9	1.20	40.5	1.20	5.4	0.0	
1600	42.4	14.4	1.28	40.3	1.28	5.2	-0.3	
1625	42.4	14.3	1.30	40.3	1.30	5.3	0.1	
1640	42.4	14.3	1.31	40.3	1.31	5.3	0.3	
1650	42.3	14.3	1.31	40.2	1.31	5.1	-0.2	
1700	42.3	14.2	1.34	40.2	1.34	5.3	-0.2	
1750	42.2	14.1	1.37	40.1	1.37	5.3	-0.1	
1800	42.1	14.0	1.40	40.0	1.40	5.3	0.0	
1810	42.1	13.9	1.41	40.0	1.40	5.3	0.7	
1825	42.1	13.9	1.42	40.0	1.40	5.3	1.4	
1850	42.0	13.9	1.43	40.0	1.40	5.0	2.1	
1900	42.0	13.8	1.46	40.0	1.40	5.0	4.3	
1950	41.9	13.8	1.49	40.0	1.40	4.7	6.4	
2000	41.8	13.7	1.53	40.0	1.40	4.5	9.3	
2050	41.8	13.7	1.56	39.9	1.44	4.7	8.0	
2100	41.7	13.7	1.59	39.8	1.49	4.7	6.8	
2150	41.6	13.6	1.63	39.7	1.53	4.7	6.3	
2200	41.6	13.6	1.67	39.6	1.58	4.9	5.8	
2250	41.5	13.6	1.70	39.6	1.62	4.9	4.8	
2300	41.4	13.6	1.74	39.5	1.67	4.9	4.4	
2350	41.3	13.6	1.78	39.4	1.71	4.9	4.0	
2400	41.3	13.6	1.82	39.3	1.76	5.1	3.7	
2450	41.2	13.6	1.86	39.2	1.80	5.1	3.3	
2500	41.1	13.6	1.90	39.1	1.85	5.0	2.5	
2550	41.0	13.7	1.94	39.1	1.91	4.9	1.6	
2600	41.0	13.7	1.98	39.0	1.96	5.1	0.8	




Figure C-3 600 – 10000 MHz Head Tissue Equivalent Matter

FCC ID A3LSMF936JPN	SAR EVALUATION REPORT	Approved by: Technical Manager	
DUT Type: Portable Handset		APPENDIX C: Page 3 of 4	



e a <u>g</u> S p

Zeughausstrasse 43, 8004 Zurich, Switzerland Phone +41 44 245 9700, Fax +41 44 245 9779 www.speag.swiss, info@speag.swiss

Measurement Certificate / Material Test

Item Name	Head Tissue Simulating Liquid (HBBL4-250V3)
Product No.	SL AAH 005 AD (Batch: 210601-1)
Manufacturer	SPEAG

Measurement Method

TSL dielectric parameters measured using calibrated DAK probe.

Setup Validation

Validation results were within $\pm 2.5\%$ towards the target values of Methanol.

Target Parameters

Target parameters as defined in the IEEE 1528 and IEC 62209 compliance standards.

Test Condition Environment temperatur (22 ± 3)°C and humidity < 70% Ambient 22°C TSL Temperature 3-Jun-21 Operator WM

49.7 49.6 77.41 75.52 73.74

49.5 49.4

49.3 49.2 70.46

49.2 67.49

48.7 60.11

48.5 58.05

220 225 49.0 64,80

66.11


63.55 48.8 62.35


61.20

59.05 0.80

Additional Information
TSL Density 1.0 TSL Density 1.042 g/cm3
TSL Heat-capacity 3.574 kJ/(kg*K)

f [MHz]	ď	o"	sigma	eps	sigma	∆-aps	∆-sigma	
5	53.0	2603.50	0.72	55.0	0.75	-3.6	-4.0	
10	52.9	1301.62	0.72	55.0	0.75	-3.8	-4.0	
15	52.9	858.41	0.72	55.0	0.75	-3.8	-4.0	
20	52.8	651.83	0,73	55.0	0.75	-4.0	-2.7	
25	52.8	521.90	0.73	55.0	0.75	-4.0	-2.7	
30	52.7	435.32	0.73	55.0	0.75	-4.2	-2.7	
35	52.6	373.51	0.73	54.9	0.75	-4.2	-2.7	
40	52.5	327.19	0.73	54.8	0.75	-4.2	-2.7	
45	52.4	291,20	0.73	54.7	0.75	-4.1	-2.7	
	500	202.44	0.73	546	0.75	-4.1	-2.7	





0.78 0.78 0.79 0.79 50.2 0.79 50.0 0.80 -1.6 -1.3 -1,1

0.79 0.79 0.80 0.60 0.80 0.80

49.7 0.80 49.5 0.80 49.3 0.81 49.0 0.81

48.8 0.81 48.6 0.82 48.3 0.82 48.1 0.82 47.9 0.83

-2.6

-0,3

0.2 -1.7

1.0

1.5

-0.6 -2.1 -1.0 -1.5

> -1.2 -1.6 -2.1

> > -2.6

-3.4




Figure C-4 4 - 250 MHz Head Tissue Equivalent Matter

FCC ID A3LSMF936JPN	SAR EVALUATION REPORT	Approved by: Technical Manager	
DUT Type: Portable Handset		APPENDIX C: Page 4 of 4	

Page 1 of 1