

APPENDIX D: SAR TISSUE SPECIFICATIONS

Measurement Procedure for Tissue verification:

- 1) The network analyzer and probe system was configured and calibrated.
- The probe was immersed in the tissue. The tissue was placed in a nonmetallic container. Trapped air bubbles beneath the flange were minimized by placing the probe at a slight angle.
- 3) The complex admittance with respect to the probe aperture was measured
- The complex relative permittivity ε can be calculated from the below equation (Pournaropoulos) and Misra):

$$Y = \frac{j2\omega\varepsilon_{r}\varepsilon_{0}}{\left[\ln(b/a)\right]^{2}} \int_{a}^{b} \int_{a}^{b} \int_{0}^{\pi} \cos\phi' \frac{\exp\left[-j\omega r(\mu_{0}\varepsilon_{r}\varepsilon_{0})^{1/2}\right]}{r} d\phi' d\rho' d\rho$$

where Y is the admittance of the probe in contact with the sample, the primed and unprimed coordinates refer to source and observation points, respectively, $r^2 = \rho^2 + \rho'^2 - 2\rho\rho'\cos\phi'$, ω is the angular frequency, and $j = \sqrt{-1}$.

3 Composition / Information on ingredients

3.2 Mixtures

Description: Aqueous solution with surfactants and inhibitors Declarable, or hazardous components:

CAS: 107-21-1	Ethanediol	>1.0-4.9%
EINECS: 203-473-3	STOT RE 2, H373;	
Reg.nr.: 01-2119456816-28-0000	Acute Tox. 4, H302	
CAS: 68608-26-4	Sodium petroleum sulfonate	< 2.9%
EINECS: 271-781-5	Eye Irrit. 2, H319	
Reg.nr.: 01-2119527859-22-0000		
CAS: 107-41-5	Hexylene Glycol / 2-Methyl-pentane-2,4-diol	< 2.9%
EINECS: 203-489-0	Skin Irrit. 2, H315; Eye Irrit. 2, H319	
Reg.nr.: 01-2119539582-35-0000		
CAS: 68920-66-1	Alkoxylated alcohol, > C ₁₆	< 2.0%
NLP: 500-236-9	Aquatic Chronic 2, H411;	
Reg.nr.: 01-2119489407-26-0000	Skin Irrit. 2, H315; Eye Irrit. 2, H319	

Additional information:

For the wording of the listed risk phrases refer to section 16.

Not mentioned CAS-, EINECS- or registration numbers are to be regarded as Proprietary/Confidential. The specific chemical identity and/or exact percentage concentration of proprietary components is

withheld as a trade secret.

Figure D-1

Note: Liquid recipes are proprietary SPEAG. Since the composition is approximate to the actual liquids utilized, the manufacturer tissue-equivalent liquid data sheets are provided below.

FCC ID A3LSMS911U	SAR EVALUATION REPORT	Approved by: Technical Manager	
DUT Type: Portable Handset		APPENDIX D: Page 1 of 4	



Schmid & Partner Engineering AG

S p e a g

5500

5500

-6.2 -1.0 5.30 -2.6 1.7

3.6

4500

3.31 -0.9 -6.3

3.55

5.36 -2.7 1.9 -2.8 2.2

5.42 5.65 -3.0 3.2 -3.2 5.77

5.88 -3.3 4.2

6.00 -3.5 4.6 -3.7 5.5

6.23

Zeughausstrasse 43, 8004 Zurich, Switzerland Phone +41 44 245 9700, Fax +41 44 245 9779 www.speag.swiss, info@speag.swiss

Measurement Certificate / Material Test

Body Tissue Simulating Liquid (MBBL600-6000V6)

Item Name SL AAM U16 BC (Batch: 210621-3) Product No.

SPEAG Manufacturer

Measurement Method

TSL dielectric parameters measured using calibrated DAK probe.

Target Parameters

Target parameters as defined in the KDB 865664 compliance standard.

Test Condition

Ambient Condition 22°C; 30% humidity

TSL Temperature 22°C Test Date 23-Jun-21

Operator WM

Additional Information
TSL Density

TSL Heat-capacity

	Measu	ired		Targe	t	Diff.to Targ	get [%]	15.0				-
[MHz]	e'	е"	sigma	eps	sigma	∆-eps	∆-sigma	10.0				
600	55.7	26.7	0.89	56.1	0.95	-0.7	-6.3	» > 5.0				
750	55.3	22.5	0.94	55.5	0.96	-0.4	-2.1	Permittivity 0.0				
800	55.1	21.5	0.96	55.3	0.97	-0.4	-1.0	-5.0		J. Link		
825	55.1	21.1	0.97	55.2	0.98	-0.3	-1.0					
835	55.1	20.8	0.97	55.1	0.99	0.0	-1.5	-10.0				
850	55.0	20.6	0.97	55.2	0.99	-0.3	-2.0	-15.0	00	1500	2500	36
900	54.9	19.9	0.99	55.0	1.05	-0.2	-5.7			1500	2500 Freque	ency M
1400	54.1	15.9	1.24	54.1	1.28	0.0	-3.1	15.0			-17	
1450	54.0	15.7	1.27	54.0	1.30	0.0	-2.3	10.0				
1600	53.8	15.3	1.36	53.8	1.39	0.0	-2.2	%		1	-	
1625	53.8	15.2	1.38	53.8	1.41	0.1	-2.1	Conductivity 0.0 2-5-		1	1	
1640	53.8	15.2	1.39	53.7	1.42	0.1	-2.1	onpuo -5.0	Λ	1	1	
1650	53.7	15.1	1.39	53.7	1.43	0.0	-2.8		1			-
1700	53.7	15.0	1.42	53.6	1.46	0.3	-2.7	-10.0 G		11144	1 301	
1750	53.6	14.9	1.45	53.4	1.49	0.3	-2.7	-15.0	00	1500	2500	3
1800	53.5	14.9	1.49	53.3	1.52	0.4	-2.0				2500 Freque	ncy M
1810	53.5	14.9	1.50	53.3	1.52	0.4	-1.3	3500	50.9	15.9	3.10	51.
1825	53.5	14.8	1.51	53.3	1.52	0.4	-0.7	3700	50.6	16.2	3.33	51.
1850	53.5	14.8	1.52	53.3	1.52	0.4	0.0	5200	47.7	18.6	5.39	49.
1900	53.4	14.8	1.56	53.3	1.52	0.2	2.6	5250	47.6	18.7	5.46	49.
1950	53.4	14.7	1.60	53.3	1.52	0.2	5.3	5300	47.5	18.8	5.54	48.
2000	53.3	14.7	1.63	53.3	1.52	0.0	7.2	5500	47.1	19.1	5.83	48.
2050	53.3	14.7	1.67	53.2	1.57	0.1	6.4	5600	46.9	19.2	5.98	48.
2100	53.2	14.7	1.71	53.2	1.62	0.1	5.6	5700	46.7	19.3	6.13	48.
2150	53.1	14.7	1.75	53.1	1.66	0.0	5.4	5800	46.5	19.4	6.27	48.
2200	53.1	14.7	1.80	53.0	1.71	0.1	5.3	6000	46.1	19.7	6.57	47.
2250	53.0	14.7	1.84	53.0	1.76	0.1	4.5	6500				
2300	52.9	14.7	1.88	52.9	1.81	0.0	3.9	7000			1	
2350	52.9	14.8	1.93	52.8	1.85	0.1	4.3	7500				
2400	52.8	14.8	1.98	52.8	1.90	0.1	4.2	8000			200	
2450	52.7	14.8	2.02	52.7	1.95	0.0	3.6	8500			100	
2500	52.6	14.9	2.07	52.6	2.02	-0.1	2.5	9000				
2550	52.5	14.9	2.12	52.6	2.09	-0.1	1.4	9500				
2600	52.5	15.0	2.16	52.5	2.16	0.0	0.0	10000			The state of	

Figure D-2 600 - 6000 MHz Body Tissue Equivalent Matter

FCC ID A3LSMS911U	SAR EVALUATION REPORT	Approved by: Technical Manager	
DUT Type: Portable Handset		APPENDIX D: Page 2 of 4	



Schmid & Partner Engineering AG



Zeughausstrasse 43, 8004 Zurich, Switzerland Phone +41 44 245 9700, Fax +41 44 245 9779 www.speag.swiss, info@speag.swiss

Measurement Certificate / Material Test

Head Tissue Simulating Liquid (HBBL600-10000V6)

Product No. SL AAH U16 BC (Batch: 210629-3)

Manufacturer SPEAG

Measurement Method

TSL dielectric parameters measured using calibrated DAK probe.

Target Parameters

Target parameters as defined in the IEEE 1528 and IEC 62209 compliance standards.

Test Condition

Ambient Condition 22°C; 30% humidity

TSL Temperature 22°C Operator WM

Additional Information

TSL Density

TSL Heat-capacity

	Measu	Measured			et	Diff.to Tar	jet [%]	
f [MHz]	e'	е"	sigma	eps	sigma	∆-eps	∆-sigma	
600	44.7	25.5	0.85	42.7	0.88	4.6	-3.6	
750	44.1	21.6	0.90	41.9	0.89	5.1	0.7	
800	44.0	20.6	0.92	41.7	0.90	5.6	2.5	
825	44.0	20.2	0.93	41.6	0.91	5.8	2.6	
835	44.0	20.0	0.93	41.5	0.91	5.9	2.0	
850	43.9	19.8	0.93	41.5	0.92	5.8	1.5	
900	43.8	19.0	0.95	41.5	0.97	5.5	-2.1	
1400	42.8	15.1	1.18	40.6	1.18	5.4	0.0	
1450	42.7	14.9	1.20	40.5	1.20	5.4	0.0	
1600	42.4	14.4	1.28	40.3	1.28	5.2	-0.3	
1625	42.4	14.3	1.30	40.3	1.30	5.3	0.1	
1640	42.4	14.3	1.31	40.3	1.31	5.3	0.3	
1650	42.3	14.3	1.31	40.2	1.31	5.1	-0.2	
1700	42.3	14.2	1.34	40.2	1.34	5.3	-0.2	
1750	42.2	14.1	1.37	40.1	1.37	5.3	-0.1	
1800	42.1	14.0	1.40	40.0	1.40	5.3	0.0	
1810	42.1	13.9	1.41	40.0	1.40	5.3	0.7	
1825	42.1	13.9	1.42	40.0	1.40	5.3	1.4	
1850	42.0	13.9	1.43	40.0	1.40	5.0	2.1	
1900	42.0	13.8	1.46	40.0	1.40	5.0	4.3	
1950	41.9	13.8	1.49	40.0	1.40	4.7	6.4	
2000	41.8	13.7	1.53	40.0	1.40	4.5	9.3	
2050	41.8	13.7	1.56	39.9	1.44	4.7	8.0	
2100	41.7	13.7	1.59	39.8	1.49	4.7	6.8	
2150	41.6	13.6	1.63	39.7	1.53	4.7	6.3	
2200	41.6	13.6	1.67	39.6	1.58	4.9	5.8	
2250	41.5	13.6	1.70	39.6	1.62	4.9	4.8	
2300	41.4	13.6	1.74	39.5	1.67	4.9	4.4	
2350	41.3	13.6	1.78	39.4	1.71	4.9	4.0	
2400	41.3	13.6	1.82	39.3	1.76	5.1	3.7	
2450	41.2	13.6	1.86	39.2	1.80	5.1	3.3	
2500	41.1	13.6	1.90	39.1	1.85	5.0	2.5	
2550	41.0	13.7	1.94	39.1	1.91	4.9	1.6	
2600	41.0	13.7	1.98	39.0	1.96	5.1	0.8	




Figure D-3 600 - 10000 MHz Head Tissue Equivalent Matter

FCC ID A3LSMS911U	SAR EVALUATION REPORT	Approved by: Technical Manager	
DUT Type: Portable Handset		APPENDIX D: Page 3 of 4	



e a 9 Ø S Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland Phone +41 44 245 9700, Fax +41 44 245 9779 www.speag.swiss, info@speag.swiss Measurement Certificate / Material Test Head Tissue Simulating Liquid (HBBL4-250V3) item Name Product No. SL AAH 005 AD (Batch: 210601-1) Manufacturer SPEAG Measurement Method
TSL dielectric parameters measured using calibrated DAK probe. $\begin{tabular}{ll} Setup Validation \\ \hline Validation results were within $\pm 2.5\%$ towards the target values of Methanol. \\ \hline \end{tabular}$ Target Parameters
Target parameters as defined in the IEEE 1528 and IEC 62209 compliance standards Test Condition
Ambient
TSL Temperature Environment temperatur (22 ± 3)°C and humidity < 70%. 3-Jun-21 Test Date WM Additional Information TSL Density 1.042 g/cm3
TSL Heat-capacity 3.574 kJ/(kg*K) 10.0 -7.5 -5.0 -2.5 -0.0 -Dev. Permitthity % -2.5 -5.0 -7.5 -10.0 5 25 45 65 7.5 5.0 2.5 -2.5 -10.0 105 125 145 165 185 205 225 245 Frequency MHz

Figure D-4
5 – 250 MHz Head Tissue Equivalent Matter

Pana 1 of 1

FCC ID A3LSMS911U	SAR EVALUATION REPORT	Approved by: Technical Manager
DUT Type: Portable Handset		APPENDIX D: Page 4 of 4