

Model Difference declaration

Centre Testing International Group Co., Ltd:

There are four models of Dynamic ECG recorder products submitted by our company for inspection, including the main inspection model [EPatch-B] and the covering models [EPatch-A, EHL-1B, EHL-1A]. The principle, structure and composition, intended use, key components, and electrical structure of the products are the same, and only the functional configuration is different. Therefore, the model with the most complete functions [EPatch-B] was selected as the main inspection model to cover the other models.

The specific differences are as follows:

Model	EHL-1A	EHL-1B	EPatch-A	EPatch-B	remark
display screen	√	√	√	√	
A single data record time	168h	168h	168h	168h	
Store the maximum number of data strips	1	10	1	10	
Battery run time	168h	168h	168h	168h	
Enclosure Color	black	black	white	white	
bluetooth	YES	YES	YES	YES	
Heart Rate measurement	YES	YES	YES	YES	

Shenzhen Creative Industry Co., Ltd.

2023-07-25

Signature: *Zhu Wanguo*