





DFT-s-OFDM Module B

NR Band n261 / 100MHz / 16QAM

Lowest Band Edge / Full RB




Highest Band Edge / Full RB




NR Band n261 / 100MHz / 64QAM

Lowest Band Edge / Full RB



Highest Band Edge / Full RB







DFT-s-OFDM Module B

NR Band n261 / 200MHz / QPSK

Lowest Band Edge / Full RB




Highest Band Edge / Full RB




NR Band n261 / 200MHz / 16QAM

Lowest Band Edge / Full RB



Highest Band Edge / Full RB






DFT-s-OFDM Module B


NR Band n261 / 200MHz / 64QAM

Lowest Band Edge / Full RB



10:46:30 23.01.2024


Highest Band Edge / Full RB



17:01:10 24.01.2024


NR Band n261 / 300MHz / QPSK

Lowest Band Edge / Full RB



11:16:23 23.01.2024

Highest Band Edge / Full RB




17:46:52 24.01.2024



DFT-s-OFDM Module B


NR Band n261 / 300MHz / 16QAM

Lowest Band Edge / Full RB



11:13:38 23.01.2024


Highest Band Edge / Full RB



17:45:30 24.01.2024


NR Band n261 / 300MHz / 64QAM

Lowest Band Edge / Full RB



11:12:32 23.01.2024

Highest Band Edge / Full RB



17:41:43 24.01.2024





DFT-s-OFDM Module B

NR Band n261 / 400MHz / QPSK

Lowest Band Edge / Full RB



Highest Band Edge / Full RB



NR Band n261 / 400MHz / 16QAM


Lowest Band Edge / Full RB

Highest Band Edge / Full RB





DFT-s-OFDM Module B






CP-OFDM Module B


NR Band n261 / 50MHz / QPSK

Lowest Band Edge / Full RB



10:11:28 23. 01. 2024


Highest Band Edge / Full RB



15:54:11 24. 01. 2024


NR Band n261 / 100MHz / QPSK

Lowest Band Edge / Full RB



06:50:55 23. 01. 2024

Highest Band Edge / Full RB




16:38:19 24. 01. 2024



CP-OFDM Module B


NR Band n261 / 200MHz / QPSK

Lowest Band Edge / Full RB



10:49:33 23.01.2024


Highest Band Edge / Full RB



17:11:58 24.01.2024


NR Band n261 / 300MHz / QPSK

Lowest Band Edge / Full RB



11:21:49 23.01.2024

Highest Band Edge / Full RB



17:48:55 24.01.2024





CP-OFDM Module B

NR Band n261 / 400MHz / QPSK

Lowest Band Edge / Full RB

Highest Band Edge / Full RB






Spurious Emission

NR Band n261 (30MHz-1GHz)


Horizontal



Site : 03CH10-HY
 Condition : -13 EIRP_WO HORIZONTAL
 Project : 3N2326
 : n261 MB

| Freq | Level | Over | Limit |
|------|-------|--------|---------------|
| MHz | dBm | dB | dBm |
| 1 | 63.95 | -55.73 | -42.73 -13.00 |

Vertical




Site : 03CH10-HY
 Condition : -13 EIRP_WO VERTICAL
 Project : 3N2326
 : n261 MB

| Freq | Level | Over | Limit |
|------|-------|--------|---------------|
| MHz | dBm | dB | dBm |
| 1 | 63.95 | -59.88 | -46.88 -13.00 |



NR Band n261 (1GHz-18GHz)


Horizontal



Site : 03CH10-HY
 Condition : -13 EIRP_WO HORIZONTAL
 Project : 3N2326
 : n261 MB

| Freq | Level | Over | Limit |
|-----------|--------|--------|--------|
| MHz | dBm | dB | dBm |
| 1 8922.00 | -48.94 | -35.94 | -13.00 |

Vertical



Site : 03CH10-HY
 Condition : -13 EIRP_WO VERTICAL
 Project : 3N2326
 : n261 MB

| Freq | Level | Over | Limit |
|-----------|--------|--------|--------|
| MHz | dBm | dB | dBm |
| 1 4791.00 | -47.86 | -34.86 | -13.00 |




Spurious emission between 18GHz to 40GHz worst case plot is reported as following.


DFT-s-OFDM Module B

NR Band n261 QPSK (18-40GHz)


Lowest Channel / 50MHz




Lowest Channel / 100MHz




Middle Channel / 50MHz




Middle Channel / 100MHz



Highest Channel / 50MHz



Highest Channel / 100MHz




Remark: In band and out of band frequencies are omitted.




DFT-s-OFDM Module B

NR Band n261 QPSK (18-40GHz)


Lowest Channel / 200MHz




Lowest Channel / 300MHz




Middle Channel / 200MHz




Middle Channel / 300MHz



Highest Channel / 200MHz



Highest Channel / 300MHz




Remark: In band and out of band frequencies are omitted.



DFT-s-OFDM Module B


NR Band n261 QPSK (18-40GHz)

Lowest Channel / 400MHz




intentionally blank

Middle Channel / 400MHz



intentionally blank

Highest Channel / 400MHz




intentionally blank

Remark: In band and out of band frequencies are omitted.



CP-OFDM Module B




Remark: In band and out of band frequencies are omitted.




CP-OFDM Module B

NR Band n261 QPSK (18-40GHz)


Lowest Channel / 200MHz




Lowest Channel / 300MHz




Middle Channel / 200MHz




Middle Channel / 300MHz



Highest Channel / 200MHz



Highest Channel / 300MHz




Remark: In band and out of band frequencies are omitted.



CP-OFDM Module B


NR Band n261 QPSK (18-40GHz)

Lowest Channel / 400MHz




intentionally blank

Middle Channel / 400MHz




intentionally blank

Highest Channel / 400MHz



intentionally blank

Remark: In band and out of band frequencies are omitted.




$$\text{Offset} = \text{Antenna Factor (dB/m)} + \text{Cable Loss (dB)} + 107 + 20\log(D) - 104.8$$
$$= 43 + 0.43 + 107 + 20\log(1) - 104.8 = 45.63(\text{dB})$$



NR Band n261

(60GHz-90GHz)




$$\text{Offset} = \text{Antenna Factor (dB/m)} + \text{Cable Loss (dB)} + 107 + 20\log(D) - 104.8$$
$$= 46.9 + 0.43 + 107 + 20\log(1) - 104.8 = 49.53 \text{ (dB)}$$



NR Band n261

(90GHz-100GHz)



$$\text{Offset} = \text{Antenna Factor (dB/m)} + \text{Cable Loss (dB)} + 107 + 20\log(D) - 104.8$$
$$= 47.92 + 0.43 + 107 + 20\log(0.5) - 104.8 = 44.53 \text{ (dB)}$$



Frequency Stability

| Test Conditions | | NR Band n261 / Middle Channel | | | Limit |
|------------------|-------------------|-------------------------------|-----------------|-----------------|---------|
| Temperature (°C) | Voltage (Volt) | CW tone | | | Note 2. |
| | | Frequency (GHz) | Deviation (kHz) | Deviation (ppm) | Result |
| 50 | Normal Voltage | 27.9249491 | 17.900 | 0.641 | Pass |
| 40 | Normal Voltage | 27.9249241 | 42.900 | 1.536 | |
| 30 | Normal Voltage | 27.9249321 | 34.900 | 1.250 | |
| 20(Ref.) | Normal Voltage | 27.924967 | 0.000 | 0.000 | |
| 10 | Normal Voltage | 27.925036 | -69.000 | 2.471 | |
| 0 | Normal Voltage | 27.9251089 | -141.900 | 5.081 | |
| -10 | Normal Voltage | 27.9251189 | -151.900 | 5.440 | |
| -20 | Normal Voltage | 27.9250709 | -103.900 | 3.721 | |
| -30 | Normal Voltage | 27.9250709 | -103.900 | 3.721 | |
| 20 | Maximum Voltage | 27.924969 | -2.000 | 0.072 | |
| 20 | Normal Voltage | 27.924966 | 1.000 | 0.036 | |
| 20 | Battery End Point | 27.924964 | 3.000 | 0.107 | |

Note:

1. Normal Voltage = 3.89 V. ; Battery End Point (BEP) = 3.6 V. ; Maximum Voltage = 4.4 V.
2. The frequency fundamental emissions stay within the operation band.



Appendix B. R&S Mixer and Horn Antenna Calibration Reports




CALIBRATION CERTIFICATE



Kalibrierschein

Certificate Number
Zertifikatsnummer

0001A300718726

| General Data | |
|---|--|
| Item Gegenstand | FS-Z60 HARMONIC MIXER 40-60GHZ |
| Manufacturer Hersteller | ROHDE & SCHWARZ |
| Type Typ | FS-Z60 |
| Material Number Materialnummer | 1048.0171.02 |
| Serial Number Seriennummer | 100986 |
| Order Number Bestellnummer | Q786007 |
| Asset Number Inventarnummer | |
| Customer Auftraggeber | Sporton International Inc. 6F., Sec. 1, Hsin Tai Wu Rd 106 221 New Taipei City TW |
| Performance | |
| Place and Date of Calibration Ort und Datum der Kalibrierung | 87700 Memmingen, Rohde-und-Schwarz-Str. 1 2023-10-31 |
| Statement of Compliance (Incoming) Konformitätsaussage (Anlieferung) | All measured values are within the data sheet specifications. |
| Statement of Compliance (Outgoing) Konformitätsaussage (Auslieferung) | All measured values are within the data sheet specifications. |
| Customers due Interval Kalibrierintervall des Kunden | |
| Extent of Calibration Document Umfang des Kalibrierdokuments | 3 Pages Certificate 6 Pages Outgoing Results 6 Pages Incoming Results |
| Date of Issue Ausstellungsdatum | Approval of the certificate by Freigabe des Kalibrierscheins durch |
| 2023-11-08 | Dr. Gerhard Rösel Johannes Negele |
| |  Laboratory management Labormanagement |
| |  Person responsible Bearbeiter |

Calibration Mark
Kalibrierzeichen

| |
|---------------------|
| 300718726 |
| D-K- 15195-01-00 |
| 2023-10 |

Member of Deutscher Kalibrierdienst
Mitglied im Deutschen Kalibrierdienst



This calibration certificate documents the metrological traceability to national standards, which realize the units of measurement according to the International System of Units (SI). The DAkkS is signatory to the multilateral agreements of the European co-operation for Accreditation (EA) and of the International Laboratory Accreditation Cooperation (ILAC) for the mutual recognition of calibration certificates. The user is obliged to have the object recalibrated at appropriate intervals. This calibration certificate may not be reproduced other than in full except with the permission of the issuing laboratory. Calibration certificates with the full name of the approval responsible person are valid without signature.

Dieser Kalibrierschein dokumentiert die metrologische Rückführbarkeit auf nationale Normale zur Darstellung der Einheiten in Übereinstimmung mit dem Internationalen Einheitensystem (SI). Die DAkkS ist Unterzeichner der multilateralen Übereinkommen der European cooperation for Accreditation (EA) und der International Laboratory Accreditation Cooperation (ILAC) zur gegenseitigen Anerkennung der Kalibrierscheine. Für die Einhaltung einer angemessenen Frist zur Wiederholung der Kalibrierung ist der Benutzer verantwortlich. Dieser Kalibrierschein darf nur vollständig weiterverbreitet werden. Auszüge oder Änderungen bedürfen der Genehmigung des ausstellenden Kalibrierlaboratoriums. Kalibrierscheine sind bei Nennung des für die Freigabe Verantwortlichen in Klarschrift auch ohne Unterschrift gültig.



Material No 1048.0171.02 **Serial No** 100986 **Certificate** 0001A300718726
Page 2/3 **Number**

Calibration Procedure

The measuring object is an RF harmonic mixer, which converts an RF signal at one frequency into a signal at another frequency (here: IF). The conversion loss was measured using a vector network analyzer. The RF output power as well as the IF input power of the corresponding ports of the VNA were traced back to a power sensor. The conversion loss is defined as the ratio of the power at the IF frequency to the power at the RF frequency with a given LO power. (IF: Intermediate frequency; LO: Local Oscillator)
The traceability is represented in the table Working Standards used.

Working Standards used

| Item | Type | Serial Number | Calibration Certificate Number | Cal. Due |
|--------------------------------|---------|---------------|--------------------------------|------------|
| Therm.Power Sensor DC-40GHz | NRP-Z55 | 130179 | 0001A300682928 | 2025-01-31 |
| Thermal Power Sensor | NRP67T | 100977 | 0001A300658977 | 2025-07-31 |
| Vector Network Analyzer 4 Port | ZVA67 | 101100 | 0001A300698540 | 2024-05-31 |

Remarks



Material No 1048.0171.02 **Serial No** 100986 **Certificate** 0001A300718726
Page 3/3 **Number**

| Environmental Conditions | | | |
|----------------------------|-------------|--------------------------|---------|
| Ambient Temperature | (23 ± 1) °C | Relative Humidity | 20%-70% |

| Comments on Measurement Results |
|---|
| <p>The reported results apply only to those items specifically listed on this calibration certificate and have been tested for compliance with the specifications. The associated uncertainty of measurement has been taken into account if not otherwise stated.</p> <p>The non-binary decision rule with guard band is used according to ILAC G8:09/2019 'Guidelines on Decision Rules and Statements of Conformity'. Pass is normally not marked. Conditional Pass is marked with UGB1, Conditional Fail with UGB2 and Fail with Fail.</p> <p>The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k such that the coverage probability corresponds to approximately 95 %. It is consistent with the EA-4/02 M:2022.</p> <p>In addition to the calibration results, the calibration certificate includes functional measurements that might have an influence on the measurement uncertainty of the calibration results.</p> <p>The functional measurement results are marked and are not intended to be used to support the further dissemination of metrological traceability. They are intended to verify the requirements on the measurement object according to manufacturer specifications and technical standards.</p> |

Outgoing Results

Designation: HARMONIC MIXER
Type: FS-Z60
Material No.: 1048.0171.02
Serial No.: 100986
Certificate No.: 0001A300718726
Referring to Test Documentation: 5038.8581.01-PB-02.00

Test Department: 3MM-P
Name: see certificate
Date: 2023-10-31



Table of contents


| | |
|--|---|
| Software used for measurement | 3 |
| 1. Conversion Loss (4. Harmonic) | 4 |
| 1.1 Conversion Loss (IF = 404.4 MHz) | 4 |
| 1.2 Conversion Loss (IF = 729 MHz) | 5 |
| 1.3 Conversion Loss (IF = 1330 MHz) | 6 |
| 1.4 Continuity response within 1 GHz | 6 |

| Software used for measurement | | | |
|--------------------------------------|-------------|----------------|-----------------------------|
| Item | Type | Version | Remark |
| Suite | Setup | V12.37.04 | Test Management Software G5 |
| Test Program (7012.8706.00_) | Component | V01.05 | |

1. Conversion Loss (4. Harmonic)

1.1 Conversion Loss (IF = 404.4 MHz)

IF = 404.4 MHz, 4. Harmonic




Conversion Loss (IF = 404.4 MHz)

| Frequency /GHz | Conversion Loss /dB | Uncertainty /dB |
|----------------|---------------------|-----------------|
| 40.0 | 10.0 | ±1.3 |
| 41.0 | 10.4 | ±1.3 |
| 42.0 | 11.9 | ±1.3 |
| 43.0 | 13.0 | ±1.3 |
| 44.0 | 13.3 | ±1.3 |
| 45.0 | 14.0 | ±1.3 |
| 46.0 | 15.1 | ±1.3 |
| 47.0 | 14.3 | ±1.3 |
| 48.0 | 14.9 | ±1.3 |
| 49.0 | 14.7 | ±1.3 |
| 50.0 | 12.5 | ±1.3 |
| 51.0 | 11.2 | ±1.3 |
| 52.0 | 11.1 | ±1.3 |
| 53.0 | 10.1 | ±1.3 |
| 54.0 | 10.0 | ±1.3 |
| 55.0 | 10.0 | ±1.3 |
| 56.0 | 10.7 | ±1.3 |
| 57.0 | 11.2 | ±1.3 |
| 58.0 | 11.7 | ±1.3 |
| 59.0 | 12.5 | ±1.3 |
| 60.0 | 14.0 | ±1.3 |

1.2 Conversion Loss (IF = 729 MHz)

IF = 729 MHz, 4. Harmonic




Conversion Loss (IF = 729 MHz)

| Frequency /GHz | Conversion Loss /dB | Uncertainty /dB |
|----------------|---------------------|-----------------|
| 40.0 | 9.9 | ±1.3 |
| 41.0 | 10.2 | ±1.3 |
| 42.0 | 11.6 | ±1.3 |
| 43.0 | 12.6 | ±1.3 |
| 44.0 | 12.9 | ±1.3 |
| 45.0 | 13.4 | ±1.3 |
| 46.0 | 15.1 | ±1.3 |
| 47.0 | 13.9 | ±1.3 |
| 48.0 | 15.0 | ±1.3 |
| 49.0 | 14.7 | ±1.3 |
| 50.0 | 12.9 | ±1.3 |
| 51.0 | 11.4 | ±1.3 |
| 52.0 | 11.2 | ±1.3 |
| 53.0 | 10.0 | ±1.3 |
| 54.0 | 10.0 | ±1.3 |
| 55.0 | 10.0 | ±1.3 |
| 56.0 | 10.7 | ±1.3 |
| 57.0 | 11.2 | ±1.3 |
| 58.0 | 11.6 | ±1.3 |
| 59.0 | 12.6 | ±1.3 |
| 60.0 | 13.9 | ±1.3 |

1.3 Conversion Loss (IF = 1330 MHz)

IF = 1330 MHz, 4. Harmonic



Conversion Loss (IF = 1330 MHz)

| Frequency /GHz | Conversion Loss /dB | Uncertainty /dB |
|----------------|---------------------|-----------------|
| 40.0 | 10.3 | ±1.3 |
| 41.0 | 10.5 | ±1.3 |
| 42.0 | 11.9 | ±1.3 |
| 43.0 | 13.2 | ±1.3 |
| 44.0 | 13.1 | ±1.3 |
| 45.0 | 13.4 | ±1.3 |
| 46.0 | 14.8 | ±1.3 |
| 47.0 | 14.1 | ±1.3 |
| 48.0 | 15.0 | ±1.3 |
| 49.0 | 16.1 | ±1.3 |
| 50.0 | 13.9 | ±1.3 |
| 51.0 | 12.8 | ±1.3 |
| 52.0 | 11.8 | ±1.3 |
| 53.0 | 10.5 | ±1.3 |
| 54.0 | 10.5 | ±1.3 |
| 55.0 | 10.4 | ±1.3 |
| 56.0 | 11.1 | ±1.3 |
| 57.0 | 11.4 | ±1.3 |
| 58.0 | 12.2 | ±1.3 |
| 59.0 | 13.0 | ±1.3 |
| 60.0 | 14.6 | ±1.3 |

1.4 Continuity response within 1 GHz

Continuity response within any 1 GHz Band, 4. Harmonic

| | DUL /dB | Continuity /dB |
|-------------------------|---------|----------------|
| max. at IF = 404.4 MHz: | 6.0 | 2.37 |
| max. at IF = 729 MHz: | 6.0 | 2.19 |
| max. at IF = 1330 MHz: | 6.0 | 2.19 |

Incoming Results

Designation: HARMONIC MIXER
Type: FS-Z60
Material No.: 1048.0171.02
Serial No.: 100986
Certificate No.: 0001A300718726
Referring to Test Documentation: 5038.8581.01-PB-02.00

Test Department: 3MM-P
Name: see certificate
Date: 2023-10-31



Table of contents

| | |
|---|----------|
| Software used for measurement | 3 |
| 1. Conversion Loss (4. Harmonic) | 4 |
| 1.1 Conversion Loss (IF = 404.4 MHz) | 4 |
| 1.2 Conversion Loss (IF = 729 MHz) | 5 |
| 1.3 Conversion Loss (IF = 1330 MHz) | 6 |
| 1.4 Continuity response within 1 GHz | 6 |


Incoming Results

| Software used for measurement | | | |
|-------------------------------|-----------|-----------|-----------------------------|
| Item | Type | Version | Remark |
| Suite | Setup | V12.37.04 | Test Management Software G5 |
| Test Program (7012.8706.00_) | Component | V01.05 | |

1. Conversion Loss (4. Harmonic)

1.1 Conversion Loss (IF = 404.4 MHz)

IF = 404.4 MHz, 4. Harmonic




Conversion Loss (IF = 404.4 MHz)

| Frequency /GHz | Conversion Loss /dB | Uncertainty /dB |
|----------------|---------------------|-----------------|
| 40.0 | 10.2 | ±1.3 |
| 41.0 | 10.6 | ±1.3 |
| 42.0 | 12.0 | ±1.3 |
| 43.0 | 13.1 | ±1.3 |
| 44.0 | 13.5 | ±1.3 |
| 45.0 | 14.2 | ±1.3 |
| 46.0 | 15.3 | ±1.3 |
| 47.0 | 14.4 | ±1.3 |
| 48.0 | 15.1 | ±1.3 |
| 49.0 | 14.8 | ±1.3 |
| 50.0 | 12.7 | ±1.3 |
| 51.0 | 11.4 | ±1.3 |
| 52.0 | 11.3 | ±1.3 |
| 53.0 | 10.3 | ±1.3 |
| 54.0 | 10.2 | ±1.3 |
| 55.0 | 10.2 | ±1.3 |
| 56.0 | 10.9 | ±1.3 |
| 57.0 | 11.3 | ±1.3 |
| 58.0 | 11.9 | ±1.3 |
| 59.0 | 12.6 | ±1.3 |
| 60.0 | 14.2 | ±1.3 |

Incoming Results

1.2 Conversion Loss (IF = 729 MHz)

IF = 729 MHz, 4. Harmonic




Conversion Loss (IF = 729 MHz)

| Frequency /GHz | Conversion Loss /dB | Uncertainty /dB |
|----------------|---------------------|-----------------|
| 40.0 | 10.1 | ±1.3 |
| 41.0 | 10.4 | ±1.3 |
| 42.0 | 11.8 | ±1.3 |
| 43.0 | 12.9 | ±1.3 |
| 44.0 | 13.1 | ±1.3 |
| 45.0 | 13.6 | ±1.3 |
| 46.0 | 15.4 | ±1.3 |
| 47.0 | 14.1 | ±1.3 |
| 48.0 | 15.2 | ±1.3 |
| 49.0 | 14.9 | ±1.3 |
| 50.0 | 13.1 | ±1.3 |
| 51.0 | 11.6 | ±1.3 |
| 52.0 | 11.4 | ±1.3 |
| 53.0 | 10.2 | ±1.3 |
| 54.0 | 10.1 | ±1.3 |
| 55.0 | 10.1 | ±1.3 |
| 56.0 | 10.8 | ±1.3 |
| 57.0 | 11.4 | ±1.3 |
| 58.0 | 11.8 | ±1.3 |
| 59.0 | 12.7 | ±1.3 |
| 60.0 | 14.0 | ±1.3 |

Incoming Results

1.3 Conversion Loss (IF = 1330 MHz)

IF = 1330 MHz, 4. Harmonic



Conversion Loss (IF = 1330 MHz)

| Frequency /GHz | Conversion Loss /dB | Uncertainty /dB |
|----------------|---------------------|-----------------|
| 40.0 | 10.5 | ±1.3 |
| 41.0 | 10.6 | ±1.3 |
| 42.0 | 12.1 | ±1.3 |
| 43.0 | 13.3 | ±1.3 |
| 44.0 | 13.2 | ±1.3 |
| 45.0 | 13.5 | ±1.3 |
| 46.0 | 15.0 | ±1.3 |
| 47.0 | 14.2 | ±1.3 |
| 48.0 | 15.1 | ±1.3 |
| 49.0 | 16.2 | ±1.3 |
| 50.0 | 14.0 | ±1.3 |
| 51.0 | 12.9 | ±1.3 |
| 52.0 | 11.8 | ±1.3 |
| 53.0 | 10.6 | ±1.3 |
| 54.0 | 10.6 | ±1.3 |
| 55.0 | 10.5 | ±1.3 |
| 56.0 | 11.2 | ±1.3 |
| 57.0 | 11.4 | ±1.3 |
| 58.0 | 12.3 | ±1.3 |
| 59.0 | 13.0 | ±1.3 |
| 60.0 | 14.7 | ±1.3 |

1.4 Continuity response within 1 GHz

Continuity response within any 1 GHz Band, 4. Harmonic

| | DUL /dB | Continuity /dB |
|-------------------------|---------|----------------|
| max. at IF = 404.4 MHz: | 6.0 | 2.35 |
| max. at IF = 729 MHz: | 6.0 | 2.18 |
| max. at IF = 1330 MHz: | 6.0 | 2.22 |

Incoming Results



CALIBRATION CERTIFICATE



Kalibrierschein

Certificate Number
Zertifikatsnummer

0001A300718729

| General Data | |
|---|--|
| Item Gegenstand | FS-Z90 HARMONIC MIXER 60-90GHZ |
| Manufacturer Hersteller | ROHDE & SCHWARZ |
| Type Typ | FS-Z90 |
| Material Number Materialnummer | 1048.0371.02 |
| Serial Number Seriennummer | 101811 |
| Order Number Bestellnummer | Q786007 |
| Asset Number Inventarnummer | |
| Customer Auftraggeber | Sporton International Inc. 6F., Sec. 1, Hsin Tai Wu Rd 106 221 New Taipei City TW |
| Performance | |
| Place and Date of Calibration Ort und Datum der Kalibrierung | 87700 Memmingen, Rohde-und-Schwarz-Str. 1 2023-10-31 |
| Statement of Compliance (Incoming) Konformitätsaussage (Anlieferung) | All measured values are within the data sheet specifications. |
| Statement of Compliance (Outgoing) Konformitätsaussage (Auslieferung) | All measured values are within the data sheet specifications. |
| Customers due Interval Kalibrierintervall des Kunden | |
| Extent of Calibration Document Umfang des Kalibrierdokuments | 3 Pages Certificate 7 Pages Outgoing Results 7 Pages Incoming Results |
| Date of Issue Ausstellungsdatum | Approval of the certificate by Freigabe des Kalibrierscheins durch |
| 2023-11-08 | Dr. Gerhard Rösel Johannes Negele |
| |   |
| | Laboratory management Labormanagement Person responsible Bearbeiter |

Calibration Mark
Kalibrierzeichen

| |
|---------------------|
| 300718729 |
| D-K- 15195-01-00 |
| 2023-10 |

Member of Deutscher Kalibrierdienst
Mitglied im Deutschen Kalibrierdienst



This calibration certificate documents the metrological traceability to national standards, which realize the units of measurement according to the International System of Units (SI). The DAkkS is signatory to the multilateral agreements of the European co-operation for Accreditation (EA) and of the International Laboratory Accreditation Cooperation (ILAC) for the mutual recognition of calibration certificates. The user is obliged to have the object recalibrated at appropriate intervals. This calibration certificate may not be reproduced other than in full except with the permission of the issuing laboratory. Calibration certificates with the full name of the approval responsible person are valid without signature.

Dieser Kalibrierschein dokumentiert die metrologische Rückführbarkeit auf nationale Normale zur Darstellung der Einheiten in Übereinstimmung mit dem Internationalen Einheitensystem (SI). Die DAkkS ist Unterzeichner der multilateralen Übereinkommen der European cooperation for Accreditation (EA) und der International Laboratory Accreditation Cooperation (ILAC) zur gegenseitigen Anerkennung der Kalibrierscheine. Für die Einhaltung einer angemessenen Frist zur Wiederholung der Kalibrierung ist der Benutzer verantwortlich. Dieser Kalibrierschein darf nur vollständig weiterverbreitet werden. Auszüge oder Änderungen bedürfen der Genehmigung des ausstellenden Kalibrierlaboratoriums. Kalibrierscheine sind bei Nennung des für die Freigabe Verantwortlichen in Klarschrift auch ohne Unterschrift gültig.