



**Material No** 1048.0371.02    **Serial No** 101811    **Certificate** 0001A300718729  
**Page** 2/3    **Number**

**Calibration Procedure**

The measuring object is an RF harmonic mixer, which converts an RF signal at one frequency into a signal at another frequency (here: IF). The conversion loss was measured using a vector network analyzer. The RF output power as well as the IF input power of the corresponding ports of the VNA were traced back to a power sensor. The conversion loss is defined as the ratio of the power at the IF frequency to the power at the RF frequency with a given LO power. (IF: Intermediate frequency; LO: Local Oscillator)  
 The traceability is represented in the table Working Standards used.

**Working Standards used**

Item	Type	Serial Number	Calibration Certificate Number	Cal. Due
Therm.Power Sensor DC-40GHz	NRP-Z55	130179	0001A300682928	2025-01-31
Thermal Waveguide Power Sensor	NRP90TWG	910001	20A1129129	2026-04-30
Vector Network Analyzer 4 Port	ZVA67	101100	0001A300698540	2024-05-31

**Remarks**



**Material No** 1048.0371.02    **Serial No** 101811    **Certificate** 0001A300718729  
**Page** 3/3    **Number**

Environmental Conditions			
<b>Ambient Temperature</b>	(23 ± 1) °C	<b>Relative Humidity</b>	20%-70%

Comments on Measurement Results
<p>The reported results apply only to those items specifically listed on this calibration certificate and have been tested for compliance with the specifications. The associated uncertainty of measurement has been taken into account if not otherwise stated.</p> <p>The non-binary decision rule with guard band is used according to ILAC G8:09/2019 'Guidelines on Decision Rules and Statements of Conformity'. Pass is normally not marked. Conditional Pass is marked with UGB1, Conditional Fail with UGB2 and Fail with Fail.</p> <p>The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k such that the coverage probability corresponds to approximately 95 %. It is consistent with the EA-4/02 M:2022.</p> <p>In addition to the calibration results, the calibration certificate includes functional measurements that might have an influence on the measurement uncertainty of the calibration results.</p> <p>The functional measurement results are marked and are not intended to be used to support the further dissemination of metrological traceability. They are intended to verify the requirements on the measurement object according to manufacturer specifications and technical standards.</p>

# Outgoing Results

**Designation:** HARMONIC MIXER  
**Type:** FS-Z90  
**Material No.:** 1048.0371.02  
**Serial No.:** 101811  
**Certificate No.:** 0001A300718729  
**Referring to Test Documentation:** 5038.8323.01-PB-02.00


**Test Department:** 3MM-P  
**Name:** see certificate  
**Date:** 2023-10-31



**The following abbreviations may be used in this document**

- {a} No measurement uncertainty stated because the errors always add together. So it is sure that a measurement result evaluated as "PASS" is pass.
  - {b} The measurement uncertainty depends on the measurement result. The stated measurement uncertainty is valid for the close area around the specification. Measurement results outside the close area have a higher measurement uncertainty but are within the specification.
  - {c} Functional test, therefore no measurement uncertainty is stated.
  - {d} Typical value, refer to performance test.
  - {e} The measurement uncertainty is taken into account when setting the measuring system.
  - {g} Verification of specified requirements, non-accredited measurements. Technical operations that consist of the determination of one or more characteristics to a specified procedure (formerly {f}).
- DL or DT Data Limit for symmetrical tolerance limits
  - DLL Datasheet Lower Limit
  - DUL Datasheet Upper Limit
  - MU Symmetrical Measurement Uncertainty
  - MLL or MLV Measurement Uncertainty Lower Value
  - MUL or MUV Measurement Uncertainty Upper Value
  - Nom. Nominal Value
  - Dev. Deviation
  - Act. Actual Value
  - UGB Uncertainty Guard Band: Measuring uncertainty violates the data (spec.) limit.
  - UGB1 A compliance statement may be possible where a confidence level of less than 95 % is acceptable.
  - UGB2 A non-compliance statement may be possible where a confidence level of less than 95 % is acceptable.
  - DU Datasheet Uncertainty

**Explanation of charts**



## Table of contents


Software used for measurement .....	4
<b>1. Conversion Loss (6. Harmonic) .....</b>	<b>5</b>
1.1 Conversion Loss (IF = 404.4 MHz) .....	5
1.2 Conversion Loss (IF = 729 MHz) .....	6
1.3 Conversion Loss (IF = 1330 MHz) .....	7
1.4 Continuity response within 1 GHz .....	7

<b>Software used for measurement</b>			
<b>Item</b>	<b>Type</b>	<b>Version</b>	<b>Remark</b>
Suite	Setup	V12.37.04	Test Management Software G5
Test Program (7012.8706.00_)	Component	V01.05	

## 1. Conversion Loss (6. Harmonic)

### 1.1 Conversion Loss (IF = 404.4 MHz)

IF = 404.4 MHz, 6. Harmonic




Conversion Loss (IF = 404.4 MHz)

Frequency /GHz	Conversion Loss /dB	Uncertainty /dB
60.0	24.9	±1.3
61.0	21.5	±1.3
62.0	20.4	±1.3
63.0	19.4	±1.3
64.0	16.9	±1.3
65.0	18.3	±1.3
66.0	19.9	±1.3
67.0	18.9	±1.3
68.0	18.0	±1.3
69.0	19.1	±1.3
70.0	19.3	±1.3
71.0	18.7	±1.3
72.0	18.0	±1.3
73.0	18.1	±1.3
74.0	17.6	±1.3
75.0	17.4	±1.3
76.0	16.9	±1.3
77.0	15.5	±1.3
78.0	14.7	±1.3
79.0	14.5	±1.3
80.0	14.1	±1.3
81.0	14.0	±1.3
82.0	14.1	±1.3
83.0	14.2	±1.3
84.0	14.6	±1.3
85.0	15.5	±1.3
86.0	16.5	±1.3
87.0	17.9	±1.3
88.0	19.8	±1.3
89.0	21.0	±1.3
90.0	21.1	±1.3

**1.2 Conversion Loss (IF = 729 MHz)**

IF = 729 MHz, 6. Harmonic




Conversion Loss (IF = 729 MHz)

Frequency /GHz	Conversion Loss /dB	Uncertainty /dB
60.0	26.2	±1.3
61.0	22.4	±1.3
62.0	21.1	±1.3
63.0	20.1	±1.3
64.0	17.4	±1.3
65.0	18.5	±1.3
66.0	20.3	±1.3
67.0	19.0	±1.3
68.0	18.0	±1.3
69.0	19.0	±1.3
70.0	19.2	±1.3
71.0	18.6	±1.3
72.0	18.1	±1.3
73.0	18.1	±1.3
74.0	17.3	±1.3
75.0	17.2	±1.3
76.0	17.0	±1.3
77.0	15.9	±1.3
78.0	14.8	±1.3
79.0	14.5	±1.3
80.0	14.3	±1.3
81.0	14.0	±1.3
82.0	14.3	±1.3
83.0	14.3	±1.3
84.0	14.5	±1.3
85.0	15.5	±1.3
86.0	16.5	±1.3
87.0	17.7	±1.3
88.0	20.0	±1.3
89.0	21.0	±1.3
90.0	21.0	±1.3



**1.3 Conversion Loss (IF = 1330 MHz)**

IF = 1330 MHz, 6. Harmonic



Conversion Loss (IF = 1330 MHz)

Frequency /GHz	Conversion Loss /dB	Uncertainty /dB
60.0	25.7	±1.3
61.0	25.0	±1.3
62.0	24.0	±1.3
63.0	22.8	±1.3
64.0	19.4	±1.3
65.0	19.5	±1.3
66.0	20.9	±1.3
67.0	19.8	±1.3
68.0	18.4	±1.3
69.0	19.3	±1.3
70.0	19.7	±1.3
71.0	19.0	±1.3
72.0	18.5	±1.3
73.0	18.5	±1.3
74.0	17.5	±1.3
75.0	17.5	±1.3
76.0	17.5	±1.3
77.0	17.0	±1.3
78.0	15.6	±1.3
79.0	15.1	±1.3
80.0	15.1	±1.3
81.0	14.7	±1.3
82.0	15.0	±1.3
83.0	15.1	±1.3
84.0	14.7	±1.3
85.0	15.9	±1.3
86.0	17.0	±1.3
87.0	18.3	±1.3
88.0	20.5	±1.3
89.0	21.5	±1.3
90.0	21.4	±1.3

**1.4 Continuity response within 1 GHz**

Continuity response within any 1 GHz Band, 6. Harmonic

	DUL /dB	Continuity /dB
max. at IF = 404.4 MHz:	6.0	3.42
max. at IF = 729 MHz:	6.0	3.82
max. at IF = 1330 MHz:	6.0	3.35

# Incoming Results

**Designation:** HARMONIC MIXER  
**Type:** FS-Z90  
**Material No.:** 1048.0371.02  
**Serial No.:** 101811  
**Certificate No.:** 0001A300718729  
**Referring to Test Documentation:** 5038.8323.01-PB-02.00


**Test Department:** 3MM-P  
**Name:** see certificate  
**Date:** 2023-10-31



**The following abbreviations may be used in this document**

- {a} No measurement uncertainty stated because the errors always add together. So it is sure that a measurement result evaluated as "PASS" is pass.
  - {b} The measurement uncertainty depends on the measurement result. The stated measurement uncertainty is valid for the close area around the specification. Measurement results outside the close area have a higher measurement uncertainty but are within the specification.
  - {c} Functional test, therefore no measurement uncertainty is stated.
  - {d} Typical value, refer to performance test.
  - {e} The measurement uncertainty is taken into account when setting the measuring system.
  - {g} Verification of specified requirements, non-accredited measurements. Technical operations that consist of the determination of one or more characteristics to a specified procedure (formerly {f}).
- DL or DT Data Limit for symmetrical tolerance limits
  - DLL Datasheet Lower Limit
  - DUL Datasheet Upper Limit
  - MU Symmetrical Measurement Uncertainty
  - MLL or MLV Measurement Uncertainty Lower Value
  - MUL or MUV Measurement Uncertainty Upper Value
  - Nom. Nominal Value
  - Dev. Deviation
  - Act. Actual Value
  - UGB Uncertainty Guard Band: Measuring uncertainty violates the data (spec.) limit.
  - UGB1 A compliance statement may be possible where a confidence level of less than 95 % is acceptable.
  - UGB2 A non-compliance statement may be possible where a confidence level of less than 95 % is acceptable.
  - DU Datasheet Uncertainty

**Explanation of charts**



Incoming Results

## Table of contents

Software used for measurement .....	4
1. Conversion Loss (6. Harmonic) .....	5
1.1 Conversion Loss (IF = 404.4 MHz) .....	5
1.2 Conversion Loss (IF = 729 MHz) .....	6
1.3 Conversion Loss (IF = 1330 MHz) .....	7
1.4 Continuity response within 1 GHz .....	7


**Incoming Results**

Software used for measurement			
Item	Type	Version	Remark
Suite	Setup	V12.37.04	Test Management Software G5
Test Program (7012.8706.00_)	Component	V01.05	

**1. Conversion Loss (6. Harmonic)**

**1.1 Conversion Loss (IF = 404.4 MHz)**

IF = 404.4 MHz, 6. Harmonic




Conversion Loss (IF = 404.4 MHz)

Frequency /GHz	Conversion Loss /dB	Uncertainty /dB
60.0	24.9	±1.3
61.0	21.5	±1.3
62.0	20.4	±1.3
63.0	19.3	±1.3
64.0	16.9	±1.3
65.0	18.2	±1.3
66.0	19.9	±1.3
67.0	18.8	±1.3
68.0	18.0	±1.3
69.0	19.0	±1.3
70.0	19.2	±1.3
71.0	18.7	±1.3
72.0	18.0	±1.3
73.0	18.1	±1.3
74.0	17.5	±1.3
75.0	17.3	±1.3
76.0	16.9	±1.3
77.0	15.4	±1.3
78.0	14.7	±1.3
79.0	14.4	±1.3
80.0	14.1	±1.3
81.0	14.0	±1.3
82.0	14.0	±1.3
83.0	14.2	±1.3
84.0	14.6	±1.3
85.0	15.5	±1.3
86.0	16.4	±1.3
87.0	17.8	±1.3
88.0	19.7	±1.3
89.0	20.9	±1.3
90.0	21.1	±1.3

Incoming Results

**1.2 Conversion Loss (IF = 729 MHz)**

IF = 729 MHz, 6. Harmonic




Conversion Loss (IF = 729 MHz)

Frequency /GHz	Conversion Loss /dB	Uncertainty /dB
60.0	26.2	±1.3
61.0	22.5	±1.3
62.0	21.1	±1.3
63.0	20.1	±1.3
64.0	17.4	±1.3
65.0	18.4	±1.3
66.0	20.3	±1.3
67.0	19.0	±1.3
68.0	18.1	±1.3
69.0	19.1	±1.3
70.0	19.3	±1.3
71.0	18.7	±1.3
72.0	18.2	±1.3
73.0	18.2	±1.3
74.0	17.3	±1.3
75.0	17.3	±1.3
76.0	17.0	±1.3
77.0	16.0	±1.3
78.0	14.8	±1.3
79.0	14.5	±1.3
80.0	14.4	±1.3
81.0	14.0	±1.3
82.0	14.3	±1.3
83.0	14.3	±1.3
84.0	14.5	±1.3
85.0	15.6	±1.3
86.0	16.5	±1.3
87.0	17.7	±1.3
88.0	20.0	±1.3
89.0	21.0	±1.3
90.0	21.1	±1.3

Incoming Results

**1.3 Conversion Loss (IF = 1330 MHz)**

IF = 1330 MHz, 6. Harmonic



Conversion Loss (IF = 1330 MHz)

Frequency /GHz	Conversion Loss /dB	Uncertainty /dB
60.0	26.2	±1.3
61.0	25.5	±1.3
62.0	24.5	±1.3
63.0	23.2	±1.3
64.0	19.9	±1.3
65.0	19.9	±1.3
66.0	21.3	±1.3
67.0	20.3	±1.3
68.0	18.8	±1.3
69.0	19.7	±1.3
70.0	20.1	±1.3
71.0	19.5	±1.3
72.0	18.9	±1.3
73.0	19.0	±1.3
74.0	17.9	±1.3
75.0	17.9	±1.3
76.0	18.0	±1.3
77.0	17.5	±1.3
78.0	16.0	±1.3
79.0	15.5	±1.3
80.0	15.6	±1.3
81.0	15.1	±1.3
82.0	15.4	±1.3
83.0	15.5	±1.3
84.0	15.2	±1.3
85.0	16.3	±1.3
86.0	17.4	±1.3
87.0	18.8	±1.3
88.0	20.9	±1.3
89.0	22.0	±1.3
90.0	21.8	±1.3

**1.4 Continuity response within 1 GHz**

Continuity response within any 1 GHz Band, 6. Harmonic

	DUL /dB	Continuity /dB
max. at IF = 404.4 MHz:	6.0	3.46
max. at IF = 729 MHz:	6.0	3.75
max. at IF = 1330 MHz:	6.0	3.34

Incoming Results





# Calibration Certificate

Certificate Number **24-0140-101128-03**

Kalibrierschein

Zertifikatsnummer

## Unit Data

Item  
Gegenstand **Harmonic Mixer, 90 GHz to 140 GHz**

Manufacturer  
Hersteller **RPG Radiometer physics GmbH**

Type  
Typ **RPG FS-Z140**

Material Number  
Materialnummer **3622.0708.02**      Serial Number  
Seriennummer **101128**

Asset Number  
Inventarnummer

This calibration certificate documents, that the named item is tested and measured against defined specifications. Measurement results are located usually in the corresponding interval with a probability of approx. 95% (coverage factor  $k = 2$ ). Calibration is performed with test equipment and standards directly or indirectly traceable by means of approved calibration techniques to the PTB/DKD or other national/international standards, which realize the physical units of measurement according to the International System of Units (SI). In all cases where no standards are available, measurements are referenced to standards of the R&S laboratories. Principles and methods of calibration correspond with EN ISO/IEC 17025. This calibration certificate may not be reproduced other than in full. Calibration certificates without signatures are not valid. The user is obliged to have the object recalibrated at appropriate intervals.

## Order Data

Customer  
Auftraggeber **Sporton International Inc.**  
**6F., Sec. 1, Hsin Tai Wu Rd 106**  
**Xizhi Dist.**  
**221 New Taipei City**  
**Taiwan**

Order Number  
Bestellnummer **8800047909**

Date of Receipt  
Eingangdatum **2023-10-27**

Dieser Kalibrierschein dokumentiert, dass der genannte Gegenstand nach festgelegten Vorgaben geprüft und gemessen wurde. Die Messwerte lagen im Regelfall mit einer Wahrscheinlichkeit von annähernd 95% im zugeordneten Wertintervall (Erweiterte Messunsicherheit mit  $k = 2$ ). Die Kalibrierung erfolgte mit Messmitteln und Normalen, die direkt oder indirekt durch Ableitung mittels anerkannter Kalibriertechniken rückgeführt sind auf Normale der PTB/DKD oder anderer nationaler/internationaler Standards zur Darstellung der physikalischen Einheiten in Übereinstimmung mit dem internationalen Einheitensystem (SI). Wenn keine Normale existieren, erfolgt die Rückführung auf Bezugsnormale der R&S-Laboratorien. Grundsätze und Verfahren der Kalibrierung beziehen sich auf EN ISO/IEC 17025. Dieser Kalibrierschein darf nur vollständig und unverändert weiterverbreitet werden. Kalibrierscheine ohne Unterschriften sind ungültig. Für die Einhaltung einer angemessenen Frist zur Wiederholung der Kalibrierung ist der Benutzer verantwortlich.

## Performance

Place and Date of Calibration  
Ort und Datum der Kalibrierung **Meckenheim, 2023-11-03**

Scope of Calibration  
Umfang der Kalibrierung **Standard Calibration**

Statement of Compliance  
(Incoming)  
Konformitätsaussage  
(Anlieferung) **All measured values are within the data sheet specifications.**

Statement of Compliance  
(Outgoing)  
Konformitätsaussage  
(Auslieferung) **All measured values are within the data sheet specifications.**

Extend of Calibration Documents  
Umfang des Kalibrierdokuments **2 pages Calibration Certificate**  
**4 pages Outgoing Results**  
**3 pages Incoming Results**

**RPG Radiometer physics GmbH; Meckenheim**


Date of Issue  
Ausstellungsdatum

**2023-11-06**

Head of Laboratory  
Laborleitung

  
**Emons**

Person Responsible  
Bearbeiter

  
**Gottbehüt**

Page (Seite) 1/2  
Vers2010-05-05/  
RPG2014-02-28

Calibration Method  
Kalibrieranweisung

RPG-PAQA-TN-2014-002

Relative Humidity 20 % - 80 %  
Relative Luftfeuchte

Ambient Temperature  
Umgebungstemperatur

(23 <sup>+7</sup>/<sub>.3</sub>) °C

Working standards used (having a significant effect on the accuracy) Verwendete Gebrauchsnormale (mit signifikantem Einfluss auf die Genauigkeit)				
Item Gegenstand	Type Typ	Serial Number Seriennummer	Calibration Certificate Number Kalibrierscheinnummer	Cal. Due Kalibr. bis
Vector Network Analyzer	R&S® ZVA40	100103	0001A300715526	2026-10-04
Powersensor	R&S® NRP-Z55	140093	0001-300700427	2025-06-06

**UGB1** A compliance statement may be possible where a confidence level of less than 95 % is acceptable.  
Die Bestätigung der Konformität ist möglich, sofern ein Grad des Vertrauens von weniger als 95 % akzeptabel ist.

**UGB2** A non-compliance statement may be possible where a confidence level of less than 95 % is acceptable.  
Die Bestätigung der Nicht-Konformität ist möglich, sofern ein Grad des Vertrauens von weniger als 95 % akzeptabel ist.

Ref.: ILAC-G8:03/2009 'Guidelines on the Reporting of Compliance with Specification'.

**Notes**  
Anmerkungen


If the new product is stored under the climate conditions as specified in the data sheet upon delivery, the product's accuracy is not significantly affected within 12 month after its calibration in our factory. In this case, the recommended calibration interval starts on the date when the product is actually put into operation.

## Outgoing Results

The following abbreviations may be used in this document

{a}	No measurement uncertainty stated because the errors always add together. So it is sure that a measurement result evaluated as "PASS" is pass.
{b}	The measurement uncertainty depends on the measurement result. The stated measurement uncertainty is valid for the close area around the specification. Measurement results outside the close area have a higher measurement uncertainty but are within the specification.
{c}	Functional test, therefore no measurement uncertainty is stated.
{d}	Typical value, refer to performance test.
{e}	The measurement uncertainty is taken into account when setting the measuring system.
DL or DT	Data Limit for symmetrical tolerance limits
DLL	Datasheet Lower Limit
DUL	Datasheet Upper Limit
MU	Measurement Uncertainty
MLL or MLV	Measurement Uncertainty Lower Value
MUL or MUV	Measurement Uncertainty Upper Value
Nom.	Nominal Value
Dev.	Deviation
MErr.	Measurement Error
Act.	Actual Value
UGB	Uncertainty Guard Band: Measuring uncertainty violates the data (spec.) limit.
UGB1	Measurement results marked as UGB1 show conformity with a probability of >50 % and <95 %.
UGB2	Measurement results marked as UGB2 show non-conformity with a probability of >50 % and <95 %.
DU	Datasheet Uncertainty

### Explanation of charts



**Software used for measurement**

**Item Type**

Measurement Studio Professional Edition  
MixerCertification

**Version**


2013  
7\_15

**Remark**

## 1.1 Conversion loss

LO level                    +14 dBm nominal  
Bias                        0 A

Measurement uncertainty:    3.5 dB



**Note:** Numeric calibration data can be found attached to the PDF file of the calibration certificate. Click the “paper clip” symbol to display the file.

The file has been renamed for safety reasons.  
When downloading the file onto your PC, please delete the “.file” extension and unzip the data.

## 1.2 Frequency response within 1 GHz


	DUL	Actual (worst case)	Evaluation
IF = 404.4 MHz, 10th harmonic	6 dB	2.49 dB	PASS
IF = 729 MHz, 10th harmonic	6 dB	2.09 dB	PASS
IF = 1330 MHz, 10th harmonic	6 dB	2.24 dB	PASS

# 1 Incoming Results

The following abbreviations may be used in this document

- {a) No measurement uncertainty stated because the errors always add together.  
So it is sure that a measurement result evaluated as "PASS" is pass.
- {b) The measurement uncertainty depends on the measurement result. The stated measurement uncertainty is valid for the close area around the specification. Measurement results outside the close area have a higher measurement uncertainty but are within the specification.
- {c) Functional test, therefore no measurement uncertainty is stated.
- {d) Typical value, refer to performance test.
- {e) The measurement uncertainty is taken into account when setting the measuring system.
- DL or DT Data Limit for symmetrical tolerance limits
- DLL Datasheet Lower Limit
- DUL Datasheet Upper Limit
- MU Measurement Uncertainty
- MLL or MLV Measurement Uncertainty Lower Value
- MUL or MUV Measurement Uncertainty Upper Value
- Nom. Nominal Value
- Dev. Deviation
- MErr. Measurement Error
- Act. Actual Value
- UGB Uncertainty Guard Band: Measuring uncertainty violates the data (spec.) limit.
- UGB1 Measurement results marked as UGB1 show conformity with a probability of >50 %and <95 %.
- UGB2 Measurement results marked as UGB2 show non-conformity with a probability of >50 %and <95 %.
- DU Datasheet Uncertainty

## Explanation of charts



Material Number 3622.0708.02

Serial Number 101128

Certificate Number 24-0140-101128-03

(Incoming)

---

**Software used for measurement**

**Item Type**

Measurement Studio Professional Edition  
MixerCertification

**Version**


2013  
7\_15

**Remark**

Incoming Report



### 1.1 Deviation between actual and previous conversion loss



Incoming Report



Calibration Method  
Kalibrieranweisung

RPG-PAQA-TN-2014-002

Relative Humidity 20 % - 80 %  
Relative Luftfeuchte

Ambient Temperature  
Umgebungstemperatur

(23 <sup>+7</sup> / <sub>-3</sub>) °C

Working standards used (having a significant effect on the accuracy) Verwendete Gebrauchsnormale (mit signifikantem Einfluss auf die Genauigkeit)				
Item Gegenstand	Type Typ	Serial Number Seriennummer	Calibration Certificate Number Kalibrierscheinnummer	Cal. Due Kalibr. bis
Vector Network Analyzer	R&S® ZVA40	100103	0001A300715526	2026-10-04
Powersensor	R&S® NRP-Z55	140093	0001-300700427	2025-06-06

<b>UGB1</b>	<b>A compliance statement may be possible where a confidence level of less than 95 % is acceptable.</b> Die Bestätigung der Konformität ist möglich, sofern ein Grad des Vertrauens von weniger als 95 % akzeptabel ist.
<b>UGB2</b>	<b>A non-compliance statement may be possible where a confidence level of less than 95 % is acceptable.</b> Die Bestätigung der Nicht-Konformität ist möglich, sofern ein Grad des Vertrauens von weniger als 95 % akzeptabel ist.
Ref.: ILAC-G8:03/2009 'Guidelines on the Reporting of Compliance with Specification'.	

**Notes**  
Anmerkungen


If the new product is stored under the climate conditions as specified in the data sheet upon delivery, the product's accuracy is not significantly affected within 12 month after its calibration in our factory. In this case, the recommended calibration interval starts on the date when the product is actually put into operation.

## Outgoing Results

The following abbreviations may be used in this document

{a}	No measurement uncertainty stated because the errors always add together. So it is sure that a measurement result evaluated as "PASS" is pass.
{b}	The measurement uncertainty depends on the measurement result. The stated measurement uncertainty is valid for the close area around the specification. Measurement results outside the close area have a higher measurement uncertainty but are within the specification.
{c}	Functional test, therefore no measurement uncertainty is stated.
{d}	Typical value, refer to performance test.
{e}	The measurement uncertainty is taken into account when setting the measuring system.
DL or DT	Data Limit for symmetrical tolerance limits
DLL	Datasheet Lower Limit
DUL	Datasheet Upper Limit
MU	Measurement Uncertainty
MLL or MLV	Measurement Uncertainty Lower Value
MUL or MUJ	Measurement Uncertainty Upper Value
Nom.	Nominal Value
Dev.	Deviation
MErr.	Measurement Error
Act.	Actual Value
UGB	Uncertainty Guard Band: Measuring uncertainty violates the data (spec.) limit.
UGB1	Measurement results marked as UGB1 show conformity with a probability of >50 % and <95 %.
UGB2	Measurement results marked as UGB2 show non-conformity with a probability of >50 % and <95 %.
DU	Datasheet Uncertainty

### Explanation of charts




**Software used for measurement**

<b>Item Type</b>	<b>Version</b>	<b>Remark</b>
Measurement Studio Professional Edition	2013	
MixerCertification	7_15	

## 1.1 Conversion loss

LO level                    +13 dBm nominal  
Bias                        0 A

Measurement uncertainty:    4 dB



**Note:** Numeric calibration data can be found attached to the PDF file of the calibration certificate. Click the “paper clip” symbol to display the file.

The file has been renamed for safety reasons.  
When downloading the file onto your PC, please delete the “.file” extension and unzip the data.

## 1.2 Frequency response within 1 GHz


	DUL	Actual (worst case)	Evaluation
IF = 404.4 MHz, 16th harmonic	6 dB	2.61 dB	PASS
IF = 729 MHz, 16th harmonic	6 dB	2.54 dB	PASS
IF = 1330 MHz, 14th harmonic	6 dB	2.39 dB	PASS

# 1 Incoming Results

## The following abbreviations may be used in this document

{a}	No measurement uncertainty stated because the errors always add together. So it is sure that a measurement result evaluated as "PASS" is pass.
{b}	The measurement uncertainty depends on the measurement result. The stated measurement uncertainty is valid for the close area around the specification. Measurement results outside the close area have a higher measurement uncertainty but are within the specification.
{c}	Functional test, therefore no measurement uncertainty is stated.
{d}	Typical value, refer to performance test.
{e}	The measurement uncertainty is taken into account when setting the measuring system.
DL or DT	Data Limit for symmetrical tolerance limits
DLL	Datasheet Lower Limit
DUL	Datasheet Upper Limit
MU	Measurement Uncertainty
MLL or MLV	Measurement Uncertainty Lower Value
MUL or MUV	Measurement Uncertainty Upper Value
Nom.	Nominal Value
Dev.	Deviation
MErr.	Measurement Error
Act.	Actual Value
UGB	Uncertainty Guard Band: Measuring uncertainty violates the data (spec.) limit.
UGB1	Measurement results marked as UGB1 show conformity with a probability of >50 %and <95 %.
UGB2	Measurement results marked as UGB2 show non-conformity with a probability of >50 %and <95 %.
DU	Datasheet Uncertainty

## Explanation of charts





Material Number 3593.3250.02

Serial Number 101014

Certificate Number 24-0220-101014-03

(Incoming)

---

**Software used for measurement**

**Item Type**

Measurement Studio Professional Edition

MixerCertification

**Version**


2013

7\_15

**Remark**

Incoming Report

### 1.1 Deviation between actual and previous conversion loss



Incoming Report

Calibration Laboratory of Microwave Measuring Equipment  
of MWMLab



Calibration certificate

ISO 17025  
ACCREDITED LABORATORY



Accreditation certificate No. № BY/112 5.0065 of 09.01.2015

Certificate number 36-21 Date when calibrated 06.07.2021 Page 1 of 2

Item calibrated Antenna QWH-UPRR00 # 1410300003

Customer Sporton International Inc.

Method of calibration GOST 20271.1, MK KL 8.2-16

*All measurements are traceable to the SI units which are realized by national measurement standards of NMI and state standards of RF. Gain measurements above 178 GHz are to confirm operation functionality and traceable only to MWMLab standards and OML. This certificate shall not be reproduced, except in full. Any publication extracts from the calibration certificate requires written permission of the issuing calibration laboratory of microwave measuring equipment.*

Authorising  
signature



/ Technical manager Date of issue 06.07.2021

# Calibration Certificate

Certificate number **36-21**

Page 2 of 2

## Calibration is performed by using

Model	Model Description	Equipment ID	Cal Due Date	Certificate Number	Trace Value
M1-11	Calibrator of power with wattmeter M3-22A	841202/ 037410	08 December 2021	3882-43	RF Power
M 568	Reference power meter	164	24 March 2022	1/111-175-20	RF Power
G4-161	Signal generator	3	12 October 2021	20-20	RF Power
MG3694C	Signal generator	133805	11 September 2021	2726-43	RF Power Frequency
V7-34	Universal voltmeter	0067787	23 September 2021	2742-42	DC Voltage
RCH3-72	Frequency meter	931200	18 September 2021	2822-43	Frequency
P6-133	Horn antenna	15005	23 September 2021	2374-43	Gain
P6-11B	Measuring horn antenna	08051	23 September 2021	2370-43	Gain

## Calibration conditions

Temperature: 23.8 °C.  
Humidity: 43.2 %.  
Pressure: 100.1 kPa.

## Calibration results are given in the measurement report # 36-21

#	Parameter	Specifications required	Specifications tested and measured
1	Frequency range	40 – 60 GHz	Corresponds
2	Antenna Gain	22.5* dBi	Corresponds (Table 1)
3	Antenna Factor	42 dB/m	Corresponds (Table 1)

\* – Expanded uncertainty of measurements 2.0 dB.

*The uncertainty evaluation has been performed in accordance with ISO/IEC Guide 98-3:2008 (GUM). The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor  $k$  such that the coverage probability corresponds to approximately 95 %. This probability corresponds to a coverage factor of  $k=2$  for a normal distribution.*

**Signature of the person who has performed calibration**

  
/ Engineer

**Calibration Laboratory of  
Microwave Measuring Equipment**

Accreditation certificate

No. BY/112 5.0065

Address: 6, P. Brovki str., Minsk  
220013, Belarus

Phone/Fax: +375 17 2938496



Technical Manager

July 6, 2021

**MEASUREMENT REPORT # 36-21**

July 6, 2021

Customer:	Sporton International Inc.
Item calibrated:	<b>Antenna QWH-UPRR00 # 1410300003</b>
Method of calibration:	GOST 20271.1, MK KL 8.2-16
Number of samples:	One
Delivery date of the sample:	21.06.2021
Date of calibration:	From 21.06.2021 to 06.07.2021

## MEASUREMENT CONDITIONS

Temperature: 23.8 °C	Humidity: 43.2 %	Pressure: 100.1 kPa
----------------------	------------------	---------------------

## MEASUREMENT EQUIPMENT

Model	Model Description	Equipment ID	Cal Due Date	Certificate Number	Trace Value
M1-11	Calibrator of power with wattmeter M3-22A	841202/ 037410	08 December 2021	3882-43	RF Power
M 568	Reference power meter	164	24 March 2022	1/111-175-20	RF Power
G4-161	Signal generator	3	12 October 2021	20-20	RF Power
MG3694C	Signal generator	133805	11 September 2021	2726-43	RF Power Frequency
V7-34	Universal voltmeter	0067787	23 September 2021	2742-42	DC Voltage
RCH3-72	Frequency meter	931200	18 September 2021	2822-43	Frequency
P6-133	Horn antenna	15005	23 September 2021	2374-43	Gain
P6-11B	Measuring horn antenna	08051	23 September 2021	2370-43	Gain

## MEASUREMENT RESULTS

Distance between tested and generating antenna 2.0 m.

Table 1

Frequency, GHz	40	50	60
Power density of electromagnetic field, W/m <sup>2</sup>	0.050	0.070	0.070
Maximum level of measured power, dBm	-14.3	-14.2	-15.8
Gain, dBi	22.3	22.8	22.8
Expanded uncertainty, dB	2.0	2.0	2.0
Antenna Factor, dB/m	40.0	41.5	43.0

*The uncertainty evaluation has been performed in accordance with ISO/IEC Guide 98-3:2008 (GUM). The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor  $k$  such that the coverage probability corresponds to approximately 95 %. This probability corresponds to a coverage factor of  $k=2$  for a normal distribution.*

Engineer




This measurement report issued in duplicate and sent to:

1. Sporton International Inc.
2. Calibration Laboratory of Microwave Measuring Equipment

Duplication of Measurement report (complete or partial) must be authorized by the laboratory.

**Calibration Laboratory of Microwave Measuring Equipment  
of MWMLab**



**Calibration certificate**

**ISO 17025**  
ACCREDITED LABORATORY



**Accreditation certificate No. № BY/112 5.0065 of 09.01.2015**

**Certificate number 38-21 Date when calibrated 06.07.2021 Page 1 of 2**

**Item calibrated** Antenna QWH-EPRR00 # 1372000000

**Customer** Sporton International Inc.

**Method of calibration** GOST 20271.1, MK KL 8.2-16

*All measurements are traceable to the SI units which are realized by national measurement standards of NMI and state standards of RF. Gain measurements above 178 GHz are to confirm operation functionality and traceable only to MWMLab standards and OML. This certificate shall not be reproduced, except in full. Any publication extracts from the calibration certificate requires written permission of the issuing calibration laboratory of microwave measuring equipment.*

**Authorising signature**



**/ Technical manager Date of issue 06.07.2021**

# Calibration Certificate

Certificate number **38-21**

Page 2 of 2

## Calibration is performed by using

Model	Model Description	Equipment ID	Cal Due Date	Certificate Number	Trace Value
M 546	Reference power meter	163	24 March 2022	1/111-174-20	RF Power
M 534	Reference power meter	161	24 March 2022	1/111-173-20	RF Power
RG4-14	Signal generator	22	12 October 2021	22-20	RF Power
G4-186	Signal generator	5	12 October 2021	21-20	RF Power
V7-34	Universal voltmeter	0067787	23 September 2021	2742-42	DC Voltage
RCH3-72	Frequency meter	931200	18 September 2021	2822-43	Frequency
P6-31A	Measuring horn antenna	35864	23 September 2021	2368-43	Gain
P6-134	Measuring horn antenna	14002	23 September 2021	2372-43	Gain

## Calibration conditions

Temperature: 23.8 °C.

Humidity: 43.2 %.

Pressure: 100.1 kPa.

## Calibration results are given in the measurement report # 38-21

#	Parameter	Specifications required	Specifications tested and measured
1	Frequency range	60 – 90 GHz	Corresponds
2	Antenna Gain	22.5* dBi	Corresponds (Table 1)
3	Antenna Factor	45.5 dB/m	Corresponds (Table 1)

\* – Expanded uncertainty of measurements 2.0 dB.

*The uncertainty evaluation has been performed in accordance with ISO/IEC Guide 98-3:2008 (GUM). The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor  $k$  such that the coverage probability corresponds to approximately 95 %. This probability corresponds to a coverage factor of  $k=2$  for a normal distribution.*

Signature of the person who has performed calibration



/ Engineer



**Calibration Laboratory of  
Microwave Measuring Equipment**

Accreditation certificate

No. BY/112 5.0065

Address: 6, P. Brovki str., Minsk  
220013, Belarus

Phone/Fax: +375 17 2938496



Technical Manager

July 6, 2021

**MEASUREMENT REPORT # 38-21**

July 6, 2021

Customer:	Sporton International Inc.
Item calibrated:	<b>Antenna QWH-EPRR00 # 1372000000</b>
Method of calibration:	GOST 20271.1, MK KL 8.2-16
Number of samples:	One
Delivery date of the sample:	21.06.2021
Date of calibration:	From 21.06.2021 to 06.07.2021