Annex D. Tissue & System Verification

The measuring results for tissue simulating liquid and system check are shown as below.

Note:

- 1. For Section 4.3, the dielectric properties of the tissue simulating liquid have been measured within 24 hours before the SAR testing and within ±10 % of the target values. Liquid temperature during the SAR testing has kept within ±2 ℃.
- 2. For Section 4.4, The SAR measurement system was validated according to procedures in KDB 865664 D01. The validation status in tabulated summary is as below.
- 3. For Section 4.5, Comparing to the reference SAR value provided by SPEAG in dipole calibration certificate, the deviation of system check results is within its specification of 10 %. The result indicates the system check can meet the variation criterion and the plots please refer to Annex A of this report.

Tissue Verification									Validation for CW			Validation for Modulation				System Validation					Note		
Plot No.	Frequency (MHz)	Liquid Temp. (°C)	Conductivity (ơ)	Permittivity (εr)	Targeted Conductivity (σ)	Targeted Permittivity (εr)	Deviation Conductivity (σ)	Deviation Permittivity (ɛr)	Sensitivity Range	Probe Linearity	Probe Isotropy	Modulation Type	Duty Factor	PAR	Date	Frequency (MHz)	Targeted 1g SAR (W/kg)	Measured 1g SAR (W/kg)	Normalized 1g SAR (W/kg)	Deviation (%)	Dipole S/N	Probe S/N	DAE S/N
S01	900	23.4	0.959	40.127	0.97	41.5	-1.13	-3.31	Pass	Pass	Pass	N/A	N/A	N/A	Sep. 11, 2021	900	10.90	0.494	9.88	-9.36	1d121	3820	1305
S02	900	23.4	0.959	40.127	0.97	41.5	-1.13	-3.31	Pass	Pass	Pass	N/A	N/A	N/A	Sep. 11, 2021	900	10.90	0.494	9.88	-9.36	1d121	3820	1305
S03	900	23.4	0.959	40.127	0.97	41.5	-1.13	-3.31	Pass	Pass	Pass	N/A	N/A	N/A	Sep. 11, 2021	900	10.90	0.494	9.88	-9.36	1d121	3820	1305
S04	900	23.4	0.959	40.127	0.97	41.5	-1.13	-3.31	Pass	Pass	Pass	N/A	N/A	N/A	Sep. 11, 2021	900	10.90	0.494	9.88	-9.36	1d121	3820	1305

Tissue Verification										Validation for CW			Validation for Modulation			System Validation				Note			
Plot No.	Frequency (MHz)	Liquid Temp. (°C)	Conductivity (σ)	Permittivity (ɛr)	Targeted Conductivity (σ)	Targeted Permittivity (εr)	Deviation Conductivity (σ)	Deviation Permittivity (ɛr)	Sensitivity Range	Probe Linearity	Probe Isotropy	Modulation Type	Duty Factor	PAR	Date	Frequency (MHz)	Targeted 10g SAR (W/kg)	Measured 10g SAR (W/kg)	Normalized 10g SAR (W/kg)	Deviation (%)	Dipole S/N	Probe S/N	DAE S/N
S05	900	23.4	0.959	40.127	0.97	41.5	-1.13	-3.31	Pass	Pass	Pass	N/A	N/A	N/A	Sep. 11, 2021	900	7.12	0.326	6.52	-8.43	1d121	3820	1305
S06	900	23.4	0.959	40.127	0.97	41.5	-1.13	-3.31	Pass	Pass	Pass	N/A	N/A	N/A	Sep. 11, 2021	900	7.12	0.326	6.52	-8.43	1d121	3820	1305