

74, Seoicheon-ro 578beon-gil, Majang-myeon, Icheon-si, Gyeonggi-do, 17383 KOREA Tel. +82 31 634 6300 Fax. +82 31 645 6401

SAR TEST REPORT

Applicant Name:

Panasonic Corporation of North America Two Riverfront Plaza, 9th Floor, Newark, NJ 07102-5490, USA Date of Issue: Dec. 17, 2021 Test Report No.: HCT-SR-2112-FC005-R1 Test Site: HCT CO., LTD.

FCC ID:

ACJ9TGWL20B

Equipment Type:	Radio Module
Application Type	Certification
FCC Rule Part(s):	47CFR §2.1093
Model Name:	WL20B
Date of Test:	Dec. 01, 2021 ~ Dec. 02, 2021

This device has been shown to be capable of compliance for localized specific absorption rate (SAR) for uncontrolled environment/general population exposure limits specified in FCC KDB procedures and had been tested in accordance with the measurement procedures specified in FCC KDB procedures.

I attest to the accuracy of data. All measurements reported herein were performed by me or were made under my supervision and are correct to the best of my knowledge and belief. I assume full responsibility for the completeness of these measurements and vouch for the qualifications of all persons taking them.

Tested By

e ;:-

Jee-III, Lee Test Engineer SAR Team Certification Division

Reviewed By

Yun-jeang, Heo Technical Manager SAR Team Certification Division

This report only responds to the tested sample and may not be reproduced, except in full, without written approval of the HCT Co., Ltd.



REVISION HISTORY

The revision history for this test report is shown in table.

Revision No.	Date of Issue	Description
0	Dec. 17, 2021	Initial Release
1	Jan. 17, 2022	Revised Antenna Distance

This test results were applied only to the test methods required by the standard.

The above Test Report is not related to the accredited test result by (KS Q) ISO/IEC 17025 and KOLAS(Korea Laboratory Accreditation Scheme), which signed the ILAC-MRA.



Table of Contents



1. ATTESTATION OF TEST RESULT OF DEVICE UNDER TEST

Test Laboratory						
Company Name:	HCT Co., LTD					
Address:	74, Seoicheon-ro 578beon-gil, Majang-myeon, Icheon-si, Gyeonggi-do, 17383, Rep. of Korea					
Telephone:	+82 31 645 6300					
Fax.:	+82 31 645 6401					
Attestation of SAR test	Attestation of SAR test result					
Applicant Name:	Panasonic Corporation of North America					
FCC ID:	ACJ9TGWL20B					
Model:	WL20B					
Host Model Name	FZ-40					
EUT Type:	Radio Module					
Application Type:	Certification					



Dec. 1	Tx. Frequency	Equipment	Reported SAR	
Band	(MHz)	Class	1g SAR (W/Kg)	
2.4GHz WLAN	2 412 ~ 2 472	DTS	< 0.1	
U-NII-1	5 180 ~ 5 250	NII	N/A	
U-NII-2A	5 260 ~ 5 320	NII	0.19	
U-NII-2C	5 500 ~ 5 720	NII	0.17	
U-NII-3	5 745 ~ 5 825	NII	< 0.1	
Simultaneous Transmission Analysis 0.99				
Date(s) of Tests:	Dec. 01, 2021 ~ Dec. 02, 2021			



2. DEVICE UNDER TEST DESCRIPTION

2.1 DUT specification

Device Wireless specification overview					
Band & Mode	Operating Mode	Tx Frequency			
U-NII-1	Data	5 180 MHz ~ 5	5 250 MHz		
U-NII-2A	Data	5 260 MHz ~ 5 320 MHz			
U-NII-2C	Data	5 500 MHz ~ 5 720 MHz			
U-NII-3	Data	5 745 MHz ~ 5 825 MHz			
2.4 GHz WLAN	Data	2 412 MHz ~ 2 472 MHz			
Bluetooth / LE 5.0	Data	2 402 MHz ~ 2	2 480 MHz		
Device Serial Numbers	Mode		Serial Number		
	2.4 GHz WLAN, 5 GHz WLAN, Bluetooth		S0P-21-01894		



2.2 Output Power Specifications

This device operates using the following maximum output power specifications. SAR values were scaled to the maximum allowed power to determine compliance per the IEEE1528-2013.

2.2.1 Maximum 2.4 GHz, 5 GHz WIFI output power

			Maximum Target Power for Host Approval(dBm)				
Band	To shore to suc	Central	SISO		MIMO		
Danu	Technology	Channel	Main	Aux Antenna	Main	Aux	Main+Aux
			Antenna	Aux Antenna	Antenna	Antenna	Antenna
2.4 GHz	20MHz DSSS	1	19.5	19.5			
	(802.11b)	2-10	21.0	21.0			
		11	20.0	20.0			
		12	18.5	19.0			
		13	15.5	15.5			
	20MHz OFDM	1	17.0	17.0			
	(802.11g)	2	19.0	19.0			
		3-9	20.25	20.25			
		10	19.25	19.25			
		11	17.0	16.75			
		12	15.0	15.5			
		13	1.50	2.00			
	20MHz OFDM	1	17.0	17.0	14.0	14.0	17.0
	(802.11n)	2-10	20.5	20.5	18.0	18.0	21.0
		11	16.5	16.0	14.5	14.5	17.5
		12	15.0	15.5	13.0	13.0	16.0
		13	1.5	2.0	-1.0	-1.0	2.0
	20MHz OFDM	1	17.0	17.5	14.0	14.0	17.0
	(802.11ax)	2-10	20.0	20.0	17.0	17.0	20.0
		11	16.0	16.0	14.0	14.0	17.0
		12	15.5	15.5	12.5	12.5	15.5
		13	1.5	2.0	-1.0	-1.0	2.0
	40MHz OFDM	3	17.0	16.5	14.5	14.5	17.5
	(802.11n)	4-6	16.0	16.5	14.0	14.0	17.0
		7-9	16.00	16.00	13.5	13.5	16.5
		10	12.5	12.5	10.0	10.0	13.0
		11	5.0	5.0	3.0	3.0	6.0
	40MHz OFDM	3	16.5	16.5	14.0	14.0	17.0
	(802.11ax)	4-6	16.0	16.0	14.5	14.5	17.5
		7-9	16.0	16.0	13.5	13.5	16.5
		10	12.0	12.5	10.5	10.5	13.5
		11	4.5	5.5	3.0	3.0	6.0



Band	Technology	Central	al Maximum Target Power for Host Approval(dBm))
		Channel	SI	SO		MIMO	
			Main	Aux Antenna	Main	Aux	Main+Aux
			Antenna		Antenna	Antenna	Antenna
Wi-Fi 5GHz	20MHz OFDM	36	18.5	18.0			
Band I,II	(802.11a)	40	21.0	21.0			
		44	21.0	21.0			
		48	21.0	21.0			
		52	21.5	21.0			
		56	21.0	21.0			
		60	21.0	21.0			
		64	17.5	17.5			
	20MHz OFDM	36	18.0	18.5	16.0	16.0	19.0
	(802.11n)	40	21.0	20.5	17.5	17.5	20.5
	(802.11ac)	44	21.0	20.5	17.5	17.5	20.5
		48	21.0	19.5	18.0	18.0	21.0
		52	21.0	21.0	18.0	18.0	21.0
		56	21.0	21.0	18.0	18.0	21.0
		60	21.0	21.0	18.0	18.0	21.0
		64	17.5	17.5	13.5	13.5	16.5
	20MHz OFDM	36	18.0	18.0	16.0	16.0	19.0
	(802.11ax)	40	21.0	20.5	18.0	18.0	21.0
		44	21.0	21.0	18.0	18.0	21.0
		48	21.0	21.0	18.0	18.0	21.0
		52	21.0	21.0	18.0	18.0	21.0
		56	21.0	21.0	18.0	18.0	21.0
		60	21.0	21.0	18.0	18.0	21.0
		64	17.5	17.5	13.5	13.5	16.5
	40MHz OFDM	38	18.5	18.0	15.5	15.5	18.5
	(802.11n)	46	21.0	19.5	17.5	17.5	20.5
	(802.11ac)	54	20.5	20.0	16.5	16.5	19.5
		62	16.5	16.5	13.0	13.0	16.0
	40MHz OFDM	38	18.0	18.0	15.0	15.0	18.0
	(802.11ax)	46	21.0	19.5	17.5	17.5	20.5
		54	20.5	20.0	16.5	16.5	19.5
		62	16.5	16.5	13.0	13.0	16.0
	80MHz OFDM	42	18.5	18.0	15.5	15.5	18.5
	(802.11ac)	58	17.5	17.5	14.0	14.0	17.0
	80MHz OFDM	42	18.5	18.0	15.5	15.5	18.5
	(802.11ax)	58	17.5	17.5	14.0	14.0	17.0
	160MHz OFDM (802.11ac)	50	15.0	15.0	12.0	12.0	15.0
	160MHz OFDM (802.11ax)	50	15.0	15.0	12.5	12.5	15.5



Band	Technology	Central		Maximum Target	Power for Host	Approval(dBm))
		Channel	00	SISO		MIMO	
			Main	Aux Antenna	Main	Aux	Main+Aux
			Antenna		Antenna	Antenna	Antenna
Wi-Fi 5GHz	20MHz OFDM	100	17.5	17.5			
Band III	(802.11a)	104-136	21.0	20.0			
		140	18.0	18.0			
	20MHz OFDM	100	17.5	17.5	14.0	14.0	17.0
	(802.11n)	104-136	21.0	20.0	18.5	18.5	21.5
	(802.11ac)	140	18.0	18.0	14.5	14.5	17.5
		144	20.5	20.0	18.0	18.0	21.0
	20MHz OFDM	100	17.5	17.5	14.0	14.0	17.0
	(802.11ax)	104-136	21.0	20.0	18.5	18.5	21.5
		140	17.5	18.0	14.0	14.0	17.0
		144	20.5	20.0	18.5	18.5	21.5
	40MHz OFDM	102	17.5	18.0	14.0	14.0	17.0
	(802.11n)	110-126	20.5	20.0	18.0	18.0	21.0
	(802.11ac)	134	19.0	19.0	17.0	17.0	20.0
		142	20.5	20.0	18.5	18.5	21.5
	40MHz OFDM	102	17.5	18.0	14.0	14.0	17.0
	(802.11ax)	110-126	20.5	20.0	18.0	18.0	21.0
		134	19.0	19.5	17.0	17.0	20.0
		142	21.0	20.0	18.5	18.5	21.5
	80MHz OFDM	106	18.0	18.0	15.0	15.0	18.0
	(802.11ac)	122	20.5	20.0	18.5	18.5	21.5
		138	21.0	20.0	19.0	19.0	22.0
	80MHz OFDM	106	18.00	17.5	15.0	15.0	18.0
	(802.11ax)	122	19.5	19.5	18.5	18.5	21.5
		138	21.0	20.0	18.5	18.5	21.5
	160MHz OFDM (802.11ac)	114	14.5	15.0	12.0	12.0	15.0
	160MHz OFDM (802.11ax)	114	14.5	14.5	12.0	12.0	15.0
Wi-Fi 5GHz Band IV	20MHz OFDM (802.11a)	149-165	21.0	19.5	-	-	-
	20MHz OFDM	149	21.0	19.5	18.0	18.0	21.0
	(802.11n)	157	21.0	19.5	18.0	18.0	21.0
	(802.11ac)	165	21.0	19.5	18.0	18.0	21.0
	20MHz OFDM	149	21.0	19.5	17.5	17.5	20.5
	(802.11ax)	157	21.0	19.5	18.0	18.0	21.0
		165	21.0	19.5	18.0	18.0	21.0
	40MHz OFDM	151	21.0	19.6	18.0	18.0	21.0
	(802.11n) (802.11ac)	159	21.0	19.6	18.0	18.0	21.0
	40MHz OFDM	151	21.0	19.5	18.0	18.0	21.0
	(802.11ax)	159	21.0	19.5	18.0	18.0	21.0
	80MHz OFDM (802.11ac)	155	19.5	19.5	16.5	16.5	19.5
	80MHz OFDM (802.11ax)	155	19.0	19.0	16.5	16.5	19.5



2.2.2 Nominal and Maximum Bluetooth Power

Mode / Band			Modulated Average (dBm)	
	DH-5	Maximum	10.5	
	DL-9	Nominal	9.5	
			Maximum	7.0
Divisionath	2-DH5	Nominal	6.0	
Bluetooth		Maximum	7.0	
	3-DH5	Nominal	6.0	
		Maximum	7.0	
	Low Energy	Nominal	6.0	



2.3 Test Methodology and Procedures

The tests documented in this report were performed in accordance with IEEE Standard 1528-2013 and the following published KDB procedures.

- FCC KDB Publication 248227 D01 802.11 Wi-Fi SAR v02r02
- FCC KDB Publication 616217 D04 SAR for Laptop and tablets v01r02
- FCC KDB Publication 447498 D01 General SAR Guidance v06
- FCC KDB Publication 865664 D01 SAR measurement 100 MHz to 6 GHz v01r04
- FCC KDB Publication 865664 D02 SAR Reporting v01r02

In Addition to the above, the following information was used.


- April 2015 TCB Workshop Notes (Simultaneous transmission summation clarified)
- October 2016 TCB Workshop Notes (Bluetooth Duty Factor)
- November 2017 TCBC Workshop Notes (LTE Carrier Aggregation)
- April 2018 TCBC Workshop Notes (LTE DL CA SAR Test Exclusion)
- May 2017 TCBC Workshop Notes (LTE 4x4 Downlink MIMO)
- April 2019 TCBC Workshop Notes (IEEE 802.11 ax)



2.4 DUT Antenna Locations [Host Model]

Antenna Distance

<Front View>





2.5 SAR Summation Scenario for Host Model



Simultaneous transmission paths

This device contains multiple transmitters that may operate simultaneously, and therefore requires a simultaneous transmission analysis according to FCC KDB 447498 D01v06.

Simultaneous Transmission Scenarios						
Applicable Combination	Body Exposure Condition					
UMTS+2.4 GHz Wif 1(Main) + 2.4 GHz Wifi 2(Aux)	Yes					
LTE+2.4 GHz Wif 1(Main) + 2.4 GHz Wifi 2(Aux)	Yes					
UMTS+2.4 GHz WiFi #1(Main) + 2.4 GHz Bluetooth	Yes					
LTE+2.4 GHz WiFi #1(Main) + 2.4 GHz Bluetooth	Yes					
UMTS+5 GHz WiFi #1(Main) + 5GHz WLAN #2(Aux)	Yes					
LTE+5 GHz WiFi #1(Main) + 5GHz WLAN #2(Aux)	Yes					
UMTS +5 GHz WiFi #1(Main)+ 2.4 GHz Bluetooth	Yes					
LTE+5 GHz WiFi #1(Main)+ 2.4 GHz Bluetooth	Yes					

1. All licensed modes share the same antenna path and cannot transmit simultaneously.

2. The highest reported SAR for each exposure condition is used for SAR summation purpose.



2.6 WLAN Test Considerations.

Since U-NII-1 and U-NII-2A bands have the same maximum output power and the highest reported SAR for U-NII-2A is less than 1.2 W/kg for 1g SAR and is less than 3.0 W/kg for 10g SAR, SAR is not required for U-NII-1 band according to FCC KDB 248227D01v02r02.

This device supports IEEE 802.11 ac with the following features:

- a) Up to 80 MHz Bandwidth only
- b) No aggregate channel configurations
- c) 2Tx antenna output
- d) 256 QAM is supported
- e) TDWR channels are supported.
- f) Straddle channels are supported.

Per April 2019 TCB Workshop Notes, SAR testing was not required for 802.11ax when applying the initial test configuration procedures of KDB 248227, with 802.11ax considered a higher order 802.11 mode



3. INTRODUCTION

The FCC has adopted the guidelines for evaluating the environmental effects of radio frequency radiation in ET Docket 93-62 on Aug. 6, 1996 to protect the public and workers from the potential hazards of RF emissions due to FCC-regulated portable devices.

The safety limits used for the environmental evaluation measurements are based on the criteria published by the American National Standards Institute (ANSI) for localized specific absorption rate (SAR) in IEEE/ANSI C95.1-1992 Standard for Safety Levels with Respect to Human Exposure to Radio Frequency Electromagnetic Fields, 3 kHz to 300 GHz. 1992 by the Institute of Electrical and Electronics Engineers, Inc., New York 10017. The measurement procedure described in IEEE/ANSI C95.3-1992 Recommended Practice for the Measurement of Potentially Hazardous Electromagnetic Fields - RF and Microwave is used for guidance in measuring SAR due to the RF radiation exposure from the Equipment Under Test (EUT). These criteria for SAR evaluation are similar to those recommended by the National Council on Radiation Protection and Measurements (NCRP) in Biological Effects and Exposure Criteria for Radio Frequency Electromagnetic Fields," NCRP Report No. 86 NCRP, 1986, Bethesda, MD 20814. SAR is a measure of the rate of energy absorption due to exposure to an RF transmitting source. SAR values have been related to threshold levels for potential biological hazards.

SAR Definition

Specific Absorption Rate (SAR) is defined as the time derivative of the incremental electromagnetic energy (dW) absorbed by (dissipated in) an incremental mass (dm) contained in a volume element (dV) of a given density (r). It is also defined as the rate of RF energy absorption per unit mass at a point in an absorbing body.

$$SAR = \frac{d}{d t} \left(\frac{d U}{d m} \right)$$

Figure 1. SAR Mathematical Equation SAR is expressed in units of Watts per Kilogram (W/kg)

 $SAR = \sigma E^2 / \rho$

NOTE: The primary factors that control rate of energy absorption were found to be the wavelength of the incident field in relations to the dimensions and geometry of the irradiated organism, the orientation of the organism in relation to the polarity of field vectors, the presence of reflecting surfaces, and whether conductive contact is made by the organism with a ground plane.



4. DESCRIPTION OF TEST EQUIPMENT

4.1 SAR MEASUREMENT SETUP

These measurements are performed using the DASY5 automated dosimetric assessment system. It is made by Schmid & Partner Engineering AG (SPEAG) in Zurich, Switzerland. It consists of high precision robotics system (Staubli), robot controller, Pentium III computer, near-field probe, probe alignment sensor, and the generic twin phantom containing the brain equivalent material. The robot is a six-axis industrial robot performing precise movements to position the probe to the location (points) of maximum electromagnetic field (EMF) (see Figure.2).

A cell controller system contains the power supply, robot controller, teach pendant (Joystick), and remote control, is used to drive the robot motors. The PC with Windows 7 is working with SAR Measurement system DASY5, A/D interface card, monitor, mouse, and keyboard. The Staubli Robot is connected to the cell controller to allow software manipulation of the robot. A data acquisition electronic (DAE) circuit performs the signal amplification, signal multiplexing, AD-conversion, offset measurements, mechanical surface detection, collision detection, etc. is connected to the Electro-optical coupler (EOC). The EOC performs the conversion from the optical into digital electric signal of the DAE and transfers data to the PC plug-in card.




Figure 2. HCT SAR Lab. Test Measurement Set-up

The DAE consists of a highly sensitive electrometer-grade preamplifier with auto-zeroing, a channel and gainswitching multiplexer, a fast 16 bit AD-converter and a command decoder and control logic unit. Transmission to the PC-card is accomplished through an optical downlink for data and status information and an optical uplink for commands and clock lines. The mechanical probe mounting device includes two different sensor systems for frontal and sidewise probe contacts. They are also used for mechanical surface detection and probe collision detection. The robot uses its own controller with a built in VME-bus computer. The system is described in detail in.



5. SAR MEASUREMENT PROCEDURE

The evaluation was performed using the following procedure compliant to FCC KDB Publication 865664 D01v01r04 and IEEE 1528-2013.

- 1. The SAR distribution at the exposed side of the head or body was measured at a distance no more than 5.0 mm from the inner surface of the shell. The area covered the entire dimension of the DUT's head and body area and the horizontal grid resolution was depending on the FCC KDB 865664 D01v01r04 table 4-1 & IEEE 1528-2013.
- Based on step, the area of the maximum absorption was determined by sophisticated interpolations routines implemented in DASY software. When an Area Scan has measured all reachable point. DASY system computes the field maximal found in the scanned are, within a range of the maximum. SAR at this fixed point was measured and used as a reference value.
- 3. Around this point, a volume was assessed according to the measurement resolution and volume size requirements of FCC KDB 865664 D01v01r04 table 4-1 and IEEE 1528-2013. On the basis of this data set, the spatial peak SAR value was evaluated with the following procedure (reference from the DASY manual.)
 - a. The data at the surface were extrapolated, since the center of the dipoles is no more than 2.7 mm away from the tip of the probe (it is different from the probe type) and the distance between the surface and the lowest measuring point is 1.2 mm. The extrapolation was based on a least square algorithm. A polynomial of the fourth order was calculated through the points in z-axes. This polynomial was then used to evaluate the points between the surface and the probe tip.

b. The maximum interpolated value was searched with a straight-forward algorithm. Around this maximum the SAR values averaged over the spatial volumes (1 g or 10 g) were computed using the 3D-Spline interpolation algorithm. The 3D-spline is composed of three one-dimensional splines with the "Not a knot" condition (in x, y, and z directions. The volume was integrated with the trapezoidal algorithm. One thousand points (10 x 10 x 10) were interpolated to calculate the average.

c. All neighboring volumes were evaluated until no neighboring volume with a higher average value was found.

4. The SAR reference value, at the same location as step 2, was re-measured after the zoom scan. If the value changed by more than 5 %, the SAR evaluation and drift measurements were repeated.



Area scan and zoom scan resolution setting follow KDB 865664 D01v01r04 quoted below.

		≤ 3 GHz	> 3 GHz	
	n closest measurement point bbe sensors) to phantom	5±1 mm	·δ·ln(2)±0.5 mm	
Maximum probe angle surface normal at the measurer	from probe axis to phantom nent location	30° ±1 °	20 °±1 °	
		≤ 2 GHz: ≤15 mm 2-3 GHz: ≤12 mm	3-4	
Maximum area scan S	patial resolution : Δx _{Area} , Δy _{Area}	When the x or y dimension of the test device, in the measurement plane orientation, is smaller than the above, the measurement resolution must be \leq the corresponding x or y dimension of the test device with at least one measurement point on the test device.		
Maximum zoom scan S	patial resolution : Δx _{zoom} , Δy _{zoom}	≤ 2 GHz: ≤8mm 2-3 GHz: ≤5mm*	3-4 ଖłz: ≤5 mm* 4-6 ଖłz: ≤4 mm*	
	uniform grid : Δz_{zoom}(n)	≤ 5 mm	3-4 GHz: ≤4 mm 4-5 GHz: ≤3 mm 5-6 GHz: ≤2 mm	
Maximum zoom scan Spatial resolution normal to phantom surface	dz _{zoom} (1): between 1 st two Points closest graded to phantom surface	≤ 4 mm	3-4 GHz: ≤3 mm 4-5 GHz: ≤2.5 mm 5-6 GHz: ≤2 mm	
	Δz _{zoom} (n>1): between subsequent Points	≤1.5 ·Δz _{zoom} (n-1)		
Minimum zoom scan volume	х, у, z	≥ 30 mm	3-4 Głz: ≥28 mm 4-5 Głz: ≥25 mm 5-6 Głz: ≥22 mm	

Note: δ is the penetration depth of a plane-wave at normal incidence to the tissue medium; see draft standard IEEE P1528-2011 for details.

* When zoom scan is required and the reported SAR from the area scan based 1-g SAR estimation procedures of KDB 447498 is \leq 1.4 W/kg, \leq 8 mm, \leq 7 mm and \leq 5 mm zoom scan resolution may be applied, respectively, for 2 GHz to 3 GHz, 3 GHz to 4 GHz and 4 GHz to 6 GHz.




6. EUT Testing Position

6.1 Laptop host platform test requirements Per KDB Publication 447498 D01 and 616217 D04v01r02

The required minimum test separation distance for incorporating transmitters and antennas into laptop, notebook and netbook computer displays is determined with the display screen opened at an angle of 90° to the keyboard compartment. When antennas are incorporated in the keyboard section of a laptop computer, SAR is required for the bottom surface of the keyboard. Provided tablet use conditions are not supported by the laptop computer, SAR tests for bystander exposure from the edges of the keyboard and display screen of laptop computers are generally not required

While users would normally keep the display open at angles greater than 90° degrees, for SAR testing purposes and to maintain conservativeness, we keep the display open at 90° degrees from the keyboard to perform the SAR measurement.



Considering the distance between the antenna of the WLAN, WWAN and the bottom side of the DUT, the SAR test of WLAN, WWAN were omitted according to KDB 447498 D01 and KDB 616217.



SAR Test Configurations

Antenna	Band	Freq.	Max.	Max. Power Separation Distances (mm)		SAR Test Exclusion Thresholds (test separation distances > 50 mm) mW	Device Conifigurations for SAR Testing	
		[MHz]	dBm	mW	Bottom	Bottom	Bottom	
	BT	2480	10.5	11.2	276.8	2363.25	0.4W/Kg (N/A)	
WLAN Aux	Wifi 2.4GHz	2472	22	158.5	276.8	2363.40	0.4W/Kg (N/A)	
	Wifi 5GHz	5.825	22	158.5	276.8	2330.15	0.4W/Kg (N/A)	

Per FCC KDB447498 D01 General RF Exposure Guidance v06 Sec 4.3.2 b) When an antenna qualifies for the standalone SAR test exclusion of 4.3.1 and also transmits simultaneously with other antennas, the standalone SAR value must be estimated according to the following to determine the simultaneous transmission SAR test exclusion criteria.

1) [(max. power of channel, including tune-up tolerance, mW) / (min. test separation distance, mm)]·[$\sqrt{f}(GHz)/x$] W/kg, for test separation distances \leq 50 mm;

where x = 7.5 for 1-g SAR and x = 18.75 for 10-g SAR.

2) 0.4 W/kg for 1-g SAR and 1.0 W/kg for 10-g SAR, when the test separation distance is > 50 mm.

For simultaneous transmission analysis with the WLAN Aux Ant 2 within the host device, the SAR evaluation value of the WWAN antenna was 0.4 W/kg according to FCC KDB 448498.D01



Per FCC KDB Publication 616217 D04v01r02, the bottom surface should be tested for SAR compliance with the laptop touching the phantom. The SAR Exclusion Threshold in KDB 447498 D01v06 can be applied to determine SAR test exclusion for adjacent edge configurations. The closet distance from the antenna to an adjacent tablet edge is used to determine if SAR testing is required for the adjacent edges, with the adjacent edge positioned against the phantom and the edge containing the antenna positioned perpendicular to the phantom.

Test Configurations for the WLAN Aux Antenna

Test Configurations	Test Antenna-to- SAR onfigurations edge/surface Required		Note		
Bottom Side (Laptop Mode)	276.8 mm	No	SAR is not required since separation distance from antenna to user is more far away		

According to FCC KDB 447498 D01 Sec 4.3 b), SAR measurement on the bottom side is excluded, and according to FCC KDB 447498 D01 Sec 4.3.2,b) ,the estimated 1g SAR value applied for simultaneous transmission evaluation is as follows.

Configuration	Estimated 1g SAR
2.4 GHz WLAN Aux	0.4
Bluetooth	0.4
5GHz WLAN Aux	0.4

For simultaneous transmission analysis with the WLAN Aux within the host device, the SAR evaluation value of the W WLAN Aux was 0.4 W/kg according to FCC KDB 448498.D01



7. RF EXPOSURE LIMITS

HUMAN EXPOSURE	UNCONTROLLED ENVIRONMENT General Population W/kg	CONTROLLED ENVIRONMENT Occupational W/kg	
The SAR averaged over the whole body mass.	0.08	0.4	
The peak spatially-averaged SAR for the head, neck and trunk, averaged over any 1 g of tissue*	1.6	8	
The peak spatially-averaged SAR in the limbs, averaged over any 10 g of tissue*	4	20	

NOTES:

- * The Spatial Peak value of the SAR averaged over any 1 g of tissue (defined as a tissue volume in the shape of a cube) and over the appropriate averaging time.
- ** The Spatial Average value of the SAR averaged over the whole-body.
 - *** The Spatial Peak value of the SAR averaged over any 10 g of tissue (defined as a tissue volume in the shape of a cube) and over the appropriate averaging time.

Uncontrolled Environments are defined as locations where there is the exposure of individuals who have no knowledge or control of their exposure. The general population/uncontrolled exposure limits are applicable to situations in which the general public may be exposed or in which persons who are exposed as a consequence of their employment may not be mad fully aware of the potential for exposure or cannot exercise control over their exposure. Members of the general public would come under this category when exposure is not employment-related; for example, in the case of a wireless transmitter that exposes persons in its vicinity.

Controlled Environments are defined as locations where there is exposure that may be incurred by persons who are aware of the potential for exposure, (i.e.as a result of employment or occupation). In general, occupational/controlled exposure limits are applicable to situations in which persons are exposed as a consequence of their employment, who have been made fully aware of the potential for exposure and can exercise control over their exposure. This exposure category is also applicable when the exposure is of a transient nature due to incidental passage through a location where the exposure levels may be higher than the general population/uncontrolled limits, but the exposed person is fully aware of the potential for exposure and can exercise control over his or her exposure by leaving the area or by some other appropriate means.



8. SAR General Measurement Procedures

Power Measurements for licensed transmitters are performed using a base simulator under digital average power.

8.1 Measured and Reported SAR

Per FCC KDB Publication 447498 D01v06, when SAR is not measured at the maximum power level allowed for production units, the results must be scaled to the maximum tune-up tolerance limit according to the power applied to the individual channels tested to determine compliance. For simultaneous transmission, the measured aggregate SAR must be scaled according to the sum of the differences between the maximum tune-up tolerance and actual power used to test each transmitter. When SAR is measured at or scaled to the maximum tune-up tolerance limit, the results are referred to as Reported SAR. The highest reported SAR results are identified on the grant of equipment authorization according to procedures in KDB 690783 D01v01r03.

8.2 SAR Testing with 802.11 Transmitters

The normal network operating configurations of 802.11 transmitters are not suitable for SAR measurements. Unpredictable fluctuations in network traffic and antenna diversity conditions can introduce undesirable variations in SAR results. The SAR for these devices should be measured using chipset based test mode software to ensure the results are consistent and reliable. See KDB Publication 248227 D01v02r02 for more details.

8.3 General Device Setup

Chipset based test mode software is hardware dependent and generally varies among manufacturers. The device operating parameters established in test mode for SAR measurements must be identical to those programmed in production units, including output power levels, amplifier gain settings and other RF performance tuning parameters.

A periodic duty factor is required for current generation SAR system to measure SAR. When 802.11 frame gaps are accounted for in the transmission, a maximum transmission duty factor of 92-96% is typically achievable in most test mode configurations. A minimum transmission duty factor of 85% is required to avoid certain hardware and device implementation issues related to wide range SAR scaling. The reported SAR is scaled to 100% transmission duty factor to determine compliance at the maximum tune-up tolerance limit.

8.4 U-NII-1 and U-NII-2A

For devices that operate in both U-NII-1 and U-NII2A bands, when the same maximum output power is specified for both bands, SAR measurement using OFDM SAR test procedures is not required for U-NII-1 unless the highest reported SAR for U-NII-2A is > 1.2 W/kg for 1g SAR or > 3.0 W/kg for 10g SAR. When different maximum output powers are specified for the bands, SAR measurement for the U-NII band with the lower maximum output power is not required unless the highest reported SAR for the U-NII band with the higher maximum output power, adjusted by the ratio of lower to higher specified maximum output power for the two bands, is > 1.2 W/kg for 1g SAR or > 3.0 W/kg for 1g SAR.

8.5 U-NII-2C and U-NII-3

The frequency range covered by U-NII-2C and U-NII-3 is 380 Mt (5.47 GHz ~ 5.85 GHz), which requires a minimum of at least two SAR probe calibration frequency points to support SAR measurements. When Terminal Doppler Weather Radar (TDWR) restriction applies, the channels at 5.60 GHz ~ 5.65 GHz in U-NII-2C band must be disabled with acceptable mechanisms and documented in the equipment certification. Unless band gap channels are permanently disabled, SAR must be considered for these channels.

8.6 2.4 GHz SAR test Requirements

SAR is measured for 2.4 GHz 802.11b DSSS using either the fixed test position 2.4 GHz 802.11 g/n OFDM are additionally evaluated for SAR if the highest reported SAR for 802.11b, adjusted by the ratio of the OFDM to DSSS specified maximum output power, is > 1.2 W/kg.



8.7 OFDM Transmission Mode and SAR Test Channel Selection

For the 2.4 GHz and 5 GHz bands, when the same maximum output power was specified for multiple OFDM transmission mode configurations in a frequency band or aggregated band, SAR is measured using the configuration with the largest channel bandwidth, lowest order modulation and lowest data rate and lowest order 802.11 a/g/n/ac mode. When the maximum output power of a channel is the same for equivalent OFDM configurations; for example, 802.11a, 802.11n and 802.11 ac or 802.11g and 802.11n with the same channel bandwidth, modulation and data rate etc., the lower order 802.11 mode i.2., 802.11a, then 802.11n and 802.11ac or 802.11g then 802.11n, is used for SAR measurement. When the maximum output power are the same for multiple test channels, either according to the default or additional power measurement requirements, SAR is measured using the channel closest to the middle of the frequency band or aggregated band. When there are multiple channels with the same maximum output power, SAR is measured using the higher number channel.

8.8 Initial Test Configuration Procedure

For OFDM, in both 2.4 (Hz and 5 (Hz bands, an initial test configuration is determined for each frequency band and aggregated band, according to the transmission mode with the highest maximum output power specified for SAR measurements. When the same maximum output power is specified for multiple OFDM transmission mode configurations in a frequency band or aggregated band, SAR is measured using the configuration(s) with the largest channel bandwidth, lowest order modulation, and lowest data rate. If the average RF output powers of the highest identical transmission modes are within 0.25 dB of each other, mid channel of the transmission mode with highest average RF output power is the initial test channel. Otherwise, the channel of the transmission mode with the highest average RF output conducted power will be the initial test configuration.

When the reported SAR is ≤ 0.8 W/kg, no additional measurements on other test channels are required. Otherwise, SAR is evaluated using the subsequent highest average RF output channel until the reported SAR result is 1.2 W/kg or all channels are measured. When there are multiple untested channels having the same subsequent highest average RF output power, the channel with higher frequency from the lowest 802.11 mode is considered for SAR measurements.

8.9 Subsequent Test Configuration Procedures

For OFDM configurations in each frequency band and aggregated band, SAR is evaluated for initial test configuration using the fixed test position or the initial test position on procedure. When the highest reported SAR (for the initial test configuration), adjusted by the ratio of the specified maximum output power of the subsequent test configuration to initial test configuration, is ≤ 1.2 W/kg for 1g SAR and ≤ 3.0 W/kg for 10g SAR, no additional SAR tests for the subsequent test configurations are required.

8.10 MIMO SAR considerations

Per KDB Publication 248227 D01v02r02, the simultaneous SAR provisions in KDB Publication 447498 D01v06 should be applied to determine simultaneous transmission SAR test exclusion for WIFI MIMO. If the sum of 1g single transmission chain SAR measurements is <1.6 W/kg, no additional SAR measurements for MIMO are required Alternatively, SAR for MIMO can be measured with all antennas transmitting simultaneously at the specified maximum output power of MIMO operation. When 10g SAR measurement is considered, a factor of 2.5 is applied to the thresholds above.



9. OUTPUT POWER SPECIFICATIONS

This device operates using the following maximum output power specifications. SAR values were scaled to the maximum allowed power to determine compliance per KDB publication 447498 D01v06.

9.1 WIFI Conducted Power measurement

9.1.1 IEEE 802.11 (2.4 GHz) Maximum Conducted Power For Laptop Mode

SISO Ant

Mode	Frequency [MHz]	Channel	IEEE 802.11 (2.4 GHz) SISO RF Conducted Power [dBm]				
			Main Ant.	Aux Ant.			
	2 412	1	19.46	19.67			
	2 417	2	20.91	20.57			
	2 437	6	21.00	20.83			
802.11b	2 457	10	20.92	20.91			
	2 462	11	19.67	19.37			
	2 467	12	18.62	18.86			
	2 472	13	15.70	15.96			
	2 422	3	20.53	20.61			
802.11g	2 437	6	20.16	20.15			
-	2 452	9	20.38	20.20			
	2 417	2	20.32	20.35			
802.11n	2 437	6	20.04	20.08			
	2 457	10	20.30	20.10			

MIMO Mode

Mode	Frequency [Mz]	Channel	IEEE 802.11 (2.4 배) MIMO RF Conducted Power [dBm]				
			Main Ant.	Aux Ant.	MIMO		
	2 417	2	17.23	17.36	20.29		
802.11n	2 437	6	17.36	17.55	20.45		
	2 457	10	17.41	17.54	20.47		



9.1.2 IEEE 802.11 (5 GHz) Maximum Conducted Power For Laptop Mode

			IEEE 802.11 (5 GHz)	Reduced Average
Mode	Frequency [MHz]	Channel	Conducted F	Power [dBm]
			Main Ant.	Aux Ant.
	5 180	36	18.44	17.88
	5 200	40	20.96	20.92
	5 220	44	21.06	20.77
	5 240	48	20.96	20.72
802.11a	5 260	52	21.00	20.75
	5 280	56	20.65	20.69
	5 300	60	20.53	20.73
	5 320	64	17.08	17.67
	5 180	36	17.71	18.33
	5 200	40	20.71	20.78
	5 220	44	20.79	20.61
802.11n20	5 240	48	20.77	19.55
	5 260	52	20.65	20.56
	5 280	56	20.53	20.53
	5 300	60	20.45	20.45
	5 320	64	16.98	16.80
	5 755	151	20.60	19.60
802.11n40	5 795	159	20.69	19.53
	5 180	36	17.74	17.82
	5 200	40	20.71	20.83
	5 220	44	20.77	20.70
	5 240	48	20.69	19.61
802.11ac20	5 260	52	20.57	21.28
	5 280	56	20.45	21.21
	5 300	60	20.35	20.68
	5 320	64	16.89	17.08
802.11ac40	5 755	151	20.59	19.56
	5 795	159	20.46	19.65
	5 530	106	18.19	18.07
802.11ac80	5 610	122	20.99	19.92
	5690	138	21.00	20.00
	5 180	36	17.51	17.68
	5 200	40	20.44	20.61
	5 220	44	20.53	20.46
	5 240	48	20.43	20.37
802.11ax_20	5 260	52	20.36	20.43
	5 280	56	20.22	20.38
	5 300	60	20.13	20.47
	5 320	64	16.67	16.92



			IEEE 802.11 (5 GHz) F	Reduced Average		
Mode	Frequency [MHz]	Channel	Conducted Power [dBm]			
			Main Ant.	Aux Ant.		
	5 180	36	15.40	15.62		
	5 200	40	17.14	17.34		
	5 220	44	17.13	17.23		
802.11n20	5 240	48	17.59	17.67		
	5 260	52	17.42	17.63		
	5 280	56	17.42	17.61		
	5 300	60	17.24	17.58		
	5 320	64	13.16	13.40		
802.11n40	5 755	151	18.00	18.09		
	5 795	159	17.55	18.02		
	5 180	36	15.62	15.86		
	5 200	40	17.13	17.34		
	5 220	44	17.16	17.25		
802 1122 20	5 240	48	17.64	17.71		
802.11ac 20	5 260	52	17.53	17.74		
802.11ac 20	5 280	56	17.49	17.69		
	5 300	60	17.42	17.79		
	5 320	64	13.35	13.40		
802.11ac40	5 755	151	17.95	18.08		
	5 795	159	17.53	18.07		
000.44	5 530	106	14.81	14.87		
802.11ac80	5 610	122	18.71	18.99		
	5690	138	18.90	19.00		
	5 180	36	15.58	15.59		
	5 200	40	17.53	17.76		
	5 220	44	17.67	17.72		
	5 240	48	17.47	17.60		
802.11ax_20	5 260	52	17.47	17.60		
	5 280	56	17.31	17.58		
	5 300	60	17.22	17.50		
	5 320	64	13.07	13.30		

Justification for test configurations for WLAN per KDB Publication 248227 D01v02r02:

• Power measurements were performed for the transmission mode configuration with the highest maximum

output power specified for production units.

• For transmission mode with the same maximum output power specification, powers were measured for the largest channel bandwidth, lowest order modulation and lowest data rate.

• For transmission modes with identical maximum specified output power, channel bandwidth, modulation and data rates, power measurements were required for all identical configurations.

• For each transmission mode configuration, powers were measured for the highest and lowest channels; and at the mid-band channel(s) when there were at least 3 channels supported. For configurations with multiple mid-band channels, due to an even number of channels, both channels were measured.

Test Configuration



9.1.3 Bluetooth Maximum Conducted Power

Mode	Channel	Ant	Bluetooth Power [dBm]
	0	Ant.2 (Aux)	9.09
DH5	39	Ant.2 (Aux)	9.47
	78	Ant.2 (Aux)	9.3
	0	Ant.2 (Aux)	5.82
2-DH5	39	Ant.2 (Aux)	6.01
	78	Ant.2 (Aux)	5.80
	0	Ant.2 (Aux)	5.83
3-DH5	39	Ant.2 (Aux)	6.02
	78	Ant.2 (Aux)	5.82

The Burst averaged-conducted power

Per October 2016 TCB Workshop Notes:

When call box and Bluetooth protocol are used for Bluetooth SAR measurement, time-domain plot is required to identify duty factor for supporting the test setup and result.

Bluetooth duty cycle was measured using Bluetooth tester equipment (CBT / R&S) with Bluetooth DH5 mode.

Bluetooth

arker 3 ∆ 3.75000 ms	PNO: Fast -+-	SPIGE DUT	ALEWALTO Avg Type: Log-Pwr	07:30:33 AM Dec 15, 2021 TRACE 1 2 3 4 TVRE 001 1 2 3 4 Det 8 70 1 1 1 1	Marker
	IFGain:Low	#Atten: 16 dB			Select Marker
Ref Offset 30.83 dB	ş		Δ	Mkr3 3.750 ms -0.02 dB	3
00 140		Ka	1∆2 ∳ ³	364	Norma
6.0	Roll-aview		byranistaja		Delta
60 60					Fixed
enter 2.441000000 GHz es BW 8 MHz	VBW 8	.0 MHz	Sweep 1	Span 0 Hz 0.00 ms (1001 pts)	no
KRI MODE TRC SCL X			UNCTION FUNCTION WIDTH	PUNCTION VALUE	
$ \begin{array}{c ccccccccccccccccccccccccccccccccccc$	2 890 ms (Δ) 4.100 ms 3.750 ms (Δ) 4.100 ms	-0.36 dB 9.65 dBm -0.02 dB 9.65 dBm			Properties
9 0					More 1 of 2
			AV (A)	3	

Duty Cycle

= (BT-On time /BT-Full time) =(2.890/3.750) = 0.771 (DH5)

Duty factor= 1/Duty cycle : 1.300



10. SYSTEM VERIFICATION

10.1 Tissue Verification

The body simulating material is calibrated by HCT using the DAKS 3.5 to determine the conductivity and permittivity.

	Table for Head Tissue Verification											
Date of Tests	Tissue Temp. (°C)	Tissue Type	Freq. (MHz)	Measured Conductivit y σ (S/m)	Measured Dielectric Constant, ε	Target Conductivit y σ (S/m)	Target Dielectric Constant, ε	% dev σ	% dev ε			
			2400	1.803	39.413	1.756	39.290	2.68%	0.31%			
12/01/2021	22.4	2450H	2450	1.856	39.161	1.800	39.200	3.11%	-0.10%			
			2500	1.903	38.963	1.855	39.140	2.59%	-0.45%			
		54001	5180	4.535	36.696	4.635	36.010	-2.16%	1.91%			
			5250	4.758	36.877	4.706	35.930	1.10%	2.64%			
			5280	4.682	36.699	4.737	35.894	-1.16%	2.24%			
			5320	4.754	36.988	4.778	35.846	-0.50%	3.19%			
12/02/2021	20.8	5180H- 5825H	5500	4.926	36.941	4.963	35.640	-0.75%	3.65%			
		3023H	5600	5.022	36.439	5.065	35.530	-0.85%	2.56%			
			5750	5.253	36.127	5.219	35.360	0.65%	2.17%			
			5800	5.296	36.863	5.270	35.300	0.49%	4.43%			
			5825	5.270	36.126	5.296	35.270	-0.49%	2.43%			

10.2 System Check

Prior to assessment, the system is verified to the \pm 10 % of the specifications at 2 450 MHz/ 5 250 MHz/ 5 600 MHz/ 5 750 MHz by using the system Check kit. (Graphic Plots Attached)

Prior the SAR assessment, the system is verified using the procedures and dipole sources as defined by IEC 62209-1 and IEC 62209-2.

The system is verified to ± 10 % of the SAR measurement on the reference dipole at the time of calibration by the calibration facility.

Input Power: 50 mW

Freq. [Mtz]	Date	Probe (S/N)	Dipole (S/N)	Liquid	Amb. Temp. [°C]	Liquid Temp. [°C]	1 W Target SAR _{1g} (SPEAG) [W/kg]	50mW Measured SAR _{1g} [W/kg]	1 W Normalized SAR _{1g} [W/kg]	Deviation [%]	Limit [%]
2 450	12/01/2021	7679	965	Head	22.5	22.4	53.3	2.89	57.8	+ 8.44	± 10
5 250	12/02/2021	7679		Head	20.9	20.8	80.6	4.14	82.8	+ 2.73	± 10
5 600	12/02/2021	7679	1107	Head	20.9	20.8	84.2	4.27	85.4	+ 1.43	± 10
5 750	12/02/2021	7679		Head	20.9	20.8	80.9	3.99	79.8	- 1.36	± 10



11. SAR TEST DATA SUMMARY

Within the Host model, the SAR test of the WLAN AUX Ant(2GHz/5GHzWLAN /BT)was excluded in accordance with FCC KDB 447498 D01 and KDB 616217.

11.1 Body SAR Measurement Results For Laptop mode

	2.4 GHz WLAN Body SAR_Main Ant																
Freque MHz	ency Ch.	Mode	Band width (₩z)	Rate	Up Limit		Power Drift (dB)	Test	Ant Config.	· · ·	Distance (mm)	Area Scan Peak SAR (W/kg)	Meas. SAR (W/kg)	Scaling Factor	Scaling Factor (Duty)	Reported SAR (W/kg)	Plot No.
2 437	6	802.11b	20	1	21	21.00	0.05	Bottom	Main	99.5	0	0.0362	0.012	1.000	1.005	0.012	1
ANSI/ IEEE C95.1 - 2005 – Safety Limit Spatial Peak Uncontrolled Exposure/ General Population								Body 1.6 W/kg (mW/g) Averaged over 1 gram									

	5 GHz WLAN Body SAR																
Freque MHz	ncy Ch.	Mode	Band width (Mtz)	Data Rate (Mbps)	Tune- Up Limit (dBm)	Meas. Power (dBm)	Power Drift (dB)	Test Position	Ant Config.		Distance	Area Scan Peak SAR (W/kg)	Meas. SAR (W/kg)	Scaling Factor	Scaling Factor (Duty)	Reported SAR (W/kg)	Plot No.
5 260	52	802.11a	20	6	21.5	21.00	0.11	Bottom	Main	98.1	0	0.344	0.163	1.122	1.019	0.186	2
5 690	138	802.11ac	80	MCS0	21.0	21.00	0.10	Bottom	Main	98.5	0	0.443	0.166	1.000	1.015	0.168	3
5 795	159	802.11n	40	MCS0	21.0	20.69	0.09	Bottom	Main	99.0	0	0.265	0.069	1.074	1.010	0.075	-
	ANSI/ IEEE C95.1 - 2005 – Safety Limit										Body						
	Spatial Peak										1.6 W/kg (mW/g)						
			Un	control	led E	xposu	re/ G	eneral Populatio	n				Averaged over 1 gram				



11.2 SAR Test Notes

General Notes:

- 1. The test data reported are the worst-case SAR values according to test procedures specified in FCC KDB Publication 616217 D04v01r02 and KDB Publication 447498 D01v06
- 2. Batteries are fully charged at the beginning of the SAR measurements.
- 3. Liquid tissue depth was at least 15.0 cm for all frequencies.
- 4. The manufacturer has confirmed that the device(s) tested have the same physical, mechanical and thermal characteristics and are within operational tolerances expected for production units.
- 5 SAR results were scaled to the maximum allowed power to demonstrate compliance per FCC KDB 447498 D01v06.
- 6.. Per FCC KDB 865664 D01v01r04, variability SAR measurement were performed when the measured SAR results for a frequency band were greater than or equal to 0.8 W/kg for 1g SAR and >2 for 10g SAR Please see Section 15 for variability analysis. the maximum tune-up tolerance limit.
- FCC KDB Publication 616217 D04v01r02 Section 4.2, SAR tests for Laptop are required for the rear side with the DUT touching the phantom with the display screen opened at an angle of 90°

WLAN Notes:

- 1.Per KDB 2482227 D01v02r02 justification for test configurations of 2.4 ^{GHz} WiFi Single transmission chain operations, the highest measured maximum output power channel for DSSS was selected for SAR measurement. SAR for OFDM modes (2.4 ^{GHz} 802.11 g/n/ac) was not required due to the maximum allowed powers and the highest reported DSSS SAR.
- 2.Per KDB 2482227 D01v02r02 justification for test configurations of 5 ^{GHz} WiFi Single transmission chain operations, the initial test configuration was selected according to the transmission mode with the highest maximum allowed powers. Other transmission mode were not investigated since the highest reported SAR for initial test configuration adjusted by the ration of maximum output powers is less than 1.2 W/kg for 1g SAR and less than 3.0 W/kg for 10 g SAR.
- 3.When the maximum reported 1g averaged SAR is \leq 0.8 W/kg, SAR testing on additional channels was not required. Otherwise, SAR for the next highest output power channel was required until the reported SAR result was \leq 1.20 W/kg or all test channels were measured.
- 4. The device was configured to transmit continuously at the required data rated, channel bandwidth and signal modulation, using the highest transmission duty factor supported by the test mode tools. The reported SAR was scaled to the 100% transmission duty factor to determine compliance. Procedures used to measure the duty factor are identical to that in the associated WLAN test reports.



12. SIMULTANEOUS SAR ANALYSIS

12.1 Simultaneous Transmission Summation.

The highest reported SAR for each exposure condition is used for SAR summation purpose.

The highest reported SAR for each exposure condition is used for SAR summation purpose. The WLAN/BT SAR testing results were used to perform transmission simultaneous analysis from SAR Test Report[HCT-SR-2112-FC006-R1],Module model: WW21B with FCC: ACJ9TGWW21B

	Simultaneous Transmission Summation Scenario									
Exposure	Distance	Band	WWAN SAR	2.4GHz WLAN Main	BT	∑ 1-g SAR				
condition	(mm)		(W/kg)	(W/kg)	(W/kg)	(W/kg)				
		UMTS Band 5	0.4	0.012	0.4	0.812				
		UMTS Band 4	0.4	0.012	0.4	0.812				
		UMTS Band 2	0.4	0.012	0.4	0.812				
		LTE Band 7	0.4	0.012	0.4	0.812				
		LTE Band 12	0.4	0.012	0.4	0.812				
		LTE Band 13	0.4	0.012	0.4	0.812				
Body	0	LTE Band 14	0.4	0.012	0.4	0.812				
Body	0	LTE Band 25	0.4	0.012	0.4	0.812				
		LTE Band 26	0.4	0.012	0.4	0.812				
		LTE Band 41	0.4	0.012	0.4	0.812				
		LTE Band 42	0.4	0.012	0.4	0.812				
		LTE Band 48	0.4	0.012	0.4	0.812				
		LTE Band 66	0.4	0.012	0.4	0.812				
		LTE Band 71	0.4	0.012	0.4	0.812				

	Simultaneous Transmission Summation Scenario										
Exposure	Distance	Band	WWAN SAR	5GHz WLAN Main	BT	∑ 1-g SAR					
condition	(mm)		(W/kg)	(W/kg)	(W/kg)	(W/kg)					
		UMTS Band 5	0.4	0.186	0.4	0.986					
		UMTS Band 4	0.4	0.186	0.4	0.986					
		UMTS Band 2	0.4	0.186	0.4	0.986					
		LTE Band 7	0.4	0.186	0.4	0.986					
		LTE Band 12	0.4	0.186	0.4	0.986					
		LTE Band 13	0.4	0.186	0.4	0.986					
Pody	0	LTE Band 14	0.4	0.186	0.4	0.986					
Body	0	LTE Band 25	0.4	0.186	0.4	0.986					
		LTE Band 26	0.4	0.186	0.4	0.986					
		LTE Band 41	0.4	0.186	0.4	0.986					
		LTE Band 42	0.4	0.186	0.4	0.986					
		LTE Band 48	0.4	0.186	0.4	0.986					
		LTE Band 66	0.4	0.186	0.4	0.986					
		LTE Band 71	0.4	0.186	0.4	0.986					





		Simulta	neous Transi	mission Summa	ation Scenario	
Exposure	Distance	Band	WWAN SAR	2.4GHz WLAN Aux	2.4GHz WLAN Main	∑ 1-g SAR
condition	(mm)		(W/kg)	(W/kg)	(W/kg)	(W/kg)
		UMTS Band 5	0.4	0.4	0.012	0.812
		UMTS Band 4	0.4	0.4	0.012	0.812
		UMTS Band 2	0.4	0.4	0.012	0.812
		LTE Band 7	0.4	0.4	0.012	0.812
		LTE Band 12	0.4	0.4	0.012	0.812
		LTE Band 13	0.4	0.4	0.012	0.812
Body	0	LTE Band 14	0.4	0.4	0.012	0.812
Бойу	0	LTE Band 25	0.4	0.4	0.012	0.812
		LTE Band 26	0.4	0.4	0.012	0.812
		LTE Band 41	0.4	0.4	0.012	0.812
		LTE Band 42	0.4	0.4	0.012	0.812
		LTE Band 48	0.4	0.4	0.012	0.812
		LTE Band 66	0.4	0.4	0.012	0.812
		LTE Band 71	0.4	0.4	0.012	0.812

	Simultaneous Transmission Summation Scenario										
Exposure	Distance	Band	WWAN SAR	5GHz WLAN Aux	5GHz WLAN Main	∑ 1-g SAR					
condition	(mm)		(W/kg)	(W/kg)	(W/kg)	(W/kg)					
		UMTS Band 5	0.4	0.4	0.186	0.986					
		UMTS Band 4	0.4	0.4	0.186	0.986					
		UMTS Band 2	0.4	0.4	0.186	0.986					
		LTE Band 7	0.4	0.4	0.186	0.986					
		LTE Band 12	0.4	0.4	0.186	0.986					
		LTE Band 13	0.4	0.4	0.186	0.986					
Body	0	LTE Band 14	0.4	0.4	0.186	0.986					
DOUY	0	LTE Band 25	0.4	0.4	0.186	0.986					
		LTE Band 26	0.4	0.4	0.186	0.986					
		LTE Band 41	0.4	0.4	0.186	0.986					
		LTE Band 42	0.4	0.4	0.186	0.986					
		LTE Band 48	0.4	0.4	0.186	0.986					
		LTE Band 66	0.4	0.4	0.186	0.986					
		LTE Band 71	0.4	0.4	0.186	0.986					

Note:

 Since antenna separation distance of Bottom side was >50mm, an estimated 1g SAR for Bottom side of 0.4 /kg was used to the simultaneous transmission SAR anaysis per FCC KDB Publication 447498D01v06.



12.2 Simultaneous Transmission Conclusion

The above numerical summed SAR Results are sufficient to determine that simultaneous transmission cases will not exceed the SAR Limit and therefore no measured volumetric simultaneous SAR summation is required per FCC KDB Publication 447498 D01v06 and the IEEE1528-2013.



13. MEASUREMENT UNCERTAINTY

The measured SAR was <1.5 W/Kg for 1g SAR and <3.75 W/Kg For 10g SAR for all frequency Bands. Therefore, per KDB Publication 865664 D01v01r04, the extended measurement uncertainty analysis per IEEE1528-2013 was not required.



14. SAR TEST EQUIPMENT

Manufacturer	Type / Model	S/N	Calib. Date	Calib.Interval	Calib.Due
SPEAG	SAM Phantom	-	N/A	N/A	N/A
HP	SAR System Control PC	-	N/A	N/A	N/A
Staubli	CS8Cspeag-TX90	F17/ 59CHA1/ C/ 01	N/A	N/A	N/A
Staubli	TX90 XIspeag	F17/ 59CHA1/ A/ 01	N/A	N/A	N/A
Staubli	Teach Pendant (Joystick)	010963	N/A	N/A	N/A
SPEAG	DAE4	1417	02/22/2021	Annual	02/22/2022
SPEAG	E-Field Probe EX3DV4	7679	09/10/2021	Annual	09/10/2022
SPEAG	Dipole D2450V2	965	06/15/2021	Annual	06/15/2022
SPEAG	Dipole D5GHzV2	1107	07/22/2021	Annual	07/22/2022
Agilent	Power Meter E4419B	MY41291386	10/06/2021	Annual	10/06/2022
Agilent	Power Meter N1911A	MY45101406	07/08/2021	Annual	07/08/2022
Agilent	Power Sensor 8481A	SG1091286	10/06/2021	Annual	10/06/2022
Agilent	Power Sensor 8481A	MY41090675	10/06/2021	Annual	10/06/2022
Agilent	Power Sensor N1921A	MY55220026	08/05/2021	Annual	08/05/2022
SPEAG	DAKS 3.5	1038	03/17/2021	Annual	03/17/2022
SPEAG	DAKS_VNA R140	0141013	04/07/2021	Annual	04/07/2022
Agilent	Signal Generator N5182A	MY47070230	05/10/2021	Annual	05/10/2022
Agilent	11636B/Power Divider	58698	02/26/2021	Annual	02/26/2022
TESTO	175-H1/Thermometer	40331915309	01/26/2021	Annual	01/26/2022
EMPOWER	RF Power Amplifier	1084	06/25/2021	Annual	06/25/2022
EMPOWER	RF Power Amplifier	1041D/C0508	06/24/2021	Annual	06/24/2022
MICRO LAB	LP Filter / LA-30N	-	10/06/2021	Annual	10/06/2022
MICRO LAB	LP Filter / LA-60N	32011	10/06/2021	Annual	10/06/2022
HP	Attenuator (3dB) 333340A	02427	09/06/2021	Annual	09/06/2022
HP	Attenuator (20dB) 8493C	09271	09/06/2021	Annual	09/17/2022
Agilent	Directional Bridge	3140A03878	05/28/2021	Annual	05/28/2022
Agilent	MXA Signal Analyzer N9020A	MY50510407	10/20/2021	Annual	10/20/2022
HP	Dual Directional Coupler	16072	10/05/2021	Annual	10/05/2022
R&S	Bluetooth CBT	100272	02/26/2021	Annual	02/26/2022

* The E-field probe was calibrated by SPEAG, by the waveguide technique procedure. Dipole Verification measurement is performed by HCT Lab. before each test. The brain/body simulating material is calibrated by HCT using the DAKS 3.5 to determine the conductivity and permittivity (dielectric constant) of the brain/body-equivalent material.



15. CONCLUSION

The SAR measurement indicates that the EUT complies with the RF radiation exposure limits of the ANSI/ IEEE C95.1 - 2005.

These measurements are taken to simulate the RF effects exposure under worst-case conditions. Precise laboratory measures were taken to assure repeatability of the tests.



16. REFERENCES

[1] Federal Communications Commission, ET Docket 93-62, Guidelines for Evaluating the Environmental Effects of Radio frequency Radiation, Aug. 1996.

[2] ANSI/IEEE C95.1 - 2005, American National Standard safety levels with respect to human exposure to radio frequency electromagnetic fields, 300 kHz to 300 GHz, New York: IEEE, Sept. 1992

[3] ANSI/IEEE C 95.1 - 2005, American National Standard safety levels with respect to human exposure to radio frequency electromagnetic fields, 3 kHz to 300 GHz, New York: IEEE, 2006

[4 ANSI/IEEE C95.3 - 2002, IEEE Recommended Practice for the Measurement of Potentially Hazardous Electromagnetic Fields - RF and Microwave, New York: December 2002.

[5] IEEE Standards Coordinating Committee 34 – IEEE Std. 1528-2013, IEEE Recommended Practice or Determining the Peak Spatial-Average Specific Absorption Rate (SAR) in the Human Body from Wireless Communications Devices

[6] NCRP, National Council on Radiation Protection and Measurements, Biological Effects and Exposure Criteria for Radio Frequency Electromagnetic Fields, NCRP Report No. 86, 1986. Reprinted Feb. 1995.

[7] T. Schmid, O. Egger, N. Kuster, Automated E-field scanning system for dosimetric assessments, IEEE Transaction on Microwave Theory and Techniques, vol. 44, Jan. 1996, pp. 105-113.

[8] K. Pokovic, T. Schmid, N. Kuster, Robust setup for precise calibration of E-field probes in tissue simulating liquids at mobile communications frequencies, ICECOM97, Oct. 1997, pp. 120-124.

[9] K. Pokovic, T. Schmid, and N. Kuster, E-field Probe with improved isotropy in brain simulating liquids, Proceedings of the ELMAR, Zadar, Croatia, June 23-25, 1996, pp. 172-175.

[10] Schmid & Partner Engineering AG, Application Note: Data Storage and Evaluation, June 1998, p2.

[11] V. Hombach, K. Meier, M. Burkhardt, E. Kuhn, N. Kuster, The Dependence of EM Energy Absorption upon Human Head Modeling at 900 ^{Mtz}, IEEE Transaction on Microwave Theory and Techniques, vol. 44 no. 10, Oct. 1996, pp. 1865-1873.

[12] N. Kuster and Q. Balzano, Energy absorption mechanism by biological bodies in the near field of dipole antennas above 300 Mz, IEEE Transaction on Vehicular Technology, vol. 41, no. 1, Feb. 1992, pp. 17-23.

[13] G. Hartsgrove, A. Kraszewski, A. Surowiec, Simulated Biological Materials for Electromagnetic Radiation Absorption Studies, University of Ottawa, Bioelectro magnetics, Canada: 1987, pp. 29-36.

[14] Q. Balzano, O. Garay, T. Manning Jr., Electromagnetic Energy Exposure of Simulated Users of Portable Cellular Telephones, IEEE Transactions on Vehicular Technology, vol. 44, no.3, Aug. 1995.

[15] W. Gander, Computer mathematick, Birkhaeuser, Basel, 1992.

[16] W.H. Press, S.A. Teukolsky, W.T. Vetterling, and B.P. Flannery, Numerical Recepies in C, The Art of Scientific Computing, Second edition, Cambridge University Press, 1992.

[17] N. Kuster, R. Kastle, T. Schmid, Dosimetric evaluation of mobile communications equipment with known precision, IEEE Transaction on Communications, vol. E80-B, no. 5, May 1997, pp. 645-652.

[18] CENELEC CLC/SC111B, European Prestandard (prENV 50166-2), Human Exposure to Electromagnetic Fields High-frequency: 10 kHz-300 GHz, Jan. 1995.

[19] Prof. Dr. Niels Kuster, ETH, EidgenØssische Technische Hoschschule Zorich, Dosimetric Evaluation of the Cellular Phone.

[20] IEC 62209-1, Human exposure to radio frequency fields from hand-held and body-mounted wireless communication devices – Human models, instrumentation and procedures – Part 1:Procedure to determine the specific absorption rate (SAR) for hand-held devices used in close proximity to the ear (frequency range of 300 ^{Mtz} to 3 GHz), July. 2016.



[21] IEC 62209-2, Human exposure to radio frequency fields from hand-held and body-mounted wireless communication devices – Human models, instrumentation, and procedures – Part 2: Procedure to determine the specific absorption rate (SAR) for wireless communication devices used in close proximity to the human body (frequency range of 30 ^{Mtz} to 6 GHz) Mar. 2010.

[22] Industry Canada RSS-102 Radio Frequency Exposure Compliance of Radio Communication Apparatus (All Frequency Band) Issue 5, March 2015.

[23] Health Canada Safety Code 6 Limits of Human Exposure to Radio Frequency Electromagnetic Fields in the Frequency Rage from 3 kHz – 300 GHz, 2009

[24] FCC SAR Test procedures for 2G-3G Devices, Mobile Hotspot and UMPC Device KDB 941225 D01.

[25] SAR Measurement Guidance for IEEE 802.11 transmitters, KDB 248227 D01v02r02

[26] SAR Evaluation of Handsets with Multiple Transmitters and Antennas KDB 648474 D03, D04.

[27] SAR Evaluation for Laptop, Notebook, Netbook and Tablet computers KDB 616217 D04.

[28] SAR Measurement and Reporting Requirements for 100 ^{Mlz} − 6 GHz, KDB 865664 D01, D02.

[29] FCC General RF Exposure Guidance and SAR procedures for Dongles, KDB 447498 D01,D02.



Appendix A. DUT Ant. Information & SETUP PHOTO

Please refer to test DUT Ant. Information & setup photo file no. as follows:

No.	Description
0	HCT-SR-2112-FC005-P



Appendix B. – SAR Test Plots



Test Laboratory:	HCT CO., LTD
EUT Type:	Radio Module
Liquid Temperature:	22.4 °C
Ambient Temperature:	22.5 °C
Test Date:	12/01/2021
Plot No.:	1

Communication System: UID 0, 2450MHz FCC (0); Frequency: 2437 MHz;Duty Cycle: 1:1 Medium parameters used (interpolated): f = 2437 MHz; σ = 1.843 S/m; ϵ_r = 39.223; ρ = 1000 kg/m³ Phantom section: Flat Section


DASY5 Configuration:

- Probe: EX3DV4 SN7679; ConvF(7.96, 7.96, 7.96) @ 2437 MHz; Calibrated: 2021-09-10
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn1417; Calibrated: 2021-02-22
- Phantom: SAM with CRP v5.0(Front); Type: QD000P40CD; Serial: TP:xxxx
- Measurement SW: DASY52, Version 52.10 (4); SEMCAD X Version 14.6.14 (7483)

802.11b Body Bottom 1Mbps 6ch/Area Scan (9x13x1): Measurement grid: dx=12mm, dy=12mm Maximum value of SAR (measured) = 0.0275 W/kg

802.11b Body Bottom 1Mbps 6ch/Zoom Scan (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 0.8830 V/m; Power Drift = 0.05 dB Peak SAR (extrapolated) = 0.0340 W/kg SAR(1 g) = 0.012 W/kg; SAR(10 g) = 0.00415 W/kg Maximum value of SAR (measured) = 0.0206 W/kg



0 dB = 0.0206 W/kg = -16.86 dBW/kg



Test Laboratory:	HCT CO., LTD
EUT Type:	Radio Module
Liquid Temperature:	20.8 °C
Ambient Temperature:	20.9 °C
Test Date:	12/02/2021
Plot No.:	2

Communication System: UID 0, WIFI 5GHz (0); Frequency: 5260 MHz;Duty Cycle: 1:1 Medium parameters used: f = 5260 MHz; σ = 4.712 S/m; ϵ_r = 36.739; ρ = 1000 kg/m³ Phantom section: Flat Section


DASY5 Configuration:

- Probe: EX3DV4 SN7679; ConvF(5.55, 5.55, 5.55) @ 5260 MHz; Calibrated: 2021-09-10
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn1417; Calibrated: 2021-02-22
- Phantom: SAM with CRP v5.0(Right)_2014_03_05; Type: QD000P40CD; Serial: TP:xxxx
- Measurement SW: DASY52, Version 52.10 (4); SEMCAD X Version 14.6.14 (7483)

802.11a Body Bottom 6Mbps 52ch/Area Scan (11x14x1): Measurement grid: dx=10mm, dy=10mm Maximum value of SAR (measured) = 0.344 W/kg

802.11a Body Bottom 6Mbps 52ch/Zoom Scan (7x7x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm

Reference Value = 7.023 V/m; Power Drift = 0.11 dB Peak SAR (extrapolated) = 0.620 W/kg SAR(1 g) = 0.163 W/kg; SAR(10 g) = 0.059 W/kg Maximum value of SAR (measured) = 0.380 W/kg



⁰ dB = 0.380 W/kg = -4.20 dBW/kg



Test Laboratory:	HCT CO., LTD
EUT Type:	Radio Module
Liquid Temperature:	20.8 °C
Ambient Temperature:	20.9 °C
Test Date:	12/02/2021
Plot No.:	3

Communication System: UID 0, WIFI 5GHz (0); Frequency: 5690 MHz;Duty Cycle: 1:1 Medium parameters used: f = 5690 MHz; σ = 5.197 S/m; ϵ_r = 35.946; ρ = 1000 kg/m³ Phantom section: Flat Section


DASY5 Configuration:

- Probe: EX3DV4 SN7679; ConvF(5.05, 5.05, 5.05) @ 5690 MHz; Calibrated: 2021-09-10
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn1417; Calibrated: 2021-02-22
- Phantom: SAM with CRP v5.0(Right)_2014_03_05; Type: QD000P40CD; Serial: TP:xxxx
- Measurement SW: DASY52, Version 52.10 (4); SEMCAD X Version 14.6.14 (7483)

802.11ac80 Body Bottom VHT0 138ch/Area Scan (11x14x1): Measurement grid: dx=10mm, dy=10mm Maximum value of SAR (measured) = 0.375 W/kg

802.11ac80 Body Bottom VHT0 138ch/Zoom Scan (7x7x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm

Reference Value = 5.901 V/m; Power Drift = 0.10 dB Peak SAR (extrapolated) = 0.903 W/kg SAR(1 g) = 0.166 W/kg; SAR(10 g) = 0.055 W/kg Maximum value of SAR (measured) = 0.411 W/kg



⁰ dB = 0.411 W/kg = -3.86 dBW/kg



Appendix C. – Dipole Verification Plots



■ Verification Data (2 450 Mb Body)

Test Laboratory:HCT CO., LTDInput Power0.05 WLiquid Temp:22.4 °CTest Date:12/01/2021

DUT: Dipole 2450 MHz D2450V2; Type: D2450V2;

Communication System: UID 0, CW (0); Frequency: 2450 MHz;Duty Cycle: 1:1 Medium parameters used: f = 2450 MHz; σ = 1.856 S/m; ϵ_r = 39.161; ρ = 1000 kg/m³ Phantom section: Flat Section


DASY5 Configuration:

- Probe: EX3DV4 SN7679; ConvF(7.96, 7.96, 7.96) @ 2450 MHz; Calibrated: 2021-09-10
- Sensor-Surface: 3mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn1417; Calibrated: 2021-02-22
- Phantom: SAM with CRP v5.0(Front); Type: QD000P40CD;
- Measurement SW: DASY52, Version 52.10 (4);
- SEMCAD X Version 14.6.14 (7483)

Dipole/2450MHz Head Verification/Area Scan (6x7x1): Measurement grid: dx=12mm, dy=12mm Maximum value of SAR (measured) = 3.18 W/kg

Dipole/2450MHz Head Verification/Zoom Scan (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 45.98 V/m; Power Drift = 0.01 dB Peak SAR (extrapolated) = 6.47 W/kg SAR(1 g) = 2.89 W/kg; SAR(10 g) = 1.29 W/kg Maximum value of SAR (measured) = 3.83 W/kg



0 dB = 3.83 W/kg = 5.83 dBW/kg



■ Verification Data (5 250 Mz Body)

Test Laboratory:	HCT CO., LTD
Input Power	0.05 W
Liquid Temp:	20.8 ℃
Test Date:	12/02/2021

DUT: Dipole D5GHzV2; Type: D5GHzV2;

Communication System: UID 0, CW (0); Frequency: 5250 MHz;Duty Cycle: 1:1 Medium parameters used: f = 5250 MHz; σ = 4.758 S/m; ϵ_r = 36.877; ρ = 1000 kg/m³ Phantom section: Flat Section


DASY5 Configuration:

- Probe: EX3DV4 SN7679; ConvF(5.55, 5.55, 5.55) @ 5250 MHz; Calibrated: 2021-09-10
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn1417; Calibrated: 2021-02-22
- Phantom: SAM with CRP v5.0(Right)_2014_03_05; Type: QD000P40CD;
- Measurement SW: DASY52, Version 52.10 (4);
- SEMCAD X Version 14.6.14 (7483)

Dipole/5250MHz Head Verification/Area Scan (7x7x1): Measurement grid: dx=10mm, dy=10mm Maximum value of SAR (measured) = 9.41 W/kg

Dipole/5250MHz Head Verification/Zoom Scan (7x7x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm

Reference Value = 49.37 V/m; Power Drift = -0.09 dB Peak SAR (extrapolated) = 16.5 W/kg SAR(1 g) = 4.14 W/kg; SAR(10 g) = 1.22 W/kg Maximum value of SAR (measured) = 10.4 W/kg



0 dB = 10.4 W/kg = 10.17 dBW/kg





Verification Data (5 600 Mb Body)

Test Laboratory:	HCT CO., LTD
Input Power	0.05 W
Liquid Temp:	20.8 ℃
Test Date:	12/02/2021

DUT: Dipole D5GHzV2; Type: D5GHzV2;

Communication System: UID 0, CW (0); Frequency: 5600 MHz;Duty Cycle: 1:1 Medium parameters used: f = 5600 MHz; σ = 5.022 S/m; ϵ_r = 36.439; ρ = 1000 kg/m³ Phantom section: Flat Section


DASY5 Configuration:

- Probe: EX3DV4 SN7679; ConvF(4.95, 4.95, 4.95) @ 5600 MHz; Calibrated: 2021-09-10
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn1417; Calibrated: 2021-02-22
- Phantom: SAM with CRP v5.0(Right)_2014_03_05; Type: QD000P40CD;
- Measurement SW: DASY52, Version 52.10 (4); SEMCAD X Version 14.6.14 (7483)

Dipole/5600MHz Head Verification/Area Scan (7x7x1): Measurement grid: dx=10mm, dy=10mm Maximum value of SAR (measured) = 9.83 W/kg

Dipole/5600MHz Head Verification/Zoom Scan (7x7x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm

Reference Value = 48.98 V/m; Power Drift = -0.04 dB Peak SAR (extrapolated) = 18.0 W/kg SAR(1 g) = 4.27 W/kg; SAR(10 g) = 1.25 W/kg Maximum value of SAR (measured) = 11.0 W/kg



0 dB = 11.0 W/kg = 10.41 dBW/kg



Verification Data (5 750 Mt Body)

Test Laboratory:	HCT CO., LTD
Input Power	0.05 W
Liquid Temp:	20.8 ℃
Test Date:	12/02/2021

DUT: Dipole D5GHzV2; Type: D5GHzV2;

Communication System: UID 0, CW (0); Frequency: 5750 MHz;Duty Cycle: 1:1 Medium parameters used: f = 5750 MHz; σ = 5.253 S/m; ϵ_r = 36.127; ρ = 1000 kg/m³ Phantom section: Flat Section


DASY5 Configuration:

- Probe: EX3DV4 SN7679; ConvF(5.05, 5.05, 5.05) @ 5750 MHz; Calibrated: 2021-09-10
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn1417; Calibrated: 2021-02-22
- Phantom: SAM with CRP v5.0(Right)_2014_03_05; Type: QD000P40CD;
- Measurement SW: DASY52, Version 52.10 (4); SEMCAD X Version 14.6.14 (7483)

Dipole/5750MHz Head Verification/Area Scan (7x7x1): Measurement grid: dx=10mm, dy=10mm Maximum value of SAR (measured) = 9.09 W/kg

Dipole/5750MHz Head Verification/Zoom Scan (7x7x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm

Reference Value = 46.13 V/m; Power Drift = -0.01 dB Peak SAR (extrapolated) = 17.2 W/kg SAR(1 g) = 3.99 W/kg; SAR(10 g) = 1.17 W/kg Maximum value of SAR (measured) = 10.2 W/kg



0 dB = 10.2 W/kg = 10.09 dBW/kg



Appendix D. – SAR Tissue Characterization

The brain and muscle mixtures consist of a viscous gel using hydrox-ethyl cellulose (HEC) gelling agent and saline solution (see Table 3.1). Preservation with a bacteriacide is added and visual inspection is made to make sure air bubbles are not trapped during the mixing process. The mixture is calibrated to obtain proper dielectric constant (permittivity) and conductivity of the desired tissue. The mixture characterizations used for the brain and muscle tissue simulating liquids are according to the data by C. Gabriel and G. Harts grove.

Ingredients		Frequency (Mz)
(% by weight)	2 450 – 2 700	3 500 - 5 800
Tissue Type	Body	Body
Water	73.2	78.66
Salt (NaCl)	0.1	0.0
Sugar	0.0	0.0
HEC	0.0	0.0
Bactericide	0.0	0.0
Triton X-100	0.0	10.67
DGBE	26.7	0.0
Diethylene glycol hexyl ether	-	-

Salt:	99 % Pure Sodium ChlorideSugar:98 % Pure SucroseDe-ionized, 16M resistivityHEC:Hydroxyethyl Cellulose99 % Di(ethylene glycol) butyl ether,[2-(2-butoxyethoxy) ethanol]			
Water:	De-ionized, 16M resistivity	HEC:	Hydroxyethyl Cellulose	
DGBE:	99 % Di(ethylene glycol) butyl eth	er,[2-(2-butc	oxyethoxy) ethanol]	
Triton X-100(ultra-pure):	Polyethylene glycol mono[4-(1,1,3	3,3-tetrameth	nylbutyl)phenyl] ether	
Composition of the Tissue Equivalent Matter				

Appendix E. – SAR system validation

Per the IEEE1528-2013.,SAR system validation status should be document to confirm measurement accuracy. The SAR systems (including SAR probes, system components and software versions) used for this device were validated against its performance specifications prior to the SAR measurements. Reference dipoles were used with the required tissue- equivalent media for system validation, according to the procedures outlined in the IEEE1528-2013. Since SAR probe calibrations are frequency dependent, each probe calibration point was validated at a frequency within the valid frequency range of the probe calibration point, using the system that normally operates with the probe for routine SAR measurements and according to the required tissue-equivalent media.

A tabulated summary of the system validation status including the validation date(s), measurement frequencies, SAR probes and tissue dielectric parameters has been included.

	Probe	Prot	ре			Dielectric	Parameters	CW	/ Validati	on	Modula	ation Va	lidation
Probe	Туре	Calibratio Point		Dipole	Date	Measured Permittivity	Measured Conductivity	Sensitivit y		Probe Isotropy	MOD. Type	Duty Factor	PAR
7679	EX3DV4	Head	2450	965	2021-11-22	39.4	1.81	PASS	PASS	PASS	OFDM	N/A	PASS
7679	EX3DV4	Head	5250	1107	2021-10-22	35.6	4.71	PASS	PASS	PASS	OFDM	N/A	PASS
7679	EX3DV4	Head	5600	1107	2021-10-22	35.3	5.04	PASS	PASS	PASS	OFDM	N/A	PASS
7679	EX3DV4	Head	5750	1107	2021-10-22	35.8	5.25	PASS	PASS	PASS	OFDM	N/A	PASS

SAR System Validation Summary 1g

Note;

All measurement were performed using probes calibrated for CW signal only. Modulations in the table above represent test configurations for which the measurement system has been validated the IEEE1528-2013. SAR system were validated for modulated signals with a periodic duty cycle, such as GMSK, or with a high peak to average ratio (>5 dB), such as OFDM according to the IEEE1528-2013.



Appendix F. – Probe Calibration Data



Calibration Laboratory of Schmid & Partner Engineering AG Zoughausstrasse 43, 8004 Zurich, Switzerland



Schweizerischer Kalibrierdienst Service suisse d'étalonnage Servizio svizzoro di taratura Swiss Calibration Service

Accreditation No.: SCS 0108

s

C

S

Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Certificate No: EX3-7679_Sep21 HCT (Dymstec) Client CALIBRATION CERTIFICATE EX3DV4 - SN:7679 Object QA CAL-01.v9, QA CAL-14.v6, QA CAL-23.v5, QA CAL-25.v7 Calibration procedure(s) Calibration procedure for dosimetric E-field probes September 10, 2021 Calibration date: This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate. All calibrations have been conducted in the closed laboratory facility: environment temperature (22 ± 3)*C and humidity < 70% Calibration Equipment used (M&TE critical for calibration) Primary Standards 1D Cal Date (Certificate No.) Scheduled Calibration 09-Apr-21 (No. 217-03291/03292) Power meter NRP SN: 104778 Apr-21 Power sensor NRP-291 SN: 103244 09-Apr-21 (No. 217-03291) Apr-21 Power sensor NRP-Z91 SN: 103245 09-Apr-21 (No. 217-03292) Apr-21 SN: CC2552 (20x) 09-Apr-21 (No. 217-03343) Reference 20 dB Attenuator Apr-21 DAE4 SN: 660 23-Dec-20 (No. DAE4-660_Dec20) Dec-21 30-Dec-20 (No. ES3-3013_Dec20) Reference Probe ES30V2 SN: 3013 Dec-21 Secondary Standards ID. Check Date (In house) Scheduled Check Power meter E4419B SN: GB41293874 06-Apr-16 (in house check Jun-20) in house check: Jun-22 Power sensor E4412A SN: MY41498087 06-Apr-16 (in house check Jun-20) In house check; Jun-22 Power sensor E4412A SN: 000110210 06-Apr-16 (in house check Jun-20) In house check: Jun-22 RF generator HP 8848C 04-Aug-69 (in house check Jun-20) SN: US3642U01700 In house check: Jun-22 Network Analyzer E8358A SN: US41080477 31-Mar-14 (in house check Oct-20) In house check: Oct-21 Function Name Signatur Jeton Kastrati Calibrated by: Laboratory Technician Approved by: Katja Pokovic Technical Manager EC.L Issued: September 14, 2021 This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: EX3-7679_Sep21

Page 1 of 22

겸

재

담당자 확인자 이니 / Ut lite (3/ 195226 1-11, 0,05 121, 29, 55



Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland



Schweizerischer Kalibrierdienst Service suisse d'étalonnage C Servizio svizzero di taratura S

Swiss Calibration Service

Accreditation No.: SCS 0108

S

Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilataral Agreement for the recognition of calibration certificates

Gloceany

Glossaly.		
TSL	tissue simulating liquid	
NORMx, y, z	sensitivity in free space	
ConvF	sensitivity in TSL / NORMx,y,z	
DCP	diade compression point	
CF	crest factor (1/duty_cycle) of the RF signal	
A, B, C, D	modulation dependent linearization parameters	
Polarization ϕ	o rotation around probe axis	
Polarization &	3 rotation around an axis that is in the plane normal to probe axis (at measurement center),	
	i.e., 8 = 0 is normal to probe axis	
Connector Angle	information used in DASY system to align probe sensor X to the robot coordinate system	

Calibration is Performed According to the Following Standards:

- a) IEC/IEEE 62209-1528, "Measurement Procedure For The Assessment Of Specific Absorption Rate Of Human Exposure To Radio Frequency Fields From Hand-Held And Body-Worn Wireless Communication Devices -Part 1528: Human Models, Instrumentation And Procedures (Frequency Range of 4 MHz to 10 GHz)*, October 2020
- b) KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

Methods Applied and Interpretation of Parameters:

- NORMx, y, z: Assessed for E-field polarization 9 = 0 (f < 900 MHz in TEM-cell; f > 1800 MHz: R22 waveguide). NORMx,y,z are only intermediate values, i.e., the uncertainties of NORMx,y,z does not affect the E²-field uncertainty inside TSL (see below ConvF).
- NORM(f)x,y,z = NORMx,y,z * frequency_response (see Frequency Response Chart). This linearization is implemented in DASY4 software versions later than 4.2. The uncertainty of the frequency response is included In the stated uncertainty of ConvF.
- DCPx,y,z: DCP are numerical linearization parameters assessed based on the data of power sweep with CW signal (no uncertainty required). DCP does not depend on frequency nor media.
- PAR: PAR is the Peak to Average Ratio that is not calibrated but determined based on the signal characteristics
- Ax,y,z; Bx,y,z; Cx,y,z; Dx,y,z; VRx,y,z: A, B, C, D are numerical linearization parameters assessed based on the data of power sweep for specific modulation signal. The parameters do not depend on frequency nor media. VR is the maximum calibration range expressed in RMS voltage across the diode.
- ConvF and Boundary Effect Parameters: Assessed in flat phantom using E-field (or Temperature Transfer Standard for f ≤ 800 MHz) and inside waveguide using analytical field distributions based on power measurements for f > 800 MHz. The same setups are used for assessment of the parameters applied for boundary compensation (alpha, depth) of which typical uncertainty values are given. These parameters are used in DASY4 software to improve probe accuracy close to the boundary. The sensitivity in TSL corresponds to NORMx, y, z * ConvF whereby the uncertainty corresponds to that given for ConvF. A frequency dependent ConvF is used in DASY version 4.4 and higher which allows extending the validity from ± 50 MHz to ± 100 MHz.
- Spherical isotropy (3D deviation from isotropy): in a field of low gradients realized using a flat phantom exposed by a patch antenna.
- Sensor Offset: The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No tolerance required.
- Connector Angle: The angle is assessed using the information gained by determining the NORMx (no uncertainty required).

Certificate No: EX3-7679_Sep21

Page 2 of 22

on used in DASY sys



EX3DV4 - SN:7679

September 10, 2021

DASY/EASY - Parameters of Probe: EX3DV4 - SN:7679

Basic Calibration Parameters

	Sensor X	Sensor Y	Sensor Z	Unc (k=2)
Norm (µV/(V/m) ²) ^A	0.65	0.49	0.63	± 10.1 %
DCP (mV) ⁶	105.9	103.2	99.7	

Calibration Results for Modulation Response

UID	Communication System Name		A dB	B dBõV	C	D dB	VR mV	Max dev.	Max Unc ^E (k=2)
0	CW	X	0.00	0.00	1.00	0.00	141.2	± 3.5 %	± 4.7 %
	570	Y	0.00	0.00	1.00		146.5	1	
		Z	0.00	0.00	1.00		138.9		
10352-	Pulse Waveform (200Hz, 10%)	X	1.43	60.13	6.05	10.00	60.0	±2.8 %	± 9.6 %
AAA	2 2 3	Y	1.61	61.06	6.62		60.0		
		Z	1.43	60.30	6.32	·	60.0	Anna anna anna	
10353-	Pulse Waveform (200Hz, 20%)	X	0.82	60.00	4.86	6:99	80.0	±2.4 %	± 9.6 %
AAA		Y	22.00	78.00	11.00		80.0	1	
		Z	44.00	80.00	11.00		80.0	in the second	THE OVER 1
10354-	Pulse Waveform (200Hz, 40%)	X	0.04	124.71	0.20	3.98	95.0	±2.5 %	± 9.6.%
AAA	A REAL POINT AND A REAL	Y	8.00	70.00	7.00		95.0		1000000
	and the second second second second	Z	0.02	120.23	1.90		95.0	1 contraction of the second	
10355-	Pulse Waveform (200Hz, 60%)	X	8.65	159.63	2.67	2.22	120.0	± 1.5 %	± 9.6 %
AAA	A DESERVED AND AND A DESCRIPTION OF A DESCRIPTION	Y	10.53	157.49	11.62	1 100 2012	120.0		
		Z	0.07	157.61	0.00		120.0		
10387-	QPSK Waveform, 1 MHz	X	0.54	61.55	10.71	1.00	150.0	± 4.2 %	± 9.6 %
AAA	Contract Contract Contract Contract	Y	0.60	64.34	12.88	0.022	150.0	1000000	
		Z	0.71	63.50	12.19	1	150.0	1	
10388-	QPSK Waveform, 10 MHz	X	1.25	63.84	12.80	0.00	150.0	±1.3 %	±9.6%
AAA.	CALIFORNIA DA	Y	1.40	66.19	14.18	1.3325	150.0	1-00000	12000
		Z	1.42	64.67	13.67	1	150.0	1	
10396-	64-QAM Waveform, 100 kHz	X	1.55	63.00	15.00	3.01	150.0	=1.2%	± 9.6 %
AAA	A CANCELLE CONTRACTOR CONTRACTOR	Y	1.70	64.47	15.77	1.152	150.0	1	1212.0
		Z	1.58	62.85	15.42	-	150.0	1	_
10399-	64-QAM Waveform, 40 MHz	X	2.74	65.26	14.43	0.00	150.0	± 1.7 %	± 9.6 %
AAA		Y	2.87	66.40	15.17		150.0		
		Z	2.89	65.57	14.82		150.0		
10414-	WLAN CCDF, 64-QAM, 40MHz	X	3.97	65.95	15.18	0.00	150.0	± 3.3 %	±9.6 %
AAA		Y	3.85	65.99	15.29		150.0		
		Z	4.17	66.04	15.46		150.0		

Note: For details on UID parameters see Appendix

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

⁶ The uncertainties of Norm X,Y,Z do not affect the E¹-field uncertainty inside TSL (see Page 5). ⁸ Numerical linearization parameter: uncertainty not required. ⁹ Uncertainty's determined using the max. deviation from linear response applying rectangular distribution and is expressed for the square of the field value.

Certificate No: EX3-7679_Sep21

Page 3 of 22





EX3DV4- SN:7679

September 10, 2021

DASY/EASY - Parameters of Probe: EX3DV4 - SN:7679

Sensor Model Parameters

	C1 fF	C2 fF	α V ⁻¹	T1 ms.V ^{-z}	T2 ms.V ⁻¹	T3 ms	T4 V-3	T5 V ⁻¹	T6
Х	11.4	82.32	33.03	3.55	0.00	4.90	0.25	0.00	1.00
Y	10.0	72.14	33.08	3.60	0.00	4.90	0.50	0.00	1.00
Ż	13.6	100.64	34.72	1.47	0.00	4.90	0.00	0.01	1.01

Other Probe Parameters

Sensor Arrangement	Triangular
Connector Angle (*)	-128.3
Mechanical Surface Detection Mode	enabled
Optical Surface Detection Mode	disabled
Probe Overall Length	337 mm
Probe Body Diameter	10 mm
Tip Length	9 mm
Tip Diameter	2.5 mm
Probe Tip to Sensor X Calibration Point	1 mm
Probe Tip to Sensor Y Calibration Point	1 mm
Probe Tip to Sensor Z Calibration Point	1 mm
Recommended Measurement Distance from Surface	1.4 mm

Note: Measurement distance from surface can be increased to 3-4 mm for an Area Scan job.

Certificate No: EX3-7679_Sep21

Page 4 of 22





EX3DV4-- SN:7679

September 10, 2021

DASY/EASY - Parameters of Probe: EX3DV4 - SN:7679

f (MHz) ^c	Relative Permittivity	Conductivity (S/m)	ConvF X	ConvF Y	ConvF Z	Alpha ^G	Depth [®] (mm)	Unc (k=2)
750	41.9	0.89	9.96	9.96	9.96	0.58	0.80	± 12.0 %
835	41.5	0.90	9.79	9,79	9.79	0.54	0.80	± 12.0 %
900	41.5	0.97	9.59	9.59	9.59	0.45	0.90	± 12.0 %
1750	40.1	1.37	8.94	8.94	8.94	0.37	0.86	± 12.0 %
1900	40.0	1.40	8.63	8.63	8.63	0.34	0.86	± 12.0 %
2300	39.5	1.67	8.27	8.27	8.27	0.31	0.90	± 12.0 %
2450	39.2	1.80	7.96	7,96	7.96	0.38	0.90	± 12.0 %
2600	39.0	1.96	7.87	7.87	7.87	0.37	0.90	± 12.0 %
5250	35.9	4.71	5.55	5.55	5.55	0.40	1.80	± 13.1 %
5600	35.5	5.07	4.95	4.95	4.95	0.40	1.80	± 13.1 %
5750	35.4	5.22	5.05	5.05	5.05	0,40	1.80	± 13.1 %

Calibration Parameter Determined in Head Tissue Simulating Media

⁶ Frequency validity above 300 MHz of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to ± 50 MHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is ± 10, 25, 40, 50 and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Validity of ConvF assessed at 8 MHz is 4-9 MHz, and ConvF assessed at 13 MHz is 9-10 MHz for Foquency validity of convF assessed at 8 MHz is 4-9 MHz, and ConvF assessed at 13 MHz is 9-10 MHz. So ConvF assessed at 0.5 GHz frequency validity of issue parameters (a and o) can be relaxed to ± 10% if liquid compensation formula is applied to measured SAR values. At frequencies below 3 GHz, the validity of issue parameters (a and o) is restricted to ± 5%. The uncertainty is the RSS of the ConvF uncertainty for indicated target lissue parameters.

Certificate No: EX3-7679_Sep21


Page 5 of 22



EX3DV4-- SN:7679

September 10, 2021

Frequency Response of E-Field (TEM-Cell:ifi110 EXX, Waveguide: R22)





Certificate No: EX3-7679_Sep21

Page 6 of 22

September 10, 2021

EX3DV4- SN:7679

Error [d5]

HCT

f=600 MHz,TEM f=1800 MHz,R22 . 8 • . . Tut Tot 0.5 0.0 278 -0.5 -150 150 -100 100 Rol ["] 100 MHz TROO MHO 2500 MHz 600 MHz Uncertainty of Axial Isotropy Assessment: ± 0.5% (k=2)


Certificate No: EX3-7679_Sep21

Page 7 of 22



EX3DV4- 5N:7679

September 10, 2021



Uncertainty of Linearity Assessment: ± 0.6% (k=2)

Certificate No: EX3-7679_Sep21

Page 8 of 22






HCT

September 10, 2021

Conversion Factor Assessment





EX3DV4-- SN:7679

September 10, 2021

UID	Rev	odulation Calibration Parameters Communication System Name	Group	PAR (dB)	Unc* (k=2)
0	-	CW	CW	0.00	± 4.7 %
10010	CAA	SAR Validation (Square, 100ms, 10ms)	Test	10.00	± 9.6 %
10011	CAB	UMTS-FDD (WCOMA)	WCDMA	2.91	± 9.6 %
10012	CAB	IEEE 802.11b WIFI 2.4 GHz (DSSS, 1 Mbps)	WLAN	1.87	± 9.6 %
10013	CAB	IEEE 802.11g WIFI 2.4 GHz (DSSS-OFDM, 6 Mbps)	WLAN	9.46	±9.69
10021	DAC	GSM-FDD (TDMA, GMSK)	GSM	9.39	±9.69
10023	DAC	GPRS-FDD (TDMA, GMSK, TN 0)	GSM	9.57	± 9.6 %
10024	DAC	GPRS-FDD (TDMA, GMSK, TN 0-1)	GSM	6.56	± 9.6.9
10025	DAC	EDGE-FDD (TDMA, 8PSK, TN 0)	GSM	12.62	± 9.6 %
10026	DAC	EDGE-FDD (TDMA, BPSK, TN 0-1)	GSM	9,55	±9.6 %
10027	DAC	GPRS-FDD (TDMA, GMSK, TN 0-1-2)	GSM	4.80	±9.6 %
10028	DAC	GPRS-FDD (TDMA, GMSK, TN 0-1-2-3)	GSM	3.55	±9.63
10029	DAC	EDGE-FDD (TDMA, 8PSK, TN 0-1-2)	GSM	7.78	±9.63
10030	CAA	IEEE 802.15.1 Bluetooth (GFSK, DH1)	Bluetooth	5.30	± 9.6 %
10031	CAA	IEEE 802.15.1 Bluetooth (GFSK, DH3)	Bluetooth	1.87	± 9.6 %
10032	CAA	IEEE 802.15.1 Bluetooth (GFSK, DH5)	Bluetooth	1.16	± 9.6.9
10033	CAA	IEEE 802.15.1 Bluetooth (PI/4-DQPSK, DH1)	Bluetooth	7.74	± 9.6 %
10034	CAA	IEEE 802.15.1 Bluetooth (PI/4-DQPSK, DH3)	Bluetooth	4.53	± 9.6 %
10035	CAA	IEEE 802.15.1 Bluetooth (PI/4-DQPSK, DH5)	Bluetooth	3.83	± 9.6 %
10036	CAA	IEEE 802.15.1 Bluetooth (8-DPSK, DH1)	Bluetooth	8.01	± 9.6 %
10037	CAA	IEEE 802.15.1 Bluetooth (8-DPSK, DH3)	Bluetooth	4,77	± 9.6 %
10038	CAA	IEEE 802.15.1 Bluetooth (8-DPSK, DH5)	Bluetooth	4,10	± 9.6 %
10039	CAB	CDMA2000 (1xRTT, RC1)	CDMA2000	4.57	± 9,6 9
10042	CAB	IS-54 / IS-136 FDD (TDMA/FDM, PI/4-DQPSK, Haifrate)	AMPS	7.78	± 9,6 %
10044	CAA	IS-91/EIA/TIA-553 FDD (FDMA, FM)	AMPS	0.00	± 9.6 %
10048	CAA	DECT (TDD, TDMA/FDM, GFSK, Full Slot, 24)	DECT	13.80	± 9.6 %
10049	CAA	DECT (TDD, TDMA/FDM, GFSK, Double Slot, 12)	DECT	10.79	± 9.6 %
10056	CAA	UMTS-TDD (TD-SCDMA, 1.28 Mcps)	TD-SCDMA	11.01	± 9.6 %
10058	DAC	EDGE-FDD (TOMA, 8PSK, TN 0-1-2-3)	GSM	6.52	± 9.6 %
10059	CAB	IEEE 802.11b WIFI 2.4 GHz (DSSS, 2 Mbps)	WLAN	2.12	± 9.6 °
10060	CAB	IEEE 802.11b WIFI 2.4 GHz (DSSS, 5.5 Mbps)	WLAN	2.83	± 9.6 1
10061	CAB	IEEE 802.11b WIFI 2.4 GHz (DSSS, 11 Mbps)	WLAN	3,60	±9.6 °
10062	CAD	IEEE 802.11a/h WiFi 5 GHz (OFDM, 6 Mbps)	WLAN	8.68	± 9.6 *
10063	CAD	IEEE 802.11a/h WiFi 5 GHz (OFDM, 9 Mbps)	WLAN	8.63	± 9.6 *
10064	CAD	IEEE 802.11a/h WiFi 5 GHz (OFDM, 12 Mbps)	WLAN	9.09	± 9.6 %
10065	CAD	IEEE 802.11a/h WiFi 5 GHz (OFDM, 18 Mbps)	WLAN	9.00	± 9.6 %
10066	CAD	IEEE 802.11a/h WIFI 5 GHz (OFDM, 24 Mbps)	WLAN	9.38	± 9.6 %
10067	CAD	IEEE 802.11a/h WIFI 5 GHz (OFDM, 36 Mbps)	WLAN	10.12	± 9.6 !
10068	CAD	IEEE 802.11a/h WIFI 5 GHz (OFDM, 48 Mbps)	WLAN	10.24	± 9.6 %
10069	CAD	IEEE 802.11a/h WiFi 5 GHz (OFDM, 54 Mbps)	WLAN	10.56	+9.6 1
10071	CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 9 Mbps)	WLAN	9.83	± 9.6 %
10072	CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 12 Mbps)	WLAN	9.62	± 9.6 *
10073	CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 18 Mbps)	WLAN	9,94	± 9.6 *
10074	CAB	IEEE 802.11g WiFI 2.4 GHz (DSSS/OFDM, 24 Mbps)	WLAN	10.30	± 9.6 %
10075	CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 36 Mbps)	WLAN	10.77	± 9.6 %
10076	CAB	IEEE 802.11g WIFI 2.4 GHz (DSSS/OFDM, 48 Mbps)	WLAN	10,94	± 9.6 %
10077	CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 54 Mbps)	WLAN	11.00	± 9.6 1
10081	CAB	CDMA2000 (1xRTT, RC3)	CDMA2000	3.97	±9.65
10082	CAB	IS-54 / IS-138 FDD (TDMA/FDM, PI/4-DQPSK, Fullrate)	AMPS	4.77	±9.61
10090	DAC	GPRS-FDD (TDMA, GMSK, TN 0-4)	GSM	6.56	± 9,6 %
10097	CAB	UMTS-FDD (HSDPA)	WCDMA	3.98	± 9.6 *
10098	CAB	UMTS-FDD (HSUPA, Subtest 2)	WCDMA	3,98	± 9.6 °
10099	DAC	EDGE-FDD (TDMA, 8PSK, TN 0-4)	GSM	9.55	± 9.6 *

Certificate No: EX3-7679_Sep21

Page 10 of 22



EX3DV4- SN:7679

September 10, 2021

10100	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	5.67	±9.6%
10101	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, 16-QAM)	LTE-FDD	6.42	± 9.6 %
10102	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, 64-QAM)	LTE-FDD	6.60	± 9.6 %
10103	CAG	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-TDD	9.29	± 9.6 %
10104	CAG	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, 16-QAM)	LTE-TDD	9.97	± 9.6 %
10105	CAG	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, 64-QAM)	LTE-TDD	10.01	± 9.6 %
10108	CAG	LTE-FDD (SC-FDMA, 100% RB, 10 MHz, QPSK)	LTE-FDD	5.80	± 9.6 %
10109	CAG	LTE-FDD (SC-FDMA, 100% RB, 10 MHz, 16-QAM)	LTE-FDD	6.43	± 9.6 %
10110	CAG	LTE-FDD (SC-FDMA, 100% RB, 5 MHz, QPSK)	LTE-FDD	5.75	± 9.6 %
10111	CAG	LTE-FDD (SC-FDMA, 100% RB, 5 MHz, 18-QAM)	LTE-FDD	6.44	± 9.6.9
10112	CAG	LTE-FDD (SC-FDMA, 100% RB, 10 MHz, 64-QAM)	LTE-FDD	6.59	± 9.6 %
10113	CAG	LTE-FDD (SC-FDMA, 100% RB, 5 MHz, 64-QAM)	LTE-FDD	6.62	± 9.6 %
10114	CAD	IEEE 802.11n (HT Greenfield, 13.5 Mbps, BPSK)	WLAN	8.10	± 9.6 %
10115	CAD	IEEE 802.11n (HT Greenfield, 81 Mbps, 16-QAM)	WLAN	8.46	± 9.6 %
10116	CAD	IEEE 802.11n (HT Greenfield, 135 Mbps, 64-QAM)	WLAN	8.15	± 9.6 %
10117	CAD	IEEE 802.11n (HT Mixed, 13.5 Mbps, BPSK)	WLAN	8.07	±9.6 %
10115	CAD	IEEE 802.11n (HT Mixed, 81 Mbps, 16-QAM)	WLAN	8.59	± 9.6 %
10119	CAD	IEEE 802,11n (HT Mixed, 135 Mbps, 64-QAM)	WLAN	8.13	± 9.6 %
10140	CAE	LTE-FDD (SC-FDMA, 100% RB, 15 MHz, 16-QAM)	LTE-FDD	6.49	±9.6 %
10141	CAE	LTE-FDD (SC-FDMA, 100% RB, 15 MHz, 64-QAM)	LTE-FDD	6.53	± 9.6 %
10142	CAE	LTE-FDD (SC-FDMA, 100% RB, 3 MHz, QPSK)	LTE-FDD	5.73	± 9.6 %
10143	CAE	LTE-FDD (SC-FDMA, 100% RB, 3 MHz, 16-QAM)	LTE-FDD	6.35	± 9.6 %
10144	CAE	LTE-FDD (SC-FDMA, 100% RB, 3 MHz, 64-QAM)	LTE-FDD	6.65	± 9.6 %
10145	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.4 MHz, QPSK)	LTE-FDD	5.76	± 9.6 °
10146	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.4 MHz, 16-QAM)	LTE-FDD	6.41	± 9.6 *
10147	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.4 MHz, 64-QAM)	LTE-FDD	6.72	± 9.6 1
10149	CAE	LTE-FDD (SC-FDMA, 50% RB, 20 MHz, 16-QAM)	LTE-FDD	6.42	# 9.6 1
10150	CAE	LTE-FDD (SC-FDMA, 50% RB, 20 MHz, 64-QAM)	LTE-FDD	6.60	± 9.6 *
10151	CAG	LTE-TDD (SC-FDMA, 50% RB, 20 MHz, QPSK)	LTE-TDD	9.28	± 9.6 1
10/152	CAG	LTE-TDD (SC-FDMA, 50% RB, 20 MHz, 16-QAM)	LTE-TDD	9.92	#9.6
10153	CAG	LTE-TDD (SC-FDMA, 50% RB, 20 MHz, 64-QAM)	LTE-TDD	10.05	± 9.6 4
10154	CAG	LTE-FDD (SC-FDMA, 50% RB, 10 MHz, QPSK)	LTE-FDD	5.75	± 9.6 1
10155	CAG	LTE-FDD (SC-FDMA, 50% RB, 10 MHz, 16-QAM)	LTE-FDD	6.43	± 9.6 °
10156	CAG	LTE-FDD (SC-FDMA, 50% RB, 5 MHz, QPSK)	LTE-FDD	5.79	±9.6 °
10157	CAG	LTE-FDD (SC-FDMA, 50% RB, 5 MHz, 16-QAM)	LTE-FDD	6.49	± 9.6 *
10158	CAG	LTE-FDD (SC-FDMA, 50% R8, 10 MHz, 64-QAM)	LTE-FDD	6.62	± 9.6.
10159	CAG	LTE-FDD (SC-FDMA, 50% RB, 5 MHz, 64-QAM)	LTE-FDD	6.56	± 9.6 9
10160	CAE	LTE-FDD (SC-FDMA, 50% RB, 15 MHz, QPSK)	LTE-FDD	5.82	± 9.6 '
10161	CAE	LTE-FDD (SC-FDMA, 50% RB, 15 MHz, 16-QAM)	LTE-FDD	6.43	± 9.6 1
10162	CAE	LTE-FDD (SC-FDMA, 50% RB, 15 MHz, 64-QAM)	LTE-FDD	6.58	± 9.6
10166	CAF	LTE-FDD (SC-FDMA, 50% RB, 1.4 MHz, QPSK)	LTE-FDD	5.46	± 9.6
10166	CAF	LTE-FDD (SC-FDMA, 50% RB, 1.4 MHz, 16-QAM)	LTE-FDD	6.21	± 9.6 °
10168	CAF	LTE-FDD (SC-FDMA, 50% RB, 1.4 MHz, 64-QAM)	LTE-FDD	6.79	± 9.6
10169	CAE	LTE-FDD (SC-FDMA, 1 RB, 20 MHz, OPSK)	LTE-FDD	5.73	± 9.6 °
10170	CAE	LTE-FDD (SC-FDMA, 1 RB, 20 MHz, 16-QAM)	LTE-FDD	6.52	± 9.6
10171	AAE	LTE-FDD (SC-FDMA, 1 RB, 20 MHz, 64-QAM)	LTE-FDD	6.49	± 9.6
10172		LTE-TDD (SC-FDMA, 1 RB, 20 MHz, QPSK)	LTE-TDD	9.21	± 9.6
10173	and the second s	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, 16-QAM)	LTE-TDD	9.48	± 9.6 °
10174	CAG	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, 64-QAM)	LTE-TDD	10.25	± 9.6
10175	CAG	LTE-FDD (SC-FDMA, 1 RB, 10 MHz, QPSK)	LTE-FDD	5.72	± 9.6
10176	CAG	LTE-FDD (SC-FDMA, 1 RB, 10 MHz, 16-DAM)	LTE-FDD	6.52	± 9.6
10177	CAL	LTE-FDD (SC-FDMA, 1 RB, 5 MHz, QPSK)	LTE-FDD	5.73	± 9.6
10178	CAG	LTE-FDD (SC-FDMA, 1 RB, 5 MHz, 16-QAM)	LTE-FDD	6.52	± 9.6
10178	CAG	LTE-FDD (SC-FDMA, 1 RB, 10 MHz, 64-QAM)	LTE-FDD	6.50	± 9.6
101/9		LTE-FDD (SC-FDMA, 1 RB, 5 MHz, 64-QAM)	LTE-FDD	6.50	± 9.6
1111100	L'UNUS	LINE AND DUCTION TO A DIST. OF LOW	LIE-FUU	0.00	I I 8.0

Certificate No: EX3-7679_Sep21

Page 11 of 22



EX3DV4- SN:7679

September 10, 2021

10182	CAE	LTE-FDD (SC-FDMA, 1 RB, 15 MHz, 16-QAM)	LTE-FDD	6.52	±9,6 %
10183	AAD	LTE-FDD (SC-FDMA, 1 RB, 15 MHz, 64-QAM)	LTE-FDD	6.50	±9.6 %
10184	CAE	LTE-FDD (SC-FDMA, 1 RB, 3 MHz, QPSK)	LTE-FDD	5.73	± 9.6 5
10185	CAE	LTE-FDD (SC-FDMA, 1 RB, 3 MHz, 16-QAW)	LTE-FDD	6.51	± 9.6 %
10186	AAE	LTE-FDD (SC-FDMA, 1 RB, 3 MHz, 64-QAM)	LTE-FDD	6.50	± 9.6 5
10187	CAF	LTE-FDD (SC-FDMA, 1 RB, 1.4 MHz, QPSK)	LTE-FDD	5.73	± 9.6 %
10188	CAF	LTE-FDD (SC-FDMA, 1 RB, 1.4 MHz, 16-QAM)	LTE-FDD	6.52	± 9.6 %
10189	AAF	LTE-FDD (SC-FDMA, 1 RB, 1.4 MHz, 64-QAM)	LTE-FDD	6.50	± 9.6 %
10193	CAD	IEEE 802.11n (HT Greenfield, 6.5 Mbps, BPSK)	WLAN	8.09	± 9.6 %
10194	CAD	IEEE 802.11n (HT Greenfield, 39 Mbps, 16-QAM)	WLAN	8.12	± 9.6 %
10195	CAD	IEEE 802.11n (HT Greenfield, 65 Mbps, 64-QAM)	WLAN	8.21	± 9.6 %
10196	CAD	IEEE 802.11n (HT Mixed, 6.5 Mbps, BPSK)	WLAN	8.10	± 9.6 %
10197	CAD	IEEE 802.11n (HT Mixed, 39 Mbps, 16-QAM)	WLAN	8.13	± 9.6 1
10198	CAD	IEEE 802.11n (HT Mixed, 65 Mbps, 64-QAM)	WLAN	8.27	± 9.6 1
10219	CAD	IEEE 802.11n (HT Mixed, 7.2 Mbps, BPSK)	WLAN	8.03	± 9.6 9
10220	CAD	IEEE 802.11n (HT Mixed, 43.3 Mbps, 16-QAM)	WLAN	8.13	± 9.6 %
10221	CAD	IEEE 802.11n (HT Mixed, 72.2 Mbps, 64-QAM)	WLAN	8.27	±9.6 %
10222	CAD	IEEE 802.11n (HT Mixed, 15 Mbps, BPSK)	WLAN	8.06	± 9.6 9
10223	CAD	IEEE 802.11n (HT Mixed, 90 Mbps, 16-QAM)	WLAN	8.48	± 9.6.9
10224	CAD	IEEE 802.11n (HT Mixed, 150 Mbps, 64-QAM)	WLAN	8.08	± 9.6 %
10225	CAB	UMTS-FDD (HSPA+)	WCDMA	5.97	± 9.6 %
10226	CAB	LTE-TDD (SC-FDMA, 1 RB, 1.4 MHz, 16-QAM)	LTE-TDD	9.49	± 9.6 %
10227	CAB	LTE-TDD (SC-FDMA, 1 RB, 1.4 MHz, 64-QAM)	LTE-TDD	10.26	± 9.6 %
10228	CAB	LTE-TDD (SC-FDMA, 1 RB, 1.4 MHz, QPSK)	LTE-TDD	9.22	± 9.6.9
10229	CAD	LTE-TOD (SC-FDMA, 1 RB, 3 MHz, 16-QAM)	LTE-TDD	9.48	± 9.6 %
10230	CAD	LTE-TDD (SC-FDMA, 1 RB, 3 MHz, 64-QAM)	LTE-TDD	10.25	± 9.6 °
10231	CAD	LTE-TDD (SC-FDMA, 1 RB, 3 MHz, QPSK)	LTE-TDD	9.19	± 9.6 1
10232		LTE-TDD (SC-FDMA, 1 RB, 5 MHz, 16-QAM)	LTE-TDD	9.48	± 9.6 1
10233		LTE-TDD (SC-FDMA, 1 R8, 5 MHz, 64-QAM)	LTE-TDD	10.25	± 9.6 1
10234		LTE-TDD (SC-FDMA, 1 RB, 5 MHz, QPSK)	LTE-TDD	9.21	± 9.61
10235	CAG	LTE-TDD (SC-FDMA, 1 RB, 10 MHz, 16-QAM)	LTE-TDD	9.48	± 9.6 *
10236	CAG	LTE-TDD (SC-FDMA, 1 RB, 10 MHz, 64-QAM)	LTE-TDD	10.25	± 9.6 1
10237	CAG	LTE-TDD (SC-FDMA, 1 RB, 10 MHz, QPSK)	LTE-TDD	9.21	± 9.6 °
10238	CAF	LTE-TDD (SC-FDMA, 1 RB, 15 MHz, 16-QAM)	LTE-TDD	9.48	± 9.6 *
10239	CAF	LTE-TDD (SC-FDMA, 1 RB, 15 MHz, 64-QAM)	LTE-TDD	10.25	± 9.6 °
10240	CAF	LTE-TDD (SC-FDMA, 1 R8, 15 MHz, QPSK)	LTE-TDD	9.21	± 9.6 *
10241	CAB	LTE-TDD (SC-FDMA, 50% RB, 1.4 MHz, 16-QAM)	LTE-TDD	9.82	± 9.6 °
10242	CAB	LTE-TDD (SC-FDMA, 50% RB, 1.4 MHz, 64-QAM)	LTE-TDD	9.85	± 9.6 1
10243	CAB	LTE-TDD (SC-FDMA, 50% RB, 1.4 MHz, QPSK)	LTE-TDD	9.46	± 9.6 1
10244	CAD	LTE-TDD (SC-FDMA, 50% RB, 3 MHz, 16-QAM)	LTE-TDD	10.06	± 9.6 *
10245	CAD	LTE-TDD (SC-FDMA, 50% RB, 3 MHz, 64-QAM)	LTE-TDD	10.06	± 9.6 1
10246	CAD	LTE-TDD (SC-FDMA, 50% RB, 3 MHz, OPSK)	LTE-TDD	9.30	±9.61
10247	CAG	LTE-TDD (SC-FDMA, 50% RB, 5 MHz, 16-QAM)	LTE-TDD	9.91	± 9.6 1
10248	CAG	LTE-TDD (SC-FDMA, 50% RB, 5 MHz, 64-QAM)	LTE-TDD	10.09	±9.6*
10249	CAG	LTE-TDD (SC-FDMA, 50% RB, 5 MHz, QPSK)	LTE-TDD	9.29	± 9.6 1
10250	CAG	LTE-TDD (SC-FDMA, 50% RB, 10 MHz, 16-QAM)	LTE-TDD	9.81	± 9.6 *
10251	CAG	LTE-TDD (SC-FDMA, 50% RB, 10 MHz, 64-QAM)	LTE-TDD	10.17	± 9.6
10252	CAG	LTE-TDD (SC-FDMA, 50% RB, 10 MHz, QPSK)	LTE-TDD	9.24	± 9.6 1
10253	CAF	LTE-TDD (SC-FDMA, 50% RB, 15 MHz, 18-QAM)	LTE-TDD	9.90	± 9.6
10254	CAF	LTE-TDD (SC-FDMA, 50% RB, 15 MHz, 64-QAM)	LTE-TDD	10.14	± 9.6 1
10255	and the second second	LTE-TDD (SC-FDMA, 50% RB, 15 MHz, QPSK)	LTE-TDD	9.20	± 9.6
10256	and the second second	LTE-TDD (SC-FDMA, 100% RB, 1.4 MHz, 16-QAM)	LTE-TDD	9.96	± 9.6 3
10257	CAB	LTE-TDD (SC-FDMA, 100% RB, 1.4 MHz, 64-QAM)	LTE-TDD	10.08	± 9.6 4
10258	CAB	LTE-TDD (SC-FDMA, 100% RB, 1.4 MHz, QPSK)	LTE-TDD	9.34	± 9.6 9
10259	CAD	LTE-TDD (SC-FDMA, 100% RB, 3 MHz, 16-QAM)	LTE-TDD	9.98	± 9.6 1
10260	CAD	LTE-TDD (SC-FDMA, 100% RB, 3 MHz, 64-QAM)	LTE-TDD	9.97	± 9.6

Certificate No: EX3-7679_Sep21

Page 12 of 22





EX3DV4-- SN:7679

September 10, 2021

10261	CAD	LTE-TOD (SC-FDMA, 100% RB, 3 MHz, QPSK)	LTE-TDD	9.24	±9.6 %
10262	CAG	LTE-TDD (SC-FDMA, 100% RB, 5 MHz, 16-QAM)	LTE-TDD	9.83	± 9.6 %
10263	CAG	LTE-TDD (SC-FDMA, 100% RB, 5 MHz, 84-QAM)	LTE-TDD	10.16	± 9.6 %
10264	CAG	LTE-TDD (SC-FDMA, 100% RB, 5 MHz, QPSK)	LTE-TDD	9.23	± 9.6 %
10265	CAG	LTE-TDD (SC-FDMA, 100% RB, 10 MHz, 16-QAM)	LTE-TDD	9,92	± 9.6 %
10266	CAG	LTE-TDD (SC-FDMA, 100% RB, 10 MHz, 64-QAM)	LTE-TDD	10.07	± 9.6 %
10267	CAG	LTE-TOD (SC-FDMA, 100% RB, 10 MHz, QPSK)	LTE-TDD	9.30	± 9.6 %
10268	CAF	LTE-TDD (SC-FDMA, 100% RB, 15 MHz, 16-QAM)	LTE-TDD	10.06	± 9.6 %
10269	CAF	LTE-TOD (SC-FDMA, 100% RB, 15 MHz, 84-QAM)	LTE-TDD	10.13	± 9.6 %
10270	CAF	LTE-TOD (SC-FDMA, 100% RB, 15 MHz, QPSK)	LTE-TDD	9.58	± 9.6 %
10274	CAB	UMTS-FDD (HSUPA, Subtest 5, 3GPP Rel8.10)	WCDMA	4.87	± 9.6 %
10275	CAB	UMTS-FDD (HSUPA, Subtest 5, 3GPP Rel8.4)	WCDMA	3.96	± 9.6 %
10277	CAA	PHS (QPSK)	PHS	11.81	± 9.6 %
10278	CAA	PHS (QPSK, BW 884MHz, Rolloff 0.5)	PHS	11.81	± 9.6 %
10279	CAA	PHS (QPSK, BW 884MHz, Rolloff 0.38)	PHS	12.18	± 9.6 %
10290	AAB	CDMA2000, RC1, SO55, Full Rate	CDMA2000	3.91	± 9.6 %
10291	AAB	CDMA2000, RC3, SO55, Full Rate	CDMA2000	3.46	± 9.6 %
10291	AAB	CDMA2000, RC3, SO32, Full Rate	CDMA2000	3.39	± 9.6 %
10292	AAB	CDMA2000, RC3, SO3, Full Rate	CDMA2000	3.50	± 9.6 %
10295	AAB	CDMA2000, RC1, SO3, 1/8th Rate 25 fr.	CDMA2000	12.49	±9.6 %
10295	AAD	LTE-FDD (SC-FDMA, 50% RB, 20 MHz, QPSK)	LTE-FDD	5.81	±9.69
	AAD	LTE-FDD (SC-FDMA, 50% RB, 3 MHz, QPSK)	LTE-FDD	5.72	±9.69
10298		LTE-FDD (SC-FDMA, 50% R8, 3 MHz, 4FGA)	LTE-FDD	6.39	±9.6 9
10299	AAD	LTE-FDD (SC-FDMA, 50% RB, 3 MHz, 10-04M)	LTE-FDD	6.60	±9.69
10300	Contraction of the local division of the loc	IEEE 802.16e WMAX (29:18, 5ms, 10MHz, QPSK, PUSC)	WIMAX	12.03	110,000,00
10301	AAA	IEEE 802.166 WMAA (29.18, 5ms, 10MHz, QPSK, PUSC) IEEE 802.166 WMAA (29.18, 5ms, 10MHz, QPSK, PUSC, 3CTRL)	and the second		±9.6 %
10302	AAA		WIMAX	12.57	a second s
10303	AAA	IEEE 802.16e WIMAX (31:15, 5ms, 10MHz, 64QAM, PUSC)	WIMAX	12.52	± 9.6 %
10304	AAA	IEEE 802.160 WIMAX (29:18, 5ms, 10MHz, 64QAM, PUSC)	WIMAX	11.86	± 9,6 %
10305	AAA	IEEE 802.16e WIMAX (31:15, 10ms, 10MHz, 64QAM, PUSC)	WIMAX	15.24	± 9.6 %
10306	AAA	IEEE 802.16e WIMAX (29:18, 10ms, 10MHz, 64QAM, PUSC)	WiMAX	14.67	± 9.6.9
10307	AAA	IEEE 802.16e WIMAX (29.18, 10ms, 10MHz, QPSK, PUSC)	XAMEW	14.49	±9.69
10308	AAA	IEEE 802.16e WIMAX (29:18, 10ms, 10MHz, 16QAM, PUSC)	WIMAX	14.46	± 9.6.9
10309	AAA	IEEE 802.16e WIMAX (29:18, 10ms, 10MHz, 16QAM,AMC 2x3)	WIMAX	14.58	± 9.6 %
10310	AAA	IEEE 802.16e WIMAX (29:18, 10ms, 10MHz, QPSK, AMC 2x3	WIMAX	14.57	± 9.6 %
10311	AAD	LTE-FDD (SC-FDMA, 100% RB, 15 MHz, QPSK)	LTE-FDD	6.06	± 9.6 %
10313	AAA	IDEN 1:3	IDEN	10.51	± 9.6 %
10314	AAA	IDEN 1:8	IDEN	13.48	±9.69
10315	AAB	IEEE 802.11b WiFi 2.4 GHz (DSSS, 1 Mbps, 96pc dc)	WLAN	1.71	± 9.6 %
10316	AAB.	IEEE 802.11g WiFI 2.4 GHz (ERP-OFDM, 6 Mbps, 96pc dc)	WLAN	8.36	± 9.6 %
10317	AAD	IEEE 802.11a WIFI 5 GHz (OFDM, 6 Mbps, 96pc dc)	WLAN	8.36	± 9.6 %
10352	AAA	Pulse Waveform (200Hz, 10%)	Generic	10.00	± 9.6 %
10353	AAA	Pulse Waveform (200Hz, 20%)	Generic	6.99	± 9.6 %
10354	AAA	Pulse Waveform (200Hz, 40%)	Generic	3.98	± 9.6 %
10355	AAA	Pulse Waveform (200Hz, 60%)	Generic	2.22	± 9.6.9
10356	AAA	Pulse Waveform (200Hz, 80%)	Generic	0.97	± 9.6 %
10387	AAA	QPSK Waveform, 1 MHz	Generic	5.10	± 9.6 %
10388	AAA	GPSK Waveform, 10 MHz	Generic	5.22	± 9.6 %
10396	AAA	64-QAM Waveform, 100 kHz	Generic	6.27	± 9.6 %
10399	AAA	64-QAM Waveform, 40 MHz	Generic	6.27	± 9.6 %
10400	AAE	IEEE 802.11ac WiFi (20MHz, 64-QAM, 99pc dc)	WLAN	8.37	± 9.6 %
10401	AAE	IEEE 802.11ac WiFi (40MHz, 64-QAM, 99pc dc)	WLAN	8.60	± 9.6 %
10402	AAE	IEEE 802.11ac WIFI (80MHz, 64-QAM, 99pc dc)	WLAN	8.53	± 9.6 %
10403	AAB	CDMA2000 (1xEV-DO, Rev. 0)	CDMA2000	3.76	± 9.6.9
10404	AAB	CDMA2000 (1xEV-DO, Rev. A)	CDMA2000	3.77	± 9.6 %
10406	AAB	CDMA2000, RC3, SO32, SCH0, Full Rate	CDMA2000	5.22	± 9.6.9
10410		LTE-TDD (SC-FDMA, 1 RB, 10 MHz, QPSK, UL Sub=2,3,4,7,8,9)	LTE-TDD	7.82	± 9.6 %

Certificate No: EX3-7679_Sep21

Page 13 of 22



EX3DV4- SN:7679

September 10, 2021

10414	AAA	WLAN CCDF, 64-QAM, 40MHz	Generic	8.54	± 9.6 %
10415	AAA	IEEE 802.11b WIFI 2.4 GHz (DSSS, 1 Mbps, 99pc dc)	WLAN	1.54	± 9.6 %
10416	AAA	IEEE 802.11g WIFI 2.4 GHz (ERP-OFDM, 6 Mbps, 99pc dc)	WLAN	8.23	± 9.6 %
10417	AAC	IEEE 802.11a/h WiFi 5 GHz (OFDM, 6 Mbps, 99pc dc)	WLAN	8.23	± 9.6 %
10418	AAA	IEEE 802 11g WIFI 2.4 GHz (DSSS-OFDM, 6 Mbps, 99pc, Long)	WLAN	8.14	± 9.6 %
10419	AAA	IEEE 802,11g WIFI 2.4 GHz (DSSS-OFDM, 6 Mbps, 99pc, Short)	WLAN	8.19	± 9.6 %
10422	AAC	IEEE 802.11n (HT Greenfield, 7.2 Mbps, BPSK)	WLAN	8.32	± 9.6 %
10423	AAC	IEEE 802.11n (HT Greenfield, 43.3 Mbps, 16-QAM)	WLAN	8.47	± 9.6 %
10424	AAC	IEEE 802.11n (HT Greenfield, 72.2 Mbps, 64-QAM)	WLAN	8.40	± 9.6 %
10425	AAC	IEEE 802,11n (HT Greenfield, 15 Mbps, BPSK)	WLAN	8.41	± 9.6 %
10426	AAC	IEEE 802.11n (HT Greenfield, 90 Mbps, 16-QAM)	WLAN	8.45	± 9.6 %
10427	AAC	IEEE 802.11n (HT Greenfield, 150 Mbps, 64-QAM)	WLAN	8.41	± 9.6 %
10430	AAD	LTE-FDD (OFDMA, 5 MHz, E-TM 3.1)	LTE-FDD	8.28	± 9.6 %
10431	AAD	LTE-FOD (OFDMA, 10 MHz, E-TM 3.1)	LTE-FDD	8.38	± 9.6 %
10432	AAC	LTE-FDD (OFDMA, 15 MHz, E-TM 3.1)	LTE-FDD	8.34	± 9.6 %
10433	AAC	LTE-FDD (OFDMA, 20 MHz, E-TM 3.1)	LTE-FDD	8.34	± 9.6 %
10434	AAA	W-CDMA (BS Test Model 1, 64 DPCH)	WCDMA	8.60	± 9.6 %
10435	AAF	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, QPSK, UL Sub)	LTE-TDD	7.82	± 9.6 %
10447	AAD	LTE-FDD (OFDMA, 5 MHz, E-TM 3.1, Clipping 44%)	LTE-FDD	7.56	± 9.6 %
10448	AAD	LTE-FDD (OFDMA, 10 MHz, E-TM 3.1, Clippin 44%)	LTE-FDD	7.53	± 9.6 %
10449	AAC	LTE-FDD (OFDMA, 15 MHz, E-TM 3.1, Cliping 44%)	LTE-FDD	7.51	± 9.6 %
10450	AAC	LTE-FDD (OFDMA, 20 MHz, E-TM 3.1, Clipping 44%)	LTE-FDD	7.48	± 9.6.%
10451	AAA	W-CDMA (BS Test Model 1, 64 DPCH, Clipping 44%)	WCDMA	7.59	± 9.6 9
10453	AAD	Validation (Square, 10ms, 1ms)	Test	10.00	±9.65
10456	AAC	IEEE 802.11ac WiFi (160MHz, 64-QAM, 99pc dc)	WLAN	B.63	±9.69
10457	AAA	UMTS-FDD (DC-HSDPA)	WCDMA	6.62	±9.69
10458	AAA	CDMA2000 (1xEV-DO, Rev. B, 2 carriers)	CDMA2000	6.55	±9.6.9
10459	AAA	CDMA2000 (1xEV-DO, Rev. B, 3 carriers)	CDMA2000	8.25	±9.6 %
10460	AAA	UMTS-FDD (WCDMA, AMR)	WCDMA	2.39	±9.6 9
10461	AAB	LTE-TDD (SC-FDMA, 1 RB, 1.4 MHz, QPSK, UL Sub)	LTE-TDD	7.82	19.6 %
10462	AAB	LTE-TDD (SC-FDMA, 1 RB, 1.4 MHz, 16-QAM, UL Sob)	LTE-TDD	8.30	± 9.6 %
10463	AAB	LTE-TDD (SC-FDMA, 1 RB, 1.4 MHz, 64-QAM, UL Sub)	LTE-TDD	8.56	± 9.6 %
10464	AAC	LTE-TDD (SC-FDMA, 1 RB, 3 MHz, QPSK, UL Sub)	LTE-TDD	7.82	± 9.6 %
10465	AAC	LTE-TDD (SC-FDMA, 1 RB, 3 MHz, 16-QAM, UL Sub)	LTE-TDD	8.32	± 9.6 %
10466	AAC	LTE-TDD (SC-FDMA, 1 RB, 3 MHz, 64-QAM, UL Sub)	LTE-TDD	8.57	± 9.6 9
10467	AAF	LTE-TDD (SC-FDMA, 1 RB, 5 MHz, QPSK, UL Sub)	LTE-TDD	7.82	± 9.6 %
10468	AAF	LTE-TOD (SC-FDMA, 1 RB, 5 MHz, 16-QAM, UL Sub)	LTE-TDD	8.32	±9.6 9
10469	AAF	LTE-TOD (SC-FDMA, 1 RB, 5 MHz, 64-QAM, UL Sub)	LTE-TDD	8.56	± 9.6 %
10400	AAF	LTE-TDD (SC-FDMA, 1 RB, 10 MHz, QPSK, UL Sub)	LTE-TDD	7.82	±9.69
10471	AAF	LTE-TDD (SC-FDMA, 1 RB, 10 MHz, 16-QAM, UL Sub)	LTE-TDD	8.32	± 9.6 %
10472	AAF	LTE-TDD (SC-FDMA, 1 RB, 10 MHz, 64-QAM, UL Sub)	LTE-TDD	8.57	±9.6 9
10473	AAE	LTE-TDD (SC-FDMA, 1 RB, 15 MHz, QPSK, UL Sub)	LTE-TDD	7.82	±9.69
10474	AAE	LTE-TOD (SC-FDMA, 1 RB, 15 MHz, 16-QAM, UL Sub)	LTE-TDD	8.32	±9.69
10474	AAE	LTE-TDD (SC-FDMA, 1 RB, 15 MHz, 64-QAM, UL Sub)	LTE-TDD	8.57	± 9.6 %
10475	AAF	LTE-TDD (SC-FDMA, 1 RB, 15 MHz, 64-QAM, UL Sub)	LTE-TDD	8.32	±9.69
10478	AAF	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, 64-QAM, UL Sub)	LTE-TDD	8.57	±9.69
10479	AAB	LTE-TDD (SC-FDMA, 50% RB, 1.4 MHz, QPSK, UL Sub)	LTE-TDD	7.74	± 9.6 9
10479	AAB	LTE-TDD (SC-FDMA, 50% RB, 1.4 MHz, 16-QAM, UL Sub)	LTE-TDD	8.18	±9.6 9
10480	AAB	LTE-TDD (SC-FDMA, 50% RB, 1.4 MHz, 64-QAM, UL Sub)	LTE-TDD	8.45	±9.6 %
10482	AAC	LTE-TOD (SC-FDMA, 50% RB, 1.4 kHz, 64-04W, 0L 500)	LTE-TDD	7.71	± 9.6 %
10483	AAC	LTE-TDD (SC-FDMA, 50% RB, 3 MHz, 16-QAM, Sub)	the second se	8.39	± 9.6 %
10483	AAC	LTE-TDD (SC-FDMA, 50% RB, 3 MHz, 16-QAM, 5ub) LTE-TDD (SC-FDMA, 50% RB, 3 MHz, 64-QAM, UL Sub)	LTE-TDD	8.39	±9.69
10484	AAG	LTE-TOD (SC-FDMA, 50% RB, 5 MHz, 64-GAM, 6L Sub) LTE-TOD (SC-FDMA, 50% RB, 5 MHz, QPSK, UL Sub)	LTE-TDD		
the second second states of the	Charles Streeting	LTE-TDD (SC-FDMA, 50% RB, 5 MHz, GPSK, UL Sub) LTE-TDD (SC-FDMA, 50% RB, 5 MHz, 16-QAM, UL Sub)	LTE-TDD	7.59	±9.69
10486	AAF	LTE-TDD (SC-FDMA, 50% RB, 5 MHZ, 16-QAM, UL SUB) LTE-TDD (SC-FDMA, 50% RB, 5 MHZ, 64-QAM, UL SUB)	LTE-TDD LTE-TDD	8.38	± 9.6 %
10487					

Certificate No: EX3-7679_Sep21

Page 14 of 22



EX3DV4- SN:7679

September 10, 2021

D489	AAF	LTE-TOD (SC-FDMA, 50% RB, 10 MHz, 16-QAM, UL Sub)	LTE-TDD	8.31	# 9.6 %
0490	AAF	LTE-TDD (SC-FDMA, 50% RB, 10 MHz, 64-QAM, UL Sub)	LTE-TDD	8.54	± 9.6 %
0491	AAE	LTE-TDD (SC-FDMA, 50% RB, 15 MHz, QPSK, UL Sub)	LTE-TDD	7.74	± 9.6 %
0492	AAE	LTE-TDD (SC-FDMA, 50% R8, 15 MHz, 18-QAM, UL Sub)	LTE-TDD	8.41	± 9.6 %
0493	AAE	LTE-TDD (SC-FDMA, 50% RB, 15 MHz, 64-QAM, UL Sub)	LTE-TDD	8.55	± 9.6 %
0494	AAF	LTE-TDD (SC-FDMA, 50% RB, 20 MHz, QPSK, UL Sub)	LTE-TDD	7.74	± 9.6 %
0495	AAF	LTE-TDD (SC-FDMA, 50% RB, 20 MHz, 16-QAM, UL Sub)	LTE-TDD	8.37	± 9.6 %
0496	AAF	LTE-TDD (SC-FDMA, 50% RB, 20 MHz, 64-QAM, UL Sub)	LTE-TDD	8.54	± 9.6 %
0490	AAB	LTE-TOD (SC-FDMA, 100% RB, 1.4 MHz, QPSK, UL Sub)	LTE-TDD	7.67	± 9.6 %
0498	AAB	LTE-TDD (SC-FDMA, 100% RB, 1.4 MHz, 16-QAM, UL Sub)	LTE-TDD	8,40	± 9.6 %
0499	AAB	LTE-TDD (SC-FDMA, 100% RB, 1.4 MHz, 64-QAM, UL Sub)	LTE-TDD	8.68	± 9.6 %
0500	AAC	LTE-TDD (SC-FDMA, 100% RB, 3 MHz, QPSK, UL Sub)	LTE-TDD	7.67	± 9.6 %
0501	AAC	LTE-TDD (SC-FDMA, 100% RB, 3 MHz, 16-QAM, UL Sub)	LTE-TDD	8.44	± 9.6 %
0502	AAC	LTE-TDD (SC-FDMA, 100% RB, 3 MHz, 64-QAM, UL Sub)	LTE-TDD	8.52	±9.6.9
0502	AAF	LTE-TDD (SC-FDMA, 100% RB, 5 MHz, QPSK, UL Sub)	LTE-TDD	7.72	± 9.6 %
0504	AAF	LTE-TDD (SC-FDMA, 100% RB, 5 MHz, 16-QAM, UL Sub)	LTE-TDD	8.31	± 9.6 %
0505	AAF	LTE-TDD (SC-FDMA, 100% RB, 5 MHz, 64-QAM, UL Sub)	LTE-TDD	8.54	± 9.6 %
0506	AAF	LTE-TDD (SC-FDMA, 100% RB, 10 MHz, QPSK, UL Sub)	LTE-TDD	7.74	± 9.6 %
0507	AAF	LTE-TDD (SC-FDMA, 100% RB, 10 MHz, 16-QAM, UL Sub)	LTE-TDD	8.36	± 9.6 %
0508	AAF	LTE-TDD (SC-FDMA, 100% RB, 10 MHz, 64-QAM, UL Sub)	LTE-TDD	8.55	± 9.6 %
0509	AAE	LTE-TDD (SC-FDMA, 100% RB, 15 MHz, QPSK, UL Sub)	LTE-TDD	7.99	± 9.6 %
0510	AAE	LTE-TDD (SC-FDMA, 100% RB, 15 MHz, 16-QAM, UL Sub)	LTE-TDD	8.49	± 9.6 9
0511	AAE	LTE-TDD (SC-FDMA, 100% RB, 15 MHz, 64-QAM, UL Sub)	LTE-TDD	8.51	± 9.6 %
0512	AAF	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, QPSK, UL Sub)	LTE-TDD	7.74	± 9.6 %
0513	AAF	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, 16-QAM, UL Sub)	LTE-TDD	8.42	# 9.6 5
0513	AAF	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, 64-QAM, UL Sub)	LTE-TDD	8.45	± 9.6 %
0515	AAA	IEEE 802.11b WiFi 2.4 GHz (DSSS, 2 Mbps, 99pc dc)	WLAN	1.58	±9.63
0516	AAA	IEEE 802.11b WiFi 2.4 GHz (DSSS, 5.5 Mbps, 99pc dc)	WLAN	1.57	19.6 %
0517	AAA	IEEE 802.11b WiFi 2.4 GHz (DSSS, 11 Mbps, 99pc dc)	WLAN	1.58	±9.6 9
10518	AAC	IEEE 802.11a/h WIFI 5 GHz (OFDM, 9 Mbps, 99pc dc)	WLAN	8.23	± 9.6.9
10519	AAC	IEEE 802.11a/h WiFi 5 GHz (OFDM, 12 Mbps, 99pc dc)	WLAN	8.39	± 9.6 9
10520	AAC	IEEE 802.11a/h WIFI 5 GHz (OFDM, 18 Mbps, 99pc dc)	WLAN	8.12	±9.6 9
10521	AAC	IEEE 802.11a/h WIFI 5 GHz (OFDM, 24 Mbps, 99pc dc)	WEAN	7.97	± 9.6 9
10522	AAC	IEEE 802.11a/h WIFI 5 GHz (OFDM, 36 Mbps, 99pc dc)	WLAN	8.45	± 9.6 %
10523	AAC	IEEE 802.11a/h WIFI 5 GHz (OFDM, 48 Mbps, 99pc dc)	WLAN	8.08	± 9.6 9
10524	AAC	IEEE 802.11a/h WIFI 5 GHz (OFDM, 54 Mbps, 99pc dc)	WLAN	8.27	± 9.6 "
10525	AAC	IEEE 802,11ac WIFI (20MHz, MCS0, 99pc dc)	WLAN	8.36	± 9.6 9
0526	AAC	IEEE 802.11ac WIFI (20MHz, MCS1, 99pc dc)	WLAN	8.42	± 9.6 %
10527	AAC	IEEE 802.11ac WiFi (20MHz, MCS2, 99pc dc)	WLAN	8.21	± 9.6 %
10528	AAC	IEEE 802.11ac WiFi (20MHz, MCS3, 99pc dc)	WLAN	8.36	± 9.6 9
10529	AAC	IEEE 802.11ac WiFi (20MHz, MCS4, 99pc dc)	WLAN	8.36	± 9.6 %
10531	AAC	IEEE 802.11ac WIFi (20MHz, MCS8, 99pc dc)	WLAN	8.43	± 9.6.9
10532	AAC	IEEE 802.11ac WIFI (20MHz, MCS7, 99pc dc)	WLAN	8.29	± 9.6 9
10533	AAC	IEEE 802.11ac WIFI (20MHz, MCS8, 99pc dc)	WLAN	8.38	± 9.6 9
10534	AAC	IEEE 802.11ac WIFI (40MHz, MCS0, 99pc dc)	WLAN	8.45	±9.6 9
10535	AAC	IEEE 802.11ac WIFI (40MHz, MCS1, 99pc dc)	WLAN	8.45	± 9.6 °
10536	AAC	IEEE 802.11ac WiFi (40MHz, MCS2, 99pc dc)	WLAN	8.32	± 9.6 9
10537	AAC	IEEE 802.11ac WiFi (40MHz, MCS3, 99pc dc)	WLAN	8.44	± 9.6 9
10538	AAC	IEEE 802.11ac WiFi (40MHz, MCS3, 86pc dc)	WLAN	8.54	± 9.6 %
10540	AAC	IEEE 802.11ac WIFI (40MHz, MCS6, 99pc dc)	WLAN	8.39	± 9.6 9
10541	AAC	IEEE 802,11ac WiFi (40MHz, MCS7, 99pc dc)	WLAN	8.46	± 9.6 °
10542	AAC	IEEE 802.11ac WiFi (40MHz, MCS8, 99pc dc)	WLAN	8.65	± 9.6 9
10543	AAC	IEEE 802.11ac WiFi (40MHz, MCS9, 99pc dc)	WLAN	8.65	± 9.6
10544	AAC	IEEE 802.11ac WiFi (80MHz, MCS0, 99pc dc)	WLAN	8.47	± 9.6 °
and the second	AAC	IEEE 802.11ac WiFi (80MHz, MCS1, 99pc dc)	WLAN	8.55	± 9.6
10545			1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1		

Certificate No: EX3-7679_Sep21

Page 15 of 22



EX3DV4-- SN:7679

September 10, 2021

10547	AAC	IEEE 802.11ac WiFI (80MHz, MCS3, 99pc dc)	WLAN	8.49	± 9.6 %
10548	AAC	IEEE 802.11ac WiFi (80MHz, MCS4, 99pc dc)	WLAN	8.37	± 9.6 %
10550	AAC	IEEE 802.11ac WiFi (80MHz, MCS6, 99pc dc)	WLAN	8.39	± 9.6 %
10551	AAC	IEEE 802.11ac WiFi (80MHz, MCS7, 99pc dc)	WLAN	8.50	± 9.6 %
10552	AAC	IEEE 802.11ac WIFI (80MHz, MCS8, 99pc dc)	WLAN	8.42	± 9.6 %
10553	AAC	IEEE 802.11ac WiFi (80MHz, MCS9, 99pc dc)	WLAN	.8.45	± 9.6 %
10554	AAD	IEEE 802.11ac WIFI (160MHz, MCS0, (89pc dc)	WLAN	8.48	± 9.6 %
10555	AAD	IEEE 802.11ac WiFi (160MHz, MCS1, 99pc dc)	WLAN	8.47	± 9.6 %
10556	AAD	IEEE 802.11ac WiFi (160MHz, MCS2, 99pc dc)	WLAN	8:50	± 9.6 %
10557	AAD	IEEE 802.11ac WiFi (160MHz, MCS3, 99pc dc)	WLAN	8.52	± 9.6 %
10558	AAD	IEEE 802.11ac WiFi (160MHz, MCS4, 99pc dc)	WLAN	8.61	± 9.6 %
10560	AAD	IEEE 802.11ac WiFi (160MHz, MCS6, 99pc dc)	WLAN	8.73	± 9.6 %
10561	AAD	IEEE 802.11ac WiFI (160MHz, MCS7, 99pc dc)	WLAN	8.56	± 9.6 %
10562	AAD	IEEE 802,11ac WiFi (160MHz, MCS8, 99pc dc)	WLAN	8.69	± 9.6 %
10563	AAD	IEEE 802,11ac WiFI (160MHz, MCSB, 99pc dc)	WLAN	8.77	± 9.6 %
10564	AAA	IEEE 802,11g WiFi 2.4 GHz (DSSS-OFDM, 9 Mbps, 99pc dc)	WLAN	8.25	± 9.6 %
10565	AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 12 Mbps, 99pc dc)	WLAN	8.45	± 9.6 %
10566	AAA	IEEE 802.11g WIFi 2.4 GHz (DSSS-OFDM, 18 Mbps, 99pc dc)	WLAN	8.13	± 9.6 %
10567	AAA	IEEE 802.11g WIFI 2.4 GHz (DSSS-OFDM, 24 Mbps, 99pc dc)	WLAN	8.00	± 9.6 %
10568	AAA	IEEE 802.11g WiFI 2.4 GHz (DSSS-OFDM, 36 Mbps, 99pc dc)	WLAN	8.37	± 9.6 %
10569	AAA	IEEE 802.11g WIFI 2.4 GHz (DSSS-OFDM, 48 Mbps, 99pc dc)	WLAN	8.10	± 9.6 %
10570	AAA	IEEE 802.11g WIFI 2.4 GHz (DSSS-OFDM, 54 Mbps, 99pc dc)	WLAN	8.30	± 9.6 %
10571	AAA	IEEE 802.11b WIFI 2.4 GHz (DSSS, 1 Mbps, 90pc dc)	WLAN	1.99	± 9.6 %
10572	AAA	IEEE 802.11b WiFi 2.4 GHz (DSSS, 2 Mbps, 90pc dc)	WLAN	1.99	± 9.6 %
10573	AAA	IEEE 802.11b WiFi 2.4 GHz (DSSS, 5.5 Mbps, 90pc dc)	WLAN	1.98	± 9.6 %
10574	AAA	IEEE 802.11b WIFI 2.4 GHz (DSSS, 0.0 Mips, superior)	WLAN	1.98	19.6 %
10575	AAA	IEEE 802.11p WIFI 2.4 GHz (DSSS-OFDM, 6 Mbps, 90pc dc)	WLAN	8.59	± 9.6 9
10576	AAA	IEEE 802.11g WFI 2.4 GHz (DSSS-OFDM, 9 Mbps, 90pc dc)	WLAN	8.60	± 9.6 %
10577	AAA	IEEE 802.11g WFI 2.4 GHz (DSSS-OFDM, 12 Mbps, 50pc dc)	WLAN	8.70	±9.69
10578	AAA	IEEE 802.11g WFI 2.4 GHz (DSSS-OFDM, 12 Mbps, 90pc dc)	WLAN	B.49	± 9.6 %
10579	AAA	IEEE 802.11g WIFI 2.4 GHz (DSSS-OFDM, 14 Mbps, solid dd)	WLAN	8.36	± 9.6 9
10580	AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 36 Mbps, 90pc dc)	WLAN	8.76	±9.6 %
10581	AAA	IEEE 802.11g WIFI 2.4 GHz (DSSS-OFDM, 48 Mbps, 90pc dc)	WLAN	8.35	± 9.6 %
10582	AAA	IEEE 802.11g WIFI 2.4 GHz (DSSS-OFDM, 54 Mbps, 90pc dc)	WLAN	8.67	± 9.6 %
10583	AAC	IEEE 802.11a/h WIFI 5 GHz (OFDM, 6 Mbps, 90pc dc)	WLAN	8.59	± 9.6 %
10584	AAC	IEEE 802,11a/h WIFI 5 GHz (OFDM, 9 Mbps, 90pc dc)	WLAN	8.60	± 9.6 %
10585	AAC	IEEE 802.11a/h WiFi 5 GHz (OFDM, 12 Mbps, 90pc dc)	WLAN	8.70	± 9.6 %
10586	AAC	IEEE 802.11a/h WIFI 5 GHz (OFDM, 18 Mbps, 90pc dc)	WLAN	8.49	± 9.6 %
10587	AAC	IEEE 802.11a/h WiFI 5 GHz (OFDM, 10 Mbps, 90pc dc)	WLAN	8.36	± 9.6 %
10588	AAC	IEEE 802.11a/h WiFi 5 GHz (OFDM, 36 Mbps, 90pc dc)	WLAN	8.76	± 9.6 %
10589	AAC	IEEE 802,11a/h WIFI 5 GHz (OFDM, 48 Mbps, 90pc dc)	WLAN	8.35	± 9.6 9
10589	AAC	IEEE 802, 11am WIFI 5 GHz (OFDM, 46 Mbps, 90pc dc) IEEE 802, 11am WIFI 5 GHz (OFDM, 54 Mbps, 90pc dc)	WLAN	8.67	± 9.6 %
10591	AAC	IEEE 802.11n (HT Mixed, 20MHz, MCS0, 90pc dc)	WLAN	8.63	± 9.6 %
10591	AAC	IEEE 802.11n (HT Mixed, 20MHz, MC30, 90pc dc)	WLAN	8,79	± 9.6 %
10592	AAC	IEEE 802.11n (HT Mixed, 20MHz, MCS1, 90pc dc)	WLAN	8.64	± 9.6 9
10593	AAC	IEEE 802.11n (HT Mixed, 20MHz, MCS2, 90pc dc)	WLAN	8.74	± 9.6 9
10594	AAC	IEEE 802.11n (HT Mixed, 20MHz, MCS3, 90pc dc)	WLAN	8.74	± 9.6 %
10596	AAC	IEEE 802.11n (HT Mixed, 20MHz, MCS4, 90pc dc)	WLAN	8.71	± 9.6 %
10596	AAC	IEEE 802.11n (HT Mixed, 20MHz, MCS6, 90pc dc)	WLAN	8.71	± 9.6 %
10597	AAC	IEEE 802.11n (H1 Mixed, 20MHz, MCS6, Bupc dc) IEEE 802.11n (HT Mixed, 20MHz, MCS7, 90pc dc)	WLAN	8.50	
		IEEE 802.11n (H1 Mixed, 20MHz, MCS1, 90pc dc)			± 9.6 9
10599	AAC		WLAN	8.79	± 9.6 9
10600	AAC	IEEE 802.11n (HT Mixed, 40MHz, MCS1, 90pc dc) IEEE 802.11n (HT Mixed, 40MHz, MCS2, 90pc dc)	WLAN	8.88	± 9.6 %
10501	AAC		WLAN	8.82	± 9.6 %
10802	AAC	IEEE 802.11n (HT Mixed, 40MHz, MCS3, 90pc dc)	WLAN	8.94	± 9.6 %
10603	AAC	IEEE 802.11n (HT Mixed, 40MHz, MCS4, 90pc dc)	WLAN	9.03	± 9.6 %
10604	AAC	IEEE 802.11n (HT Mixed, 40MHz, MCS5, 90pc dc)	WLAN	8.76	±9.6 %

Certificate No: EX3-7679_Sep21

Page 16 of 22





			1000	
H X 7	1111	r.a	-54D2	7679

0605	AAC	IEEE 802.11n (HT Mixed, 40MHz, MCS6, 90pc dc)	WLAN	8.97	± 9.6 %
3606	AAC	IEEE 802.11n (HT Mixed, 40MHz, MCS7, 90pc dc)	WLAN	8.82	± 9.6 %
0607	AAC	IEEE 802.11ac WiFi (20MHz, MCS0, 90pc dc)	WLAN	8.64	± 9.6 %
0608	AAC	IEEE 802.11ac WIFI (20MHz, MCS1, 90pc do)	WLAN	8.77	± 9.6 %
0609	AAC	IEEE 802.11ac WiFI (20MHz, MCS2, 90pc dc)	WLAN	8.57	±.9.6 %
0610	AAC	IEEE 802.11ac WIFI (20MHz, MCS3, 90pc dc)	WLAN	8.78	± 9.6 %
0611	AAC	IEEE 802.11ac WiFi (20MHz, MCS4, 90pc dc)	WLAN	8.70	± 9.6 %
0612	AAC	IEEE 802.11ac WiFi (20MHz, MCS5, 90pc dc)	WLAN	8.77	± 9.6 %
0613	AAC	IEEE 802.11ac WIFI (20MHz, MCS6, 90pc dc)	WLAN	8.94	± 9.6 %
0614	AAC	IEEE 802.11ac WiFi (20MHz, MCS7, 90pc dc)	WLAN	8.59	±9.6 %
0615	AAC	IEEE 802.11ac WIFI (20MHz, MCS8, 90pc dc)	WLAN	8.82	± 9.6 %
0616	AAC	IEEE 802.11ac WiFi (40MHz, MCS0, 90pc dc)	WLAN	8.82	± 9.6 %
0617	AAC	IEEE 802.11ac WIFI (40MHz, MCS1, 90pc dc)	WLAN	8.81	± 9.6 %
0618	AAC	IEEE 802.11ac WiFI (40MHz, MCS2, 90pc dc)	WLAN	8.58	±9.6%
0619	AAC	IEEE 802.11ac WIFI (40MHz, MCS3, 90pc dc)	WLAN	8.86	± 9.6 %
0620	AAC	IEEE 802,11ac WIFI (40MHz, MCS4, 90pc dc)	WLAN	8.87	± 9.6 %
0621	AAC	IEEE 802.11ac WIFI (40MHz, MCS5, 90pc dc)	WLAN	8.77	± 9.6 %
0622	AAC	IEEE 802.11ac WiFi (40MHz, MCS6, 90pc dc)	WLAN	8.68	± 9.6 %
0623	AAC	IEEE 802.11ac WiFi (40MHz, MCS7, 90pc dc)	WLAN	8.82	± 9.6 %
0624	AAC	IEEE 802.11ac WiFi (40MHz, MCS8, 90pc dc)	WLAN	8.96	± 9.6 %
0625	AAC	IEEE 802.11ac WiFi (40MHz, MCS9, 90pc dc)	WLAN	8.96	± 9.6 %
0626	AAC	IEEE 802.11ac WiFi (80MHz, MCS0, 90pc dc)	WLAN	8.83	±9.6 %
0627	AAC	IEEE 802.11ac WIFI (80MHz, MCS1, 90pc dc)	WLAN	8.88	± 9.6 %
0628	AAC	IEEE 802.11ac WiFi (80MHz, MCS2, 90pc dc)	WLAN	8.71	± 9.6 %
0629	AAC	IEEE 802.11ac WIFi (80MHz, MCS3, 90pc dc)	WLAN	8.85	±9.6 %
0630	AAC	IEEE 802.11ac WiFi (80MHz, MCS4, 90pc dc)	WLAN	8.72	± 9.6 %
0631	AAC	IEEE 802.11ac WiFi (80MHz, MCS5, 90pc dc)	WLAN	8.81	±9.6%
0632	AAC	IEEE 802.11ac WiFi (80MHz, MCS8, 90pc dc)	WLAN	8.74	± 9.6 %
0633	AAC	IEEE 802.11ac WIFI (80MHz, MCS7, 90pc dc)	WLAN	8.83	± 9.6 %
0634	AAC	IEEE 802,11ac WIFI (80MHz, MCS8, 90pc dc)	WLAN	8.80	±9.6 %
0635	AAC	IEEE 802.11ac WIFi (80MHz, MCS9, 90pc dc)	WLAN	8.81	± 9.6 %
0636	AAD	IEEE 802.11ac WIFI (160MHz, MCS0, 90pc dc)	WLAN	8.83	± 9.6 %
0637	AAD	IEEE 802.11ac WiFi (160MHz, MCS1, 90pc dc)	WLAN	8.79	± 9.6 %
0638	AAD	IEEE 802.11ac WIFI (160MHz, MCS2, 90pc dc)	WLAN	8.86	± 9.6 %
0639	AAD	IEEE 802.11ac WiFi (160MHz, MCS3, 90pc dc)	WLAN	8.85	± 9.6 %
0640	AAD	IEEE 802.11ac WIFI (160MHz, MCS4, 90pc dc)	WLAN	8.98	± 9.6 %
0641	AAD	IEEE 802.11ac WIFI (160MHz, MCS5, 90pc dc)	WLAN	9.06	+9.6 %
0642	AAD	IEEE 802.11ac WIFI (160MHz, MCS6, 90pc dc)	WLAN	9.06	±9.6 %
0643	AAD	IEEE 802.11ac WIFI (160MHz, MC57, 90pc dc)	WLAN	8.89	± 9.6 %
0644	AAD	IEEE 802.11ac WIFI (160MHz, MCS8, 90pc dc)	WLAN	9.05	±9.6%
0645	AAD	IEEE 802.11ac WiFi (160MHz, MCS9, 90pc dc)	WLAN	9.11	± 9.6 %
0646	AAG	LTE-TDD (SC-FDMA, 1 R8, 5 MHz, QPSK, UL Sub#2,7)	LTE-TDD	11.96	±9.6%
0647	AAF	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, QPSK, UL Sub=2,7)	LTE-TDD	11.96	± 9.6 %
0648	AAA	CDMA2000 (1x Advanced)	CDMA2000	3.45	±9.6%
0652	AAE	LTE-TDD (OFDMA, 5 MHz, E-TM 3.1, Clipping 44%)	LTE-TDD	6.91	± 9.6 %
0653		LTE-TDD (OFDMA, 10 MHz, E-TM 3.1, Clipping 44%)	LTE-TDD	7.42	± 9.6 %
0654	AAD	LTE-TDD (OFDMA, 15 MHz, E-TM 3.1, Clipping 44%)	LTE-TDD	6.96	± 9.6 %
	AAE	LTE-TDD (OFDMA, 20 MHz, E-TM 3.1, Clipping 44%)	LTE-TDD	7.21	± 9.6 %
0658	AAA	Pulse Waveform (200Hz, 10%)	Test	10.00	± 9.6 %
0659	AAA	Pulse Waveform (200Hz, 20%)	Test	6.99	± 9.6 %
0660	AAA	Pulse Waveform (200Hz, 40%)	Test	3.98	±9.6 %
0661	AAA	Pulse Waveform (200Hz, 60%)	Test	2.22	± 9.6 %
0662	AAA	Pulse Waveform (200Hz, 80%)	Test	0.97	± 9.6 %
0670	AAA	Bluetooth Low Energy	Bluetooth	2.19	± 9.6 %
0671	AAC	IEEE 802.11ax (20MHz, MCS0, 90pc dc)	WLAN	9,09	± 9.6 %

Certificate No: EX3-7679_Sep21

Page 17 of 22



EX3DV4- SN:7679

10673	AAC	IEEE 802.11ax (20MHz, MCS2, 90pc dc)	WLAN	8.78	± 9.6 %
10674	AAC	IEEE 802,11ax (20MHz, MCS3, 90pc dc)	WLAN	8.74	± 9.6 %
0675	AAC	IEEE 802.11ax (20MHz, MCS4, 90pc dc)	WLAN	8.90	± 9.6 %
0676	AAC	IEEE 802.11ax (20MHz, MCS5, 90pc dc)	WLAN	8.77	± 9.6 %
0677	AAC	IEEE 802.11ax (20MHz, MCS6, 90pc dc)	WLAN	8.73	± 9.6 %
0678	AAC	IEEE 802.11ax (20MHz, MCS7, 90pc dc)	WLAN	8,78	± 9.6 %
0679	AAC	IEEE 802.11ax (20MHz, MCS8, 90pc dc)	WLAN	8.89	± 9.6 %
0680	AAC	IEEE 802.118x (20MHz, MCS9, 90pc dc)	WLAN	8.80	± 9.6 %
10681	AAC	IEEE 802.11ax (20MHz, MCS10, 90pc dc)	WLAN	8.62	± 9.6 %
0682	AAC	IEEE 802,11ax (20MHz, MCS11, 90pc dc)	WLAN	8.83	±9.6 %
0683	AAC	IEEE 802.11ax (20MHz, MCS0, 99pc dc)	WLAN	8.42	± 9.6 %
0684	AAC	IEEE 802.11ax (20MHz, MCS1, 99pc dc)	WLAN	8.26	± 9.6 %
10685	AAC	IEEE 802.11ax (20MHz, MCS2, 99pc dc)	WLAN	8.33	± 9.6 %
0686	AAC	IEEE 802.11ax (20MHz, MCS3, 99pc do)	WLAN	8.28	± 9.6 %
0687	AAC	IEEE 802.11ax (20MHz, MCS4, 99pc dc)	WLAN	8.45	± 9.6 %
88301	AAC	IEEE 802.11ax (20MHz, MCS5, 99pc do)	WLAN	8.29	± 9.6 %
10689	AAC	IEEE 802.11ax (20MHz, MCS6, 99pc do)	WLAN	8.55	± 9.6 %
0690	AAC	IEEE 802.11ax (20MHz, MCS7, 99pc dc)	WLAN	8.29	± 9.6 %
10691	AAC	IEEE 802.11ax (20MHz, MCS8, 99pc dc)	WLAN	8.25	± 9.6 %
0692	AAC	IEEE 802.11ax (20MHz, MCS9, 99pc dc)	WLAN	B,29	± 9.6 %
10693	AAC	IEEE 802.11ax (20MHz, MCS10, 99pc dc)	WLAN	8.25	± 9.6 %
10694	AAC	IEEE 802.11ax (20MHz, MCS11, 99pc dc)	WLAN	8.57	± 9.6 %
10695	AAC	IEEE 802.11ax (40MHz, MCS0, 90pc dc)	WLAN	8.78	±9.6 %
10696	AAC	IEEE 802.11ax (40MHz, MCS1, 90pc dc)	WLAN	8.91	± 9.6 %
10897	AAC	IEEE 802.11ax (40MHz, MCS2, 90pc dc)	WLAN	8.61	± 9.6 %
10698	AAC	IEEE 802.11ax (40MHz, MCS3, 90pc dc)	WLAN	8.89	± 9.6 %
10699	AAC	IEEE 802.11ax (40MHz, MCS4, 90pc do)	WLAN	8.82	± 9.6 %
10700	AAC	IEEE 802.11ax (40MHz, MCS5, 90pc dc)	WLAN	8.73	± 9.6 %
10701	AAC	IEEE 802.11ax (40MHz, MCS6, 90pc dc)	WLAN	8.86	± 9.6 %
10702	AAC.	IEEE 802.11ax (40MHz, MCS7, 90pc dc)	WLAN	8.70	± 9.6 %
10703	AAC	IEEE 802.11ax (40MHz, MCS8, 90pc.dc)	WLAN	8.82	± 9,6 %
10704	AAC	IEEE 802.11ax (40MHz, MCS9, 90pc dc)	WLAN	8.56	± 9.6.%
10705	AAC	IEEE 802.11ax (40MHz, MCS10, 90pc dc)	WLAN	8.69	± 9.6 %
10706	AAC	IEEE 802.11ax (40MHz, MCS11, 90pc dc)	WLAN	8.66	± 9,6 %
10707	AAC	IEEE 802.11ax (40MHz, MCS0, 99pc dc)	WLAN	8.32	± 9.6 %
10708	AAC	IEEE 602.11ax (40MHz, MCS1, 99pc dc)	WLAN	8.55	± 9.6 %
10709	AAC	IEEE 802.11ax (40MHz, MCS2, 99pc dc)	WLAN	8.33	± 9.6 %
10710	AAC	IEEE 802.11ax (40MHz, MCS3, 99pc dc)	WLAN	8.29	± 9.6 %
10711	AAC	IEEE 802.11ax (40MHz, MCS4, 99pc dc)	WLAN	8.39	± 9.6 %
10712	AAC	IEEE 802.11ax (40MHz, MCS5, 99pc dc)	WLAN	8.67	± 9.6 %
10713	AAC	IEEE 802.11ax (40MHz, MCS6, 99pc dc)	WLAN	8.33	± 9.6 %
10714	And a subsection of the local diversion of th	IEEE 802.11ax (40MHz, MCS7, 99pc dc)	WLAN	8.26	± 9.6 %
10715	AAC	IEEE 802.11ax (40MHz, MCS8, 99pc dc)	WLAN	8.45	± 9.6 %
10716		IEEE 802.11ax (40MHz, MCS9, 99pc dc)	WLAN	8.30	±9.6 %
10717		IEEE 802.11ax (40MHz, MCS10, 99pc dc)	WLAN	8.48	± 9.6 %
10718	and the second second second	IEEE 802.11ax (40MHz, MCS11, 99pc dc)	WLAN	8.24	± 9.6 %
10719		IEEE 802.11ax (80MHz, MCS0, 90pc dc)	WLAN	8.81	± 9.6 %
	AAC	IEEE 802.11ax (80MHz, MCS1, 90pc dc)	WLAN	8.87	19.6 %
10721	and the second se	IEEE 802.11ax (80MHz, MCS2, 90pc dc)	WLAN	8.76	± 9.6 %
10722		IEEE 802.11ax (80MHz, MCS3, 90pc dc)	WLAN	B.55	± 9.6 %
10723		IEEE 802.11ax (80MHz, MCS4, 90pc dc)	WLAN	8.70	± 9.6 %
10724		IEEE 802.11ax (80MHz, MCS5, 90pc dc)	WLAN	8,90	± 9.6 9
10725		IEEE 802.11ax (80MHz, MCS6, 90pc dc)	WLAN	8.74	± 9.6 %
10726	AAC	IEEE 802.11ax (80MHz, MCS7, 90pc dc)	WLAN	8.72	+9.69
10727		IEEE 802.11ax (80MHz, MCS8, 90pc dc)	WLAN	8.66	± 9.6 %

Certificate No: EX3-7679_Sep21

Page 18 of 22



123.00	in	14	COM:	7070
ヒスル	หมา	14	212	7679

0729	AAC	IEEE 802.11ax (80MHz, MCS10, 90pc dc)	WLAN	8.64	± 9.6 %
0730	AAC	IEEE 802.11ax (80MHz, MCS11, 90pc dc)	WLAN	8.67	± 9.6 %
0731	AAC	IEEE 802.11ax (80MHz, MCS0, 99pc dc)	WLAN	8.42	± 9.6 %
0732	AAC	IEEE 802.11ax (80MHz, MCS1, 99pc dc)	WLAN	8.46	± 9.6 %
0733	AAC	IEEE 802.11ax (80MHz, MCS2, 99pc dc)	WLAN	8.40	± 9.6 %
0734	AAC	IEEE 802.11ax (80MHz, MC53, 99pc dc)	WLAN	8.25	±9.6 %
0735	AAC	IEEE 802.11ax (80MHz, MCS4, 99pc dc)	WLAN	8.33	± 9.6 %
0736	AAC	IEEE 802.11ax (80MHz, MCS5, 99pc dc)	WLAN	8.27	± 9.6 %
0737	AAC	IEEE 802.11ax (80MHz, MCS6, 99pc dc)	WLAN	8.36	±9.6 %
0738	AAC	IEEE 802.11ax (80MHz, MCS7, 99pc dc)	WLAN	8.42	± 9.6 %
0739	AAC	IEEE 802.11ax (80MHz, MCS8, 99pc dc)	WLAN	8.29	±9.6%
0740	AAC	IEEE 802.11ax (80MHz, MCS9, 99pc dc)	WLAN	8.48	± 9.6 %
0741	AAC	IEEE 802.11ax (80MHz, MCS10, 99pc dc)	WLAN	8.40	± 9.6 %
3742	AAC	IEEE 802.11ax (80MHz, MCS11, 99pc dc)	WLAN	8.43	±9.6 %
0743	AAC	IEEE 802.11sx (160MHz, MCS0, 90pc dc)	WLAN	8.94	± 9.6 %
0744	AAC	IEEE 802.11ax (160MHz, MCS1, 90pc dc)	WLAN	9.16	± 9.6 %
0745	AAC	IEEE 802.11ax (160MHz, MCS2, 90pc dc)	WLAN	8.93	± 9.6 %
0746	AAC	IEEE 802.11ax (160MHz, MCS3, 90pc dc)	WLAN	9.11	± 9.6 %
0747	AAC	IEEE 802.11ax (180MHz, MCS4, 90pc dc)	WLAN	9.04	± 9.6 %
0748	AAC	IEEE 802.11ax (160MHz, MCS5, 90pc dc)	WLAN	8.93	± 9.6 %
0749	AAC	IEEE 802.11ax (160MHz, MCS6, 90pc dc)	WLAN	8.90	± 9.6 %
0750	AAC	IEEE 802.11ax (160MHz, MCS7, 90pc dc)	WLAN	8.79	± 9.6 %
0751	AAC	IEEE 802.11ax (160MHz, MCS8, 90pc dc)	WLAN	8.82	±9.6 %
0752	AAC	IEEE 802.11ax (160MHz, MCS9, 90pc dc)	WLAN	8.81	±9.6 %
0753	AAC	IEEE 802.11ax (160MHz, MCS10, 90pc dc)	WLAN	9.00	±9.6 %
0754	AAC	IEEE 802.11ax (160MHz, MCS11, 90pc dc)	WLAN	8.94	± 9.6 %
0755	AAC	IEEE 802.11ax (160MHz, MCS0, 99pc dc)	WLAN	8.64	± 9.6 %
0756	AAC	IEEE 802.11ax (160MHz, MCS1, 99pc dc)	WLAN	8.77	± 9.6 %
0757	AAC	IEEE 802.11ax (160MHz, MCS2, 99pc dc)	WLAN	8.77	± 9.6 %
0758	AAC	IEEE 802.11ax (160MHz, MCS3, 99pc dc)	WLAN	8.69	± 9.6 %
0759	AAC	IEEE 802.11ax (160MHz, MCS4, 99pc dc)	WLAN	8.58	± 9.6 %
0760	AAC	IEEE 802.11ax (160MHz, MCS5, 99pc dc)	WLAN	8.49	± 9.6 %
0761	AAC	IEEE 802.11ax (160MHz, MCS6, 99pc dc)	WLAN	8.58	± 9.6 %
0762	AAC	IEEE 802.11ax (160MHz, MCS7, 99pc dc)	WLAN	8.49	± 9.6 %
0763	AAC	IEEE 802.11ax (160MHz, MCS8, 99pc dc)	WLAN	8.53	± 9.6 %
0764	AAC	IEEE 802.11ax (160MHz, MCS9, 99pc dc)	WLAN	8.54	± 9.6 %
0765	AAC	IEEE 802.11ax (160MHz, MCS10, 99pc dc)	WLAN	8.54	± 9.6 %
0766	AAC	IEEE 802.11ax (160MHz, MCS11, 99pc dc)	WLAN	8.51	± 9.6 %
0767	AAE	5G NR (CP-OFDM, 1 RB, 5 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	7.99	± 9.6 %
0768	AAD	5G NR (CP-OFDM, 1 RB, 10 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.01	± 9.6 %
0769	AAD	5G NR (CP-OFDM, 1 RB, 15 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.01	± 9.6 %
0770	AAD	5G NR (CP-OFDM, 1 RB, 20 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.02	± 9.6 %
0771	AAD	5G NR (CP-OFDM, 1 RB, 25 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.02	± 9.6 %
0772	AAD	SG NR (CP-OFDM, 1 RB, 30 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.23	± 9.6 %
0773	AAD	5G NR (CP-OFDM, 1 RB, 40 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.03	± 9.6 %
0774	AAD	5G NR (CP-OFDM, 1 RB, 50 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.02	± 9.6 %
0775	AAD	5G NR (CP-OFDM, 50% RB, 5 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.31	± 9.6 %
0776	AAD	5G NR (CP-OFDM, 50% RB, 10 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.30	± 9.6 %
0777	AAC	5G NR (CP-OFDM, 50% RB, 15 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.30	± 9.6 %
0778	AAD	5G NR (CP-OFDM, 50% RB, 20 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.34	± 9.6 %
0779	AAC	5G NR (CP-OFDM, 50% RB, 25 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.42	± 9.6 %
0780	AAD	5G NR (CP-OFDM, 50% RB, 30 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.38	± 9.6 %
0781	AAD	5G NR (CP-OFDM, 50% RB, 40 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.38	± 9.6 %
0782	AAD	5G NR (CP-OFDM, 50% RB, 50 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.43	± 9.6 %
0783	AAE	5G NR (CP-OFDM, 100% RB, 5 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.31	± 9.6 %
0784	AAD	5G NR (CP-OFDM, 100% RB, 10 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.29	± 9.6 %

Certificate No: EX3-7679_Sep21

Page 19 of 22



EX30V4- SN:7679

September 10, 2021

10785	AAD	5G NR (CP-OFDM, 100% RB, 15 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.40	± 9.6 %
10786	AAD	5G NR (CP-OFDM, 100% RB, 20 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.35	±9.6%
10787	AAD	5G NR (CP-OFDM, 100% RB, 25 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.44	± 9.6 %
10788	AAD	5G NR (CP-OFDM, 100% RB, 30 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.39	± 9.6 %
10789	AAD	5G NR (CP-OFDM, 100% RB, 40 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.37	± 9.6 %
10790	AAD	5G NR (CP-OFDM, 100% RB, 50 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.39	± 9.6 %
10791	AAE	5G NR (CP-OFDM, 1 RB, 5 MHz, QP5K, 30 kHz)	5G NR FR1 TDD	7.83	± 9.6 %
10792	AAD	5G NR (CP-OFDM, 1 RB, 10 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	7.92	± 9.6 %
10793	AAD	5G NR (CP-OFDM, 1 RB, 15 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	7.95	± 9.6 %
10794	AAD	5G NR (CP-OFDM, 1 RB, 20 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	7.82	± 9.6 %
10795	AAD	5G NR (CP-OFDM, 1 R8, 25 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	7.84	± 9.6 %
10796	AAD	5G NR (CP-OFDM, 1 RB, 30 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	7.82	±9.6 %
10797	AAD	5G NR (CP-OFDM, 1 RB, 40 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.01	± 9.6 %
10798	AAD	5G NR (CP-OFDM, 1 RB, 50 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	7.89	± 9.6 %
10799	AAD	5G NR (CP-OFDM, 1 RB, 60 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	7.93	± 9.6 %
10801	AAD	5G NR (CP-OFDM, 1 RB, 80 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	7.89	± 9.6 %
10802	AAD	5G NR (CP-OFDM, 1 RB, 90 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	7.87	±9.69
10803	AAD	5G NR (CP-OFDM, 1 RB, 100 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	7.93	±9.6 %
10805	AAD	5G NR (CP-OFDM, 50% RB, 10 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.34	± 9.6 %
10806	AAD	5G NR (CP-OFDM, 50% RB, 15 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.37	± 9.6 %
10809	AAD	5G NR (CP-OFDM, 50% RB, 30 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.34	± 9.6 %
10810	AAD	5G NR (CP-OFDM, 50% RB, 40 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.34	± 9.6 %
10812	AAD	5G NR (CP-OFDM, 50% RB, 60 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.35	± 9.6 %
10817	AAE	5G NR (CP-OFDM, 100% RB, 5 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.35	±9.6 %
10818	AAD	5G NR (CP-OFDM, 100% RB, 10 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.34	± 9.6 %
10819	AAD	5G NR (CP-OFDM, 100% RB, 15 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.33	±9.63
10820	AAD	5G NR (CP-OFDM, 100% RB, 20 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.30	± 9.6 %
10821	AAD	5G NR (CP-OFDM, 100% RB, 25 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.41	± 9.6.9
10822	and the second second	5G NR (CP-OFDM, 100% RB, 30 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	B.41	±9.6 %
10823		5G NR (CP-OFDM, 100% RB, 40 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.36	± 9.6 %
10824	AAD	5G NR (CP-OFDM, 100% RB, 50 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.39	± 9.6 9
10825	AAD	5G NR (CP-OFDM, 100% RB, 60 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.41	± 9.6 9
10827	AAD	5G NR (CP-OFDM, 100% RB, 80 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.42	± 9.6.9
10828	AAD	5G NR (CP-OFDM, 100% RB, 90 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.43	± 9.6 %
10629	AAD	5G NR (CP-OFDM, 100% RB, 100 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	B.40	± 9.6 %
10830	and the second second	5G NR (CP-OFDM, 1 RB, 10 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	7.63	± 9.6 %
10831	AAD	5G NR (CP-OFDM, 1 RB, 15 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	7.73	± 9.6 %
10832	AAD	5G NR (CP-OFDM, 1 RB, 20 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	7.74	± 9.6 %
10633		5G NR (CP-OFDM, 1 RB, 25 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	7.70	± 9.6.9
10834	AAD	5G NR (CP-OFDM, 1 RB, 30 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	7.75	± 9.6 3
10835	AAD	5G NR (CP-OFDM, 1 RB, 40 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	7.70	± 9.6 9
10836	AAD	5G NR (CP-OFDM, 1 RB, 50 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	7.66	± 9.6 3
10837	AAD	5G NR (CP-OFDM, 1 RB, 60 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	7.68	# 9.6 9
10839	AAD	5G NR (CP-OFDM, 1 RB, 60 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	7.70	± 9.6 3
10840	AAD	5G NR (CP-OFDM, 1 RB, 90 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	7.67	± 9.6 5
10841	AAD	5G NR (CP-OFDM, 1 RB, 100 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	7.71	29.6 9
10843		5G NR (CP-OFDM, 50% RB, 15 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	8.49	±9.65
10844	AAD	5G NR (CP-OFDM, 50% RB, 20 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	8.34	± 9.6 %
10846	AAD	5G NR (CP-OFDM, 50% RB, 30 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	8,41	± 9.6 %
10854	AAD	5G NR (CP-OFDM, 100% RB, 10 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	8.34	± 9.6 5
10855	AAD	5G NR (CP-OFDM, 100% RB, 15 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	8.36	± 9.6 5
10856	AAD	5G NR (CP-OFDM, 100% RB, 20 MHz, QP5K, 60 kHz)	5G NR FR1 TDD	8.37	±9.65
10857	AAD	5G NR (CP-OFDM, 100% R8, 25 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	8.35	± 9.6 %
10858	AAD	5G NR (CP-OFDM, 100% RB, 30 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	8.36	± 9.6 %
10859	AAD	5G NR (CP-OFDM, 100% RB, 40 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	8.34	±9.6 5
10860	AAD	5G NR (CP-OFDM, 100% RB, 50 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	8.41	± 9.6 1

Certificate No: EX3-7679_Sep21

Page 20 of 22



EX3DV4-- SN:7679

September 10, 2021

10861	AAD	5G NR (CP-OFDM, 100% RB, 60 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	8.40	± 9.6 %
10863	AAD	5G NR (CP-OFDM, 100% RB, 80 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	8.41	±9.6 %
10864	AAD	5G NR (CP-OFDM, 100% RB, 90 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	8.37	± 9.6 %
10865	AAD	5G NR (CP-OFDM, 100% RB, 100 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	8,41	± 9.6 %
10866	AAD	5G NR (DFT-s-OFDM, 1 RB, 100 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.68	±9.6 %
10868	AAD	5G NR (DFT-s-OFDM, 100% RB, 100 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.89	± 9.6 %
10869	AAD	5G NR (DFT-5-OFDM, 1 RB, 100 MHz, QPSK, 120 kHz)	5G NR FR2 TDD	5.75	± 9.6 %
10870	AAD	5G NR (DFT-s-OFDM, 100% RB, 100 MHz, QPSK, 120 kHz)	5G NR FR2 TDD	5.86	± 9.6 %
10871	AAD	5G NR (DFT-8-OFDM, 1 RB, 100 MHz, 16QAM, 120 kHz)	5G NR FR2 TDD	5.75	± 9.6 %
10872	AAD	5G NR (DFT-6-OFDM, 100% RB, 100 MHz, 16QAM, 120 kHz)	5G NR FR2 TDD	6.52	± 9.6 %
10873	AAD	5G NR (DFT-s-OFDM, 1 RB, 100 MHz, 64QAM, 120 kHz)	5G NR FR2 TDD	6.61	± 9.6 %
10874	AAD	5G NR (DFT-s-OFDM, 100% RB, 100 MHz, 64QAM, 120 kHz)	5G NR FR2 TDD	6.65	±9.6 %
10875	AAD	5G NR (CP-OFDM, 1 RB, 100 MHz, QPSK, 120 kHz)	5G NR FR2 TDD	7.78	± 9.6 %
10876	AAD	5G NR (CP-OFDM, 100% RB, 100 MHz, QPSK, 120 kHz)	5G NR FR2 TDD	8.39	±9.6 %
10877	AAD	5G NR (CP-OFDM, 1 RB, 100 MHz, 18QAM, 120 kHz)	5G NR FR2 TDD	7.95	± 9.6 %
10878	AAD	5G NR (CP-OFDM, 100% RB, 100 MHz, 16QAM, 120 kHz)	5G NR FR2 TDD	8.41	±9.6 %
10879	AAD	5G NR (CP-OFDM, 1 RB, 100 MHz, 64QAM, 120 kHz)	5G NR FR2 TDD	8.12	± 9.6 %
10880	AAD	5G NR (CP-OFDM, 100% RB, 100 MHz, 64QAM, 120 kHz)	5G NR FR2 TDD	8.38	± 9.6 %
10881	AAD	5G NR (DFT-s-OFDM, 1 RB, 50 MHz, QPSK, 120 kHz)	5G NR FR2 TDD	5.75	±9.6 %
10882	AAD	5G NR (DFT-s-OFDM, 100% RB, 50 MHz, QPSK, 120 kHz)	5G NR FR2 TDD	5.96	± 9.6 %
10883	AAD	5G NR (DFT-s-OFDM, 1 RB, 50 MHz, 16QAM, 120 kHz)	5G NR FR2 TDD	6.57	± 9.6 %
10884	AAD	5G NR (DFT-s-OFDM, 100% RB, 50 MHz, 16QAM, 120 kHz)	5G NR FR2 TDD	6.53	± 9.6 %
10885	AAD	5G NR (DFT-s-OFDM, 1 RB, 50 MHz, 64QAM, 120 kHz)	5G NR FR2 TDD	6.61	± 9.6 %
10886	AAD	5G NR (DFT-s-OFDM, 100% RB, 50 MHz, 64QAM, 120 kHz)	5G NR FR2 TDD	6.65	19.6 %
10887	AAD	5G NR (CP-OFDM, 1 RB, 50 MHz, QPSK, 120 kHz)	5G NR FR2 TDD	7.78	± 9.6 %
10888	AAD	5G NR (CP-OFDM, 100% RB, 50 MHz, QPSK, 120 kHz)	5G NR FR2 TDD	8.35	± 9.6 %
10889	AAD	5G NR (CP-OFDM, 1 RB, 50 MHz, 16QAM, 120 kHz)	5G NR FR2 TDD	8.02	± 9.6 5
10890	AAD	5G NR (CP-OFDM, 100% RB, 50 MHz, 160AM, 120 kHz)	5G NR FR2 TDD	8.40	±9.6 %
10891	AAD	5G NR (CP-OFDM, 1 RB, 50 MHz, 64QAM, 120 kHz)	5G NR FR2 TDD	8.13	± 9.6 %
10892	AAD	5G NR (CP-OFDM, 100% RB, 50 MHz, 64QAM, 120 kHz)	5G NR FR2 TDD	8.41	± 9.6 %
10897	AAC	5G NR (DFT-s-OFDM, 1 RB, 5 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.66	± 9.6 9
10898	AAB	5G NR (DFT-s-OFDM, 1 RB, 10 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.67	± 9.6 %
10899	AAB	5G NR (DFT-s-OFDM, 1 RB, 15 MHz, QP5K, 30 kHz)	5G NR FR1 TDD	5.67	± 9.6 %
10900	AAB	5G NR (DFT-s-OFDM, 1 RB, 20 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.68	± 9.6 %
10901	AAB	5G NR (DFT-s-OFDM, 1 RB, 25 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.68	± 9.6 %
10902	AAB	5G NR (DFT-s-OFDM, 1 RB, 30 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.68	± 9.6 %
10903	AAB	5G NR (DFT-s-OFDM, 1 RB, 40 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.68	± 9.6 9
10904	AAB	5G NR (DFT-s-OFDM, 1 RB, 50 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.68	± 9.6 %
10905	AAB	5G NR (DFT-s-OFDM, 1 RB, 60 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.68	± 9.6 %
10906	AAB	5G NR (DFT-s-OFDM, 1 RB, 60 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.68	± 9.6 %
10907	AAC	5G NR (DFT-8-OFDM, 50% RB, 5 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.78	± 9.6 %
10908	AAB	5G NR (DFT-8-OFDM, 50% RB, 10 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.93	± 9.6 9
10909	AAB	5G NR (DFT-s-OFDM, 50% RB, 15 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.96	± 9.6 %
10910	and the second second	5G NR (DFT-s-OFDM, 50% RB, 20 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.83	± 9.6 9
10911	AAB	5G NR (DFT-s-OFDM, 50% RB, 25 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.93	± 9.6 %
10912	-	5G NR (DFT-s-OFDM, 50% RB, 30 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.84	± 9.6 9
10913		5G NR (DFT-s-OFDM, 50% RB, 40 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.84	± 9.6 %
10914		5G NR (DFT-s-OFDM, 50% RB, 50 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.85	± 9.6 %
10915	AAB	5G NR (DFT-8-OFDM, 50% RB, 60 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.83	± 9.6 1
10916	And in case of the local division of	5G NR (DFT-s-OFDM, 50% RB, 80 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.87	± 9.6 1
10917	-	5G NR (DFT-s-OFDM, 50% RB, 100 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.94	± 9.6 %
10918		5G NR (DFT-s-OFDM, 100% RB, 5 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.86	± 9.6 *
10919		5G NR (DFT-s-OFDM, 100% RB, 10 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.86	± 9.6 5
10920		5G NR (DFT-s-OFDM, 100% RB, 15 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.87	± 9.6 9
10921	AAB	5G NR (DFT-s-OFDM, 100% RB, 20 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.84	± 9.6 %
10922	AAB	5G NR (DFT-s-OFDM, 100% RB, 25 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.82	± 9.6 %

Certificate No: EX3-7679_Sep21

Page 21 of 22





EX3DV4- SN:7679

September 10, 2021

0923	AAB	5G NR (DFT-5-OFDM, 100% RB, 30 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5,84	±9.6 %
10924	AAB	5G NR (DFT-s-OFDM, 100% RB, 40 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.84	± 9.6 %
10925	AAB	5G NR (DFT-s-OFDM, 100% RB, 50 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.95	± 9.6 %
10926	AAB	5G NR (DFT-5-OFDM, 100% RB, 60 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.84	± 9.6 %
10927	AAB	5G NR (DFT-5-OFDM, 100% RB, 80 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.94	± 9.6 %
10928	AAC	5G NR (DFT-s-OFDM, 1 RB, 5 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.52	± 9.6 %
10929	AAC	5G NR (DFT-s-OFDM, 1 RB, 10 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.52	± 9.6 %
10930	AAC	5G NR (DFT-s-OFDM, 1 RB, 15 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.52	± 9.6 %
10931	AAC	5G NR (DFT-s-OFDM, 1 RB, 20 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.51	± 9.6 %
10932	AAC	5G NR (DFT-s-OFDM, 1 RB, 25 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.51	± 9.6 %
10933	AAC	5G NR (DFT-s-OFDM, 1 RB, 30 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.51	± 9.6 %
10934	AAC	5G NR (DFT-s-OFDM, 1 RB, 40 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.51	± 9.6 %
10935	AAD	5G NR (DFT-s-OFDM, 1 RB, 50 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.51	± 9.6 %
10936	AAC	5G NR (DFT-s-OFDM, 50% RB, 5 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.90	±9.6.9
10937	AAC	5G NR (DFT-s-OFDM, 50% RB, 10 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.77	± 9.6.5
10938	AAC	5G NR (DFT-s-OFDM, 50% RB, 15 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.90	± 9.6 %
10939	AAC	5G NR (DFT-s-OFDM, 50% RB, 20 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.82	± 9.6 %
10940	AAC	5G NR (DFT-s-OFDM, 50% RB, 25 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.89	± 9.6 5
10941	AAC	5G NR (DFT-s-OFDM, 50% RB, 30 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.83	±9.6 %
10942	AAC	5G NR (DFT-s-OFDM, 50% RB, 40 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.85	±9.65
10943	AAD	5G NR (DFT-s-OFDM, 50% RB, 50 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.95	±9.6 %
10944	AAC	5G NR (DFT-s-OFDM, 100% RB, 5 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.81	±9.6 %
10945	AAC	5G NR (DFT-s-OFDM, 100% RB, 10 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.85	±9.6%
10946	AAC	5G NR (DFT-6-OFDM, 100% RB, 15 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.83	± 9.6 1
10947	AAC	5G NR (DFT-s-OFDM, 100% RB, 20 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.87	±9.6 %
10948	AAC	5G NR (DFT-s-OFDM, 100% RB, 25 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.94	± 9.6 9
10949	AAC	5G NR (DFT-s-OFDM, 100% RB, 30 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.87	± 9.6 9
10950	AAC	5G NR (DFT-s-OFDM, 100% RB, 40 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.94	± 9.6 %
10951	AAD	5G NR (DFT-s-OFDM, 100% RB, 50 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.92	± 9.6 %
10952	AAA	5G NR DL (CP-OFDM, TM 3.1, 5 MHz, 64-QAM, 15 kHz)	5G NR FR1 FDD	8.25	± 9.6 9
10953	AAA	5G NR DL (CP-OFDM, TM 3.1, 10 MHz, 64-QAM, 15 kHz)	5G NR FR1 FDD	8.15	± 9.6.9
10954	AAA	5G NR DL (CP-OFDM, TM 3.1, 15 MHz, 64-QAM, 15 kHz)	5G NR FR1 FDD	8.23	± 9.6.9
10955	AAA	5G NR DL (CP-OFDM, TM 3.1, 20 MHz, 64-QAM, 15 kHz)	5G NR FR1 FDD	8.42	± 9.6 1
10956	AAA	5G NR DL (CP-OFDM, TM 3.1, 5 MHz, 64-QAM, 30 kHz)	5G NR FR1 FDD	8.14	± 9.6 1
10957	AAA	5G NR DL (CP-OFDM, TM 3.1, 10 MHz, 64-QAM, 30 kHz)	5G NR FR1 FDD	8.31	≥ 9.6 5
10958	AAA	5G NR DL (CP-OFDM, TM 3.1, 15 MHz, 64-QAM, 30 kHz)	5G NR FR1 FDD	8.61	± 9.6 1
10959	AAA	5G NR DL (CP-OFDM, TM 3.1, 20 MHz, 64-QAM, 30 kHz)	5G NR FR1 FDD	8.33	± 9.6 °
10960	AAC	5G NR DL (CP-OFDM, TM 3.1, 5 MHz, 64-QAM, 15 kHz)	5G NR FR1 TDD	9.32	± 9.6 %
10961	AAB	5G NR DL (CP-OFDM, TM 3.1, 10 MHz, 64-QAM, 15 kHz)	5G NR FR1 TDD	9.36	± 9.6 %
10962	AAB	5G NR DL (CP-OFDM, TM 3.1, 15 MHz, 64-QAM, 15 kHz)	5G NR FR1 TDD	9.40	± 9.6 %
10963	AAB	5G NR DL (CP-OFDM, TM 3.1, 20 MHz, 64-QAM, 15 kHz)	5G NR FR1 TDD	9.55	± 9.6 %
10964	AAC	5G NR DL (CP-OFDM, TM 3.1, 5 MHz, 64-QAM, 30 kHz)	5G NR FR1 TDD	9.29	± 9.6 9
10965	AAB	5G NR DL (CP-OFDM, TM 3.1, 10 MHz, 64-QAM, 30 kHz)	5G NR FR1 TDD	9.37	± 9.6 %
10966	AAB	5G NR DL (CP-OFDM, TM 3.1, 15 MHz, 64-QAM, 30 kHz)	5G NR FR1 TDD	9.55	± 9.6 %
10967	AAB	5G NR DL (CP-OFDM, TM 3.1, 20 MHz, 64-QAM, 30 kHz)	5G NR FR1 TDD	9.42	± 9.6 °
10968	AAB	5G NR DL (CP-OFDM, TM 3.1, 100 MHz, 64-QAM, 30 kHz)	5G NR FR1 TDD	9.49	± 9.6 %
10972	AAB	5G NR (CP-OFDM, 1 RB, 20 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	11.59	± 9.6 %
10973	AAB	5G NR (DFT-s-OFDM, 1 RB, 100 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	9.06	± 9.6 %
10974	AAB	5G NR (CP-OFDM, 100% RB, 100 MHz, 256-QAM, 30 kHz)	5G NR FR1 TDD	10.28	± 9.6 %
10978	AAA	ULLA BDR	ULLA	2.23	± 9.6 9
10979	AAA	ULLA HDR4	ULLA	7.02	± 9.6 %
10980	AAA	ULLA HDR8	ULLA	8.82	± 9.6 9
10981	AAA	ULLA HDRp4	ULLA	1.50	± 9.6 7
10982	AAA	ULLA HDRp8	ULLA	1.44	± 9.6 1


¹² Uncertainty is determined using the max, deviation from linear response applying rectangular distribution and is expressed for the square of the field value.

Certificate No: EX3-7679_Sep21

Page 22 of 22



Appendix G. – Dipole Calibration Data





Engineering AG sughausstrasse 43, 8004 Zurich,	of Switzerland		S Schweizerischer Kalibrierdienst Service suisse d'étalonnage Servizie svizzero di taratura S Swiss Calibration Service
ccredited by the Swiss Accreditation The Swiss Accreditation Service in Nutlitateral Agreement for the rec	is one of the signatorie	s to the EA	Accreditation No.: SCS 0108
lient HCT (Dymstec)		NATIVISIONS	No: D2450V2-965_Jun21
CALIBRATION C	ERTIFICATE	걸 담당지	확 인 자
Dbject	D2450V2 - SN:96	35 XI 141 252	1 1 41 3 M2 2010705
Calibration procedure(is)	QA CAL-05.v11 Calibration Proce	dure for SAR Validation Source	- Martine -
Calibration date:	June 15, 2021		
The measurements and the uncert		robability are given on the following pages vy facility: environment temperature (22 ± 3	
The measurements and the uncert W calibrations have been conducts Calibration Equipment used (M&TE	ed in the closed laborato E critical for calibration)	robability are given on the following pages ry facility: environment temperature (22 ± 3	and are part of the certificate.)°C and humidity < 70%.
The measurements and the uncert III calibrations have been conducts Calibration Equipment used (M&TE Primary Standards	ed in the closed laborato 5 critical for calibration) ID #	robability are given on the following pages ry facility: environment temperature (22 ± 3 Cal Date (Certificate No.)	and are part of the certificate.)*C and humidity < 70%. Scheduled Calibration
he measurements and the uncert Il calibrations have been conducts alibration Equipment used (M&TE rimary Standards ower meter NRP	ed in the closed laborato critical for calibration) ID # SN: 104778	Cal Date (Certificate No.) 09-Apr-21 (No. 217-03291/03292)	and are part of the certificate. /*C and humidity < 70%. Scheduled Calibration Apr-22
he measurements and the uncert Il calibrations have been conducte alibration Equipment used (M&TE rimary Standards ower meter NRP ower sensor NRP-Z91	d in the closed laborato critical for calibration) ID # SN: 104778 SN: 103244	Cal Date (Certificate No.) 09-Apr-21 (No. 217-03291/03292) 09-Apr-21 (No. 217-03291/03292)	and are part of the certificate. /*C and humidity < 70%. Scheduled Calibration Apr-22 Apr-22
he measurements and the uncert Il calibrations have been conducts calibration Equipment used (M&TE rimary Standards 'ower meter NRP 'ower sensor NRP-Z91 'ower sensor NRP-Z91	ID # SN: 104778 SN: 103244 SN: 103245	Cal Date (Certificate No.) 09-Apr-21 (No. 217-03291/03292) 09-Apr-21 (No. 217-03291/03292) 09-Apr-21 (No. 217-03291)	and are part of the certificate. /*C and humidity < 70%. Scheduled Calibration Apr-22 Apr-22 Apr-22
The measurements and the uncert alibrations have been conductor calibration Equipment used (M&TE calibration Equipment used (M&TE cover sensor NRP-291 cover sensor NRP-291 Reference 20 dB Attenuator	ID # SN: 104778 SN: 104778 SN: 103244 SN: 103245 SN: BH9394 (20k)	robability are given on the following pages ry facility: environment temperature (22 ± 3 Cal Date (Certificate No.) 09-Apr-21 (No. 217-03291/03292) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03292) 09-Apr-21 (No. 217-03343)	and are part of the certificate. (*C and humidity < 70%. Scheduled Calibration Apr-22 Apr-22 Apr-22 Apr-22 Apr-22
The measurements and the uncert all calibrations have been conducts Calibration Equipment used (M&TE "rimary Standards "ower meter NRP "ower sensor NRP-291 "ower sensor NRP-291 Reference 20 dB Attenuator "ype-N mismatch combination	ed in the closed laborato critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245 SN: BH9394 (20k) SN: 310982 / 06327	robability are given on the following pages ry facility: environment temperature (22 ± 3 Cal Date (Certificate No.) 09-Apr-21 (No. 217-03291/03292) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03343) 09-Apr-21 (No. 217-03343)	and are part of the certificate.)*C and humidity < 70%. Scheduled Calibration Apr-22 Apr-22 Apr-22 Apr-22 Apr-22 Apr-22
The measurements and the uncert alibration Equipment used (M&TE Primary Standards Primary Standards Prower meter NRP Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Pype-N mismatch combination Reference Probe EX3DV4	ID # SN: 104778 SN: 104778 SN: 103244 SN: 103245 SN: BH9394 (20k)	robability are given on the following pages ry facility: environment temperature (22 ± 3 Cal Date (Certificate No.) 09-Apr-21 (No. 217-03291/03292) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03292) 09-Apr-21 (No. 217-03343)	and are part of the certificate. (*C and humidity < 70%. Scheduled Calibration Apr-22 Apr-22 Apr-22 Apr-22 Apr-22
The measurements and the uncert All calibrations have been conducts Calibration Equipment used (M&TE Primary Standards Power meter NRP Power sensor NRP-291 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards	ed in the closed laborato critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245 SN: 8H9394 (20k) SN: 310992 / 06327 SN: 7349 SN: 801 ID #	robability are given on the following pages ry facility: environment temperature (22 ± 3 Cal Date (Certificate No.) 09-Apr-21 (No. 217-03291/03292) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03292) 09-Apr-21 (No. 217-03343) 09-Apr-21 (No. 217-03343) 09-Apr-21 (No. 217-03344) 28-Dec-20 (No. DAE4-601_Nov20) 02-Nov-20 (No. DAE4-601_Nov20) Check Date (In house)	and are part of the certificate.)*C and humidity < 70%. Scheduled Calibration Apr-22 Apr-22 Apr-22 Apr-22 Apr-22 Dec-21 Nov-21 Scheduled Check
The measurements and the uncert all calibrations have been conducts calibration Equipment used (M&TE Primary Standards Power meter NRP Power sensor NRP-291 Power sensor NRP-291 Isterance 20 dB Attenuator ype-N mismatch combination Vaference Probe EX3DV4 DAE4 Power meter E4419B	ad in the closed laborato Critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245 SN: 103245 SN: 8H9394 (20k) SN: 310992 / 06327 SN: 7349 SN: 801 ID # SN: 0B39512475	robability are given on the following pages ry facility: environment temperature (22 ± 3 Cal Date (Certificate No.) 09-Apr-21 (No. 217-03291/03292) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03243) 09-Apr-21 (No. 217-03343) 09-Apr-21 (No. 217-03343) 09-Apr-21 (No. 217-03344) 28-Dec-20 (No. DAE4-601_Nov20) 02-Nov-20 (No. DAE4-601_Nov20) Check Date (In house) 30-Oct-14 (in house check Oct-20)	and are part of the certificate. (*C and humidity < 70%. Scheduled Calibration Apr-22 Apr-22 Apr-22 Apr-22 Apr-22 Der-21 Nov-21 Scheduled Check In house check: Oct-22
The measurements and the uncert All calibrations have been conducts Calibration Equipment used (M&TE Primary Standards "ower meter NRP "ower sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter E4419B Power meter E4419B	ad in the closed laborato Critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245 SN: 8199394 (20k) SN: 310982 / 06327 SN: 7349 SN: 801 ID # SN: 0B39512475 SN: US37292783	robability are given on the following pages ry facility: environment temperature (22 ± 3 <u>Cal Date (Certificate No.)</u> 09-Apr-21 (No. 217-03291/03292) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03292) 09-Apr-21 (No. 217-03343) 09-Apr-21 (No. 217-03343) 09-Apr-21 (No. 217-03344) 28-Dec-20 (No. EX3-7349_Dec20) 02-Nov-20 (No. DAE4-601_Nov20) <u>Check Date (in house)</u> 30-Oct-14 (in house check Oct-20) 07-Oct-15 (in house check Oct-20)	and are part of the certificate. (*C and humidity < 70%. Scheduled Calibration Apr-22 Apr-22 Apr-22 Apr-22 Apr-22 Der-21 Nov-21 Scheduled Check In house check: Oct-22 In house check: Oct-22
The measurements and the uncert All calibrations have been conducts Calibration Equipment used (M&TE Primary Standards Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Fype-N mismatch combination Reference Probe EX3DV4 DAE4 Becondary Standards Power meter E44198 Power sensor HP 8481A Power sensor HP 8481A	ad in the closed laborato critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245 SN: 103245 SN: 310982 / 06327 SN: 7349 SN: 801 ID # SN: GB39512475 SN: US37292783 SN: WY41092317	robability are given on the following pages ty facility: environment temperature (22 ± 3 Cal Date (Certificate No.) 09-Apr-21 (No. 217-03291/03292) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03243) 09-Apr-21 (No. 217-03343) 09-Apr-21 (No. 217-03243) 09-Apr-21 (No. 217-03243) 09-A	and are part of the certificate. (*C and humidity < 70%. Scheduled Calibration Apr-22 Apr-22 Apr-22 Apr-22 Apr-22 Apr-22 Dec-21 Nov-21 Scheduled Check In house check: Oct-22 In house check: Oct-22 In house check: Oct-22
The measurements and the uncert all calibrations have been conducts Calibration Equipment used (M&TE Trimary Standards Tower meter NRP Tower sensor NRP-291 Tower sensor NRP-291 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 NAE4 Recondary Standards Tower meter E4419B Tower sensor HP 8481A Tower sensor HP 8481A Regenerator R&S SMT-06	ad in the closed laborato critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245 SN: 819394 (20k) SN: 310982 / 06327 SN: 7349 SN: 301 ID # SN: 0B39512475 SN: 0B39512475 SN: 0B39512475 SN: 0B392512475 SN: 0B392512475 SN: 0B392512475 SN: 0B392512475 SN: 0037292783 SN: 400972	robability are given on the following pages ty facility: environment temperature (22 ± 3 Cal Date (Certificate No.) 09-Apr-21 (No. 217-03291/03292) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03292) 09-Apr-21 (No. 217-03343) 09-Apr-21 (No. 217-03292) 09-Apr-21 (No. 217-03292) 09-A	and are part of the certificate.)*C and humidity < 70%. Scheduled Calibration Apr-22 Apr-22 Apr-22 Apr-22 Dec-21 Nov-21 Scheduled Check In house check: Oct-22 In house check: Oct-22
The measurements and the uncert All calibrations have been conducts Calibration Equipment used (M&TE Primary Standards Power meter NRP Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter E44198 Power sensor HP 8481A Power sensor HP 8481A Regenerator R&S SMT-06	ad in the closed laborato critical for calibration) ID # SN: 104778 SN: 103245 SN: 103245 SN: 310982 / 06327 SN: 310982 / 06327 SN: 7349 SN: 601 ID # SN: GB39512475 SN: US37292783 SN: US37292783 SN: 100972 SN: 100972 SN: US41080477	robability are given on the following pages ry facility: environment temperature (22 ± 3 Cal Date (Certificate No.) 09-Apr-21 (No. 217-03291/03292) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03242) 09-Apr-21 (No. 217-03343) 09-Apr-21 (No. 217-03344) 28-Dec-20 (No. EX3-7349_Dec-20) 02-Nov-20 (No. DAE4-601_Nov-20) Check Date (In house check Oct-20) 07-Oct-15 (In house check Oct-20) 15-Jun-15 (In house check Oct-20) 31-Mar-14 (In house check Oct-20)	and are part of the certificate. (*C and humidity < 70%. Scheduled Calibration Apr-22 Apr-22 Apr-22 Apr-22 Apr-22 Dec-21 Nov-21 Scheduled Check In house check: Oct-22 In house check: Oct-21
The measurements and the uncert All calibrations have been conducte Calibration Equipment used (M&TE Primary Standards Prower sensor NRP-291 Prower sensor NRP-291 Reference 20 dB Attenuator Type-N mismatch combination Reference 20 dB Attenuator Prower sensor NRP-291 Reference 20 dB Attenuator Prower sensor NRP-291 Reference 20 dB Attenuator Prower sensor HP 8481A Recondary Standards Prower sensor HP 8481A Re generator R&S SMT-06 Retwork Analyzer Agilent E8358A	ad in the closed laborato critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245 SN: 310982 / 06327 SN: 310982 / 06327 SN: 7349 SN: 601 ID # SN: GB39512475 SN: US37292783 SN: US3729778 SN: US37297778 SN: US37297778	robability are given on the following pages ry facility: environment temperature (22 ± 3 Cal Date (Certificate No.) 09-Apr-21 (No. 217-03291/03292) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03292) 09-Apr-21 (No. 217-03343) 09-Apr-21 (No. 217-03343) 09-Apr-21 (No. 217-03344) 28-Dec-20 (No. EX3-7349_Dec20) 02-Nov-20 (No. EX3-7349_Dec20) 02-Nov-20 (No. DAE4-601_Nov20) Check Date (in house check Oct-20) 07-Oct-15 (in house check Oct-20) 07-Oct-15 (in house check Oct-20) 07-Oct-15 (in house check Oct-20) 15-Jun-15 (in house check Oct-20) 31-Mar-14 (in house check Oct-20) Function	and are part of the certificate.)*C and humidity < 70%. Scheduled Calibration Apr-22 Apr-22 Apr-22 Apr-22 Dec-21 Nov-21 Scheduled Check In house check: Oct-22 In house check: Oct-22
The measurements and the uncert All calibrations have been conducte Calibration Equipment used (M&TE Primary Standards Power sensor NRP-291 Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Pype-N mismatch combination Reference Probe EX3DV4 DAE4 Recondary Standards Power sensor HP 8481A Power sensor HP 8481A Researd P 8481A Resea	ad in the closed laborato critical for calibration) ID # SN: 104778 SN: 103245 SN: 103245 SN: 310982 / 06327 SN: 310982 / 06327 SN: 7349 SN: 601 ID # SN: GB39512475 SN: US37292783 SN: US37292783 SN: 100972 SN: 100972 SN: US41080477	robability are given on the following pages ry facility: environment temperature (22 ± 3 Cal Date (Certificate No.) 09-Apr-21 (No. 217-03291/03292) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03242) 09-Apr-21 (No. 217-03343) 09-Apr-21 (No. 217-03344) 28-Dec-20 (No. EX3-7349_Dec-20) 02-Nov-20 (No. DAE4-601_Nov-20) Check Date (In house check Oct-20) 07-Oct-15 (In house check Oct-20) 15-Jun-15 (In house check Oct-20) 31-Mar-14 (In house check Oct-20)	and are part of the certificate. (*C and humidity < 70%. Scheduled Calibration Apr-22 Apr-22 Apr-22 Apr-22 Apr-22 Dec-21 Nov-21 Scheduled Check In house check: Oct-22 In house check: Oct-21
The measurements and the uncert All calibrations have been conducts Calibration Equipment used (M&TE Primary Standards "ower sensor NRP-291 Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Fype-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter E4419B Power sensor HP 8481A Power sensor HP 8481A	ad in the closed laborato critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245 SN: 310982 / 06327 SN: 310982 / 06327 SN: 7349 SN: 601 ID # SN: GB39512475 SN: US37292783 SN: US3729778 SN: US37297778 SN: US37297778	robability are given on the following pages ry facility: environment temperature (22 ± 3 Cal Date (Certificate No.) 09-Apr-21 (No. 217-03291/03292) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03292) 09-Apr-21 (No. 217-03343) 09-Apr-21 (No. 217-03343) 09-Apr-21 (No. 217-03344) 28-Dec-20 (No. EX3-7349_Dec20) 02-Nov-20 (No. EX3-7349_Dec20) 02-Nov-20 (No. DAE4-601_Nov20) Check Date (in house check Oct-20) 07-Oct-15 (in house check Oct-20) 07-Oct-15 (in house check Oct-20) 07-Oct-15 (in house check Oct-20) 15-Jun-15 (in house check Oct-20) 31-Mar-14 (in house check Oct-20) Function	and are part of the certificate. (*C and humidity < 70%. Scheduled Calibration Apr-22 Apr-22 Apr-22 Apr-22 Apr-22 Dec-21 Nov-21 Scheduled Check In house check: Oct-22 In house check: Oct-21

Certificate No: D2450V2-965_Jun21

Page 1 of 6



Calibration Laboratory of Schmid & Partner





Schweizerischer Kalibrierdienst S Service suisse d'étalonnage C Servizio svizzero di taratura Ś Swiss Calibration Service

Accreditation No.: SCS 0108

Accredited by the Swiss Accreditation Service (SAS)

The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Closean

TSL	tissue simulating liquid
ConvF	sensitivity in TSL / NORM x,y,z
N/A	not applicable or not measured

Calibration is Performed According to the Following Standards:

- a) IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013
- b) IEC 62209-1, "Measurement procedure for the assessment of Specific Absorption Rate (SAR) from hand-held and body-mounted devices used next to the ear (frequency range of 300 MHz to 6 GHz)", July 2016
- c) IEC 62209-2, "Procedure to determine the Specific Absorption Rate (SAR) for wireless communication devices used in close proximity to the human body (frequency range of 30 MHz to 6 GHz)*, March 2010
- KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

Additional Documentation:

e) DASY4/5 System Handbook

Methods Applied and Interpretation of Parameters:

- Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss: These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured: SAR measured at the stated antenna input power.
- SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

Certificate No: D2450V2-965 Jun21

Page 2 of 6



Construction in the



Measurement Conditions

DASY system configuration, as far as not given on page 1.

DASY Version	DASY5	V52.10.4
Extrapolation	Advanced Extrapolation	
Phantom	Modular Flat Phantom	
Distance Dipole Center - TSL	10 mm	with Spacer
Zoom Scan Resolution	dx, dy, dz = 5 mm	
Frequency	2450 MHz ± 1 MHz	

Head TSL parameters

The following parameters and calculations were app	lied.
The tent of the second s	CITED IN THE REAL PROPERTY OF

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	39.2	1.80 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) "C	$37.7\pm6~\%$	1.87 mho/m ± 6 %
Head TSL temperature change during test	< 0.5 °C		44900

SAR result with Head TSL

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	
SAR measured	250 mW input power	13.7 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	53.3 W/kg ± 17.0 % (k=2)
	anndition	
SAR averaged over 10 cm ³ (10 g) of Head TSL	condition	
SAR averaged over 10 cm ³ (10 g) of Head TSL SAR measured	condition 250 mW input power	6.30 W/kg

Certificate No: D2450V2-965_Jun21

Page 3 of 6



Appendix (Additional assessments outside the scope of SCS 0108)

Antenna Parameters with Head TSL

Impedance, transformed to feed point	57.8 Ω + 6.6 jΩ
Return Loss	- 20.5 dB

General Antenna Parameters and Design

Electrical Delay (one direction)	1.153 ns
----------------------------------	----------

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.

No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

Additional EUT Data

Manufactured by	SPEAG

Certificate No: D2450V2-965_Jun21

Page 4 of 6



DASY5 Validation Report for Head TSL

Date: 15.06.2021

Test Laboratory: SPEAG, Zurich, Switzerland

DUT: Dipole 2450 MHz; Type: D2450V2; Serial: D2450V2 - SN:965


Communication System: UID 0 - CW; Frequency: 2450 MHz Medium parameters used: f = 2450 MHz; $\sigma = 1.87$ S/m; $\varepsilon_r = 37.7$; $\rho = 1000$ kg/m³ Phantom section: Flat Section Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2011)

DASY52 Configuration:

- Probe: EX3DV4 SN7349; ConvF(7.96, 7.96, 7.96) @ 2450 MHz; Calibrated: 28.12.2020
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 02.11.2020
- Phantom: Flat Phantom 5.0 (front); Type: QD 000 P50 AA; Serial: 1001
- DASY52 52.10.4(1527); SEMCAD X 14.6.14(7483)

Dipole Calibration for Head Tissue/Pin=250 mW, d=10mm/Zoom Scan (7x7x7)/Cube 0:

Measurement grid: dx=5mm, dy=5mm, dz=5mm Reference Value = 116.7 V/m; Power Drift = -0.01 dB Peak SAR (extrapolated) = 27.3 W/kg SAR(1 g) = 13.7 W/kg; SAR(10 g) = 6.30 W/kg Smallest distance from peaks to all points 3 dB below = 9 mm Ratio of SAR at M2 to SAR at M1 = 50% Maximum value of SAR (measured) = 22.4 W/kg




Certificate No: D2450V2-965_Jun21

Page 5 of 6



Impedance Measurement Plot for Head TSL



Page 6 of 6



alibration Laboratory chmid & Partner Engineering AG sughausatrasse 43, 8004 Zurich,			S Schweizerischer Kalibrierdienst C Service suisse d'étalonnage Servizio svizzero di taratura S Swiss Calibration Service
ccredited by the Swiss Accreditatio he Swiss Accreditation Service is ultilateral Agreement for the rec	s one of the signatories		Accreditation No.: SCS 0108
Bent HCT (Dymstec)		Certificati	• No: D5GHzV2-1107_Jul21
CALIBRATION C	ERTIFICATE		
Dbject	D5GHzV2 - SN:1	107	
Calibration procedure(s)	QA CAL-22.v6		
	Calibration Proce	dure for SAR Validation Sour	ces between 3-10 GHz
Calibration date:	July 22, 2021		
This calibration certificate document	nts the traceability to nati	onal standards, which realize the physica	al units of measurements (SI).
The measurements and the uncert	ainties with confidence p	robability are given on the following page	is and are part of the certificate.
All calibrations have been conducte	ad in the closed laborator	y facility: environment temperature (22 ±	s3)°C and humidity < 70%.
Calibration Equipment used (M&TE	E critical for calibration)		
	1		
Primary Standards	ID #	Cal Date (Certificate No.)	Scheduled Celibration
	ID # SN: 104778	Cal Date (Certificate No.) 09-Apr-21 (No. 217-03291/03292)	Scheduled Calibration Apr-22
Power meter NRP		en e	
Power meter NRP Power sensor NRP-291	SN: 104778	09-Apr-21 (No. 217-03291/03292) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03292)	Apr-22
Power meter NRP Power sensor NRP-291 Power sensor NRP-291	SN: 104778 SN: 103244	09-Apr-21 (No. 217-03291/03292) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03292) 09-Apr-21 (No. 217-03292)	Apr-22 Apr-22
Power mater NRP Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Type-N mismatch combination	SN: 104778 SN: 103244 SN: 103245 SN: 8H9394 (20k) SN: 310892 / 06327	09-Apr-21 (No. 217-03291/03292) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03292) 09-Apr-21 (No. 217-03343) 09-Apr-21 (No. 217-03344)	Арг-22 Арг-22 Арг-22 Арг-22 Арг-22 Арг-22
Power mater NRP Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4	SN: 104778 SN: 103244 SN: 103245 SN: 8H9394 (20k) SN: 310862 / 06327 SN: 3503	08-Apr-21 (No. 217-03291/03292) 08-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03292) 09-Apr-21 (No. 217-03343) 09-Apr-21 (No. 217-03344) 30-Dec-20 (No. EX3-3503_Dec20)	Αρι-22 Αρι-22 Αρι-22 Αρι-22 Αρι-22 Αρι-22 Dec-21
Power mater NRP Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4	SN: 104778 SN: 103244 SN: 103245 SN: 8H9394 (20k) SN: 310892 / 06327	09-Apr-21 (No. 217-03291/03292) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03292) 09-Apr-21 (No. 217-03343) 09-Apr-21 (No. 217-03344)	Арг-22 Арг-22 Арг-22 Арг-22 Арг-22 Арг-22
Power mater NRP Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4	SN: 104778 SN: 103244 SN: 103245 SN: 8H9394 (20k) SN: 310882 / 06327 SN: 3503 SN: 601	09-Apr-21 (No. 217-03291/03292) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03292) 09-Apr-21 (No. 217-03343) 09-Apr-21 (No. 217-03344) 30-Dec-20 (No. EX3-3503_Dec20) 02-Nov-20 (No. DAE4-601_Nov20)	Apr-22 Apr-22 Apr-22 Apr-22 Apr-22 Dec-21 Nov-21
Power mater NRP Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards	SN: 104778 SN: 103244 SN: 103245 SN: 8H9394 (20k) SN: 310882 / 06327 SN: 3503 SN: 601	08-Apr-21 (No. 217-03291/03292) 08-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03292) 09-Apr-21 (No. 217-03343) 09-Apr-21 (No. 217-03344) 30-Dec-20 (No. EX3-3503_Dec20) 02-Nov-20 (No. DAE4-601_Nov20) Check Date (in house)	Apr-22 Apr-22 Apr-22 Apr-22 Apr-22 Dec-21 Nov-21 Scheduled Check
Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter E4419B	SN: 104778 SN: 103244 SN: 103245 SN: 8H9394 (20k) SN: 310882 / 06327 SN: 3603 SN: 601 ID:# SN: G839512475	08-Apr-21 (No. 217-03291/03292) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03292) 09-Apr-21 (No. 217-03343) 09-Apr-21 (No. 217-03344) 30-Dec-20 (No. 217-03344) 30-Dec-20 (No. EX3-3503_Dec20) 02-Nov-20 (No. DAE4-601_Nov20) Check Date (in house) 30-Oct-14 (in house check Oct-20)	Apr-22 Apr-22 Apr-22 Apr-22 Apr-22 Dec-21 Nov-21 Scheduled Check In house check: Oct-22
Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter E4419B Power sensor HP 8481A	SN: 104776 SN: 103244 SN: 103245 SN: 8H9394 (20k) SN: 310882 / 06327 SN: 3603 SN: 601 ID # SN: G839512475 SN: US37292783	08-Apr-21 (No. 217-03291/03292) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03292) 09-Apr-21 (No. 217-03243) 09-Apr-21 (No. 217-03344) 30-Dec-20 (No. EX3-3503_Dec20) 02-Nov-20 (No. DAE4-601_Nov20) Check Date (in house) 30-Oct-14 (in house check Oct-20) 07-Oct-15 (in house check Oct-20)	Apr-22 Apr-22 Apr-22 Apr-22 Apr-22 Dec-21 Nov-21 Scheduled Check In house check: Oct-22 In house check: Oct-22
Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter E4419B Power sensor HP 8481A Power sensor HP 8481A	SN: 104776 SN: 103244 SN: 103245 SN: 8H9394 (20k) SN: 310892 / 06327 SN: 3603 SN: 601 ID:# SN: G839512475 SN: US37282783 SN: MY41092317	09-Apr-21 (No. 217-03291/03292) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03293) 09-Apr-21 (No. 217-03343) 09-Apr-21 (No. 217-03344) 30-Dec-20 (No. EX3-3503_Dec20) 02-Nov-20 (No. DAE4-601_Nov20) Check Date (in house) 30-Oct-14 (in house check Oct-20) 07-Oct-15 (in house check Oct-20) 07-Oct-15 (in house check Oct-20)	Apr-22 Apr-22 Apr-22 Apr-22 Apr-22 Dec-21 Nov-21 Scheduled Check In house check: Oct-22 In house check: Oct-22 In house check: Oct-22
Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter E44198 Power sensor HP 8481A Power sensor HP 8481A RF generator R&S SMT-06	SN: 104778 SN: 103244 SN: 103245 SN: 8H9394 (20k) SN: 310982 / 06327 SN: 3603 SN: 601 ID # SN: G839512475 SN: US37292783 SN: MY41092317 SN: 100972	09-Apr-21 (No. 217-03291/03292) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03343) 09-Apr-21 (No. 217-03344) 30-Dec-20 (No. EX3-3503_Dec20) 02-Nov-20 (No. DAE4-601_Nov20) Check Date (in house) 30-Oct-14 (in house check Oct-20) 07-Oct-15 (in house check Oct-20) 07-Oct-15 (in house check Oct-20) 15-Jun-15 (in house check Oct-20)	Apr-22 Apr-22 Apr-22 Apr-22 Apr-22 Dec-21 Nov-21 Scheduled Check In house check: Oct-22 In house check: Oct-22 In house check: Oct-22 In house check: Oct-22 In house check: Oct-22
Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter E44198 Power sensor HP 8481A Power sensor HP 8481A RF generator R&S SMT-06	SN: 104776 SN: 103244 SN: 103245 SN: 8H9394 (20k) SN: 310892 / 06327 SN: 3603 SN: 601 ID:# SN: G839512475 SN: US37282783 SN: MY41092317	09-Apr-21 (No. 217-03291/03292) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03293) 09-Apr-21 (No. 217-03343) 09-Apr-21 (No. 217-03344) 30-Dec-20 (No. EX3-3503_Dec20) 02-Nov-20 (No. DAE4-601_Nov20) Check Date (in house) 30-Oct-14 (in house check Oct-20) 07-Oct-15 (in house check Oct-20) 07-Oct-15 (in house check Oct-20)	Apr-22 Apr-22 Apr-22 Apr-22 Apr-22 Dec-21 Nov-21 Scheduled Check In house check: Oct-22 In house check: Oct-22 In house check: Oct-22
Power meter NRP Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter E44198 Power sensor HP 8481A Power sensor HP 8481A RF generator R&S SMT-06	SN: 104778 SN: 103244 SN: 103245 SN: 8H9394 (20k) SN: 310982 / 06327 SN: 3503 SN: 601 ID # SN: G839512475 SN: US37292783 SN: MY41092317 SN: 100972 SN: 100972 SN: US41080477	09-Apr-21 (No. 217-03291/03292) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03343) 09-Apr-21 (No. 217-03344) 30-Dec-20 (No. EX3-3503_Dec20) 02-Nov-20 (No. DAE4-601_Nov20) Check Date (in house) 30-Oct-14 (in house check Oct-20) 07-Oct-15 (in house check Oct-20) 07-Oct-15 (in house check Oct-20) 15-Jun-15 (in house check Oct-20) 31-Mar-14 (in house check Oct-20)	Apr-22 Apr-22 Apr-22 Apr-22 Apr-22 Dec-21 Nov-21 Scheduled Check In house check: Oct-22 In house check: Oct-22 In house check: Oct-22 In house check: Oct-22 In house check: Oct-22
Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter E4419B Power sensor HP 8481A Power sensor HP 8481A RF generator R&S SMT-06 Network Analyzer Agilent EB358A	SN: 104778 SN: 103244 SN: 103245 SN: 8H9394 (20k) SN: 310982 / 06327 SN: 3503 SN: 601 ID # SN: G839512475 SN: US37282783 SN: US37282783 SN: MY41092317 SN: 100972 SN: US41080477 Name	09-Apr-21 (No. 217-03291/03292) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03343) 09-Apr-21 (No. 217-03344) 30-Dec-20 (No. EX3-3303, Dec20) 02-Nov-20 (No. DAE4-601_Nov20) Check Date (in house) 30-Oct-14 (in house check Oct-20) 07-Oct-15 (in house check Oct-20) 07-Oct-15 (in house check Oct-20) 15-Jun-15 (in house check Oct-20) 31-Mar-14 (in house check Oct-20) Function	Apr-22 Apr-22 Apr-22 Apr-22 Apr-22 Dec-21 Nov-21 Scheduled Check In house check: Oct-22 In house check: Oct-22 In house check: Oct-22 In house check: Oct-22 In house check: Oct-22
Power meter NRP Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power sensor HP 8481A Power sensor HP 8481A RF generator R&S SMT-06 Network Analyzer Agilent EB358A	SN: 104778 SN: 103244 SN: 103245 SN: 8H9394 (20k) SN: 310982 / 06327 SN: 3503 SN: 601 ID # SN: G839512475 SN: US37292783 SN: MY41092317 SN: 100972 SN: 100972 SN: US41080477	09-Apr-21 (No. 217-03291/03292) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03343) 09-Apr-21 (No. 217-03344) 30-Dec-20 (No. EX3-3503_Dec20) 02-Nov-20 (No. DAE4-601_Nov20) Check Date (in house) 30-Oct-14 (in house check Oct-20) 07-Oct-15 (in house check Oct-20) 07-Oct-15 (in house check Oct-20) 15-Jun-15 (in house check Oct-20) 31-Mar-14 (in house check Oct-20)	Apr-22 Apr-22 Apr-22 Apr-22 Apr-22 Dec-21 Nov-21 Scheduled Check In house check: Oct-22 In house check: Oct-22 In house check: Oct-22 In house check: Oct-22 In house check: Oct-22
Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter E4419B Power sensor HP 8481A Power sensor HP 8481A RF generator R&S SMT-06 Network Analyzer Agilent EB358A Calibrated by:	SN: 104776 SN: 103244 SN: 103245 SN: 8H9394 (20k) SN: 310892 / 06327 SN: 3603 SN: 601 ID # SN: G839512475 SN: US37292783 SN: MY41092317 SN: 100972 SN: US41080477 Name Claudio Laubler	09-Apr-21 (No. 217-03291/03292) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03292) 09-Apr-21 (No. 217-03343) 09-Apr-21 (No. 217-03344) 30-Dec-20 (No. 217-03343) Dec-20 (No. 217-03343) Dec-20 (No. 217-03244) 30-Dec-20 (No. 217	Apr-22 Apr-22 Apr-22 Apr-22 Apr-22 Dec-21 Nov-21 Scheduled Check In house check: Oct-22 In house check: Oct-22 In house check: Oct-22 In house check: Oct-22 In house check: Oct-22
Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter E4419B Power sensor HP 8481A Power sensor HP 8481A RF generator R&S SMT-06 Network Analyzer Agilent EB358A Calibrated by:	SN: 104778 SN: 103244 SN: 103245 SN: 8H9394 (20k) SN: 310982 / 06327 SN: 3503 SN: 601 ID # SN: G839512475 SN: US37282783 SN: US37282783 SN: MY41092317 SN: 100972 SN: US41080477 Name	09-Apr-21 (No. 217-03291/03292) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03343) 09-Apr-21 (No. 217-03344) 30-Dec-20 (No. EX3-3303, Dec20) 02-Nov-20 (No. DAE4-601_Nov20) Check Date (in house) 30-Oct-14 (in house check Oct-20) 07-Oct-15 (in house check Oct-20) 07-Oct-15 (in house check Oct-20) 15-Jun-15 (in house check Oct-20) 31-Mar-14 (in house check Oct-20) Function	Apr-22 Apr-22 Apr-22 Apr-22 Apr-22 Dec-21 Nov-21 Scheduled Check In house check: Oct-22 In house check: Oct-22 In house check: Oct-22 In house check: Oct-22 In house check: Oct-22
Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter E4419B Power sensor HP 8481A Power sensor HP 8481A RF generator R&S SMT-06 Network Analyzer Agilent EB358A Calibrated by:	SN: 104776 SN: 103244 SN: 103245 SN: 8H9394 (20k) SN: 310892 / 06327 SN: 3603 SN: 601 ID # SN: G839512475 SN: US37292783 SN: MY41092317 SN: 100972 SN: US41080477 Name Claudio Laubler	09-Apr-21 (No. 217-03291/03292) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03292) 09-Apr-21 (No. 217-03343) 09-Apr-21 (No. 217-03344) 30-Dec-20 (No. 217-03343) Dec-20 (No. 217-03343) Dec-20 (No. 217-03244) 30-Dec-20 (No. 217	Apr-22 Apr-22 Apr-22 Apr-22 Apr-22 Dec-21 Nov-21 Scheduled Check In house check: Oct-22 In house check: Oct-21
Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter E4419B Power sensor HP 8481A Power sensor HP 8481A RF generator R&S SMT-06 Network Analyzer Agilent E8358A Calibrated by:	SN: 104776 SN: 103244 SN: 103245 SN: 8H9394 (20k) SN: 310892 / 06327 SN: 3503 SN: 601 ID# SN: G839512475 SN: US37292783 SN: MY41092317 SN: 100972 SN: 10972 SN: US41080477 Name Cilaudio Laubler	09-Apr-21 (No. 217-03291/03292) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03292) 09-Apr-21 (No. 217-03292) 09-Apr-21 (No. 217-03344) 30-Dec-20 (No. EX3-3503_Dec20) 02-Nov-20 (No. DAE4-601_Nov20) Check Date (in house) 30-Oct-14 (in house check Oct-20) 07-Oct-15 (in house check Oct-20) 07-Oct-15 (in house check Oct-20) 31-Mar-14 (in house check Oct-20) 31-Mar-14 (in house check Oct-20) Function Laboratory Technician	Apr-22 Apr-22 Apr-22 Apr-22 Apr-22 Dec-21 Nov-21 Scheduled Check In house check: Oct-22 In house check: Oct-21 Signature House check: Oct-21 In house check: Oct-21 Signature House check: Oct-21
Power meter NRP Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power sensor HP 8481A Power sensor HP 8481A RF generator R&S SMT-06 Network Analyzer Agilent EB358A Calibrated by:	SN: 104776 SN: 103244 SN: 103245 SN: 8H9394 (20k) SN: 310892 / 06327 SN: 3503 SN: 601 ID# SN: G839512475 SN: US37292783 SN: MY41092317 SN: 100972 SN: 10972 SN: US41080477 Name Cilaudio Laubler	09-Apr-21 (No. 217-03291/03292) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03293) 09-Apr-21 (No. 217-03343) 09-Apr-21 (No. 217-03344) 30-Dec-20 (No. EX3-3503_Dec20) 02-Nov-20 (No. DAE4-601_Nov20) Check Date (in house) 30-Oct-14 (in house check Oct-20) 07-Oct-15 (in house check Oct-20) 07-Oct-15 (in house check Oct-20) 31-Mar-14 (in house check Oct-20) 31-Mar-14 (in house check Oct-20) Function Laboratory Technician Technical Manager	Apr-22 Apr-22 Apr-22 Apr-22 Apr-22 Dec-21 Nov-21 Scheduled Check In house check: Oct-22 In house check: Oct-21 Signature House check: Oct-21 In house check: Oct-21 Signature House check: Oct-21
	SN: 104776 SN: 103244 SN: 103245 SN: 8H9394 (20k) SN: 310882 / 06327 SN: 3603 SN: 601 ID # SN: GB39512475 SN: US37292783 SN: MY41092317 SN: 100972 SN: US41080477 Name Claudio Laubler Katja: Pokovic	09-Apr-21 (No. 217-03291/03292) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03292) 09-Apr-21 (No. 217-03292) 09-Apr-21 (No. 217-03343) 09-Apr-21 (No. 217-03344) 30-Dec-20 (No. EX3-3503_Dec20) 02-Nov-20 (No. DAE4-601_Nov20) Check Date (In house) 30-Oct-14 (In house check Oct-20) 07-Oct-15 (In house check Oct-20) 07-Oct-15 (In house check Oct-20) 15-Jun-15 (In house check Oct-20) 15-Jun-15 (In house check Oct-20) 31-Mar-14 (In house check Oct-20) 31-Mar-14 (In house check Oct-20) Function Laboratory Technician	Apr-22 Apr-22 Apr-22 Apr-22 Apr-22 Dec-21 Nov-21 Scheduled Check In house check: Oct-22 In house check: Oct-21 Signature Jack of the second s
Power meter NRP Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power sensor HP 8481A Power sensor HP 8481A RF generator R&S SMT-06 Network Analyzer Agilent EB358A Calibrated by: Approved by: This calibration certificate shall not	SN: 104776 SN: 103244 SN: 103245 SN: 8H9394 (20k) SN: 310882 / 06327 SN: 3603 SN: 601 ID # SN: GB39512475 SN: US37292783 SN: MY41092317 SN: 100972 SN: US41080477 Name Claudio Laubler Katja: Pokovic	09-Apr-21 (No. 217-03291/03292) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03292) 09-Apr-21 (No. 217-03292) 09-Apr-21 (No. 217-03344) 30-Dec-20 (No. EX3-3503_Dec20) 02-Nov-20 (No. DAE4-601_Nov20) Check Date (in house) 30-Oct-14 (in house check Oct-20) 07-Oct-15 (in house check Oct-20) 07-Oct-15 (in house check Oct-20) 15-Jun-15 (in house check Oct-20) 15-Jun-15 (in house check Oct-20) 31-Mar-14 (in house check Oct-20) Function Laboratory Technician Technical Manager 144 without written approval of the labor	Apr-22 Apr-22 Apr-22 Apr-22 Apr-22 Dec-21 Nov-21 Scheduled Check In house check: Oct-22 In house check: Oct-21 Signature House check: Oct-21 In house check: Oct-21 Signature House check: Oct-21
Power meter NRP Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power sensor HP 8481A Power sensor HP 8481A RF generator R&S SMT-06 Network Analyzer Agilent EB358A Calibrated by:	SN: 104776 SN: 103244 SN: 103245 SN: 8H9394 (20k) SN: 310882 / 06327 SN: 3603 SN: 601 ID # SN: GB39512475 SN: US37292783 SN: MY41092317 SN: 100972 SN: US41080477 Name Claudio Laubler Katja: Pokovic	09-Apr-21 (No. 217-03291/03292) 09-Apr-21 (No. 217-03291) 09-Apr-21 (No. 217-03292) 09-Apr-21 (No. 217-03292) 09-Apr-21 (No. 217-03343) 09-Apr-21 (No. 217-03344) 30-Dec-20 (No. EX3-3503_Dec20) 02-Nov-20 (No. DAE4-601_Nov20) Check Date (In house) 30-Oct-14 (In house check Oct-20) 07-Oct-15 (In house check Oct-20) 07-Oct-15 (In house check Oct-20) 15-Jun-15 (In house check Oct-20) 15-Jun-15 (In house check Oct-20) 31-Mar-14 (In house check Oct-20) 31-Mar-14 (In house check Oct-20) Function Laboratory Technician	Apr-22 Apr-22 Apr-22 Apr-22 Apr-22 Dec-21 Nov-21 Scheduled Check In house check: Oct-22 In house check: Oct-21 Signature Jacoba Jacoba J



Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland



Schweizerlscher Kallbrierdienst Service suisse d'étalonnage Servizio svizzero di taratura Swiss Calibration Service

Accreditation No.: SCS 0108

S

С

S

Accredited by the Swiss Accreditation Service (SAS)

The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Glossary:

TSL	tissue simulating liquid
ConvF	sensitivity in TSL / NORM x,y,z
N/A	not applicable or not measured

Calibration is Performed According to the Following Standards:

- a) IEC/IEEE 62209-1528, "Measurement Procedure For The Assessment Of Specific Absorption Rate Of Human Exposure To Radio Frequency Fields From Hand-Held And Body-Worn Wireless Communication Devices - Part 1528: Human Models, Instrumentation And Procedures (Frequency Range of 4 MHz to 10 GHz)", October 2020.
- b) KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

Additional Documentation:

c) DASY System Handbook

Methods Applied and Interpretation of Parameters:

- Measurement Conditions: Further details are available from the Validation Report at the end
 of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL: The source is mounted in a touch configuration below the center marking of the flat phantom.
- Return Loss: This parameter is measured with the source positioned under the liquid filled phantom (as described in the measurement condition clause). The Return Loss ensures low reflected power. No uncertainty required.
- SAR measured: SAR measured at the stated antenna input power.
- SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

Page 2 of 8





Measurement Conditions

DASY system configuration, as far as not given on page 1.

DASY Version	DASY52	V52.10.4
Extrapolation	Advanced Extrapolation	
Phantom	Modular Flat Phantom V5.0	
Distance Dipole Center - TSL	10 mm	with Spacer
Zoom Scan Resolution	dx, dy = 4.0 mm, dz = 1.4 mm	Graded Ratio = 1.4 (Z direction)
Frequency	5250 MHz ± 1 MHz 5600 MHz ± 1 MHz 5750 MHz ± 1 MHz	

Head TSL parameters at 5250 MHz

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	35.9	4.71 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	35.6 ± 6 %	4.60 mho/m ± 6 %
Head TSL temperature change during test	< 0.5 °C	1000	

SAR result with Head TSL at 5250 MHz

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	
SAR measured	100 mW input power	8.08 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	80.6 W/kg ± 19.9 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Head TSL	condition	
SAR measured	100 mW input power	2.33 W/kg

Head TSL parameters at 5600 MHz

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	35.5	5.07 mha/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	35.1 ± 6 %	4.95 mha/m ± 6 %
Head TSL temperature change during test	< 0.5 °C		

SAR result with Head TSL at 5600 MHz

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	
SAR measured	100 mW input power	8.44 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	84.2 W/kg ± 19.9 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Head TSL	condition	
SAR averaged over 10 cm ³ (10 g) of Head TSL SAR measured	condition 100 mW input power	2.43 W/kg

Certificate No: D5GHzV2-1107_Jul21

Page 3 of 8



Head TSL parameters at 5750 MHz The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	35.4	5.22 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	34.8 ± 6 %	5.11 mho/m ± 6 %
Head TSL temperature change during test	< 0.5 °C		

SAR result with Head TSL at 5750 MHz

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	
SAR measured	100 mW input power	8.13 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	80.9 W/kg ± 19.9 % (k=2)
SAP supraged over 10 em3 (10 e) of Meed TCI	and distant	
SAR averaged over 10 cm ³ (10 g) of Head TSL	condition	
SAR averaged over 10 cm ³ (10 g) of Head TSL SAR measured	condition 100 mW input power	2.34 W/kg

Page 4 of 8



Appendix (Additional assessments outside the scope of SCS 0108)

Antenna Parameters with Head TSL at 5250 MHz

Impedance, transformed to feed point	48.4 Ω - 6.5 jΩ
Return Loss	- 23.4 dB

Antenna Parameters with Head TSL at 5600 MHz

Impedance, transformed to feed point	54.2 Ω - 2.6 jΩ	
Return Loss	- 26.5 dB	

Antenna Parameters with Head TSL at 5750 MHz

Impedance, transformed to feed point	56.5 Ω - 2.9 jΩ	
Return Loss	- 23.5 dB	

General Antenna Parameters and Design

Electrical Delay (one direction)	1.199 ns
----------------------------------	----------

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.

No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

Additional EUT Data

Manufactured by	SPEAG
-----------------	-------

Certificate No: D5GHzV2-1107_Jul21

Page 5 of 8



DASY5 Validation Report for Head TSL

Date: 22.07.2021

Test Laboratory: SPEAG, Zurich, Switzerland

DUT: Dipole D5GHzV2; Type: D5GHzV2; Serial: D5GHzV2 - SN:1107

Communication System: UID 0 - CW; Frequency: 5250 MHz, Frequency: 5600 MHz, Frequency: 5750 MHz Medium parameters used: f = 5250 MHz; σ = 4.6 S/m; ε_r = 35.6; ρ = 1000 kg/m³. Medium parameters used: f = 5600 MHz; σ = 4.95 S/m; ε_r = 35.1; ρ = 1000 kg/m³ Medium parameters used: f = 5750 MHz; σ = 5.11 S/m; ε_r = 34.8; ρ = 1000 kg/m³ Phantom section: Flat Section Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2011)

DASY52 Configuration:

- Probe: EX3DV4 SN3503; ConvF(5.5, 5.5, 5.5) @ 5250 MHz, ConvF(5.1, 5.1, 5.1) @ 5600 MHz, ConvF(5.08, 5.08, 5.08) @ 5750 MHz; Calibrated: 30.12.2020
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 02.11.2020
- Phantom: Flat Phantom 5.0 (front); Type: QD 000 P50 AA; Serial: 1001
- DASY52 52.10.4(1535); SEMCAD X 14.6.14(7501)

Dipole Calibration for Head Tissue/Pin=100mW, dist=10mm, f=5250 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 77.05 V/m; Power Drift = 0.03 dB Peak SAR (extrapolated) = 27.4 W/kg SAR(1 g) = 8.08 W/kg; SAR(10 g) = 2.33 W/kg Smallest distance from peaks to all points 3 dB below = 7.5 mm Ratio of SAR at M2 to SAR at M1 = 71.4% Maximum value of SAR (measured) = 18.3 W/kg

Dipole Calibration for Head Tissue/Pin=100mW, dist=10mm, f=5600 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 76.80 V/m; Power Drift = 0.03 dB Peak SAR (extrapolated) = 31.0 W/kg SAR(1 g) = 8.44 W/kg; SAR(10 g) = 2.43 W/kg Smallest distance from peaks to all points 3 dB below = 7.5 mm Ratio of SAR at M2 to SAR at M1 = 68.7% Maximum value of SAR (measured) = 19.7 W/kg


Dipole Calibration for Head Tissue/Pin=100mW, dist=10mm, f=5750 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 74.42 V/m; Power Drift = 0.02 dB Peak SAR (extrapolated) = 31.4 W/kg SAR(1 g) = 8.13 W/kg; SAR(10 g) = 2.34 W/kg Smallest distance from peaks to all points 3 dB below = 7.6 mm

Certificate No: D5GHzV2-1107_Jul21

Page 6 of 8



Ratio of SAR at M2 to SAR at M1 = 66.9% Maximum value of SAR (measured) = 19.3 W/kg




0 dB = 18.3 W/kg = 12.62 dBW/kg

Certificate No: D5GHzV2-1107_Jul21

Page 7 of 8



Impedance Measurement Plot for Head TSL



Page 8 of 8