

SPEAG 900 MHz Dipole; Model D900V2, SN 084; Test Date: 8/20/03

Motorola CGISS EME Lab

Run #: System perf R2-030820-01

TX Freq: 900MHz

Simulated tissue temp: 21.1C

Start Power; 250mW

Target: 11.99W/g for 1g SAR, 7.58mW/g for 10g SAR, +/- 10% from system performance target.

SAR calculated 1g is 11.39 mW/g percent from target (including drift) is -5.0%


SAR Calculated 10g is 7.26 mW/g Percent from target (including drift) is -4.3%

Flat; Probe: ET3DV6 - SN1547 (Cal Date 09-28-2002); Probe Cal Date: 28/09/02 ConvF(6.20,6.20,6.20); Crest factor: 1.0; FCC

Body 900: $\sigma = 1.04$ mho/m $\epsilon = 52.4$ $\rho = 1.00$ g/cm³; DAE: DAE3V1 SN406; Cal 11/11/02

Cubes (3): Peak: 2.95 mW/g $\pm$ 17.21 dB, SAR (1g): 1.89 mW/g $\pm$ 16.71 dB, SAR (10g): 1.21 mW/g $\pm$ 16.25 dB, (Worst-case extrapolation) Penetration depth: 12.2 (11.3, 13.5) [mm]

Power drift: -0.01 dB



SPEAG 900 MHz Dipole; Model D900V2, SN 084; Test Date: 8/20/03

Motorola CGISS EME Lab

Run #: System perf R2-030820-02

TX Freq: 900MHz

Simulated tissue temp: 20.9 C

Start Power; 250mW

Target: 11.66 mW/g for 1g SAR, 7.26 mW/g for 10g SAR, +/- 10% from system performance target.

SAR calculated 1g is 11.72 mW/g percent from target (including drift) is 0.5 %


SAR Calculated 10g is 7.35 mW/g Percent from target (including drift) is 1.2%

SAM; Probe: ET3DV6 - SN1547 (Cal Date 09-28-2002);Probe Cal Date: 28/09/02ConvF(6.50,6.50,6.50); Crest factor: 1.0;

IEEE HEAD 900MHz: $\sigma = 1.01$ mho/m $\epsilon = 41.3$ $\rho = 1.00$ g/cm³; DAE: DAE3V1 SN406; Cal 11/11/02

Cubes (3): Peak: 3.15 mW/g $\pm$ 18.12 dB, SAR (1g): 1.97 mW/g $\pm$ 0.00 dB, SAR (10g): 1.24 mW/g $\pm$ 0.00 dB, (Worst-case extrapolation) Penetration depth: 11.3 (10.5, 12.6) [mm]

Power drift: 0.03 dB



SPEAG 835 MHz Dipole; Model D835V2, SN 426; Test Date: 8/21/03

Motorola CGISS EME Lab

Run #: System perf R2-030821-01

TX Freq: 900MHz

Simulated tissue temp: 21.0 C

Start Power; 250mW

Target: 11.66 mW/g for 1g SAR, 7.26 mW/g for 10g SAR, +/- 10% from system performance target.

SAR calculated 1g is 11.89 mW/g percent from target (including drift) is 2.0 %


SAR Calculated 10g is 7.45 mW/g Percent from target (including drift) is 2.6%

SAM; Probe: ET3DV6 - SN1547 (Cal Date 09-28-2002);Probe Cal Date: 28/09/02ConvF(6.50,6.50,6.50); Crest factor: 1.0;

IEEE HEAD 900MHz: $\sigma = 1.01$ mho/m $\epsilon = 41.3$ $\rho = 1.00$ g/cm³

Cubes (3): Peak: 3.19 mW/g $\pm$ 20.28 dB, SAR (1g): 2.00 mW/g $\pm$ 22.83 dB, SAR (10g): 1.26 mW/g $\pm$ 0.00 dB, (Worst-case extrapolation) Penetration depth: 11.3 (10.5, 12.6) [mm]

Power drift: 0.04 dB



SYSTEM PERFORMANCE CHECK TARGET SAR

Date: 05/22/03 Frequency (MHz): 900
Lab Location: CGISS Mixture Type: 900-FCC Body
Robot System: CGISS-2 Ambient Temp.(°C): 20.9
Probe Serial #: 1547 Tissue Temp.(°C): 21.0
DAE Serial #: DAE3V1 SN401

Tissue Characteristics Phantom Type/SN: 80302002A/S8
Permittivity: 53.9 Distance (mm): 15
Conductivity: 1.07

Reference Source: Dipole (Dipole/Handset)
Reference SN: 084

Power to Dipole: 250 mW

Measured SAR Value: 2.99 mW/g, 1.89 mW/g (10g avg.)
Power Drift: -0.01 dB

New Target/Measured
SAR Value: 11.99 mW/g, 7.58 mW/g (10g avg.)
(normalized to 1.0 W,
with drift compensation)

Test performed by: Kim Uong Initial: KU

05/22/03

Dipole D900V2 SN084; Test Date: 05/22/03

Run #: 0305022-06 Phantom #: 80302002A/S8

Model#: SPEAG dipole D900V2 SN084

Robot#: CGISS-2 DAE: DAE3V1 SN374 (2/11/02) Tester: Kim Uong

Tx Freq: 900MHz Simulated tissue temp: 21.0C, Ambient :20.9C

Start power: 250mW

Target:

11.8mW/g for 1g-SAR, 7.52mW/g for 10g-SAR, +/-12% from SPEAG Dipole certificate 2/11/02

Flat Phantom; Section;


Probe: ET3DV6 - SN1547 (Cal Date 09-28-2002); ConvF(6.20,6.20,6.20); Probe cal date: 28/09/02; Crest

factor: 1.0; FCC Body 900: $\sigma = 1.07$ mho/m $\epsilon_r = 53.9$ $\rho = 1.00$ g/cm³

Cubes (2): SAR (1g): 2.99 mW/g ± 0.03 dB, SAR (10g): 1.89 mW/g ± 0.03 dB, (Worst-case extrapolation)

Coarse: Dx = 15.0, Dy = 15.0, Dz = 10.0; Max at 28.5, 60.0, 4.0

Powerdrift: -0.01 dB



SAR_{Tot} [mW/g]



Motorola CGISS EME Lab

SYSTEM PERFORMANCE CHECK TARGET SAR

Date: 05/05/03 Frequency (MHz): 900
Lab Location: CGISS Mixture Type: 900-IEEE Head
Robot System: CGISS-2 Ambient Temp.(°C): 22.4
Probe Serial #: 1547 Tissue Temp.(°C): 20.3
DAE Serial #: DAE3V1 SN401

Tissue Characteristics Phantom Type/SN: SAMTP1209
Permittivity: 40.5 Distance (mm): 15
Conductivity: 0.98

Reference Source: Dipole (Dipole/Handset)
Reference SN: 084

Power to Dipole: 250 mW

Measured SAR Value: 2.97 mW/g, 1.85 mW/g (10g avg.)
Power Drift: 0.08 dB

New Target/Measured
SAR Value: 11.66 mW/g, 7.26 mW/g (10g avg.)
(normalized to 1.0 W,
with drift compensation)

Test performed by: Kim Uong Initial: KU

05/05/03

Dipole D900V2 SN084; Test date:05/05/03

Run #: 030505-01 Phantom #:SAMTP1209

Model#: SPEAG dipole D900V2 SN084

Robot#: CGISS-2

DAE: DAE3V1 SN401 (8/26/02)

Tester: Kim Uong

Tx Freq:900MHz

Simulated tissue temp: 20.3C

Start power: 250mW

Target:

11.2mW/g for 1g-SAR, 7.12mW/g for 10g-SAR, +/-12% from SPEAG Dipole certificate 2/11/02.


11.66mW/g for 1g-SAR, 7.26mW/g for 10g-SAR, +/- 10% from system performance target 05/05/03.

SAM; Probe: ET3DV6 - SN1547 (Cal Date 09-28-2002); ConvP(6.50,6.50,6.50); Crest factor: 1.0; IEEE HEAD 900MHz: $\sigma = 0.98$ mho/m $\epsilon_r = 40.5$ $\rho = 1.00$ g/cm³


Cubes (2): Peak: 4.85 mW/g ± 0.02 dB, SAR (1g): 2.97 mW/g ± 0.02 dB, SAR (10g): 1.85 mW/g ± 0.02 dB, (Worst-case extrapolation)

Penetration depth: 11.1 (9.9, 12.7) [mm]

Powerdrift: 0.08 dB



SAR_{tot} [mW/g]



Motorola CGISS EME Lab

APPENDIX D

Calibration Certificates

Schmid & Partner Engineering AG

Zeughausstrasse 43, 8004 Zurich, Switzerland, Phone +41 1 245 97 00, Fax +41 1 245 97 79

Calibration Certificate

Dosimetric E-Field Probe

Type:

ET3DV6

Serial Number:

1547

Place of Calibration:

Zurich

Date of Calibration:

September 28, 2002

Calibration Interval:

12 months

Schmid & Partner Engineering AG hereby certifies, that this device has been calibrated on the date indicated above. The calibration was performed in accordance with specifications and procedures of Schmid & Partner Engineering AG.

Wherever applicable, the standards used in the calibration process are traceable to international standards. In all other cases the standards of the Laboratory for EMF and Microwave Electronics at the Swiss Federal Institute of Technology (ETH) in Zurich, Switzerland have been applied.

Calibrated by:

N. Vetterli

Approved by:

Philippe Kappeler

DASY - Parameters of Probe: ET3DV6 SN:1547**Sensitivity in Free Space****Diode Compression**

NormX	1.37 $\mu\text{V}/(\text{V}/\text{m})^2$	DCP X	90	mV
NormY	1.23 $\mu\text{V}/(\text{V}/\text{m})^2$	DCP Y	90	mV
NormZ	1.23 $\mu\text{V}/(\text{V}/\text{m})^2$	DCP Z	90	mV

Sensitivity in Tissue Simulating Liquid

Head	835 MHz	$\epsilon_r = 41.5 \pm 5\%$	$\sigma = 0.90 \pm 5\% \text{ mho/m}$
Head	900 MHz	$\epsilon_r = 41.5 \pm 5\%$	$\sigma = 0.97 \pm 5\% \text{ mho/m}$
ConvF X	6.5 $\pm 9.5\%$ (k=2)	Boundary effect:	
ConvF Y	6.5 $\pm 9.5\%$ (k=2)	Alpha	1.47
ConvF Z	6.5 $\pm 9.5\%$ (k=2)	Depth	1.19
Head	1900 MHz	$\epsilon_r = 40.0 \pm 5\%$	$\sigma = 1.40 \pm 5\% \text{ mho/m}$
Head	1800 MHz	$\epsilon_r = 40.0 \pm 5\%$	$\sigma = 1.40 \pm 5\% \text{ mho/m}$
ConvF X	5.2 $\pm 9.5\%$ (k=2)	Boundary effect:	
ConvF Y	5.2 $\pm 9.5\%$ (k=2)	Alpha	0.63
ConvF Z	5.2 $\pm 9.5\%$ (k=2)	Depth	2.21

Boundary Effect

Head	835 MHz	Typical SAR gradient: 5 % per mm		
Probe Tip to Boundary			1 mm	2 mm
SAR _{be} [%]	Without Correction Algorithm		7.1	2.9
SAR _{be} [%]	With Correction Algorithm		0.0	0.1
Head	1900 MHz	Typical SAR gradient: 10 % per mm		
Probe Tip to Boundary			1 mm	2 mm
SAR _{be} [%]	Without Correction Algorithm		13.2	8.5
SAR _{be} [%]	With Correction Algorithm		0.1	0.4

Sensor Offset

Probe Tip to Sensor Center	2.7	mm
Optical Surface Detection	1.3 $\pm$ 0.2	mm

Schmid & Partner Engineering AG

Zeughausstrasse 43, 8004 Zurich, Switzerland, Phone +41 1 245 97 00, Fax +41 1 245 97 79

Additional Conversion Factors for Dosimetric E-Field Probe

Type:

ET3DV6

Serial Number:

1547

Place of Assessment:

Zurich

Date of Assessment:


September 30, 2002

Probe Calibration Date:

September 28, 2002

Schmid & Partner Engineering AG hereby certifies that conversion factor(s) of this probe have been evaluated on the date indicated above. The assessment was performed using the FDTD numerical code SEMCAD of Schmid & Partner Engineering AG. Since the evaluation is coupled with measured conversion factors, it has to be recalculated yearly, i.e., following the re-calibration schedule of the probe. The uncertainty of the numerical assessment is based on the extrapolation from measured value at 900 MHz or at 1800 MHz.

Assessed by:




Dosimetric E-Field Probe ET3DV6 SN:1547

Conversion factor ($\pm$ standard deviation)

150 MHz	ConvF	$8.0 \pm 8\%$	$\epsilon_r = 61.9 \pm 5\%$ $\sigma = 0.80 \pm 5\% \text{ mho/m}$ (body tissue)
236 MHz	ConvF	$7.8 \pm 8\%$	$\epsilon_r = 59.8 \pm 5\%$ $\sigma = 0.87 \pm 5\% \text{ mho/m}$ (body tissue)
300 MHz	ConvF	$7.7 \pm 8\%$	$\epsilon_r = 58.2 \pm 5\%$ $\sigma = 0.92 \pm 5\% \text{ mho/m}$ (body tissue)
350 MHz	ConvF	$7.6 \pm 8\%$	$\epsilon_r = 57.7 \pm 5\%$ $\sigma = 0.93 \pm 5\% \text{ mho/m}$ (body tissue)
450 MHz	ConvF	$7.3 \pm 8\%$	$\epsilon_r = 56.7 \pm 5\%$ $\sigma = 0.94 \pm 5\% \text{ mho/m}$ (body tissue)
784 MHz	ConvF	$6.4 \pm 8\%$	$\epsilon_r = 55.4 \pm 5\%$ $\sigma = 0.97 \pm 5\% \text{ mho/m}$ (body tissue)
1450 MHz	ConvF	$5.3 \pm 8\%$	$\epsilon_r = 54.0 \pm 5\%$ $\sigma = 1.30 \pm 5\% \text{ mho/m}$ (body tissue)


Conversion Factor Assessment



Head	MHz	$\epsilon_r = 39.2 \pm 5\%$	$\sigma = 1.80 \pm 5\% \text{ mho/m}$
ConvF X	4.8 $\pm 8.9\%$ (k=2)	Boundary effect:	
ConvF Y	4.8 $\pm 8.9\%$ (k=2)	Alpha	1.00
ConvF Z	4.8 $\pm 8.9\%$ (k=2)	Depth	1.81
Body	MHz	$\epsilon_r = 52.7 \pm 5\%$	$\sigma = 1.95 \pm 5\% \text{ mho/m}$
ConvF X	4.5 $\pm 8.9\%$ (k=2)	Boundary effect:	
ConvF Y	4.5 $\pm 8.9\%$ (k=2)	Alpha	1.00
ConvF Z	4.5 $\pm 8.9\%$ (k=2)	Depth	1.50

Deviation from Isotropy n HSL

Error ϕ , f 900 MHz



Frequency Response of E-Field

(TEM-Cell:ifi110, Waveguide R22)

