10607- AAA	IEEE 802.11ac WiFi (20MHz, MCS0, 90pc duty cycle)	X	4.59	66.33	16.38	0.46	130.0	± 9.6 %
		Y	4.63	65.88	16.06		130.0	1
		Z	4.60	65.82	15.94		130.0	
10608- AAA	IEEE 802.11ac WiFi (20MHz, MCS1, 90pc duty cycle)	X	4.76	66.70	16.54	0.46	130.0	± 9.6 %
		Y	4.81	66.28	16.23		130.0	
		Z	4.78	66.21	16.10		130.0	
10609- AAA	IEEE 802.11ac WiFi (20MHz, MCS2, 90pc duty cycle)	X	4.65	66.54	16.37	0.46	130.0	± 9.6 %
		Y	4.70	66.12	16.06		130.0	
		Z	4.67	66.05	15.94		130.0	
10610- AAA	IEEE 802.11ac WiFi (20MHz, MCS3, 90pc duty cycle)	X	4.70	66.70	16.54	0.46	130.0	± 9.6 %
		Y	4.75	66.28	16.22	-	130.0	
		Z	4.72	66.20	16.09		130.0	
10611- AAA	IEEE 802.11ac WiFi (20MHz, MCS4, 90pc duty cycle)	X	4.61	66.50	16.38	0.46	130.0	± 9.6 %
		Y	4.67	66.08	16.07		130.0	
		Z	4.64	66.01	15.94		130.0	
10612- AAA	IEEE 802.11ac WiFi (20MHz, MCS5, 90pc duty cycle)	X	4.62	66.66	16.43	0.46	130.0	±9.6 %
		Y	4.67	66.22	16.10		130.0	
		Z	4.64	66.16	15.98		130.0	
10613- AAA	IEEE 802.11ac WiFi (20MHz, MCS6, 90pc duty cycle)	X	4.61	66.49	16.29	0.46	130.0	± 9.6 %
		Y	4.68	66.10	15.99		130.0	
		Z	4.64	66.04	15.87		130.0	
10614- AAA	IEEE 802.11ac WiFi (20MHz, MCS7, 90pc duty cycle)	X	4.57	66.72	16.55	0.46	130.0	± 9.6 %
100 C		Y	4.62	66.30	16.22		130.0	
		Z	4.59	66.21	16.08		130.0	
10615- AAA	IEEE 802.11ac WiFi (20MHz, MCS8, 90pc duty cycle)	X	4.61	66.33	16.15	0.46	130.0	± 9.6 %
		Y	4.66	65.90	15.84	15-1 - 15-	130.0	
		Z	4.64	65.86	15.73		130.0	
10616- AAA	IEEE 802.11ac WiFi (40MHz, MCS0, 90pc duty cycle)	X	5.23	66.65	16.52	0.46	130.0	± 9.6 %
		Y	5.28	66.36	16.27		130.0	
distant and		Z	5.25	66.30	16.15		130.0	
10617- AAA	IEEE 802.11ac WiFi (40MHz, MCS1, 90pc duty cycle)	X	5.29	66.84	16.59	0.46	130.0	± 9.6 %
		Y	5.35	66.53	16.32		130.0	1
-		Z	5.31	66.46	16.20		130.0	
10618- AAA	IEEE 802.11ac WiFi (40MHz, MCS2, 90pc duty cycle)	X	5.19	66.89	16.63	0.46	130.0	± 9.6 %
		Y	5.23	66.54	16.34		130.0	
		Z	5.19	66.46	16.22		130.0	
10619- AAA	IEEE 802.11ac WiFi (40MHz, MCS3, 90pc duty cycle)	X	5.20	66.65	16.45	0.46	130.0	± 9.6 %
		Y	5.25	66.34	16.18		130.0	
		Z	5.21	66.28	16.06		130.0	
10620- AAA	IEEE 802.11ac WiFi (40MHz, MCS4, 90pc duty cycle)	X	5.27	66.66	16.50	0.46	130.0	± 9.6 %
		Y	5.34	66.38	16.25		130.0	
		Z	5.30	66.32	16.13		130.0	
10621- AAA	IEEE 802.11ac WiFi (40MHz, MCS5, 90pc duty cycle)	X	5.29	66.82	16.70	0.46	130.0	± 9.6 %
		Y	5.34	66.52	16.44		130.0	
		Z	5.30	66.45	16.31		130.0	
10622- AAA	IEEE 802.11ac WiFi (40MHz, MCS6, 90pc duty cycle)	X	5.29	66.94	16.76	0.46	130.0	± 9.6 %
		Y	5.35	66.68	16.51		130.0	
		Z	5.31	66.59	16.38		130.0	

10623-	IEEE 802.11ac WiFi (40MHz, MCS7,	X	5.17	66.47	16.39	0.46	130.0	± 9.6 %
AAA	90pc duty cycle)			00.00	10.11			
		Y	5.23	66.20	16.14		130.0	
10624-	IEEE 802.11ac WiFi (40MHz, MCS8,	Z	5.20	66.15	16.03	0.40	130.0	10.0.01
AAA	90pc duty cycle)	X	5.36	66.68	16.56	0.46	130.0	± 9.6 %
		Y	5.42	66.41	16.31		130.0	
10005		Z	5.38	66.35	16.19		130.0	
10625- AAA	IEEE 802.11ac WiFi (40MHz, MCS9, 90pc duty cycle)	X	5.58	67.22	16.88	0.46	130.0	± 9.6 %
		Y	5.76	67.31	16.81		130.0	
40000		Z	5.70	67.18	16.66		130.0	
10626- AAA	IEEE 802.11ac WiFi (80MHz, MCS0, 90pc duty cycle)	X	5.54	66.68	16.46	0.46	130.0	± 9.6 %
		Y	5.58	66.43	16.23		130.0	
		Z	5.54	66.38	16.12		130.0	
10627- AAA	IEEE 802.11ac WiFi (80MHz, MCS1, 90pc duty cycle)	X	5.77	67.26	16.72	0.46	130.0	± 9.6 %
		Y	5.81	66.98	16.47		130.0	
10000		Z	5.76	66.89	16.33		130.0	
10628- AAA	IEEE 802.11ac WiFi (80MHz, MCS2, 90pc duty cycle)	X	5.55	66.70	16.37	0.46	130.0	± 9.6 %
		Y	5.61	66.50	16.16		130.0	
10000		Z	5.57	66.45	16.05		130.0	
10629- AAA	IEEE 802.11ac WiFi (80MHz, MCS3, 90pc duty cycle)	X	5.63	66.79	16.41	0.46	130.0	± 9.6 %
		Y	5.68	66.54	16.18		130.0	
		Z	5.64	66.50	16.07		130.0	
10630- AAA	IEEE 802.11ac WiFi (80MHz, MCS4, 90pc duty cycle)	X	5.96	67.98	17.01	0.46	130.0	± 9.6 %
_		Y	6.09	67.95	16.88		130.0	
		Z	5.99	67.74	16.70		130.0	
10631- AAA	IEEE 802.11ac WiFi (80MHz, MCS5, 90pc duty cycle)	X	5.91	67.94	17.18	0.46	130.0	± 9.6 %
		Y	6.00	67.81	17.01		130.0	
		Z	5.94	67.67	16.84		130.0	
10632- AAA	IEEE 802.11ac WiFi (80MHz, MCS6, 90pc duty cycle)	X	5.75	67.36	16.91	0.46	130.0	± 9.6 %
		Y	5.78	67.05	16.65		130.0	
		Z	5.74	66.95	16.50		130.0	
10633- AAA	IEEE 802.11ac WiFi (80MHz, MCS7, 90pc duty cycle)	X	5.62	66.91	16.51	0.46	130.0	± 9.6 %
		Y	5.67	66.67	16.28		130.0	
		Z	5.64	66.63	16.17		130.0	
10634- AAA	IEEE 802.11ac WiFi (80MHz, MCS8, 90pc duty cycle)	X	5.60	66.94	16.58	0.46	130.0	± 9.6 %
		Y	5.65	66.70	16.35		130.0	
		Z	5.62	66.65	16.24		130.0	
10635- AAA	IEEE 802.11ac WiFi (80MHz, MCS9, 90pc duty cycle)	X	5.46	66.20	15.94	0.46	130.0	± 9.6 %
		Y	5.53	66.01	15.74		130.0	
10000		Z	5.51	66.01	15.66		130.0	
10636- AAA	IEEE 1602.11ac WiFi (160MHz, MCS0, 90pc duty cycle)	X	5.96	67.02	16.54	0.46	130.0	± 9.6 %
		Y	5.99	66.79	16.32		130.0	
		Z	5.95	66.74	16.21		130.0	
10637- AAA	IEEE 1602.11ac WiFi (160MHz, MCS1, 90pc duty cycle)	X	6.11	67.38	16.70	0.46	130.0	± 9.6 %
-		Y	6.14	67.16	16.49		130.0	
		Z	6.10	67.09	16.37		130.0	
10638- AAA	IEEE 1602.11ac WiFi (160MHz, MCS2, 90pc duty cycle)	X	6.11	67.37	16.67	0.46	130.0	± 9.6 %
		Y	6.14	67.14	16.45		130.0	

10639- AAA	IEEE 1602.11ac WiFi (160MHz, MCS3, 90pc duty cycle)	X	6.08	67.29	16.68	0.46	130.0	±9.6 %
		Y	6.12	67.09	16.47		130.0	
		Z	6.08	67.03	16.36	-	130.0	
10640- AAA	IEEE 1602.11ac WiFi (160MHz, MCS4, 90pc duty cycle)	X	6.07	67.28	16.61	0.46	130.0	± 9.6 %
and the second second		Y	6.12	67.09	16.41		130.0	
		Z	6.08	67.04	16.31		130.0	
10641- AAA	IEEE 1602.11ac WiFi (160MHz, MCS5, 90pc duty cycle)	X	6.14	67.25	16.62	0.46	130.0	± 9.6 %
		Y	6.17	67.01	16.39		130.0	
		Z	6.13	66.96	16.29		130.0	
10642- AAA	IEEE 1602.11ac WiFi (160MHz, MCS6, 90pc duty cycle)	X	6.17	67.47	16.89	0.46	130.0	± 9.6 %
		Y	6.21	67.26	16.69		130.0	
		Z	6.17	67.20	16.57		130.0	
10643- AAA	IEEE 1602.11ac WiFi (160MHz, MCS7, 90pc duty cycle)	X	6.01	67.17	16.64	0.46	130.0	± 9.6 %
		Y	6.05	66.94	16.42		130.0	
		Z	6.01	66.89	16.32		130.0	
10644- AAA	IEEE 1602.11ac WiFi (160MHz, MCS8, 90pc duty cycle)	X	6.11	67.48	16.82	0.46	130.0	± 9.6 %
		Y	6.20	67.40	16.67		130.0	
		Z	6.15	67.33	16.56		130.0	
10645- AAA	IEEE 1602.11ac WiFi (160MHz, MCS9, 90pc duty cycle)	X	6.23	67.49	16.78	0.46	130.0	± 9.6 %
		Y	6.46	67.80	16.83		130.0	
		Z	6.39	67.66	16.68		130.0	
10646- AAC	LTE-TDD (SC-FDMA, 1 RB, 5 MHz, QPSK, UL Subframe=2,7)	X	21.90	112.90	37.92	9.30	60.0	± 9.6 %
		Y	18.12	104.94	34.87		60.0	
		Z	20.93	109.66	36.61	Sec. 1	60.0	
10647- AAB	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, QPSK, UL Subframe=2,7)	X	18.90	110.39	37.33	9.30	60.0	± 9.6 %
		Y	16.61	103.75	34.63		60.0	
		Z	18.58	107.78	36.19		60.0	
10648- AAA	CDMA2000 (1x Advanced)	X	0.78	66.29	12.06	0.00	150.0	± 9.6 %
		Y	0.70	63.41	10.80		150.0	
		Z	0.67	62.80	10.34		150.0	

<sup>E</sup> Uncertainty is determined using the max. deviation from linear response applying rectangular distribution and is expressed for the square of the field value.

## Calibration Laboratory of Schmid & Partner

Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland



Schweizerischer Kalibrierdienst

Service suisse d'étalonnage

Servizio svizzero di taratura

Swiss Calibration Service

Accreditation No.: SCS 0108

Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Client UL CCS USA

Certificate No: EX3-3686\_Aug16

S

С

S

## **CALIBRATION CERTIFICATE**

Object	EX3DV4 - SN:3686
Calibration procedure(s)	QA CAL-01.v9, QA CAL-14.v4, QA CAL-23.v5, QA CAL-25.v6 Calibration procedure for dosimetric E-field probes
Calibration date:	August 25, 2016
	uments the traceability to national standards, which realize the physical units of measurements (SI). Incertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature (22 ± 3)°C and humidity < 70%.

Calibration Equipment used (M&TE critical for calibration)

Primary Standards	ID	Cal Date (Certificate No.)	Scheduled Calibration
Power meter NRP	SN: 104778	06-Apr-16 (No. 217-02288/02289)	Apr-17
Power sensor NRP-Z91	SN: 103244	06-Apr-16 (No. 217-02288)	Apr-17
Power sensor NRP-Z91	SN: 103245	06-Apr-16 (No. 217-02289)	Apr-17
Reference 20 dB Attenuator	SN: S5277 (20x)	05-Apr-16 (No. 217-02293)	Apr-17
Reference Probe ES3DV2	SN: 3013	31-Dec-15 (No. ES3-3013_Dec15)	Dec-16
DAE4	SN: 660	23-Dec-15 (No. DAE4-660_Dec15)	Dec-16
Secondary Standards	ID	Check Date (in house)	Scheduled Check
Power meter E4419B	SN: GB41293874	06-Apr-16 (in house check Jun-16)	In house check: Jun-18
Power sensor E4412A	SN: MY41498087	06-Apr-16 (in house check Jun-16)	In house check: Jun-18
Power sensor E4412A	SN: 000110210	06-Apr-16 (in house check Jun-16)	In house check: Jun-18
RF generator HP 8648C	SN: US3642U01700	04-Aug-99 (in house check Jun-16)	In house check: Jun-18
Network Analyzer HP 8753E	SN: US37390585	18-Oct-01 (in house check Oct-15)	In house check: Oct-16

	Name	Function	Signature
Calibrated by:	Leif Klysner	Laboratory Technician	Seef Thep
			20 1
Approved by:	Katja Pokovic	Technical Manager	filly
			Issued: August 26, 2016

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

### **Calibration Laboratory of**

Schmid & Partner **Engineering AG** Zeughausstrasse 43, 8004 Zurich, Switzerland



Schweizerischer Kalibrierdienst

S Service suisse d'étalonnage С

Servizio svizzero di taratura

S Swiss Calibration Service

Accreditation No.: SCS 0108

Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Glossary:	
TSL	tissue simulating liquid
NORMx,y,z	sensitivity in free space
ConvF	sensitivity in TSL / NORMx,y,z
DCP	diode compression point
CF	crest factor (1/duty_cycle) of the RF signal
A, B, C, D	modulation dependent linearization parameters
Polarization $\phi$	φ rotation around probe axis
Polarization 9	9 rotation around an axis that is in the plane normal to probe axis (at measurement center),
	i.e., $\vartheta = 0$ is normal to probe axis
Connector Angle	information used in DASY system to align probe sensor X to the robot coordinate system

## Calibration is Performed According to the Following Standards:

- a) IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013
- b) IEC 62209-1, "Procedure to measure the Specific Absorption Rate (SAR) for hand-held devices used in close proximity to the ear (frequency range of 300 MHz to 3 GHz)", February 2005
- IEC 62209-2, "Procedure to determine the Specific Absorption Rate (SAR) for wireless communication devices used in close proximity to the human body (frequency range of 30 MHz to 6 GHz)", March 2010
- d) KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

## Methods Applied and Interpretation of Parameters:

- NORMx, y, z: Assessed for E-field polarization  $\vartheta = 0$  (f  $\leq 900$  MHz in TEM-cell; f > 1800 MHz: R22 waveguide). NORMx, y, z are only intermediate values, i.e., the uncertainties of NORMx, y, z does not affect the E<sup>2</sup>-field uncertainty inside TSL (see below ConvF).
- NORM(f)x,y,z = NORMx,y,z \* frequency\_response (see Frequency Response Chart). This linearization is . implemented in DASY4 software versions later than 4.2. The uncertainty of the frequency response is included in the stated uncertainty of ConvF.
- DCPx,y,z: DCP are numerical linearization parameters assessed based on the data of power sweep with CW . signal (no uncertainty required). DCP does not depend on frequency nor media.
- PAR: PAR is the Peak to Average Ratio that is not calibrated but determined based on the signal characteristics
- Ax,y,z; Bx,y,z; Cx,y,z; Dx,y,z; VRx,y,z: A, B, C, D are numerical linearization parameters assessed based on the data of power sweep for specific modulation signal. The parameters do not depend on frequency nor media. VR is the maximum calibration range expressed in RMS voltage across the diode.
- ConvF and Boundary Effect Parameters: Assessed in flat phantom using E-field (or Temperature Transfer Standard for f ≤ 800 MHz) and inside waveguide using analytical field distributions based on power measurements for f > 800 MHz. The same setups are used for assessment of the parameters applied for boundary compensation (alpha, depth) of which typical uncertainty values are given. These parameters are used in DASY4 software to improve probe accuracy close to the boundary. The sensitivity in TSL corresponds to NORMx, y, z \* ConvF whereby the uncertainty corresponds to that given for ConvF. A frequency dependent ConvF is used in DASY version 4.4 and higher which allows extending the validity from ± 50 MHz to ± 100 MHz.
- Spherical isotropy (3D deviation from isotropy): in a field of low gradients realized using a flat phantom exposed by a patch antenna.
- Sensor Offset: The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No tolerance required.
- Connector Angle: The angle is assessed using the information gained by determining the NORMx (no uncertainty required).

# Probe EX3DV4

## SN:3686

Manufactured: Calibrated: March 10, 2009 August 25, 2016

Calibrated for DASY/EASY Systems (Note: non-compatible with DASY2 system!)

#### **Basic Calibration Parameters**

	Sensor X	Sensor Y	Sensor Z	Unc (k=2)
Norm $(\mu V/(V/m)^2)^A$	0.46	0.48	0.42	± 10.1 %
DCP (mV) <sup>B</sup>	99.0	96.8	100.7	

## **Modulation Calibration Parameters**

UID	Communication System Name		A dB	B dB√μV	С	D dB	VR mV	Unc <sup>⊏</sup> (k=2)
0	CW	X	0.0	0.0	1.0	0.00	160.0	±3.3 %
-		Y	0.0	0.0	1.0		141.5	
-		Z	0.0	0.0	1.0		149.7	

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

<sup>A</sup> The uncertainties of Norm X,Y,Z do not affect the E<sup>2</sup>-field uncertainty inside TSL (see Pages 5 and 6).

<sup>B</sup> Numerical linearization parameter: uncertainty not required.

<sup>E</sup> Uncertainty is determined using the max. deviation from linear response applying rectangular distribution and is expressed for the square of the field value.

f (MHz) <sup>C</sup>	Relative Permittivity <sup>F</sup>	Conductivity (S/m) <sup>F</sup>	ConvF X	ConvF Y	ConvF Z	Alpha <sup>G</sup>	Depth <sup>G</sup> (mm)	Unc (k=2)
750	41.9	0.89	9.47	9.47	9.47	0.46	0.80	± 12.0 %
900	41.5	0.97	9.22	9.22	9.22	0.30	1.09	± 12.0 %
1750	40.1	1.37	8.22	8.22	8.22	0.35	0.80	± 12.0 %
1900	40.0	1.40	7.90	7.90	7.90	0.34	0.80	± 12.0 %
2300	39.5	1.67	7.47	7.47	7.47	0.36	0.80	± 12.0 %
2450	39.2	1.80	7.04	7.04	7.04	0.40	0.80	± 12.0 %
2600	39.0	1.96	6.96	6.96	6.96	0.37	0.88	± 12.0 %
5250	35.9	4.71	5.18	5.18	5.18	0.40	1.80	± 13.1 %
5600	35.5	5.07	4.44	4.44	4.44	0.50	1.80	± 13.1 %
5750	35.4	5.22	4.58	4.58	4.58	0.50	1.80	± 13.1 %

#### Calibration Parameter Determined in Head Tissue Simulating Media

<sup>C</sup> Frequency validity above 300 MHz of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to ± 50 MHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is ± 10, 25, 40, 50 and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Above 5 GHz frequency validity can be extended to ± 110 MHz.

<sup>F</sup> At frequencies below 3 GHz, the validity of tissue parameters ( $\varepsilon$  and  $\sigma$ ) can be relaxed to  $\pm$  10% if liquid compensation formula is applied to measured SAR values. At frequencies above 3 GHz, the validity of tissue parameters ( $\varepsilon$  and  $\sigma$ ) is restricted to  $\pm$  5%. The uncertainty is the RSS of the ConvF uncertainty for indicated target tissue parameters.

The ConvF uncertainty for indicated target tissue parameters. <sup>G</sup> Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than  $\pm$  1% for frequencies below 3 GHz and below  $\pm$  2% for frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the boundary.


f (MHz) <sup>C</sup>	Relative Permittivity <sup>F</sup>	Conductivity (S/m) <sup>F</sup>	ConvF X	ConvF Y	ConvF Z	Alpha <sup>G</sup>	Depth <sup>G</sup> (mm)	Unc (k=2)
750	55.5	0.96	9.12	9.12	9.12	0.56	0.80	± 12.0 %
900	55.0	1.05	9.12	9.12	9.12	0.43	0.89	± 12.0 %
1750	53.4	1.49	7.74	7.74	7.74	0.42	0.84	± 12.0 %
1900	53.3	1.52	7.46	7.46	7.46	0.45	0.80	± 12.0 %
2300	52.9	1.81	7.37	7.37	7.37	0.40	0.80	± 12.0 %
2450	52.7	1.95	7.12	7.12	7.12	0.40	0.80	± 12.0 %
2600	52.5	2.16	6.97	6.97	6.97	0.29	0.80	± 12.0 %
5250	48.9	5.36	4.34	4.34	4.34	0.50	1.90	± 13.1 %
5600	48.5	5.77	3.67	3.67	3.67	0.55	1.90	± 13.1 %
5750	48.3	5.94	3.87	3.87	3.87	0.60	1.90	± 13.1 %

## Calibration Parameter Determined in Body Tissue Simulating Media

<sup>c</sup> Frequency validity above 300 MHz of  $\pm$  100 MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to  $\pm$  50 MHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is  $\pm$  10, 25, 40, 50 and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Above 5 GHz frequency validity validity can be extended to  $\pm$  110 MHz.


F At frequencies below 3 GHz, the validity of tissue parameters ( $\varepsilon$  and  $\sigma$ ) can be relaxed to  $\pm$  10% if liquid compensation formula is applied to measured SAR values. At frequencies above 3 GHz, the validity of tissue parameters ( $\varepsilon$  and  $\sigma$ ) is restricted to  $\pm$  5%. The uncertainty is the RSS of the Compensation formula target tissue parameters.

the ConvF uncertainty for indicated target tissue parameters. <sup>6</sup> Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than  $\pm$  1% for frequencies below 3 GHz and below  $\pm$  2% for frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the boundary.




## Frequency Response of E-Field (TEM-Cell:ifi110 EXX, Waveguide: R22)

Uncertainty of Frequency Response of E-field: ± 6.3% (k=2)




## Receiving Pattern ( $\phi$ ), $\vartheta = 0^{\circ}$

Uncertainty of Axial Isotropy Assessment: ± 0.5% (k=2)



Dynamic Range f(SAR<sub>head</sub>) (TEM cell , f<sub>eval</sub>= 1900 MHz)

Uncertainty of Linearity Assessment: ± 0.6% (k=2)



## **Conversion Factor Assessment**

## **Other Probe Parameters**

Sensor Arrangement	Triangular
Connector Angle (°)	34.5
Mechanical Surface Detection Mode	enabled
Optical Surface Detection Mode	disabled
Probe Overall Length	337 mm
Probe Body Diameter	10 mm
Tip Length	9 mm
Tip Diameter	2.5 mm
Probe Tip to Sensor X Calibration Point	1 mm
Probe Tip to Sensor Y Calibration Point	1 mm
Probe Tip to Sensor Z Calibration Point	1 mm
Recommended Measurement Distance from Surface	1.4 mm

## Calibration Laboratory of Schmid & Partner

Client

Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland

**UL CCS USA** 



Schweizerischer Kalibrierdienst Service suisse d'étalonnage Servizio svizzero di taratura Świss Calibration Service

Accreditation No.: SCS 0108

Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Certificate No: EX3-3749\_Jan17

S

С

S

## CALIBRATION CERTIFICATE

Object	EX3DV4 - SN:3	749	
Calibration procedure(s)		QA CAL-14.v4, QA CAL-23.v5, QA edure for dosimetric E-field probes	
Calibration date:	January 23, 201	7	NT ST - TOTAL
This calibration certificate do The measurements and the	ocuments the traceability to na uncertainties with confidence	tional standards, which realize the physical units probability are given on the following pages and	s of measurements (SI). are part of the certificate.
All calibrations have been co	onducted in the closed laborate	bry facility: environment temperature (22 $\pm$ 3)°C -	and humidity < 70%.
Calibration Equipment used	(M&TE critical for calibration)		
Primary Standards	ID	Cal Date (Certificate No.)	Scheduled Calibration
Power meter NRP	SN: 104778	06-Apr-16 (No. 217-02288/02289)	Apr-17

Filinary Stanuarus	ID	Cal Date (Certificate No.)	Scheduled Calibration
Power meter NRP	SN: 104778	06-Apr-16 (No. 217-02288/02289)	Apr-17
Power sensor NRP-Z91	SN: 103244	06-Apr-16 (No. 217-02288)	Apr-17
Power sensor NRP-Z91	SN: 103245	06-Apr-16 (No. 217-02289)	Apr-17
Reference 20 dB Attenuator	SN: S5277 (20x)	05-Apr-16 (No. 217-02293)	Apr-17
Reference Probe ES3DV2	SN: 3013	31-Dec-16 (No. ES3-3013_Dec16)	Dec-17
DAE4	SN: 660	7-Dec-16 (No. DAE4-660_Dec16)	Dec-17
Secondary Standards	ID	Check Date (in house)	Scheduled Check
Power meter E4419B	SN: GB41293874	06-Apr-16 (in house check Jun-16)	In house check: Jun-18
Power sensor E4412A	SN: MY41498087	06-Apr-16 (in house check Jun-16)	In house check: Jun-18
Power sensor E4412A	SN: 000110210	06-Apr-16 (in house check Jun-16)	In house check: Jun-18
RF generator HP 8648C	SN: US3642U01700	04-Aug-99 (in house check Jun-16)	In house check: Jun-18
Network Analyzer HP 8753E	SN: US37390585	18-Oct-01 (in house check Oct-16)	In house check: Oct-17

	Name	Function	Signature
Calibrated by:	Jeton Kastrati	Laboratory Technician	92 192
Approved by:	Katja Pokovic	Technical Manager	Al les
			Issued: January 26, 2017

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

## Calibration Laboratory of

Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland



S Schweizerischer Kalibrierdienst

C Service suisse d'étalonnage

S Servizio svizzero di taratura

Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: SCS 0108

Multilateral Agreemer	t for the recognition of calibration certificates
<b>Glossary:</b> TSL NORMx,y,z ConvF	tissue simulating liquid sensitivity in free space sensitivity in TSL / NORMx,y,z
DCP CF A, B, C, D Polarization φ	diode compression point crest factor (1/duty_cycle) of the RF signal modulation dependent linearization parameters φ rotation around probe axis
Polarization &	$\vartheta$ rotation around an axis that is in the plane normal to probe axis (at measurement center), i.e., $\vartheta = 0$ is normal to probe axis

Connector Angle information used in DASY system to align probe sensor X to the robot coordinate system

## Calibration is Performed According to the Following Standards:

- a) IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013
- b) IEC 62209-1, "Procedure to measure the Specific Absorption Rate (SAR) for hand-held devices used in close proximity to the ear (frequency range of 300 MHz to 3 GHz)", February 2005
- c) IEC 62209-2, "Procedure to determine the Specific Absorption Rate (SAR) for wireless communication devices used in close proximity to the human body (frequency range of 30 MHz to 6 GHz)", March 2010
- d) KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

## Methods Applied and Interpretation of Parameters:

- NORMx,y,z: Assessed for E-field polarization 
   9 = 0 (f ≤ 900 MHz in TEM-cell; f > 1800 MHz: R22 waveguide). NORMx,y,z are only intermediate values, i.e., the uncertainties of NORMx,y,z does not affect the E<sup>2</sup>-field uncertainty inside TSL (see below *ConvF*).
- *NORM(f)x,y,z* = *NORMx,y,z* \* *frequency\_response* (see Frequency Response Chart). This linearization is implemented in DASY4 software versions later than 4.2. The uncertainty of the frequency response is included in the stated uncertainty of *ConvF*.
- *DCPx,y,z*: DCP are numerical linearization parameters assessed based on the data of power sweep with CW signal (no uncertainty required). DCP does not depend on frequency nor media.
- *PAR*: PAR is the Peak to Average Ratio that is not calibrated but determined based on the signal characteristics
- *Ax,y,z; Bx,y,z; Cx,y,z; Dx,y,z; VRx,y,z: A, B, C, D* are numerical linearization parameters assessed based on the data of power sweep for specific modulation signal. The parameters do not depend on frequency nor media. *VR* is the maximum calibration range expressed in RMS voltage across the diode.
- ConvF and Boundary Effect Parameters: Assessed in flat phantom using E-field (or Temperature Transfer Standard for f ≤ 800 MHz) and inside waveguide using analytical field distributions based on power measurements for f > 800 MHz. The same setups are used for assessment of the parameters applied for boundary compensation (alpha, depth) of which typical uncertainty values are given. These parameters are used in DASY4 software to improve probe accuracy close to the boundary. The sensitivity in TSL corresponds to NORMx,y,z \* ConvF whereby the uncertainty corresponds to that given for ConvF. A frequency dependent ConvF is used in DASY version 4.4 and higher which allows extending the validity from ± 50 MHz to ± 100 MHz.
- Spherical isotropy (3D deviation from isotropy): in a field of low gradients realized using a flat phantom exposed by a patch antenna.
- Sensor Offset: The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No tolerance required.
- Connector Angle: The angle is assessed using the information gained by determining the NORMx (no uncertainty required).

# Probe EX3DV4

## SN:3749

Manufactured: Calibrated: March 26, 2010 January 23, 2017

Calibrated for DASY/EASY Systems (Note: non-compatible with DASY2 system!)

#### **Basic Calibration Parameters**

	Sensor X	Sensor Y	Sensor Z	Unc (k=2)
Norm $(\mu V/(V/m)^2)^A$	0.48	0.44	0.41	± 10.1 %
DCP (mV) <sup>B</sup>	101.3	99.7	101.1	

#### Modulation Calibration Parameters

UID	Communication System Name		A dB	B dB√μV	С	D dB	VR mV	Unc <sup>E</sup> (k=2)
0	CW	X	0.0	0.0	1.0	0.00	132.7	±3.3 %
		Y	0.0	0.0	1.0		128.4	
		Z	0.0	0.0	1.0		144.1	

Note: For details on UID parameters see Appendix.

#### Sensor Model Parameters

	C1 fF	C2 fF	α V <sup>-1</sup>	T1 ms.V <sup>-2</sup>	T2 ms.V <sup>-1</sup>	T3 ms	T4 V <sup>-2</sup>	T5 V <sup>-1</sup>	Т6
Х	51.64	379.7	34.79	20.44	1.301	5.012	1.211	0.325	1.006
Y	51.62	377.8	34.39	17.07	1.562	4.975	1.591	0.342	1.005
Z	51.15	375.1	34.5	19.05	1.412	4.994	1.783	0.23	1.006

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

<sup>A</sup> The uncertainties of Norm X,Y,Z do not affect the E<sup>2</sup>-field uncertainty inside TSL (see Pages 5 and 6).

<sup>B</sup> Numerical linearization parameter: uncertainty not required. <sup>E</sup> Uncertainty is determined using the max. deviation from linear response applying rectangular distribution and is expressed for the square of the field value.

f (MHz) <sup>C</sup>	Relative Permittivity <sup>F</sup>	Conductivity (S/m) <sup>F</sup>	ConvF X	ConvF Y	ConvF Z	Alpha <sup>G</sup>	Depth <sup>G</sup> (mm)	Unc (k=2)
750	41.9	0.89	9.51	9.51	9.51	0.55	0.80	± 12.0 %
900	41.5	0.97	8.95	8.95	8.95	0.43	0.90	± 12.0 %
1750	40.1	1.37	7.93	7.93	7.93	0.33	0.80	± 12.0 %
1900	40.0	1.40	7.81	7.81	7.81	0.36	0.80	± 12.0 %
2300	39.5	1.67	7.31	7.31	7.31	0.31	0.80	± 12.0 %
2450	39.2	1.80	7.01	7.01	7.01	0.24	0.99	± 12.0 %
2600	39.0	1.96	6.72	6.72	6.72	0.39	0.80	± 12.0 %
5250	35.9	4.71	4.73	4.73	4.73	0.35	1.80	± 13.1 %
5600	35.5	5.07	4.41	4.41	4.41	0.35	1.80	± 13.1 %
5750	35.4	5.22	4.45	4.45	4.45	0.40	1.80	± 13.1 %

## Calibration Parameter Determined in Head Tissue Simulating Media

<sup>C</sup> Frequency validity above 300 MHz of  $\pm$  100 MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to  $\pm$  50 MHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is  $\pm$  10, 25, 40, 50 and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Above 5 GHz frequency validity validity can be extended to  $\pm$  110 MHz.

<sup>F</sup> At frequencies below 3 GHz, the validity of tissue parameters ( $\varepsilon$  and  $\sigma$ ) can be relaxed to ± 10% if liquid compensation formula is applied to measured SAR values. At frequencies above 3 GHz, the validity of tissue parameters ( $\varepsilon$  and  $\sigma$ ) is restricted to ± 5%. The uncertainty is the RSS of the ConvF uncertainty for indicated target tissue parameters. <sup>G</sup> Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is

<sup>G</sup> Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than ± 1% for frequencies below 3 GHz and below ± 2% for frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the boundary.


f (MHz) <sup>C</sup>	Relative Permittivity <sup>F</sup>	Conductivity (S/m) <sup>F</sup>	ConvF X	ConvF Y	ConvF Z	Alpha <sup>G</sup>	Depth <sup>G</sup> (mm)	Unc (k=2)
750	55.5	0.96	9.13	9.13	9.13	0.39	0.91	± 12.0 %
900	55.0	1.05	9.07	9.07	9.07	0.44	0.80	± 12.0 %
1750	53.4	1.49	7.61	7.61	7.61	0.39	0.80	± 12.0 %
1900	53.3	1.52	7.42	7.42	7.42	0.43	0.80	± 12.0 %
2300	52.9	1.81	7.12	7.12	7.12	0.37	0.80	± 12.0 %
2450	52.7	1.95	7.07	7.07	7.07	0.42	0.80	± 12.0 %
2600	52.5	2.16	6.78	6.78	6.78	0.34	0.80	± 12.0 %
5250	48.9	5.36	4.66	4.66	4.66	0.40	1.90	± 13.1 %
5600	48.5	5.77	3.98	3.98	3.98	0.45	1.90	± 13.1 %
5750	48.3	5.94	4.16	4.16	4.16	0.50	1.90	± 13.1 %

## Calibration Parameter Determined in Body Tissue Simulating Media

<sup>C</sup> Frequency validity above 300 MHz of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to ± 50 MHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is ± 10, 25, 40, 50 and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Above 5 GHz frequency validity can be extended to ± 110 MHz. <sup>F</sup> At frequencies below 3 GHz, the validity of tissue parameters ( $\epsilon$  and  $\sigma$ ) can be relaxed to ± 10% if liquid compensation formula is applied to


<sup>L</sup> At frequencies below 3 GHz, the validity of tissue parameters ( $\varepsilon$  and  $\sigma$ ) can be relaxed to ± 10% if liquid compensation formula is applied to measured SAR values. At frequencies above 3 GHz, the validity of tissue parameters ( $\varepsilon$  and  $\sigma$ ) is restricted to ± 5%. The uncertainty is the RSS of the ConvF uncertainty for indicated target tissue parameters.

<sup>G</sup> Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than ± 1% for frequencies below 3 GHz and below ± 2% for frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the boundary.




## Frequency Response of E-Field (TEM-Cell:ifi110 EXX, Waveguide: R22)


Uncertainty of Frequency Response of E-field: ± 6.3% (k=2)



Uncertainty of Axial Isotropy Assessment: ± 0.5% (k=2)



Uncertainty of Linearity Assessment: ± 0.6% (k=2)



## **Conversion Factor Assessment**

## **Other Probe Parameters**

Triangular
119.8
enabled
disabled
337 mm
10 mm
9 mm
2.5 mm
1 mm
1 mm
1 mm
1.4 mm

## **Appendix: Modulation Calibration Parameters**

UID	Communication System Name		A dB	B dBõV	С	D dB	VR mV	Max Unc <sup>E</sup> (k=2)
0	CW	Х	0.00	0.00	1.00	0.00	132.7	± 3.3 %
		Y	0.00	0.00	1.00		128.4	
10010	CAD Validation (Orwans, 400ms, 40ms)	Z	0.00	0.00	1.00	10.00	144.1	
10010- CAA	SAR Validation (Square, 100ms, 10ms)	X	11.00	70.00	30.00	10.00	20.0	±9.6 %
		Y	3.60	68.72	12.74	_	20.0	
10011		Z	3.79	69.72	13.18	0.00	20.0	
10011- CAB	UMTS-FDD (WCDMA)	X	1.19	70.16	17.01	0.00	150.0	± 9.6 %
		Y	1.01	66.90	15.05		150.0	_
10012-	IEEE 802.11b WiFi 2.4 GHz (DSSS, 1	Z X	1.02 1.25	66.92 65.05	15.04 16.05	0.44	150.0 150.0	1000
CAB	Mbps)					0.41		± 9.6 %
_		Y	1.19	63.80	14.98		150.0	
10013-	IEEE 802.11g WiFi 2.4 GHz (DSSS-	Z X	1.21 4.94	63.93 66.80	15.07 17.11	1.46	150.0 150.0	± 9.6 %
CAB	OFDM, 6 Mbps)	× Y		m		1.40		± 9.6 %
			4.89 4.90	66.50 66.58	16.77 16.86		150.0 150.0	
10021- DAC	GSM-FDD (TDMA, GMSK)	Z X	4.90 56.48	107.04	26.68	9.39	50.0	± 9.6 %
DAC		Y	10.25	83.02	19.63		50.0	
		Z	17.13	90.49	22.09		50.0	
10023- DAC	GPRS-FDD (TDMA, GMSK, TN 0)	X	35.11	100.53	25.02	9.57	50.0	± 9.6 %
		Y	9.17	81.33	19.09		50.0	
		Z	14.08	87.68	21.26		50.0	
10024- DAC	GPRS-FDD (TDMA, GMSK, TN 0-1)	X	100.00	111.69	26.09	6.56	60.0	± 9.6 %
		Y	14.12	87.54	19.63		60.0	
		Z	57.30	104.44	24.25		60.0	
10025- DAC	EDGE-FDD (TDMA, 8PSK, TN 0)	X	11.53	96.65	37.16	12.57	50.0	± 9.6 %
		Y	5.18	71.49	24.88		50.0	
		Z	8.33	85.94	32.26		50.0	
10026- DAC	EDGE-FDD (TDMA, 8PSK, TN 0-1)	X	15.21	100.48	34.73	9.56	60.0	± 9.6 %
_		Y	10.06	89.00	29.89		60.0	
10027-	GPRS-FDD (TDMA, GMSK, TN 0-1-2)	Z X	12.39 100.00	94.78 110.91	32.42 24.99	4.80	60.0 80.0	± 9.6 %
DAC				1.00		4.00		1 0.0 /0
		Y	46.17	100.76	22.18		80.0	
10000		Z	100.00	109.78	24.47	2.55	80.0	+0.0.0
10028- DAC	GPRS-FDD (TDMA, GMSK, TN 0-1-2-3)	X	100.00	111.51	24.59	3.55	100.0	± 9.6 %
		Y	100.00	108.83	23.33		100.0	
10020	EDGE-FDD (TDMA, 8PSK, TN 0-1-2)	Z X	100.00 9.13	109.84 88.92	23.83 29.52	7.80	80.0	± 9.6 %
10029- DAC	EDGE-FDD (IDIVIA, OFSK, IN U-I-2)	X Y	6.93	81.68	29.52	1.00	80.0	1 9.0 %
		Z	7.91	85.12	27.78		80.0	
10030- CAA	IEEE 802.15.1 Bluetooth (GFSK, DH1)	X	100.00	109.99	24.86	5.30	70.0	± 9.6 %
57.5.		Y	11.39	84.94	18.15		70.0	
		Z	49.41	101.39	22.66		70.0	
10031- CAA	IEEE 802.15.1 Bluetooth (GFSK, DH3)	X	100.00	112.44	23.69	1.88	100.0	± 9.6 %
of states		Y	100.00	108.11	21.78		100.0	
		Z	100.00	109.39	22.40	1	100.0	

10061- CAB	IEEE 802.11b WiFi 2.4 GHz (DSSS, 11 Mbps)	X	5.79	89.35	24.77	2.04	110.0	± 9.6 %
		Y	2.79	76.39	19.54		110.0	
		Z	3.32	79.25	20.78		110.0	
10062- CAB	IEEE 802.11a/h WiFi 5 GHz (OFDM, 6 Mbps)	X	4.74	66.83	16.60	0.49	100.0	±9.6 %
		Y	4.70	66.57	16.32		100.0	
		Z	4.70	66.59	16.35		100.0	
10063- CAB	IEEE 802.11a/h WiFi 5 GHz (OFDM, 9 Mbps)	X	4.76	66.92	16.70	0.72	100.0	± 9.6 %
		Y	4.71	66.64	16.39		100.0	
		Z	4.72	66.68	16.44		100.0	
10064- CAB	IEEE 802.11a/h WiFi 5 GHz (OFDM, 12 Mbps)	X	5.05	67.18	16.91	0.86	100.0	± 9.6 %
		Y	5.01	66.90	16.60		100.0	
		Z	5.02	66.95	16.66		100.0	
10065- CAB	IEEE 802.11a/h WiFi 5 GHz (OFDM, 18 Mbps)	X	4.93	67.10	17.00	1.21	100.0	± 9.6 %
		Y	4.88	66.79	16.66		100.0	
		Z	4.89	66.85	16.74		100.0	
10066- CAB	IEEE 802.11a/h WiFi 5 GHz (OFDM, 24 Mbps)	X	4.95	67.13	17.16	1.46	100.0	± 9.6 %
		Y	4.90	66.80	16.80		100.0	
		Z	4.91	66.88	16.89		100.0	
10067- CAB	IEEE 802.11a/h WiFi 5 GHz (OFDM, 36 Mbps)	X	5.24	67.22	17.54	2.04	100.0	± 9.6 %
		Y	5.18	66.89	17.17		100.0	
		Z	5.20	67.00	17.28		100.0	
10068- CAB	IEEE 802.11a/h WiFi 5 GHz (OFDM, 48 Mbps)	X	5.31	67.38	17.80	2.55	100.0	± 9.6 %
		Y	5.26	67.01	17.39		100.0	
		Z	5.27	67.14	17.52		100.0	
10069- CAB	IEEE 802.11a/h WiFi 5 GHz (OFDM, 54 Mbps)	X	5.39	67.34	17.97	2.67	100.0	± 9.6 %
		Y	5.33	66.97	17.56		100.0	
		Z	5.35	67.11	17.70		100.0	
10071- CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 9 Mbps)	Х	5.04	66.91	17.41	1.99	100.0	± 9.6 %
		Y	4.99	66.58	17.03		100.0	
		Z	5.01	66.68	17.14		100.0	
10072- CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 12 Mbps)	X	5.04	67.30	17.63	2.30	100.0	± 9.6 %
		Y	4.98	66.92	17.22		100.0	
		Z	5.01	67.04	17.35		100.0	
10073- CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 18 Mbps)	X	5.12	67.49	17.95	2.83	100.0	± 9.6 %
		Y	5.06	67.07	17.51		100.0	
		Z	5.08	67.22	17.66		100.0	
10074- CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 24 Mbps)	X	5.12	67.43	18.12	3.30	100.0	± 9.6 %
		Y	5.05	67.00	17.65		100.0	
		Z	5.08	67.16	17.82		100.0	
10075- CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 36 Mbps)	X	5.19	67.66	18.47	3.82	90.0	± 9.6 %
		Y	5.12	67.20	17.97		90.0	
		Z	5.15	67.38	18.16		90.0	
10076- CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 48 Mbps)	X	5.20	67.44	18.57	4.15	90.0	± 9.6 %
		Y	5.13	66.99	18.06		90.0	
		Z	5.16	67.18	18.27		90.0	
10077- CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 54 Mbps)	X	5.22	67.51	18.66	4.30	90.0	± 9.6 %
	,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,	Y	5.16	67.06	18.15		90.0	
								1

the second states and strengthered

10112- CAD	LTE-FDD (SC-FDMA, 100% RB, 10 MHz, 64-QAM)	X	3.14	68.03	16.34	0.00	150.0	± 9.6 %
		Y	3.06	67.40	15.87		150.0	
		Z	3.05	67.34	15.83		150.0	
10113- CAD	LTE-FDD (SC-FDMA, 100% RB, 5 MHz, 64-QAM)	X	2.92	69.24	16.89	0.00	150.0	± 9.6 %
		Y	2.82	68.40	16.32		150.0	
		Z	2.80	68.24	16.22		150.0	
10114- CAB	IEEE 802.11n (HT Greenfield, 13.5 Mbps, BPSK)	X	5.17	67.38	16.55	0.00	150.0	± 9.6 %
		Y	5.14	67.19	16.34		150.0	
		Z	5.14	67.17	16.34		150.0	
10115- CAB	IEEE 802.11n (HT Greenfield, 81 Mbps, 16-QAM)	X	5.49	67.57	16.65	0.00	150.0	± 9.6 %
		Y	5.47	67.41	16.46		150.0	
100		Z	5.46	67.38	16.45		150.0	
10116- CAB	IEEE 802.11n (HT Greenfield, 135 Mbps, 64-QAM)	X	5.28	67.60	16.59	0.00	150.0	± 9.6 %
_		Y	5.25	67.41	16.38		150.0	
		Z	5.24	67.38	16.37		150.0	
10117- CAB	IEEE 802.11n (HT Mixed, 13.5 Mbps, BPSK)	X	5.15	67.29	16.52	0.00	150.0	± 9.6 %
		Y	5.12	67.12	16.32		150.0	
		Z	5.12	67.09	16.31		150.0	
10118- CAB	IEEE 802.11n (HT Mixed, 81 Mbps, 16- QAM)	X	5.56	67.74	16.74	0.00	150.0	± 9.6 %
		Y	5.53	67.56	16.54		150.0	
		Z	5.52	67.53	16.53		150.0	
10119- CAB	IEEE 802.11n (HT Mixed, 135 Mbps, 64- QAM)	X	5.25	67.53	16.56	0.00	150.0	± 9.6 %
		Y	5.22	67.34	16.36		150.0	
		Z	5.21	67.32	16.35		150.0	
10140- CAC	LTE-FDD (SC-FDMA, 100% RB, 15 MHz, 16-QAM)	X	3.49	68.11	16.33	0.00	150.0	± 9.6 %
		Y	3.42	67.56	15.90		150.0	
		Z	3.41	67.51	15.88		150.0	
10141- CAC	LTE-FDD (SC-FDMA, 100% RB, 15 MHz, 64-QAM)	X	3.61	68.17	16.48	0.00	150.0	± 9.6 %
		Y	3.54	67.66	16.08		150.0	
		Z	3.53	67.61	16.05		150.0	
10142- CAD	LTE-FDD (SC-FDMA, 100% RB, 3 MHz, QPSK)	X	2.22	70.38	16.88	0.00	150.0	± 9.6 %
		Y	2.03	68.55	15.80		150.0	
		Z	2.02	68.47	15.74		150.0	
10143- CAD	LTE-FDD (SC-FDMA, 100% RB, 3 MHz, 16-QAM)	X	2.71	70.41	16.83	0.00	150.0	± 9.6 %
	and the second se	Y	2.55	69.11	16.05		150.0	
		Z	2.52	68.88	15.91		150.0	
10144- CAD	LTE-FDD (SC-FDMA, 100% RB, 3 MHz, 64-QAM)	X	2.42	67.73	15.07	0.00	150.0	± 9.6 %
	Constant in the State of the latter of	Y	2.30	66.72	14.40		150.0	
		Z	2.30	66.66	14.35		150.0	
10145- CAD	LTE-FDD (SC-FDMA, 100% RB, 1.4 MHz, QPSK)	X	1.55	68.40	14.03	0.00	150.0	± 9.6 %
		Y	1.36	66.23	12.82		150.0	
		Z	1.33	65.94	12.62		150.0	
10146- CAD	LTE-FDD (SC-FDMA, 100% RB, 1.4 MHz, 16-QAM)	X	2.77	70.59	14.09	0.00	150.0	± 9.6 %
		Y	2.42	68.28	12.84		150.0	
		Z	2.39	68.32	12.82		150.0	
10147- CAD	LTE-FDD (SC-FDMA, 100% RB, 1.4 MHz, 64-QAM)	X	3.90	75.00	16.05	0.00	150.0	± 9.6 %
		Y	3.03	71.13	14.25		150.0	
		Z	3.00	71.14	14.20		150.0	

10168- CAD	LTE-FDD (SC-FDMA, 50% RB, 1.4 MHz, 64-QAM)	X	5.51	76.77	21.79	3.01	150.0	±9.6 %
		Y	5.66	76.63	21.45		150.0	
		Z	5.58	76.75	21.56		150.0	
10169- CAC	LTE-FDD (SC-FDMA, 1 RB, 20 MHz, QPSK)	X	3.26	70.95	19.84	3.01	150.0	± 9.6 %
		Y	3.37	70.86	19.43		150.0	
		Z	3.29	70.89	19.57		150.0	
10170- CAC	LTE-FDD (SC-FDMA, 1 RB, 20 MHz, 16-QAM)	X	5.20	79.62	23.12	3.01	150.0	± 9.6 %
		Y	5.57	79.74	22.74		150.0	
		Z	5.48	80.18	23.03		150.0	
10171- AAC	LTE-FDD (SC-FDMA, 1 RB, 20 MHz, 64-QAM)	X	3.96	73.86	19.76	3.01	150.0	± 9.6 %
		Y	4.09	73.35	19.13		150.0	
		Z	4.07	73.99	19.52		150.0	
10172- CAC	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, QPSK)	X	17.73	102.33	31.17	6.02	65.0	± 9.6 %
_		Y	8.06	85.83	25.07		65.0	
		Z	10.50	91.75	27.46		65.0	
10173- CAC	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, 16-QAM)	X	30.94	106.72	30.32	6.02	65.0	± 9.6 %
		Y	14.00	91.41	25.10		65.0	
		Z	22.19	100.06	27.98		65.0	
10174- CAC	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, 64-QAM)	X	20.10	98.04	27.29	6.02	65.0	± 9.6 %
		Y	9.09	83.72	22.14		65.0	
		Z	12.38	89.55	24.31		65.0	
10175- CAD	LTE-FDD (SC-FDMA, 1 RB, 10 MHz, QPSK)	X	3.21	70.56	19.56	3.01	150.0	± 9.6 %
		Y	3.31	70.41	19.12		150.0	
		Z	3.24	70.50	19.28		150.0	
10176- CAD	LTE-FDD (SC-FDMA, 1 RB, 10 MHz, 16-QAM)	Х	5.21	79.65	23.13	3.01	150.0	± 9.6 %
		Y	5.58	79.78	22.76		150.0	
		Z	5.49	80.21	23.04		150.0	
10177- CAF	LTE-FDD (SC-FDMA, 1 RB, 5 MHz, QPSK)	X	3.24	70.75	19.67	3.01	150.0	± 9.6 %
		Y	3.35	70.63	19.25		150.0	
		Z	3.27	70.69	19.40		150.0	
10178- CAD	LTE-FDD (SC-FDMA, 1 RB, 5 MHz, 16- QAM)	X	5.12	79.30	22.97	3.01	150.0	± 9.6 %
		Y	5.47	79.36	22.56		150.0	
		Z	5.39	79.84	22.87		150.0	
10179- CAD	LTE-FDD (SC-FDMA, 1 RB, 10 MHz, 64-QAM)	X	4.51	76.53	21.28	3.01	150.0	± 9.6 %
		Y	4.70	76.16	20.70		150.0	
		Z	4.67	76.79	21.08		150.0	
10180- CAD	LTE-FDD (SC-FDMA, 1 RB, 5 MHz, 64- QAM)	X	3.94	73.75	19.70	3.01	150.0	± 9.6 %
		Y	4.07	73.22	19.05		150.0	
		Z	4.05	73.88	19.45		150.0	
10181- CAC	LTE-FDD (SC-FDMA, 1 RB, 15 MHz, QPSK)	X	3.24	70.73	19.66	3.01	150.0	± 9.6 %
		Y	3.34	70.61	19.23		150.0	
		Z	3.27	70.67	19.39		150.0	
10182- CAC	LTE-FDD (SC-FDMA, 1 RB, 15 MHz, 16-QAM)	X	5.11	79.27	22.95	3.01	150.0	± 9.6 %
		Y	5.46	79.32	22.55		150.0	
		Z	5.38	79.80	22.85		150.0	
10183- AAB	LTE-FDD (SC-FDMA, 1 RB, 15 MHz, 64-QAM)	X	3.93	73.72	19.68	3.01	150.0	± 9.6 %
AAB		Y	4.06	73.20	19.04	1	150.0	
		1	4.00	10.20	15.04		100.0	

10223- CAB	IEEE 802.11n (HT Mixed, 90 Mbps, 16- QAM)	X	5.42	67.45	16.61	0.00	150.0	± 9.6 %
		Y	5.40	67.29	16.42		150.0	
		Z	5.39	67.26	16.41		150.0	
10224- CAB	IEEE 802.11n (HT Mixed, 150 Mbps, 64- QAM)	X	5.17	67.42	16.51	0.00	150.0	± 9.6 %
		Y	5.15	67.24	16.31		150.0	
		Z	5.14	67.21	16.30		150.0	
10225- CAB	UMTS-FDD (HSPA+)	X	2.89	66.64	15.77	0.00	150.0	± 9.6 %
		Y	2.84	66.14	15.35		150.0	17
		Z	2.83	66.09	15.31		150.0	
10226- CAA	LTE-TDD (SC-FDMA, 1 RB, 1.4 MHz, 16-QAM)	X	34.49	108.77	30.98	6.02	65.0	± 9.6 %
		Y	15.06	92.73	25.61		65.0	
		Z	24.40	101.80	28.58		65.0	
10227- CAA	LTE-TDD (SC-FDMA, 1 RB, 1.4 MHz, 64-QAM)	Х	26.20	102.36	28.58	6.02	65.0	± 9.6 %
		Y	12.96	89.09	23.92		65.0	
		Z	19.35	96.43	26.42		65.0	
10228- CAA	LTE-TDD (SC-FDMA, 1 RB, 1.4 MHz, QPSK)	X	20.54	105.44	32.18	6.02	65.0	± 9.6 %
		Y	10.78	91.28	26.98		65.0	
		Z	14.62	98.02	29.51		65.0	
10229- CAB	LTE-TDD (SC-FDMA, 1 RB, 3 MHz, 16- QAM)	X	31.17	106.84	30.37	6.02	65.0	± 9.6 %
		Y	14.11	91.53	25.14		65.0	
		Z	22.38	100.18	28.03		65.0	
10230- CAB	LTE-TDD (SC-FDMA, 1 RB, 3 MHz, 64- QAM)	X	24.04	100.81	28.06	6.02	65.0	± 9.6 %
		Y	12.21	88.06	23.51		65.0	
		Z	17.94	95.12	25.95		65.0	
10231- CAB	LTE-TDD (SC-FDMA, 1 RB, 3 MHz, QPSK)	X	19.14	103.97	31.66	6.02	65.0	± 9.6 %
		Y	10.25	90.27	26.56		65.0	
		Z	13.79	96.83	29.06		65.0	
10232- CAC	LTE-TDD (SC-FDMA, 1 RB, 5 MHz, 16- QAM)	X	31.15	106.83	30.36	6.02	65.0	± 9.6 %
		Y	14.08	91.51	25.13		65.0	
		Z	22.35	100.17	28.02		65.0	
10233- CAC	LTE-TDD (SC-FDMA, 1 RB, 5 MHz, 64- QAM)	X	24.01	100.81	28.06	6.02	65.0	± 9.6 %
		Y	12.19	88.05	23.51		65.0	
		Z	17.92	95.11	25.95		65.0	
10234- CAC	LTE-TDD (SC-FDMA, 1 RB, 5 MHz, QPSK)	X	17.91	102.50	31.11	6.02	65.0	± 9.6 %
		Y	9.77	89.27	26.12		65.0	
_		Z	13.05	95.63	28.57		65.0	
10235- CAC	LTE-TDD (SC-FDMA, 1 RB, 10 MHz, 16-QAM)	X	31.24	106.90	30.38	6.02	65.0	± 9.6 %
-		Y	14.09	91.53	25.14		65.0	
		Z	22.39	100.21	28.03		65.0	
10236- CAC	LTE-TDD (SC-FDMA, 1 RB, 10 MHz, 64-QAM)	X	24.31	100.98	28.10	6.02	65.0	± 9.6 %
		Y	12.28	88.14	23.53		65.0	
		Z	18.10	95.24	25.98		65.0	
10237- CAC	LTE-TDD (SC-FDMA, 1 RB, 10 MHz, QPSK)	X	19.26	104.12	31.71	6.02	65.0	± 9.6 %
		Y	10.26	90.32	26.58		65.0	
		Z	13.84	96.92	29.09		65.0	
10238- CAC	LTE-TDD (SC-FDMA, 1 RB, 15 MHz, 16-QAM)	X	31.11	106.82	30.36	6.02	65.0	± 9.6 %
		Y	14.05	91.48	25.12	1	65.0	
		Z	22.31	100.15	28.01		65.0	

January 23, 2017

10255- CAC	LTE-TDD (SC-FDMA, 50% RB, 15 MHz, QPSK)	X	7.56	78.91	21.67	3.98	65.0	± 9.6 %
		Y	6.49	75.72	20.07		65.0	
		Z	6.89	76.92	20.67		65.0	
10256- CAA	LTE-TDD (SC-FDMA, 100% RB, 1.4 MHz, 16-QAM)	X	5.64	73.17	16.36	3.98	65.0	± 9.6 %
		Y	4.82	70.46	14.91		65.0	-
		Z	5.10	71.44	15.40		65.0	
10257- CAA	LTE-TDD (SC-FDMA, 100% RB, 1.4 MHz, 64-QAM)	X	5.48	72.43	15.96	3.98	65.0	± 9.6 %
		Y	4.76	70.00	14.63		65.0	
		Z	5.01	70.87	15.08		65.0	
10258- CAA	LTE-TDD (SC-FDMA, 100% RB, 1.4 MHz, QPSK)	X	5.45	75.66	17.80	3.98	65.0	± 9.6 %
		Y	4.22	71.56	15.84		65.0	
		Z	4.56	72.76	16.42		65.0	
10259- CAB	LTE-TDD (SC-FDMA, 100% RB, 3 MHz, 16-QAM)	X	6.49	76.26	19.97	3.98	65.0	± 9.6 %
		Y	5,64	73.48	18.52		65.0	
		Ζ	5.94	74.46	19.02		65.0	
10260- CAB	LTE-TDD (SC-FDMA, 100% RB, 3 MHz, 64-QAM)	X	6.50	75.98	19.87	3.98	65.0	± 9.6 %
		Y	5.69	73.33	18.48		65.0	
		Z	5.98	74.27	18.96		65.0	
10261- CAB	LTE-TDD (SC-FDMA, 100% RB, 3 MHz, QPSK)	X	8.03	82.10	22.17	3.98	65.0	± 9.6 %
		Y	6.12	77.01	19.88		65.0	
		Z	6.73	78.72	20.66		65.0	
10262- CAC	LTE-TDD (SC-FDMA, 100% RB, 5 MHz, 16-QAM)	X	7.07	77.77	21.51	3.98	65.0	± 9.6 %
		Y	6.22	75.02	20.07		65.0	
		Z	6.53	76.01	20.58		65.0	
10263- CAC	LTE-TDD (SC-FDMA, 100% RB, 5 MHz, 64-QAM)	X	6.66	75.56	20.31	3.98	65.0	± 9.6 %
		Y	5.98	73.22	19.02		65.0	the second second
- 10 C		Z	6.26	74.15	19.52		65.0	
10264- CAC	LTE-TDD (SC-FDMA, 100% RB, 5 MHz, QPSK)	X	8.40	82.43	22.73	3.98	65.0	± 9.6 %
		Y	6.60	77.71	20.58		65.0	
		Z	7.20	79.36	21.34		65.0	
10265- CAC	LTE-TDD (SC-FDMA, 100% RB, 10 MHz, 16-QAM)	X	6.83	75.25	20.50	3.98	65.0	± 9.6 %
		Y	6.22	73.13	19.29		65.0	
		Z	6.48	74.02	19.78		65.0	
10266- CAC	LTE-TDD (SC-FDMA, 100% RB, 10 MHz, 64-QAM)	X	7.24	76.18	21.24	3.98	65.0	± 9.6 %
		Y	6.61	74.11	20.07		65.0	
		Z	6.87	74.95	20.53		65.0	
10267- CAC	LTE-TDD (SC-FDMA, 100% RB, 10 MHz, QPSK)	X	7.90	79.42	21.64	3.98	65.0	± 9.6 %
		Y	6.71	76.11	20.01		65.0	-
		Z	7.14	77.34	20.61		65.0	
10268- CAC	LTE-TDD (SC-FDMA, 100% RB, 15 MHz, 16-QAM)	X	7.40	75.01	20.79	3.98	65.0	± 9.6 %
		Y	6.89	73.32	19.78		65.0	
		Z	7.12	74.04	20.20		65.0	
10269- CAC	LTE-TDD (SC-FDMA, 100% RB, 15 MHz, 64-QAM)	X	7.35	74.58	20.68	3.98	65.0	± 9.6 %
		Y	6.87	72.99	19.71		65.0	
		Z	7.09	73.69	20.12		65.0	
10270- CAC	LTE-TDD (SC-FDMA, 100% RB, 15 MHz, QPSK)	X	7.52	76.67	20.72	3.98	65.0	± 9.6 %
		Y	6.77	74.44	19.51		65.0	
			0.11		10.01		00.0	

10303- AAA	IEEE 802.16e WiMAX (31:15, 5ms, 10MHz, 64QAM, PUSC)	X	5.21	66.48	18.46	4.96	50.0	± 9.6 %
		Y	5.09	65.83	17.99		50.0	
der sin seiter		Z	5.15	66.13	18.18		50.0	
10304- AAA	IEEE 802.16e WIMAX (29:18, 5ms, 10MHz, 64QAM, PUSC)	X	4.98	66.21	17.88	4.17	50.0	±9.6 %
		Y	4.87	65.61	17.45		50.0	
		Z	4.91	65.86	17.59		50.0	
10305- AAA	IEEE 802.16e WiMAX (31:15, 10ms, 10MHz, 64QAM, PUSC, 15 symbols)	X	5.08	70.26	21.12	6.02	35.0	± 9.6 %
		Y	4.89	69.02	20.28		35.0	
		Z	4.98	69.62	20.66		35.0	
10306- AAA	IEEE 802.16e WiMAX (29:18, 10ms, 10MHz, 64QAM, PUSC, 18 symbols)	X	5.14	68.13	20.13	6.02	35.0	± 9.6 %
		Y	5.02	67.29	19.49		35.0	
		Z	5.08	67.70	19.78		35.0	
10307- AAA	IEEE 802.16e WiMAX (29:18, 10ms, 10MHz, QPSK, PUSC, 18 symbols)	X	5.10	68.60	20.24	6.02	35.0	± 9.6 %
	and a second back of the second second second	Y	4.97	67.70	19.57		35.0	
100		Z	5.03	68.12	19.86	0.01	35.0	
10308- AAA	IEEE 802.16e WiMAX (29:18, 10ms, 10MHz, 16QAM, PUSC)	X	5.10	68.90	20.42	6.02	35.0	± 9.6 %
		Y	4.96	67.94	19.73		35.0	
10000		Z	5.03	68.40	20.04	0.00	35.0	
10309- AAA	IEEE 802.16e WiMAX (29:18, 10ms, 10MHz, 16QAM, AMC 2x3, 18 symbols)	X	5.22	68.40	20.29	6.02	35.0	± 9.6 %
		Y	5.08	67.52	19.63		35.0	
		Z	5.15	67.95	19.93		35.0	
10310- AAA	IEEE 802.16e WiMAX (29:18, 10ms, 10MHz, QPSK, AMC 2x3, 18 symbols)	X	5.11	68.30	20.15	6.02	35.0	± 9.6 %
		Y	4.99	67.44	19.51		35.0	
_		Z	5.05	67.86	19.79		35.0	
10311- AAB	LTE-FDD (SC-FDMA, 100% RB, 15 MHz, QPSK)	X	3.35	70.31	16.97	0.00	150.0	± 9.6 %
		Y	3.15	69.05	16.21		150.0	
		Z	3.13	68.96	16.17	0.00	150.0	
10313- AAA	iDEN 1:3	X	4.87	74.46	16.49	6.99	70.0	± 9.6 %
		Y	3.59	70.25	14.63		70.0	11 h
		Z	4.01	71.74	15.33	10.00	70.0	
10314- AAA	iDEN 1:6	X	6.67	81.47	21.78	10.00	30.0	± 9.6 %
		Y	4.37	74.10	18.77		30.0	
10315- AAB	IEEE 802.11b WiFi 2.4 GHz (DSSS, 1 Mbps, 96pc duty cycle)	Z X	<u>4.87</u> 1.14	76.14 64.89	19.70 16.01	0.17	30.0 150.0	± 9.6 %
	mbps, oope duty cycle/	Y	1.10	63.69	14.96		150.0	
11114		Z	1.10	63.74	14.99		150.0	
10316- AAB	IEEE 802.11g WiFi 2.4 GHz (ERP- OFDM, 6 Mbps, 96pc duty cycle)	X	4.64	66.85	16.40	0.17	150.0	± 9.6 %
		Y	4.60	66.59	16.12		150.0	
1178		Z	4.60	66.61	16.15		150.0	
10317- AAB	IEEE 802.11a WiFi 5 GHz (OFDM, 6 Mbps, 96pc duty cycle)	X	4.64	66.85	16.40	0.17	150.0	± 9.6 %
		Y	4.60	66.59	16.12		150.0	
		Z	4.60	66.61	16.15		150.0	
10400- AAC	IEEE 802.11ac WiFi (20MHz, 64-QAM, 99pc duty cycle)	X	4.76	67.21	16.42	0.00	150.0	± 9.6 %
		Y	4.73	66.99	16.18		150.0	
		Z	4.73	66.97	16.18		150.0	
10401- AAC	IEEE 802.11ac WiFi (40MHz, 64-QAM, 99pc duty cycle)	X	5.42	67.27	16.50	0.00	150.0	± 9.6 %
	1	Y	5.39	67.08	16.29		150.0	
		Z	0.00	01.00	10.20		150.0	

10427- AAA	IEEE 802.11n (HT Greenfield, 150 Mbps, 64-QAM)	X	5.41	67.50	16.61	0.00	150.0	± 9.6 %
		Y	5.38	67.33	16.41		150.0	
		Z	5.37	67.31	16.40		150.0	
10430- AAA	LTE-FDD (OFDMA, 5 MHz, E-TM 3.1)	X	4.42	71.44	18.64	0.00	150.0	± 9.6 %
		Y	4.41	71.31	18.50		150.0	
		Z	4.31	70.78	18.19		150.0	
10431- AAA	LTE-FDD (OFDMA, 10 MHz, E-TM 3.1)	X	4.30	67.50	16.44	0.00	150.0	± 9.6 %
		Y	4.26	67.20	16.17		150.0	
		Z	4.25	67.17	16.15		150.0	
10432- AAA	LTE-FDD (OFDMA, 15 MHz, E-TM 3.1)	X	4.59	67.33	16.47	0.00	150.0	± 9.6 %
		Y	4.55	67.09	16.22		150.0	
		Z	4.55	67.06	16.21		150.0	
10433- AAA	LTE-FDD (OFDMA, 20 MHz, E-TM 3.1)	X	4.83	67.29	16.52	0.00	150.0	± 9.6 %
-		Y	4.80	67.09	16.29		150.0	
		Z	4.79	67.06	16.28		150.0	
10434- AAA	W-CDMA (BS Test Model 1, 64 DPCH)	X	4.58	72.50	18.73	0.00	150.0	± 9.6 %
		Y	4.56	72.31	18.57		150.0	
		Z	4.42	71.67	18.20		150.0	
10435- AAB	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	X	100.00	117.61	28.57	3.23	80.0	± 9.6 %
		Y	11.08	87.62	20.24		80.0	
		Z	28.78	99.94	23.77		80.0	
10447- AAA	LTE-FDD (OFDMA, 5 MHz, E-TM 3.1, Clipping 44%)	X	3.62	67.69	15.94	0.00	150.0	± 9.6 %
		Y	3.56	67.23	15.58		150.0	
-		Z	3.55	67.18	15.54		150.0	
10448- AAA	LTE-FDD (OFDMA, 10 MHz, E-TM 3.1, Clippin 44%)	X	4.13	67.28	16.31	0.00	150.0	± 9.6 %
		Y	4.09	66.98	16.03		150.0	
		Z	4.08	66.95	16.01		150.0	
10449- AAA	LTE-FDD (OFDMA, 15 MHz, E-TM 3.1, Cliping 44%)	X	4.39	67.17	16.38	0.00	150.0	± 9.6 %
		Y	4.36	66.92	16.13		150.0	
and the second		Z	4.35	66.90	16.11		150.0	
10450- AAA	LTE-FDD (OFDMA, 20 MHz, E-TM 3.1, Clipping 44%)	X	4.58	67.07	16.38	0.00	150.0	± 9.6 %
		Y	4.55	66.86	16.15		150.0	
		Z	4.55	66.83	16.14		150.0	
10451- AAA	W-CDMA (BS Test Model 1, 64 DPCH, Clipping 44%)	X	3.55	68.00	15.66	0.00	150.0	± 9.6 %
		Y	3.47	67.45	15.26		150.0	
		Z	3.45	67.39	15.21		150.0	
10456- AAA	IEEE 802.11ac WiFi (160MHz, 64-QAM, 99pc duty cycle)	X	6.24	68.03	16.74	0.00	150.0	± 9.6 %
_		Y	6.21	67.90	16.57		150.0	
		Z	6.21	67.88	16.57		150.0	
10457- AAA	UMTS-FDD (DC-HSDPA)	X	3.82	65.50	16.09	0.00	150.0	± 9.6 %
		Y	3.80	65.31	15.86		150.0	
14 Jun 1		Z	3.80	65.29	15.85		150.0	
10458- AAA	CDMA2000 (1xEV-DO, Rev. B, 2 carriers)	X	3.36	67.29	15.09	0.00	150.0	± 9.6 %
		Y	3.29	66.74	14.69		150.0	
		Z	3.28	66.73	14.66		150.0	
10459- AAA	CDMA2000 (1xEV-DO, Rev. B, 3 carriers)	X	4.44	65.40	15.86	0.00	150.0	± 9.6 %
		Y	4.28	64.66	15.37		150.0	
		Z	4.30	64.77	15.42		150.0	

10477- AAB	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, 16- QAM, UL Subframe=2,3,4,7,8,9)	X	2.35	67.50	11.72	3.23	80.0	± 9.6 %
		Y	1.36	61.70	8.84		80.0	
		Z	1.42	62.54	9.22		80.0	
10478- AAB	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, 64- QAM, UL Subframe=2,3,4,7,8,9)	X	1.49	62.74	9.24	3.23	80.0	± 9.6 %
		Y	1.16	60.00	7.61		80.0	
		Z	1.15	60.32	7.73		80.0	
10479- AAA	LTE-TDD (SC-FDMA, 50% RB, 1.4 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	X	9.37	87.95	23.43	3.23	80.0	± 9.6 %
		Y	4.46	75.85	18.76		80.0	
		Ζ	5.38	78.99	20.06		80.0	
10480- AAA	LTE-TDD (SC-FDMA, 50% RB, 1.4 MHz, 16-QAM, UL Subframe=2,3,4,7,8,9)	X	8.95	82.27	19.78	3.23	80.0	± 9.6 %
_		Y	4.36	72.11	15.79		80.0	
40404		Z	5.23	74.80	16.88	0.00	80.0	
10481- AAA	LTE-TDD (SC-FDMA, 50% RB, 1.4 MHz, 64-QAM, UL Subframe=2,3,4,7,8,9)	X	6.96	78.34	18.11	3.23	80.0	± 9.6 %
		Y	3.79	69.93	14.62		80.0	
40400		Z	4.38	72.09	15.53	0.00	80.0	
10482- AAA	LTE-TDD (SC-FDMA, 50% RB, 3 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	X	3.90	74.48	17.74	2.23	80.0	± 9.6 %
		Y	2.52	68.00	14.77		80.0	
10400		Z	2.76	69.32	15.40	0.00	80.0	1000
10483- AAA	LTE-TDD (SC-FDMA, 50% RB, 3 MHz, 16-QAM, UL Subframe=2,3,4,7,8,9)	X	5.37	75.29	17.59	2.23	80.0	± 9.6 %
_		Y	3.50	69.05	14.77		80.0	
10101	LTC TOD (CO EDMA 500/ DD O MUL	Z	3.87	70.55	15.47	0.00	80.0	
10484- AAA	LTE-TDD (SC-FDMA, 50% RB, 3 MHz, 64-QAM, UL Subframe=2,3,4,7,8,9)	X	5.01	74.15	17.17	2.23	80.0	± 9.6 %
and the second		Y	3.42	68.54	14.57	-	80.0	1000
40405	LTE TOD (00 FDMA 500/ DD 5 MU	Z	3.74	69.90	15.21	0.00	80.0	
10485- AAB	LTE-TDD (SC-FDMA, 50% RB, 5 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	X	4.23	75.74	19.09	2.23	80.0	± 9.6 %
-		Y	2.89	69.57	16.23		80.0	-
10486-	LTE-TDD (SC-FDMA, 50% RB, 5 MHz,	Z X	<u>3.17</u> 3.76	70.97	16.90 16.75	2.23	80.0 80.0	1069/
AAB	16-QAM, UL Subframe=2,3,4,7,8,9)	^ Y	3.02	67.29	14.91	2.23		± 9.6 %
		Z	3.18	68.11	15.33		80.0	
10487-	LTE-TDD (SC-FDMA, 50% RB, 5 MHz,	X	3.74	70.39	16.55	2.23	80.0	+060/
AAB	64-QAM, UL Subframe=2,3,4,7,8,9)	^ Y	3.05	67.09	14.82	2.23		± 9.6 %
				67.85	14.02		80.0	
10488- AAB	LTE-TDD (SC-FDMA, 50% RB, 10 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	Z X	3.20 4.37	74.52	19.25	2.23	80.0 80.0	± 9.6 %
		Y	3.36	69.96	17.01		80.0	1.2.1
P. D.M. D.		Z	3.61	71.11	17.58	1.18	80.0	
10489- AAB	LTE-TDD (SC-FDMA, 50% RB, 10 MHz, 16-QAM, UL Subframe=2,3,4,7,8,9)	X	3.96	70.27	17.64	2.23	80.0	± 9.6 %
		Y	3.45	67.74	16.21		80.0	
		Z	3.59	68.41	16.58		80.0	
10490- AAB	LTE-TDD (SC-FDMA, 50% RB, 10 MHz, 64-QAM, UL Subframe=2,3,4,7,8,9)	X	4.04	70.03	17.56	2.23	80.0	± 9.6 %
		Y	3.56	67.67	16.21		80.0	
		Z	3.69	68.31	16.56		80.0	
10491- AAB	LTE-TDD (SC-FDMA, 50% RB, 15 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	X	4.44	72.57	18.61	2.23	80.0	± 9.6 %
		Y	3.71	69.31	16.90		80.0	
		Z	3.90	70.18	17.36		80.0	
10492- AAB	LTE-TDD (SC-FDMA, 50% RB, 15 MHz, 16-QAM, UL Subframe=2,3,4,7,8,9)	X	4.25	69.36	17.55	2.23	80.0	± 9.6 %
		Y	3.86	67.48	16.41		80.0	
		Z	3.98	68.02	16.73		80.0	2

10508- AAB	LTE-TDD (SC-FDMA, 100% RB, 10 MHz, 64-QAM, UL	X	4.35	69.42	17.61	2.23	80.0	± 9.6 %
	Subframe=2,3,4,7,8,9)							
		Y	3.96	67.59	16.50		80.0	
40500		Z	4.08	68.11	16.81		80.0	
10509- AAB	LTE-TDD (SC-FDMA, 100% RB, 15 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	X	5.05	72.50	18.42	2.23	80.0	± 9.6 %
		Y	4.32	69.71	16.95		80.0	
		Z	4.52	70.46	17.35		80.0	
10510- AAB	LTE-TDD (SC-FDMA, 100% RB, 15 MHz, 16-QAM, UL Subframe=2,3,4,7,8,9)	X	4.75	69.38	17.65	2.23	80.0	± 9.6 %
		Y	4.39	67.79	16.67		80.0	
		Z	4.50	68.26	16.95		80.0	
10511- AAB	LTE-TDD (SC-FDMA, 100% RB, 15 MHz, 64-QAM, UL Subframe=2,3,4,7,8,9)	X	4.79	69.08	17.57	2.23	80.0	± 9.6 %
		Y	4.45	67.61	16.65		80.0	
		Z	4.56	68.05	16.92		80.0	
10512- AAB	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	Х	5.40	74.31	18.97	2.23	80.0	± 9.6 %
		Y	4.40	70.76	17.22		80.0	
		Z	4.66	71.68	17.68		80.0	
10513- AAB	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, 16-QAM, UL Subframe=2,3,4,7,8,9)	X	4.66	69.73	17.78	2.23	80.0	± 9.6 %
		Y	4.26	67.98	16.72		80.0	
		Z	4.38	68.48	17.02		80.0	-
10514- AAB	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, 64-QAM, UL Subframe=2,3,4,7,8,9)	X	4.65	69.25	17.64	2.23	80.0	± 9.6 %
		Y	4.30	67.67	16.66		80.0	
		Z	4.41	68.13	16.94		80.0	
10515- AAA	IEEE 802.11b WiFi 2.4 GHz (DSSS, 2 Mbps, 99pc duty cycle)	X	1.00	64.08	15.52	0.00	150.0	± 9.6 %
		Y	0.97	63.05	14.55		150.0	
		Z	0.97	63.02	14.52		150.0	
10516- AAA	IEEE 802.11b WiFi 2.4 GHz (DSSS, 5.5 Mbps, 99pc duty cycle)	X	0.87	78.01	21.44	0.00	150.0	± 9.6 %
		Y	0.55	68.46	16.40		150.0	
		Z	0.56	68.55	16.39		150.0	
10517- AAA	IEEE 802.11b WiFi 2.4 GHz (DSSS, 11 Mbps, 99pc duty cycle)	X	0.88	66.85	16.68	0.00	150.0	± 9.6 %
		Y	0.82	64.71	15.07		150.0	
		Ζ	0.82	64.69	15.03		150.0	1.000
10518- AAA	IEEE 802.11a/h WiFi 5 GHz (OFDM, 9 Mbps, 99pc duty cycle)	X	4.58	66.95	16.36	0.00	150.0	± 9.6 %
		Y	4.55	66.74	16.13		150.0	
		Z	4.55	66.72	16.12		150.0	
10519- AAA	IEEE 802.11a/h WiFi 5 GHz (OFDM, 12 Mbps, 99pc duty cycle)	X	4.77	67.19	16.47	0.00	150.0	± 9.6 %
		Y	4.75	66.98	16.25		150.0	
		Z	4.74	66.96	16.24		150.0	
10520- AAA	IEEE 802.11a/h WiFi 5 GHz (OFDM, 18 Mbps, 99pc duty cycle)	X	4.63	67.17	16.41	0.00	150.0	± 9.6 %
		Y	4.60	66.95	16.17		150.0	
10.00		Z	4.59	66.92	16.16	0.00	150.0	
10521- AAA	IEEE 802.11a/h WiFi 5 GHz (OFDM, 24 Mbps, 99pc duty cycle)	X	4.56	67.18	16.40	0.00	150.0	± 9.6 %
		Y	4.53	66.95	16.16		150.0	
10505		Z	4.52	66.92	16.15	0.00	150.0	10.00
10522- AAA	IEEE 802.11a/h WiFi 5 GHz (OFDM, 36 Mbps, 99pc duty cycle)	X	4.62	67.24	16.47	0.00	150.0	± 9.6 %
		Y	4.59	67.01	16.23		150.0	
		Z	4.58	66.99	16.22		150.0	

10541- AAA	IEEE 802.11ac WiFi (40MHz, MCS7, 99pc duty cycle)	X	5.17	66.65	16.20	0.00	150.0	±9.6 %
		Y	5.14	66.47	16.00		150.0	
		Z	5.13	66.44	15.99	1.000	150.0	
10542- AAA	IEEE 802.11ac WiFi (40MHz, MCS8, 99pc duty cycle)	X	5.32	66.70	16.24	0.00	150.0	±9.6 %
		Y	5.29	66.53	16.04		150.0	
		Z	5.28	66.50	16.03		150.0	
10543- AAA	IEEE 802.11ac WiFi (40MHz, MCS9, 99pc duty cycle)	X	5.40	66.73	16.27	0.00	150.0	±9.6 %
		Y	5.37	66.56	16.07		150.0	
		Z	5.36	66.53	16.06		150.0	
10544- AAA	IEEE 802.11ac WiFi (80MHz, MCS0, 99pc duty cycle)	X	5.48	66.76	16.16	0.00	150.0	± 9.6 %
		Y	5.45	66.59	15.97		150.0	
		Z	5.44	66.56	15.96		150.0	
10545- AAA	IEEE 802.11ac WiFi (80MHz, MCS1, 99pc duty cycle)	X	5.66	67.13	16.29	0.00	150.0	± 9.6 %
-		Y	5.63	66.94	16.09		150.0	
		Z	5.62	66.92	16.08		150.0	
10546- AAA	IEEE 802.11ac WiFi (80MHz, MCS2, 99pc duty cycle)	X	5.55	66.98	16.24	0.00	150.0	± 9.6 %
		Y	5.52	66.81	16.04		150.0	
1.1.1		Z	5.51	66.78	16.03		150.0	
10547- AAA	IEEE 802.11ac WiFi (80MHz, MCS3, 99pc duty cycle)	X	5.62	67.01	16.24	0.00	150.0	± 9.6 %
-	in the second	Y	5.59	66.84	16.05		150.0	
		Z	5.58	66.81	16.04		150.0	
10548- AAA	IEEE 802.11ac WiFi (80MHz, MCS4, 99pc duty cycle)	X	5.84	67.83	16.62	0.00	150.0	± 9.6 %
		Y	5.78	67.57	16.39		150.0	
		Z	5.77	67.56	16.39		150.0	
10550- AAA	IEEE 802.11ac WiFi (80MHz, MCS6, 99pc duty cycle)	X	5.57	66.97	16.24	0.00	150.0	± 9.6 %
		Y	5.54	66.80	16.05		150.0	
		Z	5.53	66.77	16.04		150.0	
10551- AAA	IEEE 802.11ac WiFi (80MHz, MCS7, 99pc duty cycle)	X	5.58	67.03	16.23	0.00	150.0	± 9.6 %
		Y	5.55	66.85	16.04		150.0	
		Z	5.54	66.83	16.03		150.0	
10552- AAA	IEEE 802.11ac WiFi (80MHz, MCS8, 99pc duty cycle)	X	5.50	66.84	16.15	0.00	150.0	± 9.6 %
		Y	5.47	66.67	15.96		150.0	
		Z	5.46	66.64	15.95		150.0	
10553- AAA	IEEE 802.11ac WiFi (80MHz, MCS9, 99pc duty cycle)	X	5.58	66.87	16.19	0.00	150.0	± 9.6 %
		Y	5.56	66.71	16.01		150.0	
		Z	5.55	66.68	16.00		150.0	
10554- AAA	IEEE 1602.11ac WiFi (160MHz, MCS0, 99pc duty cycle)	X	5.88	67.10	16.24	0.00	150.0	± 9.6 %
		Y	5.85	66.95	16.06		150.0	
		Z	5.84	66.92	16.05		150.0	
10555- AAA	IEEE 1602.11ac WiFi (160MHz, MCS1, 99pc duty cycle)	X	6.00	67.39	16.35	0.00	150.0	± 9.6 %
		Y	5.97	67.22	16.17		150.0	
		Z	5.96	67.19	16.16		150.0	
10556- AAA	IEEE 1602.11ac WiFi (160MHz, MCS2, 99pc duty cycle)	X	6.02	67.43	16.37	0.00	150.0	± 9.6 %
		Y	5.99	67.26	16.18		150.0	
		Z	5.98	67.24	16.18		150.0	
10557- AAA	IEEE 1602.11ac WiFi (160MHz, MCS3, 99pc duty cycle)	X	6.00	67.36	16.36	0.00	150.0	± 9.6 %
		Y	5.96	67.20	16.17		150.0	
		Z	5.96	67.17	16.16		150.0	

January 23, 2017

10575- AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS- OFDM, 6 Mbps, 90pc duty cycle)	X	4.68	66.75	16.49	0.46	130.0	± 9.6 %
	Or Divi, o wops, sope duty cycle)	Y	4.65	66.49	16.21		130.0	
		Z	4.65	66.52	16.24		130.0	_
10576-	IEEE 802.11g WiFi 2.4 GHz (DSSS-	X	4.03	66.92	16.56	0.46	130.0	± 9.6 %
AAA	OFDM, 9 Mbps, 90pc duty cycle)	^	4.71	00.52	10.00	0.40	130.0	1 5.0 %
	Of Divi, 9 Mbps, 90pc duty cycle)	Y	4.68	66.67	16.29		130.0	
-		Z	4.68	66.69	16.31	-	130.0	
10577			4.00	67.22	16.73	0.46	130.0	± 9.6 %
10577- AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS- OFDM, 12 Mbps, 90pc duty cycle)	X				0.40		I 9.0 %
_		Y	4.89	66.98	16.47	_	130.0	
		Z	4.89	66.99	16.49	0.40	130.0	
10578- AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS- OFDM, 18 Mbps, 90pc duty cycle)	X	4.82	67.40	16.84	0.46	130.0	± 9.6 %
and an effective		Y	4.79	67.15	16.58		130.0	
		Z	4.78	67.14	16.59		130.0	
10579- AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS- OFDM, 24 Mbps, 90pc duty cycle)	Х	4.58	66.67	16.14	0.46	130.0	± 9.6 %
		Y	4.54	66.37	15.83		130.0	
		Z	4.54	66.42	15.89		130.0	
10580- AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS- OFDM, 36 Mbps, 90pc duty cycle)	X	4.63	66.69	16.16	0.46	130.0	± 9.6 %
		Y	4.58	66.38	15.84		130.0	
		Z	4.59	66.44	15.90		130.0	
10581- AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS- OFDM, 48 Mbps, 90pc duty cycle)	Х	4.72	67.44	16.78	0.46	130.0	± 9.6 %
7001		Y	4.68	67.17	16.50		130.0	
		Z	4.68	67.17	16.51		130.0	
10582- AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS- OFDM, 54 Mbps, 90pc duty cycle)	X	4.52	66.41	15.93	0.46	130.0	± 9.6 %
~~~		Y	4.48	66.10	15.60		130.0	
-		Z	4.49	66.17	15.67		130.0	
10583- AAA	IEEE 802.11a/h WiFi 5 GHz (OFDM, 6 Mbps, 90pc duty cycle)	X	4.68	66.75	16.49	0.46	130.0	± 9.6 %
AAA	Mbps, sope duty cycle)	Y	4.65	66.49	16.21		130.0	
		Z	4.65	66.52	16.24		130.0	
10584-	IEEE 802.11a/h WiFi 5 GHz (OFDM, 9 Mbps, 90pc duty cycle)	X	4.71	66.92	16.56	0.46	130.0	± 9.6 %
AAA	Mibbs, sope duty cycle)	Y	4.68	66.67	16.29		130.0	
_		Z	4.68	66.69	16.31		130.0	
10585-	IEEE 802.11a/h WiFi 5 GHz (OFDM, 12 Mbps, 90pc duty cycle)	X	4.00	67.22	16.73	0.46	130.0	± 9.6 %
AAA	wibps, sope duty cycle)	Y	4.89	66.98	16.47		130.0	
-			4.89	66.99	16.49		130.0	
10586- AAA	IEEE 802.11a/h WiFi 5 GHz (OFDM, 18 Mbps, 90pc duty cycle)	X	4.82	67.40	16.84	0.46	130.0	± 9.6 %
	inspo, cope daty byold	Y	4.79	67.15	16.58		130.0	
		Z	4.78	67.14	16.59		130.0	
10587- AAA	IEEE 802.11a/h WiFi 5 GHz (OFDM, 24 Mbps, 90pc duty cycle)	X	4.58	66.67	16.14	0.46	130.0	± 9.6 %
7.0-0-1	mopo, oopo daty oyoioj	Y	4.54	66.37	15.83		130.0	
		Z	4.54	66.42	15.89		130.0	-
10588- AAA	IEEE 802.11a/h WiFi 5 GHz (OFDM, 36 Mbps, 90pc duty cycle)	X	4.63	66.69	16.16	0.46	130.0	± 9.6 %
////	wopa, oope duty cycle)	Y	4.58	66.38	15.84		130.0	
		Z	4.59	66.44	15.90		130.0	
10589- AAA	IEEE 802.11a/h WiFi 5 GHz (OFDM, 48 Mbps, 90pc duty cycle)	X	4.72	67.44	16.78	0.46	130.0	± 9.6 %
		Y	4.68	67.17	16.50	-	130.0	
		Z	4.68	67.17	16.51		130.0	
10590-	IEEE 802.11a/h WiFi 5 GHz (OFDM, 54	X	4.52	66.41	15.93	0.46	130.0	± 9.6 %
AAA	Mbps, 90pc duty cycle)	Y	4.48	66.10	15.60		130.0	

THE R. LEWIS CO., Name and Address of the Owner, Name and Name and Name and Name and Name and Name and Name and

10607- AAA	IEEE 802.11ac WiFi (20MHz, MCS0, 90pc duty cycle)	X	4.67	66.14	16.22	0.46	130.0	± 9.6 %
		Y	4.63	65.87	15.94		130.0	
		Z	4.64	65.89	15.96		130.0	
10608- AAA	IEEE 802.11ac WiFi (20MHz, MCS1, 90pc duty cycle)	X	4.87	66.55	16.38	0.46	130.0	± 9.6 %
		Y	4.82	66.28	16.11		130.0	
1.1.1.1		Z	4.82	66.29	16.13		130.0	
10609- AAA	IEEE 802.11ac WiFi (20MHz, MCS2, 90pc duty cycle)	X	4.76	66.41	16.23	0.46	130.0	± 9.6 %
		Y	4.71	66.11	15.94		130.0	
		Z	4.71	66.14	15.97		130.0	
10610- AAA	IEEE 802.11ac WiFi (20MHz, MCS3, 90pc duty cycle)	X	4.81	66.57	16.39	0.46	130.0	± 9.6 %
		Y	4.76	66.29	16.11		130.0	
		Z	4.76	66.30	16.13		130.0	
10611- AAA	IEEE 802.11ac WiFi (20MHz, MCS4, 90pc duty cycle)	X	4.72	66.37	16.24	0.46	130.0	± 9.6 %
		Y	4.68	66.08	15.94		130.0	
		Z	4.68	66.10	15.97		130.0	
10612- AAA	IEEE 802.11ac WiFi (20MHz, MCS5, 90pc duty cycle)	X	4.74	66.53	16.28	0.46	130.0	± 9.6 %
		Y	4.68	66.20	15.97		130.0	
		Z	4.69	66.25	16.01		130.0	
10613- AAA	IEEE 802.11ac WiFi (20MHz, MCS6, 90pc duty cycle)	X	4.74	66.42	16.17	0.46	130.0	± 9.6 %
		Y	4.69	66.11	15.86		130.0	
-		Z	4.69	66.14	15.90		130.0	
10614- AAA	IEEE 802.11ac WiFi (20MHz, MCS7, 90pc duty cycle)	X	4.68	66.62	16.41	0.46	130.0	± 9.6 %
		Y	4.64	66.33	16.12		130.0	
		Z	4.64	66.33	16.14		130.0	
10615- AAA	IEEE 802.11ac WiFi (20MHz, MCS8, 90pc duty cycle)	X	4.72	66.19	16.01	0.46	130.0	± 9.6 %
		Y	4.67	65.88	15.70		130.0	
		Z	4.68	65.93	15.75		130.0	
10616- AAA	IEEE 802.11ac WiFi (40MHz, MCS0, 90pc duty cycle)	X	5.31	66.60	16.39	0.46	130.0	± 9.6 %
		Y	5.27	66.38	16.14		130.0	
		Z	5.28	66.39	16.16		130.0	
10617- AAA	IEEE 802.11ac WiFi (40MHz, MCS1, 90pc duty cycle)	X	5.37	66.74	16.42	0.46	130.0	± 9.6 %
		Y	5.33	66.50	16.17		130.0	
		Z	5.33	66.51	16.19		130.0	
10618- AAA	IEEE 802.11ac WiFi (40MHz, MCS2, 90pc duty cycle)	X	5.27	66.78	16.47	0.46	130.0	± 9.6 %
		Y	5.22	66.55	16.21		130.0	
		Z	5.22	66.55	16.23	-	130.0	
10619- AAA	IEEE 802.11ac WiFi (40MHz, MCS3, 90pc duty cycle)	X	5.28	66.58	16.30	0.46	130.0	± 9.6 %
		Y	5.24	66.34	16.04		130.0	
		Z	5.24	66.36	16.07		130.0	
10620- AAA	IEEE 802.11ac WiFi (40MHz, MCS4, 90pc duty cycle)	X	5.38	66.63	16.37	0.46	130.0	± 9.6 %
		Y	5.34	66.40	16.12		130.0	
		Z	5.34	66.42	16.15		130.0	
10621- AAA	IEEE 802.11ac WiFi (40MHz, MCS5, 90pc duty cycle)	×	5.37	66.76	16.55	0.46	130.0	± 9.6 %
		Y	5.34	66.56	16.33		130.0	
		Z	5.34	66.55	16.33		130.0	
10622- AAA	IEEE 802.11ac WiFi (40MHz, MCS6, 90pc duty cycle)	X	5.38	66.90	16.61	0.46	130.0	± 9.6 %
		Y	5.34	66.67	16.37	1	130.0	
		Z	5.34	66.67	16.38		130.0	

A DESCRIPTION OF THE OWNER OWNER

10639-	IEEE 1602.11ac WiFi (160MHz, MCS3,	X	6.14	67.32	16.57	0.46	130.0	± 9.6 %
AAA	90pc duty cycle)						-	
		Y	6.10	67.13	16.35		130.0	
		Z	6.10	67.13	16.37		130.0	
10640- AAA	IEEE 1602.11ac WiFi (160MHz, MCS4, 90pc duty cycle)	X	6.14	67.33	16.51	0.46	130.0	19.6%
		Y	6.10	67.11	16.28		130.0	
		Z	6.10	67.13	16.31		130.0	
10641- AAA	IEEE 1602.11ac WiFi (160MHz, MCS5, 90pc duty cycle)	X	6.18	67.19	16.46	0.46	130.0	± 9.6 %
		Y	6.13	66.98	16.23		130.0	
		Z	6.14	67.01	16.27		130.0	
10642- AAA	IEEE 1602.11ac WiFi (160MHz, MCS6, 90pc duty cycle)	X	6.24	67.50	16.79	0.46	130.0	± 9.6 %
		Y	6.20	67.33	16.58		130.0	
		Z	6.20	67.32	16.59		130.0	
10643- AAA	IEEE 1602.11ac WiFi (160MHz, MCS7, 90pc duty cycle)	X	6.06	67.15	16.51	0.46	130.0	± 9.6 %
		Y	6.02	66.94	16.28		130.0	
NH I		Z	6.02	66.96	16.31		130.0	
10644- AAA	IEEE 1602.11ac WiFi (160MHz, MCS8, 90pc duty cycle)	X	6.23	67.67	16.79	0.46	130.0	± 9.6 %
		Y	6.18	67.44	16.55		130.0	
		Z	6.18	67.45	16.58		130.0	
10645- AAA	IEEE 1602.11ac WiFi (160MHz, MCS9, 90pc duty cycle)	X	6.60	68.36	17.08	0.46	130.0	± 9.6 %
		Y	6.53	68.07	16.81		130.0	
		Z	6.53	68.07	16.84		130.0	
10646- AAC	LTE-TDD (SC-FDMA, 1 RB, 5 MHz, QPSK, UL Subframe=2,7)	X	28.99	114.83	37.66	9.30	60.0	± 9.6 %
		Y	14.23	96.89	31.19		60.0	
		Z	20.31	105.98	34.61		60.0	
10647- AAB	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, QPSK, UL Subframe=2,7)	X	26.72	113.84	37.52	9.30	60.0	± 9.6 %
		Y	13.29	96.12	31.05		60.0	
		Z	18.78	105.01	34.45		60.0	
10648- AAA	CDMA2000 (1x Advanced)	X	0.81	65.75	12.47	0.00	150.0	± 9.6 %
		Y	0.72	63.77	11.22		150.0	
		Z	0.71	63.60	11.09		150.0	

<sup>E</sup> Uncertainty is determined using the max. deviation from linear response applying rectangular distribution and is expressed for the square of the field value.