

## Annex E. Calibration Certificates

---

ID	Device	Type/Model	Serial Number	Manufacturer	Calibration Certificate
003-007	Dosimetric E-field Probe	EX3DV4	7465	SPEAG	See attached files
071-000	750 MHz System Validation Dipole	D750V3	1136	SPEAG	See attached files

### Dipole calibration

According to the KDB 865664 D01, a dipole must be calibrated using a fully validated SAR system according to the tissue dielectric parameters and SAR probe calibration frequency required for device testing. However, instead of the typical annual calibration recommended by measurement standards, longer calibration intervals of up to three years may be considered when it is demonstrated that the SAR target, impedance and return loss of a dipole have remain stable according to the following requirements.

When the most recent return-loss result, measured at least annually, deviates by less than 20% from the previous measurement (i.e. value in dB  $\times$  0.2) or not meeting the required 20 dB minimum return-loss requirement.

When the most recent measurement of the real or imaginary parts of the impedance, measured at least annually, deviates by less than 5  $\Omega$  from the previous measurement

The below results show the latest return loss and impedance measurements for each dipole performed by the lab:

ID #071-000 Dipole 750 MHz Body TSL			
	Return Loss [dB]	Impedance [ $\Omega$ ]	Date
Original Calibration	-27.9	49.4 – 4.0 j	2021-01-21
WRF Lab verification	-29.9	49.1 – 3.1 j	2022-01-17