CERTIFICATE OF CALIBRATION

ISSUED BY UL INTERNATIONAL (UK) LTD

DATE OF ISSUE: 06/May/2022

CERTIFICATE NUMBER: 12345678JD01A



UL INTERNATIONAL (UK) LTD UNIT 1-3 HORIZON KINGSLAND PARK, WADE ROAD BASINGSTOKE, HAMPSHIRE RG24 8AH, UK

TEL: +44 (0) 1256 312100 FAX: +44 (0) 1256 312001

Email: LST.UK.Calibration@ul.com

(UL)

Page 1 of 10

APPROVED SIGNATORY

Harmohan Sahota

Customer:

UL VS Inc 47173 Benicia Street Fremont, CA 94538, USA

Equipment Details:

Description:

Dipole Validation Kit

Date of Receipt:

25/April/2022

Manufacturer:

Speag

Type/Model Number:

D3700V2

Serial Number:

1039

Calibration Date:

06/May/2022

Calibrated By:

Masood Khan

Laboratory Test Engineer

Signature:

Monay

All Calibration have been conducted in the closed laboratory facility: Lab Temperature (22±3) °C and humidity < 70%

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

Use of the UKAS mark demonstrates that compliance with the requirements of BS/EN/ISO/IEC 17025:2017 has been independently assessed. The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

The results relate only to the item calibrated.

CERTIFICATE OF CALIBRATION ISSUED BY UL INTERNATIONAL (UK) LTD

CERTIFICATE NUMBER: 12345678JD01A

UKAS Accredited Calibration Laboratory No. 5772

Page 3 of 6

SAR System Specification

obot System Positioner: Stäubli Unimation Corp. Robot Model: TX60L		
Robot Serial Number:	F17/5ENYG1/A/01	
DASY Version:	cDASY6.14.0.959	
Phantom:	Flat section of SAM Twin Phantom	
Distance Dipole Centre:	10 mm (with spacer)	
Frequency:	3700 MHz	

Dielectric Property Measurements – Head Simulating Liquid (HSL)

Simulant Liquid	Frequency (MHz)	Room Temp		Liquid Temp		Parameters	Target	Measured	Uncertainty
		Start	End	Start	End	1 didifictors	Value	Value	(%)
Head	3700	18.5 °C 18.30 °C	10 20 %	18.8 ℃	18.2 °C	εr	37.7	39.0	± 5%
			10.0 %	10.2 1	σ	3.12	3.05	± 5%	

SAR Results – Head Simulating Liquid (HSL)

Simulant Liquid	SAR Measured	250 mW input Power	Normalised to 1.00 W	Uncertainty (%)
Head	SAR averaged over 1g	17.400 W/Kg	69.271 W/Kg	+20.70 / -20.50
	SAR averaged over 10g	6.450 W/Kg	25.678 W/Kg	+20.62 / -20.45

Antenna Parameters – Head Simulating Liquid (HSL)

Simulant Liquid	Parameter	Measured Level	Uncertainty
Head	Impedance	52.3 + 3.15j Ω	± 10.83 %
	Return Loss	28.39	± 1.37 dB


CERTIFICATE OF CALIBRATION ISSUED BY UL INTERNATIONAL (UK) LTD

CERTIFICATE NUMBER: 12345678JD01A

UKAS Accredited Calibration Laboratory No. 5772

Page 5 of 6

Impedance Measurement Plot for Head Stimulating Liquid (HSL)



Calibration Certificate Label:



UL INTERNATIONAL (UK) LTD Tel: +44 (0) 1256312000

Certificate Number: 12345678JD01A

Instrument ID:

Calibration Date: 06/May/2022

Calibration Due Date:



UL INTERNATIONAL (UK) LTD Tel: +44 (0) 1256312000

Certificate Number: 12345678JD01A

Instrument ID: 1039

Calibration Date: 06/May/2022

Calibration Due Date:



UL INTERNATIONAL (UK) LTD Tel: +44 (0) 1256312000

Certificate Number: 12345678JD01A

Instrument ID: 1039

Calibration Date: 06/May/2022

Calibration Due Date: