

# RF System Check Test Report for CD1880:SN 1122

## Hardware Setup

Probe | Calibration Date EF3DV3 - SN4028 | September 04, 2023  
DAE | Calibration Date DAE4 Sn1546 | March 11, 2024  
Software Version 1.2.0.1713

## Communication Systems

Band Name CD1880  
Channel | Frequency [MHz] 50 | 1880.0  
Communication Systems' Name CW

## Grid Settings

Extent X [mm]	20.0	Step X [mm]	5.0
Extent Y [mm]	90.0	Step Y [mm]	5.0
Distance [mm]	15.0		

## Results

E <sub>max</sub> [V/m]	83.9
Drift [dB]	0.03

