

RF System Check Test Report for CD1880:SN 1122

Hardware Setup

Probe | Calibration Date EF3DV3 - SN4028 | September 04, 2023
DAE | Calibration Date DAE4 Sn1546 | March 11, 2024
Software Version 1.2.0.1713

Communication Systems

Band Name CD1880
Channel | Frequency [MHz] 50 | 1880.0
Communication Systems' Name CW

Grid Settings

| | | | |
|---------------|------|-------------|-----|
| Extent X [mm] | 20.0 | Step X [mm] | 5.0 |
| Extent Y [mm] | 90.0 | Step Y [mm] | 5.0 |
| Distance [mm] | 15.0 | | |

Results

E_{max} [V/m] **82.4**
Drift [dB] **0.01**

