

RF System Check Test Report for CD1880:SN 1159

Hardware Setup

Probe Calibration Date	EF3DV3 - SN4041 March 11, 2024
DAE Calibration Date	DAE4 Sn1798 May 22, 2024
Software Version	1.2.6.2199

Communication Systems

Band Name	CD1880
Channel Frequency [MHz]	50 1880.0
Communication Systems' Name	CW

Grid Settings

Extent X [mm]	20.0	Step X [mm]	5.0
Extent Y [mm]	90.0	Step Y [mm]	5.0
Distance [mm]	15.0		

Results

E _{max} [V/m]	85.0
Drift [dB]	0.01

