

Philips Healthcare

May 29, 2018

Telecommunication Certification Body UL VS Ltd Unit 3, Horizon Wade Road Kingsland Business Park Basingstoke Hampshire RG24 8AH United Kingdom

Subject: Output Power Certification Application FCC ID: BDZ319PHB

To whom it may concern

We state that the maximum conducted output power of this device is +5dBm.

The device is calibrated in the factory during production and the output power cannot be altered once the device has left the factory.

Yours truly,

Clyde Dottin

Clyde Dottin Manager, Configuration and Data Management Philips Healthcare, Home Healthcare Solutions, Home Monitoring 111 Lawrence Street Framingham, MA 01702-8156 Phone: 508.988.1313 Fax: 240.536.3263 clyde.dottin@philips.com

Philips Lifeline 111 Lawrence Street Framingham, MA 01702



www.lifelinesys.com