

July 25, 2018

RF Exposure Analysis

FCC ID: BDZ319PHB

Analysis for FCC portable use

Standalone SAR test exclusion considerations are defined in KDB 447498D01 (v06) Chapter 4.3.1 where the 1-g head or body and 10-g extremity SAR exclusion threshold is defined by the following formula:

$[(\text{max. power of channel, including tune-up tolerance, mW}) / (\text{min. test separation distance, mm})] * [\sqrt{f(\text{GHz})}] \leq 3.0$ for 1-g SAR and ≤ 7.5 for 10-g extremity SAR

- $f(\text{GHz})$ is the RF channel transmit frequency in GHz
- Power and distance are rounded to the nearest mW and mm before calculation
- The result is rounded to one decimal place for comparison

For the 319 Personal Help Button, the maximum conducted output power is +5.0 dBm (3.16mW).

Applying the above data using the given KDB 447498 D01 formula, and minimum separation distance of 5mm, the following results:

$(3.16\text{mW} / 5 \text{ mm}) \times \sqrt{0.319 \text{ GHz}} = 0.36$ (i.e.: ≤ 3.0 for 1-g SAR and ≤ 7.5 for 10-g extremity SAR)

This demonstrates the 319 Personal Help Button meets the criteria for 1-g head/ body and 10-g extremity SAR test exemption.

Conclusion

The 319 Personal Help Button can be used for Portable (or Mobile) usage.

Clyde Dottin

Manager, Configuration and Data Management
Philips Healthcare, Home Healthcare Solutions, Home Monitoring
111 Lawrence Street
Framingham, MA 01702-8156
Phone: 508.988.1313
Fax: 240.536.3263
clyde.dottin@philips.com