



September 16, 2013

TUV SUD BABT
Octagon House, Concord Way
Segensworth Rd N, Fareham
P015 5RL

Attention: Director of Certification

RE: SAR Exclusion for Portable Device

IC ID: 655C-7000AHB
FCC ID: BDZ-7000AHB

Hereby declare that the Lifeline Systems Inc, Troy2 Auto-alert Help Button Device Model: 7000AHB which is defined as a portable device under the Code of Federal Regulations 47 CFR paragraph 2.1093 is categorically excluded from routine evaluation of radiofrequency exposure evaluation for portable device. The device satisfies the SAR Exclusion Threshold as per Clause 4.3.1 of KDB 447498 D01v05 using the formula:

$$3.0 \geq \left(\frac{\text{max power in mW}}{\text{min. test separation distance in mm}} \right) \left(\sqrt{f_{(\text{GHz})}} \right)$$

$$3.0 \geq \left(\frac{0.42 \text{ mW}}{5 \text{ mm}} \right) (\sqrt{0.917})$$

$$3.0 \geq 0.08$$

Sincerely,

A handwritten signature in blue ink, appearing to read 'Juan', written over a horizontal line.

Juan Manuel Gonzalez September 16, 2013

Name

Authorized Signatory

Title: EMC/Wireless Test Engineer

TÜV SÜD America (San Diego)