



September 17, 2013

TUV SUD BABT  
Octagon House, Concord Way  
Segensworth Rd N, Fareham  
P015 5RL

Attention: Director of Certification

RE: SAR Exclusion for Portable Device

IC ID: 655C-7000PHB  
FCC ID: BDZ-7000PHB

Hereby declare that the Lifeline Systems Inc, Halifax Personal Help Button Device Model: 7000PHB which is defined as a portable device under the Code of Federal Regulations 47 CFR paragraph 2.1093 is categorically excluded from routine evaluation of radiofrequency exposure evaluation for portable device. The device satisfies the SAR Exclusion Threshold as per Clause 4.3.1 of KDB 447498 D01v05 using the formula:

$$3.0 \geq \left( \frac{\text{max power in mW}}{\text{min. test separation distance in mm}} \right) \left( \sqrt{f_{(\text{GHz})}} \right)$$

$$3.0 \geq \left( \frac{0.32 \text{ mW}}{5 \text{ mm}} \right) (\sqrt{0.921})$$

$$3.0 \geq 0.061$$

Sincerely,

A handwritten signature in blue ink, appearing to read 'Juan', written over a horizontal line.

Juan Manuel Gonzalez September 17, 2013

Name

Authorized Signatory

Title: EMC/Wireless Test Engineer

TÜV SÜD America (San Diego)