




Appendix A. Integrated Device Variant Declaration

Please complete the declaration for each Parent-Variant pair and upload to the PTCRB input request in the Supporting Documentation area.

Product Details	
Company name	Philips Healthcare
Parent product name/model	7100MHB
PTCRB Parent request #	61574
Parent module manufacturer/model	SARA-U260-03
Parent module PTCRB request #	47664
Integrated Device Variant name/model	7150MHB
PTCRB Integrated Device Variant request #	62411
Integrated Device Variant module manufacturer/model	SARA-U260-03
Integrated Device Variant module PTCRB request #	47664
Description of change from Parent to Integrated Device Variant	
<p>The 7150MHB help button is physically identical to the 7100MHB. They both have the exactly same hardware, including cellular module and IMEI TAC type. The only difference is in the firmware configuration on turning ON/OFF the ISM transceiver.</p> <p>The 7100MHB is configured to use the ISM transceiver to report alarm and device status via 7000C or 7000L communicator when the 7100MHB user is at home. The 7150MHB is configured NOT to use the ISM transceiver, and report alarm and device status ONLY through the cell network. Note that when the 7100MHB is out of the 7000C or 7000L communicators range, it behaves exactly the same as 7150MHB.</p>	
Compliance Declaration	
DECLARATION <i>(signature)</i>	
Date of Declaration	May 31, 2017
Contact Name	Bill Bekdash
Email Address	Bill.bekdash@philips.com