Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland





S Schweizerischer Kalibrierdienst
C Service suisse d'étalonnage
Servizio svizzero di taratura
Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)

The Swiss Accreditation Service is one of the signatories to the EA

Multilateral Agreement for the recognition of calibration certificates

CALIBRATION CERTIFICATE

Accreditation No.: SCS 0108

Client

Audix (Auden)

Certificate No: 5G-Veri10-2014_Dec22

5G Verification Source 10 GHz - SN: 2014 Object **QA CAL-45.v4** Calibration procedure(s) Calibration procedure for sources in air above 6 GHz December 07, 2022 Calibration date: This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate. All calibrations have been conducted in the closed laboratory facility: environment temperature (22 \pm 3)°C and humidity < 70%. Calibration Equipment used (M&TE critical for calibration) **Primary Standards** ID# Cal Date (Certificate No.) Scheduled Calibration Reference Probe EUmmWV3 SN: 9374 2021-12-21(No. EUmmWV3-9374_Dec21) Dec-22 DAE4ip SN: 1602 2022-06-27 (No. DAE4ip-1602_Jun22) Jun-23 Secondary Standards ID# Check Date (in house) Scheduled Check RF generator R&S SMF100A SN: 100184 19-May-22 (in house check Nov-22) In house check: Nov-23 Power sensor R&S NRP18S-10 SN: 101258 31-May-22 (in house check Nov-22) In house check: Nov-23

Calibrated by:

Name Leif Klysner

Function

Signature

Approved by:

Sven Kühn

Technical Manager

Laboratory Technician

Issued: December 13, 2022

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland





S Schweizerischer Kalibrierdienst
C Service suisse d'étalonnage
Servizio svizzero di taratura
Swiss Calibration Service

Accreditation No.: SCS 0108

Accredited by the Swiss Accreditation Service (SAS)

The Swiss Accreditation Service is one of the signatories to the EA

Multilateral Agreement for the recognition of calibration certificates

Glossary

CW

Continuous wave

Calibration is Performed According to the Following Standards

- Internal procedure QA CAL-45, Calibration procedure for sources in air above 6 GHz.
- IEC/IEEE 63195-1, "Assessment of power density of human exposure to radio frequency fields from wireless devices in close proximity to the head and body (frequency range of 6 GHz to 300 GHz)", May 2022

Methods Applied and Interpretation of Parameters

- Coordinate System: z-axis in the waveguide horn boresight, x-axis is in the direction of the E-field, y-axis normal to the others in the field scanning plane parallel to the horn flare and horn flange.
- Measurement Conditions: (1) 10 GHz: The radiated power is the forward power to the horn antenna minus ohmic and mismatch loss. The forward power is measured prior and after the measurement with a power sensor. During the measurements, the horn is directly connected to the cable and the antenna ohmic and mismatch losses are determined by farfield measurements. (2) 30, 45, 60 and 90 GHz: The verification sources are switched on for at least 30 minutes. Absorbers are used around the probe cub and at the ceiling to minimize reflections.
- Horn Positioning: The waveguide horn is mounted vertically on the flange of the waveguide source to allow vertical positioning of the EUmmW probe during the scan. The plane is parallel to the phantom surface. Probe distance is verified using mechanical gauges positioned on the flare of the horn.
- E- field distribution: E field is measured in two x-y-plane (10mm, 10mm + λ/4) with a vectorial E-field probe. The E-field value stated as calibration value represents the E-field-maxima and the averaged (1cm² and 4cm²) power density values at 10mm in front of the horn.
- Field polarization: Above the open horn, linear polarization of the field is expected. This is verified graphically in the field representation.

Calibrated Quantity

 Local peak E-field (V/m) and average of peak spatial components of the poynting vector (W/m²) averaged over the surface area of 1 cm² and 4cm² at the nominal operational frequency of the verification source. Both square and circular averaging results are listed.

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

Certificate No: 5G-Veri10-2014_Dec22

Page 2 of 7

Measurement Conditions

DASY system configuration, as far as not given on page 1.

DASY Version	DASY8 Module mmWave	V3.0
Phantom	5G Phantom	
Distance Horn Aperture - plane	10 mm	
XY Scan Resolution	dx, dy = 7.5 mm	
Number of measured planes	2 (10mm, 10mm + λ/4)	
Frequency	10 GHz ± 10 MHz	

Calibration Parameters, 10 GHz

Circular Averaging

Distance Horn Aperture to Measured Plane	Prad¹ (mW)	Max E-field (V/m)	Uncertainty (k = 2)	Avg (psPDi	er Density n+, psPDtot+, mod+) /m²)	Uncertainty (k = 2)
				1 cm ²	4 cm ²	
10 mm	124	270	1.27 dB	188	152	1.28 dB

Square Averaging

Distance Horn Aperture to Measured Plane	Prad¹ (mW)	Max E-field (V/m)	Uncertainty (k = 2)	Avg (psPD psPD	er Density n+, psPDtot+, mod+) /m ²)	Uncertainty (k = 2)
				1 cm ²	4 cm ²	
10 mm	124	270	1.27 dB	190	152	1.28 dB

 $^{^{\}rm 1}$ Assessed ohmic and mismatch loss plus numerical offset: 0.95 dB

Measurement Report for 5G Verification Source 10 GHz, UID 0 -, Channel 10000 (10000.0MHz)

Device under Test Properties

Name, Manufacturer	Dimensions [mm]	IMEI	DUT Type
5G Verification Source 10 GHz	100.0 x 100.0 x 100.0	SN: 2014	Воттуре

Exposure Conditions

Phantom Section	Position, Test Distance [mm]	Band	Group,	Frequency [MHz], Channel Number	Conversion Factor
5G -	10.0 mm	Validation band	CW	10000.0, 10000	1.0


Hardware Setup

Phantom	Medium	Probe, Calibration Date	DAE, Calibration Date
mmWave Phantom - 1002	Air	EUmmWV3 - SN9374_F1-55GHz, 2021-12-21	DAE4ip Sn1602, 2022-06-27

Scan Setup

EG Coon
5G Scan 2022-12-07, 15:57 1.00 188 188 189 270 -0.02

Measurement Results



Measurement Report for 5G Verification Source 10 GHz, UID 0 -, Channel 10000 (10000.0MHz)

Device under Test Properties

Name, Manufacturer	Dimensions [mm]	IMEI	DUT Type	
5G Verification Source 10 GHz	100.0 x 100.0 x 100.0	SN: 2014	-	

Exposure Conditions


Phantom Section	Position, Test Distance [mm]	Band	Group,	Frequency [MHz], Channel Number	Conversion Factor
5G -	10.0 mm	Validation band	CW	10000.0, 10000	1.0

Hardware Setup

Phantom	Medium	Probe, Calibration Date	DAE, Calibration Date
mmWave Phantom - 1002	Air	EUmmWV3 - SN9374_F1-55GHz, 2021-12-21	DAE4ip Sn1602, 2022-06-27

Scan Setup

	Measurement Results	
5G Scan		5G Scan
120.0 x 120.0	Date	2022-12-07, 15:57
0.25 x 0.25	Avg. Area [cm ²]	4.00
10.0	psPDn+ [W/m ²]	150
MAIA not used	psPDtot+ [W/m ²]	151
	psPDmod+ [W/m ²]	154
	E _{max} [V/m]	270
	Power Drift [dB]	-0.02
	120.0 x 120.0 0.25 x 0.25 10.0	5G Scan 120.0 x 120.0 Date 0.25 x 0.25 Avg. Area [cm²] 10.0 psPDn+ [W/m²] MAIA not used psPDtot+ [W/m²] psPDmod+ [W/m²] Emax [V/m]



Measurement Report for 5G Verification Source 10 GHz, UID 0 -, Channel 10000 (10000.0MHz)

Device under Test Properties

Name, Manufacturer	Dimensions [mm]	IMEI	DUT Type	
5G Verification Source 10 GHz	100.0 x 100.0 x 100.0	SN: 2014	-	

Exposure Conditions

Phantom Section	Position, Test Distance [mm]	Band	Group,	Frequency [MHz], Channel Number	Conversion Factor	
5G -	10.0 mm	Validation band	CW	10000.0, 10000	1.0	


Hardware Setup

Phantom	Medium	Probe, Calibration Date	DAE, Calibration Date
mmWave Phantom - 1002	Air	EUmmWV3 - SN9374_F1-55GHz, 2021-12-21	DAE4ip Sn1602, 2022-06-27

Scan Setup

Scan Setup		Measurement Results	
	5G Scan		5G Scan
Grid Extents [mm]	120.0 x 120.0	Date	2022-12-07, 15:57
Grid Steps [lambda]	0.25 x 0.25	Avg. Area [cm ²]	1.00
Sensor Surface [mm]	10.0	psPDn+ [W/m ²]	189
MAIA	MAIA not used	psPDtot+ [W/m ²]	190
		psPDmod+ [W/m ²]	191
		E _{max} [V/m]	270
		Power Drift [dB]	-0.02

-0.02



Measurement Report for 5G Verification Source 10 GHz, UID 0 -, Channel 10000 (10000.0MHz)

Device under Test Properties

Name, Manufacturer	Dimensions [mm]	IMEI	DUT Type	
5G Verification Source 10 GHz	100.0 x 100.0 x 100.0	SN: 2014	-	

Exposure Conditions


Phantom Section	Position, Test Distance [mm]	Band	Group,	Frequency [MHz], Channel Number	Conversion Factor
5G -	10.0 mm	Validation band	CW	10000.0, 10000	1.0

Hardware Setup

Phantom	Medium	Probe, Calibration Date	DAE, Calibration Date
mmWave Phantom - 1002	Air	EUmmWV3 - SN9374_F1-55GHz, 2021-12-21	DAE4ip Sn1602, 2022-06-27

Scan Setup

Scan Setup		Measurement Results	
	5G Scan		5G Scan
Grid Extents [mm] Grid Steps [lambda] Sensor Surface [mm] MAIA	120.0 x 120.0 0.25 x 0.25 10.0 MAIA not used	Date Avg. Area [cm²] psPDn+ [W/m²] psPDtot+ [W/m²] psPDmod+ [W/m²] E _{max} [V/m] Power Drift [dB]	2022-12-07, 15:57 4.00 151 151 154 270 -0.02



Calibration Laboratory of

Schmid & Partner Engineering AG

Zeughausstrasse 43, 8004 Zurich, Switzerland





S Schweizerischer Kalibrierdienst

C Service suisse d'étalonnage Servizio svizzero di taratura

S Swiss Calibration Service

Accreditation No.: SCS 0108

Accredited by the Swiss Accreditation Service (SAS)

The Swiss Accreditation Service is one of the signatories to the EA

Multilateral Agreement for the recognition of calibration certificates

Client

Audix

New Taipei City, Taiwan

Certificate No.

EUmm-9544_Apr23

CALIBRATION CERTIFICATE

Object

EUmmWV4 - SN:9544

Calibration procedure(s)

QA CAL-02.v9, QA CAL-25.v8, QA CAL-42.v3

Calibration procedure for E-field probes optimized for close near field

evaluations in air

Calibration date

April 19, 2023

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature (22±3) ℃ and humidity < 70%.

Calibration Equipment used (M&TE critical for calibration)

Primary Standards	ID	Cal Date (Certificate No.)	Scheduled Calibration
Power sensor NRP110T	SN: 101244	12-Apr-23 (No. 0001A300692178)	Apr-24
Spectrum analyzer FSV40	SN: 101832	23-Jan-23 (No. 4030-315005314)	Jan-24
Ref. Probe EUmmWV3	SN: 9374	03-Jan-23 (No. EUmmWV3-9374_Jan23)	Jan-24
DAE4	SN: 789	03-Jan-23 (No. DAE4-789_Jan23)	Jan-24

0	Lie		
Secondary Standards	ID	Check Date (in house)	Scheduled Check
Generator APSIN26G	SN: 669	28-Mar-17 (in house check May-22)	
Congretor Agilent FOOTAA			In house check: May-23
Generator Agilent E8251A	SN: US41140111	28-Mar-17 (in house check May-22)	In house check: May-23


	Name	Function	Signature
Calibrated by	Leif Klysner	Laboratory Technician	Seif Tellen
Approved by	Sven Kühn	Technical Manager , a.	1. kelled

Issued: April 25, 2023

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Calibration Laboratory of

Schmid & Partner **Engineering AG**







S

Schweizerischer Kalibrierdienst

Service suisse d'étalonnage Servizio svizzero di taratura **Swiss Calibration Service**

Accreditation No.: SCS 0108

Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Glossary

NORMx,y sensitivity in free space DCP diode compression point

CF crest factor (1/duty_cycle) of the RF signal A, B, C, D modulation dependent linearization parameters

Polarization φ φ rotation around probe axis

 θ rotation around an axis that is in the plane normal to probe axis (at measurement center), i.e., $\theta = 0$ is Polarization ∂

normal to probe axis

Connector Angle information used in DASY system to align probe sensor X to the robot coordinate system Sensor Angles sensor deviation from the probe axis, used to calculate the field orientation and polarization

is the wave propagation direction

Calibration is Performed According to the Following Standards:

a) IEEE Std 1309-2005, "IEEE Standard for calibration of electromagnetic field sensors and probes, excluding antennas, from 9 kHz to 40 GHz", December 2005

Methods Applied and Interpretation of Parameters:

- NORMx,y: Assessed for E-field polarization $\theta = 0$ ($f \le 900 \,\text{MHz}$ in TEM-cell; $f > 1800 \,\text{MHz}$: R22 waveguide). For frequencies > 6 GHz, the far field in front of waveguide horn antennas is measured for a set of frequencies in various waveguide bands up to 110 GHz.
- DCPx,y: DCP are numerical linearization parameters assessed based on the data of power sweep with CW signal. DCP does not depend on frequency nor media.
 - Note: As the field is measured with a diode detector sensor, it is warrantied that the probe response is linear (E2) below the documented lowest calibrated value.
- PAR: PAR is the Peak to Average Ratio that is not calibrated but determined based on the signal characteristics
- The frequency sensor model parameters are determined prior to calibration based on a frequency sweep (sensor model involving resistors R, Rp, inductance L and capacitors C, Co).
- Ax,y; Bx,y; Cx,y; Dx,y; VRx,y: A, B, C, D are numerical linearization parameters assessed based on the data of power sweep for specific modulation signal. The parameters do not depend on frequency nor media. VR is the maximum calibration range expressed in RMS voltage across the diode.
- · Sensor Offset: The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No tolerance required.
- · Connector Angle: The angle is assessed using the information gained by determining the NORMx (no uncertainty required).
- Equivalent Sensor Angle: The two probe sensors are mounted in the same plane at different angles. The angles are assessed using the information gained by determining the NORMx (no uncertainty required).
- Spherical isotropy (3D deviation from isotropy): in a locally homogeneous field realized using an open waveguide / horn setup.

EUmmWV4 - SN:9544 April 19, 2023

Parameters of Probe: EUmmWV4 - SN:9544

Basic Calibration Parameters

	Sensor X	Sensor Y	Unc (k = 2)
Norm $(\mu V/(V/m)^2)$	0.01849	0.02029	±10.1%
DCP (mV) B	105.0	105.0	±4.7%
Equivalent Sensor Angle	-61.3	35.3	

Calibration Results for Frequency Response (750 MHz - 110 GHz)

Frequency GHz	Target E-Field V/m	Deviation Sensor X dB	Deviation Sensor Y dB	Unc (k = 2) dB
0.75	77.2	-0.03	-0.35	±0.43
1.8	140.4	-0.04	-0.06	±0.43
2.0	133.0	0.12	0.14	±0.43
2.2	124.8	-0.07	-0.06	±0.43
2.5	123.0	0.09	0.14	±0.43
3.5	256.2	-0.18	-0.05	±0.43
3.7	249.8	0.03	0.13	±0.43
6.6	76.1	0.29	0.22	±0.98
8.0	68.3	0.14	0.05	±0.98
10.0	67.5	0.06	0.06	±0.98
15.0	55.3	0.37	0.23	±0.98
26.6	114.9	-0.04	-0.03	±0.98
30.0	121.2	-0.00	0.02	±0.98
35.0	119.8	0.14	0.11	±0.98
40.0	105.8	0.22	0.13	±0.98
50.0	60.5	0.54	0.57	±0.98
55.0	75.8	-0.06	-0.06	±0.98
60.0	80.0	0.09	0.08	±0.98
65.0	77.7	0.31	0.32	±0.98
70.0	73.8	0.29	0.25	±0.98
75.0	73.2	-0.20	-0.22	±0.98
75.0	80.8	0.10	0.09	±0.98
80.0	79.9	-0.41	-0.37	±0.98
85.0	47.6	-0.53	-0.50	±0.98
90.0	72.3	-0.30	-0.30	±0.98
92.0	72.0	-0.18	-0.17	±0.98
95.0	66.6	-0.01	-0.03	±0.98
97.0	57.0	0.04	0.03	±0.98
100.0	55.0	0.06	0.04	±0.98
105.0	53.0	-0.19	-0.18	±0.98
110.0	61.1	-0.03	-0.02	±0.98

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

Certificate No: EUmm-9544_Apr23 Page 3 of 18

^B Linearization parameter uncertainty for maximum specified field strength.

EUmmWV4 - SN:9544 April 19, 2023

Parameters of Probe: EUmmWV4 - SN:9544

Calibration Results for Modulation Response

UID	Communication System Name		Α	В	С	D	VR	Max	Max
			dB	$dB\sqrt{\mu V}$		dB	mV	dev.	UncE
				V P				401.	k=2
0	CW	X	0.00	0.00	1.00	0.00	128.3	±3.5%	±4.7%
		Y	0.00	0.00	1.00		63.4	10.070	1.770
10352	Pulse Waveform (200Hz, 10%)	Х	2.39	60.00	14.56	10.00	6.0	±0.9%	±9.6%
		Y	1.47	60.00	16.32	1	6.0	20.070	±0.070
10353	Pulse Waveform (200Hz, 20%)	X	1.65	60.00	13.37	6.99	12.0	±1.0%	±9.6%
		Y	10.00	86.00	23.00		12.0		20.070
10354	Pulse Waveform (200Hz, 40%)	Х	0.98	60.00	12.07	3.98	23.0	±1.4%	±9.6%
		Y	0.69	60.00	13.88		23.0	,	20.070
10355	Pulse Waveform (200Hz, 60%)	Х	0.57	60.00	11.37	2.22	27.0	±1.1%	±9.6%
	200	Y	0.53	60.00	12.47		27.0		±0.070
10387	QPSK Waveform, 1 MHz	X	1.25	60.00	12.21	1.00	22.0	±1.5%	±9.6%
		Y	1.32	60.00	11.80		22.0	211070	20.070
10388	QPSK Waveform, 10 MHz	Х	1.28	60.00	11.88	0.00	22.0	±0.9%	±9.6%
		Y	1.53	60.00	11.62		22.0		201070
10396	64-QAM Waveform, 100 kHz	Х	3.42	65.98	16.24	3.01	17.0	±0.6%	±9.6%
		Υ	2.16	60.00	13.90		17.0		±0.070
10399	64-QAM Waveform, 40 MHz	Х	2.09	60.00	12.39	0.00	19.0	±1.1%	±9.6%
		Υ	2.29	60.00	12.25	2.00	19.0	270	±0.070
10414	WLAN CCDF, 64-QAM, 40 MHz	X	3.27	60.00	12.82	0.00	12.0	±0.8%	±9.6%
		Υ	3.43	60.00	12.71	0.00	12.0	_0.070	±0.076

Note: For details on UID parameters see Appendix

E Uncertainty is determined using the max. deviation from linear response applying rectangular distribution and is expressed for the square of the field value.

EUmmWV4 - SN:9544 April 19, 2023

Parameters of Probe: EUmmWV4 - SN:9544

Calibration Results for Linearity Response

Frequency GHz	Target E-Field V/m	Deviation Sensor X dB	Deviation Sensor Y dB	Unc (k = 2) dB
0.9	50.0	0.08	0.14	±0.2
0.9	100.0	-0.02	0.07	±0.2
0.9	500.0	-0.00	-0.03	±0.2
0.9	1000.0	0.03	-0.01	±0.2
0.9	1500.0	0.03	-0.01	±0.2
0.9	2100.0	-0.02	-0.01	±0.2

Sensor Frequency Model Parameters (750 MHz – 55 GHz)

	Sensor X	Sensor Y
R (Ω)	114.18	87.21
$R_p(\Omega)$	124.58	97.49
L (nH)	0.15031	0.10470
C (pF)	0.1747	0.2822
Cp (pF)	0.0595	0.0784

Sensor Frequency Model Parameters (55 GHz - 110 GHz)

	Sensor X	Sensor Y
R (Ω)	25.25	47.36
$R_p(\Omega)$	135.16	244.33
L (nH)	0.07195	0.13584
C (pF)	0.0584	0.0320
C _p (pF)	0.0673	0.0341

Sensor Model Parameters

	C1 fF	C2 fF	α V ⁻¹	T1 ms V ⁻²	T2 ms V ⁻¹	T3 ms	T4 V ⁻²	T5 V ⁻¹	Т6
Х	61.2	445.10	33.78	0.92	5.59	5.00	0.00	1.72	1.01
У	45.2	323.56	32.87	0.92	3.49	5.05	0.00	1.98	1.01

Other Probe Parameters

Sensor Arrangement	Rectangular
Connector Angle	-175.0°
Mechanical Surface Detection Mode	enabled
Optical Surface Detection Mode	disabled
Probe Overall Length	320 mm
Probe Body Diameter	8 mm
Tip Length	23 mm
Tip Diameter	8.0 mm
Probe Tip to Sensor X Calibration Point	1.5 mm
Probe Tip to Sensor Y Calibration Point	1.5 mm

Certificate No: EUmm-9544_Apr23