

<b>Date / Time:</b>	6/05/2003 5:12:20 PM	<b>Position:</b>	verification
<b>Filename:</b>	*.txt	<b>Phantom:</b>	HeadBox_spout.csv
<b>Device Tested:</b>	system	<b>Head Rotation:</b>	0
<b>Antenna:</b>	2450 dipole	<b>Test Frequency:</b>	2450
<b>Shape File:</b>	Dipole1800.csv	<b>Power Level:</b>	maximum

<b>Probe:</b>	0123																
<b>Cal File:</b>	123_2450_HEAD																
<b>Cal Factors:</b>	<table border="1"> <thead> <tr> <th></th> <th>X</th> <th>Y</th> <th>Z</th> </tr> </thead> <tbody> <tr> <td><b>Air</b></td> <td>346</td> <td>318</td> <td>386</td> </tr> <tr> <td><b>DCP</b></td> <td>20</td> <td>20</td> <td>20</td> </tr> <tr> <td><b>Lin</b></td> <td>.768</td> <td>.768</td> <td>.768</td> </tr> </tbody> </table>		X	Y	Z	<b>Air</b>	346	318	386	<b>DCP</b>	20	20	20	<b>Lin</b>	.768	.768	.768
		X	Y	Z													
	<b>Air</b>	346	318	386													
	<b>DCP</b>	20	20	20													
<b>Lin</b>	.768	.768	.768														
<b>Amp Gain:</b>	2																
<b>Averaging:</b>	1																
<b>Batteries Replaced:</b>	05/13/03																

<b>Liquid:</b>	2450
<b>Type:</b>	head
<b>Conductivity:</b>	1.89
<b>Relative Permittivity:</b>	39.40
<b>Liquid Temp (deg C):</b>	22.0
<b>Ambient Temp (deg C):</b>	22.0
<b>Ambient RH (%):</b>	50
<b>Density (kg/m3):</b>	1000
<b>Software Version:</b>	0.420

**ZOOM SCAN RESULTS:**

<b>Spot SAR (W/kg):</b>	<b>Start Scan</b>	<b>End Scan</b>
	5.686	5.558
<b>Change during Scan (%):</b>	-2.26	
<b>Max E-field (V/m):</b>	135.70	
<b>Max SAR (W/kg)</b>	<b>1g</b>	<b>10g</b>
	26.365	12.301

  

<b>Location of Max (mm):</b>	<b>X</b>	<b>Y</b>	<b>Z</b>
	77.1	-17.6	-136.6

**AREA SCAN:**

<b>Scan Extent:</b>		<b>Min</b>	<b>Max</b>	<b>Steps</b>
	<b>Y</b>	-40.0	40.0	10.0
	<b>Z</b>	-180.0	-110.0	10.0