

Section 15.231 and ANSI C63.4

This is a list of all test equipment used.

Test Equipment list for Honeywell OATS:

Equipment	Mfg	Model	Cal Date	Cal Due
Spectrum Analyzer	Rohde & Schwarz	FSEA20	10/12/09	10/12/10
Antenna ('Biconilog')	ETS (EMCO)Lindgren	3149	07/10/09	07/10/10
Transient Limiter	HP	11947A	5/12/10	5/12/11
LISN	EMCO	3825/2	10/16/09	10/16/10

PLEASE SEE PAGE 2-6 FOR TEST EQUIPMENT TRACEABILITY

If you need any additional information from Honeywell please contact:

Greg Barbato RF Engineer
(Acting for Ken Eskildsen)
Phone (Direct): (516) 577-5863
Email: greg.barbato@honeywell.com

Certificate of Calibration

Issue Date: 10/12/2009



General Calibration, Inc.

2 Mars Court, Boonton, New Jersey 07005
Phone (973) 299-2950 Fax (973) 299-0595

Certificate #: 15395MR
Purchase Order: 5105648
Work Order #: MR309
Customer #: 001464

Performed By:
GENERAL CALIBRATION, INC.
2 MARS COURT

MONTVILLE, NJ 07045

Equipment Information

Job No.: 018675

Manufacturer: R&S

Description: SPECTRUM ANALYZER

Department: ALARMNET

Temp./RH: 22 C / 45 %

Cal. Interval: 12 MONTHS

Cal Date: 10/12/2009

Location of Calibration:
HONEYWELL SECURITY (001464)
2 CORPORATE CENTER DRIVE

MELVILLE, NY 11747

Asset Tag No.: 10506
Model Number: FSEA20
Serial Number: DE23427
Inspected By: MR1
Job Title: METROLOGIST
Calibration Result: PASS
Cal. Due Date: 10/12/2010

Calibration Notes

Condition: Found In Tolerance and Left In Tolerance

Procedure #GCP: R&S FSEA20

Standards Used To Calibrate Equipment

Company	I.D.	Description	Cal. Due Date
GENERAL CALIBRATION	1031	SYNTHESIZED SWEEPER	02/02/2010
GENERAL CALIBRATION	434	POWER SPLITTER	09/18/2010
GENERAL CALIBRATION	650	POWER SPLITTER	12/08/2009
GENERAL CALIBRATION	774	POWER SENSOR	07/13/2010
GENERAL CALIBRATION	967	FUNCTION GENERATOR	09/15/2010

The above instrument has been checked and calibrated against the above working standard(s) which are traceable to the NIST. The test limits stated in the report correspond to the published specifications of the equipment, at the points tested. Also, the collective uncertainties of measurement standards do not exceed 25% of the tolerance of the characteristics being calibrated, where possible. The metrology procedures utilized conform to and satisfy the requirements set forth in ANSI/NCSL Z540-1-1994, 10 CFR part 21, ISO 9001-2000, ISO 10012-2003, and MIL-STD 45662A.

Approved By

General Calibration, Inc. - Q. A. Manager



Lab ID:1207.01-Calibration



ETS-LINDGREN
An ESCO Technologies Company
Track# S000016846 Ltd Cal
By SL Date 10-Jul-09
Next Cal Due _____
www.ets-lindgren.com

Cert I.D.: 74136

1301 Arrow Point Drive
Cedar Park, Texas 78613
(512) 531-6498

Certificate of Calibration Conformance

Page 1 of 5

The instrument identified below has been individually calibrated in compliance with the following standard(s):

SAE, ARP-958 - 2003, Electromagnetic Interference Measurement Antennas; Standard Calibration Method, Society of Automotive Engineers, Aerospace Recommended Practice. Fixed height, three antenna rotation, 1 meter separation. 3 meter separation performed per Annex C. Vertical calibration performed per above listed methodology.

Environment: Laboratory MTE is maintained in a temperature controlled environment with ambient conditions from 18 to 28 C, relative humidity less than 90%. The instrument under test has been calibrated on an open air test site (OATS) with environment temperature conditions ranging from 0 to 40 C which has no known influences on measurement quality.

Manufacturer:	EMCO	Operating Range:	80 MHz - 6 GHz
Model Number:	3149	Instrument Type:	Biconilog (Type 5)
Serial Number/ ID:	00045682	Date Code:	
Tracking Number:	S000016846	Alternate ID:	11242
Date Completed:	10-Jul-09	Customer:	HONEYWELL (NY)
Test Type:	3 meter, Horizontal and Vertical		
Calibration Uncertainty:	01m k=2, (95% Confidence Level)	80 - 1000 MHz, +/-0.9 dB; 1000 - 2000 MHz, +/-0.8 dB; 2000 - 6000 MHz, +/-1.2 dB	
	03m	80 - 1000 MHz, +/-0.9 dB; 1000 - 2000 MHz, +/-0.8 dB; 2000 - 6000 MHz, +/-1.3 dB	
	10m	80 - 1000 MHz, +/-1.0 dB; 1000 - 2000 MHz, +/-1.4 dB; 2000 - 6000 MHz, +/-2.3 dB	

Test Remarks: Unit was calibrated down to 26 MHz.

Calibration Traceability: All Measuring and Test Equipment (M/TE) identified below are traceable to the National Institute for Standards and Technology (NIST). Calibration Laboratory and Quality System controls are compliant with ISO/IEC 17025-2005.

Standards and Equipment Used:

Make / Model / Name / S/N / Recall Date

Hewlett Packard	8753C	Network Analyzer	3029A01587	11-Feb-10	Condition of Instrument Upon Receipt:
Hewlett Packard	85047A	S-parameter Test Set	3033A02186	11-Feb-10	In Tolerance to Internal Quality Standards
Rohde & Schwarz	ZVK	Vector Network Analyzer	1127.8651.60	26-Feb-10	On Release:
					In Tolerance to Internal Quality Standards

Calibration Completed By
Slav Ligai, Calibration Technician

Attested and Issued on 10-Jul-09
Justin Tarr, Calibration Supervisor

Certificate of Calibration

Issue Date: 05/12/2009

GEN-CAL

General Calibration, Inc.

2 Mars Court, Boonton, New Jersey 07005
Phone (973) 299-2950 Fax (973) 299-0595

Certificate #: 14309MR
Purchase Order: 5105648
Work Order #: MR267
Customer #: 001464

Performed By:
GENERAL CALIBRATION, INC.
2 MARS COURT

MONTVILLE, NJ 07045

Equipment Information

BarCode: 096188

Manufacturer: HP

Description: TRANSIENT LIMITER

Department:

Temp./RH: 22 C / 45 %

Cal. Interval: 12 MONTHS

Cal Date: 05/12/2009

Location of Calibration:
HONEYWELL SECURITY (1464)
2 CORPORATE CENTER DRIVE

MELVILLE, NY 11747

Instrument I.D.: 10131

Model Number: 11947A

Serial Number: 3107A02782

Inspected By: MR1

Job Title: METROLOGIST

Calibration Result: PASS

Cal. Due Date: 5/12/2010

Calibration Notes

Condition: Found In Tolerance and Left In Tolerance

Standards Used To Calibrate Equipment

Company	I.D.	Description	Cal. Due Date
GENERAL CALIBRATION	503	SPECTRUM ANALYZER	03/11/2010
GENERAL CALIBRATION	838	SIGNAL GENERATOR	08/25/2009

The above instrument has been checked and calibrated against the above working standard(s) which are traceable to the NIST. The test limits stated in the report correspond to the published specifications of the equipment, at the points tested. Also, the collective uncertainties of measurement standards do not exceed 25% of the tolerance of the characteristics being calibrated, where possible. The metrology procedures utilized conform to and satisfy the requirements set forth in ANSI/NCSL Z540-1-1994, 10 CFR part 21, ISO 9001-2000, ISO 10012-2003, and MIL-STD 45662A.

Approved By

Robert A. Ziffner

General Calibration, Inc. - Q. A. Manager

Certificate of Calibration

Issue Date: 5/12/2010

GEN-CAL

General Calibration, Inc.

2 Mars Court, Boonton, New Jersey 07005
Phone (973) 299-2950 Fax (973) 299-0595

Certificate #: 16533MR

Purchase Order: 5172133

Work Order #: MR361

Customer #: 001464

Performed By:

GENERAL CALIBRATION, INC.
2 MARS COURT

BOONTON, NJ 07005

Equipment Information

Job No.: 096188

Manufacturer: HP

Description: TRANSIENT LIMITER

Department:

Temp./RH: 22 C / 45 %

Cal. Interval: 12 MONTHS

Cal Date: 05/12/2010

Location of Calibration:

HONEYWELL SECURITY (001464)
2 CORPORATE CENTER DRIVE

MELVILLE, NY 11747

Asset Tag No.: 10131

Model Number: 11947A

Serial Number: 3107A02782

Inspected By: MR1

Job Title: METROLOGIST

Calibration Result: PASSED

Cal. Due Date: 05/12/2011

Calibration Notes

Condition: Found In Tolerance and Left In Tolerance

Standards Used To Calibrate Equipment

Company	I.D.	Description	Cal. Due Date
GENERAL CALIBRATION	1039	FUNCTION GENERATOR	03/11/2011
GENERAL CALIBRATION	778	POWER SENSOR	08/11/2010
GENERAL CALIBRATION	790	POWER METER	10/19/2010
GENERAL CALIBRATION	906	SYNTHESIZED SWEEPER	09/09/2010

The above instrument has been checked and calibrated against the above working standard(s) which are traceable to the NIST. The test limits stated in the report correspond to the published specifications of the equipment, at the points tested. Also, the collective uncertainties of measurement standards do not exceed 25% of the tolerance of the characteristics being calibrated, where possible. The metrology procedures utilized conform to and satisfy the requirements set forth in ANSI/NCSL Z540-1-1994, 10 CFR part 21, ISO 9001-2008, ISO 10012-2003, and MIL-STD 45662A.

Approved By

Peter D. Hoff

General Calibration, Inc. - Q. A. Manager

Certificate of Calibration

Issue Date: 10/16/2009

GEN-CAL

General Calibration, Inc.

2 Mars Court, Boonton, New Jersey 07005
Phone (973) 299-2950 Fax (973) 299-0595

Certificate #: 15434MR

Purchase Order: 5105648

Work Order #: MR309

Customer #: 001464

Performed By:

GENERAL CALIBRATION, INC.
2 MARS COURT

MONTVILLE, NJ 07045

Equipment Information

Job No.: 004345
Manufacturer: EMCO
Description: LISN
Department: BLDG. 163
Temp./RH: 22 C / 45 %
Cal. Interval: 12 MONTHS
Cal Date: 10/16/2009

Location of Calibration:

HONEYWELL SECURITY (001464)
2 CORPORATE CENTER DRIVE

MELVILLE, NY 11747

Asset Tag No.: 04832
Model Number: 3825/2
Serial Number: 1076
Inspected By: MR1
Job Title: METROLOGIST
Calibration Result: PASS
Cal. Due Date: 10/16/2010

Calibration Notes

Condition: Found In Tolerance and Left In Tolerance

Procedure #GCP: EMCO 3825/2

Standards Used To Calibrate Equipment

Company	I.D.	Description	Cal. Due Date
GENERAL CALIBRATION	1004	DIGITAL MULTIMETER	05/28/2010
GENERAL CALIBRATION	700	DIGITAL MULTIMETER	01/16/2010

The above instrument has been checked and calibrated against the above working standard(s) which are traceable to the NIST. The test limits stated in the report correspond to the published specifications of the equipment, at the points tested. Also, the collective uncertainties of measurement standards do not exceed 25% of the tolerance of the characteristics being calibrated, where possible. The metrology procedures utilized conform to and satisfy the requirements set forth in ANSI/NCSL Z540-1-1994, 10 CFR part 21, ISO 9001-2000, ISO 10012-2003, and MIL-STD 45662A.

Approved By

Robert A. Ziffare

General Calibration, Inc. - Q. A. Manager