

## **Spacelabs**

Telemetry Transmitters
96281-A05N (narrowband),
96281-A05W (wideband)
Report #: SPAC0475



Report Prepared By Northwest EMC Inc.

NORTHWEST EMC - (888) 364-2378 - www.nwemc.com

California – Minnesota – Oregon – New York – Washington



22975 NW Evergreen Parkway Suite 400 Hillsboro, Oregon 97124

## **Certificate of Test**

Last Date of Test: January 25, 2012 Spacelabs

Model: Telemetry Transmitters 96281-A05N (narrowband), 96281-A05W (wideband)

#### **Emissions**

Test Description	Specification	Test Method	Pass/Fail
Frequency Stability	FCC 95H:2012	ANSI/TIA/EIA-603-C-2004	Pass
Occupied Bandwidth	FCC 95H:2012	ANSI/TIA/EIA-603-C-2004	Pass
Field Strength of Spurious Emissions	FCC 95H:2012	ANSI/TIA/EIA-603-C-2004	Pass
Field Strength of Fundamental	FCC 95H:2012	ANSI/TIA/EIA-603-C-2004	Pass

#### **Deviations From Test Standards**

None

Approved By:

Tim O'Shea, Operations Manager

NV(AA)

**NVLAP Lab Code: 200630-0** 

#### Test Facility

The measurement facility used to collect the data is located at:

Northwest EMC, Inc. 22975 NW Evergreen Parkway, Suite 400 Hillsboro, OR 97124

Phone: (503) 844-4066 Fax: 844-3826

This site has been fully described in a report filed with and accepted by the FCC (Federal Communications Commission) and Industry Canada (Site filing #2834D-2).

This report must not be used to claim product certification, approval, or endorsement by NVLAP, NIST, or any agency of the federal government of the United States of America.

Product compliance is the responsibility of the client, therefore the tests and equipment modes of operation represented in this report were agreed upon by the client, prior to testing. This Report may only be duplicated in its entirety. The results of this test pertain only to the sample(s) tested. The specific description is noted in each of the individual sections of the test report supporting this certificate of test.

## **Revision History**

Revision Number	Description	Date	Page Number
00	None		

Revision 09/01/11



# Accreditations and Authorizations

#### **FCC**

Accredited by NVLAP for performance of FCC radio, digital, and ISM device testing. Our Open Area Test Sites, certification chambers, and conducted measurement facilities have been fully described in reports filed with the FCC and accepted by the FCC in letters maintained in our files. Northwest EMC has been accredited by ANSI to ISO / IEC Guide 65 as a product certifier. We have been designated by the FCC as a Telecommunications Certification Body (TCB). This allows Northwest EMC to certify transmitters to FCC specifications in accordance with 47 CFR 2.960 and 2.962.

#### **NVLAP**

Northwest EMC, Inc. is accredited under the National Voluntary Laboratory Accreditation Program (NVLAP) for satisfactory compliance with the requirements of ISO/IEC 17025 for Testing Laboratories. NVLAP is administered by the National Institute of Standards and Technology (NIST), an agency of the U.S. Commerce Department. The NVLAP accreditation encompasses Electromagnetic Compatibility Testing in accordance with the European Union EMC Directive 2004/108/EC, and ANSI C63.4. Additionally, Northwest EMC is accredited by NVLAP to perform radio testing in accordance with the European Union R&TTE Directive 1999/5/EEC, the requirements of FCC, and the RSS radio standards for Industry Canada.

## **Industry Canada**

Accredited by NVLAP for performance of Industry Canada RSS and ICES testing. Our Open Area Test Sites and certification chambers comply with RSS-Gen, Issue 2 and have been filed with Industry Canada and accepted. Northwest EMC has been accredited by ANSI to ISO / IEC Guide 65 as a product certifier. We have been designated by NIST and recognized by Industry Canada as a Certification Body (CB) per the APEC Mutual Recognition Arrangement (MRA). This allows Northwest EMC to certify transmitters to Industry Canada technical requirements. (Site Filing Numbers - Hillsboro: 2834D-1, 2834D-2, Sultan: 2834C-1, Irvine: 2834B-1, 2834B-2, Brooklyn Park: 2834E-1)

### **CAB**

Designated by NIST and validated by the European Commission as a Conformity Assessment Body (CAB) to conduct tests and approve products to the EMC directive and transmitters to the R&TTE directive, as described in the U.S. - EU Mutual Recognition Agreement.

## Australia/New Zealand

The National Association of Testing Authorities (NATA), Australia has been appointed by the ACA as an accreditation body to accredit test laboratories and competent bodies for EMC standards. Accredited test reports or assessments by competent bodies must carry the NATA logo. Test reports made by an overseas laboratory that has been accredited for the relevant standards by an overseas accreditation body that has a Mutual Recognition Agreement (MRA) with NATA are also accepted as technical grounds for product conformity. The report should be endorsed with the respective logo of the accreditation body (NVLAP).

Revision 09/01/11

# Accreditations and Authorizations

#### **VCCI**

Accepted as an Associate Member to the VCCI, Acceptance No. 564. Conducted and radiated measurement facilities have been registered in accordance with Regulations for Voluntary Control Measures, Article 8. (Registration Numbers. - Hillsboro: C-1071, R-1025, G-84, C-2687, T-1658, and R-2318, Irvine: R-1943, G-85, C-2766, and T-1659, Sultan: R-871, G-83, C-3265, and T-1511, Brooklyn Park: R-3125, G-86, G-141, C-3464, and T-1634).

#### **BSMI**

Northwest EMC has been designated by NIST and validated by C-Taipei (BSMI) as a CAB to conduct tests as described in the APEC Mutual Recognition Agreement (US0017).

#### **GOST**

Northwest EMC, Inc. has been assessed and accredited by the Russian Certification bodies Certinform VNIINMASH, CERTINFO, SAMTES, and Federal CHEC, to perform EMC and Hygienic testing for Information Technology Products. As a result of their laboratory assessment, they will accept test results from Northwest EMC, Inc. for product certification

### **KCC**

Northwest EMC, Inc is a CAB designated by MRA partners and recognized by Korea. (Assigned Lab Numbers: Hillsboro: US0017, Irvine: US0158, Sultan: US0157, Brooklyn Park: US0175)

## **VIETNAM**

Vietnam MIC has approved Northwest EMC as an accredited test lab. Per Decision No. 194/QD-QLCL (dated December 15, 2009), Northwest EMC test reports can be used for Vietnam approval submissions.

#### SCOPE

For details on the Scopes of our Accreditations, please visit: http://www.nwemc.com/accreditations/



## Locations






Oregon Labs EV01-EV12 22975 NW Evergreen Pkwy Suite 400 Hillsboro, OR 97124 (503) 844-4066

California Labs OC01-OC13 41 Tesla Irvine, CA 92618 (949) 861-8918 Minnesota Labs MN01-MN08 9349 W Broadway Ave. Brooklyn Park, MN 55445 (763) 425-2281 Washington Labs SU01-SU07 14128 339<sup>th</sup> Ave. SE Sultan, WA 98294 (360) 793-8675

New York Labs WA01-WA04 4939 Jordan Rd. Elbridge, NY 13060 (315) 685-0796









## **Product Description**

#### **Client and Equipment Under Test (EUT) Information**

Company Name:	Spacelabs Medical Inc.
Address:	5150 220th Ave SE
City, State, Zip:	Issaquah, WA 98027-7018
Test Requested By:	Steve Cantwell
Model:	Telemetry Transmitters 96281-A05N (narrowband), 96281-A05W (wideband)
First Date of Test:	January 19, 2012
Last Date of Test:	January 25, 2012
Receipt Date of Samples:	January 19, 2012
Equipment Design Stage:	Preproduction
Equipment Condition:	No Damage

## Information Provided by the Party Requesting the Test

#### **Functional Description of the EUT (Equipment Under Test):**

The 96281-A05 is a body-worn telemetry transmitter operating at 600 MHz and also contains a Bluetooth radio.

#### **Testing Objective:**

To demonstrate compliance with FCC 95H requirements for a medical telemetry device operating in the 600 MHz band.



## Configurations

## **Configuration 1 SPAC0475**

Software/Firmware Running during test				
Description	Version			
STT Driver Interface	1.1.4392.26483			

EUT			
Description	Manufacturer	Model/Part Number	Serial Number
Telemetry Transmitter (no display)	Spacelabs	96281-A05	6004

Peripherals in test setup boundary						
Description	Model/Part Number	Serial Number				
ECG lead wires shorting bars	Spacelabs	None	None			

Cables								
Cable Type	Shield	Length (m)	Ferrite	Connection 1	Connection 2			
ECG lead wires	No	0.6m	No	Telemetry Transmitter (no display)	Shorting bar			
PA = Cable is permanently attached to the device. Shielding and/or presence of ferrite may be unknown.								



## **Modifications**

## **Equipment Modifications**

Item	Date	Test	Modification	Note	Disposition of EUT	
1	1/19/2012 Strength of deliv		Tested as delivered to Test Station.	No EMI suppression devices were added or modified during this test.	EUT remained at Northwest EMC following the test.	
2	1/23/2012	Field Strength of Spurious Emissions	Tested as delivered to Test Station.	No EMI suppression devices were added or modified during this test.	EUT remained at Northwest EMC following the test.	
3	1/24/2012	Occupied Bandwidth	Tested as delivered to Test Station.	No EMI suppression devices were added or modified during this test.	EUT remained at Northwest EMC following the test.	
4	1/25/2012	Frequency Stability	Tested as delivered to Test Station.	No EMI suppression devices were added or modified during this test.	Scheduled testing was completed.	



Testing was performed using the mode(s) of operation and configuration(s) noted within the report. The individuals and/or the organization requesting the test provided the modes, configurations and settings used to complete the evaluation. The actual test parameters are specified in the test data, this includes items such as investigated frequency range (scanned) and test levels. The testing methods and performance specifications, as well as the test site used for the evaluation are indicated in the test data.

#### **TEST EQUIPMENT**

Description	Manufacturer	Model	ID	Last Cal.	Interval
Spectrum Analyzer	Agilent	E4440A	AFD	7/5/2011	12
DC Power Supply	Topward	TPS-2000	TPD	NCR	0
Multimeter	Tektronix	DMM912	MMH	1/28/2011	24
Near Field Probe	EMCO	7405	IPD	NCR	0
Chamber Temp. & Humidity	ESZ / Eurotherm	Dimension II	TBC	NCR	0
Controller					
Chamber, Temp./Humidity Chamber	Cincinnati Sub Zero (CSZ)	ZH-32-2-2-H/AC	TBA	8/20/2010	24

#### **MEASUREMENT UNCERTAINTY**

A measurement uncertainty estimation has been performed for each test per our internal quality document WP 342. The estimation is used to compare the measured result with its "true" or theoretically correct value. The expanded measurement uncertainty for radiated emissions measurements is less than +/- 4 dB, and for conducted emissions measurements is less than +/- 2.7 dB. Our measurement data meets or exceeds the measurement uncertainty requirements of CISPR 16-4; therefore, the test data can be compared directly to the specification limit to determine compliance. The calculations for measurement uncertainty are available upon request.

#### **TEST DESCRIPTION**

Per FCC 95.1115(e) Frequency stability. 'Manufacturers of wireless medical telemetry devices are responsible for ensuring frequency stability such that an emission is maintained within the band of operation under all of the manufacturer's specified conditions."

The Frequency Stability was measured using a near-field probe and a spectrum analyzer. The spectrum analyzer is configured with a precision frequency reference that exceeds the stability requirement of the transmitter.

#### Variation of Supply Voltage

The primary supply voltage was varied from 85% to 115% of the nominal voltage. A DC lab supply was used to vary the supply voltage.


#### Variation of Ambient Temperature


The EUT was placed inside a temperature / humidity chamber. The near-field probe was placed near the transmitter. A low-loss coaxial cable connected the near-field probe to the spectrum analyzer outside of the chamber.

The transmit frequency was recorded at the extremes of the specified temperature range (-30° C to +50° C) and at 10°C intervals.





	Telemetry Transmitter 96281-A05				Work Order: SP/	
Serial Number					Date: 01/2	
	: Spacelabs				Temperature: 23°	
Attendees					Humidity: 36%	
Project Tosted by	: Rod Peloquin	Power: Battery			Barometric Pres.: 101  Job Site: EV	
TEST SPECIFICAT		Test Method			Job Site. EV	99
FCC 95H:2012	10110	ANSI/TIA/EIA-60	3-C-2004			
1 00 3311.2012		7440071742174 00	10 O 2004			
COMMENTS						
	es, ECD shorting bar					
	· ·					
DEVIATIONS FRO	M TEST STANDARD					
		10120				
Configuration #	1	Rolly le Religs				
	Signature	0				
			Measured	Assigned	Error	
			Value (MHz)	Value (MHz)	(ppm)	Result
Narrow Band	Low, 1241, 608.0125					
	3.45 VDC. 115%		608.012557	608.0125	0.09	Pass
	3.0 VDC, 110%		608.01254	608.0125	0.03	Pass
	2.55 VDC 85%		608.01254	608.0125	0.07	Pass
	+50°		608.012523	608.0125	0.04	Pass
	+40°		608.012506	608.0125	0.01	Pass
	+30°		608.012473	608.0125	0.04	Pass
	+20° +10°		608.012523	608.0125 608.0125	0.04 0.02	Pass Pass
	0°		608.01249 608.01212	608.0125	0.62	Pass
	-10°		608.01212	608.0125	0.62	Pass
	-20°		608.011987	608.0125	0.84	Pass
	-30°		608.012203	608.0125	0.49	Pass
	High, 1480, 613.9875					
	3.45 VDC, 115%		613.987553	613.9875	0.09	Pass
	3.0 VDC, 100% 2.55 VDC 85%		613.987554 613.987553	613.9875 613.9875	0.09 0.09	Pass Pass
	+50°		613.98752	613.9875	0.03	Pass
	+40°		613.987503	613.9875	0	Pass
	+30°		613.987485	613.9875	0.02	Pass
	+20°		613.98752	613.9875	0.03	Pass
	+10° 0°		613.987484	613.9875	0.03	Pass
	-10°		613.987134 613.987117	613.9875 613.9875	0.6 0.62	Pass Pass
	-20°		613.986968	613.9875	0.87	Pass
	-30°		613.987217	613.9875	0.46	Pass
Wide Band						
	Low, 2242, 608.375					
	3.45 VDC, 115%		608.037554	608.0375	0.09	Pass
	3.0 VDC, 100% 2.55 VDC 85%		608.03754 608.03754	608.0375 608.0375	0.07 0.07	Pass Pass
	2.55 VDC 85% +50°		608.03754	608.0375	0.07	Pass
	+40°		608.03749	608.0375	0.02	Pass
	+30°		608.037473	608.0375	0.04	Pass
	+20°		608.037523	608.0375	0.04	Pass
	+10°		608.037506	608.0375	0.01	Pass
	0°		608.03712	608.0375	0.62	Pass
	-10° -20°		608.037103 608.03697	608.0375 608.0375	0.65 0.87	Pass Pass
	-20° -30°		608.03697	608.0375	0.87	Pass Pass
	High, 2478, 613.9375		555.007 107	333.3070	5.5.	1 455
	3.45 VDC, 115%		613.937554	613.9375	0.09	Pass
	3.0 VDC, 100%		613.937554	613.9375	0.09	Pass
	2.55 VDC 85%		613.937553	613.9375	0.09	Pass
	+50°		613.937503	613.9375	0	Pass
	+40° +30°		613.937485 613.93747	613.9375 613.9375	0.02 0.05	Pass Pass
	+30° +20°		613.937537	613.9375	0.06	Pass
	+10°		613.93752	613.9375	0.03	Pass
	0°		613.937118	613.9375	0.62	Pass
	-10°		613.937085	613.9375	0.68	Pass
	-20°		613.936969	613.9375	0.86	Pass
	-30°		613.937136	613.9375	0.59	Pass







Narrow Band, Low, 1241, 608.0125, 3.0 VDC, 100%								
		Me	easured	Assigned	Error			
		Val	ue (MHz)	Value (MHz)	(ppm)		Result	










ı	Narrow Band, Low, 1241, 608.0125, +50°								
Measured Assigned Error									
ı				Value (MHz)	Value (MHz)	(ppm)		Result	
ı				608.012523	608.0125	0.04	N/A	N/A	










	Narrow Band, Low, 1241, 608.0125, +30°									
Measured Assigned Error										
			Value (MHz)	Value (MHz)	(ppm)		Result			
			608.012473	608.0125	0.04	N/A	N/A			










	Narrow	Band, Low, 1241, 6	08.0125, +10°			
	Measure	d Assigned	Error			
	Value (MF	lz) Value (MHz)	(ppm)		Result	
	608 0124	0 608 0125	0.02	NI/A	N/A	1










	Narrow Band, Low, 1241, 608.0125, -10°									
Measured Assigned Error										
			Value (MHz)	Value (MHz)	(ppm)		Result			
			608.01212	608.0125	0.62	N/A	N/A			










	Narrow Band, Low, 1241, 608.0125, -30°									
Measured Assigned Error										
			Value (MHz)	Value (MHz)	(ppm)		Result			
			608.012203	608.0125	0.49	N/A	N/A			










	Narrow Band, Hi	gh, 1480, 613.987	5, 3.0 VDC, 100%		
	Measured	Assigned	Error		
	Value (MHz)	Value (MHz)	(ppm)	Result	










	Narrow Band, High, 1480, 613.9875, +50°									
Measured Assigned Error										
			Value (MHz)	Value (MHz)	(ppm)		Result			
			613.98752	613.9875	0.03	N/A	N/A			










	Narrow Band, High, 1480, 613.9875, +30°									
Measured Assigned Error										
			Value (MHz)	Value (MHz)	(ppm)		Result			
			613.987485	613.9875	0.02	N/A	N/A			










		Narrow Band	d, High, 1480, 613	3.9875, +10°		
		Measured	Assigned	Error		
	V	Value (MHz)	Value (MHz)	(ppm)	Result	









	Narrov	w Band, High, 1480, 6	13.9875, -10°		
	Measur	ed Assigned	Error		
	Value (M	Hz) Value (MHz)	(ppm)	Result	













	Narrow Ban	d, High, 1480, 610	3.9875, -30°		
	Measured	Assigned	Error		
	Value (MHz)	Value (MHz)	(ppm)	Result	










	Wide Band, Low, 2242, 608.375, 3.0 VDC, 100%									
Measured Assigned Error										
			Value (MHz)	Value (MHz)	(ppm)		Result			
			608.03754	608.0375	0.07	N/A	N/A			










	Wide Band, Low, 2242, 608.375, +50°									
Measured Assigned Error										
			Value (MHz)	Value (MHz)	(ppm)		Result			
			608.037506	608.0375	0.01	N/A	N/A			










Wide Band, Low, 2242, 608.375, +30°									
	Measured Assigned Error								
			Value (MHz)	Value (MHz)	(ppm)		Result		










Wide Band, Low, 2242, 608.375, +10°									
	Measured Assigned Error								
			Value (MHz)	Value (MHz)	(ppm)		Result		










Wide Band, Low, 2242, 608.375, -10°									
	Measured Assigned Error								
			Value (MHz)	Value (MHz)	(ppm)		Result		
			value (iiii iz)	value (iiii iz)	(PPIII)		resuit		










Wide Band, Low, 2242, 608.375, -30°									
Measured Assigned Error									
Value (MHz) Value (MHz) (ppm) Result									
608.037187 608.0375 0.51 N/A N/A									










Wide Band, High, 2478, 613.9375, 3.0 VDC, 100%									
		Measured	Assigned	Error					
		Value (MHz)	Value (MHz)	(ppm)		Result			
		613 037554	613 0375	0.00	NI/A	NI/A			










	Wide Band, High, 2478, 613.9375, +50°									
Measured Assigned Error										
Value (MHz) Value (MHz) (ppm) Result							Result			
	613.937503 613.9375 0 N/A N/A									










	Wide Band, High, 2478, 613.9375, +30°									
Measured Assigned Error										
Value (MHz) Value (MHz)							Result			
	Value (MHz)         Value (MHz)         (ppm)         Result           613.93747         613.9375         0.05         N/A         N/A									










	Wide Band, High, 2478, 613.9375, +10°									
Measured Assigned Error										
	Value (MHz) Value (MHz) (ppm) Result									
	613.93752 613.9375 0.03 N/A N/A									










	Wide Band, High, 2478, 613.9375, -10°									
Measured Assigned Error										
			Value (MHz)	Value (MHz)	(ppm)		Result			
	613.937085 613.9375 0.68 N/A N/A									







	Wide Band, High, 2478, 613.9375, -30°									
Measured Assigned Error										
			Value (MHz)	Value (MHz)	(ppm)		Result			
	613.937136 613.9375 0.59 N/A N/A									





#### OCCUPIED BANDWIDTH

Testing was performed using the mode(s) of operation and configuration(s) noted within the report. The individuals and/or the organization requesting the test provided the modes, configurations and settings used to complete the evaluation. The actual test parameters are specified in the test data, this includes items such as investigated frequency range (scanned) and test levels. The testing methods and performance specifications, as well as the test site used for the evaluation are indicated in the test data.

#### **TEST EQUIPMENT**

Description	Manufacturer	Model	ID	Last Cal.	Interval
Spectrum Analyzer	Agilent	E4440A	AFD	7/5/2011	12
Near Field Probe	EMCO	7405	IPD	NCR	0

#### **MEASUREMENT UNCERTAINTY**

A measurement uncertainty estimation has been performed for each test per our internal quality document WP 342. The estimation is used to compare the measured result with its "true" or theoretically correct value. The expanded measurement uncertainty for radiated emissions measurements is less than +/- 4 dB, and for conducted emissions measurements is less than +/- 2.7 dB. Our measurement data meets or exceeds the measurement uncertainty requirements of CISPR 16-4; therefore, the test data can be compared directly to the specification limit to determine compliance. The calculations for measurement uncertainty are available upon request.


#### **TEST DESCRIPTION**

Per 47 CFR 2.1049, the 99% bandwidth was measured utilizing the analyzer's Occupied BW measurement function. The antenna is integral to the EUT, so a radiated measurement was made using a spectrum analyzer and a near field probe. The analyzer's resolution bandwidth was sufficiently narrow to plot the actual bandwidth of the signal and not the filter response curve of the spectrum analyzer. The resolution bandwidth was >1% of the 99% bandwidth and the video bandwidth was greater than or equal to the resolution bandwidth.

The occupied bandwidth was measured with the EUT configured for continuous modulated operation at the low, mid, and high channel of the operational band.




FIIT: T	Telemetry Transmitter 96281-A05				Work Order:	SDAC0475	
Serial Number: 6						01/24/12	
Customer: S					Temperature:		
Attendees: N					Humidity:		
Project: N					Barometric Pres.:		
	Rod Peloquin		Power:	3.0 VDC	Job Site:		
TEST SPECIFICATIO				Test Method			
FCC 95H:2012				ANSI/TIA/EIA-603-C-2004			
COMMENTS							
With ECG lead wires	·		Rolly to Reley	_			
Configuration #	4		roturn le return				
	· I	Signature	0	<u> </u>	Value	Limit	Result
	<u> </u>	Signature			Value	Limit	Result
Narrow Band	Low, 1241, 608.0125	Signature		3	<b>Value</b> 8.7316	Limit N/A	Result
larrow Band	Low, 1241, 608.0125 Mid, 1360, 610.9875	Signature		7			
Jarrow Band L		Signature			8.7316	N/A	N/A
Varrow Band L N F Wide Band	Mid, 1360, 610.9875 High, 1480, 613.9875	Signature		,	8.7316 8.7532 8.7853	N/A N/A N/A	N/A N/A N/A
Narrow Band L N H Wide Band L	Mid, 1360, 610.9875 High, 1480, 613.9875 Low, 2242, 608.375	Signature		7	8.7316 8.7532 8.7853 18.1122	N/A N/A N/A	N/A N/A N/A
Narrow Band L N F Wide Band L	Mid, 1360, 610.9875 High, 1480, 613.9875	Signature		,	8.7316 8.7532 8.7853	N/A N/A N/A	N/A N/A N/A





Narrow Band, Mid, 1360, 610.9875										
					Value	Limit	Result			
					8.75 kHz	N/A	N/A			










	Wide Band, Low, 2242, 608.375										
					Value	Limit	Result				
					18.11 kHz	N/A	N/A				











Testing was performed using the mode(s) of operation and configuration(s) noted within the report. The individuals and/or the organization requesting the test provided the modes, configurations and settings used to complete the evaluation. The actual test parameters are specified in the test data, this includes items such as investigated frequency range (scanned) and test levels. The testing methods and performance specifications, as well as the test site used for the evaluation are indicated in the test data. The test data represents the configuration / operating mode/ model that produced the highest emission levels as compared to the specification limit

### **MODES OF OPERATION**

Transmitting WBT (Wideband), Low = Ch.2242, 608.0375 MHz
Transmitting WBT (Wideband), Mid = Ch. 2360, 610.9875 MHz
Transmitting WBT (Wideband), High = Ch. 2478, 613.9375 MHz
Transmitting NBT (Narrowband), Low = Ch.1241, 608.0125 MHz
Transmitting NBT (Narrowband), Mid = Ch. 1360, 610.9875 MHz
Transmitting NBT (Narrowband), High = Ch. 1480, 613.9875 MHz

# POWER SETTINGS INVESTIGATED

Battery

#### **CONFIGURATIONS INVESTIGATED**

SPAC0475 - 1

#### FREQUENCY RANGE INVESTIGATED

Start Frequency 30 MHz	Stop Frequency	7 GHz
------------------------	----------------	-------

#### SAMPLE CALCULATIONS

Radiated Emissions: Field Strength = Measured Level + Antenna Factor + Cable Factor - Amplifier Gain + Distance Adjustment Factor + External Attenuation

#### **TEST EQUIPMENT**

Description	Manufacturer	Model	ID	Last Cal.	Interval
Spectrum Analyzer	Agilent	E4440A	AFD	7/5/2011	12
High Pass Filter	Micro-Tronics	50108	HGF	1/9/2012	24
Attenuator	Pasternack	PE7005-20	AUN	7/5/2011	12
Attenuator	INMET	64671 6A-10dB	AUI	10/6/2011	12
Low Pass Filter 0-425 MHz	Micro-Tronics	LPM50003	HGL	7/14/2010	24
.5-1 GHz Notch Filter	K&L Microwave	3TNF-500/1000-N/N	HFT	1/11/2011	24
Pre-Amplifier	Miteq	AM-1616-1000	AVM	6/20/2011	12
Antenna, Biconilog	EMCO	3141	AXG	3/15/2010	24
EV12 Cables	N/A	Bilog Cables	EVS	6/20/2011	12
Antenna, Horn	ETS	3115	AIB	9/8/2010	24
Pre-Amplifier	Miteq	AMF-3D00100800-32-13P	AVF	6/20/2011	12
EV12 Cables	N/A	Double Ridge Horn Cables	EVT	10/6/2011	12

### **MEASUREMENT BANDWIDTHS**

Frequency Range	Peak Data	Quasi-Peak Data	Average Data
(MHz)	(kHz)	(kHz)	(kHz)
0.01 - 0.15	1.0	0.2	0.2
0.15 - 30.0	10.0	9.0	9.0
30.0 - 1000	100.0	120.0	120.0
Above 1000	1000.0	N/A	1000.0

Measurements were made using the IF bandwidths and detectors specified. No video filter was used, except in the case of the FCC Average Measurements above 1GHz. In that case, a peak detector with a 10Hz video bandwidth was used.

# MEASUREMENT UNCERTAINTY

A measurement uncertainty estimation has been performed for each test per our internal quality document WP 342. The estimation is used to compare the measured result with its "true" or theoretically correct value. The expanded measurement uncertainty for radiated emissions measurements is less than +/- 2.7 dB. Our measurement data meets or exceeds the measurement uncertainty requirements of CISPR 16-4; therefore, the test data can be compared directly to the specification limit to determine compliance. The calculations for measurement uncertainty are available upon request.

#### **TEST DESCRIPTION**

The highest gain of each type of antenna to be used with the EUT was tested. The EUT was configured for low, mid, and high band transmit frequencies. For each configuration, the spectrum was scanned throughout the specified range. While scanning, emissions from the EUT were maximized by rotating the EUT on a turntable, adjusting the position of the EUT and the EUT antenna in three orthogonal axis, and adjusting measurement antenna height and polarization, and manipulating the EUT antenna in 3 orthogonal planes (per ANSI C63.10:2009). A preamp was used for this test in order to provide sufficient measurement sensitivity.



# FIELD STRENGTH OF SPURIOUS EMISSIONS

EUT:	Telemetry Transmitter 96281-A05	Work Order:	SPAC0475	
Serial Number:	6004	Date:	01/20/12	
Customer:	Spacelabs	Temperature:	23	
Attendees:	Hugh Carter		Humidity:	38%
Project:	None	Barometric Pres.:	101.76 kPa	
Tested by:	Rod Peloquin	Job Site:	EV12	

**TEST SPECIFICATIONS** 

Test Method

FCC 95H·201

ANSI/TIA/EIA-603-C-2004

TEST PARAMETERS
Antenna Height(s) (m)

Antenna Height(s) (m) 1 - 4 Test Distance (m) 3

COMMENTS


With ECG lead wires, ECD shorting bar


# **EUT OPERATING MODES**

Transmitting WBT

# **DEVIATIONS FROM TEST STANDARD**

No doviations





Freq	Amplitude	Factor	Azimuth	Height	Distance	External Attenuation	Polarity	Detector	Distance Adjustment	Adjusted	Spec. Limit	Compared to Spec.	
(MHz)	(dBuV)	(dB)	(degrees)	(meters)	(meters)	(dB)			(dB)	dBuV/m	dBuV/m	(dB)	Comments
2431.343	26.6	-1.3	65.0	1.0	3.0	0.0	V-Horn	AV	0.0	25.3	54.0	-28.7	Low Channel, Wideband
2432.243	26.6	-1.3	224.0	1.0	3.0	0.0	H-Horn	AV	0.0	25.3	54.0	-28.7	Low Channel, Wideband
2454.550	26.5	-1.2	254.0	1.0	3.0	0.0	H-Horn	AV	0.0	25.3	54.0	-28.7	High Channel, Wideband
2443.387	26.5	-1.3	3.0	1.0	3.0	0.0	H-Horn	AV	0.0	25.2	54.0	-28.8	Mid Channel, Wideband
2443.860	26.5	-1.3	228.0	1.0	3.0	0.0	V-Horn	AV	0.0	25.2	54.0	-28.8	Mid Channel, Wideband
2454.317	26.4	-1.2	229.0	1.0	3.0	0.0	V-Horn	AV	0.0	25.2	54.0	-28.8	High Channel, Wideband
1841.817	27.9	-3.7	274.0	1.0	3.0	0.0	H-Horn	AV	0.0	24.2	54.0	-29.8	High Channel, Wideband
1824.150	27.8	-3.9	84.0	1.0	3.0	0.0	H-Horn	AV	0.0	23.9	54.0	-30.1	Low Channel, Wideband
1824.230	27.7	-3.9	18.0	1.0	3.0	0.0	V-Horn	AV	0.0	23.8	54.0	-30.2	Low Channel, Wideband
1841.863	27.5	-3.7	325.0	1.0	3.0	0.0	V-Horn	AV	0.0	23.8	54.0	-30.2	High Channel, Wideband
1833.510	27.5	-3.8	56.0	2.5	3.0	0.0	V-Horn	AV	0.0	23.7	54.0	-30.3	Mid Channel, Wideband
1834.557	27.5	-3.8	343.0	1.0	3.0	0.0	H-Horn	AV	0.0	23.7	54.0	-30.3	Mid Channel, Wideband
1216.063	28.9	-8.7	53.0	1.0	3.0	0.0	V-Horn	AV	0.0	20.2	54.0	-33.8	Low Channel, Wideband
1227.837	28.6	-8.5	7.0	1.0	3.0	0.0	V-Horn	AV	0.0	20.1	54.0	-33.9	High Channel, Wideband
1221.950	28.8	-8.7	82.0	1.0	3.0	0.0	V-Horn	AV	0.0	20.1	54.0	-33.9	Mid Channel, Wideband
1227.997	28.5	-8.5	340.0	1.2	3.0	0.0	H-Horn	AV	0.0	20.0	54.0	-34.0	High Channel, Wideband
1215.463	28.5	-8.7	195.0	1.0	3.0	0.0	H-Horn	AV	0.0	19.8	54.0	-34.2	Low Channel, Wideband
1221.743	28.4	-8.7	265.0	1.0	3.0	0.0	H-Horn	AV	0.0	19.7	54.0	-34.3	Mid Channel, Wideband
2454.917	39.7	-1.2	229.0	1.0	3.0	0.0	V-Horn	PK	0.0	38.5	74.0	-35.5	High Channel, Wideband
2442.970	39.7	-1.3	228.0	1.0	3.0	0.0	V-Horn	PK	0.0	38.4	74.0	-35.6	Mid Channel, Wideband
2444.613	39.6	-1.3	3.0	1.0	3.0	0.0	H-Horn	PK	0.0	38.3	74.0	-35.7	Mid Channel, Wideband
2454.827	39.5	-1.2	254.0	1.0	3.0	0.0	H-Horn	PK	0.0	38.3	74.0	-35.7	High Channel, Wideband
2431.163	39.5	-1.3	65.0	1.0	3.0	0.0	V-Horn	PK	0.0	38.2	74.0	-35.8	Low Channel, Wideband
2432.163	39.5	-1.3	224.0	1.0	3.0	0.0	H-Horn	PK	0.0	38.2	74.0	-35.8	Low Channel, Wideband
1833.490	41.7	-3.8	343.0	1.0	3.0	0.0	H-Horn	PK	0.0	37.9	74.0	-36.1	Mid Channel, Wideband
1832.667	41.5	-3.8	56.0	2.5	3.0	0.0	V-Horn	PK	0.0	37.7	74.0	-36.3	Mid Channel, Wideband

Freq	Amplitude	Factor	Azimuth	Height	Distance	External Attenuation	Polarity	Detector	Distance Adjustment	Adjusted	Spec. Limit	Compared to Spec.	
(MHz)	(dBuV)	(dB)	(degrees)	(meters)	(meters)	(dB)			(dB)	dBuV/m	dBuV/m	(dB)	Comments
1823.863	41.6	-3.9	18.0	1.0	3.0	0.0	V-Horn	PK	0.0	37.7	74.0	-36.3	Low Channel, Wideband
1823.870	41.4	-3.9	84.0	1.0	3.0	0.0	H-Horn	PK	0.0	37.5	74.0	-36.5	Low Channel, Wideband
1842.090	41.1	-3.7	274.0	1.0	3.0	0.0	H-Horn	PK	0.0	37.4	74.0	-36.6	High Channel, Wideband
1842.447	40.8	-3.7	325.0	1.0	3.0	0.0	V-Horn	PK	0.0	37.1	74.0	-36.9	High Channel, Wideband
1228.563	42.1	-8.5	7.0	1.0	3.0	0.0	V-Horn	PK	0.0	33.6	74.0	-40.4	High Channel, Wideband
1228.603	42.0	-8.5	340.0	1.2	3.0	0.0	H-Horn	PK	0.0	33.5	74.0	-40.5	High Channel, Wideband
1215.297	42.0	-8.7	195.0	1.0	3.0	0.0	H-Horn	PK	0.0	33.3	74.0	-40.7	Low Channel, Wideband
1215.997	42.0	-8.7	53.0	1.0	3.0	0.0	V-Horn	PK	0.0	33.3	74.0	-40.7	Low Channel, Wideband
1221.083	41.3	-8.7	82.0	1.0	3.0	0.0	V-Horn	PK	0.0	32.6	74.0	-41.4	Mid Channel, Wideband
1221.157	41.2	-8.7	265.0	1.0	3.0	0.0	H-Horn	PK	0.0	32.5	74.0	-41.5	Mid Channel, Wideband



# **FIELD STRENGTH OF SPURIOUS EMISSIONS**

EUT	Telemetry Transmitter 96281-A05	Work Order:	SPAC0475	
Serial Number	: 6004	Date:	01/23/12	
Customer	Spacelabs	Temperature:	23	
Attendees	: Hugh Carter		Humidity:	38%
Project	None		Barometric Pres.:	101.76 kPa
Tested by	Rod Peloquin	Power: Battery	Job Site:	EV12

**TEST SPECIFICATIONS** 

**Test Method** 

ANSI/TIA/EIA-603-C-2004


TEST PARAMETERS


Test Distance (m) Antenna Height(s) (m)

COMMENTS
With ECG lead wires, ECD shorting bar

# EUT OPERATING MODES Transmitting NBT

# **DEVIATIONS FROM TEST STANDARD**





Ī							External			Distance			Compared to	
	Freq	Amplitude	Factor	Azimuth	Height	Distance	Attenuation	Polarity	Detector	Adjustment	Adjusted	Spec. Limit	Spec.	
	(MHz)	(dBuV)	(dB)	(degrees)	(meters)	(meters)	(dB)			(dB)	dBuV/m	dBuV/m	(dB)	Comments
	2431.097	26.5	-1.3	232.0	1.0	3.0	0.0	V-Horn	AV	0.0	25.2	54.0	-28.8	Low Channel, Narrowband
	2431.703	26.5	-1.3	206.0	1.0	3.0	0.0	H-Horn	AV	0.0	25.2	54.0	-28.8	Low Channel, Narrowband
	2443.043	26.5	-1.3	83.0	1.0	3.0	0.0	H-Horn	AV	0.0	25.2	54.0	-28.8	Mid Channel, Narrowband
	2444.713	26.5	-1.3	66.0	2.1	3.0	0.0	V-Horn	AV	0.0	25.2	54.0	-28.8	Mid Channel, Narrowband
	2456.100	26.4	-1.2	35.0	2.6	3.0	0.0	V-Horn	AV	0.0	25.2	54.0	-28.8	High Channel, Narrowband
	2456.163	26.4	-1.2	354.0	1.0	3.0	0.0	H-Horn	AV	0.0	25.2	54.0	-28.8	High Channel, Narrowband
	1824.183	27.7	-3.9	85.0	1.0	3.0	0.0	H-Horn	AV	0.0	23.8	54.0	-30.2	Low Channel, Narrowband
	1840.690	27.5	-3.7	102.0	1.0	3.0	0.0	V-Horn	AV	0.0	23.8	54.0	-30.2	High Channel, Narrowband
	1840.747	27.5	-3.7	158.0	2.3	3.0	0.0	H-Horn	AV	0.0	23.8	54.0	-30.2	High Channel, Narrowband
	1824.333	27.6	-3.9	29.0	1.0	3.0	0.0	V-Horn	AV	0.0	23.7	54.0	-30.3	Low Channel, Narrowband
	1833.063	27.5	-3.8	186.0	1.0	3.0	0.0	V-Horn	AV	0.0	23.7	54.0	-30.3	Mid Channel, Narrowband
	1833.210	27.5	-3.8	144.0	1.0	3.0	0.0	H-Horn	AV	0.0	23.7	54.0	-30.3	Mid Channel, Narrowband
	1228.020	28.7	-8.5	6.0	1.0	3.0	0.0	V-Horn	AV	0.0	20.2	54.0	-33.8	High Channel, Narrowband
	1216.097	28.6	-8.7	357.0	1.0	3.0	0.0	V-Horn	AV	0.0	19.9	54.0	-34.1	Low Channel, Narrowband
	1227.800	28.4	-8.5	290.0	1.0	3.0	0.0	H-Horn	AV	0.0	19.9	54.0	-34.1	High Channel, Narrowband
	1222.127	28.5	-8.7	307.0	1.0	3.0	0.0	H-Horn	AV	0.0	19.8	54.0	-34.2	Mid Channel, Narrowband
	1222.240	28.5	-8.7	236.0	1.0	3.0	0.0	V-Horn	AV	0.0	19.8	54.0	-34.2	Mid Channel, Narrowband
	1215.973	28.4	-8.7	317.0	1.0	3.0	0.0	H-Horn	AV	0.0	19.7	54.0	-34.3	Low Channel, Narrowband
	2431.447	40.8	-1.3	206.0	1.0	3.0	0.0	H-Horn	PK	0.0	39.5	74.0	-34.5	Low Channel, Narrowband
	2443.667	40.6	-1.3	83.0	1.0	3.0	0.0	H-Horn	PK	0.0	39.3	74.0	-34.7	Mid Channel, Narrowband
	2455.877	40.5	-1.2	354.0	1.0	3.0	0.0	H-Horn	PK	0.0	39.3	74.0	-34.7	High Channel, Narrowband
	2455.990	39.9	-1.2	35.0	2.6	3.0	0.0	V-Horn	PK	0.0	38.7	74.0	-35.3	High Channel, Narrowband
	2431.110	39.8	-1.3	232.0	1.0	3.0	0.0	V-Horn	PK	0.0	38.5	74.0	-35.5	Low Channel, Narrowband
	2444.623	39.7	-1.3	66.0	2.1	3.0	0.0	V-Horn	PK	0.0	38.4	74.0	-35.6	Mid Channel, Narrowband
	1832.720	41.2	-3.8	144.0	1.0	3.0	0.0	H-Horn	PK	0.0	37.4	74.0	-36.6	Mid Channel, Narrowband
	1842.077	41.0	-3.7	102.0	1.0	3.0	0.0	V-Horn	PK	0.0	37.3	74.0	-36.7	High Channel, Narrowband

Freq	Amplitude	Factor	Azimuth	Height	Distance	External Attenuation	Polarity	Detector	Distance Adjustment	Adjusted	Spec. Limit	Compared to Spec.	
(MHz)	(dBuV)	(dB)	(degrees)	(meters)	(meters)	(dB)			(dB)	dBuV/m	dBuV/m	(dB)	Comments
1831.977	40.6	-3.8	186.0	1.0	3.0	0.0	V-Horn	PK	0.0	36.8	74.0	-37.2	Mid Channel, Narrowband
1842.807	40.5	-3.7	158.0	2.3	3.0	0.0	H-Horn	PK	0.0	36.8	74.0	-37.2	High Channel, Narrowband
1823.060	40.5	-3.9	85.0	1.0	3.0	0.0	H-Horn	PK	0.0	36.6	74.0	-37.4	Low Channel, Narrowband
1824.003	40.2	-3.9	29.0	1.0	3.0	0.0	V-Horn	PK	0.0	36.3	74.0	-37.7	Low Channel, Narrowband
1222.597	42.8	-8.7	307.0	1.0	3.0	0.0	H-Horn	PK	0.0	34.1	74.0	-39.9	Mid Channel, Narrowband
1226.983	41.7	-8.5	6.0	1.0	3.0	0.0	V-Horn	PK	0.0	33.2	74.0	-40.8	High Channel, Narrowband
1227.730	41.4	-8.5	290.0	1.0	3.0	0.0	H-Horn	PK	0.0	32.9	74.0	-41.1	High Channel, Narrowband
1222.260	41.6	-8.7	236.0	1.0	3.0	0.0	V-Horn	PK	0.0	32.9	74.0	-41.1	Mid Channel, Narrowband
1216.577	41.4	-8.7	357.0	1.0	3.0	0.0	V-Horn	PK	0.0	32.7	74.0	-41.3	Low Channel, Narrowband
1215.613	41.3	-8.7	317.0	1.0	3.0	0.0	H-Horn	PK	0.0	32.6	74.0	-41.4	Low Channel, Narrowband

Testing was performed using the mode(s) of operation and configuration(s) noted within the report. The individuals and/or the organization requesting the test provided the modes, configurations and settings used to complete the evaluation. The actual test parameters are specified in the test data, this includes items such as investigated frequency range (scanned) and test levels. The testing methods and performance specifications, as well as the test site used for the evaluation are indicated in the test data. The test data represents the configuration / operating mode/ model that produced the highest emission levels as compared to the specification limit.

#### MODES OF OPERATION

Transmitting WBT (Wideband), Low = Ch.2242, 608.0375 MHz	
Transmitting WBT (Wideband), Mid = Ch. 2360, 610.9875 MHz	
Transmitting WBT (Wideband), High = Ch. 2478, 613.9375 MHz	
Transmitting NBT (Narrowband), Low = Ch.1241, 608.0125 MHz	
Transmitting NBT (Narrowband), Mid = Ch. 1360, 610.9875 MHz	
Transmitting NBT (Narrowband), High = Ch. 1480, 613.9875 MHz	

#### **POWER SETTINGS INVESTIGATED**

Battery

### **CONFIGURATIONS INVESTIGATED**

SPAC0475 - 1

### FREQUENCY RANGE INVESTIGATED

Start Frequency 60	08 MHz	Stop Frequency	616 MHz
--------------------	--------	----------------	---------

### **SAMPLE CALCULATIONS**

Radiated Emissions: Field Strength = Measured Level + Antenna Factor + Cable Factor - Amplifier Gain + Distance Adjustment Factor + External Attenuation

#### **TEST EQUIPMENT**

Description	Manufacturer	Model	ID	Last Cal.	Interval
Spectrum Analyzer	Agilent	E4440A	AFD	7/5/2011	12
Antenna, Biconilog	EMCO	3141	AXG	3/15/2010	24
EV12 Cables	N/A	Bilog Cables	EVS	6/20/2011	12

# **MEASUREMENT BANDWIDTHS**

Frequency Range (MHz)	Peak Data (kHz)	Quasi-Peak Data (kHz)	Average Data (kHz)		
0.01 - 0.15	1.0	0.2	0.2		
0.15 - 30.0	10.0	9.0	9.0		
30.0 - 1000	100.0	120.0	120.0		
Above 1000	1000.0	N/A	1000.0		

Measurements were made using the IF bandwidths and detectors specified. No video filter was used, except in the case of the FCC Average Measurements above 1GHz. In that case, a peak detector with a 10Hz video bandwidth was used.

# **MEASUREMENT UNCERTAINTY**

A measurement uncertainty estimation has been performed for each test per our internal quality document WP 342. The estimation is used to compare the measured result with its "true" or theoretically correct value. The expanded measurement uncertainty for radiated emissions measurements is less than +/- 4 dB, and for conducted emissions measurements is less than +/- 2.7 dB. Our measurement data meets or exceeds the measurement uncertainty requirements of CISPR 16-4; therefore, the test data can be compared directly to the specification limit to determine compliance. The calculations for measurement uncertainty are available upon request.

#### **TEST DESCRIPTION**

The antennas to be used with the EUT were tested. The EUT was transmitting and while set at the lowest channel, a middle channel, and the highest channel available. While scanning, emissions from the EUT were maximized by rotating the EUT, adjusting the measurement antenna height and polarization, and manipulating the EUT and EUT antenna in 3 orthogonal planes (per ANSI C63.10:2009).



	TERRAL WIN										
EUT:	Telemetry Transmitter 96281-A05	Work Order:	SPAC0475								
Serial Number:	6004	Date:	01/19/12								
Customer:	Spacelabs	Temperature:	23								
Attendees:	Hugh Carter	Humidity:	38%								
Project:	None	Barometric Pres.:	101.76 kPa								
Tested by:	Rod Peloquin	Job Site:	EV12								

TEST SPECIFICATIONS FCC 95H:2012

Test Method

ANSI/TIA/EIA-603-C-2004


TEST PARAMETERS
Antenna Height(s) (m) Test Distance (m)

COMMENTS
With ECG lead wires, ECD shorting bar

# EUT OPERATING MODES Transmitting NBT

# DEVIATIONS FROM TEST STANDARD No deviations.





Freq (MHz)	Amplitude (dBuV)	Factor (dB)	Azimuth (degrees)	Height (meters)	Distance (meters)	External Attenuation (dB)	Polarity	Detector	Distance Adjustment (dB)	Adjusted 80.00	Spec. Limit 80.00	Compared to Spec. (dB)	Comments
608.018	67.8	29.8	112.0	1.3	3.0	0.0	H-Bilog	QP	0.0	97.6	106.0	-8.4	Low Channel, Narrowband, EUT horizontal
610.994	66.8	29.8	108.0	1.2	3.0	0.0	H-Bilog	QP	0.0	96.6	106.0	-9.4	Mid Channel, Narrowband, EUT horizontal
613.993	64.9	29.9	42.0	1.2	3.0	0.0	H-Bilog	QP	0.0	94.8	106.0	-11.2	High Channel, Narrowband, EUT horizontal
608.018	64.2	29.8	84.0	1.3	3.0	0.0	V-Bilog	QP	0.0	94.0	106.0	-12.0	Low Channel, Narrowband, EUT horizontal
610.994	63.8	29.8	84.0	1.3	3.0	0.0	V-Bilog	QP	0.0	93.6	106.0	-12.4	Mid Channel, Narrowband, EUT horizontal
613.993	62.5	29.9	85.0	1.4	3.0	0.0	V-Bilog	QP	0.0	92.4	106.0	-13.6	Mid Channel, Narrowband, EUT horizontal



EUT: Telemetry Transmitter 96281-A05
Serial Number: 6004 Work Order: SPAC0475
Date: 01/23/12 Temperature: 23
Humidity: 38%
Barometric Pres.: 101.76 kPa
Job Site: EV12 Customer: Spacelabs
Attendees: None Project: None
Tested by: Rod Peloquin Power: Battery

TEST SPECIFICATIONS FCC 95H:2012

Test Method ANSI/TIA/EIA-603-C-2004


TEST PARAMETERS
Antenna Height(s) (m)


Test Distance (m)

COMMENTS
With ECG lead wires, ECD shorting bar

EUT OPERATING MODES
Transmitting WPT

# DEVIATIONS FROM TEST STANDARD No deviations.





Freq	Amplitude	Factor	Azimuth	Height	Distance	External Attenuation	Polarity	Detector	Distance Adjustment	Adjusted	Spec. Limit	Compared to Spec.	
(MHz)	(dBuV)	(dB)	(degrees)	(meters)	(meters)	(dB)	· Oldiny	Detector	(dB)	dBuV/m	dBuV/m	(dB)	Comments
608.042	67.8	29.8	117.0	1.3	3.0	0.0	H-Bilog	QP	0.0	97.6	106.0	-8.4	Low Channel, Wideband, EUT horizontal
610.993	67.2	29.8	306.0	1.3	3.0	0.0	H-Bilog	QP	0.0	97.0	106.0	-9.0	Mid Channel, Wideband, EUT horizontal
610.992	66.9	29.8	121.0	1.3	3.0	0.0	H-Bilog	QP	0.0	96.7	106.0	-9.3	Mid Channel, Wideband, EUT on side
613.944	66.5	29.9	304.0	1.3	3.0	0.0	H-Bilog	QP	0.0	96.4	106.0	-9.6	High Channel, Wideband, EUT horizontal
610.993	64.5	29.8	91.0	1.6	3.0	0.0	V-Bilog	QP	0.0	94.3	106.0	-11.7	Mid Channel, Wideband, EUT on side
613.943	63.9	29.9	234.0	1.0	3.0	0.0	V-Bilog	QP	0.0	93.8	106.0	-12.2	High Channel, Wideband, EUT on side
610.992	63.9	29.8	99.0	2.8	3.0	0.0	V-Bilog	QP	0.0	93.7	106.0	-12.3	Mid Channel, Wideband, EUT horizontal
608.043	63.3	29.8	238.0	1.0	3.0	0.0	V-Bilog	QP	0.0	93.1	106.0	-12.9	Low Channel, Wideband, EUT on side
610.992	62.2	29.8	53.0	1.7	3.0	0.0	V-Bilog	QP	0.0	92.0	106.0	-14.0	Mid Channel, Wideband, EUT vertical
610.992	59.2	29.8	187.0	1.3	3.0	0.0	H-Bilog	QP	0.0	89.0	106.0	-17.0	Mid Channel, Wideband, EUT vertical