Measurement/Technical Report

General Information

Applicant:	Spacelabs Medical	
Address:	PO Box 97013	
City, State, Zip	Redmond, WA 98073-9713	
Test Requested By:	Steve Cantwell	
Model:	6300 and 6330	
FCC ID:	CM6010-0914-00	
First Date of Test:	July 11, 2001	
Last Date of Test:	February 18, 2002	
Receipt Date of Samples:	July 11, 2001	
Job Number	SPAC0264	

Scope

Regulatory Authority	Federal Communications Commission	
Approval Type	Class II Permissive Change	
Equipment Type	Spread Spectrum Transmitter	
Rule Part	FCC Part 15.247	
Rule Exemptions	None	
Related Submittals or Grants	The original grant was issued May 16, 1995	

Report Information

Prepared By	Vicki Albertson, Technical Report and Documentation Manager Northwest EMC, Inc.	
Signature	Vicki Albertson	
Issued By	Northwest EMC, Inc. 22975 NW Evergreen Parkway, Suite 400 Hillsboro, Oregon 97124 Ph. (503) 844-4066 Fax (503) 844-3826	
Report Number	SPAC0264	
Date Issued	February 22, 2002	

EUT Information

Frequency Range	2402-2480 MHz
Antenna Conducted Output Power	83mW
Modulation	Frequency Hopping
Channel Bandwidth	964 kHz
Antenna(s)	Armstrong/Centurion Ceiling Tile Antenna
Supply Power	5Vdc from host system

Test Facility

The measurement facility used to collect the radiated and conducted data is located at

Northwest EMC, Inc. 22975 NW Evergreen Parkway, Suite 400 Hillsboro, OR 97124 (503) 844-4066

Fax: 844-3826

This site has been fully described in a report filed with the FCC (Federal Communications Commission), and accepted by the FCC in a letter maintained in our files.

Laboratory Accreditation

Northwest EMC, Inc. is recognized under the United States Department of Commerce, National Institute of Standards and Technology, National Voluntary Laboratory Accreditation Program (NVLAP) for satisfactory compliance with criteria established in Title 15, Part 285 Code of Federal Regulations. These criteria encompass the requirements of ISO/IEC Guide 25 and the relevant requirements of ISO 9002 (ANSI/ASQC Q92-1987) as suppliers of calibration or test results. NVLAP Lab Code: 200059-0.