

Information about the Applicant

Company Name	Spacelabs Medical
Address	5250 220 th Ave. SE
City, State, Zip	Issaquah, WA 98027-7018
Requested By	Stan Wiley
Job Number	90341-5 WMTS Transmitter
Model	Spacelabs Medical
FCC ID	CM676A90341-WMTS
Agent	Northwest EMC, Inc.
Approval Type	Certification
Equipment Type	Licensed Non-Broadcast Transmitter Worn on Body
Rule Part	95H

Overview

This application is for the Spacelabs Medical, Model 90341-05 Digital Telemetry Transmitter, FCC ID: CM676A90341-WMTS. It is a Licensed Non-Broadcast Transmitter Worn on Body, biomedical telemetry transmitter operating in the 608-614 MHz band.

The technical report and exhibits included with this application demonstrate compliance with FCC rules 47 CFR 95.1115.

Recommendation

All items have been resolved and completed to my satisfaction; therefore I recommend this application for approval.



Dean Ghizzone, President

August 11, 2003

Date

Findings and Resolution

Item	The client has not submitted a signed contract.
Resolution	Resolved

Opinion

Specification Requirements	Description
95.1125, 1.1307(b) & 2.1093	RF Exposure

Opinion: The Equipment meets the intent specified by the requirements listed above.

Discussion: Compliance with 47 CFR 95.1125

The 608 -614 MHz band in Part 95 was re-farmed from Part 15.242 without any change to the field strength requirements of the fundamental. Since 15.242 did not require SAR data, neither should SAR data be required of this radio. In addition, the FCC has not previously imposed any SAR requirements on the applicant for this type of 608 - 614 MHz WMTS radio.

The applicant's radio, is compliant with the requirements of 95.1125.

Reference: RF Exposure Exhibit

Specification Requirements	Description
2.1049	Bandwidth

Opinion: The Equipment meets the intent specified by the requirements listed above.

Discussion: The Applicant has submitted test results in the form of a test report.

Reference: Technical Report

Specification Requirements	Description
95.1115(a)	Field Strength of the Fundamental

Opinion: The Equipment meets the intent specified by the requirements listed above.

Discussion: The Applicant has submitted test results in the form of a test report.

Reference: Technical Report

Specification Requirements	Description
95.1115(b)	Undesired Emissions

Opinion: The Equipment meets the intent specified by the requirements listed above.

Discussion: The Applicant has submitted test results in the form of a test report.

Reference: Technical Report

Specification Requirements	Description
95.1115(e)	Frequency Stability

Opinion: The Equipment meets the intent specified by the requirements listed above.

Discussion: The Applicant has submitted test results in the form of a test report.

Reference: Technical Report

Specification Requirements	Description
95.1115(c)	Emission Types

Opinion: The Equipment meets the intent specified by the requirements listed above.

Discussion: The Equipment is a medical telemetry device not used for video or voice.

Reference: Operation Description

Specification Requirements	Description
95.1123	Protection of Medical Equipment

Opinion: The Equipment meets the intent specified by the requirements listed above.

Discussion: The Applicant addresses this requirement in a warning section in the manual page 17-6.

Reference: User's Manual