

RF Exposure statement provided by client:

This section is not applicable.

The FCC, with full knowledge that patients within a healthcare facility wear these devices continuously in close proximity to the body, determined field strength limitations on transmitters operating in the Wireless Medical Telemetry System (WMTS).

Based upon information provided by the FCC, using the rule-of-thumb that SAR~P, for all P transferred into 1g then 1.6 mW/g requires 1.6mW, which is above the output of the device described in this certification application.