

Date: 11/21/2023

Attn: FCC Office of Engineering and Technology / UL Verification Services TCB

Ref: Class 2 Permissive Change for FCC ID: CZO23LPCR2

Original approval date: June, 7, 2019

Additional C2PC dates: None

Applicant: Physio-Control, Inc. (now a part of Stryker)

To Whom It May Concern

This is to request for a Class II Permissive Change for FCC Certified Transmitter Module FCC ID: CZO23LPCR2 to add the following Host model:

- LIFEPAK CR2

Physio-Control, Inc., now a part of Stryker, is herein submitting a Class II permissive change filing for the Cellular Module as listed in supporting exhibits. Pursuant to FCC Part 22H, 24E, 27, and 90, Physio-Control Inc. is submitting a request to change PCS Licensed Transmitter Grant. Installers and end users must be provided with antenna installation instructions and transmitter operating conditions and instructions for satisfying RF exposure compliance. The final product operating with this transmitter must include operating instructions and antenna installation instructions, for end-users and installers to satisfy RF exposure compliance requirements.

The new application for LPCR2 defibrillator as a portable host device uses RF exposure data to demonstrate compliance for the host/ module combination. These data obtained by our evaluation have been provided to support this application and the 30 mm separation distance. The antenna is the same type but lower gain than the antenna certified for use with the module. No RF parameter or transmitter power changes by hardware or firmware have been made from the original filing.

This Class II permissive change seeks to enable portable use (<20cm distance from persons) in the LPCR2 host device. The LPCR2 also contains an FCC Part 15C radio described on Physio-Control grant under FCC ID: CZO23LPCR2T. Physio-Control, Inc. here seeks to provide evidence pertinent to the listed concerns to obtain the new grant due to any changes in the hardware such as installation.

If you have any questions regarding this application, please feel free to contact me.

Sincerely yours,



Timothy Rand

Sr Compliance Standards Engineer

+1 425 867 4477

tim.rand@stryker.com

Emergency Care

11811 Willows Road NE, Redmond, WA 98052 USA | P +1 425 867 4000 | Toll-free +1 800 442 1142 | stryker.com