



*July 13, 2020*

Federal Communications Commission  
Equipment Authorization Branch  
7435 Oakland Mills Road  
Columbia, MD 21046

**FCC ID: DGFPSDTR602CX1**

To whom it may concern,

Enclosed please find FCC ID: DGFPSDTR602CX1 certification application for Class II Permissive Change. Class II PC is necessary to include improved 2.4GHz antenna output filtering with passive components (swapped the positions of L2 and L3 inductors which has minor electric characterization change and added one capacitor to minimize noise in the antenna)

The equipment covered by this submission is identical in design, power settings and construction to the originally certified 3M Versaflo™ Powered Air Purifying Respirator.

If you have any questions or comments, please do not hesitate to contact me.

Sincerely,

A handwritten signature in black ink that reads "Yuriy Litvinov".

Yuriy Litvinov  
Lead EMC Engineer-Regulatory Engineering  
3M EMC Laboratory, 76-1-01  
St. Paul, MN 55107-1208  
Office: 651 778 6336  
ylitvinov@mmm.com