

EXHIBIT P: Frequency Stability Attestation

FCC ID DV8LX5160



February 26, 2001

To whom it may concern,

Re: Temperature - Frequency stability Certification

This letter is to certify that the Temperature – Frequency stability of the LX-5160 telemetry transmitter which is satisfied it's specification over the operating temperature range.

A handwritten signature in black ink that reads "Yoshi Muraki".

**Yoshi Muraki
General Manager, Patient monitor R&D div.,
Fukuda Denshi Co., Ltd.**

Test report of Frequency Stability

EUT : LX-5160, Serial No. =00000001

Freq. Range : 608 to 614 MHz

Freq. Stability : +/-2.5 ppm

Operating temp. : +10 to +40 degrees C

Date : Feb. 26. 2001

Manufacture : Fukuda Denshi Co., Ltd.

Test engineer : Kenichi Kamisaka

Test site : Fukuda Denshi, Hongo office, ME R&D Dept.

Temp. Chamber : LHU-112, TABAI ESPEC CORP.
Serial No.1571794
Registration No. 224-046-001

Freq. Counter : TR5823H, ADVANTEST
Serial No.00570534
Registration No.202-027-001

Measurement result:

<u>Temp. degrees C</u>	<u>Freq. MHz</u>	<u>Difference, Hz</u>	<u>Difference, ppm</u>
0	611.24979	-207	-0.339
+10	611.24987	-125	-0.201
+20	611.250107	+107	+0.175
+30	611.250067	+67	+0.110
+40	611.249831	-169	-0.276
+50	611.249676	-324	-0.530