

EXHIBIT Q: Cover Letter from Agent

FCC ID DV8LX5160

Northwest EMC, INC.

22975 NW Evergreen Parkway, Suite 400
Hillsboro, OR 97124

March, 26 2001

Dear Application Examiner:

On behalf of Fukuda Denshi Co., Ltd., Northwest EMC, Inc. is submitting this application for the certification of the LX-5160 ECG and Respirator Transmitter, FCC ID: DV8LX5160. It is a portable biomedical telemetry transmitter operating in the 608- 614 MHz band. The patient cables act as an antenna for the device. The maximum operating output power is 1 milliwatt. Testing was performed with the transmitter set to maximum power.

The LX-5160 is categorically excluded from routine evaluation for RF exposure due to its use, transmit frequency, and output power.

The technical report and exhibits included with this application demonstrate compliance with FCC rules 47 CFR 95.1115.

An index of exhibits has been provided to facilitate the review process (see file "Index of Exhibits.pdf"). Please note that the field strength data of the fundamental emission is included in the spurious radiated emissions data (see file "Radiated Emissions.pdf")

Your efforts in reviewing this application are greatly appreciated.

Best regards,



Dave Tolman
Northwest EMC, Inc.