

R&D Department2
35-8, Hongo 2-chome, Bunkyo-ku, Tokyo 113-8420, Japan
Phone+81-3-5684-1300 Fax+81-3-5684-1449

Date : 2010/11/16
Company : Fukuda Denshi Co LTD
Address : 2-35-8 Hongo, Bunkyo-Ku, Tokyo 113-8420

Justification for RF exposure compliance

To whom it may concern:

In accordance with FCC section 2.1093, when the portable device is operating under WMTS/Part 95 Subpart H technical requirements, Routine RF evaluation is required to demonstrate RF exposure compliance.

The FCC, with full knowledge that patients within a healthcare facility wear these devices continuously in close proximity to the body, determined field strength limitations on transmitters operating in the Wireless Medical Telemetry System (WMTS).

Based upon information provided by the FCC, using the rule-of thumb that SAR-P, for all P transferred into 1 g then 1.6 mW/g requires 1.6mW, which is above the output power of the device described in this certification application.

LX-7120: RF output power is specified 1.6.mW or less and measured value 0.7mW, therefore maximum RF output power is less than 1.6mW. If all of the RF output power concentrates any 1g of tissue, literally the SAR does not exceed 1.6mW/g.

As the conclusion, we justify that FCC section 2.1093 is not applicable to the LX-7120.

Sincerely yours,



Yoshi Muraki
Fukuda Denshi Co Ltd